of solutes that, passing in succession along a rod or wire, are finally pyrolysed or volatilized for measurement. In connexion with column chromatography of liquids a variety of equipment for on-line quantitative determination is already available—at a price. Refractometric devices are deservedly, if somewhat belatedly, becoming popular, but will not supersede u.v. devices. Although the outlay (of the order of £500) for a u.v. device will not guarantee that meaningful values will be obtained for separated constituents, such a device at least has a valuable function as a monitor for detecting peaks, which may subsequently be examined off-line.

Kaiser, R. (1969) *Chem. Brit.* **5**, 54–61

Gas-Liquid Chromatography in Biochemical Analysis: a Review of Recent Trends in Equipment

By GEORGES GUIOCHON (*Laboratoire de Chimie Analytique Physique, École Polytechnique, 17 Rue Descartes, Paris 5, France*)

Although g.l.c. was invented by two biochemists, most people in this field have, until relatively recently, regarded this technique more as a snare than as a tool. The main reason is that, in spite of many papers by leading experts working with equipment of their own make, in spite of many advertisements for the 'new gas chromatograph X for biochemical analysis', most commercial equipment has been designed, often successfully, for the analysis of petroleum fractions. This situation is slowly changing, however, and analysis of many compounds of biochemical interest may now be carried out successfully on a routine basis.

It must be taken for granted that the conventional systems for pressure, flow-rate and temperature controls, signal recording and data acquisition and handling are working properly, since they are common to all kinds of chromatographs. Problems associated with sampling, column design and detectors will be discussed.

Many compounds of biochemical interest cannot be analysed as such because they are too polar and not volatile enough. The formation of suitable derivatives is often a good solution, and the various reagents and methods that can be used will be reviewed. Another solution, though much less frequently used, is pyrolysis g.l.c. Some examples of this will be given. The carbon-skeleton determinator will also be exemplified.

Column injection is not too difficult with conventional packed columns. It is much more of a problem with capillary columns, and various ways of solving this will be discussed, with emphasis on the direct injection of very diluted solutions.

Column packings have changed little during the last 10 years, although a few stationary phases that can be used in the 300–400°C range have been developed. A truly inert support is still to be obtained, but the art of chemical modification of solid surfaces by grafting is now fairly well advanced and can offer satisfactory solutions.

New detectors have been developed that have very high sensitivity and selectivity. The flame-photometric detector can determine separately compounds containing nitrogen, sulphur and phosphorus. The electron-capture detector is highly sensitive to halogens, though it is somewhat less predictable. The mass spectrometer is a very sensitive detector, and simultaneously offers a solution to a problem that is usually not solvable by g.l.c. alone, namely the identification of a truly unknown compound.

Finally, some recent examples of difficult analyses will be discussed, as well as the scope of the method. A few problems will be discussed in detail, this seeming to be more useful than a superficial review of the very abundant scientific and technical literature.

Equipment for High-Pressure Liquid Chromatography

By JOHN H. KNOX (*Department of Chemistry, University of Edinburgh, Edinburgh EH9 3JJ, U.K.*)

High-pressure liquid chromatography has developed within the last 5 years and has brought to liquid chromatography the speed, resolution, sensitivity and ease of operation previously associated with g.l.c. The technique promises to be of immense importance for clinical, forensic and biochemical analysis.

The essential features of a modern liquid chromatograph are: (a) a pressurization or pump system capable of delivering degassed dust-free solvent against back pressures up to 21 MN/m² (3000 lbf/in²) and at flow rates up to 10 cm³/min; (b) an injection system for placing samples of volume a few microlitres on the column head; (c) a column, generally 25–100 cm long × 2–4 mm diam., that is packed with 20–50 μm-diam. partitioning material and preferably thermostatically controlled; (d) a highly sensitive detector, of not more than 10 μl volume attached directly to the column outlet, that produces an electrical signal that can be processed by a recorder or other data-handling system.

Commercial equipment costs between £3000 and £10 000, depending on the 'sophistication' and versatility desired. Home-made equipment is cheaper but has its drawbacks. The expensive components in home-made equipment are the pressurization system or pump (£600–1000) and the detector (£600–1200). Couplings should be of the compression type, and all parts coming in contact with the eluent must be of

stainless-steel, Teflon or glass. Columns may be either of heavy walled glass [up to 7 MN/m^2 (1000 lbf/in^2)] or of stainless-steel. Injection is generally by $10\mu\text{l}$ micro-syringe for pressures up to 10 MN/m^2 (1500 lbf/in^2), and sample sizes range from 10 ng to $100\mu\text{g}$. Pressure measurement is made by a Bourdon gauge or a pressure transducer. Thermostatic control of the column can be by a water jacket. Electrical thermostatic control can be dangerous if leakage of flammable solvents occurs.

The pump, column packing and detector are the key parts of any high-pressure liquid-chromatographic system.

(1) Pumps. The use of high-pressure N_2 from a cylinder in direct contact with the solvent is the simplest way of generating a pressure up to 21 MN/m^2 (3000 lbf/in^2). The solvent is held in a long narrow tube. The system is cheap but potentially dangerous. Constant pressure may also be generated by a safer but more expensive pressure-intensifier in which a low gas pressure bears on a large-area piston rigidly connected to a small-area piston bearing on the solvent. These systems deliver liquid at a controlled pressure to the column, they are readily adjusted and they deliver $50\text{--}100\text{ cm}^3$ per stroke. Reciprocating plunger pumps with a variable stroke give a controlled flow rate, but the flow pulsates and a damper is required between the pump and column. Piston pumps have low swept volumes, and are particularly useful for gradient elution since the gradient-producing system can be operated at low pressure; with constant-pressure systems the gradient has to be produced on the high-pressure side, and this is expensive. Motor-driven syringes combine constant flow with continuous delivery: the best versions employ a pair of reciprocating small-volume syringe pumps.

(2) Column packings for high-pressure liquid chromatography must be exceptionally well graded and should have particle diameters between 20 and $50\mu\text{m}$. Many are forms of silica gel, but the most satisfactory materials have a thin absorptive layer deposited on an impervious core. Great care has to be taken with column packing to achieve good trans-column uniformity.

(3) Detection methods used are of three types: those that monitor the change in bulk property of the eluent (refractive index for instance); those that monitor a selective property of the eluent (u.v.- or visible-light absorption, fluorescence, electro-reducibility); and those that physically separate solute from solvent and subsequently determining the solute by a g.l.c. method (for example the Pye wire-transport detector). Bulk detectors generally have a sensitivity limit of between 1 in 10^6 and 1 in 10^7 of solute in mobile phase, whereas the selective detectors can detect as little as 1 in 10^9 . The separation-type detectors are potentially very sensitive, but mechanical and other problems have prevented the

attainment of sensitivity limits lower than about 1 in 10^6 .

Aids to Column Chromatography

By ALAN J. THOMAS (*Department of Biochemistry, Imperial College, London SW7 2AY, U.K.*)

Fast liquid-phase column chromatography is potentially a powerful tool for the sensitive measurement of many important biochemicals. Its greatest use may well lie in the separation and determination of the more hydrophilic compounds, which cannot be easily and quantitatively converted into stable volatile derivatives and thus determined by g.l.c. Good examples are amino acids, sugars, phosphate esters and nucleotides.

For optimum separation of a relatively large number of components at high resolution and in a short time, micro-spherical column packings are required, with narrow-bore columns and, inevitably, considerable pressures. Medium pressures, up to 7 MN/m^2 (1000 lbf/in^2), are relatively easy to handle, and the Milton Roy Minipump, at a cost of under £200, is extremely suitable. Medium pressures, combined with gradient elution and low flow rates, can be extremely effective.

Since the introduction of the multichamber Varigrad by Piez & Morris (1960) various devices have become available for producing both multi-exponential gradients and linear gradients. Although these devices are useful and can be comparatively inexpensive, they are difficult to use for repetitive work and are limited in the forms of gradient that can be generated. A completely flexible electronic gradient maker that could be programmed in an analogue form was described by Thomas (1970), and similar devices are now commercially available. Although these instruments are relatively expensive (about £1000), they permit a total control of the elution process with a very high reproducibility and allow the optimum separation of a large number of components in the minimum time. Their high cost can be frequently justified by the speeding up of the chromatographic process and therefore the amount of effective analysis that can be achieved by a single column, high-pressure pump and detection system.

Many of the compounds best separated in the liquid phase require elution with aqueous gradients most frequently composed of low-volatility organic buffers. In such circumstances physical detection systems can present problems. For example, detection methods that use refraction, conductivity or u.v. absorption will show large base-line shifts due to change of eluent composition upon which are superimposed very small changes due to the separated components. Attempts to compensate for these base-line shifts by