parts of brain (cortex, cerebellum and medulla) fell significantly after 6–24 weeks of pyridoxine deficiency. This observation, unfortunately, was based on only one rat in each group at different intervals and needs confirmation. Their results were not in agreement with our present findings showing that the mean values of total brain GSH of pyridoxine-deficient rats were almost identical with those of pyridoxine-treated rats (see Table 4). In addition, the rats receiving a cysteine-free diet had a high concentration of liver GSH when pyridoxine became a limiting factor.

SUMMARY

1. The effect of pyridoxine deficiency induced under various dietary regimens, on the concentration of non-protein SH compounds (essentially GSH) in erythrocytes and tissues was studied in rats.

2. Deficiency of pyridoxine resulted in an elevation of GSH concentrations in erythrocytes and liver, but not in kidneys, spleen and brain.

3. An increase of GSH concentration in liver was also observed when pyridoxine-deficient rats received a diet free of cysteine or glutamic acid.

4. The incorporation of [2-¹⁴C]glycine and [1-¹⁴C]glutamic acid into the GSH moiety was enhanced by pyridoxine deficiency.

5. It is suggested that pyridoxine is an important regulator in glutathione metabolism.

The authors acknowledge with thanks a grant-in-aid (no. A-2505) from U.S. Public Health Service. They also thank the Dow Chemical Co. for providing amino acids for this study.

REFERENCES

- Ashwood-Smith, M. J. & Smith, A. D. (1959). *Nature, Lond.*, **184**, 2028.
 Beaton, J. R. (1955). *Canad. J. Biochem. Physiol.* **33**, 161.
 Binkley, F., Christensen, G. M. & Jensen, W. N. (1952). *J. biol. Chem.* **194**, 109.
 Blaschko, H., Datta, S. P. & Harris, H. (1953). *Brit. J. Nutr.* **7**, 364.
 Goldzieher, J. W., Besch, P. K. & Velez, M. E. (1958). *J. biol. Chem.* **231**, 445.
 Grunert, R. R. & Phillips, P. H. (1951). *Arch. Biochem. Biophys.* **30**, 217.
 Hanes, C. S., Hird, F. R. & Isherwood, F. A. (1950). *Nature, Lond.*, **166**, 288.
 Hope, D. B. (1958). *J. Physiol.* **141**, 31.
 Hsu, J. M. & Chow, B. F. (1957). *Arch. Biochem. Biophys.* **72**, 322.
 Hsu, J. M. & Kawin, B. (1962). *Proc. Soc. exp. Biol., N.Y.*, **111**, 679.
 Patterson, J. W. & Lazarow, A. (1955). *Meth. biochem. Anal.* **2**, 274.
 Pike, R. L. & Brown, M. L. (1959). *J. Nutr.* **68**, 531.
 Reitman, S. & Frankel, S. (1957). *Amer. J. clin. Path.* **28**, 56.
 Stoerk, H. C., Eisen, H. N. & John, H. M. (1947). *J. exp. Med.* **85**, 365.
 Williams, M. A. & Hata, B. (1959). *Arch. Biochem. Biophys.* **80**, 367.

Biochem. J. (1964) **90**, 64

A Note on the Construction of a Modified Tissue Homogenizer

BY G. R. WEBSTER AND A. T. SMITH

Department of Chemical Pathology and the Medical School Workshop, Guy's Hospital Medical School, London, S.E. 1

(Received 15 July 1963)

Aldridge, Emery & Street (1960) described a tissue homogenizer, based on that of Potter & Elvehjem (1936), which incorporated modifications specifically designed so that homogenization was both efficient and reproducible and also so that if the homogenizer was broken it could be replaced by one with an identical performance.

The present note concerns further modification in the design of this instrument, which, while retaining all the functional advantages of that described by Aldridge *et al.* (1960), greatly simplifies the construction and, in our experience, has several other advantages, which are discussed below.

The instrument described by Aldridge *et al.* (1960) consists of a rotating Perspex pestle with a cylindrical body and a cone-shaped end (90° angle). The body of the pestle is machined so as to fit into a precision-bore glass tube leaving a precisely defined clearance between the two surfaces; the cone-shaped end fits accurately into a corresponding conical depression in a Perspex plug which closes the end of the tube and forms the base of the homogenizer. The Perspex base-plug is machined so as to be a sliding fit into the tube and is cemented in place with Bostik D.

In our experience it is not easy to ensure that the annular space between the Perspex base-plug and

the glass tube is completely filled with cement; if it is not, traces of homogenized tissue may be retained in it so that cleaning is difficult. Further, we have frequently found, especially if the plug is made to be a good fit into the glass tube, that breakage of the tube occurs round the base. This tendency presumably arises because the coefficient of expansion of Perspex is about 25-fold that of borosilicate glass (Kaye & Laby, 1956), so that even a relatively slight rise in ambient temperature may result in considerable stresses being exerted on the glass tube. The design of the Perspex base makes repairs of such damage, as well as the initial construction, both costly and time-consuming.

In a modified design (Fig. 1), which has been found to overcome these disadvantages, the base-plug is made from stainless steel which has a coefficient of expansion only about threefold that of borosilicate glass; the plug can therefore be made to be a relatively tight push fit into a length of Veridia tubing (Chance Brothers Ltd.) with 30 mm. bore and 2 mm. wall thickness. The outer end of this steel plug is drilled and tapped to take two countersunk screws (2 B.A. $\times \frac{3}{8}$ in.), by means of which it is attached to a Perspex disk (3 in. diam. and $\frac{1}{2}$ in. thick). The plug can then be pushed

into the glass tube until the end of this abuts on to the Perspex disk. The homogenizer is used in conjunction with the Perspex pestle described by Aldridge *et al.* (1960). The base and glass tube can be easily separated for cleaning.

In use, the glass tube is grasped in the left hand while the Perspex disk of the base is allowed to lie on the flattened palm of the right hand; the pestle is moved up and down in the tube by alternately pushing up with the right hand and then downwards with the left. In this way, although the plug is not cemented into the tube, there is no tendency for the two components to part company during homogenization, and little or no leakage of liquid occurs at the junction. When, however, the homogenizer is cooled by standing it in a beaker of ice, slight leakage of contents may occur and, further, care is needed in removing the homogenizer from the beaker to avoid separation of the base and glass tube. To guard against these possibilities R. C. Emery & B. W. Street (personal communication) have incorporated a rubber 'O' ring, carried in a groove in the Perspex disk round the steel base-plug, to provide a seating for the end of the tube; they have also fitted a Perspex clamp, grooved to contain a rubber 'O' ring, which grips the tube and is attached to the Perspex disk by tension springs.

It is advisable to make the steel base-plug to fit a length of Veridia tubing with a diameter near the outer limit of the tolerance specified by the manufacturers (30 ± 0.01 mm.); other lengths within these tolerance limits can then easily be ground out to fit the same plug. This is done with a piece of nylon rod which fits the tube leaving 0.005 in. radial clearance, and which is centred in a lathe chuck and coated with a paste of oil and fine carborundum. Replacements in the event of breakage are thus very simple.

SUMMARY

1. The construction of a tissue homogenizer with improved facilities for cleaning and repair is described.

Our thanks are due to Dr W. N. Aldridge for his helpful comments and to Mr R. C. Emery and Mr B. W. Street for permission to refer to their further modifications in design.

REFERENCES

- Aldridge, W. N., Emery, R. C. & Street, B. W. (1960). *Biochem. J.* **77**, 326.
- Kaye, G. W. C. & Laby, T. H. (1956). *Physical and Chemical Constants*, 11th ed., p. 54. London: Longman, Green and Co.
- Potter, V. R. & Elvehjem, C. A. (1936). *J. biol. Chem.* **114**, 495.

Fig. 1. Diagram of modified tissue homogenizer. A, Stainless-steel base plug; B, countersunk screws (2 B.A. $\times \frac{3}{8}$ in.); C, Perspex disk (3 in. diam. $\times \frac{1}{2}$ in. thickness); D, Veridia glass tubing (30 mm. bore, 2 mm. wall thickness).