*RACK FOR CUMULATIVE-RESPONSE RECORDERS*CHARLES A. GILL, WILLIAM FRY, AND ROGER T. KELLEHER¹

SMITH KLINE & FRENCH LABORATORIES

The use of numerous cumulative-response recorders necessitates an efficient technique for collecting the paper that is discharged from the recorders. The recorder rack shown in Fig. 1 handles the paper neatly and permits a compact, orderly arrangement of the recorders. Furthermore, it eliminates the need for added space for separate collection baskets. The paper is simply guided by the curved sheet metal piece to pass behind the recorder below it. This piece is attached in a pivotal fashion so that it swings upward when the recorder cover below it is raised for marking or adjustment. Figure 1 shows the piece in both normal and raised (dashed lines) positions.

Detailed plans of the apparatuses described here are available upon request to the authors.

¹Now at Harvard Medical School.

Fig. 1. Scale diagram of cumulative recorder rack.