Published in final edited form as:

Heart Rhythm. 2005 April ; 2(4): 339–354.

Noninvasive Electrocardiographic Imaging (ECGI): Comparison to intraoperative mapping in patients

Raja N. Ghanem, PhD^a, Ping Jia, PhD^a, Charulatha Ramanathan, PhD^a, Kyungmoo Ryu, MS^a, Alan Markowitz, MD^b, and Yoram Rudy, PhD^a

^aFrom the Cardiac Bioelectricity Research and Training Center (CBRTC), Case Western Reserve University, Cleveland, Ohio

^bDivision of Cardiothoracic Surgery, University Hospitals of Cleveland, Cleveland, Ohio.

Abstract

OBJECTIVES/BACKGROUND—Cardiac arrhythmias are a leading cause of death and disability. Electrocardiographic imaging (ECGI) is a noninvasive imaging modality that reconstructs potentials, electrograms, and isochrones on the epicardial surface from body surface measurements. We previously demonstrated in animal experiments through comparison with simultaneously measured epicardial data the high accuracy of ECGI in imaging cardiac electrical events. Here, images obtained by noninvasive ECGI are compared to invasive direct epicardial mapping in open heart surgery patients.

METHODS—Three patients were studied during sinus rhythm and right ventricular endocardial and epicardial pacing (total of five datasets). Body surface potentials were acquired preoperatively or postoperatively using a 224-electrode vest. Heart-torso geometry was determined preoperatively using computed tomography. Intraoperative mapping was performed with two 100-electrode epicardial patches.

RESULTS—Noninvasive potential maps captured epicardial breakthrough sites and reflected general activation and repolarization patterns, localized pacing sites to ~1 cm and distinguished between epicardial and endocardial origin of activation. Noninvasively reconstructed electrogram morphologies correlated moderately with their invasive counterparts (cross correlation = 0.72 ± 0.25 [sinus rhythm], 0.67 ± 0.23 [endocardial pacing], 0.71 ± 0.21 [epicardial pacing]). Noninvasive isochrones captured the sites of earliest activation, areas of slow conduction, and the general excitation pattern.

CONCLUSIONS—Despite limitations due to nonsimultaneous acquisition of the surgical and noninvasive data under different conditions, the study demonstrates that ECGI can capture important features of cardiac electrical excitation in humans noninvasively during a single beat. It also shows that general excitation patterns and electrogram morphologies are largely preserved in open chest conditions.

Keywords

Electrophysiology; Electrocardiography; Imaging; Mapping; Arrhythmia

Address reprint requests and correspondence: Dr. Yoram Rudy, Washington University in St. Louis Cardiac Bioelectricity and Arrhythmia Center (CBAC), 290 Whitaker Hall, Campus Box 1097, One Brookings Drive, St. Louis, Missouri 63130-4899. E-mail address: rudy@wustl.edu.

Dr. Rudy's current affiliation is the Cardiac Bioelectricity and Arrhythmia Center, Washington University, St. Louis, Missouri.

Introduction

Cardiac arrhythmias are a leading cause of death and disability, with more than 250,000 cases of sudden death annually in the United States alone. Despite this fact, a noninvasive imaging modality for cardiac electrophysiology and arrhythmias does not yet exist in clinical practice. Cardiac arrhythmias are diagnosed based on ECG measurements made on the body surface, far away from the heart. Because each ECG electrode reflects the electrical activity from the entire heart, geometric relationships and relative locations of active regions in the heart are not preserved on the body surface.¹ Moreover, the effects of the distance and torso volume conductor properties are to smooth the potential distribution on the body surface, resulting in a low-resolution projection of potential patterns and inability to resolve and locate multiple cardiac events.¹ In contrast, epicardial potentials reflect with high resolution the underlying cardiac electrophysiology in normal and pathophysiologic conditions.² Therefore, it is important, for clinical and research applications, to image noninvasively electrophysiologic information on the epicardium.

ECGI is an imaging modality that is under development in our laboratory. ECGI reconstructs noninvasively during a single beat potentials, electrograms, and isochrones on the epicardial surface of the heart. To date, we have validated ECGI extensively in controlled torso-tank and dog experiments in normal³⁻⁵ and abnormal⁶⁻⁸ canine hearts and during ventricular arrhythmias.^{7,9} Results from our torso-tank experiments, which allow validation against a “gold standard” measured simultaneously and under identical conditions, showed that pacing sites can be localized to within 10 mm and that dual pacing sites can be differentiated with 17-mm resolution.⁴ In the same study, correlation coefficients between noninvasively reconstructed and directly measured epicardial electrograms were >0.9 for 72% of all epicardial locations, indicating very good agreement. We recently described the application of ECGI in humans and provided examples of imaging activation and repolarization during normal sinus rhythm (SR), right bundle branch block, ventricular pacing, and atrial flutter.¹⁰ We also showed that pacing sites can be localized to within 7 and 11 mm of their actual positions (determined by computed tomography [CT]) on the right ventricle (RV) and left ventricle (LV), respectively.¹⁰ Using the generalized minimal residual iterative technique, localization of pacing sites improved to within 3 mm.¹¹ Here, we compare epicardial images obtained using noninvasive ECGI during SR and ventricular pacing to invasive epicardial mapping data from the same patient. Note that, unlike the torso-tank experiments where invasive and noninvasive data are obtained simultaneously under identical conditions, here the invasive data are obtained during open heart surgery, whereas the noninvasive data for ECGI reconstructions are obtained preoperatively and postoperatively, under closed chest and different physiologic conditions. Consequently, differences between the two protocols in the shape and geometry of the heart and the shape and magnitude of epicardial electrograms are to be expected. Additionally, the reference potential and epicardial potential magnitudes are affected by opening the chest.¹² Finally, changes are possibly caused by the surgical manipulations and cardiopulmonary bypass. In all cases, differences in geometry (an important element in the inverse computation) cannot be avoided. Because of these differences, this study cannot provide quantitative validation of the ECGI methodology, as done previously in the torso-tank experiments.^{4,5,7} However, based on animal experiments¹² and modeling studies,¹³ one expects general similarity of patterns. Therefore, the aim of this study is to provide only a qualitative comparison between noninvasive ECGI reconstructions and epicardial data measured directly from the surface of the human heart, subject to the limitations stated.

Methods

Patient population

Three patients (1 male and 2 females) undergoing open heart surgery were enrolled in the study. Informed consent was obtained according to Institutional Review Board guidelines at University Hospitals of Cleveland, which approved all protocols. Protocols were designed so that each patient served as his/her own control. Five datasets that included SR and RV pacing were analyzed. Patient OR1 suffered from coronary artery disease (50% left main occlusion, 60% narrowing in ostium of circumflex mid posterolateral branch and long left main, 80% narrowing in the proximal portion of the mid branch and severe within-stent restenosis in the proximal right coronary artery [RCA]) and an inferior myocardial infarction due to severe RCA disease. Patient OR2 had a history of chronic atrial fibrillation and had undergone two surgical procedures (a stent placement for circumflex occlusion and repair for mitral valve regurgitation). The epicardial mapping reported here was conducted during surgery for (1) mitral valve regurgitation, (2) tricuspid valve regurgitation, (3) maze procedure for atrial fibrillation repair, and (4) closure of the patent foramen ovale. Transesophageal echocardiography showed (1) severe right atrial enlargement, (2) normal left atrial size, (3) normal LV size but with abnormal segmental wall motions, and (4) mildly impaired systolic RV function with moderate enlargement of the RV. The patient also had received a dual-chamber pacemaker. Patient OR3 suffered from an ascending aortic aneurysm at the transverse section of the aortic arch and in the proximal descending aorta. The patient also had a history of aorto/bi-iliac replacement and systolic murmur over the aortic valve. In this study, he underwent surgery to repair his aortic arch aneurysm. All patients had an LV ejection fraction of at least 50%. Patient data are summarized in Table 1.

ECGI

ECGI procedure consisted of body surface potential mapping and computed CT to obtain patient specific heart-torso geometry.¹¹ A 224 Ag/AgCl dry electrode vest was used for body surface potential mapping (Figure 1B). With the patient in the supine position, body surface potentials referenced to the left leg were amplified (gain 100, 200, 1000, 2000), bandpass filtered (0.05–500 Hz), digitized (12-bit A/D converter), and stored on a local data acquisition computer for analysis and display. Except for patient OR3, CT axial scanning (Figure 1A) was used to obtain the heart surface geometry and body surface ECG electrode positions. The transfer matrix relating the heart surface to the location of each ECG electrode was formed based on the boundary element method¹¹ (Figure 1C). Body surface potentials and the transfer matrix served as input for our ECGI methodology.¹¹ Zero-order Tikhonov regularization¹¹ and the generalized minimal residual iterative scheme¹¹ were used to compute epicardial potentials (Figure 1D) on many epicardial sites (1466, 716, and 873 nodes for patients OR1, OR2, and OR3, respectively). The regularization parameter (Tikhonov) or number of iterations (generalized minimal residual) were determined using the composite residual and smoothing operator or conditional L-curve, respectively.

Patient study procedure

Patients enrolled in this study underwent the following: (1) ECGI procedure (preoperative body surface potential mapping and CT); (2) intraoperative mapping during surgery; and (3) postoperative body surface potential mapping. Data were acquired during two episodes of SR and RV epicardial pacing and one episode of RV endocardial pacing (total of five different datasets).

Preoperative phase—A few days prior to surgery, patients underwent the ECGI procedure. The vest was secured around the torso. Skin markers were affixed on the patient's torso to mark the layout of the vest, which also was documented using digital photographs (SONY DSC-

S85). Body surface potential mapping was performed during light breathing. Following body surface potential mapping and with the vest on, patients OR1 and OR2 underwent axial CT imaging. Because patient OR3 was clinically indicated for chest CT, the vest was not worn during CT imaging. In this patient, a digitizer (Microscribe G2L, Immersion Corporation, San Jose, CA) with 1-mm resolution was used to digitize the ECG electrode positions on the chest while the patient lay in the supine position during body surface potential mapping. The back electrodes were digitized from the residual impressions found on the patient's back as he sat upright after body surface potential mapping. Nine skin markers were affixed on the patient's chest and back in order to register the digitized vest electrodes with the CT heart geometry.

Intraoperative mapping—Intraoperative mapping was performed using two patches of flexible cloth material onto which 10 strips (five per patch with an interstrip spacing of ~ 1 cm) were sutured (Figure 1E). Each strip consisted of insulated flexible circuits (MicroConnex, Inc., Snoqualmie, WA) with 20 conductive Ag/AgCl (patient OR1) or gold (patients OR2 and OR3) unipolar electrodes arranged in pairs (5-mm spacing) with 1.2-cm spacing between pairs. Upon completion of intraoperative mapping, the strips were detached from the patches that were discarded after each study. The strips were water flushed with soap and cleaned, resutured to clean patches, and resterilized between patient studies. Following general anesthesia, the patient was placed in the supine position, and the chest, groins, and legs were prepped and draped in a sterile fashion. The heart was exposed through a midline sternotomy and appropriate pericardial retraction was applied. The two patches were placed on the ventricles as allowed by the surgical procedure (Figure 1F). Epicardially recorded potentials were amplified (gain 1,000 or 2,000), bandpass filtered (1–500 Hz), sampled (1 kHz and 12-bit A/D), and stored. Recorded potentials were obtained during SR (patients OR1 and OR3), RV endocardial pacing (patient OR2), and RV epicardial pacing (patients OR2 and OR3) (Table 1). For pacing, stainless steel wires were sutured onto the ventricular surface and the electrode patches slipped over them. After mapping, the wires were extracted and the surgical procedure performed. Following the procedure, the pacing wires were resutured in their approximate original position. Intraoperative recordings for patients OR1 and OR3 were obtained prior to bypass. For patient OR2, mapping was performed during normothermic cardiopulmonary bypass. All intraoperative recordings were performed without cardioplegia.

Postoperative mapping—Postoperative body surface potential mapping was conducted on patients OR2 and OR3 (Table 1) during epicardial pacing from temporary pacing wires placed during surgery. Using the skin markers and the digital photographs acquired during preoperative mapping, the vest was placed on the patient. Care was taken to ensure that the vest remained in the same orientation as that during preoperative mapping. This method enabled us to use the patient specific geometry (heart surface and ECG electrodes positions acquired preoperatively) with postoperative body surface potential mapping for ECGI reconstructions. For patient OR2, postoperative body surface potential mapping was performed 7 days after surgery. For patient OR3, postoperative body surface potential mapping was performed under anesthesia following the closure of the chest and recovery from bypass.

Data analysis

Noninvasive epicardial potentials, electrograms, and isochrones were compared to invasive data. Isochrone maps were constructed from activation times (determined from the maximum negative derivative of the electrograms) for both invasive and noninvasive data.

Localization of pacing sites—The earliest maximum (endocardial pacing) or minimum (epicardial pacing) in the epicardial potential maps was used to determine pacing site locations noninvasively.^{5,6} These locations were compared to the locations determined from CT (endocardial pacing) or from intraoperative mapping (epicardial pacing).

Signal processing—Prior to ECG imaging reconstructions, ECGs were inspected for quality. Missing (no contact) or bad ECGs (signals within baseline noise levels) were discarded. ECGs were detrended to correct for baseline drift (best linear fit, piecewise linear fit, or visual fit). In the case of patient OR2, ECGs were filtered due to measurement noise. Intraoperative electrograms were corrected for baseline drift after missing (no contact) or bad electrograms (signals within baseline noise levels) were discarded.

Time alignment—To compare noninvasive and invasive epicardial potential maps, time alignment of the two cardiac sequences was necessary, especially because invasive and noninvasive data were not obtained simultaneously. For SR data, the root mean square signals (square root of the sum of squares) of the ECGs and intraoperative electrograms were computed. These signals were time aligned along the maximum upslope of the QRS. For pacing, ECGs and intraoperative electrograms were aligned from the onset of the pacing artifact.

Three-dimensional spatial alignment—To visualize the invasive and noninvasive potential maps, the intraoperative patches were aligned spatially relative to the three-dimensional (3D) heart surface geometry obtained from CT. Digital photographs were taken during surgery to document the orientation of the anterior patch. The visible anatomic landmarks (margin of the RV, left anterior descending artery, RV AV groove) were marked on the patch with a surgical marker. The orientation of the posterior patch was estimated and documented relative to known anatomic landmarks (coronary sinus and LV AV groove). To minimize the 3D alignment error between invasive and noninvasive datasets during SR, locations of noninvasively reconstructed electrograms at the patch periphery were aligned with those measured. For pacing, epicardial potential patterns during the pacing artifact were matched. This spatial alignment was maintained unmodified for comparison of potentials, electrograms, and isochrones on the entire patches throughout the cardiac cycle for all protocols.

Comparison of electrograms

Following 3D alignment, each electrogram on the intraoperative patch (measured node) was paired with a noninvasively reconstructed one on the 3D heart (reconstructed node) based on the shortest euclidean distance between the two nodes. For each pair of electrograms, cross correlation and time lag between the noninvasive and intraoperative electrograms were computed as measures of similarity. Correlation coefficient measured similarity of morphology; time lag was the time shift at which correlation coefficient was maximum. Time lag was used to minimize time misalignment between the intraoperative and noninvasively reconstructed electrograms that were not acquired simultaneously and were recorded under different conditions. The time lag shift permitted comparison of electrogram morphologies independent of time misalignment between the two datasets.

Results

Epicardial potentials during SR

Figure 2 compares noninvasive and invasive epicardial potentials during SR for patient OR1. Invasive maps (left column) show positive potentials (red) over the RV 10 ms following onset of QRS (top row). At 21 ms (second row), the wavefront breaks through the epicardial surface resulting in negative potentials (blue). At 30 ms (bottom row), the epicardial wavefront expands, and a larger region of negativity (blue) is mapped. Noninvasive epicardial potentials (middle column) capture this progression. At 21 ms, two minima are reconstructed (second row, second column, sites 1 and 2) indicating two breakthrough sites. One of these minima corresponds to the measured minimum; the other is outside the area covered by the recorded

patch and cannot be validated by direct comparison. At that time, the body surface potential map (right column) is smooth and shows a single broad negative. The low spatial resolution is a general property of body surface potential mapping and results from the smoothing effect of the torso volume conductor.^{1,4,14,15} Similar ability to reconstruct two epicardial potential minima from a single broad body surface potential mapping minimum with ECGI was demonstrated in animal⁴ and human studies.¹⁰ Here, the two distinct minima were reconstructed by two independent algorithms (Tikhonov and generalized minimal residual) at the exact same epicardial locations. In this patient, only 60% of the anterior patch covered the RV; the remainder was over the RA and the right margin of the RV. This finding may result from the geometry of the heart during open heart surgery and the physical environment that did not enable the surgeon to cover the entire RV. This patient did not suffer from hypertrophy or enlarged ventricles (Table 1). The patient had an inferior infarct that resulted in negative potentials and a qR electrogram morphology indicating late activation over the inferior LV, as shown in Figure 3 (electrograms 57 and 54).

Electrograms during SR

Noninvasive (black) and invasive (red) electrograms during activation (QRS complex) are compared for patient OR1 (Figure 3). Maximum correlation coefficient and time lag are provided. Invasive and noninvasively reconstructed electrograms depict rS morphology over the RV breakthrough region (panel A, sites 27, 31, 41) and RV base (sites 8, 12). Over the lateral RV (sites 2, 4, 19, 20), both show rS morphology with low-level amplitude deflections superimposed over a wide S wave.

LV electrograms lateral to the interventricular septum (panel B, sites 16, 33, 31, 45) depict RS morphology. Invasive electrograms from lateral LV and base show qR (sites 38, 54) or qRs (sites 26, 41, 57) morphologies. Corresponding noninvasively reconstructed electrograms reproduce these morphologies, albeit with depressed amplitude due to ECGI smoothing. Failure to reconstruct s waves in the noninvasively reconstructed electrograms (sites 26, 41, 57) may be attributed to errors in ECGI. However, the remaining electrograms capture the correct morphologies, suggesting that these differences result from different conditions (see Study limitations) during invasive and noninvasive mapping. On LV apex, the electrogram (site 64) shows Rs morphology. Correlation coefficient and time lag are given in Table 2.

Epicardial potentials and electrograms during RV endocardial pacing

In Figure 4A, invasive (first column) and noninvasive (second column) epicardial potentials are compared for patient OR2 on anterior RV 3, 40 and 60 ms following onset of endocardial RV pacing (0.5-ms pulse width). At 3 ms (top row), positive epicardial potentials (red) are measured over the apex above the pacing site (asterisk). This maximum and two flanking minima (blue) are reconstructed noninvasively by ECGI (second column); only one minimum is captured in the invasive map due to limited size of the epicardial patch. At 40 ms, epicardial breakthrough (local negative, blue) occurs over the RV apex. At 60 ms, the wavefront develops further (larger negative, blue) activating the RV. ECGI reconstructs this temporal progression (second column). As activation spreads to LV, a single apical minimum (blue) and a single basal maximum (red) are measured at 110, 130, and 150 ms (last column). This spatial pattern and temporal progression of potentials is captured with ECGI (third column). Based on the earliest epicardial maximum,⁵ the endocardial pacing site is estimated noninvasively within 4.3 mm from its position as determined from CT (Table 2). It should be recognized that epicardial potential magnitudes differ significantly between open chest and closed chest conditions^{12,13} and that ECGI methodology acts to suppress these magnitudes.^{11,14} Therefore, low-amplitude details (Figure 4A, positive potential islands [orange]) are lost in the ECGI reconstruction. However, the global pattern of positive and negative regions is preserved.

During repolarization (panel B, T peak), positive potentials (red) over the apex and RV margin and negative potentials (blue) over the LV base are reconstructed using ECGI (second and third columns) and confirmed invasively (first and last columns).

In Figure 4C, six noninvasive (black) and invasive (red) electrograms are compared. Invasive electrograms (first column) near the pacing site (sites 8, 40, 57) show typical Q waves followed by positive T waves. Electrograms remote to the pacing site show RS (site 32) or Rs (sites 12, 43) morphologies. Noninvasively reconstructed electrograms (second column) reproduce these different morphologies. Correlation coefficient and time lag are given in Table 2. Note that mean correlation coefficient on anterior RV is relatively low (0.55), possibly due to intraoperative mapping during cardiopulmonary bypass in this patient (dictated by the surgical procedure) that increased differences in conditions (possibly due to reduced ventricular filling) under which invasive and noninvasively reconstructed electrograms were acquired.

Epicardial potentials during RV epicardial pacing

In Figure 5A, noninvasive (second and third columns) and invasive (first and last column) epicardial potentials are compared 80 ms (top row) following onset of RV epicardial pacing (2 ms, 10-V stimulus) and at T peak (bottom row) for patient OR3. At 80 ms, negative potentials (blue) with an intense minimum (minus sign) are measured at the pacing site (note that the same negative region appears in both right and anterior views due to overlap). On lateral RV, invasive negative potentials extend laterally and basally in an elliptical fashion. The local minimum (–) at the pacing site and potential gradients on anterior and lateral RV (negative at base to positive at apex) are reconstructed using ECGI. Based on the earliest intense minimum, the pacing site was localized noninvasively to within 15.8 mm (Table 2). During repolarization, the potential pattern is reversed with positive potentials (red) reconstructed at the pacing site and negative potentials (blue) at the base.

Figure 5B shows similar data for patient OR2 70 ms following onset of pacing and at T peak. At 70 ms, noninvasive potentials show elliptical negative region (blue) around the pacing site (–) on anterior RV and negative potentials (green) across the interventricular septum and positive potentials (red) on LV free wall. These potential features (minima and gradients) are confirmed with invasive data. ECGI localized the pacing site to within 18.8 mm (Table 2). During repolarization, positive potentials (red) over the pacing site, a single maximum (red) across the interventricular septum, and a single minimum (blue) over the LV free wall are reconstructed. Invasive potentials confirm these patterns.

Electrograms during RV epicardial pacing

Figure 6 shows noninvasive (black) and invasive (red) electrograms during RV epicardial pacing for patient OR3. During activation, invasive electrograms near the pacing site (asterisk) depict negative Q-wave morphology (sites 8, 20, 22, 33, 36, 38, 54, 61); electrograms remote to pacing show RS morphology (sites 32, 46). During repolarization, electrograms near the pacing site reveal positive T waves (sites 8, 20, 22, 33, 36, 38), whereas those remote to pacing show negative (site 46) or positive (site 32) T waves. Noninvasively reconstructed electrograms reproduce these QRS and T morphologies with high correlation coefficient. Invasive electrograms with small deflections superimposed on the QRS (sites 21, 23) also are reconstructed using ECGI. Small deflections over the QRS at sites 38 and 57 are within the noise level measured during the ST segment and, therefore, are not reconstructed by ECGI. Correlation coefficient and time lag are given in Table 2.

Epicardial isochrones

Figure 7 compares noninvasive and invasive isochrone maps during SR, RV endocardial pacing, and RV epicardial pacing. Activation times are referenced to onset of QRS after time

alignment (see Methods, Time Alignment). During SR (top row), noninvasive isochrones show that epicardial excitation originates from two breakthrough sites on anterior RV (second column, marked 1 and 2); only one site is contained within the intraoperative patch (first column, 1). On posterior LV (third column), activation spreads laterally toward the posterolateral LV. During RV endocardial pacing (second row), propagation spreads from the earliest breakthrough (RV apex, asterisk) toward the RV outflow tract in both invasive and noninvasive maps. On posterior LV, activation spreads from LV apex towards basal and posterolateral LV in both invasive and noninvasive maps. During epicardial pacing (bottom row), activation originates from the pacing site (asterisk) and spreads radially. Regions of slow conduction (crowded isochrones, arrows) are captured noninvasively.

Discussion

This study compares noninvasively imaged human epicardial potentials, electrograms, and isochrones, reconstructed during a single beat using ECGI, to those obtained from direct intraoperative mapping in the same patient. Unlike our previous torso-tank experiments,^{4-6, 9,16} where body surface and epicardial potentials were acquired simultaneously and under identical conditions, here, as a consequence of the inherent limitations of the human protocol, body surface and epicardial recordings were taken at different times (different heart beats) under different physiologic conditions (Body surface: closed chest with unperturbed heart position, geometry and orientation; absence of anesthesia or other effects of the surgical procedure. Epicardium: open chest, heart geometry perturbed by the surgery, presence of anesthesia and other surgical effects). Because these two datasets cannot be obtained simultaneously under identical conditions in humans, we cannot justify using quantitative assessments of similarity, similar to those we used in the torso-tank experiments. Based on animal¹² and model studies,¹⁶ one expects general patterns to be qualitatively similar for the two conditions. Therefore, we limit evaluation to a qualitative comparison (rather than quantitative validation) of measured and noninvasively reconstructed data (except for electrograms where we attempt to compare morphologies rather than absolute amplitudes). Results are discussed below.

Epicardial potentials during intramural activity

Noninvasively reconstructed epicardial potentials can reflect transmural progression of cardiac activation.^{5,15,17,18} During SR, positive potentials over anterior RV reflect intramural wavefront propagation from endocardium to epicardium.¹⁷ Upon breakthrough, negative epicardial potentials invade positive potentials and expand as the wavefront develops. These patterns were mapped in the chimpanzee¹⁵ and humans¹⁷ and in this study correlate with intraoperative recordings in the same patient.

Immediately following apical endocardial pacing (Figure 4A, 3 ms), noninvasive epicardial potentials show a positive region over the RV apex, as predicted theoretically¹⁸ and observed in direct mapping of canine hearts.⁵ This positive is flanked by two minima oriented along endocardial fibers at the pacing location.¹⁹ This pattern is reversed in polarity from the central minimum flanked by two maxima associated with intramural or epicardial pacing^{4,5} and permits differentiation between endocardial and epicardial origin of ectopic activation, an important distinction when ablation is considered. The polarity reversal results from the effect of the intracavitary blood on current flow during endocardial pacing.⁵ As activation spreads towards the epicardium, positive potentials change to negative potentials upon breakthrough over the apex. For epicardial pacing, quasi-elliptical negative potentials with major axis along epicardial fibers⁴ develop around the pacing site (Figure 5). During repolarization, this potential pattern is reversed in polarity. This reversal is expected theoretically¹⁸ and documented experimentally²⁰; it results from the dominant role of the activation sequence in

determining the repolarization sequence during pacing, when activation spread is slow (as opposed to the dominant role of intrinsic repolarization gradients during SR).²¹ It should be recognized that although in many cases of intramural pacing (or focal activation) the epicardial pattern permits determination of the initiation site (as shown here and in references^{5,10}), this becomes difficult when activation spread is complex because of (for example) fibrotic barriers or penetration of the conduction system by the propagating wavefront.

Localization of pacing sites (ectopic foci)

Noninvasive determination of ectopic arrhythmogenic foci or exit sites of reentrant pathways can uncover arrhythmia mechanisms and guide ablation procedures.²²⁻²⁴ We have shown in torso-tank experiments^{4,5} and human studies (based on CT)¹⁰ that ECGI can localize pacing sites with 10-mm accuracy. Here, the endocardial pacing site is localized within 4.3 mm (relative to CT localization) and the two epicardial sites within 15.8 and 18.8 mm. During surgery, epicardial pacing wires are placed and mapping is performed prior to the corrective procedure. The temporary wires are removed during the procedure and sutured back within 1 cm of their original position. Noninvasive ECGI is performed postoperatively for these new positions. Given the 1-cm uncertainty, the best noninvasive estimates could be within 5.8 and 8.8 mm of the actual pacing sites. At worst, the accuracy is 25.8 and 28.8 mm. If higher resolution is needed locally, ECGI could be used to guide a mapping catheter directly to the region of interest.

Characterization of abnormal substrates

ECGI reconstructs abnormal electrophysiologic properties associated with infarcted myocardium.^{6,7} Negative S-wave electrograms with superimposed low-amplitude deflections reflect slow discontinuous conduction in an infarct with surviving islands of excitable myocardium.²⁵ In patient OR1, wide S waves with superimposed small deflections are imaged noninvasively on lateral RV base (Figure 3A), probably reflecting myocardial damage due to within-stent restenosis in the proximal RCA (detected during catheterization). Dispersion of myocardial repolarization is arrhythmogenic and reflected in epicardial potential maps.⁸ The invasive potential map at T peak (Figure 5B, bottom row, left column) shows multiple islands of negative and positive potentials indicative of repolarization dispersion.⁸ The corresponding noninvasive map (second column) also reconstructs multiple positive and negative regions albeit with a smoother pattern due to ECGI methodology.⁸ This complex pattern is observed on RV, whereas LV potentials (third and fourth columns) contain a single minimum and maximum without steep gradients. Interestingly, in patient OR2, RV function is compromised due to cardiomyopathy, whereas the LV maintains a normal ejection fraction. Finally, regions of slow conduction that are important for initiation and maintenance of arrhythmias²⁶ are captured noninvasively with ECGI (Figure 7, arrows).

Study limitations

We previously validated ECGI in torso-tank experiments^{3-5,7} through direct comparison with epicardial data measured simultaneously and under identical conditions. The present human study suffers from several limitations that are unavoidable in human surgical protocols. Importantly, body surface potential mapping and epicardial potentials cannot be recorded simultaneously and under identical conditions (as could be done in our torso-tank experiments^{3-5,7}). This difficulty limits the accuracy of quantitative evaluation of ECGI using human intraoperative mapping data.

Open chest vs closed chest—Intraoperative mapping and body surface potential mapping were obtained in open chest and closed chest, respectively. Open chest epicardial potentials were augmented by the insulating air¹ and, therefore, were larger than noninvasive potentials

reconstructed in closed chest. As a result, contour plots of intraoperative potentials span larger scales than the noninvasively reconstructed potentials (as shown in Figures 2, 4, and 5). To minimize differences in perception of patterns due to different color schemes, we adjusted the color scales accordingly. Amplitudes were affected, but potential patterns remained similar.

¹² Here, potential maxima, minima and isopotential line densities were preserved in the noninvasive closed chest reconstructions. A major limitation was the inability to obtain invasive and noninvasive data simultaneously, under the same conditions and for the same beat. Time alignment between invasive and noninvasive epicardial potential maps was achieved based on the pacing spike (for pacing data) and the maximum upslope of the QRS (for SR data). Physiologic conditions differed among the preoperative, intraoperative, and postoperative heart. We made an effort to conduct the intraoperative mapping prior to cardiopulmonary bypass to reduce such differences. In one case (OR2), the surgical procedure dictated mapping during bypass, resulting in low correlation coefficient values between noninvasive and invasive electrograms on anterior RV. Images obtained by ECGI were reconstructed over the entire epicardial surface. Intraoperative mapping used two epicardial patches with only partial coverage. This was decided by the surgeon in order to minimize constriction of the heart and ensure patient safety. To optimize spatial sampling, placement of the two patches was anterior RV and posterior LV in two patients and anterior RV and lateral RV in one patient. Epicardial electrode positions during surgery were estimated from digital photographs and anatomic markers. The associated inaccuracy may account for some of the differences between the invasive and noninvasive epicardial potential patterns and isochrone maps. Despite these differences, there was good correspondence between the images obtained by noninvasive ECGI and the invasive maps.

Cardioplegia and anesthesia effects—Intraoperative recordings were performed without cardioplegia. In patient OR3, body surface potential mapping was performed postsurgery during recovery from cardioplegia and reperfusion. In a porcine model of cardiopulmonary bypass, Cohen et al²⁷ showed that cold blood cardioplegia had no effect on activation patterns, effective refractory period, or conduction velocities during sinus and epicardially paced beats. Our results were consistent with the study by Cohen et al. The differences in QT interval (mean $\Delta QT = 90$ ms) between invasive and noninvasively reconstructed electrograms (Figure 6) could be attributed to rate adaptation²⁸ (80 bpm during intraoperative mapping vs 111 bpm during body surface potential mapping). Similar adaptation range of action potential duration was measured in human myocytes.²⁸ Anesthesia slowed conduction and prolonged refractoriness.²⁹ These effects were apparent in Figure 7, where intraoperative isochrones depict delayed activation times relative to the ECGI maps in the unanesthetized state.

Reconstruction to epicardium—Noninvasive potentials, electrograms, and isochrones are reconstructed on the epicardial surface. Although most clinical ventricular tachycardia (VT) originates from the subendocardium, it almost always contains an intramural component.³⁰ We have shown that noninvasive epicardial potentials, electrograms, and isochrones provide information on intramural ectopic activation⁵ and reentry.⁷ There is increasing evidence that clinical reentrant VTs (~33%) involve epicardial components or originate in the subepicardium.³¹ Inability to ablate subepicardial targets has been associated with failed catheter ablation of clinical VT.²² To overcome this difficulty, a method for transthoracic catheter ablation of epicardial sites has been developed and applied successfully.^{23,24,31} Images obtained by ECGI can identify epicardial components of VT and guide such ablation. It also can be used to evaluate outcome of the intervention in follow-up examinations.

Acknowledgments

We thank Leslie Ciancibello for technical assistance with CT and Dr. John Haaga for access to CT facilities; Dr. Jayakumar Sahadevan for advice and assistance in intraoperative mapping; and Elena T. DuPont for assistance in transferring CT images.

This study was supported by NIH-NHLBI Grants R37-HL-33343 and R01-HL-49054 to Dr. Rudy and by a Whitaker Foundation Development Award. Dr. Rudy is the Fred Saigh Distinguished Professor of Engineering at Washington University in St Louis.

References

1. Rudy, Y. The electrocardiogram and cardiac excitation. In: Sperelakis, N.; Kurachi, Y.; Terzic, A.; Cohen, MV., editors. *Heart Physiology and Pathophysiology*. Academic Press; San Diego, CA: 2000. p. 133-48.
2. Spach MS, Barr RC. Ventricular intramural and epicardial potential distributions during ventricular activation and repolarization in the intact dog. *Circ Res* 1975;37:243–257. [PubMed: 1149199]
3. Messinger Rapport B, Rudy Y. Noninvasive recovery of epicardial potentials in a realistic heart-torso geometry: normal sinus rhythm. *Circ Res* 1990;66:1023–1039. [PubMed: 2317885]
4. Oster HS, Taccardi B, Lux RL, Ershler PR, Rudy Y. Noninvasive electrocardiographic imaging: reconstruction of epicardial potentials, electrograms, and isochrones and localization of single and multiple electrocardiac events. *Circulation* 1997;96:1012–1024. [PubMed: 9264513]
5. Oster HS, Taccardi B, Lux RL, Ershler PR, Rudy Y. Electrocardiographic imaging: noninvasive characterization of intramural myocardial activation from inverse-reconstructed epicardial potentials and electrograms. *Circulation* 1998;97:1496–1507. [PubMed: 9576431]
6. Burnes JE, Taccardi B, MacLeod RS, Rudy Y. Noninvasive ECG imaging of electrophysiologically abnormal substrates in infarcted hearts: a model study. *Circulation* 2000;101:533–540. [PubMed: 10662751]
7. Burnes JE, Taccardi B, Ershler PR, Rudy Y. Noninvasive ECG imaging of substrate and intramural ventricular tachycardia in infarcted hearts. *J Am Coll Cardiol* 2001;38:2071–2078. [PubMed: 11738317]
8. Ghanem RN, Burnes JE, Waldo AL, Rudy Y. Imaging dispersion of myocardial repolarization, II—noninvasive reconstruction of epicardial measures. *Circulation* 2001;104:1306–1312. [PubMed: 11551884]
9. Burnes JE, Taccardi B, Rudy Y. A noninvasive imaging modality for cardiac arrhythmias. *Circulation* 2000;102:2152–2158. [PubMed: 11044435]
10. Ramanathan C, Ghanem RN, Jia P, Ryu K, Rudy Y. Electrocardiographic imaging (ECGI): a noninvasive imaging modality for cardiac electrophysiology and arrhythmia. *Nat Med* 2004;10:422–428. [PubMed: 15034569]
11. Ramanathan C, Jia P, Ghanem R, Calvetti D, Rudy Y. Noninvasive electrocardiographic imaging (ECGI): application of the generalized minimal residual (GMRes) method. *Ann Biomed Eng* 2003;31:981–994. [PubMed: 12918913]
12. Green LS, Taccardi B, Ershler PR, Lux RL. Epicardial potential mapping: effects of conducting media on isopotential and isochrone distributions. *Circulation* 1991;84:2513–2521. [PubMed: 1959201]
13. Rudy Y, Plonsey R. A comparison of volume conductor and source geometry effects on body surface and epicardial potentials. *Circ Res* 1980;46:283–291. [PubMed: 6444278]
14. Rudy Y, Burnes JE. Noninvasive electrocardiographic imaging. *Ann Noninvas Electrocardiol* 1999;4:340–359.
15. Spach MS, Barr RC, Lanning CF, Tucek PC. Origin of body surface QRS and T wave potentials from epicardial potential distributions in the intact chimpanzee. *Circulation* 1977;55:268–278. [PubMed: 832342]
16. Ramanathan C, Rudy Y. Electrocardiographic imaging II: effect of torso inhomogeneities on the noninvasive reconstruction of epicardial potentials, electrograms and isochrones. *J Cardiovasc Electrophysiol* 2001;12:241–252. [PubMed: 11232625]
17. Durrer D, Van Dam RT, Freud GE, Janse MJ, Meijler FL, Arzbaecher RC. Total excitation of the isolated human heart. *Circulation* 1970;41:899–912. [PubMed: 5482907]

18. Colli Franzone P, Guerri L, Taccardi B. Potential distributions generated by point stimulation in a myocardial volume: simulation studies in a model of anisotropic ventricular muscle. *J Cardiovasc Electrophysiol* 1993;4:438–458. [PubMed: 8269311]
19. Streeter DD, Sputnitz HM, Patel DP, Ross J, Sonnenblick EH. Fiber orientation in the canine left ventricle during diastole and systole. *Circ Res* 1969;26:339–347. [PubMed: 5766515]
20. Spach MS, Barr RC. Analysis of ventricular activation and repolarization from intramural and epicardial potential distributions for ectopic beats in the intact dog. *Circ Res* 1975;37:830–843. [PubMed: 1192575]
21. Gima K, Rudy Y. Ionic current basis of electrocardiographic waveforms: a model study. *Circ Res* 2002;90:889–896. [PubMed: 11988490]
22. Schweikert RA, Saliba WI, Tomassoni G, Marrouche NF, Cole CR, Dresing TJ, Tchou PJ, Bash D, Beheiry S, Lam C, Kanagaratnam L, Natale A. Percutaneous pericardial instrumentation for endo-epicardial mapping of previously failed ablations. *Circulation* 2003;108:1329–1335. [PubMed: 12952851]
23. Sapp J, Soejima K, Couper GS, Stevenson WG. Electrophysiology and anatomic characterization of an epicardial accessory pathway. *J Cardiovasc Electrophysiol* 2001;12:1411–1414. [PubMed: 11797999]
24. Swarup V, Morton JB, Arruda M, Wilber DJ. Ablation of epicardial macroreentrant ventricular tachycardia associated with idiopathic nonischemic dilated cardiomyopathy by a percutaneous transthoracic approach. *J Cardiovasc Electrophysiol* 2002;13:1164–1168. [PubMed: 12475110]
25. Laxer C, Ideker RE, Pilkington TC. The use of unipolar epicardial QRS potentials to estimate myocardial infarction. *IEEE Trans Biomed Eng* 1985;32:64–67. [PubMed: 3980032]
26. Janse MJ, Wit AL. Electrophysiological mechanisms of ventricular arrhythmias resulting from myocardial ischemia and infarction. *Physiol Rev* 1989;69:1049–1169. [PubMed: 2678165]
27. Cohen NM, Allen CA, Hsia PW, Nixon TE, Wise RM, Damiano RJ Jr. Electrophysiologic consequences of hyperkalemic cardioplegia during surgical ischemia. *Ann Thorac Surg* 1994;57:1076–1083. [PubMed: 8179367]
28. Li GR, Yang B, Feng J, Bosch RF, Carrier M, Nattel S. Transmembrane I_{Ca} contributes to rate-dependent changes of action potentials in human ventricular myocytes. *Am J Physiol* 1999;276:H98–H106. [PubMed: 9887022]
29. Lefrant JY, Muller L, de La Coussaye JE, Lalourcey L, Ripart J, PA Peray, Mazoit X, Dauzat M, Sassine A, Eledjam JJ. Hemodynamic and cardiac electrophysiologic effects of lidocaine-bupivacaine mixture in anesthetized and ventilated piglets. *Anesthesiology* 2003;98:96–103. [PubMed: 12502985]
30. Kaltenbrunner W, Cardinal R, Dubuc M, Shenasa M, Nadeau R, Tremblay G, Vermeulen M, Savard P, Page PL. Epicardial and endocardial mapping of ventricular tachycardia in patients with myocardial infarction. Is the origin of the tachycardia always subendocardially localized? *Circulation* 1991;84:1058–1071. [PubMed: 1884439]
31. Sosa E, Scanavacca M, D'Avila A, Piccioni J, Sanchez O, Velarde JL, Silva M, Reolao B. Endocardial and epicardial ablation guided by nonsurgical transthoracic epicardial mapping to treat recurrent ventricular tachycardia. *J Cardiovasc Electrophysiol* 1998;9:229–239. [PubMed: 9580377]

Figure 1. Study design. **A:** Cross-sectional CT slice showing the heart and body surface electrodes. **B:** Vest for body surface potential mapping. **C:** Subject-specific heart-torso geometry digitized from CT (*blue circles* indicate electrode positions). **D:** Noninvasive epicardial potentials obtained from CT and body surface potential mapping data using electrocardiographic imaging. **E:** Epicardial patches used for intraoperative mapping. **F:** Snapshot of the intraoperative mapping procedure showing the anterior right ventricle (RV) with inserted epicardial patch and epicardial pacing wires.

Figure 2.

Invasive (**left column**) and noninvasively imaged (**middle column**) epicardial potential maps from body surface potential mapping (BSPM; **right column**) 10 ms (**top row**), 21 ms (**second row**), and 30 ms (**bottom row**) following QRS onset. Numbers (**second row**) indicate right ventricular breakthrough sites. Left anterior descending coronary artery (LAD) and AV groove are marked in *thick gray lines*. LV = left ventricle.

Figure 3.

Invasive (*red*) and noninvasively reconstructed (*black*) QRS electrograms with cross-correlation (CC) and time lag (in milliseconds) during sinus rhythm. **A:** Anterior right ventricle (RV). **B:** Posterior left ventricle (LV). Boundaries of epicardial patches are shown. Numbers identify electrogram locations. LAD = Left anterior descending coronary artery.

Figure 4.

A: Invasive (**first and last columns**) and noninvasively imaged (**second and third columns**) epicardial potentials during right ventricular (RV) endocardial pacing. Times are relative to onset of pacing. *Asterisk* denotes pacing site. **B:** Invasive and noninvasively imaged potentials during repolarization (T peak). **C:** Invasive (*red*) and noninvasively reconstructed (*black*) electrograms. LV = left ventricle.

Figure 5.

A: Patient OR3. Invasive potentials from lateral (**first column**) and anterior (**last column**) right ventricle (RV) with corresponding noninvasively imaged potentials (**second and third columns**), 80 ms (**top row**) following pacing and during repolarization (T peak, **bottom row**). Minus sign (–) denotes pacing site. **B:** Patient OR2. Invasive and noninvasively imaged epicardial potentials at 70 ms following pacing and during repolarization. Same format as panel A. LAD = Left anterior descending coronary artery.

Figure 6.

Invasive (*red*) and noninvasively reconstructed (*black*) electrograms with cross-correlation (CC) and time lag (in milliseconds) during right ventricular (RV) epicardial pacing. **A:** Anterior RV. Asterisk denotes reconstructed pacing site. **B:** Lateral RV.

Figure 7. Invasive (first and last columns) and noninvasively imaged (second and third columns) isochrones. **A:** Sinus rhythm. Numbers indicate right ventricular (RV) breakthrough sites. **B:** RV endocardial pacing. **C:** RV epicardial pacing. Arrows point to regions of slow conduction (crowded isochrones). Asterisk denotes pacing site. LAD = left anterior descending coronary artery; LV = left ventricle.

Table 1

Patient population and study summary

Patient	Gender	Age (yr)	Race	Disease	Geometry acquisition	Heart rhythm during mapping			LVEF
						Preoperative BSPM	Intraoperative Mapping	Postoperative BSPM	
OR1	Female	64	Black	Coronary artery disease	CT	SR	SR	—	~50%
OR2	Female	59	Black	Chronic AF Mitral valve regurgitation Tricuspid valve regurgitation PFO	CT	Endocardial pacing (65 bpm)	Endocardial pacing (70 bpm) and RV epicardial pacing (90 bpm)	RV epicardial pacing (110 bpm)	~50%
OR3	Male	71	White	Aortic Arch Aneurysm	Digitizer/CT	SR	SR and RV epicardial pacing (80 bpm)	AV pacing (111 bpm)	~60%

AF = atrial fibrillation; AV = atrioventricular; BSPM = body surface potential mapping; CT = computed tomography; LVEF = left ventricular ejection fraction; PFO = patent foramen ovale; RV = right ventricular; SR = sinus rhythm.

Table 2

Summary of results

	Electrograms CC				TL (ms)		
	Max	Min	Mean	SD	Max	Min	SD
Sinus Rhythm (OR1)							
Anterior RV	0.99	0.16	0.89	0.14	35	-1	7.85
Posterior LV	0.98	0.04	0.67	0.24	49	0	11.98
Sinus Rhythm (OR3)							
Anterior RV	0.99	0.57	0.83	0.11	5	-19	-7.06
Lateral RV	0.95	0.01	0.58	0.30	65	-20	14.01
Endocardial Pacing (OR2)—Localization accuracy 4.3 mm							
Anterior RV	0.97	0	0.55	0.23	180	-82	4.13
Posterior LV	0.95	0.13	0.77	0.17	35	-105	-10.95
Epicardial Pacing (OR2)—Localization accuracy 18.8 mm							
Anterior RV	0.97	0.12	0.64	0.20	111	-154	-31.74
Posterior LV	0.98	0.02	0.66	0.21	116	-85	-25.63
Epicardial Pacing (OR3)—Localization accuracy 15.8 mm							
Lateral RV	0.99	0.13	0.79	0.18	55	-35	7.87
Anterior RV	0.99	0.25	0.88	0.15	61	-23	18.88

CC = cross-correlation; LV = left ventricle Max = maximum; Min = minimum; RV = right ventricle; TL = time lag (in milliseconds); SD = standard deviation.