event of pain or for revisions can be technically quite challenging since various screw manufacturers use different types of screw-drivers and the screw heads can be of varying sizes.

TECHNIQUE

We have devised a simple technique for removing such screws using Steinman pin and bone cement. The senior author (NA) has used this successfully in three cases. The cross-section of the proximal (blunt) end of the Steinman pin is quadrangular or triangular. The blunt end of a pin is inserted into the screw socket to check for a reasonably snug fit and the pin is bent into an 'L'-shape to aid gripping. A small quantity of bone cement is then pressed into the screw socket, the end of the pin is tapped in, and the cement is left to set. An essential component of the technique is to allow time for the cement to set. Once the cement is set, the screw can be removed without much effort by turning on the bent end of the pin.

DISCUSSION

Cannulated screws often need to be removed if they are causing symptoms or in patients subsequently requiring hip replacement or a revision of their ACL reconstruction. This is an essential part of surgical process, which can be quite challenging and often frustrating. A set of sterile Allen keys of different diameters and lengths would be a possible solution but these are rarely available in most operating theatres. There are different techniques described in the literature for removal;¹⁻⁴ however, one often attempts to use the pliers. Removing such screws by gripping their ends with pliers is not only difficult and time-consuming, but also often necessitates a considerable amount of bone around the screw to be chiselled or nibbled out specially in the case of headless cannulated screws commonly used in ACL reconstruction.

References

1. Wilson ER, Ruoff MJ. Tips of the trade #39. Technique for removal of cannulated screws. *Orthop Rev* 1991; **20**: 807-9.
2. Cameron HU. Tips of the trade #15. Removal of hardware prior to THR. *Orthop Rev* 1989; **18**: 918-20.
3. Eyres K. An aid for the removal of compression hip screws. *Br J Clin Pract* 1990; **44**: 396.
4. Phipatanakul WP, Watson JT. Dynamic hip screw removal using blade plate extraction device. *Am J Orthop* 2003; **32**: 362-3.

Technical Tips

Knee arthroscopy: 1-ml syringe as an out-flow cannula

P MUKHERJEE

Torbay Hospital, Torquay, UK

CORRESPONDENCE TO

Dr P Mukherjee, SHO BST, c/o Doctor's Mess, Torbay Hospital, Torquay TQ2 7AA, UK

M: +44 (0)7832 136163; E: pavelmukherjee@yahoo.co.uk

A 'soft' quill can be used to drain 'arthroscopy debris'. A 1-ml syringe with the nozzle cut off at an angle of 45° is a better alternative – being stiff, it is easier to manoeuvre in the joint. To direct the entry of the syringe during insertion, slide it over the arthroscopic trochar, positioning the trochar point where the syringe will end at insertion into the joint. This 'Seldinger technique' also eases its passage through the soft tissues.

The fibre-glass spica bar

STEPHEN M BLAKE, PETER J COX

Princess Elizabeth Orthopaedic Centre, Royal Devon and Exeter Hospital, Exeter, UK

CORRESPONDENCE TO

Mr Stephen M Blake, Specialist Registrar in Orthopaedics, Princess Elizabeth Orthopaedic Centre, Royal Devon and Exeter Hospital, Barrack Road, Exeter EX2 5DS, UK
M: +44 (0)7799 146 282; E: stephenblake_uk@yahoo.co.uk

Hip spicas control hip position and are used in the management of a number of conditions, e.g. developmental dysplasia of the hip (DDH). The wooden bar between the thighs markedly

Figure 1 A fibre-glass spica bar.