Figure 1 Cortical screw inserted into the trapezium.

is often challenging due to its small dimensions and intimate ligamentous attachments. When removed piece-meal, remnants of trapezium left *in situ* may cause persistent symptoms. We describe a simple surgical technique to aid single piece trapezial excision.

TECHNIQUE

The trapezium is approached via a volar incision centred over the trapezio-metacarpal joint. The midpoint of the volar aspect of the trapezium is identified and perforated in the antero-posterior plane using a 2.5-mm drill. A standard 3.5-mm partially threaded cortical screw of 18-mm length is inserted into the drill hole utilising the full threaded length (Fig. 1). The screw is held using a pair of Kocker's forceps and sharp dissection is carried out aided by 'joysticking' and tractioning manoeuvres via the forceps. The trapezium is thereby delivered as a single piece.

DISCUSSION

Due to inaccessibility of all trapezial ligaments using conventional techniques, often piece-meal trapeziectomy is unavoidable. This may result in risk of injury to surrounding structures, increased operation time and incomplete excision of trapezial fragments. The senior author (SA) has employed this method of trapeziectomy for over 10 years without major complications. We believe this to be a safe and reproducible technique for excising the trapezium as one piece.

The Nottingham lead foot

MATTHEW K BOULTER, THOMAS K ROWLANDS, SUNIL DHAR
Department of Orthopaedics, Nottingham City Hospital,
Nottingham, UK

CORRESPONDENCE TO

Matthew K Boulter, Clinical Fellow, Department of
Orthopaedics, Nottingham City Hospital, Hucknall Road,
Nottingham NG5 1PB, UK
E: matthew167@hotmail.com

BACKGROUND

Flexor tenotomy of the toes to correct clawing is a common operation in foot surgery and is frequently performed by a surgeon operating alone. The technique involves manual dorsiflexion of each toe, exposure of the plantar surface, dissection to the underlying tendon and tenotomy followed by wound closure. The majority of the operation is commonly performed using one hand as the other is required to maintain the toe in dorsiflexion. It is, therefore, difficult to retract the wound edges, cut the tendon under direct vision and suture the wound.

TECHNIQUE

We have designed a 'lead foot' which is able to hold toes firmly in dorsiflexion, thus freeing the surgeon to operate with both hands allowing safer and quicker surgery (Fig. 1). Lead feet are simple and cheap to produce, are autoclaveable, re-usable and durable. We recommend the use of 2-mm thick roofing lead which is available from builders' merchants. The adult version is depicted (Fig. 2) but we also use a two-thirds' scale version for paediatric cases. The Nottingham lead foot is also potentially suitable for other operations on the plantar aspect of the foot and toes.

Figure 1 Cut lead sheet showing support structures.

Figure 2 The Nottingham lead foot in position.

Figure 1 Emergency technique to tighten a Jacob's chuck.

Slip on a stockinette

STIJN HERMANS, STEPHEN M BLAKE, JONATHAN R HOWELL
Princess Elizabeth Orthopaedic Centre, Royal Devon and
Exeter Hospital, Exeter, UK

CORRESPONDENCE TO

SM Blake, Princess Elizabeth Orthopaedic Centre, Royal Devon
and Exeter Hospital, Barrack Road, Exeter EX2 5DS, UK
M: +44 (0)7799 146282; E: stephenblake_uk@yahoo.co.uk

TECHNICAL TIPS

ALTERNATIVE TECHNIQUE TO TIGHTEN A JACOB'S CHUCK

D SUNDERAMOORTHY¹, N GEARY²

¹Dept of Orthopaedics, Royal Gwent Hospital, Newport, Gwent, UK

²Dept of Orthopaedics, Arrowe Park Hospital, Upton, Wirral, UK

CORRESPONDENCE TO

D Sunderamoorthy, 10 New Hampshire Close, Great Sankey,
Warrington WA5 8DD, UK
E: dsundar6@hotmail.com

What do you do if, during an operation, you drop a chuck key and there is no replacement? In emergencies, it is possible to tighten the chuck using what is to hand. Put an instrument such as the tip of a Trethowan bone lever or Steinmann pin in the round hole on the chuck that receives the chuck key. Choose a periosteal elevator, old osteotome or flat-headed screwdriver; engage this in the gear face of the chuck and lever against the Trethowan to tighten the chuck.

The manipulation and application of backslabs for limb fractures is a common emergency department and orthopaedic practice. We would like to present an alternative method of plaster application which allows easier handling, avoids excessive soiling of the plastering environment and continues to permit a high level of mouldability.

Figure 1 Rolling the stockinette over the plaster.