Although in Group A the skin infection was minor in character and there was no diarrhoea, yet the potential danger of cross-infection inherent in hospitalization was always present. It is, of course, well known that the infant born in hospital is exposed to risks of this nature even when reasonable precautions are taken. It is not so well appreciated that the advantage which the baby born at home has in being less exposed to the hazard of cross-infection may be nullified by the danger he runs from respiratory infection acquired from unmasked relatives and visitors, and this fact should be impressed on all those responsible for the care of the newborn baby in his own home.

We should like to express our thanks to the staff of the Sorrento Maternity Hospital and the domiciliary midwifery staff for their assistance in this investigation.

Medical Memoranda

A Simple Pneumoflator

Though many anaesthetists may doubt the necessity for or value of full pneumoflator anaesthesia, I think all would agree on the convenience of frequent mechanical assistance

during controlled respiration. The machine illustrated was designed to act as a "mechanical hand" on the many occasions when the anaesthetist, working alone, had to relinquish the bag to fill syringes, etc. The prototype shown here was built of pressed steel by Mr. T. Eames, of Paddington. Further

machines have been constructed in the workshops of the National Hospital, Queen Square, and are enclosed in an outer casing, thus producing almost complete silence.

The motors used have been R.A.F. disposals predictor motors of the 10 KB/565, 230 A.C. type with worm drive gear-boxes. These motors run at 3,000 r.p.m., reduced to 23 r.p.m. on the output shaft. They give ample power. Suitable motors for further machines can be obtained from Messrs. Klaxon, 201, Holland Park Avenue, W.11.

The whole machine has been kept as simple as possible, and there is only one control—a rheostat that adjusts the rate of the motor. A 1-gallon (4.5-litre) bag is used. The pressure can be varied according to the flow of gas and consequent inflation of the bag. With a manometer it has been found that a pressure of 15–20 cm. of water may be obtained, which normally is ample.

The prototype has now run for about 100 hours and has shown no weaknesses. It has been used for its proper purpose as a mechanical hand and also continuously as a full-scale pneumoflator in many cases at the London Chest Hospital and the National Hospital, Queen Square. There may be a use for such an inexpensive and simple device in various fields, including the prolonged treatment of poliomyelitis.

The motor runs absolutely cool, but at present it is not suitable for use with explosive mixtures; in practice this is no disadvantage. Should such use be envisaged, both motor and rheostat must be of suitable type, and these are easily obtainable.

I would like to express my thanks to Mr. T. Holmes Sellors, Mr. Harvey Jackson, and Mr. C. H. Bateman for their kindness in allowing me to develop this machine during treatment of their cases.

R. ATWOOD BEAVER, B.M., F.F.A.R.C.S.
Senior Anaesthetist, London Chest Hospital.

Repeated Narcosis in a Young Child

It was decided to expose a child of 16 months to deep x-ray therapy for a retinoblastoma. The exposures were to be made on 10 occasions at two- to three-day intervals, and each treatment was to be for 24 minutes. The x-ray applicator was to be held in position by a plastic mask which fitted over the child's face, and during therapy the child had to be isolated in the treatment-room, under observation from a small window. We were asked to provide narcosis at the appropriate times. It was clear that inhalation anaesthesia was unsuitable and that the desiderata were: minimal disturbance of normal feeding and routine; certain response to narcosis; low toxicity; ease of administration; and non-inflammability. We felt that these conditions were best met by administering thiopentone sodium rectally. This method of narcosis has been described by Weinstein (1939), Schotz (1942), Burnap *et al.* (1948), and others.

TECHNIQUE

Diet was normal up to three hours before the treatment. A rectal washout was given one hour before induction on the first three occasions only. Its subsequent omission caused no inconvenience. Atropine, 1/150 gr. (0.43 mg.), was given subcutaneously half an hour before induction. The injection of anaesthetic solution was given slowly, per rectum, through a rubber catheter, which was then clamped and left *in situ*. The child fell asleep in four to ten minutes and remained asleep during the subsequent therapy, at the end of which any residual solution (amounting to 0.4 ml.) was allowed to drain away through the catheter. The treatment, including the time taken for adjustment of mask and refocusing on the contralateral eye, lasted for approximately 40 minutes. A paper flag placed on the child's abdomen during therapy facilitated our observation of the respiration. Oxygen was available in case of need, but was not required.

Dosage.—The child's weight was 25 lb. (11.3 kg.). The recommended dose of 0.5 g. was made up in 2½% solution. After the first three administrations the periods of undisturbed sleep were, respectively, two and a half, three and a half, and four and a quarter hours. We then felt that a residual effect was occurring. This was attributed to accumulation of thiopentone in the body-fat, as described by Lief (1951) and by Brodie (1951). Accordingly, the dosage was gradually reduced in subsequent administrations from 20 to 12 mg. per lb. (44 to 26.4 mg. per kg.). With these doses the period of undisturbed sleep varied from two and a half to three hours.

Complications.—The child developed a cold between the second and third administrations. This was not, in the circumstances, regarded as a contraindication to narcosis. It cleared up in a few days. There was no sign of rectal irritation. Bowel movement continued regularly and normally. The child was fed half an hour after awakening. No ill effects were observed.

Eight months later a second course of deep x-ray therapy was instituted, six treatments being given in 14 days. The dose of thiopentone required at first was 61 mg. per kg. This was gradually reduced first to 53 mg. and later to 45 mg. per kg.

Conclusion.—It seems to us that the technique described is safe and satisfactory for repeated narcosis in small children.

Our acknowledgments are due to Professor W. J. B. Riddell for permission to publish this case, and to Dr. H. H. Pinkerton for his encouragement.

R. CECIL BROWN, M.B., Ch.B., D.A.,
Senior Registrar Anaesthetist, Western Infirmary, Glasgow.
J. BRIAN LISTER, M.B., Ch.B., D.A.,
Junior Hospital Medical Officer (Anaesthesia),
Western Infirmary, Glasgow.

REFERENCES

- Brodie, B. B. (1951). Int. Congress of Anaesthetists, London. Abst. *British Medical Journal*, 1951, 2, 668.
Burnap, R. W., Gain, E. A., and Watts, E. H. (1948). *Anesthesiology*, 9, 524.
Lief, P. A. (1951). Int. Congress of Anaesthetists, London. Abst. *British Medical Journal*, 1951, 2, 668.
Schotz, S. (1942). *Anesth. and Analg.*, 21, 295.
Weinstein, M. L. (1939). *Ibid.*, 18, 221.