BRITISH MEDICAL JOURNAL

LONDON SATURDAY NOVEMBER 23 1946

PHYSIOLOGICAL REST*

WITH SPECIAL REFERENCE TO ARTHRITIS AND NERVE LESIONS AND
TO THE MANUFACTURE OF APPLIANCES

BY

NORMAN CAPENER, F.R.C.S.

FIG. 1.—Shoulder platform splint—physiologically unsatisfactory.

In this country during the nineteenth century the outstanding exponents of the use of rest in the treatment of disease were John Hilton and Hugh Owen Thomas. The former in his Arris and Gale Lectures of 1859–61 coined the phrase “physiological rest,” which following Hunterian teaching he called “natural therapeutics.” Hilton aimed at dissociation of physiological activity from a diseased or injured limb. He emphasized the dependence of both growth and repair upon rest. Hugh Owen Thomas (1876) completed Hilton’s argument by showing how rest should be used. Thomas is known to many as the designer of numerous splints that go by his name. But more important than this were the principles he taught for their application.

In this century Robert Jones achieved for Thomas the acceptance which was so largely denied him during life. Thomas overstated his case, yet for the conditions of his day the emphasis was needed just where he placed it, namely, upon “rest—enforced, uninterrupted, and prolonged.” Conditions of medicine, as of life in general, have changed. Though Thomas’s views need modification, he has important lessons for us now. While neither Hilton nor Thomas was an orthopaedic surgeon in the restricted modern sense, the ideas which they taught have resulted in the establishment of orthopaedics as a differentiated (not necessarily “specialist”) field of medicine and surgery. In the aseptic evolution of technical

surgery it has been the appreciation of the meaning of physiological rest which has provided the orthopaedic principle. (“The ultimate basis of the existence of something.”—O.E.D.)

Definition

Physiological rest as we understand it to-day is a controlled process, giving immobilization—while safeguarding functional restoration—of a part or the whole of the body in the position most favourable for the biological processes of growth and repair to the maximum extent required for these processes in the particular disease or injury being treated. Physiological rest is not necessarily absolute or uninterrupted rest, though for some period it may be so. Its purposes are to promote healing by reducing the local physiological demands of the limb, by avoiding the aggravation of pathological processes, and by bringing to bear upon the lesion the whole general metabolic resources of the patient.

It must be arranged so as to prevent the deforming influences of uncontrolled or unbalanced muscular activity, of muscular atrophy, and of gravity. Relaxation and restriction of muscular activity is modulated according to the disease. In some conditions rest will alternate rhythmically with movement, the relation of one to the other varying inversely as healing proceeds. While pathological conditions are being relieved by rest, no extrinsic pathological factors should be introduced. Local treatment must be planned within the most favourable general background for physical and psychological health

FIG. 2.—The adjustable Littler-Jones abduction splint. (Wingfield-Oxford type.)

* Abstracted from an Arris and Gale Lecture delivered at the Royal College of Surgeons on Jan. 3.

FIG. 3.—(a and b) The physiological position for the hand.
(c) A plaster moulded splint in this position.

The local positions and their duration vary according to the different conditions being treated. Physiological rest in some form is an important requirement in the treatment of such varied conditions as fractures, dislocations, acute and chronic arthritis, tuberculosis, osteomyelitis, peripheral nerve injuries, infantile paralysis, burns, and the lacerated hand; but each calls for particular modifications, all of which are covered by the definition as stated. In chronic deforming arthritis in particular there is scope for improvement—for a clearer realization of what could be done to relieve local disease, to improve general well-being, to prevent deformity, and to maintain or restore function. Furthermore, in all fields there is need for better knowledge of splint design, construction, and usage.

Arthritis

Whatever may be the cause of chronic deforming arthritis, it is necessary to emphasize the interrelation of the morbid change in joints with the spasm and atrophy of muscles, and conversely of the contributory effects of deficient muscular activity upon joint degeneration, and, furthermore, the aggravation that the local process provides for the whole constitutional disease. Therefore, no matter what medical or physical therapeutic measures are applied, most joints affected with arthritis pass at various times through phases of activity which require immobilization.

While joints are acutely inflamed the most perfect immobilization is necessary and must be carried out until the inflammatory process has subsided and muscle spasm is relieved. The subsidence of pain and the diminution of local tenderness, swelling, and irritability are important guides. Defects in the application of rest are, first, not giving adequate attention to the posture in which the joint is placed; secondly, not guarding against the ill effects of movement, friction, pressure, constriction, traction, and gravity; thirdly, imperfect timing of the duration or modification of rest and, consequently, failure

FIG. 4.—Preparation of a semi-permanent block leather or plastic splint suitable for arthritis of the wrist or basal thumb-joints. (a) Plaster-of-Paris mould cut before removal. (b) Finished splint prepared on a metal casting.

adequately to safeguard the maintenance or restoration of function when permissible.

For every normal joint there is a physiological position which is optimal for the average needs of the limb—a position in which its controlling musculature is able to exert its power to best advantage, a position in which muscles are poised for function and where none of the flexible capsular tissues is being stretched. The “physiological positions” of limbs are completely different from the so-called “anatomical positions,” which merely provide standards of convenience and, clinically, are almost always unsuitable and indeed undesirable for treatment. The physiological position is not a fixed one but varies within a moderate range generally near the middle of the normal range of movement of the joint. The optimal position for the treatment of arthritic joints—modified from this physiological position—must be not only suitable for the pathological and therapeutic requirements but also such that if stiffness or ankylosis occurs it shall be in a position which will be best for function. In other words a nice compromise must be established between the optimal position for muscular competence and the later possible requirements of joint posture. These positions are enumerated in the following list:

Table showing the Optimal Positions for the Treatment of Arthritic Joints

Joint	Position	Remarks
Shoulder	Abduction 45° Flexion 30° External rotation 15°	Prevents stiffness in adduction; aids deltoid function; safeguards capsular integrity. The position is favourable for function in ankylosis
Elbow	Extended to about 100°	Safeguards the function of the elbow flexor muscles. If ankylosis is likely to occur modifications of the position recommended must be made according to the wishes of the patient, guided by his social circumstances and occupation. Where both elbows are involved one should be treated at a smaller angle
Forearm	Mid prone-supine position	Favours the restoration of either movement. If ankylosis occurs compensation for loss of pronation can be gained by abduction of the shoulder. Stiffness in supination is objectionable
Wrist	Extension 30°	The angle is measured upon the dorsum of the radius and of the third metacarpal
Digits and thumb	All joints flexed about 25° Abducted and opposed moderately	The hand as a unit is held in the position of semi-grasp, the thumb practically continuing the line of the lateral border of the radius, the phalanges of all digits in the position of natural relaxation—e.g., that assumed when the normal hand is loose at the side of the body. Capsular and ligamentous stretching is avoided. The vicious position of metacarpophalangeal hyperextension is avoided
Spine	Normal curves maintained	Flexion as well as lordosis of the lumbar spine to be avoided
Hip	Abduction only of sufficient amount to compensate for true shortening Rotation nil Flexion 20°	In arthritis extreme degrees of true shortening of a lower extremity are unlikely to be found. Flexion permits sitting if there is compensatory movement in the lumbar spine
Knee	Flexion 15°	Avoids posterior capsular stretching and resultant hyperextension instability. The competence of the quadriceps is adequately safeguarded
Ankle	Right-angle	If ankylosis appears inevitable the foot should be plantar-flexed 5° to 10° in men and a few degrees more in women to facilitate walking in high-heel shoes
Tarsal joints	Neutral as to inversion or eversion	While the arches of the foot should be preserved, it is better, should ankylosis be inevitable, to err in the direction of eversion, for the weight-bearing surface exposed to the ground is less likely to be troublesome than if the foot is inverted or cavoid
Toes	Maintain flexion of M.T.P. joints and extension of I.P. joints	Preserves intrinsic muscular activity; avoids metatarsophalangeal subluxation and metatarsal pressure points

In my view these positions are worthy of acceptance universally as the standard “physiological positions.” During the chronic stages of arthritic disease rest may still form a very important accompaniment of whatever other treatment is given. Here, however, there may be the added need to correct deformity. Finally, rest may even be used with the object of obtaining complete fixation or ankylosis of a joint, an end which may also be aided, of course, by surgical means.

FIG. 5.—Brian Thomas extensor appliance (Exeter model).

Lower Motor Neurone Lesions

All that has been said about muscular competence in arthritic disease applies also in the treatment of nerve lesions whether peripheral or poliomyelitic. There is indeed in the muscular phenomena associated with poliomyelitis a striking similarity to the muscular condition in arthritis. In the past it would appear that too great emphasis has been given to placing paralysed muscles in positions of greatest shortening, to the detriment of joint efficiency and of muscular re-education. An intermediate position is far better. Knowledge of the normal mechanism of maintaining physiological tonus (the simplest and gentlest type of reflex skeletal muscular activity) suggests this. Studies of the changes in denervated muscle (Tower, 1939; Bowden and Gutmann, 1944) indicate the same need, as well as the advisability of maintaining relative rest in good positions rather than absolute rest. The positions already listed are those suitable for the treatment of an extensive lower motor nerve lesion as in poliomyelitis. Even in the treatment of *any* individual peripheral nerve lesion the relevant positions indicated would be compatible with adequate protection of the paralysed muscles and suitable for muscular re-education. In nerve lesions the application of rest and support is associated where possible with substitution devices for replacement or assistance of deficient muscles.

Design of Orthopaedic Appliances

Physiological rest is far more than the fixation or splinting of limbs. It is perhaps unfortunate that the word "splint" has come to mean an immobilizing device, for originally a "splint" was something that permitted rather than restricted movement. In mediaeval times the movable sections of suits of armour—for example, at the elbow or knee—were called "splints."

Appliances are used for the following purposes: (a) Support; to oppose gravity. (b) Immobilization or limitation of movement. (c) The correction of deformity. (d) The mobilization of joints. (e) The substitution of defective muscular action. Appliances must be designed so as to fulfil these purposes with maximum comfort and convenience. This involves lightness and accessibility; smoothness of anatomical fit; adequate

FIG. 6.—Adaptation of same appliance for finger flexion deformities—e.g., ischaemic or arthritic.

FIG. 7.—Further modification of springs having "push" action for the treatment of metacarpo-phalangeal stiffness.

strength; simplicity of construction; and economical and speedy manufacture. Surgical appliance making has been in the past, and still is, in the hands of commercial craftsmen who have done splendid work, yet who are often divorced from adequate medical guidance. Control by the surgeon is greatly needed, and in order to bring the maximum good to the majority modern methods of manufacture should be more readily applied.

The greater mobility of populations necessitates the production of apparatus using standard parts which will receive servicing in one part of the country as in another, as in the case of the motor-car. In surgical appliance making there should be less conservatism.

Improved efficiency and economy could be achieved by an increased use of regionalization, and the setting up of workshops or the development of industries using similar materials with standard designs for certain appliances or portions thereof and with standard schedules for measuring and ordering. Standardization need not be an obstacle to initiative and advancement.

The wedding of mechanical efficiency to comfort in use is often a serious problem. The Thomas bed-knee splint, in spite of various later useful modifications, still remains as finally used by its inventor, a supreme example of the solution of this problem. Makeshift appliances are a particular anathema in orthopaedics, and with them "fitness for purpose" is "more honoured in the breach than the observance." There should be no excuse for the application to patients of appliances in the unhygienic, much-used, second-hand state so frequently found. Good plaster-of-Paris technique has made this unnecessary. Indeed, for most short-term requirements plaster-of-Paris, properly applied, provides the almost ideal material and the most readily available instrument of physiological rest. It is not necessary here to discuss this at length, although an illustration of the use of plaster in the treatment of arthritis in the hand is given.

FIG. 8.—A new device for the treatment of flexion contractures of the interphalangeal joints, using multiple strips of window-curtain springs.

FIG. 9. — The ulnar intrinsic spring splint.

elbow by means of a curved slotted rod with a bolt and nut permitting 90° variability in rotation. This appliance can be worn under the outer clothes and is equally suitable for wearing outside. Though in the model illustrated numerous straps are shown which distribute the weight of the apparatus, very few of these are essential. The lower part of the frame where it encircles the pelvis, instead of being made of metal rod, is fashioned from mild steel flat strip suitably padded to distribute pressure over a wide area. In some of the earlier versions of this Oxford modification pressure was experienced beneath the elbow, generally on the medial epicondyle or against the ulnar nerve. This difficulty has been overcome in a later model by dipping the framework downwards at this point. Here, therefore, is an appliance giving relative immobilization, and it is readily adjustable with splint benders because it is made of mild steel, thus making it possible to correct deformity. It is available for one side of the body only. Its chief defect is the presence of disagreeable vertical rods at the back and front of the trunk.

The Shoulder

A common appliance is the abduction splint used in the treatment of the stiff shoulder, in conditions due to neuro-vascular drag across the first rib, and for the support of the limb in injuries to the forearm and hand, etc. It is valuable in the correction and mobilization of deformity and for the support and re-education of weakened muscles. It is common to see the adaptation for these purposes of unsatisfactory appliances. In the one illustrated (Fig. 1) there is a mere platform with little adjustment: it fits merely where it touches and is uncomfortable and inconvenient.

The "Littler-Jones" frame (Fig. 2) is an improvement but in its original form did not provide a truly physiological position and lacked adjustability. At the workshops of the Wingfield-Morris Orthopaedic Hospital, Oxford, a great improvement was effected by division of the main frame at the elbow-joint and the addition of a forearm splint, articulating with the main frame beneath the

The Hand

In many ways the hand is at last receiving better care, particularly in fractures, wounds, and burns. The lessons from these fields should with modifications be applied to the treatment of arthritis. Various first-aid appliances of sheet metal are available, but none so far produced is satisfactory. In acute lesions of the interphalangeal joints, each digit has its own problems and must be treated very often individually. Plaster-of-Paris, which still supplies the most satisfactory material for immediate support, requires precision in application, and should have as much care devoted to it as a surgeon would expect to give to the application of a plaster jacket or spica.

I illustrate (Fig. 3) the physiological position for the hand and a plaster splint which with careful interdigital moulding will give some measure of individuality to the digits. The hand provides an example of the use of plaster moulds from which hard casts (Fig. 4) may be prepared upon which plastic material, such as leather or synthetic resins, may be moulded for more permanent appliances.

The problems of arthritis in the less acute states bear certain resemblances to those problems seen in peripheral nerve and vascular injuries, in which stiffness and deformation occur, the latter largely due to defective muscle balance and capsular contracture. Improvements in appliances for the hand have been made as the result of experiences with peripheral nerve injuries during the war. Some are illustrated (Figs. 5-9). To one particular splint I want to draw attention because it involves a principle which though not new could be applied probably to many other problems throughout the body. We are familiar with appliances which replace to some extent the toe-raising spring for paralytic drop-foot, and the M.R.C. elastic extensor appliance for radial-nerve palsy. In the recent war Brian Thomas (1944) applied the principle in a modified form, with a spring-wire elevator applied to a dorsal metal plate upon the forearm and wrist (Fig. 5). Mr. A. M. Hendry (1945) has applied the principle of adding springs to appliances

FIG. 10.—A coiled spring toe-raiser for drop-foot, and a similar device reversed and acting as a plantar flexor spring for calf-muscle paralysis.

FIG. 11.—(a) Paralytic scoliosis showing inability to sit without support. (b) The same patient with true spinal support taken from seat. (c) The same appliance in the course of manufacture.

for the replacement of intrinsic muscular activity in ulnar paralysis. In the ulnar splint illustrated (Fig. 9) the whole appliance is converted into one complete spring. The device is simple to make and apply, and has mechanical advantages in action. One advantage of all these appliances is the relative freedom of the palmar aspect for working purposes. A similar idea can be employed in the foot where, instead of an added spring, steel uprights of an ordinary short-leg appliance are converted into springs for the purpose of replacing either the dorsi-extensors or the plantar flexors of the foot (Fig. 10).

The Spine

The purpose to be achieved by a spinal support is essentially the limitation of movement. It is impossible for any appliance to limit all spinal movement. It is impossible to apply any force upon the exterior of the trunk in an ambulant patient so as to effect any material corrective influence upon structural defects. Support, however, for spinal deformity can be given to a considerable extent, as is shown in one case illustrated—a woman suffering from very severe paralytic scoliosis who, with an appliance, has been able to sit unaided with confidence and to manage her household affairs (Fig. 11).

This appliance illustrates the complexities of manufacture, the craftsmanship involved, and the need for close co-operation between craftsman and surgeon. The essential features are a spring steel pelvic band with spring steel uprights, two at the back and one at each side, the side struts supporting crutches for the axillae. These are joined by padded shoulder straps across the front of each shoulder to the back of the central posterior uprights. Pressure upon the pelvis is distributed through curved mild steel bands moulded above the iliac crests, and with downward extensions at the side of each hip to the seat. The support is strengthened by a moulded sheet of duralumin fitting into the concavity beneath the right ribs. Although such an appliance is not suitable for mass production, great benefit would accrue from standardization of some of its parts—e.g., in the provision of the special screws which are used and which, under present conditions, have to be made individually by hand on a power-driven lathe.

A simpler type of spinal "support," which is merely a brace to limit movement and assist muscular action, is the modified Taylor support made of leather-covered spring steel (Fig. 12). The advantages of this support are: simplicity of manufacture and convenience; the close relation of its uprights to the posterior surface of the spine, which controls antero-posterior movement; the firm hold at its base through the almost encircling steel band round the truss line; and the use of a suprapubic abdominal pad. Furthermore it is readily adjustable in shape by the use of bending irons. It is useful in the treatment of postural defects and adolescent kyphosis, in the convalescence of cases of tuberculosis of the spine, and in the after-care of spondylitis.

The Knee

In acute stages of arthritis of the knee nothing is better than a well-applied, cellulose-lined, long-leg guarding plaster, which can be split as soon as, on clinical grounds, the condition appears to be subsiding. It is then available as a removable splint for the rest of the programme of rehabilitation. A wedging plaster also provides a convenient method of treating flexion deformity while permitting general activity and ambulation. In the later treatment of arthritis of the knee a variety of appliances are available, all applicable to many other pathological and traumatic conditions—e.g., the Thomas weight-relieving calliper.

There is, however, frequently need for an apparatus which permits flexion. Such appliances would be used more often if they were readily available. Here is a particular field for standardization in the interests of mass production, cheapness, and efficiency. One type of joint is shown (Fig. 13) with a locking device to maintain the appliance rigidly in extension when the patient walks. It requires the highest form of craftsmanship as at present made. Such an appliance often has a particular disadvantage—namely, that when it is flexed to a right-angle for sitting the sharp upper angle of the lower section, upon which the lock depends, causes a sharp upward projection which is liable to tear the patient's clothes. This difficulty has been overcome in the splint illustrated by placing the joint

Fig. 12.—The Taylor spinal brace limiting flexion.

within a circular expansion of iron-work with an eccentric joint, so that the sharp upward angle disappears within the circle when the knee is completely flexed.

This appliance (Fig. 14) is made of hand-forged mild steel, the joint surfaces being case-hardened, so as to minimize the effects of friction. There is no reason why its parts should not be made in the rougher stages by the process of drop-forging from prefabricated dies. A suitable die would be expensive to a degree that few workshops would contemplate. If, however, some measure of uniformity could be decided upon throughout all the regional workshops of the country, perhaps with the aid of a large consumer such as the Ministry of Pensions, then a very important practical solution would be found.

Rest in Bed

Sending a patient to bed is the commonest form of rest used in treatment. As almost universally applied it is unphysiological. Bed treatment is only really properly understood and carried out in the care of the really ill patient. For the chronic sick it is sadly misused. The depression of the chest, the compression of the abdominal viscera, long-standing flexion of the lumbar spine, extension of the knees—these are a few of the common defects of the chronic case allowed to sit up in bed. Improvement in our beds for the chronic sick is much needed, but a good deal can be done to make it physiological by providing suitable support and by giving frequent changes to complete recumbency; by improving abdominal tonus and respiratory excursion; by the metabolic stimulus of fresh air, adequate diet, and a proper fluid intake; and by psychological care and occupation. Thus suitable conditions might be provided for the care of individual joint mechanics and for the associated medical treatment of the disease. The Hippocratic principle that the doctor should "do no harm" should be interpreted widely to include the prevention of such results of decubitus as the not uncommon development of renal calculus in orthopaedic patients. In applying physiological principles to the treatment of patients in bed we see the commonest illustration of the principle of physiological rest.

Materials

The work of appliance manufacture is at present hampered by shortage of certain materials, particularly in the grades required. A hope for the future is that recent metallurgical and engineering developments and new plastics will help to solve many of our problems. Of the metals used for splint framework, ferrous compounds are still the commonest, and wrought iron and steel are usual. Steel preponderates, and there are many different qualities for our purposes, ranging from tough malleability to rigid hardness, from mild steel to spring steel.

There seems to be some misunderstanding of the use of iron and steel in splint construction and this appears to have originated with the work of H. O. Thomas. The conditions of manufacture of his day justified his view-point. Mild steel, made possible by the Bessemer process, was not available until the end of Thomas's life. What he wanted was a metal frame with malleability and just sufficient rigidity to maintain the position but not interfere with the surgeon's ability to mould it with bending irons. Wrought iron was the only suitable material—steel as then known (1876–1891) was far too hard and rigid. This does not apply to-day, and yet there are those who advocate

FIG. 13.—The jointed knee calliper.

the use of wrought iron, strangely, for the purpose of its supposed "rigid, unyielding" nature and "not of steel, which would alter in shape under pressure" (McMurray, 1937).

Experimentally it can easily be shown that a mild steel rod, of equal calibre to one of wrought iron, has sufficient malleability and is far superior in rigidity. This may seem a small point, yet the recent war has shown that surgical appliances

FIG. 14.—Jointed calliper locking device. (a) The rough hand forgings. (b) The finished joint locked. (c) Semi-flexed, showing the locking spike disappearing within the circle. (d) In further flexion; complete disappearance of spike. Note eccentric movement which permits this and provides easing of pressure upon the calf band in flexion.

have to be made on occasion by people previously inexpert and who should be guided in the use of the most readily available materials if these are superior to those less easily obtained.

Lightness is a consideration of importance, and the use of duralumin has been advocated. Unfortunately, to provide adequate strength it must be relatively bulky, and it will not stand up to friction. Consequently in jointed appliances the moving surfaces must be faced with hard steel. The result is that appliances thus made are scarcely, if at all, lighter than those carefully forged from mild steel. Processes not yet fully explored are the use of tubular metals and electric welding.

Covering materials provide a difficulty at present owing to the scarcity of supply, particularly of leather. Here in any case improvements are needed in the use of materials which can be moulded upon curved apparatus without folds and which at the same time will be non-irritant to the skin. Perspex and polyvinyl chloride offer possibilities which are beginning to be used (MacGowan, 1943). A noteworthy example is in the work at the Royal National Orthopaedic Hospital at Stanmore reported by Scales and Herschell (1945) and Cholmeley (1946).

Summary and Conclusions

In discussing the present-day change in orientation upon the therapeutic use of rest I have attempted to define our conception of Hilton's term "physiological rest." I have emphasized Thomas's teaching that the surgeon must know how to apply rest physiologically and must also have an intimate knowledge of splint construction and management. In describing a few methods and appliances used in arthritis and nerve lesions I have tried to show some of the difficulties and to indicate the general lines for future research and practical improvement. Apart from improved design and materials, the co-operation of regional surgical workshops (with some measure of standardization) should lead to greater economy and efficiency as well as comfort for the patient. The amount of standardization need not thwart enterprise or originality.

Those of us who consider improvements in design should maintain an attitude of humility. The human form has not altered during the past hundred years. Before the modern advancement of surgical technique many great minds had devoted themselves to this subject. Study of the old masters (Bonnet in particular) reveals how far they had got. Most "new" ideas have been thought of before. Advance chiefly will come in the utilization of new manufacturing processes and new materials.

It is to be hoped that in the museum of the Royal College of Surgeons there may be established a collection, not of "museum pieces" but of well-proved basic surgical appliances with suitable illustrations of their use. Such a section could be of great teaching value.

In any discussion of orthopaedic appliance making acknowledgment should be given to the originality and skill of commercial craftsmen such as Critchley, Ernst, and many others. In addition, I would like to thank my assistants who have helped in the production of this paper; Mr. H. Salz, M.R.C.S., Miss E. L. Burrow, S.R.N., Miss M. Green, Mr. A. Reaney, M.S.R., and above all Mr. F. W. Suter (a "master-craftsman") of the Orthopaedic Workshops, the Princess Elizabeth Orthopaedic Hospital, Exeter.

REFERENCES

- Bowden, R. E. M., and Gutmann, E. (1944). *Brain*, **67**, 273.
- Cholmeley, J. A. (1946). *British Medical Journal*, **1**, 357.
- Hendry, A. M. (1945). Personal communication.
- Hilton, J. (1863). *Rest and Pain*. (2nd edition 1876.) London, reprinted 1930.
- MacGowan, T. J. B. A. (1943). *Lancet*, **1**, 805.
- McMurray, T. P. (1937). *A Practice of Orthopaedic Surgery*, 2nd ed., London, 1943.
- Med. Res. Cncl. Sp. Rep. Ser. No. 54 (1920). H.M.S.O., London, reprinted 1942.
- Scales, J. T., and Herschell, W. (1945). *British Medical Journal*, **2**, 423.
- Thomas, F. B. (1944). *J. Bone Jt. Surg.*, **26**, 602.
- Thomas, H. O. (1876). *Diseases of the Hip, Knee, and Ankle Joints*. 2nd ed., Liverpool.
- Tower, S. S. (1939). *Physiol. Rev.*, **19**, 1.

SOUTH PACIFIC BOARD OF HEALTH

A South Pacific Board of Health, with headquarters in Suva, is established under an agreement which has been signed by the Prime Minister of New Zealand and the Governor of Fiji and High Commissioner for the Western Pacific High Commission. The adoption of a long-range policy of progressive alignment of the medical and health services of Fiji and the Western Pacific High Commission and those of New Zealand and the Dominions' island territories was one of the recommendations made in the Watt-Lambie report on medical services in Fiji. The agreement gives practical effect to this policy, and is the result of negotiations which have been in train since the publication of the Watt-Lambie report early in 1944. It applies to Fiji, the British Solomon Islands Protectorate, the Gilbert and Ellice Islands Colony, Western Samoa, Cook Islands, and the Tokelau Islands. The Board, which will be advisory, will assist the participating administrations in the more effective control of disease and the promotion of health in their territories, and will co-ordinate research and the compilation of statistics on the incidence of disease, with particular reference to the maintenance of standardized quarantine procedure. The Board will also assist in the selection, posting, training, and service conditions of the medical staff.

The agreement contemplates the establishment of a medical centre in Fiji, including a new central medical school and nurses' training school. The secondment of New Zealand nurses to Fiji and other territories will be arranged by the Director, Division of Nursing, New Zealand.