MINOR TACTICS IN THE BLOOD BANK

BY

W. GILBERT MILLAR, M.B., Ch.B.

Director, Blood Bank, Victoria Infirmary, Glasgow

It is perhaps a little difficult to justify the production of yet another paper on the technique of blood withdrawal. The fact is, however, that in all practical matters slight variations in method may make a considerable difference to the comfort of the operator and to the facility, and hence to the efficiency, of the operation. It is with this in mind that the following account is given of some of the "minor tactics" employed in this bank.

The Apparatus

The apparatus used consists of the following component parts, and is shown in the accompanying figure. The blood bottle (1) is the standard M.R.C. pattern with two holes in the metal cap. The donor needle (2) is that supplied by the Scottish National Blood Transfusion Association, 35 mm. long

Components of the apparatus (with attached scale of 3 inches). 1, M.R.C. blood bottle with citrate. 2, Donor needle with stylet as point guard. 3, Same in test-tube, sterile. 4, Bottle needle (long) and air-outlet needle (short). 5, Same in test-tube, sterile. 6, Rubber connecting-tube with glass window, and needle adapters. 7, Same, sterile in its glass tube with protruding wire handle. 8, Citrate bottle, showing filter and nozzle for syringe on right, needle socket in glass dust shade on left. 9, Pilot tube. 10, Wassermann tube.

by 16 S.W.G. The "bottle" needle is longer and stouter—about 50 mm. by 14 S.W.G. These are used connected by red rubber tubing furnished at each end with a needle adapter and with a short length of glass tubing in the middle. The air-outlet needle is of 14 S.W.G. and about 35 mm. long. Additional apparatus comprises a vacuum bottle and a sodium citrate bottle.

The blood bottle contains 75 c.cm. of 3.8% sodium citrate in normal saline and is sterilized by autoclaving in the usual manner. It is then immediately capped with a square of cellophane held by a rubber band. The donor needle is lightly oiled with liquid paraffin and the stylet inserted. The wire is bent twice at 45 degrees so that the tip comes just beyond the needle point when the first bend engages with the bottom of the bevel. The wire thus acts as a point guard and allows the needle to be stored point down in an ordinary test-tube without danger of blunting (3). The tube is plugged with cotton-wool and sterilized in the hot-air oven. The "bottle" and air-outlet needles (4) are treated in exactly the same way, and are packed in the one tube (5). The connecting tube (6) is of ordinary red rubber of 5 mm. bore, not of pressure tubing, which is unnecessary and is cumbersome in use. The glass window is not essential, but it allows the progress of the bleeding to be observed and is of service in that short lengths of rubber tube can be used up. These shorter lengths are easier to clean than the full length of 18 inches or so. The connexion is contained in a length of glass tubing of about 15 mm. bore and about 42 cm. long (7), which is closed at one end by a vaccine-bottle cap of rubber and at the other by a wool plug. A small piece

of wire lightly hooked round the rubber tube at the end near the plug protrudes outside and provides a means of extracting the connexion in an aseptic manner. The tubing is autoclaved for sterilization. The apparatus is thus available for direct use on the donor in already sterilized unit packages—namely, (i) blood bottle, (ii) donor needle, (iii) connecting tube, and (iv) bottle needle and air-outlet needle.

The sodium citrate bottle (8) is used to blow citrate solution through the donor needle and rubber connexion before use. It consists of a standard blood bottle with a two-hole rubber bung. Through one hole passes a short glass tube bent twice at right-angles, to which is connected, through a cotton-wool filter, the bulb of a Higginson syringe. The other hole carries a long glass tube reaching nearly to the bottom of the bottle. It also is bent at two right-angles. Its outer limb passes through a rubber bung in a wider glass tube, which acts as a dust shade, and ends with the socket of an old Record needle attached to it by a short length of rubber tubing secured by wire. The lower end of the dust shade extends a little below the level of the needle socket and is closed while not in use by a second rubber bung attached to it by a string. This citrate bottle is sterilized in the autoclave.

The vacuum bottle provides a handy means of applying suction to accelerate the blood flow if this becomes too slow. It consists simply of a Winchester quart bottle provided with a rubber stopper through which passes a short glass tube. To this is attached a four-foot length of rubber tubing ending in a brass tube about three inches long. A needle adapter is sweated into the free end of the glass tube. A spring burette clip on the rubber tubing acts as an on-off valve, while a screw clip serves to adjust the strength of the suction. The bottle is evacuated before use by means of a Geryk air-pump, about two minutes being long enough. If a water-pump is employed about double this time is required.

The remainder of the apparatus actually used at the couch-side is standard, and includes pilot tube (9), tube for Wassermann sample (10), sphygmomanometer, hypodermic with novocain (now very rarely called for), swabs, skin antiseptic, etc. Equipment used by the person receiving the full bottle is mainly clerical and comprises the necessary record books or forms, labels, rubber bands of two sizes (one to fit the neck and the other to fit the waist of the bottle), and a supply of cellophane squares for re-capping the bottle. This cellophane should not be of the waterproofed variety, as the proofing varnish cracks and flakes on sterilizing. It has been found that if a packet of ordinary cellophane squares is autoclaved it sticks together in a solid mass, from which single sheets can be peeled as if from a sketch-block; the remainder being closely applied together, the contiguous surfaces are kept quite sterile.

Method of Use

When the donor has been prepared in the usual way, the operator receives a rubber connexion (less wire handle) from its tube and attaches it to the donor needle. The adapter at the other end of the connexion is pushed into the needle socket on the citrate bottle, and the assistant gives two or three pumps at the rubber bulb to blow citrate down the tube and needle. Meanwhile the blood bottle has been uncapped and the bottle needle and air-outlet needle thrust through the washer via the two holes in the metal cap. The donor needle is pushed into the vein, and when blood is flowing easily through the glass window-tube the adapter at the far end of the rubber connexion is pushed into the socket of the longer needle. It is worth while taking a little trouble to ensure that it is the long needle that is used, for then there is scarcely any chance of soiling the lower end of the air-outlet needle and so blocking the airway.

In most cases the blood runs as fast as is compatible with the well-being of the donor, but in the minority there may be some sluggishness, which not only extends the time of the operation but also increases the risk of clotting the needle. In our experience the commonest cause of stoppages is the obstruction of the end of the needle either by a flap of a valve or by the wall of the vein coming against it. In most cases a slight adjustment of the position of the needle is all that is necessary to re-establish the flow. In a few cases there is a strong suggestion that the vein goes into spasm. If it is decided that

suction is required, the adapter at the end of the long rubber tube from the vacuum bottle is pushed into the socket of the (shorter) air-outlet needle and the spring clip opened for a short time, the screw clip having been adjusted so that the suction is not too violent. The vacuum bottle may be left attached—a momentary opening of the clip being then necessary from time to time to remove the excess air from the blood bottle—or it may be detached if the flow is properly re-established. We have found it important to avoid the use of too much or too violent suction, since this is very apt to draw valve or vein wall against the point of the needle and thus very effectively to defeat its purpose. Moreover, too violent a suction very often causes a considerable frothing of the blood. The degree of this seems to vary a great deal from donor to donor, but any pronounced “head” on the blood is an unmitigated nuisance. Not only does it get into the air-outlet needle, blocking it or soiling the vacuum tube, but it usually does not get properly mixed with the citrate and therefore clots, forming a kind of foamy blancmange which floats on the surface of the blood. This is very likely to block the delivery set at transfusion or, alternatively, cause a good deal of staining of the plasma from haemoglobin laked from the suspended corpuscles and thus depreciate the value of the sample for plasma salvage.

While the blood is running into the bottle it is kept well mixed with the citrate by fairly brisk rotation and sideways shaking. Regular rotation is to be avoided, since it is apt to segregate the incoming blood into a column in the middle of the bottle. This column does not get properly mixed and may clot, with results similar to but rather worse than those produced by frothing.

When the required amount of blood has been drawn off, the adapter is removed from the bottle needle, the rubber tube being pinched, and the bottle with its attached needles passed on to the reception member of the team. Meantime the pilot tube and Wassermann tubes are filled, the donor needle withdrawn, and the donor tidied up. The operator hands the connecting tube to the attendant, who at once washes it through with water to remove the bulk of the blood and prevent blocking. It is then put into water to await proper cleaning. The bottle of blood has already been handed on to the responsible receiver, who removes the needles and at once re-caps it with sterile cellophane, labels it, and, when he receives them, attaches the pilot tube by means of the larger rubber band and labels the Wassermann tube. The pilot tube contains no anticoagulant, as we have found it preferable to let the blood clot. When a test is required a portion of the clot is fished out with a sterile wire loop and emulsified in saline. Not only do the cells keep better in their own serum than in citrate or oxalate but there is less liability to total loss of the sample if the tube is dropped or otherwise broken, since the semi-solid mass of clot can usually be retrieved more or less unharmed. The routine check-grouping is done on cells taken from the Wassermann tube and similarly clotted. For obtaining pieces of the clot we have found that a resistance wire (nichrome) loop of rather large size, with the wire at the gape flattened into a fairly sharp cutting edge, is a convenient implement. Platinum is too soft to make much impression on the average tough clot.

Conclusion

No particular novelty is claimed for the set-up described above. It is the outcome of a couple of years' experience, and works well under our particular conditions. There are one or two definite advantages. The packing of the needles and tubes in unit glass containers is more convenient for a centre where the apparatus is prepared and sterilized on the spot than the more usual packing of a complete set in a paper or cloth wrapping. For one thing it permits of different modes of sterilization appropriate to the different components, and, for another, it is simpler to remove the components from their tubes than it is to unwrap a whole withdrawal set. Moreover, it is more economical in supplies, a matter of no small moment in view of the frequent delays in obtaining requirements. The vacuum bottle is superior to the air-pump used

directly as a means of suction. It is cheap; it can be expanded rapidly and easily, to cope with any sudden call for the bleeding of large numbers of donors, in a way that would be quite impracticable if additional air-pumps had to be provided; it avoids the necessity for permanent or semi-permanent fittings (an important point if the bank has to change its quarters); it provides the necessary gentle suction; and, finally, there is no possibility of an accident such as was recently reported in the *Journal* (August 30, 1941, p. 311).

WELFARE OF THE EXPECTANT MOTHER

The Minister of Health (Mr. Ernest Brown), addressing the National Council for Maternity and Child Welfare, stated that as soon as possible there would be a full investigation into the neonatal mortality rate, which was too high. He had been told that while many factors contributed to the high death rate of newborn infants, prematurity was one of the most frequent. No effort should be spared to ensure that every expectant mother had access to services and facilities which would enable her to have a normal pregnancy and confinement. As a result of the evacuation plans which were made for expectant mothers and the provision for their care and confinement 150 emergency maternity homes were now in existence, with ante-natal hostels to serve most of them for the reception of expectant mothers during the last month of pregnancy. The homes had been fully utilized, and 30,000 women had been confined in them. Speaking of wartime nurseries for the care of the children of working women, he said that over 400 of these had been approved and 264 more were in preparation.

Lord Dawson of Penn, who presided, drew attention to the serious warning inherent in the trends of population to-day. If these trends persisted there was a prospect that the total population of England and Wales would have declined by three and a half million by 1965. Relatively older lives contributed increasingly and new lives decreasingly to its composition. There was a fear, indeed a prospect, that this generation would not during its lifetime produce sufficient children to replace itself.

Dr. W. Fletcher Shaw, President of the Royal College of Obstetricians and Gynaecologists, gave a heartening account of the progress which had been made, more particularly in child welfare, since the first decade of this century, especially the reduction in infant mortality, due in large part to the elimination of infantile diarrhoea. On the ante-natal side, however, he did not think that as much progress had been achieved. There was a disposition to take maternal mortality as the one criterion of maternity work, but for every woman who lost her life in childbirth there were many who dated ill-health and disabilities to the same process. The incidence of eclampsia, said Dr. Shaw, had fallen materially, and this must be due, he thought, to better ante-natal supervision. Fistulae and other obstetric injuries were fewer, and this also in part must be ascribed to the same cause, though another reason was the better training of medical students. The aim must be to get every pregnant woman under observation; there were still far too many whose difficulties were recognized only when they were in labour. The greatest confession of bad treatment in obstetrics was the case of “failed forceps.” He concluded with a plea for regionalization of services, whereby it would be possible to establish hospitals with trained staffs where they could best be used, without regard to arbitrary demarcations of small areas.

A. M. Snell (*Minnesota Med.*, 1941, 24, 210), who maintains that pancreatic lithiasis is not nearly so rare as was formerly believed, states that 18 cases occurred at the Mayo Clinic during the last three years. X-ray evidence of stone was found in all; in 6 the diagnosis was confirmed by operation and in 2 by necropsy. The principal symptoms are colicky pain, motor disturbances of the gastro-intestinal function, steatorrhoea, and complications such as tuberculosis, endocarditis, and pyelitis. Treatment is exclusively surgical.