ELECTRICAL IGNITION OF ANAESTHETICS

BY

H. W. SWANN

His Majesty's Senior Electrical Inspector of Factories

Two papers on electrical ignition of inflammable anaesthetics were recently read at the Institution of Electrical Engineers and, with the discussion which followed, showed that useful research work has been done towards reducing the risk. The purpose of this article is to summarize the more recent contributions to the subject and to suggest their application in practice.

Whilst explosions attended by serious results are fortunately rare the Ministry of Health in 1935 thought it desirable to issue a memorandum entitled "Precautions Against Anaesthetic Explosions in Operating Theatres."* This was of a preliminary character, and may be supplemented now that further information is available. The Ministry is aware of a number of accidents, information on which has been gathered from various sources. Although this figure may bear no relation to the actual number of cases, it indicates where the chief risk may lie.

TABLE I

Cause of Ignition (Electricity)	Non-Fatal	Fatal	Total
Static electricity	3	4	7
Surgical lamp in mouth	2	4	6
Cauteries	1	3	4
Surgical diathermy	2	1	3
Hot nozzle, dental syringe		1	1
Table lamp (switch)	1		1
X-ray apparatus	1		1
Totals	10	13	23

Table I summarizes the information available, and it will be seen that in recent years twenty-three cases of explosion are known to have occurred in this country and abroad which are attributed to electrical ignition in some form or other.

The Chief Risks

It will be noted that static electricity and the use of surgical lamps in the mouth constitute the chief risk, and although some degree of danger must inevitably attend the use of the general electrical equipment for lighting, sterilizing, ventilating, etc., attention has focused on the first two causations, especially as Coste and Chaplin† have found by experiment that the room distribution of ether vapour during operations under open-mask anaesthesia resulted in low concentrations. The air at about two inches from the patient's mouth contained less ether than would form an explosive mixture, and the proportion fell off rapidly as the distance from the patient increased.

Another bearing on the subject may be obtained from a statement by Dr. H. W. Featherstone‡ to the effect that from time to time he had heard of a minor "pop" when a small lighted electric bulb had been inserted into the mouth during ether-air anaesthesia. Dr. Featherstone described a case in which an explosion occurred in the patient's mouth and was attributed to defective contact between the lamp and the holder which caused a spark. He also discussed the possibilities of the leakage of ether into the gas stream even when the anaesthetist, in endeavouring to guard against risk, thought he had shut it off.

* Memo. 191, Medical, Ministry of Health, 1935; S.O. No. 32-9999.

† J. H. Coste, F.I.C., F.Inst.P., formerly chemist-in-chief, London County Council, and C. A. Chaplin, Ph.D., B.Sc., A.I.C., of the chemical branch of the Public Health Department, London County Council.

‡ *Proceedings of the Royal Society of Medicine*, vol. 25, Part I.

These and other opinions indicated that attention should be focused, in the first instance at least, on the electrical ignition risks in the immediate vicinity of the patient, and of these risks, those attending nose and throat operations appeared to be the most important. It must not be assumed, however, that risk in other positions can be dismissed as negligible, especially where ventilation is bad. Dr. C. F. Hadfield* particularly stresses the danger of foot switches in relation to accumulations of ether vapour at floor level.

Intrinsic Safety Surgical Lamps

There is a wide variation in the design, construction, and maintenance of the different forms of surgical lamps in use, and with the limitations necessarily imposed on size, and the need of sterilization, it becomes difficult to suggest forms of construction which could be considered gas-tight or immune from electrical defect. While there is little doubt that from the point of view of ignition risk design could be improved, it was thought that there was more hope in the direction of intrinsically safe circuits, which, if practicable, could be applied with less difficulty and delay. This principle has many successful industrial applications, but it was realized that there was no experience with mixtures possessing the inflammable characteristics of ether-oxygen. The problem has, however, been tackled by Professor W. M. Thornton†, whose researches are detailed in a paper read on February 24, 1938, at the Institution of Electrical Engineers. Thornton found that with hot wires, such as cauteries, ignition can occur within limits as wide as 3 to 80 per cent. of ether in oxygen. With sparks the limits are in general much narrower. The researches described determine these limits and also the least currents that cause ignition when broken slowly or suddenly or by fusing wires. Second only to hot wires as a possible risk are the fizzling sparks which may occur between fine wire ends, such as at an intermittent break in a fine wire flexible. Thornton found that when the current is kept below certain well-defined values ignition did not take place, and thus he was able to specify the resistance for any given voltage which would limit the current in that circuit to safe values. This resistance is, of course, located at the source of energy—for example, battery or transformer—and is independent of lamp or other circuit resistances. Its effect is to limit the energy, which appears as a spark caused by an open or short circuit anywhere in that circuit. Thornton advocated a factor of safety of 2 as being an ample margin: Table II takes account of this, and is amplified by a graph (Fig. 1).

TABLE II

Direct Current		Alternating Current	
Volts	Minimum Safe Limiting Resistance	Volts	Minimum Safe Limiting Resistance
0-2	0.25 ohms	0-2	1.25 ohms
2-4	1.25 "	2-4	6.00 "
4-6	3.25 "	4-6	14.00 "
6-8	7.00 "	6-8	25.00 "
8-12	27.00 "	8-13.5	54.00 "

The curves relate to fizzling sparks, which over most of the range of the experiments were the more dangerous. It is worthy of note, however, that the curve for plain break A.C. would cross the fizzling break at about 12.1 volts, which means that the plain break may become the more dangerous at voltages above this figure. Similarly, the curve for plain break D.C. would cross the fizzling break curve at 3 volts, and thus again plain break may be

* Discussion on Professor Thornton's and Dr. Rayner's papers, February 24, 1938; the Institution of Electrical Engineers.

† "The Electrical Ignition of Mixtures of Ether Vapour, Air, and Oxygen," by Professor W. M. Thornton, O.B.E., D.Sc., D.Eng., Past-President.

the more dangerous below that figure. Only the fizzling curves are shown, and the scale figures on the right-hand ordinate are those found experimentally by Thornton, while those on the left side allow a factor of safety of 2.

FIG. 1.—Curves for Fizzling Sparks.

Hot-wire Ignition and Spark Ignition

Thornton distinguished between hot-wire ignition and spark ignition, and the importance of the point has been brought out by Dr. H. L. Saunders, Ph.D., F.I.C., who found that although circuits incorporating the prescribed value of resistance may be safe as regards most forms of sparks, a special case arises when a circuit is completed through a single strand of fine wire which may get hot. Under certain conditions the temperature of the ether mixture may thus be raised above its ignition point, causing explosion, or sufficiently to predispose to spark ignition should the fine wire break. Intermittent contact in a partly broken flexible conductor, involving a combination of hot-wire heating and sparking, is undoubtedly the most insidious form of danger, as a much smaller spark energy may be enough to cause an explosion. This is a particularly difficult case to define accurately, and may not be fully covered by Thornton's formulae. Very finely stranded flexible is thus likely to be more dangerous than coarse strands, as the heating will be greater in the last strand to break just before severance due to wear and tear.

Saunders has suggested that the only safeguard against this dangerous combination is to test the resistance of the circuit regularly. A partial break in a flexible will show up as an increase above the usual resistance and indicate the need for examination and repair. Lamp resistances vary, and it would be best to replace lamps by a low value of constant resistance for purposes of the test advocated.

The "Cool Flame"

Another aspect of the subject, with an important bearing on the risk of ignition, has been pointed out by Dr. D. T. A. Townend* in connexion with that peculiar type of combustion usually referred to as a "cool flame." In the case of ether-air mixtures in a 2.5-cm. tube there

is a range of inflammable mixtures which at atmospheric pressure is capable of normal flame explosion between about 3 and 9 per cent. of ether in air, and a quite distinct range for the cool flame between about 19 and 34 per cent. of ether in air. Within the normal flame range the mixture can be ignited by sparks, but to initiate a cool flame it is necessary to use a hot wire—a distinction of great importance. Townend places the ignition temperature range of cool flame between *circa* 170° and 270° C., and describes cool-flame combustion as very incomplete, and invisible, except in darkness. The flames travel slowly, with a speed of something like 20 cm. per second in a tube 3.5 cm. in diameter; and at about atmospheric pressure, which is that approximately customary in anaesthesia, the ranges overlap, as the normal flame range starts at about 250° C.

Much the same type of system is formed with ether-oxygen-enriched mixtures as with ether-air mixtures, though with the former the ranges are naturally much wider. Townend showed that a cool flame initiated by hot wire could travel some distance in a tube, leaving behind it incompletely burnt products capable of normal flame explosion and ignitable from the preceding cool flame. Thus, where the velocity of the supply of anaesthetic is low, a cool flame initiated at the patient's mask during anaesthesia may travel back along the rubber tube to the gas-producing equipment, and may cause a normal explosion, either by ignition as above or on encountering a more explosive mixture.

Townend also discussed the possibility that although sparks of *normal* intensity, such as those from an induction coil or a broken wire, do not initiate cool flame, there may be a range of *weak* spark intensities capable of doing so, and the point has an obvious bearing on explosions caused by static sparks which may attract the attention of an interested investigator.

Overrunning Surgical Lamp Bulbs

Surgical lamps are often used in conjunction with some form of regulating resistance which allows of overrunning. The little bulbs are as a rule rated at between 2.5 to 3.5 volts and, when overrun to the point of burning out, the glass may reach temperatures capable of igniting ether mixtures. A valuable contribution to this subject has been made by Mr. C. T. A. Harper, who has measured bulb temperatures of typical lamps when overrun to an extent causing collapse and burning out of filaments. He employed a special fine thermocouple, which was kept in intimate contact with the bulb, thus avoiding the use of material which might act as lagging and so produce a higher temperature than would normally be obtained. Capped lamps did not reach such high temperatures as uncapped bulbs, possibly owing to the increased area of dissipation, and it is worthy of note that in no case did the collapse of the filament result in bursting or damage to the bulb.

Harper recorded his measurements on graphs, of which Table III is a summary. The graphs are reproduced as Figs. 2, 3, 4, and 5.

TABLE III

Type of Lamp	Number Tested	Rated Voltage	Bulb Temperature At Rated Voltage (Average)	At Burn-out of Filament	
				Lowest Value Noted	Highest Value Noted
Surgical uncapped	4	2.5	120° C.	305° C.	380° C.
Pharyngoscope capped	5	3.5	60° C.	158° C.	230° C.
Urethroscope capped	4	3.5	80° C.	158° C.	250° C.
Ward's flexible capped	4	3.5	70° C.	218° C.	280° C.

In most cases the filaments appear to burn out at just under double the rated voltage.

* Discussion on Professor Thornton's paper, February 24, 1938; the Institution of Electrical Engineers.

The importance of this work is seen when it is remembered that the ignition temperature range for normal ether-flame explosions lies between 250° and 350° C. and for a cool flame between 170° and 270° C. It seems likely that some explosions in anaesthesia may have been due to over-running, as it is believed that burning out of bulbs from this cause is not uncommon. The limiting resistance specified by Thornton for intrinsically safe circuits tends to diminish the risk of overrunning, but it will be obvious that it is in addition necessary that the rated voltage of the lamps should not be materially less than the voltage of the battery or transformer constituting the source of supply. An intrinsically safe 6-volt circuit would still allow a 2.5-volt lamp to be dangerously overrun.

CONCLUSIONS

The following conclusions may now be drawn as to the conditions under which surgical lamps may be used with

FIG. 2.—Relation between applied voltage and bulb temperature. Surgical lamp; 2.5 volts, uncapped.

FIG. 3.—Relation between applied voltage and bulb temperature. Pharyngoscope lamp; 3.5 volts, capped.

comparative safety in the presence of inflammable ether-oxygen mixtures:

1. The adoption of Thornton's formulae for intrinsically safe circuits will prevent any form of spark ignition on either A.C. or D.C. circuits.

2. The danger of hot-spot or hot-wire ignition, or a combination of this with sparking at the moment when the last strand of a flexible breaks, may be minimized by a regular test of the circuit resistance and the keeping of records.

3. The supply voltage should not exceed the rated voltage of capped lamps by more than 25 per cent. Uncapped lamps should not be overrun.

Experiments made to demonstrate the utility of circuits conforming to conclusions 1 and 3 above showed that a useful degree of light was obtainable, especially on D.C. circuits. At the higher values of A.C. voltage the inductive effect has led to the necessity of high values of

resistance, which diminish the light considerably. The utility of the battery-fed circuit is therefore the greater.

Static Electricity

Although the climate of this country can as a rule be depended upon to produce a degree of humidity which is incompatible with the retention of static charges, the modern practice of supplying large buildings with cleaned and conditioned air tends to neutralize the climate, and Dr. Rayner has presented a paper,* read on February 24, 1938, at the Institution of Electrical Engineers, which is devoted to a study of the manner in which static charges can be built up in operating theatres. Explosions attributed to static have occurred, and it has been found that rubber-tired vehicles used for carrying patients can be electrified by withdrawing a warm dry blanket. Bulgin† says that an ordinary standard tyre of specific resistance 10^8 to 10^9 ohms will generate enough static to raise the potential of any vehicle to about 5,000 volts at a speed of ten miles an hour on a smooth road during warm weather. If sparks are produced in these or other ways they may jump not only to an earthed object or person but also to

FIG. 4.—Relation between applied voltage and bulb temperature. Urethroscope lamp; 3.5 volts, capped.

FIG. 5.—Relation between applied voltage and bulb temperature. Lamp for Ward's Flexible; 3.5 volts, capped.

a rubber-tired vehicle carrying an anaesthetic equipment. Thus the most likely place for static appears to be in the immediate vicinity of the patient, where the risk of ether concentration is greatest.

The risk with rubber-tired vehicles is easily overcome by the use of light trailing chains, but there is a good deal of misconception as to the conditions under which these are likely to be successful. Chains are usually effective on most floor surfaces, excepting rubber and linoleum, and quite high surface resistances, of the order of 100,000 ohms, are permissible. The point is easily settled in individual cases by producing electrification *without* a trailing

* "The Risk of Explosion due to Electrification in Operating Theatres of Hospitals," by Dr. E. H. Rayner, M.A., D.Sc. Dr. Rayner's work was done at the National Physical Laboratory for the Department of Scientific and Industrial Research, in consultation with the Anaesthetics Committee jointly appointed by the Medical Research Council and the Royal Society of Medicine.

† Discussion on Dr. Rayner's paper, February 24, 1938; Institution of Electrical Engineers.

chain, and then attempting to repeat it after the chain has been attached. In the absence of a suitable instrument a charge can be detected by holding a finger-tip about one-eighth of an inch away from the metal-work of the vehicle while a dry blanket is smartly pulled away.

A form of rubber possessing conductive properties has recently been developed with special reference to the risk of static in operating theatres, and has several similar industrial applications. It can be used in the shape of tyres, or in sheets for flooring, but low values of resistance are both unnecessary and apt to be dangerous. Rubber is inherently inflammable, and if the resistance is low there is risk of fire should electrical leakage occur through it, as, for instance, from a defective vacuum cleaner or a defective pedestal-type lighting standard. At 230 volts a specific resistance of the order of 100,000 to 150,000 ohms should be free of fire risk and yet quite effective for dispersal of static.

Electrical practice has always associated *non*-conducting properties with any form of rubber, and if accidents are to be avoided it is imperative that the new conductive variety should be plainly marked in a permanent manner. It is known that the India Rubber Manufacturers' Association has both resistance standardization and marking under consideration.

Dr. Rayner also gave attention to the possibility of static being produced at the anaesthetic equipment by discharge of gas from high-pressure storage cylinders. Some earlier work* on the electrification of oxygen cylinders has been done at Birmingham Central Technical College, from which it was concluded that in the absence of foreign matter in the gas stream there appears to be no electrification. The point is no doubt relevant also to cylinders containing carbon dioxide and nitrous oxide, and as a high degree of purity is ensured in manufacturing and bottling, the risk is small and may be minimized still further by maintaining electrical conductivity between all gas cylinders and other metal work comprising anaesthetic equipment supplemented by earthing, either direct or through conductive rubber. This measure will prevent sparking between components or to other bodies, but does not deal with the rare case of combustion at the oxygen valve. The design of the latter has, however, recently been improved by the elimination of ebonite for purposes of valve seating. The usual precautions of opening valves gently and keeping them free from oil or grease are, of course, desirable.

Gas cylinders may be located at a central point, away from the operating theatres, the gas supply being distributed by permanent pipe-line. This affords an opportunity for reducing the cylinder pressure in two stages, that in the pipe-line being maintained at, say, 60 lb. per square inch. This arrangement would reduce risk, and has already been adopted in America and in a few cases in this country, not only for the distribution of oxygen but also for nitrous oxide and other gases.

In cases of exceptional dryness the use of rubber shoes may insulate the person sufficiently to allow static charges being set up by friction with the floor and of being carried. Rayner found that a similar condition may arise with leather shoes on cork carpet. This would mean the possibility of sparking from the finger-tips, a phenomenon almost unknown in this country, but well recognized in very dry climates. An effective safeguard lies in the use of conductive rubber for the shoes.

CONCLUSIONS

The following conclusions may now be drawn as to the means of dispersing static charges:

1. Rubber-tyred vehicles should be earthed by trailing chains or conductive rubber.
2. Electrical conductivity should be maintained between gas cylinders and other metal components of anaesthetic equipment, and the equipment should be earthed direct

* G. F. Partridge, B.Sc., M.I.E.E., in a report dated April 1, 1935, to the British Oxygen Co. Ltd.

or through conductive rubber. Gas cylinders may with advantage be situated at a central point outside the theatres.

3. In exceptional cases of dryness rubber shoes should be conductive.

X-Ray Apparatus and Electrical Equipment

The information recorded in Table I shows that although static and the use of surgical lamps constitute the chief risks, there are other matters which may probably be regarded as of secondary importance, such as the diagnostic use of x-ray apparatus during anaesthesia and the general electrical equipment of the theatre. Surgical diathermy or the use of cauteries in the presence of an inflammable anaesthetic would of course be a mistake, but modern forms of x-ray equipment probably introduce little risk except as regards defects, such as in flexible conductors, during use. The general electrical equipment—pendants, switches, plugs, sterilizers, etc.—is not as a rule in the immediate vicinity of the patient, but the degree of risk depends entirely on this factor and the fortunate lack of coincidence between an electrical defect and the presence of a mixture rich enough to ignite. With the exception of mercury-type switches there is little that can be suggested in connexion with the general equipment which would not involve revolutionary changes, and thus the greatest importance attaches to efficient ventilation.

Collaboration

The inquiries made in connexion with anaesthetic explosions, and the result of the work described in this paper, lead to the conclusion that some central representative body is required which could influence the design of electrical apparatus for the operating theatre, or perhaps for the hospital as a whole, and advise on its use. The employment of electro-medical apparatus is increasing as the applications of electricity become more numerous, and such a body might, without retarding progress, foresee risks and combine the views and requirements of surgeons and engineers to the best advantage.

STUDENTSHIPS AND FELLOWSHIPS IN CLINICAL SCIENCE AND EXPERIMENTAL PATHOLOGY

The Medical Research Council invite applications for a third series of studentships and fellowships intended to encourage young British medical graduates of special ability and original mind towards becoming investigators in those branches of medical science which are concerned directly with disease as it occurs in human beings. This field of research includes investigations into disease or manifestations of disease in patients, together with experimental work of an immediately relevant kind.

Postgraduate Studentships.—Postgraduate studentships are offered for medical graduates who have already held house appointments and are strongly inclined to a career in clinical science or experimental pathology. Each student chosen will receive a stipend at the rate of £200 per annum, during a period not exceeding twelve months, for personal maintenance while undertaking approved courses of study in this country such as may be regarded as best calculated to advance his training in methods of research. This study may include modern languages and such advanced physiological, pathological, and special clinical work under recognized teachers as may be regarded as forming a suitable preliminary to serious research work to be undertaken later; but the approved course may not include studies of which the purpose is to enable the student to pass further examinations.

Research Fellowships.—The Council also offers research fellowships for candidates of similar qualifications who have already had some experience in the use of research methods. Each fellowship will be tenable for one year at the ordinary value of £250 per annum, and will be renewable in approved instances at the rate of £300 per annum for a second year. These fellowships are intended as probationary appointments for research in clinical science or experimental pathology under suitable direction in this country. Research expenses may be provided in addition to stipend. Applications should be lodged with the Secretary, Medical Research Council, 38, Old Queen Street, London, S.W.1, before October 1, 1938. Forms for the purpose will be supplied on request.