among the fæces and brought to me. It was an inch and a half in length, the head being rather large and the point very sharp.

It is scarcely probable that I should have communicated this case had it not been that I desired to append to it an observation which might, I thought, possess a certain amount of practical value. I did not, although urgently requested to do so, administer to my little patient *purgative* or any other kind of medicine.

ON A PORTABLE SPIROMETER.*

By W. P. BAIN, M.D.

IN 1846 was read, at the Royal Medical and Chirurgical Society of London, a paper "On the Capacity of the Lungs and on the Respiratory Functions, with a view of Establishing a Precise and Easy Method of Detecting Disease by the Spirometer", by John Hutchinson, Surgeon.

This was one of the most brilliant papers ever read before the Society. The original views, the great research, the numerous observations, and the discoveries of the relations between respiration and the height and weight of an individual, and to health and disease, stamped the author at once as no ordinary man. An instrument which he invented—the spirometer—was one of his chief aids in these investigations, and he thoroughly established its value in the diagnosis of diseases of the chest.

His physiological views were very extensively adopted, and they are now standard in all works on physiology. But how have we benefited by these discoveries? In the treatises on medicine at this day, they seem to be entirely ignored. In Hoblyn's *Dictionary of Medical Terms*, the word "spirometer" is not to be found; neither is it mentioned by Sir Thomas Watson. Dr. Barclay's excellent work on *Medical Diagnosis* contains not a word on the subject. In one part of his book he says, "All true diagnosis is ultimately based upon inductions separately framed out of clinical and pathological investigations and experiments." And yet, while treating of phthisis, he passes over in total silence one of the most simple and accurate means of diagnosing this disease at its commencement, and ventures to say that "in its early stages accurate diagnosis is perhaps impossible with the aid of physical signs"—one of which physical signs, and perhaps the most important at this period, he entirely overlooks.

Dr. Guy, in speaking of respiration, treats very fully of the spirometer; and yet, when he treats on the diagnosis of phthisis, not a word is spoken of its great practical value. I believe, indeed, that the profession in general does not sufficiently prize the indications supplied by its use, for it is certain that the invasion of phthisis may be detected in this way long before it can be by percussion and auscultation. On this point Dr. Hutchinson, speaking of one of his cases—the American giant Freeman—states that, when he examined him during his "fighting condition", he found that his vital capacity was 434 cubic inches. Two years afterwards he came to town in ill health. "I then examined him", says Dr. Hutchinson, "twenty times at various intervals, during which his vital capacity varied from 390 to 340 cubic inches, and the mean of the observations was 344—a decrease of 90, or more than 20 per cent. At this time, I took him to two physicians well skilled in auscultation, and they both affirmed that they could not detect any organic disease." He became rapidly worse, and confirmed phthisis ended in death a year afterwards.

But it is not only in affording evidence in phthisis that the spirometer is valuable; it is also of the highest service in cases of a converse nature. What can be more satisfactory both to the patient and to the physician than to be assured thoroughly that in supposed cases of consumption there really exists no such disease? If, besides a favourable report of a patient's health through percussion and auscultation, the spirometer indicates that his vital capacity equals or exceeds the natural mean, a great additional testimony is afforded of his health, and an immense relief is given to his mind, while the physician feels doubly sure of his diagnosis.

It is possible that defects in the construction of the spirometers in use may have led to their not being employed so often as is evidently necessary. The bulk, the weight, the consequent want of portability, and general complex arrangements of stopcocks and valves, may have led to these results. In Dr. Hutchinson's, for example, the patient has to be educated, as it were, in its use. He has to be taught to turn the stopcock at the precise moment when he begins to expire, and to turn

it off again at the precise moment when his expiration ceases: a certain time is wasted in teaching this.

A simple and portable spirometer is much to be desired. As almost every one knows, Dr. Hutchinson's spirometer is on the same principle as an ordinary gasometer—one cylinder inverted in another which is partially filled with water. The smaller cylinder rises in proportion to the air which is blown into it, and the amount is marked by an index and a scale on the side.

Another instrument, sold by Mr. Coxeter of London, is of a more portable description. It consists of two air-tight bags joined by a stopcock, with another at each end. In using it, the intermediate stopcock is closed, and the patient breathes into the first bag. The stopcock at the mouth-piece is then closed, the intermediate one opened, and the air passed into the second bag, which is graduated so as to measure all the air that may be expired. This instrument, though apparently simple, has the disadvantage of three stopcocks and the repeated measurements.

Some time ago, it struck me that an instrument upon the bellows or accordion principle, having a graduated scale at the side, would fulfil all requirements. I found, however, that in practice it did not work well. One side would rise more than another, and the indications were therefore inexact. I then thought that, by fixing one end so as to form a kind of hinge, and allowing the other to be free, and making the body long and narrow, the rise would be more steady and uniform, and the scale have a longer sweep, so that it might be graduated with clearness. I have therefore had the instrument made in this manner—measuring twelve inches in length, six in breadth, and about two in depth: a graduated scale in brass, or a tap, is so placed as to indicate the amount of cubic inches expired. The weight of the top and bellows is counterbalanced by a spiral spring within the instrument, and no stopcocks are required to complicate the arrangements. It can be carried in the hand or under the arm like a book, and is, in every sense, simple and portable.

In experiments which I lately made on artificial respiration, I occasionally used one of these instruments, with satisfactory results. I found that it was very sensible, and easier in every way to manage, than other means—being portable, not requiring water, and the nuisance of valves and stopcocks being done away with.

In connection with this subject, I venture to lay before you a few practical remarks on the different modes that have been used for artificial respiration, in cases of asphyxia from drowning, poisons, gases, strangulation, etc., with a system of my own, which I believe to be simpler and more effective than any other.

The system of Marshall Hall was examined some years ago, by a committee of the Royal Medical and Chirurgical Society, and found not to realise the expectations that had been formed of it. The plan of Dr. Silvester was also tried by the same committee, and it was found that a considerably larger quantity of air could be introduced by it than by any other hitherto known, and it was accordingly recommended and adopted by the Royal Humane Society. It is so well known, that I need only say that its principle consists in causing the thorax to expand, by imitating the natural act of inspiration, by traction of the inspiratory muscles in raising the arms of the patient above his head.

Professor Mattei, of Paris, in 1867, brought before the Academy of Medicine of Paris, a method of inducing respiration in the newlyborn infant, during suspended animation. This consists in raising the infant by its shoulders and giving it some shakes, or "secousses", as he terms it. He did not, however, explain the rationale of the process, nor did he propose to extend it to the treatment of adults.

Another plan was proposed by the celebrated Pacini, of Florence, who advises that the operator should stand at the head of the patient, who is lying on a table. The head is then placed against his abdomen, the fingers passed behind the shoulders into the axillæ, and the upper portion of the trunk moved upwards and forwards, or towards the head of the operator. This plan is well thought of by its inventor; but, when tried in this country, both by myself and by others, has not given so good a result as Silvester's, whilst its performance is most fatiguing to the operator.

The method which I have discovered is founded upon that of Silvester; but whereas he adopts what I call a roundabout mode of acting on the chest, through the arms, I act immediately on it, by pulling the shoulders up at once, by the hands placed in front in the axillæ. It is three times more simple than Silvester's, as one movement only is required, instead of three, and the operator has not to bear the weight of the patient's arms, which, as he has to stoop over the body at the same time, involves immense fatigue, when his efforts are long continued.

* Read in the Physiological Section at the Annual Meeting of the British Medical Association at Leeds, July 1869.