to Baudelocque's and Scanzoni's, and greater than in most others. By this increased length the base of the skull is more surely reached, and therefore more surely broken up. The hold of the instrument is more secure, and the head less likely to slip from between the blades, and when the instrument is introduced the lock is external to the vulva, and nipping of the labia is therefore avoided. In considering the length of the blades it is to be remembered that in the cases of extreme narrowing the head is always quite above the brim of the pelvis, and the depth of the pelvis itself is often much increased. When the blades are closed, their extreme width is one inch and a-half, and the cephalic curve is formed by a curve commencing at an inch and a-half from the points. The space between the blades at the widest part when closed is one inch. The gradual cephalic curve, and the closeness of the blades, will allow of the instruments being applied in a narrower space than most others. It is true that they diminish the extracting powers of the instrument, but this I look upon as an advantage, as, for the safe use of the cephalotribe, the head should be crushed to such a degree by repeated applications that it may be extracted without force. The lock in my instrument is the reverse of that in general use. It is made so, that when the upper, or right, blade is introduced, the patient lying in the ordinary obstetric position, on her left side, the lock of that blade looks forwards. I have spoken at some length on this formation of the lock in my papers in the *Dublin Quarterly Journal of Medical Science*. I believe no change could be made in the cephalotribe, or in the midwifery forceps, that would tend more to facilitate their correct application and use than this. Formerly forceps were made in this way. I have a pair showing this, made, I believe, in 1790. In the museum of the Rotunda Hospital there is a similar pair. Dr. Radford, in his volume of *Essays* published in 1836, has devoted a section to this form of lock, but still instrument makers have not adopted it. Messrs. Fannin, of this city, however, who made the cephalotribe I exhibit, now manufacture forceps with the same form of lock, and I am greatly mistaken if any operator who has once used either instrument will ever willingly operate on a case with the head high up in the pelvis with one having the lock differently placed. The handle of my instrument is half-an-inch shorter than Simpson's or Hicks', the shortest hitherto in use. This, added to the increased length of the blades, tends, no doubt, to diminish its compressing power; but I have ascertained, by repeated experiments, that there is still abundant power to compress the most firmly ossified head, and the instrument gains somewhat in lightness and portability by the change. The screw used is the same as Simpson's, which is not only most simple and easy of application, but forms a very convenient and firm cross-handle for manipulating the instrument.

ON THE CEPHALOTRIBE.*

By SIR JAMES Y. SIMPSON, BART., M.D.,
Professor of Midwifery in the University of Edinburgh.

SIR J. Y. SIMPSON said he felt, and he was sure many of his brethren felt, that to bring any information concerning midwifery to Dublin was like bringing coals to Newcastle. The medical men of Dublin had reason to be proud of the Dublin Lying-in-Hospital at the Rotunda. He believed it was the first institution of the kind in the three kingdoms. He would address to the Section some observations in illustration of the cephalotribe.

Under certain circumstances in practice, obstetricians were often obliged to reduce the size of the child's head, and for that purpose they had recourse to a variety of measures, craniotomy being one. At present that was not a very favourable operation in hospital practice. He (Sir James Simpson), having looked at Dr. Churchill's book, found that there were 310 cases of craniotomy performed in the Dublin Hospital. Out of these 65 died, or more than one-fifth. The operation was attended with great danger for various reasons; and especially for this reason, that when we perforate the child's head and take away a piece of bone, we are extremely apt to injure the vagina; and, again, by the operation, the size of the child's head was often increased. Dr. Hamilton, many years ago, was in attendance on a young woman whose pelvis was very deformed, and, in order to deliver her, he had to remove the whole arch of the cranium; and he states that after having done it, which took him eight hours, he had to be rolled up in blankets and carried home himself; so what must have been the state of the patient when the doctor was so bad? Notwithstanding all that, they had the base of the skull quite entire; and in putting on the cephalotribe it broke the head at once. In reference to the size of the pelvis, M. Pajot stated that he delivered women with a pelvis of an inch and a-half. That,

in Sir James Simpson's opinion, was too small a pelvis to be delivered at all, and it was one where the woman ought to submit at once to the Cæsarean operation. He himself had attended a case where a very deformed woman was attempted to be delivered by craniotomy. He used that instrument (exhibiting a cephalotribe), and the delivery was over in two or three minutes. The instrument used was quite curved, and the advantage of it was, that it would not slip when a straight pair would. One of the great difficulties is this, that after fixing the head it slips backwards. The instrument needs to be pushed very far forward, and with a long straight forceps it is almost impossible to hold the head. The instrument exhibited was a very powerful one and very heavy. The cephalotribe, with teeth for fixing the scalp of the child, he regarded as a great improvement; but it is a little more complicated in its arrangement than others. It was an instrument capable of crushing any head, and rendered the operation of craniotomy easy and safe for the patient.

THE CEPHALOTRIBE.*

By J. BRAXTON HICKS, M.D., F.R.S.,

One of the Lecturers on Midwifery, and Assistant-Physician-Accoucheur at Guy's Hospital; Physician to the Royal Maternity Charity; Examiner in Midwifery at the University of London; etc.

In exhibiting the modification of the cephalotribe already described by me in the eighth volume of the *Transactions* of the Obstetrical Society of London, I may remark that it is very nearly similar to that of Sir J. Simpson as regards the blades, both as to size and form, the difference being chiefly in the mode of applying the power by the screw. I may mention that I have found it, in my own practice, the most convenient arrangement. I have employed it in four cases, and am led to the conclusion that, as an addition to the obstetric bag, it is a very valuable one.

From numerous experiments on heads of still-born children, I have arrived at the following conclusions, which I will state as concisely as possible.

1. If an endeavour be made to crush the foetal head before perforation, a very powerful instrument will be required. Besides which, from the shape of the head, there is a constant tendency in it to glide from beneath its influence. Thus a serious addition is made to danger of the operation.

2. If perforation be undertaken before the use of the cephalotribe, the vault of the cranium forms very little resistance; and then an instrument like that of Sir J. Simpson will be sufficient for crushing the head; and there is little tendency to its gliding away under these circumstances.

3. When the instrument is applied to the perforated head, it may act beneficially in two ways—(a) by completely crushing the base, thereby flattening sideways the whole head, or doubling up the base; or (b) by the slipping of one of the blades inwards a little, whereby the base is tilted edgewise, and the skull flattened by the pressing inwards on to the base of the squamous portion of temporal bone and the parietal.

4. Under either of these processes, the head can be so flattened as to allow the blades to meet.

5. Generally speaking, two applications will suffice to effect cephalotripsy.

6. The flattening is much more commonly in the transverse diameter of the foetal head, than in the longitudinal; which, however, is seldom elongated by the lateral pressure.

7. The thorax and pelvis of the foetus form no appreciable resistance to the instrument.

From the above results, I have been led to advise that it should be a general rule, that perforation should precede cephalotripsy; for, where it is impossible to perforate, it will be exceedingly difficult to apply the cephalotribe. The case would, in that case, fall under treatment by Cæsarean section. When it is said "as a general rule," it is admitted that a few instances may occur where the crushing may be tried first.

From these experiments, and from consideration of the cases in which I have used the instrument, I regard it rather as a substitute, in the severe forms of obstruction, for the other craniotomy instruments after perforation; and not so much as a substitute for Cæsarean section. It is, no doubt, very valuable both in doing away with the danger arising from spicula of bone, and from the crotchet, etc., as well as tending to shorten the duration of the operation; but, if we rely upon it as an instrument capable of accomplishing the whole delivery by itself, we shall be overrating it, and exposing ourselves to disappointment. For it is not a tractor; and therefore, should pains be absent, we shall generally have to draw down the head by some other means. When it is regarded as a substitute for the Cæsarean section, it seems that the advan-

* Read before the Midwifery Section at the Annual Meeting of the British Medical Association in Dublin, August 1867.

* Read before the Midwifery Section at the Annual Meeting of the British Medical Association in Dublin, August 1867.