purged, and on the day following an attack of acute diarrhoea; fifthly, the existence of paralysis of the bladder, which had come on without any apparent cause so far as the genito-urinary apparatus was concerned, and which has never been recorded as a concomitant of ordinary intestinal occlusion; sixthly, the neurotic temperament of the patient—I distinctly expressed the opinion that the intestinal occlusion was due to nervous paralysis, and I urgently recommended the immediate application of the constant electric current. This advice was at once endorsed by the excellent practitioner in attendance, Dr. D'Auria, but the other medical men maintained that this measure was useless, and that it was necessary to perform laparotomy without delay. Dr. Vizioli, Professor of Electro-Therapeutics in the University of Naples, was called in. The constant current which was employed was furnished by a Daniell's battery with Onimus's piles, modified as regards the graduation and the number of piles by Professor Vizioli. The intensity was measured by a milliampère galvanometer, the strength employed being 10 milliampères at every application. The positive pole was, by means of a rectal catheter, carried 20 to 25 centimetres up the bowel; and the negative pole, which was olivary in form and covered with cloth steeped in a saturated solution of chloride of sodium, was rubbed transversely over the surface of the abdomen, especially in the parts corresponding to the cæcum, the ascending, transverse, and descending colon, and the sigmoid flexure, as well as over the hypogastric region. The duration of each application was from eight to ten minutes, and there were three sittings every day. By the end of the first day the retention ceased, the patient was able to pass water freely, his general condition improved, especially subjectively, and the attacks of pain were less violent, but the bowels were still constipated.

The surprising result of the treatment shown in the cure of the retention confirmed me in my own opinion of the nature of the case; and notwithstanding pressure constantly brought to bear in the opposite sense by the other doctors, who terrified the family with pictures of imaginary dangers because they were determined at all hazards to have laparotomy performed, the electrical applications were continued, and, after the ninth sitting, the patient had spontaneous motions of the bowels, and by degrees he completely recovered.

A CONTRIBUTION TO THE PHYSICS OF THE ELECTRIC BATH.

By W. S. HEDLEY, M.D.,
Brigade-Surgeon (Retired).

I HAVE lately tried by direct experiment to throw light on some of the physical and physiological problems involved in the administration of the electric bath. Of course, in entering upon an investigation of this kind, our first solid standpoint must be measurement. In the case before us this almost resolves itself into the question, "What proportion of the main current passes through the human body when immersed under given conditions in an electric (dipolar) bath? This is a primary question but a very complex one. It covers a great deal of ground, and in glancing over the literature of the subject, there appears a tendency to "talk round" this point rather than to approach it by actual experiment.

For the present purpose the following are selected from a large number of experiments bearing on the point. Apparatus used: Oak bath 6 ft. long 2 ft. 6 in. wide at widest part, the waste pipe being insulated from earth by a short length of rubber pipe inserted near the bath. Battery of 74 Leclanché cells giving E.M.F. of 75 volts, dipolar bath one electrode (30 × 20 sq. cm.) resting at each end; plain tap water 12½ in. in bath before immersion of subject; temperature 98° F.; R. of bath water before entry of subject 165 ohms¹; subject lying in water, having head resting on strap out of water, and body from shoulders downwards completely immersed, shoulders 8 in. from + electrode, feet 3 in. from — electrode.

Experiment 1.—A current of 100 ma. passed through the

bath. Subject distinctly felt the current as a tingling sensation in his legs but not in other parts of the body. An intercepting current now arranged having one electrode (336 sq. cm.) on nape of neck, the other (95 sq. cm.) on ball of great toe and behind it, this portion of the foot being raised out of the water for the purpose. These electrodes were connected with reflecting galvanometer, having R. of 976 ohms. Deflection right off scale but estimated at 500 microampères.

Experiment 2.—Intercepting current arranged with a plain handle electrode of convenient size held in subject's mouth and making good contact with mucous membrane; the other electrode (22.5 sq. cm.) watertight on calf; Edelmann's suspension galvanometer (R. 200 ohms); other conditions as in last experiment. Deflection 0.75 ma. = (750 micamp.); constant whilst it lasted for two minutes.

Experiment 3.—Whilst the circuits were still complete the watertight covering of calf electrode was intentionally ruptured, so that the bath water came in contact with the electrode. Deflection 2.5 ma. = 2,500 micamp.

These experiments seem to afford some direct evidence of the proportion of C. passing through the patient's body in dipolar bath. The readings were carefully taken, twice verified, and found constant after C. had been passing for two minutes. They were further confirmed by an observation with alternating current and telephone. I then proceeded to test the point in other ways. After a simple experiment merely to emphasise (under conditions identical with those of the other experiments) the well-known fact that the chief lines of force do run through the bath water by the shortest and most direct route when not interfered with; the body of the subject was immersed, lying in the direct line of C. flow, as in the ordinary way of administering an electric bath, and the results were a striking contrast. Thus, without the body in the water 44 per cent. of total C. was intercepted in the direct line of C. flow, and only 40 per cent. at the sides. The presence of the body in the water caused this condition of things to be reversed, 30 per cent. only being intercepted in the direct line and 48 per cent. at the sides. This seems to point to the fact that the effect of the presence of the patient's body in the water is to deflect to a large extent the lines of force, so that the greater quantity of C. passes round him and not through him.

The next step was to divide the hydro-electric field into sections constituting shunt circuits (arranged as in following diagram), which could be separately tested without the body and with it.

Results shown in following table:—

Condition of Hydro-electric Field.	Current in Main.	Shunt.	Current in Shunt.	Condition of other Shunts.
A Without Body.	53 ma.	a a'	38 ma.	Closed through R. of 200 ohms each.
	"	"	39 "	Open.
	"	{ c c' }	2.5 each	Closed.
	"	{ b b' }	3.4 "	Open.
B With body (lying in line of col. a a').	53 ma.	c c'	32 ma.	Open.
	"	"	25 "	Closed through 200 ohms.
	"	b b'	25 "	" " "

The R. of patient's body, on emerging from the bath, after 5 minutes' immersion—one electrode (8 × 12 inches) on shoulder and standing on the other (taken by Wheatstone bridge), was 8,000 ohms, R. of shunt was 200 ohms—the C. through the shunt 39 ma. May we then apply the law of divided circuits as follows:—

$$X = \frac{200 \times 39}{8000} = 0.97$$

Therefore 0.97 ma. has passed through the patient's body.

A further consideration of this experiment shows that, as shunt a a' gave 38 ma., and the two side shunts gave 2.5 each, 38 + (2.5 × 2) = 43; and as the total main C. was 53, this leaves 10 as loss by diffusion. When the body was immersed the side shunts gave 25 ma. each; thus the utmost the body could

¹ R. of bath water, taken as it cooled (by Wheatstone bridge), at 92° F. 194 ohms, at 87° F. 231 ohms, at 70° F. 440 ohms.

receive was 3 ma., and in the face of the direct evidence of loss by diffusion we may fairly conclude that the C. received by the subject did not amount to this.

One main electrode (metal 8x12 inches) at each end of bath. *a a'*, Plain metal electrode, 8x12 inches; *b b'*, plain metal electrode, 8x12 b' 4½x12; *c c'*, plain metal electrode, c 8x12 c' 4½x12.

With reference to the experiment with the water-tight electrode—if we may assume that (owing to the absence of skin R.) the mouth electrode was inside the body and its area equal to at least that of one of the main electrodes (namely, 600 sq. cm.), and if we may further assume (as seems pretty well established) that the C. through the body is directly proportional to the contact area—it follows that, as 22.5+600=622.5 (area of shunt electrodes), and 600+600=1200 (area of main electrodes), and as C. passed through the shunt=0.75—

$$X = \frac{1200 \times 0.75}{622.5} = 1.4$$

But this was obtained with 100 ma. in main circuit, so we must divide it by 2 to compare it with the above results:

$$\frac{1.4}{2} = 0.7.$$

Thus we have two calculations that give the C. obtained by the patient as less than 1 ma., and the divergence of the current from the centre to the sides of the bath confirm this, though we cannot quite say what should be allowed for diffusion. (The question of polarisation must be the subject of future experiments.) Without affirming that the above figures nearly cover the amount of current received by the patient, I think these experiments at least justify the conclusion that the waste of current in administering a dipolar bath is in excess of what is generally supposed, and we need have no hesitation in saying that the calculation sometimes made that the proportion of main [C. the patient receives is $\frac{1}{2}$ is too high. I do not, however, for a moment use this as an argument against the use of the electric bath in proper hands. The dipolar bath, with its painless—or at worst “pleasantly painful”—evenly-applied, and widely-distributed C., must ever be an admirable method of general electrification. That it is wasteful is not a serious drawback with plenty of battery power behind us—or still less so if we are utilising electric light circuits for electro-therapeutic work.

The question of density, which is so important a factor in “dosage,” and which in ordinary electrical applications depends upon the size of the electrodes, becomes a very complicated one in the dipolar bath. Here it is evident that not only the size of the electrode is to be considered but the amount of diffusion the C. undergoes in passing through the water from the electrode to the body; and this will depend of course partly on the size of the electrode, partly on distance, and partly on the conductivity (that is the specific R.) of the water. In other words we have not only to consider the size and position of the electrodes electrifying the water but we have to look upon the whole extent of water in contact with the body as a huge electrode carrying a widely diffused current with a density of course diminished in proportion to its diffusion.

In all electro-therapeutic work it is absolutely essential to aim at clear ideas about density. In Landois and Stirling's *Textbook of Physiology*, after impressing on the mind of the reader the necessity for distinguishing between “strength” of C. and “intensity,” the writers seem to apply the term “intensity” to what it is usual to speak of as “density.” “If the size of the transverse sections of the circuit varies the electricity must be of the greater ‘intensity’ at the narrower

parts.”² This is “density,” and the fact is generally stated by saying that the density is inversely proportional to the transverse section of the conductor or $D = \frac{C}{S}$.

“The current is comparable to a girl's hair, which may be gathered up into a narrow tress or allowed to flow loosely without changing the number of its constituent parts.”³ In a recent admirable article on the Electric Bath the writer states that, “It can easily be shown that part of the C. traverses the human body when immersed in a bath, even if the feet do not touch the bottom electrode, for if one of the legs be held out of the water the current strength, as registered by the galvanometer, is reduced.”⁴ I have succeeded in obtaining the same experimental results. On the same page an experiment of Beard and Rockwell is referred to with apparent approval, bearing on the relative conductivity of the body and water. I give the quotation: “That the body conducts better than the water is proved by an experiment we have often made. Place both hands at some distance apart in a bath through which a current of considerable strength is running and a sensation will be distinctly felt in them. Bring the hands still immersed very close to each other and the sensation will be very much diminished. When the hands are far apart a considerable proportion of the current passes through the body from one hand to the other. It prefers this much longer and roundabout road to the direct path through the water.” No one doubts the experiment. It can easily be verified by the galvanometer. But I think it does not warrant the above conclusion. Does it not point to the fact that the experimenter in approximating his hands was simply removing them from the region of greater difference of potential and greater current density to that of less?

THE ELECTRIC LIGHT AND ITS EFFECTS UPON THE EYES.

By GUSTAVUS HARTRIDGE, F.R.C.S.,
Surgeon to the Royal Westminster Ophthalmic Hospital, etc.

THERE are two chief forms of electric light in use at the present time—the arc light, which is chiefly employed for outdoor illumination, for large stations, workshops, etc.; and the incandescent electric light employed for indoor illumination, railway carriages, and shops. The arc light is produced by passing an electric current between two carbon points which are separated by a slight interval, the most intense light being produced when only a short space intervenes between the two points; particles of carbon at a white heat pass over from the positive to the negative terminal, a concavity being produced in the carbon connected with the positive pole, while a deposit takes place on the carbon point to which the negative pole is attached. This passage of particles of carbon at a white heat from one pole to the other adds no doubt considerably to the illumination produced. The light is of a bluish dazzling brilliancy, which renders it unsuitable for illuminating purposes except when placed at a considerable distance from the eyes; even then the eyes should not be exposed to its direct influence. Another great disadvantage of the arc light is that it is apt to be intermittent and unsteady.

The incandescent electric light is obtained by passing the electric current through a non-conducting medium *in vacuo*. At first a platinum wire was used, but this has given way almost entirely to a carbon filament made from cocoa nut fibre. When electricity is passed through this non-conducting material the friction produced is sufficient to raise the filament itself to a white heat. If this takes place in a vacuum no combustion of the carbon occurs. The light thus produced is very suitable for general illuminating purposes;

² Landois and Stirling's *Textbook of Physiology*, p. 669, vol. ii.

³ *Electro-therapeutics*, Erb, p. 55.

⁴ *Lancet*, March 28th, 1891, p. 710.

NOTE.—“Intensity” is a much misapplied term. It is even sometimes used to express electromotive force. It is not “density,” however, and it is not “tension.” But it is well to remember that in France it is used as synonymous with our term “current strength,” thus: $I = \frac{E}{R}$.