administration of large quantities of saline under pressure or in a pulsatile fashion, which we feel appropriately for our location, resembles a bagpipe. A saline bag (akin to the reservoir bag of the bagpipe) connected to a Yankauer suction tube (the chanter) is held in the axilla between the lateral chest wall and the medial upper arm. This allows the flow rate to be controlled or pulsed by pressure exerted by abduction of the arm and the stream directed using the tube. The entire system has the advantage of being held with just one arm, leaving the other hand free. We feel this is a simple, elegant and easily reproducible method, using readily available materials, which can be used in trauma situations 'in the field'.

Reference

1. McKenzie J, Cashman JP. Wound irrigation: simple delivery by the litre. *Ann R Coll Surg Engl* 2004; **86**: 53.

A simple technique for accurate placement of patch angioplasty apex sutures

RSM DAVIES

Department of Vascular Surgery, University Hospital Birmingham NHS Foundation Trust, Birmingham, UK

CORRESPONDENCE TO

Robert SM Davies, Flat 4, 5 Meadow Road, Edgbaston, Birmingham B17 8DH, UK. E: RSMDavies@hotmail.com

Figure 1 Schematic representation of the described technique during right carotid endarterectomy. Abbreviations: CCA, common carotid artery; ECA, external carotid artery; ICA, internal carotid artery; STA, superior thyroid artery.

Haematoma formation following carotid endarterectomy is a common cause of postoperative hoarseness and often results from inexact placement of proximal angle arterial wall sutures during patch application. A simple method that permits accurate placement of each suture at the right distance and correct bite is for the assistant to apply tension to the suture in the opposite direction to its path of travel following passage through the patch rather than the artery, as is traditional (Fig. 1). This opens up the anastomosis allowing placement of the subsequent arterial wall suture under direct vision whilst maintaining tension on the anastomosis line.

DIY laparoscopic kit

DA RAPTIS, K MOUZAKI, DM GORE

Department of Surgery, University College London Hospitals, London, UK

CORRESPONDENCE TO

DM Gore, Department of Surgery, University College London Hospitals, Charles Bell House, 67–73 Riding House Street, London WC1E 6AU, UK

With the on-going reduction of junior doctors' hours, surgical trainees find it ever more difficult to establish their laparoscopic

Figure 1 The BioStatIT DIY laparoscopic kit.

skills. Basic surgical skills' courses have, typically, only 1 h dedicated to laparoscopic practice on simulators. Commercial laparoscopic kits are prohibitively costly. We suggest that this BioStatIT DIY Laparoscopic KitCCC (BDLKCCC; Fig. 1) is ideal for home laparoscopic training. By introducing trainees to two-dimensional spatial learning, important laparoscopic skills such as dissection, cutting, suturing and knot-tying can be practised. It is simple to assemble, safe, and has a relatively low cost (£25 for a night-vision camera and £2 for a plastic box). The plastic box is turned facing downwards and a night-vision camera is connected on the inner surface of the bucket. Camera cables can then be connected to a home television, PC or laptop. Trainees can seek decommissioned laparoscopic instruments from their operating department or surgical equipment representatives. A dedicated website (<www.DIYLaparoscopy.com>) has been developed for further free information and guidance to its use.

The world's longest instrument – an aid to acetabular cup extraction in revision hip surgery

MANSOOR KASSIM, SUNIL KUMAR PAI, MARK ASHWORTH
Department of Trauma and Orthopaedics, Torbay District
General Hospital, Torquay, UK

CORRESPONDENCE TO

Mark Ashworth, Consultant Orthopaedic Surgeon, Level 5,
Torbay District General Hospital, Lawes Bridge, Torquay TQ2
7AA, UK
E: mark.ashworth@nhs.net

Figure 1 Cork-screw, slap hammer and a universal femoral component extractor assembled together in series.

Component removal is a time-consuming, yet necessary, step during revision hip surgeries. Because of the diversity of the components and the methods used to secure them, an equal diversity of approaches and tools are necessary for component removal. Many instruments are described for removing the acetabular component of a total hip replacement. Cement-component debonding is an important step in acetabular component removal. Revision instruments,¹ chisels, ultrasound tools and commercially available purpose built devices, such as the 'Explant' acetabular cup removal system (Implex/Zimmer Inc.), achieve this to a greater or lesser degree. We report a modification of a technique that the senior author was instructed in at the Nuffield Orthopaedic Centre in Oxford, UK. Direct disimpaction of the cup is desirable to avoid leverage forces acting on, and potentially damaging, the acetabular bone stock when removing the cup. A cork-screw slap hammer device is ideal for this. Whilst these are available as combined or one-piece instruments, we describe a technique using instruments which should be readily available in any revision theatre, namely a cork-screw, a slap hammer and a universal femoral component

extractor, assembled together in series as shown in Figure 1. The connection between the two extraction components (the cork-screw and the slap hammer) is rigid, rather than a loop of steel wire as was used in the original assembly used in Oxford (where this device was colloquially known in theatre as 'the world's longest instrument'). The more solid arrangement described here is superior to its predecessor in several respects: our device employs a better coupling mechanism (the universal femoral extractor) which is less likely to break during use, directly imparts the disimpacting force (achieved by upwards blows of the slap hammer to the cork-screw engaged in the cup) allowing greater control in cup removal and, unlike single-purpose commercially available devices, has a versatility which allows it be used for a number of different potential purposes.

Reference

1. Paprosky WG, Weeden SH, Bowling Jr JW. Component removal in revision total hip arthroplasty. *Clin Orthop* 2001; **393**: 181–93.