Video Article

CD4⁺ T-Lymphocyte Capture Using a Disposable Microfluidic Chip for HIV

Sang Jun Moon¹, Richard Lin², Utkan Demirci¹

¹Bio-Acoustic-MEMS Laboratory in Medicine (BAMM), HST-Center for Bioengineering, Brigham and Women's Hospital

²Massachusetts Institute of Technology

Correspondence to: Richard Lin at rllin@mit.edu

URL: <http://www.jove.com/index/Details.stp?ID=315>

DOI: 10.3791/315

Citation: Moon S.J., Lin R., Demirci U. (2007). CD4⁺ T-Lymphocyte Capture Using a Disposable Microfluidic Chip for HIV. JoVE. 8. <http://www.jove.com/index/Details.stp?ID=315>, doi: 10.3791/315

Abstract

Protocol

1. Prepare microfluidic channel with anti-CD4⁺ biotinylated surface.
2. Inject 10 ml of sample blood at 5 ml/min into the microfluidic channel with syringe pump.
3. Flush out the blood at 20 ml/min for 20 min with PBS.
4. Inject diluted FACS lysing solution (diluted to 10% from stock solution with PBS) at 5 μ l/min for 15 min.
5. Incubate for 15 min.
6. Inject primary CD3⁺ antibody solution (diluted to 10% from stock solution with PBS) at 5 ml/min for 20 min.
7. Incubate for 2 hours.
8. Flush out channel with PBS at 5 ml/min for 15 min.
9. Inject 20% AF488 CD4⁺ antibody, 0.2% DAPI stock solution, and 10% secondary CD3⁺ antibody at 5 ml/min for 20 min at the same time.
10. Incubate for 20 min.
11. Flush out the channel with PBS at 5 ml/min for 15 min.
12. Count cells under fluorescent microscope:
 1. White (optical) - Black and white: cell and dust
 2. UV(DAPI) - Blue: cell only
 3. Blue(CD4⁺) - Green: T-lymphocyte and monocyte
 4. Green(CD3⁺) - Red: T-lymphocyte only

References