Diagnosis of Human Immunodeficiency Virus Type 1 Infection in Infants by Use of Dried Blood Spots and an Ultrasensitive p24 Antigen Assay[▽]

Ada Cachafeiro,¹ Gayle G. Sherman,² Annette H. Sohn,³
Consuelo Beck-Sague,⁴ and Susan A. Fiscus^{1*}

Department of Microbiology & Immunology, University of North Carolina at Chapel Hill, Chapel Hill, North Carolina¹; Department of Molecular Medicine and Haematology, University of Witwatersrand, Johannesburg, South Africa²; Division of Pediatric Infectious Diseases, University of California, San Francisco, San Francisco, California³; and Clínica de Familia MIR, La Romana, Dominican Republic⁴

Received 22 June 2008/Returned for modification 27 September 2008/Accepted 3 December 2008

We tested 617 dried blood spots (DBS) from human immunodeficiency virus-exposed infants from five countries using an ultrasensitive p24 antigen assay (Up24). The sensitivity was 94.4% (67/71) and the specificity was 100% (431/431) for infants with DBS specimens ≤ 20 months old; DBS older than 30 months demonstrated only 72.2% sensitivity (39/54) ($P < 0.001$) but displayed 100% specificity (61/61).

Several factors reduce diagnosis of human immunodeficiency virus (HIV) in infancy in resource-limited settings. Simple, rapid antibody assays cannot be used because of the presence of maternal antibodies. As a consequence, viral nucleic acids or proteins must be detected. Nucleic acid assays are technologically complex and subject to contamination and require expensive equipment and highly trained technologists, limiting availability to selected centers of excellence in resource-limited countries (1, 8, 12, 15, 28, 30).

The ultrasensitive p24 antigen assay (Up24), an alternative that overcomes the limitations of antibody assays while using an enzyme-linked immunosorbent assay (ELISA)-based platform, reduces the need for expensive equipment and personnel. It has been used successfully with plasma for over 12 years for diagnosis of HIV infection in infants (9, 11, 13, 18, 19, 26, 29, 32).

Another factor limiting diagnosis of HIV infection in infants is the need for phlebotomy to collect specimens and cold chain maintenance while the specimen is transported to a central laboratory. These have been addressed by adaptation of nucleic acid (2–7, 10, 16, 20, 21, 27, 31), and now Up24 antigen assays, for use with dried blood spots (DBS) (14, 22, 23).

Other investigators have adapted the Up24 assay for use with DBS. Table 1 compares these different extraction methods used for p24 antigen assays with DBS. Li et al. (17), following the manufacturer's instructions for plasma specimens, reported 0% specificity using DBS. Knuchel et al. (14) improved the sensitivity by modifying the elution buffer (25) and increased specificity to 100% by adding a quenching step using hydrogen peroxide after the first wash. Patton et al. (22, 23) improved elution by using the previously described buffer (25) and 0.5% Triton X-100. The reported sensitivity and specificity were 98.3 to 98.8% and 100%, respectively. Our objec-

tives were to develop and optimize a simpler method for HIV infant diagnosis using DBS and reagents found in commercially available Up24 test kits as much as possible and to assess the optimized assay performance.

(This work was presented in part at the CDC Conference on HIV Diagnostics, Atlanta, GA, 5 to 7 December 2007, and the 15th Conference on Retroviruses and Opportunistic Infections, Boston, MA, 2 to 6 February 2008.)

DBS obtained from patients in the Dominican Republic ($n = 60$), Malawi ($n = 9$), South Africa ($n = 210$), United States ($n = 264$), and Vietnam ($n = 74$) were collected on Whatman 903 paper (Whatman, Sanford, ME), air dried, stored with a desiccant in individual ziplock bags at room temperature (Dominican Republic, South Africa, and Malawi) or at 4 to 8°C (Vietnam) or at -20°C (United States), and shipped to the laboratory at the University of North Carolina, Chapel Hill (UNC). The infants had been previously tested for HIV type 1 (HIV-1) either by DNA PCR using cell pellets from whole blood using the Roche Amplicor assay, version 1.5 (Branchburg, NJ) for specimens from the United States and South Africa or an in-house HIV DNA assay (Pasteur Institute, Ho Chi Minh City, Vietnam) for specimens from Vietnam. Specimens from Malawi and the Dominican Republic were originally tested as DBS using an HIV RNA assay (NuccliSens QT; bioMérieux, Durham, NC). These results were considered the gold standard against which the DBS Up24 antigen results were compared. All infected infants had follow-up specimens to confirm HIV infection.

Most U.S. infants were receiving 6 weeks of zidovudine prophylaxis at the time of sampling. All Malawian infants had received a single dose of nevirapine and 1 week of zidovudine and lamivudine. Some Vietnamese and South African children were beyond infancy and known to be HIV infected. In addition, DBS from South African infants had been tested in a previous Up24 assay study (23). DBS ≥ 30 months old were evaluated separately.

Controls were prepared from non-HIV-1-infected whole blood samples. Negative controls consisted of 50- μl spots of whole blood on Whatman 903 paper. Positive-control DBS

* Corresponding author. Mailing address: Department of Microbiology & Immunology, University of North Carolina at Chapel Hill, Chapel Hill, NC 27599-7290. Phone: (919) 966-6872. Fax: (919) 966-9873. E-mail: fiscussa@med.unc.edu.

[▽] Published ahead of print on 10 December 2008.

TABLE 1. Comparison of different extraction methods used for p24 antigen assays with DBS^a

Assay details	Li et al. (17)	Patton et al. (22, 23)	Knuchel et al. (14)	This study
Steps in procedure	1. Elute 50 μ l DBS using 0.5% Triton X-100 2. Incubate 60 min 3. Centrifuge 13,600 $\times$ g for 3 min 4. Transfer to a screw-cap cryovial and boil for 5 min at 100°C 5. Proceed with the standard Up24 antigen assay protocol	1. Incubate duplicate 6-mm punches with 25 μ l SNCR buffer^b for 10 min at room temp 2. Add 275 μ l of 0.5% Triton X-100 3. Incubate overnight at 4°C 4. Boil for 5 min at 100°C 5. Proceed with the standard Up24 antigen assay protocol	1. Combine 16-mm DBS plus 0.8 ml of elution buffer (1 part SNCR buffer , 1 part water, and 9 parts of 0.5% Triton X-100) 2. Incubate on a shaker/rocker 8 h at room temp 3. Boil for 5 min at 100°C 4. Proceed with the Up24 antigen assay protocol using overnight incubation 5. After the first wash, quench with 0.3% H₂O₂ in 50 mM Tris-HCl for 10 min at 37°C 6. Continue with the standard Up24 antigen assay protocol	1. Incubate duplicate 6-mm punches with 50 μ l specimen preparation buffer^c for 10 min at room temp 2. Add 0.25 ml of 9 parts of 0.5% Triton X-100 diluted in PBS to 1 part of Tris reagent 3. Shake for 2 h at room temp or overnight at 4°C 4. Boil for 5 min at 100°C 5. Proceed with the standard Up24 antigen assay protocol
Assay times	~8 h	4 to 5 h on day 1, overnight incubation, 4 h on day 2	~9 h on day 1, overnight incubation, 3 h on day 2	~1 h on day 1, overnight incubation, 4 to 5 h on day 2
Sensitivity (%)	96 if viral load >1,000 copies/ml	98.3–98.8	84	94.4
Specificity (%)	0	100	100	100

^a Reagents in boldface type were not included in the manufacturer's kit.

^b **SNCR buffer** consists of 30 mM Tris-HCl (pH 7.2), 450 mM NaCl, 1.5% Triton X-100, 1.5% deoxycholic acid, 0.3% sodium dodecyl sulfate, and 10 mM EDTA.

^c **Specimen preparation buffer** consists of kit specimen diluent, 10 mM EDTA, and 30 mM Tris-HCl (pH 7.2).

were prepared by adding the PerkinElmer HIV-1 ELISA positive control to normal whole blood to a titer of 100 pg/ml. The control DBS were air dried and stored in ziplock bags at -20°C with a desiccant sachet and a humidity indicator card (Static Control Components, Inc., Sanford, NC).

Two 6-mm disks were punched from the DBS using an ordinary paper puncher cleaned between samples and placed in 1.5-ml screw-cap Sarsted tubes (Newton, NC). The disks were incubated for 10 min at room temperature with 50 μ l of a specimen preparation buffer consisting of the PerkinElmer HIV-1 core profile ELISA standard/sample diluent in the ELAST amplification kit NEP116VL (PerkinElmer Life Sciences, Boston, MA) with the addition of 10 mM EDTA and 30 mM Tris-HCl, pH 7.2. The elution buffer consisted of 9 parts of 0.5% Triton X-100 diluted 1:10 in PBS and 1 part (vol/vol) of Tris reagent. The PerkinElmer kit includes both Tris-HCl and Triton X-100. After the addition of 0.25 ml of the elution buffer, the DBS disks were rocked overnight at 4°C and removed from the sample using wooden sticks. The tubes were transferred to a heat block, boiled for 5 min at 100°C , and allowed to cool for 10 min. The Up24 protocol in the ELAST amplification kit was subsequently followed except that instead of using the QuantiKin software to read the assay, an endpoint reading at two wavelengths of 490 and 650 nm was obtained after 30 min of substrate incubation. Four methods of calculating a cutoff were explored. The cutoff selected (mean optical

densities of two negative controls plus 0.05) provided the best correlation for the data. Samples with optical densities greater than that number were considered positive. All specimens were tested once. Repeat tests were performed on several specimens; the result was always concordant with the first. UNC testing personnel were blinded as to the infant's infection status.

Of 617 DBS from HIV-exposed infants assessed, 324 were presumed to be subtype B and came from the United States and Dominican Republic; 219 from Malawi and South Africa were presumed to be subtype C, and 74 from Vietnam were presumed to be subtype A/E. DBS that were ≤ 20 months old were evaluated with the optimized assay and compared to their respective gold standard (Table 2). Among the 431 true negative specimens, there were no false-positive results (100% specificity). The sensitivity of the assay increased from 50% within the first week of life to 80% for infants 1 to 6 weeks old to $\geq 95\%$ for infants and children > 6 weeks old. Both false-negative results in the youngest age group were from Malawian infants receiving zidovudine-lamivudine-nevirapine for prevention of mother-to-child transmission. The false-negative specimen in the 1- to 6-week age group was from a 3-week-old U.S. infant receiving zidovudine prophylaxis. The false-negative result for the 6- to 26-week age group was from a 14-week-old infant from the Dominican Republic who was not receiving prophylaxis. Patton et al. (22, 23) never tested neonates in the

TABLE 2. Sensitivity and specificity of the Up24 antigen assay with DBS by age of the infant and age of DBS

Age of infant	DBS <20 mo old				DBS >30 mo old			
	Sensitivity (%) ^a	Specificity (%) ^b	PPV (%) ^c	NPV (%) ^d	Sensitivity (%)	Specificity (%)	PPV (%)	NPV (%)
<7 days	50 (2/4)	100 (80/80)	100	97.6	Not tested	Not tested	Not tested	Not tested
1–6 wks	80 (4/5)	100 (95/95)	100	98.9	100 (3/3)	100 (10/10)	100	100
>6–26 wks	95 (21/22)	100 (192/192)	100	99.5	87.5 (7/8)	100 (51/51)	100	98.1
>26 wks–72 wks	100 (24/24)	100 (63/63)	100	100	72.7 (8/11)	Not tested		
>72 wks	100 (16/16)	100 (1/1)	100	100	65.6 (21/32)	Not tested		
Total	94.4 (67/71)	100 (431/431)	100	99.1	72.2 (39/54)	100 (61/61)	100	80.3

^a The values in parentheses are the number of specimens with a positive result/total number of specimens tested.

^b The values in parentheses are the number of specimens with a negative result/total number of specimens tested.

^c PPV, positive predictive value.

^d NPV, negative predictive value.

first week of life (the youngest was 20 days old). If the age group within the first week of life is excluded, the sensitivity is 97% (65/67).

Among 115 DBS from South Africa that had been tested previously and then stored at room temperature for ≥ 30 months (23), there was a trend toward decreasing sensitivity as the age of the infant increased, although it was not significant ($P = 0.12$); the sensitivity was 100% (3/3) for DBS from infants >1 and <6 weeks old, while it was 66% (21/32) for infants of ≥ 72 weeks of age (Table 2). The specificity was 100% across all age groups. Some older infants and children may have been on antiretroviral treatment at the time of collection, but no treatment data were collected.

Our aim was to simplify the procedure and maximize the use of the reagents already in the manufacturer's kit. The final optimized procedure used a specimen preparation buffer that contained only two reagents not found in the kit—10 mM EDTA and 30 mM Tris-HCl. We have now tested this procedure on 617 DBS from HIV-exposed infected and uninfected infants and children. The assay demonstrated 94.4% sensitivity for diagnosis of HIV infection with presumed HIV-1 subtypes B, C, and AE ($n = 502$) using DBS ≤ 20 months old versus 72.2% in older specimens ($P < 0.001$), achieving 100% specificity without complicating the procedure with a quenching step.

This is the largest series of infants <7 days of age tested using the Up24 assay using DBS. Although the sensitivity was only 50% (2/4), the specificity was 100% (80/80). Others have suggested that p24 assays might be less specific in very young infants (24). However, we previously tested plasma samples with the Up24 assay and observed similar specificity in infants 0 to 7 days of age (1 of 109 [99%]) (11). The optimal time to use the Up24 antigen assay for infant diagnosis requires further investigation, especially as prevention of mother-to-child transmission programs in low-resource countries move to more-complex and longer postdelivery antiretroviral regimens.

The lengths of storage time and temperatures varied, depending on the original collection site (country), but did not appear to affect assay sensitivity when specimens were tested within 20 months of collection, in contrast to data suggesting that DBS should be tested within 6 weeks of collection if Whatman no. 1 filter paper was used (22). When Whatman 903 was employed (as done here), little loss of sensitivity was observed at 12 weeks. In most cases, samples are tested within a

few weeks of collection. DBS from infants 6 to 26 weeks of age demonstrated 94.5% sensitivity and 100% specificity, regardless of the length of room temperature storage (up to 20 months); 72% of DBS stored more than 30 months at room temperature still had detectable antigen using the optimized Up24 assay.

Our results confirm that infant HIV infection can be accurately diagnosed using DBS and an optimized Up24 antigen assay, with lower cost and less complexity.

S. A. Fiscus and A. Cachafeiro were supported by a grant from the National Institute of Allergy and Infectious Diseases for the International Maternal Pediatric Adolescent AIDS Clinical Trials Network (U01AI068632) and the UNC Center for AIDS Research (P30AI50410). A. H. Sohn was supported by a grant from the Eunice Kennedy Shriver National Institute of Child Health and Human Development (K23 HD047166). G. G. Sherman received support from the President's Emergency Plan for AIDS Relief (PEPFAR) through the U.S. Centers for Disease Control and Prevention for South Africa, and from UNICEF. PerkinElmer Life Sciences provided some of the p24 antigen kits used in the study.

We are grateful to the CDC-funded Breast Feeding and Nutrition (BAN) study for allowing us to use excess DBS previously tested in this study (Prevention Research Centers Special Interest Project SIP 13-01 U48-CCU409660-09 and SIP 26-04 U48-DP000059-01, U.S. Centers for Disease Control and Prevention). We thank Janet Patton for sharing her expertise while we optimized the assay and Takesha McMillion, Paul Alabanza, and Maria Schanz for preparing DBS for the North Carolina infants.

REFERENCES

- Benjamin, D. K., Jr., W. C. Miller, S. A. Fiscus, D. K. Benjamin, M. Morse, M. Valentine, and R. E. McKinney, Jr. 2001. Rational testing of the HIV-1-exposed infant. *Pediatrics* 108:E3.
- Biggar, R. J., W. Miley, P. Miotti, T. E. Taha, A. Butcher, J. Spadaro, and D. Waters. 1997. Blood collection on filter paper: a practical approach to sample collection for studies of perinatal HIV-1 transmission. *J. Acquir. Immune Defic. Syndr.* 14:368–373.
- Cassol, S., T. Salas, M. Arella, P. Neumann, M. T. Schechter, and M. O'Shaughnessy. 1991. Use of dried blood spot specimens in the detection of human immunodeficiency virus type 1 by the polymerase chain reaction. *J. Clin. Microbiol.* 29:667–671.
- Cassol, S., A. Butcher, S. Kinard, J. Spadaro, T. Sy, N. Lapointe, S. Read, P. Gomez, M. Fauvel, C. Major, and M. O'Shaughnessy. 1994. Rapid screening for early detection of mother-to-child transmission of human immunodeficiency virus type 1. *J. Clin. Microbiol.* 32:2641–2645.
- Cassol, S. A., N. Lapointe, T. Salas, C. Hankins, M. Arella, M. Fauvel, G. Delage, M. Boucher, J. Samson, J. Charest, et al. 1992. Diagnosis of vertical HIV-1 transmission using the polymerase chain reaction and dried blood spot specimens. *J. Acquir. Immune Defic. Syndr.* 5:113–119.
- Comeau, A. M., J. Harris, K. McIntosh, B. J. Weiblen, R. Hoff, and G. F. Grady. 1992. Polymerase chain reaction in detecting HIV-1 infection among seropositive infants: relation to clinical status and age and to results of other assays. *J. Acquir. Immune Defic. Syndr.* 5:271–278.

7. Comeau, A. M., H. W. Hsu, M. Schwerzler, G. Mushinsky, E. Walter, L. Hofman, and G. F. Grady. 1993. Identifying human immunodeficiency virus infection at birth: application of polymerase chain reaction to Guthrie cards. *J. Pediatr.* **123**:252–258.
8. Cunningham, C. K., T. T. Charbonneau, K. Song, D. Patterson, T. Sullivan, T. Cummins, and B. Poiesz. 1999. Comparison of human immunodeficiency virus 1 DNA polymerase chain reaction and qualitative and quantitative RNA polymerase chain reaction in human immunodeficiency virus 1-exposed infants. *Pediatr. Infect. Dis. J.* **18**:30–35.
9. De Baets, A. J., B. S. Edidi, M. J. Kasali, G. Beelaert, W. Schrooten, A. Litroth, P. Kolsteren, D. Denolf, and K. Fransen. 2005. Pediatric human immunodeficiency virus screening in an African district hospital. *Clin. Diagn. Lab. Immunol.* **12**:86–92.
10. Fischer, A., C. Lejczak, C. Lambert, J. Servais, N. Makombe, J. Rusine, T. Staub, R. Hemmer, F. Schneider, J. C. Schmit, and V. Arendt. 2004. Simple DNA extraction method for dried blood spots and comparison of two PCR assays for diagnosis of vertical human immunodeficiency virus type 1 transmission in Rwanda. *J. Clin. Microbiol.* **42**:16–20.
11. Fiscus, S. A., J. Weiner, E. Abrams, M. Bulterys, A. Cacahfeiro, and R. A. Respass. 2007. Ultra-sensitive p24 antigen assay for the diagnosis of perinatal human immunodeficiency virus type 1 infection. *J. Clin. Microbiol.* **45**:2274–2277.
12. Fiscus, S. A., B. Cheng, S. Crowe, L. Demeter, C. Jennings, V. Miller, R. Respass, W. Stevens, and the Forum for Collaborative HIV Research Alternative Viral Load Assay Working Group. 2006. HIV viral load assays for resource limited settings—a review. *PLoS* **3**:e417.
13. George, E., C. A. Beauharnais, E. Brignoli, F. Noel, P. De Matteis Rouzier, M. Altenor, D. Lauture, M. Hosty, S. Mehta, P. E. Wright, and J. W. Pape. 2007. Potential of a simplified p24 assay for early diagnosis of infant human immunodeficiency virus type 1 infection in Haiti. *J. Clin. Microbiol.* **45**:3416–3418.
14. Knuchel, M. C., Z. Tomasik, R. F. Speck, R. Luthy, and J. Schupbach. 2006. Ultrasensitive quantitative HIV-1 p24 antigen assay adapted to dried plasma spots to improve treatment monitoring in low-resource settings. *J. Clin. Virol.* **36**:64–67.
15. Lambert, J. S., D. R. Harris, E. R. Stiehmel, J. Moye, Jr., M. G. Fowler, W. A. Meyer III, J. Bethel, and L. M. Mofenson. 2003. Performance characteristics of HIV-1 culture and HIV-1 DNA and RNA amplification assays for early diagnosis of perinatal HIV-1 infection. *J. Acquir. Immune Defic. Syndr.* **34**:512–519.
16. Li, C. C., I. A. Beck, K. D. Seidel, and L. M. Frenkel. 2004. Persistence of human immunodeficiency virus type 1 subtype B DNA in dried-blood samples on FTA filter paper. *J. Clin. Microbiol.* **42**:3847–3849.
17. Li, C. C., K. D. Seidel, R. W. Coombs, and L. M. Frenkel. 2005. Detection and quantification of human immunodeficiency virus type 1 p24 antigen in dried whole blood and plasma on filter paper stored under various conditions. *J. Clin. Microbiol.* **43**:3901–3905.
18. Lyamuya, E., U. Bredberg-Raden, A. Massawe, E. Urassa, G. Kawo, G. Msemo, T. Kazimoto, A. Ostborn, K. Karlsson, F. Mhalu, and G. Biberfeld. 1996. Performance of a modified HIV-1 p24 antigen assay for early diagnosis of HIV-1 infection in infants and prediction of mother-to-infant transmission of HIV-1 in Dar es Salaam, Tanzania. *J. Acquir. Immune Defic. Syndr.* **12**:421–426.
19. Nadal, D., J. Boni, C. Kind, O. E. Varinier, F. Steiner, Z. Tomasik, and J. Schupbach. 1999. Prospective evaluation of amplification-boostered ELISA for heat-denatured p24 antigen for diagnosis and monitoring of pediatric HIV-1 infections. *J. Infect. Dis.* **180**:1089–1095.
20. Nyambi, P. N., K. Fransen, H. De Beenhouwer, E. N. Chomba, M. Temmerman, J. O. Ndinya-Achola, P. Piot, and G. van der Groen. 1994. Detection of human immunodeficiency virus type 1 (HIV-1) in heel prick blood on filter paper from children born to HIV-1-seropositive mothers. *J. Clin. Microbiol.* **32**:2858–2860.
21. Panteleeff, D. D., G. John, R. Nduati, D. Mbori-Ngacha, B. Richardson, J. Kreiss, and J. Overbaugh. 1999. Rapid method for screening dried blood samples on filter paper for human immunodeficiency virus type 1 DNA. *J. Clin. Microbiol.* **37**:350–353.
22. Patton, J. C., G. G. Sherman, A. H. Coovadia, W. S. Stevens, and T. M. Meyers. 2006. Ultrasensitive human immunodeficiency virus type 1 p24 antigen assay modified for use on dried whole blood spots as a reliable, affordable test for infant diagnosis. *Clin. Vaccine Immunol.* **13**:152–155.
23. Patton, J. C., A. H. Coovadia, T. M. Meyers, and G. G. Sherman. 2008. Evaluation of the ultrasensitive human immunodeficiency virus type 1 (HIV-1) p24 antigen assay performed on dried blood spots for diagnosis of HIV-1 infection in infants. *Clin. Vaccine Immunol.* **15**:388–391.
24. Rich, K. C., W. Janda, L. A. Kalish, J. Lew, D. Hofheinz, S. Landesman, J. Pitt, C. Diaz, J. Moye, and J. L. Sullivan. 1997. Immune complex dissociated p24 antigen in congenital or perinatal HIV infection: role in the diagnosis and assessment of risk of infection in infants. *J. Acquir. Immune Defic. Syndr.* **12**:198–203.
25. Schupbach, J., J. Boni, L. R. Bisset, Z. Tomasik, M. Fischer, H. F. Gunthard, B. Ledergerber, M. Opravil, and Swiss HIV-1 Cohort Study. 2003. HIV-1 p24 antigen is a significant inverse correlate of CD4 T-cell change in patients with suppressed viremia under long-term antiretroviral therapy. *J. Acquir. Immune Defic. Syndr.* **33**:292–299.
26. Sherman, G. G., G. Stevens, and W. Stevens. 2004. Affordable diagnosis of HIV-1 infection in infants by p24 antigen detection. *Pediatr. Infect. Dis. J.* **23**:173–176.
27. Sherman, G. G., G. Stevens, S. A. Jones, P. Horsfield, and W. S. Stevens. 2005. Dried blood spots for HIV-1 diagnosis of infants: improving access to care for HIV-1 affected infants in low resource settings. *J. Acquir. Immune Defic. Syndr.* **38**:615–617.
28. Simonds, R. J., T. M. Brown, D. M. Thea, S. L. Orloff, R. W. Steketee, F. K. Lee, P. E. Palumbo, and M. L. Kalish. 1998. Sensitivity and specificity of a qualitative RNA detection assay to diagnose HIV-1 infection in young infants. Perinatal AIDS Collaborative Transmission Study. *AIDS* **12**:1545–1549.
29. Sutthent, R., N. Gaudart, K. Chokpaibulkit, N. Tanliang, C. Kanoksinsombath, and P. Chaisilwatana. 2003. p24 antigen detection assay modified with a booster step for diagnosis and monitoring of human immunodeficiency virus type 1 infection. *J. Clin. Microbiol.* **41**:1016–1022.
30. Young, N. L., N. Shaffer, T. Chaowanachan, T. Chotpitayusunondh, N. Vanparapar, P. A. Mock, N. Waranawat, K. Chokephaibulkit, R. Chuachoo Wong, P. Wasinrapee, T. D. Mastro, R. J. Simonds, and Bangkok Collaborative Perinatal HIV-1 Transmission Study Group. 2000. Early diagnosis of HIV-1-infected infants in Thailand using RNA and DNA PCR assays sensitive to non-B subtypes. *J. Acquir. Immune Defic. Syndr.* **24**:401–407.
31. Yourno, J., and J. Conroy. 1992. A novel polymerase chain reaction method for detection of human immunodeficiency virus in dried blood spots on filter paper. *J. Clin. Microbiol.* **30**:2887–2892.
32. Zijenah, L. S., O. Tobaiwa, S. Rusakaniko, K. J. Nathoo, M. Nhembe, P. Matibe, and D. A. Katzenstein. 2005. Signal-boostered qualitative ultrasensitive p24 antigen assay for diagnosis of subtype C HIV-1 infection in infants under the age of 2 years. *J. Acquir. Immune Defic. Syndr.* **39**:391–394.