Published in final edited form as:

J Biomed Opt. 2008 ; 13(5): 054028. doi:10.1117/1.2976141.

Fast 3D dark-field reflection-mode photoacoustic microscopy *in vivo* with a 30-MHz ultrasound linear array

Liang Song^a, Konstantin Maslov^a, Rachel Bitton^b, K. Kirk Shung^b, and Lihong V. Wang^{a,1}

^aWashington University in St. Louis, Optical Imaging Laboratory, Department of Biomedical Engineering, St. Louis, Missouri 63130

^bUniversity of Southern California, Department of Biomedical Engineering, Los Angeles, California 90089

Abstract

We present an *in vivo* dark-field reflection-mode photoacoustic microscopy system that performs cross-sectional (B-scan) imaging at 50 Hz with realtime beamforming and 3D imaging consisting of 166 B-scan frames at 1 Hz with post-beamforming. To our knowledge, this speed is currently the fastest in photoacoustic imaging. A custom-designed light delivery system is integrated with a 30-MHz ultrasound linear array to realize dark-field reflection-mode imaging. Linear mechanical scanning of the array produces 3D images. The system has axial, lateral, and elevational resolutions of 25, 70, and 200 μm , respectively, and can image 3 mm deep in scattering biological tissues. Volumetric images of subcutaneous vasculature in rats are demonstrated *in vivo*. Fast 3D photoacoustic microscopy is anticipated to facilitate applications of photoacoustic imaging in biomedical studies that involve dynamics and clinical procedures that demand immediate diagnosis.

Keywords

biomedical imaging; realtime photoacoustic imaging; three-dimensional photoacoustic microscopy; high frequency ultrasound array; clinical diagnosis; subcutaneous vasculature; small animal imaging

1 Introduction

Photoacoustic imaging is a hybrid imaging modality with excellent optical absorption contrast and high resolution beyond the optical ballistic and quasiballistic regimes (~ 1 mm in scattering biological tissues) [1]. It has become a rapidly growing field in biomedical research because of its great potential in breast cancer diagnosis [2,3], skin melanoma detection [4-6], and functional brain imaging [7,8]. High imaging speed is essential for biomedical research involving dynamics and is highly desirable for decision-making in clinics. However, photoacoustic imaging systems using a single element ultrasound transducer are limited in speed by mechanical scanning. To improve imaging speed, several groups have employed ultrasound arrays for photoacoustic imaging [9,10]. For example, using a 7.5 MHz ultrasound array, Niederhauser et al. [9] obtained an imaging speed of 7.5 B-scan-frames/s with 0.3-0.4 mm resolution. While ultrasound arrays of clinical diagnostic ultrasound frequencies (several MHz) are commercially available, arrays of high frequencies (>12 MHz) are still at the research stage. We developed a photoacoustic microscopy system with a high frequency (30 MHz) ultrasound array [11], which provided ~ 25 μm axial and <100 μm lateral resolutions. Realtime B-scan imaging capability was demonstrated [12].

¹Phone: (314) 935-6152; Fax: (314) 935-7448; E-mail: lhwang@biomed.wustl.edu.

Besides high imaging speed, 3D imaging is also highly desirable. One advantage of 3D imaging is to provide maximum amplitude projection (MAP) images in various orientations. Another advantage is the ability to view images interactively. The flexibility to rotate, scale, and view the region of interest from various perspectives can facilitate visualization.

We present a fast 3D dark-field reflection-mode photoacoustic microscopy system with a 30-MHz ultrasound linear array. This system performs realtime cross-sectional (B-scan) imaging at 50 Hz (faster than the 30-Hz video rate) with realtime beamforming and 3D imaging of 166 B-scan frames at 1 Hz with post-beamforming. To enable dark-field reflection-mode photoacoustic imaging, we designed a novel light delivery system consisting of both fiber and free-space optic components. The dark-field laser pulse illumination configuration was known to have advantages in suppressing undesirable photoacoustic signals from the superficial layers of the skin [4]. Three-dimensional photoacoustic images of the subcutaneous vasculature in rats were demonstrated *in vivo*, which matched well with their *ex vivo* transmission optical microscopy counterparts. Our photoacoustic microscopy system is also presented as a compact prototype for handheld operation and is anticipated to accelerate preclinical and clinical applications of photoacoustic imaging.

2 Methods

A schematic of the system is shown in Fig. 1. Our system consists of a diode-pumped Q-switched Nd:YLF laser, a tunable dye laser, a 30-MHz ultrasound linear array, custom receive and control electronics, an 8-channel PCI data acquisition (DAQ) card, a multi-core PC, a custom designed light delivery system, and a motorized linear motion actuator.

2.1 Optics and light delivery

As the irradiation source, the tunable dye laser (Cobra, Sirah Laser-und Plasmatechnik GmbH, Germany) was pumped by the Nd:YLF laser (INNOSLAB, Edgewave GmbH, Germany). The Nd:YLF pump laser had a pulse duration of 7 ns and a pulse energy of 12 mJ at 523 nm. The continuous optical pumping from the diode stacks in this Q-switched Nd:YLF laser provided the flexibility of external triggering on demand at rates up to 1 kHz without compromising the pulse energy. This feature offered a significant advantage over flashlamp-pumped Q-switched lasers, which are typically maintained at a fixed low pulse repetition rate (e.g., 10 Hz). Rhodamine 6G laser dye was used to enable a peak output of 2 mJ per pulse with a pulse width of 7 ns at the 570-nm wavelength from the dye laser, with a 40-nm tuning range. For our imaging experiments, we used the peak 570-nm wavelength. This wavelength also corresponds to an isosbestic point where oxyand deoxy-hemoglobin have equal molar extinction coefficients.

Proper delivery of laser light into biological tissues for photoacoustic excitation is crucial to achieving a high signal-to-noise ratio (SNR) in photoacoustic imaging. The light delivery system was designed to provide a compact photoacoustic imaging device with sufficient SNR and robust performance. The dye laser output was coupled into a 0.6-mm-core-diameter multimode optical fiber, and was collimated by a fiber collimator at the output end of the fiber. A free space optic setup, integrated with the ultrasound array, was used to guide light further (Fig. 1). The collimated light beam, of ~6 mm diameter and ~1.2 mJ energy, was split into two beams by a 50:50 non-polarizing beamsplitter. The two beams were reflected by mirrors towards two cylindrical lenses and were coupled into a plastic slab. The 6-mm thick plastic slab, with a hollow cylindrical core for the ultrasound probe, was polished for light transmission. During experiments, the cylindrical space was filled with water and sealed by a piece of thin low-density polyethylene (LDPE) film fixed by an o-ring. In total, ~80% of the light energy reached the surface of the film. Due to finite fiber aperture the final optical

illumination patterns on the skin surface were thick-line shaped, as shown in Fig. 1. The length and width of each illumination area were 6 and 3 mm, respectively.

Dark-field laser pulse illumination was achieved through fine tuning the mirrors and cylindrical lenses, reducing the photoacoustic signals from the superficial paraxial area. However, a large dark-field area may reduce the optical fluence reaching the targeted area. The optimal illumination radius was estimated to be 7 mm using the concept of effective attenuation coefficient [13], the exponential decay rate of fluence far from the source, with typical tissue parameters. Consequently, leaving a ~ 1 mm width right below the array elements as the dark field gives approximately the best performance. In practice, the mirrors and cylindrical lenses were finely tuned to optimize the SNR.

The optical fluence on the skin surface was estimated to be ~ 2 mJ/cm² per pulse, well below the ANSI recommended Maximum Permissible Exposure (MPE) of 20 mJ/cm² for a single pulse. We acquired data in 1 s—50 frames for realtime B-scan imaging or 166 frames for 3D imaging; the time average light intensity during this 1 s was ~ 600 mW/cm² (the total illuminated surface area for one 3D image was 1.2 cm² due to the mechanical scanning), also below the ANSI recommended MPE calculated as $1.1t^{0.25}$ W/cm² (t in sec.) [14]. For prolonged illumination, the ANSI recommended MPE for average light intensity would be lower. However, for prolonged illumination during B-scan imaging, we can either pause a few seconds between acquisitions or slow down the frame rate. We also expect to reduce the delivered energy with improved SNR by system optimization. The ANSI safety limit for this pulse width region is dominantly based on the thermal mechanism; thus our compliance to the ANSI standards guarantees no thermal damage to the tissue.

2.2 Ultrasound array and beamforming

We used a unique 30-MHz ultrasound linear array fabricated from a 2-2-piezo-composite by the NIH Resource Center for Medical Ultrasonic Transducer Technology at the University of Southern California [15]. The array had 48 elements (of dimensions $82\text{ }\mu\text{m} \times 2\text{ mm}$) with 100- μm spacing. The dimension of the element in the elevation direction was 2 mm, and the elements were focused in this direction with a fixed focal length of 8.2 mm, which provides a resolution of 200 μm in the elevation direction within the ~ 3.5 mm focal zone. The pulse-echo insertion loss and element cross-talk were 19.1 and -25 dB, respectively. The mean fractional bandwidth was 50% for pulse-echo operation, which translates to $\sim 70\%$ for receiving-only operation, as used in our present photoacoustic imaging system.

While ultrasound beamforming traditionally has used dedicated hardware [16,17], we instead used multi-core processors (Dell Precision 490 with two 2.66 GHz Quad core Xeon processors), which allows off-the-shelf personal computers to perform the task and offers programming flexibility. Microsoft Visual Studio 2005 Professional Edition and Microsoft XNA Game Studio in the Visual Studio C# 2005 Express Edition environment were used to develop the software for dynamic receive beamforming and display. Details on implementation of multi-threaded parallel programming and GPU-based scan conversion and display in this software beamforming can be found elsewhere [12].

2.3 Data acquisition and volume imaging

Photoacoustic signals picked up by the ultrasound array were amplified by a custom-built RF board with a 33–73 dB variable gain and were down-multiplexed to eight channels, which were digitized at 125 mega-samples per second (MSPS) using a 14-bit 8-channel PCI DAQ card (Octopus CompuScope 8389, GaGe Applied Systems, USA). The card was used as the master clock for the entire system and was programmed to send trigger signals to the multiplexer

control and laser. The repetition rate was set at 1 kHz, which was the highest rate that the laser could work without degradation of pulse energy.

We used linear scanning to achieve 3D imaging. During the scanning, the array translated linearly over the skin surface, so that the B-scan imaging planes were all parallel to each other. This was accomplished by mounting the light delivery system and the ultrasound array in a linear motion actuator (KR20, THK CO. LTD., Japan). A bipolar stepper motor (4118S, Lin Engineering, USA) controlled by a microstep stepper motor controller (BC2D15, Peter Norberg Consulting, Inc., USA) was used to drive the linear motion actuator. The scanning system provided sufficient precision (20 μm) for our use. Six laser shots were needed to obtain one B-scan image because of the 6:1 down-multiplexing in our data acquisition. To produce one 3D image, 166 B-scan frames were acquired in 996 ms, corresponding to 996 laser shots at a 1-kHz repetition rate. During the data acquisition, the array scanned continuously at a constant speed (10 mm/s). The speed was set so that the distance the array travelled during each B-scan time period was 60 μm , less than the ultrasonic focus in the elevational direction (200 μm). Although the linear motion actuator was fixed on an optical table for scanning, handheld operation was also possible. More user-friendly handheld operation for daily clinical use can be achieved by shrinking the size of the light delivery and scanning systems through custom manufacturing.

While realtime B-scan imaging was demonstrated at 50 frames per second (FPS) [12], post-beamforming after data acquisition reached 83 FPS. This higher speed was due to less hardware communication. In post-beamforming, a B-scan movie was first played, and an MAP of the acquired 3D image was then displayed for preview immediately. In total, ~ 3 s was needed for a user to view an MAP image, representing the fastest speed among reported photoacoustic imaging systems. The user might choose to either replay the B-scan movie or display a contrast-enhanced MAP image processed by a Dynamic-link Library (DLL) written in MATLAB (Math Works, Inc., USA). All these operations could be done online by simply clicking corresponding buttons on a graphic user interface (GUI) generated by the C# program.

3 Results

The spatial resolution of our system was characterized by imaging 6- μm -diameter carbon fibers in water. The axial, lateral, and elevational resolutions (at ~ 7 mm normal depth from the transducer surface) were estimated to be 25, 70, and 200 μm , respectively. The imaging depth was demonstrated to be greater than 3 mm in scattering biological tissues.

To demonstrate the system's capability to image blood vessels *in vivo*, we imaged the upper dorsal region of a Sprague Dawley rat (Harlan Sprague Dawley, Inc., USA), ~ 95 g in weight. The rat was anesthetized by intramuscular injection of a mixture of Ketamine (85 mg/kg) and Xylazine (15 mg/kg). The hair in the imaged region was removed with commercial hair-removal lotion before imaging. Acoustic coupling gel was applied to improve acoustic coupling between the animal and the LDPE film. After the experiment, the rat recovered normally, without noticeable health problems. All experimental animal procedures were carried out in compliance with approved protocols.

Fig. 2 shows an *in vivo* MAP photoacoustic image and an *ex vivo* transmission optical microscopic image of the subcutaneous vasculature in an imaged area. The photoacoustic image is a gray-level plot of the maximum amplitude of the envelope of each time-resolved (A-scan) photoacoustic signal within a 2 mm depth from the skin. The signals from the skin surface were digitally removed. The transmission optical microscopic photograph ($\sim 4 \times 5$ mm), partially overlapping with the photoacoustic image, was acquired *ex vivo* from the dermal side of the excised skin. As shown in Fig. 2, the vascular distributions matched well between

the two images. Vessels shown in the photoacoustic image had diameters varying from ~80 to ~700 μm . The data acquisition of the photoacoustic image was completed within 1 s, and the MAP image was displayed within 3 – 4 s. An animation of the 3D image from varying perspectives (Fig. 3) was obtained from post-processing using VolView (Kitware Inc., USA). Interactive features (rotation, scaling, etc.) were available in VolView as well.

To validate the system's reproducibility, we imaged a sacrificed Sprague Dawley rat noninvasively *in situ*. The experimental procedures were similar to those described above, except that the rat was sacrificed to avoid imaging artifacts due to breathing motion. We repeated scanning the same area (in the lower dorsal region) four times, with a time interval of ~2 min. The system demonstrated robust performance, with major vessels shown consistently in all four images (Fig. 4).

4 Conclusions and discussion

We have demonstrated fast 3D photoacoustic microscopy with a 30-MHz ultrasound linear array capable of realtime B-scan imaging at 50 Hz and 3D imaging at 1 Hz. To our knowledge, this is the fastest photoacoustic imaging achieved hitherto. The system integrates a novel light delivery system and a motorized linear motion actuator with the ultrasound array, performing scans in elevational direction for 3D imaging. The light delivery system enables robust reflection-mode imaging, which is applicable to more anatomical sites than orthogonal- or transmission-mode imaging. Optical dark-field illumination is also achieved, reducing the otherwise strong interference of the extraneous photoacoustic signals from the superficial paraxial area. The imaging performance is validated to be robust and reproducible.

Previous studies have demonstrated that photoacoustic imaging has great potential in early cancer diagnosis, blood oxygenation mapping, and functional brain imaging. Because of the unique realtime 2D and fast 3D imaging capability, our system is anticipated to facilitate many of these biomedical applications. The compact system design also favors clinical practices. Particularly, the promising results shown in this report should motivate future development of photoacoustic microscopy with high frequency ultrasound arrays towards clinical applications.

Acknowledgement

This work was sponsored in part by grants R01 EB000712 and R01 NS46214 from the National Institutes of Health.

References

1. Wang, LV.; Wu, HI. Biomedical optics: principles and imaging. John Wiley & Sons; Hoboken, New Jersey: 2007.
2. Manohar S, Vaartjes SE, van Hespen JCG, Klaase JM, van den Engh FM, Steenbergen W, van Leeuwen TG. Initial results of *in vivo* non-invasive cancer imaging in the human breast using near-infrared photoacoustics. *Opt. Express* 2007;15:12277–12285.
3. Ermilov SA, Conjusteau A, Mehta K, Lacewell R, Henrichs PM, Oraevsky AA. 128-channel laser photoacoustic imaging system (LOIS-128) for breast cancer diagnostics. *Proc. SPIE* 2006;6086:68–79.
4. Zhang HF, Maslov K, Stoica G, Wang LV. Functional photoacoustic microscopy for high-resolution and noninvasive *in vivo* imaging. *Nat. Biotechnol* 2006;24:848–851. [PubMed: 16823374]
5. Oh JT, Li M, Zhang HF, Maslov K, Stoica G, Wang LV. Three-dimensional imaging of skin melanoma *in vivo* by dual-wavelength photoacoustic microscopy. *J. Biomed. Opt* 2006;11:034032.
6. Weight RM, Viator JA, Dale PS, Caldwell CW, Lisle AE. Photoacoustic detection of metastatic melanoma cells in the human circulatory system. *Opt. Lett* 2006;31:2998–3000. [PubMed: 17001379]
7. Wang X, Pang Y, Ku G, Xie X, Stoica G, Wang LV. Non-invasive laser-induced photoacoustic tomography for structural and functional imaging of the brain *in vivo*. *Nat. Biotechnol* 2003;21:803–806. [PubMed: 12808463]

8. Yang S, Xing D, Lao Y, Yang D, Zeng L, Xiang L, Chen WR. Noninvasive monitoring of traumatic brain injury and post-traumatic rehabilitation with laser-induced photoacoustic imaging. *Appl. Phys. Lett* 2007;90:243902.
9. Niederhauser J, Jaeger M, Lemor R, Weber P, Frenz M. Combined ultrasound and optoacoustic system for real-time high contrast vascular imaging *in vivo*. *IEEE Trans. Med. Imaging* 2005;24:436–440. [PubMed: 15822801]
10. Yang DW, Xing D, Yang SH, Xiang LZ. Fast full-view photoacoustic imaging by combined scanning with a linear transducer array. *Opt. Express* 2007;15:15566–15575.
11. Zemp R, Li M, Bitton R, Shung KK, Stoica G, Wang LV. Photoacoustic imaging of the microvasculature with a high-frequency ultrasound array transducer. *J. Biomed. Opt* 2007;12:010501. [PubMed: 17343475]
12. Zemp RJ, Song L, Bitton R, Shung KK, Wang LV. Realtime photoacoustic microscopy *in vivo* with a 30-MHz ultrasound array transducer. *Opt. Express*. in press
13. Wang LV, Chen WR, Nordquist RE. Optimal beam size for light delivery to absorption-enhanced tumors buried in biological tissues and effect of multiple beam delivery: a Monte Carlo study. *Appl. Opt* 1997;36:8286–8291. [PubMed: 18264369]
14. The Laser Institute of America. American National Standard for Safe Use of Lasers (ANSI Z136.1-2000). The Laser Institute of America; 2000.
15. Cannata JM, Williams JA, Zhou QF, Ritter TA, Shung KK. Development of a 35-MHz piezo-composite ultrasound array for medical imaging. *IEEE Trans. Ultrason. Ferroelectr. Freq. Control* 2006;53:224–236. [PubMed: 16471449]
16. Wall K, Lockwood GR. Modern implementation of a realtime 3D beamformer and scan converter system. *IEEE Ultrasonics Symposium* 2005:1400–1403.
17. Hu CH, Xu XC, Cannata JM, Yen JT, Shung KK. Development of a real-time, high-frequency ultrasound digital beamformer for high-frequency linear array transducers. *IEEE Trans. Ultrason. Ferroelectr. Freq. Control* 2006;53:317–323. [PubMed: 16529106]

Fig. 1.
Schematic of the 3D dark-field reflection-mode photoacoustic imaging system.

Fig. 2.

(a) *In vivo* noninvasive MAP photoacoustic image of subcutaneous blood vessels in the upper dorsal region of a Sprague Dawley rat. X and Y represent the mechanical scan and electronic beamforming directions, respectively. The grayscale represents relative optical absorption (arbitrary unit). The regions enclosed by major blood vessels are labeled using capital letters A–E. (b) Photograph taken from the dermal side of the excised skin with transmission illumination. The same regions are identified and labeled A–E. The photograph covers most of the photoacoustically imaged region.

Fig. 3. Movie (Video 1) of *in vivo* three-dimensional photoacoustic images of the upper dorsal region of a Sprague Dawley rat.

Fig. 4. Four consecutive photoacoustic-image acquisitions of the same lower dorsal area of a sacrificed Sprague Dawley rat (images were acquired noninvasively *in situ*). The grayscale represents relative optical absorption (arbitrary unit). The major vessels are shown consistently.