Published in final edited form as:

Talanta. 2009 July 15; 79(2): 280–283. doi:10.1016/j.talanta.2009.03.054.

Development of an automatic solid phase extraction and liquid chromatography mass spectrometry method by using a monolithic column for the analysis of Cyclosporin A in human plasma

Covadonga Alvarez¹ and Irving W. Wainer^{2,*}

¹ Department of Pharmaceutical Technology. Faculty of Pharmacy. Complutense University of Madrid. 28040. Madrid. Spain

² Laboratory of Bioanalytical Chemistry. Gerontology Research Center. National Institutes of Health. 5600 Nathan Shock Drive. Baltimore, MD 21224-6825. USA

Abstract

A sensitive and specific and automated liquid chromatography-electrospray mass spectrometric (LC-ESI-MS) assay for the quantification of Cyclosporin A in human plasma **was developed**. Following a simple protein precipitation step, the supernatant was extracted on line and directly injected into the system LC-ESI-MS. A relatively new type of monolithic column consisting on a silica rod with bimodal pore structure was used to achieve a retention time of 2.4 minutes with a very low backpressure at a flow rate of 1 ml/min. The assay was linear from 0.050 to 1.000 µg/ml. The mean recovery was 91%. The mean inter-day intra-day precisions were 1.85, 2.83% respectively. The combination of the automated solid phase extraction and the low retention time achieved with this columns increase the throughput and decrease the time of analysis of each sample. This technology is useful in order to improve the efficiency of the bioanalytical studies.

Keywords

Cyclosporin A; Monolithic columns; LC-ESI-MS

1. Introduction

Monolithic materials for the use in HPLC are one of the tools for bioanalysis [1,2]. Different monolithic-type HPLC column based on silica for fast separation have developed achieving speed, sensitivity and selectivity with low backpressure.

The use of the Merck Chromolith columns is described in this paper. These columns are prepared according to a sol-gel process, which is based on the hydrolysis and polycondensation of alkoxy silanes in the presence of water-soluble polymers [3,4]. The method leads to rods made of one piece of porous silica with a define porous structure. The main feature of silica

*Corresponding author: Tel: 410-558-8498, Fax: 410-558-8409, e-mail: E-mail: wainerir@grc.nia.nih.gov.

Publisher's Disclaimer: This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our customers we are providing this early version of the manuscript. The manuscript will undergo copyediting, typesetting, and review of the resulting proof before it is published in its final citable form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

rods columns is a higher total porosity, about 15% higher than for conventional particulate HPLC columns. The column pressure drop is therefore much lower, allowing to operate at higher flow rates including flow gradients. Consequently, the high speed and good separation can be achieved at the same time using this type of monolithic column. The important feature of this is its relevance in high-throughput bioanalysis [5]. The method described has been used to screen over 4000 compounds thus far and has met the sensitivity and reproducibility criteria on over 97% of the compounds tested. The advantage of this method is that development time is no longer needed to investigate new drug candidates, and little, if any, sample preparation time is required. The result is increased productivity through higher sample throughput.

As a direct result of the short analysis times offered by these monolithic columns, sample preparation has become the rate-limiting step. Much effort has been also devoted in order to automate the sample preparation step [6]. Another possible approach to overcome the problem of high pressure associated with small particles is to fabricate a column made of one piece of a porous solid with small-sized skeletons and relatively large through-pores which could provide both low pressure drop and high column efficiency. The skeletons can be meso-porous or microporous to have double-pore structures, or even nonporous. The most important features are high mechanical stability of beds and the freedom in the ratio of through-pore size to skeleton size. Several examples of such monolithic columns made of an organic polymer have been reported recently

The Cyclosporin A (CsA) (Figure 1) has been used widely as an autoimmune drug in organ transplantation, treatment of autoimmune disorders and psoriasis [7]. The analysis of CsA in biological samples by high-performance liquid chromatography has been reported before [8, 9,10]. The most sensitive method was achieved by coupling liquid chromatography to mass spectrometry (LC-MS) [11] or tandem mass spectrometry (LC-MS-MS) [12].

A first side-by-side comparison of the two main types of commercially available monolithic nano-LC column used in reversed-phase liquid chromatography by using them to analyze a real-life proteomics sample (*A. thaliana* chloroplast) has been studied [13]. Superior separations were difficult to achieve on PS-DVB when strict MS-targeted conditions were adhered to (i.e., no additional ion-pairing agents were permitted in the mobile phase). In such cases, the conclusion was that nano-LC columns made from monolithic silica outperform those made from polystyrene in terms of separation efficiency and number of peptides collected. In addition, the higher permeability of the monolith allowed higher flow rates to be used, which enabled an increase in the information throughput. This should be of particular interest to those who wish to couple RP-HPLC with separation methods without sacrificing analytical throughput.

The aim of this work was to develop a new LC-MS method for CsA determination in human plasma by using the monolithic type columns (Chromolith Performance) in combination with an automatic solid phase extraction (SPE) system in order to approach for optimizing its extraction and quantitation, after oral administration, achieving fast separation and high throughput.

2. Experimental

2.1. Chemicals

HPLC grade acetonitrile was purchased from Fisher Scientific (Fairlawn, NJ, USA). Ultra-pure water was obtained from Mili-Q water purification system (Millipore, Milford, MA, USA). Analytical grade ammonium acetate was obtained from J.T.Baker (Phillipsburg, NJ, USA). CsA was purchased from Sigma-Aldrich (St.Louis, MO,USA). Blank human plasma

(Sodium heparin as anticoagulant) was used to prepare calibration and quality control samples (Valley Biomedical).

2.2. Instrument

The HPLC equipment consisted on an HP 1100 Series with a ChemStation (Hewlett Packard, Waldbronn, Germany) with electrospray, positive mode (API-ES). The separation column was Chromolith Performance RP-18e (10×4.6 mm) from Merck (Merck KGaA, Darmstadt, Germany) was eluted by 90% acetonitrile/10% ammonium acetate buffer pH 5.1 with a flow rate of 1ml/min. The injection volume was 50µl and the column temperature was maintained at 43°C.

The LC-API-ES-MS conditions were set as follows: the fragmentor voltage was 250V, the nitrogen gas flow was maintained at 11 l/min, the nebulizer pressure was set up at 40 psig, the positive capillary voltage was 3500V and the drying gas temperature was 350°C. The quantification is based on the total peak area of the CsA in SIM chromatogram.

2.3. Automatic sample preparation

The extractions were performed in an automatic Gilson ASPEC XL (West Beltline, Hwy, USA). The system is designed for the automation and optimization of SPE in order to provide a more efficient sample preparation. In addition the system is fitted with an injection valve for performing SPE with on-line injection onto the system. Plasma samples (500 µl) were loaded onto the extraction cartridges (Oasis HLB, Waters Corporation, Milford, MA, USA), which had been preconditioned sequentially using 500 µl of methanol and 500 µl of water. The cartridge was then washed using 500 µl of water followed by elution with 500 µl of methanol with 2% of HCl. The eluted samples were directly injected into the LC-API-ES-MS system.

2.4. Methods

2.4.1. Optimization of LC-API-ES-MS experimental conditions—In order to set up the optimized conditions, the following parameters were tuned: the fragmentor voltage (50 to 250V), the capillary voltage (1000 to 4000V), the nebulizer pressure (40 to 60 psig) and the drying temperature (200 to 350°C).

2.4.2. Calibration curve—The determination of CsA was based on the external standard method. Five point calibration curves (triplicate injections) were created at 0.05, 0.125, 0.250, 0.500 and 1.000 µg/ml.

2.4.3. Recovery, precision and accuracy—The recoveries were determined by comparing the peak areas of plasma premixed with a known amount of CsA with that of an equivalent amount of standard CsA dissolved in pure methanol. For intra-day precision, spiked plasma samples with three different concentrations (0.050, 0.250 and 1.000µg/ml) were analyzed. For the inter-day precision, the above samples were analyzed on three subsequent days. Accuracy was measured using CsA standard samples (0.050, 0.125, 0.250, 0.500 and 1.000 µg/ml) and calculated as the deviation from the theoretical values.

3. Results and discussion

3.1. Optimization of the LC-API-ES-MS conditions

In order to measure the concentration of CsA in plasma with high sensitivity the API-ES interface parameters including the fragmentor voltage, the positive capillary voltage and the drying gas temperature were tested.

The fragmentor was set up at 250V. The positive capillary voltage was raised at 3500V, and the optimal conditions for the drying gas temperature was approached at 350°C.

With the optimal conditions the peak quality is shown in Figure 2. As be seen in the Figure 2 the retention time obtained is ≈ 2.4 minutes, due to the high porosity of the monolithic silica rod columns which posses a higher total porosity, allowing the use of a high flow rate achieving a low backpressure. The decrease on retention time is evident in comparison with previous studies. Salm et al. [11]. obtained a retention time of 17 minutes for CsA by using a 10 μm Bondepak C₁₈ column (3.9 $\times$ 300 mm I.D., 10 μm particle sizes, Waters, Milford, USA). The mobile phase was a mixture of acetonitrile-methanol-deionized water (55:15:30) V/V/V. The use of other type of particulate column (phenyl type) is also described in the bibliography for CsA samples [10] achieving a retention time of 8 minutes.

3.2. Linearity, accuracy and recovery of the LC-API-ES-MS method

The LOD, defined as the minimum analyte concentration that produced an instrumental signal significantly different from that for the blank, was calculated in accordance with the IUPAC's criterion, i.e. as the analyte concentration giving a signal exceeding that of the blank (y_B) by 3 times its standard deviation (S_B). The LOQ, defined as the minimum analyte concentration needed to ensure precise quantitative measurements, was determined similarly, using 10 times the standard deviation for the blank instead. Thus, LOD and LOQ were calculated from the following expression:

$$\text{LOD (LOQ)} = \frac{y_B + 3(10)S_B}{b}$$

where b is the slope of the calibration curve. The limit of detection (LOD) and the limit of quantification (LOQ) for the CsA in plasma were 0.0004 and 0.015 $\mu\text{g/ml}$ respectively.

The standard curve was created between 0.050 and 1.000 $\mu\text{g/ml}$ with a typical correlation coefficient of 0.9989 ($A = -5399,16 + 6511,32 \times C$). Where A achieves the absorbance obtained and C de concentration achieved expressed in $\mu\text{g/ml}$. The peak area increased also linearly over the range from 0.015 to 5.000 $\mu\text{g/ml}$ (correlation coefficient 0.99). Detailed accuracy data obtained by analysis of a set of spiked plasma samples are listed in Table I. The coefficient of variance (C.V.) was less than 3% over the whole analytical range and the deviations were less than 10%. Detailed precision data obtained by analysis of a set of spiked plasma samples are listed in Table II. The coefficient of variance (C.V.) for both, intra-day and inter-day data were less than 3% over the whole analytical range.

When CsA was added to blank plasma, over 90% was recovered with the 2% HCl methanol extraction in the automatized ASPEC XL system and analyzed by the optimized procedure (Table II).

Table II shows the inter-day and intra-day precisions which were less than 3% for 0.050, 0.125 and 1.000 $\mu\text{g/ml}$ samples.

The previous results shown that the developed method has a good accuracy, recovery and precision, than other existing methods described before [15], but with the main advantage of the increased throughput due to the monolithic column which allows the decrease on the retention time (2.4 min) in comparison with other methods described before [10,11] and the automated SPE system in a sequential input. This input consists in a total extraction time of 6 minutes, while the LC-MS system is analyzing the previous sample. This total and real time

of 6 minutes/sample allows the system to analyze 10 samples/hour. At this point the blood samples under this study were analyzed following this possibility.

Other methods has been described in literature with the similar purpose [16]. It has been shown that the use of short monolithic silica columns coupled to mass spectrometry provides reduced analytical run times for the purpose of metabolite identification from different samples compared to conventional analytical chromatography, with no loss in chromatographic performance.

3.3. Application of the new method

Previous studies have shown that the absorption of oral CsA was shown to be slow and incomplete [17]. Peak blood CsA concentrations occurred between 1 and 8 hours after dosing. The peak blood concentration ranged from 0.165 to 3.431 µg/ml. Therefore our method is perfectly up to be used in pharmacokinetic studies after oral administration of CsA. This is the beginning for further studies that will be developed in the future.

4. Conclusions

The results shown in this report describe a simple and sensible analytical method for the CsA in biological samples after oral administration. The monolithic type column Chromolith decreased the retention time of the CsA from 12 minutes with phenyl columns [10] into 2.4 minutes, and in combination with the highly automated nature of the method significantly improved the sample analysis throughput (6min/sample). The LC-API-ES-MS method can detect samples in the range of 0.015 to 5.000 µg/ml with linear response.

Acknowledgments

This work was supported in part by the Intramural Research Program of the National Institute on Aging/NIH (IWW). We would like to thank Dra. Danuta Siluk for her help in preparing this manuscript and to Dr. Michael Bayhamn for his technical support in the LC-MS system. We would like to thank Merck for the kind gift of the Chromolith Performance RP 18-e column. We also thank the Complutense University and Madrid Community Administration and its research group number 910939.

References

1. Hjerten S, Liao JL, Zhang R. *J Chromatogr* 1989;473:273.
2. Hjerten S, Li YM, Liao JL, Mohammad J, Nakazato K, Petterson G. *Nature* 1992;356:810.
3. Minakuchi H, Nakanishi K, Soga N, Ishizuka N, Tanaka N. *Anal Chem* 1996;68:3498.
4. Minakuchi H, Nakanishi K, Soga N, Ishizuka N, Tanaka N. *J Chromatogr* 1997;762:135.
5. Plumb R, Dear G, Mallet D, Ayrton J. *Rapid Commun Mass Spectrom* 2001;15:986. [PubMed: 11400208]
6. Allanson JP, Biddlecombe RA, Jones AE, Pleasance S. *Rapid Commun Mass Spectrom* 1996;10:811.
7. Faulds D, Goa KL, Benfield P. *Drugs* 1993;45:953. [PubMed: 7691501]
8. Tuominen J, Suortti T, Ishikawa K, Lundell J, Koga Y. *Rapid Commun Mass Spectrom* 1998;12:1085. [PubMed: 9737014]
9. Holt DW, Johnston A, Kahan BD, Morris RG, Oellerich M, Shaw LM. *Clin Chem* 2000;46:872. [PubMed: 10839783]
10. Van Tellingen O, Kemper M, Tijssen F, Van Asperen J, Nooijen WJ, Beijnen JH. *J Chromatogr B Biomed Sci Appl* 1998;20:251. [PubMed: 9869389]
11. Safarik K, Brozmanova H, Bartos V, Jegorov A, Grundmann M. *Clin Chim Acta* 2001;20:165. [PubMed: 11498082]
12. Simpson J, Zhang Q, Ozaeta P, Aboleneen H. *Ther Drug Monit* 1998;20:294. [PubMed: 9631926]

13. Guryča' V, Kieffer-Jaquinod S, Garin' J, Masselon C. *Anal Bioanal Chem* 2008;392:1291. [PubMed: 18949466]
14. Zhou L, Tan D, Theng J, Lim L, Liu YP, Lam KW. *J Chromatogr B Biomed Sci Appl* 2001;15:201. [PubMed: 11318416]
15. Ptachcinski JR, Venkataramanan R, Rosenthal JT, Buckart GU, Taylor RJ, Hakala TR. *Clin Pharmacol Ther* 1985;38:296. [PubMed: 3896612]
16. Dear G, Plumb R, Mallett D. *Rapid Com Mass Spect* 2000;15(2):152.
17. Lee YJ, Chung SJ, Shim CK. *J Pharm Biomed Anal* 2000 Feb;22(1):183. [PubMed: 10727138]

Figure 1.
Chemical structure of Cyclosporin A.

Figure 2.
Chromatographic profile of cyclosporin A (CsA) after optimization of the chromatographic system.

Table I

Accuracy of the LC-ESI-MS method (n=3)

Theoretical concentration(g/ml)	Measured concentration(g/ml)	C.V. (%)	Accuracy(%)	Deviation(%)
0.050	0.046	4.70	91.86	−8.14
0.125	0.124	2.07	99.88	−0.12
0.250	0.259	1.92	103.86	+3.86
0.500	0.497	3.60	99.43	−0.57
1.000	1.092	4.30	109.18	+9.18

Table II

Recovery (%) of CsA (n=3) from human plasma and precision of the LC-ESI-MS method

Concentration($\mu\text{g/ml}$)	Recovery (% $\pm$ SD)	Precision (C.V. %)	
		Intra-day (Average)	Inter-day (Average)
0.050	94.58 $\pm$ 2.48	(1.68, 1.69, 1.73) 1.70	(2.62, 2.59, 2.64) 2.62
0.125	92.37 $\pm$ 6.46	(2.06, 2.10, 2.06) 2.07	(3.10, 2.97, 3.21) 3.09
1.000	92.39 $\pm$ 4.73	(1.93, 1.79, 1.80) 1.80	(2.86, 2.97, 2.60) 2.80
		<i>Mean Average (CV %)</i>	
		1.84	2.81