Further research is of paramount importance if trends in work-related compensation claims in the UK follow those reported from the US.

References

1. Hayward AC, Bradley MJ, Burke FD. Primary care referral protocol for carpal tunnel syndrome. *Postgrad Med J* 2002; **78**: 149–52.
2. Ellis RA, Novak CB, Mackinnon SE, Cheng CJ. Workers' compensation, return to work, and patient satisfaction after carpal tunnel decompression. *Am J Orthop* 2007; **36**: E63–6.

COMMENT ON

doi 10.1308/003588408X285720

PJ Tomlinson, WJ Harries. The use of a surgical glove in the prevention of tourniquet-associated chemical burns. *Ann R Coll Surg Engl* 2008; **90**: 255

doi 10.1308/003588408X321846

Prevention of chemical burns

D BAYNE

Department of Plastic Surgery, The Mountbatten Plastic Surgery Unit, Royal Hospital Haslar, Gosport, Hampshire, UK

CORRESPONDENCE TO

D Bayne, Department of Plastic Surgery, The Mountbatten Plastic Surgery Unit, Royal Hospital Haslar, Gosport, Hampshire PO12 2AA, UK. E: paperscorrespondence@inbox.com

The authors' tip to prevent chemical burns fails to emphasise that alcoholic skin preparation is the greatest risk factor, yet is exchangeable for the much safer aqueous povidone-iodine.¹ The chemical burn below the unprotected tourniquet is well documented and is still encountered in our practice as plastic surgeons.²

The Atlantech tourniquet drape (Arthrocare) costs about £3.50 and can protect large thighs and upper limbs providing an adhesive seal to the skin, integrated padding and fold-over protection of the tourniquet from soiling. A pair of surgical gloves costs about £1.15 with a less effective elastic, there is also the possible disadvantage of a venous tourniquet effect when the pneumatic tourniquet is deflated.

We suggest that Bruner's *Ten Rules for Safe Tourniquet Use* should be updated to include:^{3,4} (i) using an occluding dressing or drape to prevent fluid spread into the padding or onto the tourniquet; and (ii) a new eleventh rule to avoid alcoholic skin preparations for paediatric and elderly patients and its restricted use on adults.

The patient morbidity and cost to the NHS from adverse tourniquet events will be greatly reduced with adherence to these rules.

References

1. Edwards PS, Lipp A, Holmes A. Preoperative skin antiseptics for preventing surgical wound infections after clean surgery. *Cochrane Database Syst Rev* 2004; **3**: CD003949.
2. Dickinson JC, Bailey BN. Chemical burns beneath tourniquets *BMJ* 1988; **294**: 1513.
3. Bruner JM. Safety factors in the use of the pneumatic tourniquet for haemostasis in surgery of the hand. *J Bone Joint Surg Am* 1951; **33**: 221–4.
4. Braithwaite I, Klenerman L. Burns under tourniquets – Bruner's ten rules revisited. *J Med Def Unions* 1996; **12**: 14–5.

AUTHORS' RESPONSE

doi 10.1308/003588408X358738

PJ TOMLINSON, WJ HARRIES

Department of Orthopaedic Surgery, Frenchay Hospital, Bristol, UK

CORRESPONDENCE TO

PJ Tomlinson E: petetomlinson1@hotmail.com

Although it is well recognised that tourniquet-associated chemical burns occur more commonly with alcohol, rather than aqueous-based preparations, this complication is not exclusive to such solutions. Nahlieli *et al.*¹ described three cases of burns occurring in association with povidone-iodine skin preparation fluids as analogous to an acute accelerated decubitus ulcer, arising as a result of a combination of irritation, maceration, friction and pressure.

Commercially available tourniquet covers such as the Atlantech drape (Arthrocare) are often superior to non-commercial alternatives, having been specifically designed for the purpose. However, in the climate of a publicly funded health-care system, such commercial covers are not always obtainable. The purpose of our article was to highlight a viable alternative that is cheap and readily accessible in theatre. The surgical glove shown in the article (see Fig. 2) is a non-sterile latex glove. These retail at £23.40 (excluding VAT) for 1000 gloves. This equates to £0.02 per use as opposed to £1.15 as suggested.

With respect to Bruner's rules regarding safe tourniquet use,^{2–4} we accept that the avoidance of alcohol-based skin preparations should be included. However, using some form of barrier to occlude a tourniquet in order to prevent the soaking of wool already forms one of the ten rules.

References

1. Nahlieli O, Baruchin AM, Levi D, Shapira Y, Yoffe B. Povidone-iodine related burns. *Burns* 2001; **27**: 185–8.
2. Kutty S, McElwain JP. Padding under tourniquets in tourniquet controlled surgery: Bruner's ten rules revisited. *Injury* 2003; **33**: 75.
3. Bruner JM. Safety factors in the use of the pneumatic tourniquet for haemostasis in surgery of the hand. *J Bone Joint Surg Am* 1951; **33**: 221–4.
4. Braithwaite I, Klenerman L. Burns under tourniquets – Bruner's ten rules revisited. *J Med Def Unions* 1996; **12**: 14–5.