Published in final edited form as:

Eur J Pediatr. 2008 March ; 167(3): 287–291. doi:10.1007/s00431-007-0477-0.

Ventilator assessment of respiratory mechanics in paediatric intensive care

Gopinathannair Harikumar, Anne Greenough, and Gerrard F Rafferty

King's College London, MRC-Asthma Centre Division of Asthma, Allergy and Lung Biology, Kings' College London School of Medicine at Guy's, King's College and St Thomas' Hospitals London, UK

Abstract

Many modern “paediatric” mechanical ventilators have in-built features for estimation of respiratory mechanics which could be useful in the management of ventilated infants and children. The aim of this study was to determine if such measurements were reproducible and accurate. Ventilator (Draeger Evita 4) displayed compliance (Cvent) and resistance (Rvent) values were assessed and compared to the results of respiratory system mechanics (respiratory system compliance (Crs) and resistance (Rrs)) measurements obtained using a single breath occlusion technique. Seventeen children (median age 5.1; range 0.3 to 16 yrs) were studied on 24 occasions. The mean coefficients of variations for the techniques were similar (Cvent 13%; Crs 11%; Rvent 16%; Rrs 14%). The mean (SD) Crs (22.8 (12.3) ml/cmH₂O) did not differ significantly from Cvent (22.1 (12.7) ml/cm H₂O) but the mean Rrs 21.0 (12.7) cmH₂O/l/sec was significantly higher than the mean Rvent 32.0 (32.0) cmH₂O/l/sec ($p=0.03$). Bland and Altman analysis demonstrated a mean difference of -10.94 cmH₂O/l/sec (SD 24.1) between Rrs and Rvent; the agreement between Rrs and Rvent decreased as Rrs increased ($p=0.008$).

Conclusions: Ventilator assessment of compliance, but not resistance, using the Evita 4 is reproducible and reliable.

Keywords

Respiratory mechanics; intensive care; ventilation; children; single breath lung mechanics

INTRODUCTION

The respiratory system status of ventilated paediatric patients is usually monitored by clinical assessment and blood gas measurements. Measurement of respiratory mechanics in ventilated children, however, could provide useful information about the physiological status of the respiratory system, the course of disease and response to treatment interventions [15]. Specialised equipment and expertise have been required to make such measurements, which may explain why a survey of paediatric intensive care units in the UK [11] revealed that only two of the 22 responding units used lung function measurements. Modern mechanical ventilators now available, however, have in-built features for estimation of respiratory mechanics, which may mean more widespread use of lung function measurements in the PICU. As a consequence, it is important and the aim of this study to determine if such measurements are reproducible and if they are accurate, by assessing their relationship to the results of standard techniques of assessing respiratory mechanics.

MATERIALS AND METHODS

Children who were mechanically ventilated for more than 24 hours were eligible for entry into this study. The study was approved by the Research Ethics Committee of King's College Hospital NHS Trust and parents gave informed written consent for their children to take part. Measurements were performed in the recovery stage of the subjects' illnesses with the subjects mechanically ventilated at the settings determined by the medical team. All the children were receiving sedation, but not muscle relaxants and all were making spontaneous respiratory efforts. Measurements were performed with the patient supine and haemodynamically stable with stable blood gases. Endotracheal tubes were suctioned 15 to 20 minutes prior to performing the study.

All the children were supported by the Draeger Evita 4 ventilator in a pressure control mode. The ventilator estimated compliance (C_{vent}) by dividing the tidal volume by the difference between the PIP and PEEP (ΔP) and resistance (R_{vent}) by dividing the pressure difference (ΔP) by the expiratory flow. Flow and tidal volume were measured by the ventilator at the expiratory valve and breath-by-breath measurements of compliance and resistance displayed. Ten consecutive breath-to-breath estimates of C_{vent} and R_{vent} displayed by the ventilator were recorded and the mean was calculated.

Respiratory system compliance (C_{rs}) and resistance (R_{rs}) were then measured using the single breath occlusion technique. Flow was measured using a pneumotachograph (4500 series, Hans Rudolf inc, Kansas, USA) and differential pressure transducer (MP45, Validyne Corporation, Northridge CA, USA). The pneumotachograph was attached to a manually operated, three-way sliding valve (2820 Series, Hans Rudolf inc, Kansas, USA) and inserted between the ETT and the ventilator circuit. Airway pressure was measured from a side port on the pneumotachograph using a differential pressure transducer (MP 45, Validyne Corporation, Northridge CA, USA). The signals from both transducers were amplified (CD280 carrier amplifier, Validyne Corporation, Northridge CA, USA) and displayed in real time on a computer running Labview software (version 5.1, National Instruments, Austin TX, USA) with 100 Hz analog to digital sampling (DAQ 16XE-50, National Instruments, Austin TX USA). Tidal volume was obtained by digital integration of the respiratory flow signal by the software. The three-way valve was actuated at end inspiration, briefly occluding the airway, evoking the Hering Breuer inflation reflex, and subsequently diverting the resulting passive exhalation to atmosphere. The resulting relaxed expiratory flow volume curve and the pressure time trace were analysed to calculate the expiratory time constant (Trs), C_{rs} and R_{rs} . C_{rs} was calculated by dividing the change in airway pressure by expired volume. Trs was measured as the slope of a best fit line fitted to the expiratory flow volume curve and R_{rs} by dividing C_{rs} by Trs . Occlusions were maintained for at least 400 ms. Only those with a pressure plateau of at least 100 ms were used. The flow volume curve was analysed using the software and only those with a linearity of >0.99 for at least 40% of the curve were used. Eight to ten occlusions were recorded and the mean value calculated. At least 10 breaths were allowed between occlusions. The resistances of the pneumotachograph and slide valve were subtracted from the total measured resistance to obtain R_{rs} .

Statistical analysis

Data were tested (Shapiro-Wilk) and shown to be normally distributed, hence differences were assessed for statistical significance using a paired t-test. The level of agreement between the techniques was assessed using Bland and Altman analysis [3] and Spearman Rank analysis used to further assess this relationship.

RESULTS

Seventeen infants and children (mean (SD) age 6.3 (6.0) years (Table 1) were measured on 24 separate occasions are reported. Their mean (SD) duration of ventilator days was 4.3 (1.8). Fifteen patients were intubated using cuffed ETT and had no leak, ie no difference between inspiratory and expiratory tidal volumes measured from the pneumotachograph. The other two children had less than 5% difference between inspiratory and expiratory tidal volumes as measured from the pneumotachograph.

The mean intra-subject coefficients of variation were similar for both techniques for compliance (Cvent 13% and CRS 11%) and resistance (Rvent 16% and Rrs 14%). There was no statistically significant difference between the mean (SD) Cvent 22.1 (12.7) ml/cmH₂O and Crs 22.8 (12.3) ml/cmH₂O ($p=0.4$), but there was a statistically significant difference between the mean Rvent 32.0 (32.0) cmH₂O/l/sec and Rrs 21.0 ((SD) 12.7) cmH₂O/l/sec ($p=0.03$) (Table 2). There was a mean difference of 0.7 ml/cmH₂O (SD 4.2) between the Crs and Cvent (Fig 1) and -10.94 cmH₂O/l/sec (SD 24.1) between the Rrs and Rvent results. The agreement between Rrs and Rvent tended to decrease as Rrs increased (Fig 1); with a significant association between the mean of Rvent and Rrs and mean difference between Rvent and Rrs ($r=-0.53$, $p = 0.008$).

DISCUSSION

We have demonstrated that the compliance and resistance results estimated by the ventilator were reproducible, but only the compliance results were similar to those determined by single breath mechanics. We used the single breath occlusion technique, as this can be easily applied in ventilated infants and does not require passage of balloon catheters. Key to obtaining accurate measurements using this technique is achieving relaxed expiration. In the current study, all the children were receiving sedation. Careful examination of the resulting flow volume loop using the criteria defined above was performed to ensure relaxed expirations had been obtained following airway occlusions. Other techniques have been employed to evaluate the respiratory mechanics of ventilated infants and children, including passive expiration to PEEP [17] and the modified Mead-Whittenbeger technique [13]. We have previously shown [12] that measures of compliance and resistance were similar for all techniques. The patients studied had a variety of diagnoses. Thirty percent had a primary lung disease but others had lung involvement due to generalised sepsis, chest wall oedema, ascites or increased post operative intra abdominal pressure affecting their respiratory status. Thus we were able to study children with a range of lung function abnormalities and hence our results are generalisable.

Respiratory diseases were the primary or admitting diagnosis in only 30% of subjects, but respiratory system involvement was present in the majority of our patients. Most children were studied in the recovery stage of their illness, that is, shortly before extubation and as a consequence few had very abnormal Crs and Rrs results (Table 2). It is important to note, however, that the children had a range of Crs and Rrs results and we were able to demonstrate that the difference between Rrs and Rvent increased as resistance increased, indicating that in patients in the acute phase with greater degrees of underlying lung disease a greater discrepancy between measurements would exist

The patients studied in the current study had little or no leak around the endotracheal tube. Endotracheal tube leak, however, can occur in ventilated neonates and children, especially when straight, uncuffed ETTs are used [2, 10]. The influence of the ETT leak on Crs and Rrs has been investigated [8-10, 16], demonstrating the presence of a leak around the endotracheal tube leads to volume loss and higher airflow which in turn affect the accuracy

of measurements of compliance and resistance respectively. The ventilatory mode compound the errors produced by endotracheal tube leak [8-10, 16]. Under such circumstances the accuracy, and hence reliability, of ventilator derived measurements would be further affected.

Care was taken to avoid changing the physiological status of the patients during the study. The three-way sliding valve and pneumotachograph were inserted into the ventilator circuit quickly and with minimal disruption to the patient's ventilatory support. Similarly, end inspiratory airway occlusions were brief, approximately 400ms, and the valve returned to its standby position immediately at the end of expiration. No changes in blood pressure, respiratory rate or blood gas levels were observed during equipment insertion or airway occlusion. In addition, the close agreement between Crs and Cvent suggest Crs measurement had minimal influence on patient status.

Single breath occlusion involves passive measurements, whereas the ventilator makes dynamic measurements, thus we did not expect to see the same results, but, if the ventilator assessments were reliable then the results should have been similar to those of single breath mechanics. Although CRS and Cvent results were similar, there was a highly significant difference between the Rvent and RRS results and most importantly the differences increased at higher resistance values, reflecting inaccuracy of the ventilator measurements.

We only assessed one ventilator type, but previous studies have demonstrated the inaccuracy of ventilator derived measurements, particularly with the Servo 300, which also measures tidal volume at the expiratory valve, [4, 5, 14]. Castle et al [5] demonstrated that the expired tidal volume displayed by Servo 300, after correcting for gas compression, underestimated the true expired volume by up to 64% in the smallest infants but overestimated by as much as 29% in older children. We now demonstrate using the Evita 4 estimates of respiratory resistance are inaccurate.

Randomized controlled studies are required to evaluate whether routine assessment of these measurements of respiratory function improve the quality of treatment and outcome in children receiving mechanical ventilation. The value of such measurements in predicting the outcome of weaning and extubation has been limited [7] [1] [6], but respiratory system compliance and resistance, provide information about underlying lung pathology and the load that is imposed on the respiratory muscles. If there are improvements in the accuracy and range of measurements available including common weaning indices and measures of respiratory muscle function, then ventilator derived measurements may become of increasing usefulness.

We conclude compliance estimated by the Draeger Evita 4 mechanical ventilator is reproducible and reliable, and thus could provide a useful alternative to standard measurements in ventilated children. Our results, however, emphasize resistance estimates using this ventilator should be interpreted with caution. Many modern mechanical ventilators provide measures of respiratory mechanics, we would recommend, however, exercising caution before using such information in research and particularly if used in clinical decision making and patient management. Ventilator derived measurements should be compared to those of a standard technique before utilizing such results for research or clinical applications.

Acknowledgments

Funding: Dr Harikumar was supported by the Wellcome Trust.

ABBREVIATIONS

Cr_s	Respiratory system compliance
C_{vent}	Ventilator estimated compliance
PIP	Peak inflating pressure
PEEP	Positive end expiratory pressure
R_{vent}	Ventilator estimated resistance
R_s	Respiratory system resistance
Tr_s	Expiratory time constant

REFERENCES

1. Baumeister BL, el-Khatib M, Smith PG, Blumer JL. Evaluation of predictors of weaning from mechanical ventilation in pediatric patients. *Pediatr. Pulmonol.* 1997; 24:344–352. [PubMed: 9407568]
2. Bernstein G, Knodel E, Heldt GP. Airway leak size in neonates and autocycling of three flow-triggered ventilators. *Crit. Care Med.* 1995; 23:1739–1744. [PubMed: 7587241]
3. Bland JM, Altman DG. Statistical methods for assessing agreement between two methods of clinical measurement. *Lancet.* 1986; 1:307–310. [PubMed: 2868172]
4. Cannon ML, Cornell J, Tripp-Hamel DS, Gentile MA, Hubble CL, Meliones JN, Cheifetz IM. Tidal volumes for ventilated infants should be determined with a pneumotachometer placed at the endotracheal tube. *Am J Respir Crit Care Med.* 2000; 162:2109–2112. [PubMed: 11112123]
5. Castle RA, Dunne CJ, Mok Q, Wade AM, Stocks J. Accuracy of displayed values of tidal volume in the pediatric intensive care unit. *Crit. Care Med.* 2002; 30:2566–2574. [PubMed: 12441771]
6. Jubran A, Tobin MJ. Passive mechanics of lung and chest wall in patients who failed or succeeded in trials of weaning. *Am J Respir Crit Care Med.* 1997; 155:916–921. [PubMed: 9117026]
7. Khan N, Brown A, Venkataraman ST. Predictors of extubation success and failure in mechanically ventilated infants and children. *Crit. Care Med.* 1996; 24:1568–1579. [PubMed: 8797633]
8. Kondo T, Matsumoto I, Lanteri CJ, Sly PD. Respiratory mechanics during mechanical ventilation: a model study on the effects of leak around a tracheal tube. *Pediatr. Pulmonol.* 1997; 24:423–428. [PubMed: 9448234]
9. Kuo CY, Gerhardt T, Bolivar J, Claure N, Bancalari E. Effect of leak around the endotracheal tube on measurements of pulmonary compliance and resistance during mechanical ventilation: a lung model study. *Pediatr. Pulmonol.* 1996; 22:35–43. [PubMed: 8856802]
10. Main E, Castle R, Stocks J, James I, Hatch D. The influence of endotracheal tube leak on the assessment of respiratory function in ventilated children. *Intensive Care Med.* 2001; 27:1788–1797. [PubMed: 11810124]
11. Manczur TI, Greenough A, Rafferty GF. Survey of current practices of weaning from mechanical ventilation in paediatric intensive care. *British Journal of Intensive Care.* 2000; 10:114–120.
12. Manczur TI, Greenough A, Rafferty GF, Milner AD. Measurement of pulmonary mechanics in the paediatric intensive care unit: a comparison of techniques during pressure- and volume-limited ventilation. *Pediatr. Pulmonol.* 2000; 30:265–268. [PubMed: 10973046]
13. Naik S, Greenough A, Giffin FJ, Baker A. The effect of changes in PEEP pressure on gas exchange - influence of baseline lung function abnormalities. *British Journal of Intensive Care.* 1996; 62:82–88.
14. Neve V, Leclerc F, Noizet O, Vernoux S, Leteurtre S, Forget P, Sadik A, Riou Y. Influence of respiratory system impedance on volume and pressure delivered at the Y piece in ventilated infants. *Pediatr Crit Care Med.* 2003; 4:418–425. [PubMed: 14525635]
15. Newth CJ, Stretton M, Deakers TW, Hammer J. Assessment of pulmonary function in the early phase of ARDS in pediatric patients. *Pediatr. Pulmonol.* 1997; 23:169–175. [PubMed: 9094724]

16. Nikischin W, Lange M. Correction of compliance and resistance altered by endotracheal tube leaks. *Pediatr Crit Care Med*. 2003; 4:344–352. [PubMed: 12831418]
17. Storme L, Riou Y, Logier R, Dubos JP, Kacet N, Rousseau S, Lequien P. A new application of an old method for respiratory mechanics measurements: the passive inflation method in newborn infants during pressure-controlled ventilation. *Pediatr. Pulmonol*. 1994; 18:244–254. [PubMed: 7838624]

Fig 1. Bland and Altman analysis of Crs and Cvent and Rrs and Rvent. The mean difference and ($\pm 2SD$) are shown

Table 1

Age, weight, height and diagnosis of the patients in the study

Study no	AGE (years)	WEIGHT (kg)	HEIGHT (cm)	DIAGNOSES
1	8	30.4	130.6	Cerebral palsy, Aspiration pneumonia
2	16	38	140	CF, Pneumonia
3	0.5	5.1	65	Sepsis, ARDS, Liver failure Pulmonary Haemorrhage
4	15	65	165	Cardiomyopathy
5	0.8	12.7	78.5	Meningitis
6	5.2	20	110	Sepsis, Multiorgan failure, ARDS
7	15	50	160	Head Injury
8	1.5	10.7	76	Hepatic failure, Transplant
9	5.1	14	98	Encephalopathy
10	1	9.8	78	Liver Transplant
11	12	41	160	Hepatic failure
12	15	50	152	Gastro-intestinal Haemorrhage
13	1.2	14	82	LRTI, Adenovirus
14	1.2	13	77	Liver Transplant
15	5.5	22	115	Liver transplant
16	0.33	6.8	66	Head Injury
17	3	18	98	Liver transplant

Table 2

Compliance and resistance results obtained from the ventilator and using single breath lung mechanics

Study no	Cvent (ml/cmH ₂ O)	Cr _s (ml/cmH ₂ O)	Rvent (cmH ₂ O/l/sec)	R _{rs} (cmH ₂ O/l/sec)
1	30.19	25.89	25.63	11.59
2	45.68	48.04	11.85	7.72
3	22.27	24.15	12.24	8.82
4	29.99	28.12	9.95	11.81
5	29.13	20.90	10.38	14.31
6	5.59	6.75	55.55	47.70
7	20.84	26.35	10.28	8.89
8	10.30	13.06	30.45	32.20
9	25.40	19.48	65.59	22.86
10	4.16	5.29	131.07	36.02
11	5.54	7.49	80.38	41.48
12	7.55	7.20	23.73	33.79
13	14.79	16.99	19.18	18.09
14	19.16	17.64	11.64	30.57
15	30.46	30.00	13.15	9.31
16	27.95	34.70	20.08	18.89
17	36.39	44.03	12.94	9.43
18	10.30	19.92	30.45	21.87
19	35.04	35.64	6.13	8.33
20	39.81	34.99	7.95	12.04
21	43.73	40.97	16.26	11.36
22	9.54	11.09	94.05	46.52
23	13.48	13.55	52.04	25.54
24	13.28	15.33	16.27	15.56
Mean (SD)	22.1 (12.7)	22.8 (12.3)	32.0 (32.0)	21.0 (12.8)