Microwave Oven for Melting Laboratory Media

CHARLES W. HANSON* AND WILLIAM J. MARTIN

Clinical Laboratories, Microbiology Section, Department of Pathology, University of California at Los Angeles, Center for the Health Sciences, Los Angeles, California 90024

Received for publication 21 December 1977

A microwave oven was used to melt microbiological media, resulting in a substantial reduction of heat generation and considerable savings in time.

The media kitchen is often under pressure from other sections of the laboratory to reduce the excessive amounts of heat generated from autoclaves, steam ovens, and stoves. Furthermore, it is often necessary to heat less commonly used tubed media at the bench (1), resulting in additional heat generation and uncomfortable working conditions. The introduction and use of a microwave oven in our laboratory resulted in a partial solution to the problem of heat generation. Using an Amana Radarange, model RC-7PB (2450 MHz) with inside dimensions of 30 by 23 by 32 cm, we demonstrated to our satisfaction that a microwave oven could effectively replace most operations carried out under steam or over a gas burner, with the distinct advantage of producing substantially less heat and saving considerable amounts of time. In addition, Latimer and Matsen have recently demonstrated how a microwave oven could be used for decontamination procedures in the laboratory (4).

For purposes of dissolving dehydrated agar media preparatory to autoclaving, 1 liter of brain heart infusion broth containing 2.5% agar (Difco Laboratories, Detroit, Mich.) (2, 3) was thoroughly dissolved after 7 min of heating in the microwave oven. Routinely, 100-ml portions of such media are dispensed into 200-ml-capacity square bottles (Kimble, Vineland, N.J.), autoclaved, stored at 4°C, and melted as needed. Approximately 25 to 30 min are required to melt 1 liter (10 bottles) of agar medium in a steam oven, whereas the same result can be achieved in 12 to 14 min in a microwave oven. However, melting conditions tend to be uneven under continuous microwave exposure, even though the Amana model RC-7PB is equipped with a "stirrer fan" designed to distribute the microwaves evenly throughout the oven. This problem of uneven heating was overcome by the use of intermittent periods of microwave exposure, followed by short periods of heat equilibration. For example, 10 bottles, each containing 100 ml of agar medium (2, 3), were melted by using four 3-min microwave exposure periods. Each exposure

was followed by 90 s of heat equilibration, during which time the tray holding the bottles was rotated 90° with gentle agitation. Visual inspection of the bottles between each exposure period determined which bottles showed complete melting, and these were removed before the next exposure period was initiated. Some microwave ovens come equipped with a rotating platform. This accessory would eliminate the necessity to manually rotate the bottles in the oven between exposure periods.

Plating media not used with enough regularity to justify pouring plates on a routine basis can be prepared in 500-ml amounts, and 15-ml portions can be dispensed to screw-capped tubes (20 by 150 mm; Kimble). Tubes of the appropriate media are melted as needed in a boiling-water bath (7 by 8 by 7 inches [ca. 17.78 by 20.32 by 17.78 cm]; Wilson Diagnostics Inc., Glenwood, Ill.) and cooled, heat-labile constituents are added if required, and plates are poured. The procedure requires 15 to 20 min to bring the water bath to boiling and another 5 min to melt the agar. By using a microwave oven, the same procedure can be accomplished within 2.5 min. For this purpose, a 600-ml beaker containing 300 ml of water was brought to boiling in a microwave oven (about 3 min). The tubes of agar medium were added to the boiling water, and the entire system was given 30 s of microwave exposure, followed by a 30- to 60-s period of heat equilibration. Two additional 15-s microwave exposures, separated by a 30-s period of heat equilibration, were sufficient to melt the tubed agar medium.

In anaerobic bacteriology, it is recommended that all liquid media be heated 8 to 10 min in a boiling-water bath to drive off dissolved oxygen (1), particularly if one is not using freshly prepared or reduced media. However, when tubes of broth were heated directly in a microwave oven, problems of uneven heating and boiling were encountered. As described above for the tubed agar medium, tubes of liquid media were immersed in a beaker of water, and the entire

system was brought to boiling (about 5 to 6 min of microwave exposure). The temperature of the tubes was then maintained with periodic 30-s periods of microwave exposure.

Although these experiments were conducted in a microwave oven designed to heat foods, we feel that certain modifications could be incorporated to make it more effective for use in the microbiology laboratory. For example, a rotating-shaker platform and a non-microwave-absorbing mixing system (a plastic version of a Waring blender) would tend to curtail the problem of uneven heating and isolated areas of boiling. It also would be useful to incorporate a programmable timing device so that alternate periods of heating and cooling (to allow for better heat redistribution) could be achieved automatically.

For several months we have been using the procedures described in this report in our anaerobic bacteriology laboratory and have found no deleterious effects on the ability of the media to support adequate bacterial growth. The growth patterns and reactions of organisms used to quality control plating media, broth media, and agar dilution susceptibility testing media have remained unchanged from what they were when the media were melted or heated by more conventional methods. The exact timing of microwave exposure will vary depending on the conditions used in a given laboratory. For example,

plastic containers require longer periods of microwave exposure than glass containers. Moreover, the length of exposure time increases with the total volume of media being melted. Exposure time also increases with increased agar content, i.e., a given volume of 1% agar will melt faster than the same volume containing 2.5% agar. Due to these variables, the values given in this report are given only as guidelines, and exact conditions should be worked out empirically depending on containers, volumes, and types of media used for each laboratory. The above notwithstanding, we have found that use of the microwave oven effectively reduces heat output and saves considerable time in the preparation of microbiological media.

LITERATURE CITED

1. Dowell, V. R., and T. M. Hawkins. 1974. Laboratory methods in anaerobic bacteriology, p. 4. *In* CDC laboratory manual, publication no. (CDC) 74-8272. U.S. Government Printing Office, Washington, D.C.
2. Hanson, C. W., and W. J. Martin. 1976. Evaluation of enrichment, storage, and age of blood agar media in relation to its ability to support growth of anaerobic bacteria. *J. Clin. Microbiol.* 4:394-399.
3. Holdeman, L. V., and W. E. C. Moore. 1972. Anaerobic bacteriology manual, p. 124. Anaerobe Laboratory, Virginia Polytechnic Institute and State University, Blacksburg, Va.
4. Latimer, J. M., and J. M. Matsen. 1977. Microwave oven irradiation as a method for bacterial decontamination in a clinical microbiology laboratory. *J. Clin. Microbiol.* 6:340-342.