Published in final edited form as:

Magn Reson Med. 2009 December ; 62(6): 1597–1608. doi:10.1002/mrm.22106.

Complex-Valued Analysis of Arterial Spin Labeling Based FMRI Signals

Luis Hernandez-Garcia^{a,*}, Alberto L. Vazquez^b, and Daniel B. Rowe^c

^a Functional MRI laboratory, University of Michigan, MI 48109, USA

^b Department of Radiology, University of Pittsburgh, PA 15206, USA

^c Department of Biophysics, Medical College of Wisconsin, WI 53226, USA

Abstract

Cerebral blood flow dependent phase differences between tagged and control ASL images are reported in this article. A biophysical model is presented to explain this phase difference as being of vascular origin. An ASL study indicated that the phase difference is largest when the arterial component of the signals is preserved, but is greatly reduced as the arterial contribution is suppressed by post inversion delays or flow crushing gradients. Arterial vasculature imaging by saturation (AVIS) data of activation and hypercapnia conditions, showed increases in phase accompanying blood flow increases.

An ASL FMRI study yielded significant activation by magnitude-only, phase-only, and complex analyses when preserving the whole ASL signal. After suppression of the arterial signal by post inversion delays, magnitude-only and complex models yielded similar activation levels, but the phase-only model detected nearly no activation. When flow crushers were used for arterial suppression, magnitude-only activation was slightly lower and fluctuations in phase were dramatically higher than when post inversion delays were used.

Although the complex analysis performed did not improve detection, a simulation study indicated that the complex-valued activation model exhibits combined magnitude and phase detection power and thus maximizes sensitivity under ideal conditions. This suggests that, as ASL imaging and image correction methods develop, the complex-valued detection model may play a crucial role in signal detection.

Keywords

complex-valued; phase; ASL

Introduction

Until recently, functional magnetic resonance imaging (FMRI) data analysis was concerned only with the magnitude of the image time courses although the data are intrinsically complex-valued. Indeed, most magnetic resonance (MR) images are reconstructed from complex-valued signals collected by quadrature detectors and most reconstruction techniques yield complex-valued images. Yet, the phase information is routinely discarded, retaining only the magnitude images (1). Recently, complex-valued FMRI activation models with combined magnitude and

*Correspondence to: Luis Hernandez-Garcia, Ph.D., FMRI Laboratory, 2360 Bonisteel Blvd., Ann Arbor, Michigan, 48109-2108, (734) 763 9254, hernan@umich.edu.

phase detection have been developed by Rowe and colleagues for blood oxygen level dependent (BOLD) contrast FMRI time series (1,2). Although our mechanism is not BOLD based, we aim to take advantage of these models for ASL FMRI. The complex-valued models are designed to detect task related phase changes (TRPCs) in addition to task related magnitude changes. The complex model's ability to detect TRPCs increases the possible circumstances under which activation is detected for given signal-to-noise ratio (SNR) and contrast-to-noise ratio (CNR). Changes in the signal's phase were described by Menon (3), who put forth a model explaining how the physiological changes that give rise to the BOLD effect manifest themselves in the magnitude and the phase of the signal, particularly in areas of larger venous vasculature. This effect was utilized to reduce the contribution of the further downstream venous vasculature to the magnitude activation (4). Arterial spin labeling (ASL) time series are typically used as indicators of perfusion (a real-valued quantity), but ASL time series can be designed to be BOLD weighted as well (5-7).

The SNR and CNR in ASL techniques are inherently low (8-11), making ASL-based FMRI more challenging than BOLD-based FMRI. For example, Perthen et al. (10) reports SNR maps in the range of 1-2 for ASL, and more than 100 for BOLD data. Yang et al. (11) reports t-scores of approximately 5 for ASL data, while BOLD data produce t-scores greater than 10, reflecting the CNR difference. It is therefore imperative that statistical analyses of ASL data be as efficient and sensitive as possible. Typical ASL analysis (ordinary least squares (OLS) analysis of the subtracted magnitude data) results in reduced estimation efficiency (6) and “wastes” information by using only the magnitude of the signals, as potentially valuable phase information is routinely discarded. Complex-valued analysis techniques should also benefit regression analysis of ASL data. Furthermore, the presence of the inversion labeled ASL data can also induce changes in the phase of the labeled images, also called tagged images, relative to the control images, as we discuss below in greater detail.

In this article we examine the biophysical basis for phase changes in the ASL signal (independent of the BOLD effect related phase changes), and develop an analysis approach for ASL time series that includes generalized least-squares analysis of the complex-valued, undifferenced ASL data. This analysis scheme models the changes in perfusion and BOLD contrast as complex-valued (magnitude and phase) quantities. The model and techniques in this work are demonstrated on continuous ASL data, but are equally applicable to pulsed ASL. Complex-valued analysis of ASL data can potentially increase the activation of functional studies and also help identify vascular sources of noise.

Vector Model of ASL Signals

Pairs of images are acquired in a typical ASL experiment: one is preceded by a tagging period, in which the inflowing arterial spins are inverted (tagged), and a second control image without the inversion label (control). Figure 1 depicts the magnetization vectors for the tissue and blood pool under labeling and control conditions. Within the voxel, a fraction of the spins are in the tissue and their movement is governed by random diffusion. Indistinguishable from tissue is also the isotropic capillary perfusion, although their relaxation rates are slightly different. A smaller fraction of the spins is in the vasculature and their motion is governed by anisotropic, coherent flow (3,12).

In the control images, the coherently-moving vascular spins acquire some net phase, ϕ , that does not get refocused by the imaging pulse sequence. This net phase gain is due to motion in the presence of the imaging magnetic field gradients and has been studied extensively both as a source of artifacts (12,13), and contrast to identify vasculature (14). The net phase gain of the incoherently moving spins (i.e., those of the tissue and capillaries), on the other hand, is zero because of their isotropic motion.

In the tagged case, the arterial magnetization is inverted by the labeling RF pulses and gradually relaxes back along the z-axis toward its equilibrium state. If the arrival time is short enough, the incoming arterial magnetization is still in the negative longitudinal axis at the time the imaging pulses of the imaging sequence are applied. The magnetization vector is tipped onto the transverse plane by the imaging RF pulses, but 180 degrees from the tipped location of the control case (for a visual example of this effect, please see Figure 3 of reference (15)). The net phase gain of the vascular spins is then $\phi + 180$. If the arrival time is sufficiently long for the labeled spins to have relaxed past the null point ($M_z=0$), but are not yet fully relaxed, the net phase from the coherent compartment is the same as in the control case. The magnitude of the vascular magnetization vector, however, is shorter than in the control case.

In both control and tagged cases, the tissue's net magnetization vector does not acquire any net phase, as the movement of the spins due to perfusion and diffusion is mostly isotropic. Hence, when there is a significant arterial contribution in the ASL signal, the inversion label should cause the net (vascular plus tissue) magnetization vector to change in both magnitude and phase. When there is not a significant arterial contribution to the signal, only the magnitude of the signal is expected to change.

The phase gained by the coherently flowing vascular spins can vary widely between 0 and 360 degrees depending on their velocity and the gradients used in the acquisition sequence. In order to estimate an upper bound for the size of this effect, let us assume ideal conditions of a 90 degree phase gain in arterial vasculature relative to tissue, 5% of a voxel's volume to be occupied by arterioles and 15% to be occupied by large arteries, and perfect inversion labels. This would produce a phase difference of 6.8 degrees between control and tagged images because of the arterioles, and 22 degrees if the signal from large arteries is also included (not suppressed). It should be noted that the observed effects reported below are much smaller than this theoretical upper bound, but still observable.

It should be noted that in Menon's model for the BOLD effect, the phase differences are produced by local resonance frequency shifts in venous areas due to increases in the amount of oxygenated blood. The BOLD phase differences are a function of the T_2^* weighting of the imaging sequence, and can be minimized by spin echo imaging. While Menon's model focused on the venous contribution to the MR signal and the phase changes associated with neuronal activation, the vector model in this article is mainly concerned with the arterial and arteriolar contribution, as no significant amount of the inversion label reaches the veins, and BOLD effects are assumed to be significantly diminished by spin-echo acquisition. A complete model for T_2^* weighted ASL data, can be built by including both arterial and venous contributions to the phase separately.

In this article, we measure the phase contributions of the vascular compartment to the MR signal in ASL data by gradually suppressing their contribution to the perfusion signal using post inversion delays and diffusion gradient pairs. A significant post-labeling delay allows the labeled signal in blood to reach and exchange into tissue. Diffusion gradient pairs are designed to introduce an intra-voxel phase accrual in coherently moving spins along the direction(s) of the gradient(s) that aims to cancel their overall signal, while stationary spins accrue no net phase. We also examine the phase content of arterial-only images under rest activation and hypercapnia conditions. We then examine the usefulness of complex-valued analysis for ASL-based fMRI with and without vascular suppression.

Methods

All scanning protocols were performed using a 3.0 T Signa LX Excite scanner (General Electric, Waukesha, WI) in accordance to the University of Michigan's IRB regulations.

Arterial Suppression Study

ASL images were collected from four human subjects. The degree of arterial contribution to the signal was varied in each experiment as follows. ASL imaging was done using the pseudo-CASL sequence (16,17) (spin-echo spiral acquisition with TR= 4000 ms, TE =15 ms., slice thickness = 7 mm, FOV = 24 cm, 1 slice, 96 time frames). The labeling pulses consisted of a train of Hanning window shaped pulses (pulse width = 500 μ s, pulse spacing = 290 μ s, flip angle 22.5 degrees, net gradient moment = 3×10^{-5} G/cm/s) applied for 3600 ms. The experiment was repeated using post inversion delays (18) of 0, 1200 and 1800 ms without any flow crushers. The net tagging time was reduced by the amount of the post inversion delay, accordingly. Similarly, a second set of scans was collected with no post-inversion delay, but in the presence of a pair of flow crushers (pulse width = 2 ms, separation = 8.8 ms, amplitude= 0, 1, 2 and 4 G/cm). The VENC values of this pair were *Inf* (i.e., no crushers), 6.6 and 3.3 cm/s along the Y gradient axis.

All images were reconstructed to yield magnitude and phase images. Motion parameters were estimated from the magnitude images using MCFLIRT to ensure that there was no movement greater than half a voxel shift or one degree rotation. The complex-valued time courses were phase centered by subtracting their angular mean (4) to prevent phase wraps. The images were then pairwise subtracted and the mean and variance of the magnitude and phase subtractions were computed. T-statistics were calculated on the magnitude and phase differences. Maps of the correlation between magnitude and phase time courses were computed. Statistical significance of the difference images as well as the correlation images was assessed by the p-values corresponding to the measured statistics.

Arterial Vasculature Studies Under Activation and Hypercapnia Conditions

An activation study was performed using the AVIS (Arterial Vasculature Imaging using Saturation) ASL technique (19). In summary, the AVIS technique images only the blood signal by completely saturating the background tissue signal (gray matter, white matter and CSF) while fresh and fully relaxed blood spins enter the imaging slice. Multiple inversion pulses are then used to maintain the undesired tissue signal nulled while imaging the blood signal after a pre-determined time has elapsed (saturation-to-imaging time). This method was determined to be biased to the arterial volume signal. The imaging parameters of the spin-echo spiral acquisition used were: FOV = 20 cm, Matrix = 40 $\times$ 40 (reconstructed to 64 $\times$ 64), TE = 40 ms, Saturation-to-imaging Time = 800 ms, TR = 4000 ms. Visually cued finger tapping task was used with activation and rest blocks each of 60 s in duration. During this study the subject wore a non-rebreathable mask and 100% air was administered to the subject. A hyperoxic-hypercapnia experiment followed the stimulation experiment. Prior to the start of the acquisition, the administered gas was changed to 100% oxygen and a few minutes were allowed to elapse. A total of 100 images were then acquired using the same parameters as the stimulation experiment. The administered gas was switched to 95% oxygen and 5% carbon dioxide 50 s into the acquisition. This gas mixture was supplied for 150 s and then switched back to 100% oxygen. After this experiment finished, the gas mixture was switched back to 100% air.

ASL fMRI Activation Study

ASL images were collected from another human subject using the same pseudo CASL sequence described for the resting state experiments, while they experienced a visual stimulation paradigm (8 Hz flashing checkerboard: six cycles of 50 s rest – 50 s active). Five slices were prescribed encompassing the visual cortex and 150 time frames were collected. The experiment was performed two additional times using two different methods of arterial signal suppression: a pair of flow crusher gradients (VENC = 3.3 cm/s), and a post inversion delay of 1200 ms. In this article, we refer to the data collected without arterial suppression or

post inversion delays as “whole,” those collected with flow crushers as “crush” and those collected with post inversion delays as “PID.”

Activation maps were calculated by testing three general linear models for the activation experiments. The first is a linear regression model (linear x with linear y) applied to the magnitude-only (MO) data (20,21) (after discarding the phase portion of the data. This model has been shown to be equivalent to a complex-valued model with an unrestricted phase (4). The second is an angular regression model (linear x with angular y) applied to the phase-only (PO) data (22) after discarding the magnitude portion of the data. This model is equivalent to the usual general least squares model for the phase time series when the signal-to-noise (SNR) ratio is high and no phase transitions are present. The third is a complex-valued nonlinear regression model that fully utilizes the magnitude and phase (MP) data (2,23). The details of the models and their underlying distributions are described in the Appendix.

The design matrix utilized for these unsubtracted analyses was composed of a baseline, the BOLD response, an ASL control regressor, and an ASL tag regressor, as previously described (9). Together, they make up the columns of the design matrix X while β and γ are regression coefficients in the magnitude and phase respectively. A contrast $C=(0,0,-1,1)$ was utilized to test hypotheses. For the magnitude-only model, parameter estimates and a likelihood ratio based test statistic are derived from the high SNR normal approximation to Eq. [A3] considering $H_0: C\beta=0$ versus $H_1: C\beta\neq 0$ (4). For the phase-only model, parameter estimates and a likelihood ratio based test statistic are derived from the high SNR normal approximation to Eq. [A4] considering $H_0: C\gamma=0$ versus $H_1: C\gamma\neq 0$ (19). For the complex-valued model, parameter estimates and a likelihood ratio based test statistic are derived from Eq. [A2] considering $H_0: C\beta=0, C\gamma=0$ versus $H_1: C\beta\neq 0, C\gamma\neq 0$ (5,6). These hypothesis statements use the traditional abbreviated notation in which $C\beta=0$ and $C\gamma=0$ are also included in the parameter space for the alternative hypotheses. All hypothesis pairs are to detect a difference between tag and control images.

Simulation Study

To examine the detection properties of these three models under ideal conditions, a simulation study was performed. For this simulation, relevant coefficient values were taken to approximate the values from the real data. The $SNR=\beta_0/\sigma$ was selected to be 130 and the error standard deviation to be $\sigma=1$. In our experience, these values represent typical complex fMRI data well. The $CNR=C\beta/\sigma$ is varied from 0 to 0.75 from left to right in 100 equal steps and Task Related Phase Change (TRPC), $TRPC=C\gamma$ is varied from 0 to 21.5 degrees from top to bottom in 100 equal steps. The same design matrix was used as for the real activation study and $n=150$. The coefficient values were $\beta=(SNR\cdot\sigma, 5, 0, CNR\cdot\sigma)$ in AU, $\gamma=(30, 0.25, 0, TRPC)$ in degrees. A total of 1500 simulated data sets were generated to make power surfaces.

Results

Arterial Suppression Study

The effects of removing arterial signals from resting ASL images in a representative subject are illustrated in Figure 2. Figure 2a shows the results of using post inversion delays for arterial suppression while Figure 2b shows the results obtained using flow crushers. The resting state experiments indicated a small phase difference between control and tagged images when no crushers or post inversion delays were employed. The mean value across the brain of the absolute phase difference was 0.3 degrees, with a standard deviation of 0.4 degrees, but it ranged up to 7 degrees in some regions. Maps of the statistical significance of the differences, expressed as $-\log_{10}(p)$ were calculated (data not shown). This difference was much more prominent in voxels dominated by arterial signals, verified by the anatomical images obtained

for this study. The first row shows the magnitude difference, while the second one shows the phase difference. The bottom row shows the correlation of the unsubtracted magnitude and phase time courses across the slice. Both methods reduced the arterial contribution to the ASL signal, as evidenced by the reduction in magnitude of the subtraction images (top row). All subjects showed significant differences in magnitude and phase, and correlations between magnitude and phase when the arterial signal was preserved. Those effects were reduced by the removal of arterial contribution in all cases.

As expected, introducing arterial suppression techniques reduced the mean gray matter ASL signal (magnitude subtraction of control and tagged images). On average, the inclusion of crushers (VENC of 6.6 and 3.3 cm/s) in the acquisition sequence reduced the magnitude of the ASL subtraction signal by 44.9% and 46.8% indicating the removal of the signals arising from the major arteries in the field of view. The inclusion of post inversion delays of 1200 and 1800 ms resulted in a magnitude reduction of up to 44.8% and 67.8% respectively.

The difference in phase between control and tagged images was dramatically reduced in both cases as well (middle row). However, as the flow crusher gradient strength was increased, a spatial gradient in the phase difference image could be observed along the direction of the crusher gradient. We attribute this to eddy currents and imperfect gradient stability.

The temporal variance of the pairwise (control minus tag) phase differences (averaged over the tissue) increased dramatically when flow crushers were included (600% and 3100% relative to no crushers). The changes were much smaller when post inversion delays were used to suppress the arterial signal (14%, -16%).

Since both the signal's magnitude and phase in the arteries are modulated by the same labeling function (i.e., the control-tag scheme), there is a significant degree of correlation between them. The correlation between magnitude and phase was significantly reduced by the use of post inversion delays and flow crusher gradients (bottom row), supporting an arterial origin for the correlation. The reduction in correlation was much stronger in the case with diffusion crushers, which increase the phase noise and remove the arterial signal altogether. In the case where crusher gradients were used, we attribute the reduction in correlation to the drop in SNR at the arteries. In the case of post inversion delays, however, the arterial signal is not destroyed in the unsubtracted time courses. However, the modulation of the arterial signal by the labeling function is no longer present, as the labeled blood has been cleared the arteries during the post inversion delay, and thus the correlation is also reduced.

Arterial Vasculature Studies Under Activation and Hypercapnia Conditions

As discussed elsewhere (19), the AVIS magnitude data revealed increases in arterial blood flow and volume in the motor cortex during visually cued finger tapping stimulation. Such magnitude increases were also observed globally during the hypercapnia challenge. In the present analysis, phase increases were also observed along with magnitude changes. Time courses extracted from active regions in the motor activation paradigm can be seen in Figure 3a. The average phase time course during the CO₂ challenge can be seen in Figure 3b. Both indicate phase changes closely corresponding to increases in blood flow. These AVIS results lend evidence to support that the phase changes of ASL signals stems from arterial signal contributions.

ASL Activation Study

Only one of the five slices was utilized for analysis. Preprocessing prior to statistical analysis entailed quadratic detrending the magnitude and phase time series then the first four frequencies ideally filtered. The MO, PO, and MP activations were converted to the $-\log_{10}(\text{p-value})$ scale

and thresholded at $p=0.05$ (21). Activations were superimposed on each scan's magnitude of the complex-valued average of controls minus tags. An ROI (outlined by the magenta box) within the visual cortex is defined for reference with additional results. The specific location of the box was chosen to include an area that was highly active for the magnitude-only analysis in all three data cases. Magnitude-only activation in the visual cortex was observed using all three sets of data and presented in Figures 4a-c. Statistically significant activation was detected by the phase-only model in the visual cortex when the arterial signals were preserved (whole data) in Figure 4d but not in the crush and PID cases in Figures 4d and 3e. This is indicative that activation induced perfusion changes are accompanied by changes in the phase difference between tagged and control images. These phase changes in turn are indicative of arterial volume and/or velocity changes during activation. Magnitude and phase activation is clear in the whole case (Figure 4i) and similar in the PID case (Figure 4h) while much reduced in the crush case (Figure 4g). While the whole and PID cases may have benefited from analysis using a complex model, the crush data did not. We attribute this to the increase in phase noise introduced by the crusher gradients (as seen in Figure 5j and Table 1).

The contrasts of parameter estimates in the magnitude-only and phase-only models were computed along with the model variance, in order to investigate the details underlying the activation maps. These are displayed in Figure 5 and the same ROI within the visual cortex (outlined by the black box) is examined. The MP coefficient estimates and their contrasts are nearly identical to those for the MO and PO models and are not shown in Figure 5. In Figure 5c the magnitude-only model, whole data (i.e. acquired without arterial suppression) produced larger contrast between active vs. rest coefficients in the ROI than the flow-crushed data in Figure 5a and both of these cases yielded a larger contrast than the PID data in Figure 5b. Magnitude-only model variance in Figure 5d-f was similar for the three data types but slightly lower for the crushed data in Figure 5d. Thus the pattern of magnitude-only activation of whole data in Figure 4c was larger than flow-crushed in Figure 4a, and PID data in Figure 4a because of differences in the tag vs. control contrast coefficients with little difference in residual model variance in Figures 5d-f.

In the case of flow-crushed data, the phase-only model produced a parameter estimate map (Figure 5g) in which pair of large, positive clusters could be seen qualitatively at the same location as the active region detected by the magnitude-only model (see figures 4a, 4b, and 4c). In the PID and whole data in Figures 5h and 5j, the contrast was negative and smaller than in the flow-crushed data. The residual variance of the phase-only model was much larger for the crush data (Figure 5j) than for the PID or whole data (Figures 5k and 5h, as previously described). While a few pixels in the region were statistically detectable by the PO model in the case of PID data, the use of crusher gradients did not yield statistically significant activation by the phase-only model.

Some phase difference between rest and activation was detected in the PID data, but it was more prominent in the whole data. The MP coefficient estimates and their contrast were nearly identical to those for the MO and PO models. However, the MP residual variance was much larger for the flow-crushed data than for the PID or whole data due to the large phase variation. This is quantitatively examined using average statistics for the three activation models in the 3×3 ROI indicated by the black box in Figure 5 and displayed in Table 1.

From Table 1 it can be seen that when no significant phase contrast is present as in the crush data, the MO model performs better than the MP model that detects magnitude and phase contrast (although the coefficient estimates are similar). In the PID and whole data the MO model performs slightly better than the MP model (although the coefficient estimates are similar). This is due to temporal phase variations in the real data most likely of physiologic

origin that increases the residual variance and decreases detection power. Approaches to reduce the phase noise are the subject of future research and are in the Discussion section.

Simulation Study

Figure 6 displays surfaces for $p=0.05$ thresholded detection power for the MO, PO, and MP models. In these simulations, the MO model exhibits equivalent CNR detection power regardless of TRPC as seen in Figure 6a. The PO model exhibits equivalent TRPC contrast detection power regardless of CNR as seen in Figure 6b. The MP model exhibits combined CNR and TRPC detection power as seen in Figure 6c. When CNR is present and no TRPC is present, the MP model exhibits slightly lower detection power than the MO model while when TRPC is present and no CNR is present, the MP model exhibits slightly lower detection power than the PO model. Hence, the MP model detects MO and/or PO contrast.

Discussion

FMRI data are intrinsically complex-valued, although the phase data are routinely discarded. It is possible to “improve” activation under ideal conditions with the MP model over the MO model by including the phase as in the simulation. The potential improvement with the MP model is due to its ability to detect changes in either the magnitude or the phase of the signal. However, as seen in Figure 6, when there is only a task related magnitude change (phase change), a model that is specifically designed only to detect magnitude changes (phase changes) is slightly better than one that is designed to also detect task related phase changes (magnitude changes) due to the increase in the dimensionality of the contrast space from adding dimensions that contribute zero signal.

Furthermore, our data indicate that there is a phase difference between control and tagged images in arterial spin labeling experiments, and that this phase difference is related to flow since the phase difference increases with activation and also hypercapnia, a known paradigm that causes systemic increases in blood flow. The ASL data also indicate that the phase difference comes from arterial sources rather than capillary or tissue sources and that removing the arterial contributions reduce the phase differences dramatically. Under ideal conditions, these analyses could render ASL FMRI a more competitive alternative to BOLD FMRI by either increasing the detection power of typical ASL FMRI experiments or, more importantly, by selecting acquisition parameters that will increase the ASL signal (by minimizing the use of post-inversion delays or flow-crushing gradients and/or selecting shorter TRs) and adjusting for larger vascular contributions on the post-acquisition data analysis. The current data presented demonstrates a detectable activation signal in the phase of ASL signals. The complex activation model that we are using follows the classical thermal noise model and does not account for physiologic noise in the phase. One approach to reduce their contribution is to include nuisance regressors in the phase design matrix (5). It can be difficult to provide valid regressors for phase, and even when properly modeled, dynamic image warping artifacts still remain, currently making this a less attractive solution (22). However, it has been shown that this phase noise can be significantly decreased to near ideal conditions with the use of a dynamic field mapping method and that phase sensitive FMRI analysis methods can be reliably used (22). The dynamic field mapping method (22) was developed for Cartesian k-space trajectories and need to be extended to spiral k-space trajectories as used for our data.

These results must be viewed in the context of the vascular anatomy and hemodynamics of the brain. Duong and Kim reported arterial CBV of 29% in rats including the capillary fraction (24). Cortical surface arteries have been reported to contribute to the cerebral blood volume approximately 15% in cats (25), 21% in human subjects (26) and their velocity ranges from 10 to 60 cm/s. Penetrating arterioles from the cortical surface into the deeper layer of the cortex have a smaller diameter and run mainly parallel to each other (27) and hence the blood within

them flows in a coherent manner. These penetrating arterioles are particularly dense in layers 1 through 4 of the cortex.

In ASL experiments, the venous contributions are negligible because the inversion label relaxes well before reaching the venous compartment. The arterial contribution is routinely treated as contamination of the perfusion signal and consequently suppressed by using post inversion delays or flow crushing gradients (failure to do so results in overestimation of perfusion in quantitative maps). Hence, the observed ASL signal arises from the labeled water within the arterioles and capillaries as well as those water spins that have exchanged into the tissue.

The use of post inversion delays allows the bulk of the labeled blood to leave the arterial compartment but a fraction always remains in the arterioles because of the arterial dispersion (18). Flow crushers can also be used to reduce the arterial contribution from the ASL signal, but since arteriolar velocities have been reported to be in the range of 1 cm/s (28), the arteriolar signal is unlikely to be suppressed by them. Instead, flow crushers increase the phase gained in the ASL signal at the arteriolar level. In short, the flow crushers and post inversion delays remove the ASL signal from the larger arteries, but the signal from the smaller penetrating arterioles is preserved. These arteriolar contributions to the arterial cerebral blood volume are small, approximately 3-5% of the CBV in humans (29), and are considered part of the blood supply to the voxel of interest (30) and hence are not considered as part of the arterial contamination.

We observed in the resting state data that the phase difference between control and tagged images decreased dramatically as the arterial signal was removed. In the activation case, subtle changes in the phase difference between control and tagged could be detected by the phase-only and magnitude-phase models when the arterial contribution was preserved, but they were much smaller when the arterial contribution was removed by post inversion delays or by crushers.

A potential confound in the correlation analysis is that the SNR of the raw time courses is reduced by the use of flow crushers that null the arterial signal. Reduced SNR would decrease the correlation coefficient of the signals. This is not so much the case of post-inversion delays because the arterial signal is preserved in the unsubtracted data, even if the labeled blood has cleared the arteries. It must be noted that while the total ASL signal (the difference between control and tagged images) is reduced by the clearance of the arterial label during the post inversion delay, the label content in the tissue compartment remains unchanged until the arterial compartment clears. The amount of label in the tissue indeed decreases when the post inversion delay exceeds the transit time, since tagging time is also reduced, but at a slower rate than the arterial compartment (30). Since the magnitude of the subtracted ASL signal was reduced by the crusher gradients and the first post inversion delay (1200 ms.) by approximately the same amount (44, 46, and 45%), we can infer that the post inversion delay of 1200 ms. accounts mostly for the label clearing the arterial compartment, rather than a reduction of tissue label because of shortened tagging time. On the other hand, the signal reduction of 67% due to the 1800 ms. post inversion delay implies that the loss in tissue label could account for the observed reduction of the phase difference.

The use of a pair of flow crushing gradients dramatically increased the variance of the phase over time, as gradient pairs are known to do. This increase in phase noise was very detrimental to the phase-only and magnitude-phase analyses, essentially yielding no active voxels in the visual cortex.

Including post inversion delays to remove the arterial label's contribution to the signal also reduced the sensitivity to activation-related perfusion changes, but activation could still be detected at lower threshold values since the post inversion delays had no significant effect on

the phase noise. We hypothesize that including flow crushers in addition to the post inversion delays should not change the ASL signal, as they both eliminate the arterial compartment. However, we speculate that the phase noise would increase dramatically, as observed in the flow-crushed data.

Activation induced changes in perfusion are accompanied by changes in blood volume (31) and arterial velocity (32,33). Two-photon, bolus injection studies on rats indicated significant changes in transit time from the arteries to the capillaries and venules, but not significant from the arteries to the arterioles (34). It is well known that changes in velocity of blood affect the MR signal's phase, but our data indicate that the arteriolar volume is too small to show large phase changes in the phase-only analysis. While the effect is still observable in the PID data, the number of active voxels is reduced relative to the other models because of the large phase noise, as Figure 5 and Table 1 indicate. However, the magnitude-phase analysis of the PID and whole data did show significant activation in the visual cortex while reducing the number of active voxels outside the visual cortex, indicating that including the phase information into the model was helpful to decrease the activation region and lead to an apparent increase in the focality of the analysis.

Although absolute quantification of perfusion is not necessary for most cognitive psychology fMRI experiments, it is certainly beneficial. Current perfusion models are based solely on the magnitude of the ASL signal and discard the phase. While these quantification models may not be fully efficient from a statistical point of view, it must be noted that they are still valid under the model described in Figure 1 and the phase effects described reported in the present work. Including phase information in quantification of perfusion from the ASL signal is desirable, but not straightforward. Because blood flow is proportional to the product of both blood velocity and blood volume (assuming no changes in vessel length). Since no large changes in transit time have been reported between arteries and arterioles, the phase of the ASL signal can be seen as an indicator of vascular volume, particularly in the label images. The phase information can be used to estimate velocity along the images vascular volume and/or to differentiate the velocity contribution to the increases in CBF. This might be advantageous considering some of the current methods to measure cerebral blood volume (CBV) and mean transit time (MTT). For example, dynamic susceptibility contrast (DSC) bolus injection experiments quantify perfusion as cerebral blood volume (CBV) over the mean transit time (MTT) through the capillary. However, a robust implementation of velocity estimates would require encoding velocity in all cardinal directions. More problematic is that this calculation is likely to be confounded by multiple vessels in a voxel with different orientations. Thus, a more complete model that includes the phase of the signal in order to efficiently use all the information available is beyond the scope of this work but currently under investigation.

Finally, while preserving the whole arterial signal (by not using flow crushers or post-inversion delays) prevents the classical quantification of perfusion, it enhances the ability to detect neuronal activity. Thus, when designing an ASL based fMRI experiment, investigators should consider whether their main goal is to detect or quantify changes in blood flow. In experiments where CNR is low, it may be useful (or even necessary) to sacrifice quantification for the sake of sensitivity. For example, a multi-session study where BOLD imaging is problematic because of scanner drifts would still benefit from non-quantitative ASL techniques. Although outside the scope of the present work, ASL techniques that leverage transit time changes to amplify activation responses (35,36) should also benefit from complex-valued analysis.

In summary, there is valuable phase information in the phase of ASL based fMRI time series that could be used to increase activation if modeled properly. The complex activation model that we are using follows the classical thermal noise model and does not account for physiologic noise in the phase. The source of this phase information is arterial in nature. Since there is

arteriolar signal still present even after arterial suppression, the phase information is greatly reduced but still detectable in some of the pixels of the active region (see Figure 4). One must keep in mind, though, that if arterial suppression is achieved through diffusion crushers, the phase noise overwhelms the signal. Thus, post inversion delays are a preferable method of suppression when analyzing the phase of ASL images.

Acknowledgments

This work was funded by the National Institute of Health (R01 EB004346, R01EB007827, and R01EB00215)

Appendix: Magnitude-only, Phase-only, and Magnitude-Phase Activation Models

Here, we describe the activation models which are necessary for the maximum likelihood estimation process. These three models can be described from a complex-valued perspective. Let the complex-valued measurement in a voxel $y_{Rt} + iy_{It}$ at time t be described as

$$\begin{pmatrix} y_{Rt} \\ y_{It} \end{pmatrix} = \begin{pmatrix} \rho_t \cos \theta_t \\ \rho_t \sin \theta_t \end{pmatrix} + \begin{pmatrix} \eta_{Rt} \\ \eta_{It} \end{pmatrix}, \quad [A1]$$

where ρ_t and θ_t are the true magnitude and phase while η_{Rt} and η_{It} are additive statistically independent normally distributed noise with zero mean and variance σ^2 . From Eq. [A1] a joint distribution $p(y_{Rt}, y_{It} | \rho_t, \theta_t, \sigma^2)$ and likelihood can be presented. This distribution can be converted to magnitude and phase polar coordinates

$$p(m_t, \phi_t | \rho_t, \theta_t, \sigma^2) = \frac{m_t}{2\pi\sigma^2} \exp \left\{ -\frac{1}{2\sigma^2} \left[m_t^2 + \rho_t^2 - 2\rho_t m_t \cos(\phi_t - \theta_t) \right] \right\} \quad [A2]$$

Upon discarding the phase portion of the data (equivalent to marginalizing Eq. [A2] with respect to ϕ_t), the observed magnitude m_t follows the Ricean distribution

$$p(m_t | \rho_t, \sigma^2) = \frac{m_t}{\sigma^2} \exp \left\{ -\frac{m_t^2 + \rho_t^2}{2\sigma^2} \right\} \int_{\phi_t=-1}^1 \frac{1}{2} \exp \left\{ \frac{\rho_t m_t}{\sigma^2} \cos(\phi_t - \sigma_t) \right\} d\phi_t \quad [A3]$$

as has been previously described (2-4). The true underlying magnitude in Eq. [A3] is then modeled as $m_t = x_t' \beta = \beta_0 + \beta_1 x_{1t} + \dots + \beta_{q1} x_{q1t}$ where x_t' is the t^{th} row of a magnitude design matrix X and β is a $q1$ dimensional vector of magnitude regression coefficients. This Ricean distribution is well approximated by the normal distribution with mean ρ_t and variance σ^2 when the temporal SNR is high, above 7.5 (3). The usual magnitude-only linear regression model $m_t = x_t' \beta + \varepsilon_t$, parameter estimates, and test statistics follow where ε_t are additive statistically independent normally distributed noise with zero mean and variance σ^2 (3).

Upon discarding the magnitude portion of the data (equivalent to marginalizing Eq. [A2] with respect to m_t), the observed phase ϕ_t follows the distribution

$$p(\phi_t|\rho_t, \theta_t, \sigma^2) = \frac{1}{2^I} \exp \left\{ -\frac{\rho_t^2}{2\sigma^2} \right\} \left[1 + \frac{\rho_t \sqrt{2^I}}{\sigma} \cos(\phi_t - \theta_t) \exp \left\{ \frac{\rho_t^2 \cos(\phi_t - \theta_t)}{2\sigma^2} \right\} \Phi \left(\frac{\rho_t^2 \cos(\phi_t - \theta_t)}{\pi} \right) \right] \quad [\text{A4}]$$

as has been previously described (2,4). The true underlying phase in Eq. [A4] is then modeled as $\theta_t = \mathbf{x}_t' \boldsymbol{\gamma} = \gamma_0 + \gamma_1 x_{1t} + \dots + \gamma_{q_2} x_{q_2t}$ where $\boldsymbol{\gamma}$ is a q_2 dimensional vector of phase regression coefficients. We use the same design matrix in magnitude and phase although this is not required (5,6). This distribution is well approximated by the normal distribution with mean θ_t and variance $(\sigma/\rho_t)^2$ when the temporal SNR is reasonable (19). The usual phase-only linear regression model $\phi_t = \mathbf{x}_t' \boldsymbol{\beta} + \delta_t$, parameter estimates, and test statistics follow where δ_t are additive statistically independent normally distributed noise with zero mean and variance $(\sigma/\rho_t)^2$ (19).

Alternatively, the magnitude and phase can be modeled simultaneously as

$\rho_t = \mathbf{x}_t' \boldsymbol{\beta} = \beta_0 + \beta_1 x_{1t} + \dots + \beta_{q_1} x_{q_1t}$ and $\theta_t = \mathbf{x}_t' \boldsymbol{\gamma} = \gamma_0 + \gamma_1 x_{1t} + \dots + \gamma_{q_2} x_{q_2t}$ from Eq. [A1] and Eq. [A2]. Upon doing so, a likelihood can be formed from Eq. [A2], maximized under null and alternative hypotheses, and a likelihood ratio statistic formed (5,6).

References

1. Rowe DB, Logan BR. A complex way to compute fMRI activation. *Neuroimage* 2004;23(3):1078–1092. [PubMed: 15528108]
2. Rowe DB. Modeling both the magnitude and phase of complex-valued fMRI data. *Neuroimage* 2005;25(4):1310–1324. [PubMed: 15850748]
3. Menon RS. Postacquisition suppression of large-vessel BOLD signals in high-resolution fMRI. *Magn Reson Med* 2002;47(1):1–9. [PubMed: 11754436]
4. Nencka AS, Rowe DB. Reducing the unwanted draining vein BOLD contribution in fMRI with statistical post-processing methods. *Neuroimage* 2007;37(1):177–188. [PubMed: 17560130]
5. Liu T, Wong E. A signal processing model for arterial spin labeling functional MRI. *Neuroimage* 2005;24(1):207–215. [PubMed: 15588612]
6. Mumford JA, Hernandez-Garcia L, Lee GR, Nichols TE. Estimation efficiency and statistical power in arterial spin labeling fMRI. *Neuroimage* 2006;33(1):103–114. [PubMed: 16860577]
7. Wong E, Buxton R, Frank L. Implementation of quantitative perfusion imaging techniques for functional brain mapping using pulsed arterial spin labeling. *NMR Biomed* 1997;10(45):237–249. [PubMed: 9430354]
8. Noam H, Ugurbil K, Uludag K, Yacoub E. Frontiers of brain mapping using MRI. *Journal of Magnetic Resonance Imaging* 2006;23(6):945–957. [PubMed: 16649202]
9. Pandey, K.; Noll, D. Characterization and Correction of Artifacts from Dynamic Interaction Between Motion and Position Dependent Off-Resonance Patterns. Toronto, ON: 2008. p. 3121
10. Perthen JE, Bydder M, Restom K, Liu TT. SNR and functional sensitivity of BOLD and perfusion-based fMRI using arterial spin labeling with spiral SENSE at 3 T. *Magnetic Resonance Imaging* 2008;26(4):513–522. [PubMed: 18158226]
11. Yang Y, Gu H, Stein AE. Simultaneous MRI acquisition of blood volume, blood flow, and blood oxygenation information during brain activation. *Magnetic Resonance in Medicine* 2004;52(6):1407–1417. [PubMed: 15562477]
12. Le Bihan D, Breton E, Lallemand D, Grenier P, Cabanis E, Laval-Jeantet M. MR imaging of intravoxel incoherent motions: application to diffusion and perfusion in neurologic disorders. *Radiology* 1986;161(2):401–407. [PubMed: 3763909]
13. Haacke EM, Lenz GW. Improving MR image quality in the presence of motion by using rephasing gradients. *AJR Am J Roentgenol* 1987;148(6):1251–1258. [PubMed: 3495155]
14. Dumoulin CL. Phase contrast MR angiography techniques. *Magnetic resonance imaging clinics of North America* 1995;3(3):399–411. [PubMed: 7584246]

15. Hernandez-Garcia L, Lewis D, Moffat B, Branch C. Magnetization transfer effects on the efficiency of flow-driven adiabatic fast passage inversion of arterial blood. *NMR Biomed* 2007;20:733–742. [PubMed: 17304639]
16. Garcia, D.; de Bazelaire, C.; Alsop, D. Pseudo-continuous Flow Driven Adiabatic Inversion for Arterial Spin Labeling. Miami, FLA: 2005. p. 37
17. Wu WC, Fernandez-Seara M, Detre JA, Wehrli FW, Wang J. A theoretical and experimental investigation of the tagging efficiency of pseudocontinuous arterial spin labeling. *Magn Reson Med* 2007;58(5):1020–1027. [PubMed: 17969096]
18. Alsop D, Detre J. Reduced transit-time sensitivity in noninvasive magnetic resonance imaging of human cerebral blood flow. *J Cereb Blood Flow Metab* 1996;16(6):1236–1249. [PubMed: 8898697]
19. Vazquez AL, Lee GR, Hernandez-Garcia L, Noll DC. Application of selective saturation to image the dynamics of arterial blood flow during brain activation using magnetic resonance imaging. *Magn Reson Med* 2006;55(4):816–825. [PubMed: 16506156]
20. Rowe DB, Logan BR. Complex fMRI analysis with unrestricted phase is equivalent to a magnitude-only model. *Neuroimage* 2005;24(2):603–606. [PubMed: 15627605]
21. Bandettini PA, Jesmanowicz A, Wong EC, Hyde JS. Processing strategies for time-course data sets in functional MRI of the human brain. *Magn Reson Med* 1993;30(2):161–173. [PubMed: 8366797]
22. Rowe DB, Meller CP, Hoffmann RG. Characterizing phase-only fMRI data with an angular regression model. *J Neurosci Methods* 2007;161(2):331–341. [PubMed: 17157916]
23. Rowe, DB.; Hernandez-Garcia, L.; Lee, GR. Complex analysis of ASL fMRI data yields more focal activation. Berlin, Germany: 2007. p. 1422
24. Duong TQ, Kim SG. In vivo MR measurements of regional arterial and venous blood volume fractions in intact rat brain. *Magn Reson Med* 2000;43(3):393–402. [PubMed: 10725882]
25. Tomita M, Gotoh F, Amano T, Tanahashi N, Kobari M, Shinohara T, Mihara B. Transfer function through regional cerebral cortex evaluated by a photoelectric method. *Am J Physiol* 1983;245(3):H385–398. [PubMed: 6614188]
26. Leenders KL, Perani D, Lammertsma AA, Heather JD, Buckingham P, Healy MJ, Gibbs JM, Wise RJ, Hatazawa J, Herold S, et al. Cerebral blood flow, blood volume and oxygen utilization. Normal values and effect of age. *Brain* 1990;113(Pt 1):27–47. [PubMed: 2302536]
27. Duvernoy HM, Delon S, Vannson JL. Cortical blood vessels of the human brain. *Brain Res Bull* 1981;7(5):519–579. [PubMed: 7317796]
28. Nishimura N, Schaffer CB, Friedman B, Lyden PD, Kleinfeld D. Penetrating arterioles are a bottleneck in the perfusion of neocortex. *Proc Natl Acad Sci U S A* 2007;104(1):365–370. [PubMed: 17190804]
29. Berne, RM.; Levy, MN. Cardiovascular physiology. St. Louis: Mosby-Year Book; 1992. p. vip. 298
30. Buxton, RB. Introduction to functional magnetic resonance imaging : principles and techniques. Cambridge, UK ; New York: Cambridge University Press; 2002. p. xip. 523
31. Grubb RL Jr, Raichle ME, Eichling JO, Ter-Pogossian MM. The effects of changes in PaCO₂ on cerebral blood volume, blood flow, and vascular mean transit time. *Stroke* 1974;5(5):630–639. [PubMed: 4472361]
32. Lee, G.; Hernandez-Garcia, L.; Noll, D. Effects of Activation Induced Transit Time Changes On Functional Turbo ASL Imaging. ISMRM P. , editor. Kyoto, Japan: 2004. p. 519.9
33. Gonzalez-At J, Alsop D, Detre J. Cerebral perfusion and arterial transit time changes during task activation determined with continuous arterial spin labeling. *Magn Reson Med* 2000;43(5):739–746. [PubMed: 10800040]
34. Hutchinson EB, Stefanovic B, Koretsky AP, Silva AC. Spatial flow-volume dissociation of the cerebral microcirculatory response to mild hypercapnia. *Neuroimage* 2006;32(2):520–530. [PubMed: 16713717]
35. Hernandez-Garcia L, Lee GR, Vazquez AL, Noll DC. Fast, pseudo-continuous arterial spin labeling for functional imaging using a two-coil system. *Magn Reson Med* 2004;51(3):577–585. [PubMed: 15004800]
36. Hernandez-Garcia L, Lee GR, Vazquez AL, Yip CY, Noll DC. Quantification of perfusion fMRI using a numerical model of arterial spin labeling that accounts for dynamic transit time effects. *Magn Reson Med* 2005;54(4):955–964. [PubMed: 16155868]

FIG 1.

Simple cartoon (not drawn to scale) illustrating the sources of magnetization and their phase gain during the acquisition segment of an ASL experiment. Blue arrows indicate the magnetization due to the tissue spins, and red arrows indicate the magnetization from the arterial sources. The net magnetization vectors under control and labeled conditions are shown in black and green, respectively. Note that the tissue magnetization vector is shortened as the tagged blood exchanges into the tissue.

FIG 2.

Arterial contamination removed by (a) two different post inversion delays (PID) and (b) two different flow crusher amplitudes. The top row of images shows the mean magnitude subtraction of the time course, the middle row shows the mean phase difference, while the bottom row shows the correlation between magnitude and phase of the unsubtracted data. In (a) the PIDs from left to right were 0, 1200, and 1800 ms. In (b) the VENCs from Left to right were *Inf* (i.e., no crushers), 6.6, and 3.3 cm/s along the Y gradient axis.

FIG 3.

Magnitude and phase time courses for the AVIS data extracted from (a) the motor cortex during an activation paradigm (a reference activation function overlaid to indicate activation periods and (b) a gray matter mask during the CO₂ inhalation experiment (the black bar indicates the timing of the inhalation).

FIG 4. Activation maps for the three models and three experiments (a) MO Crush (b) MO PID (c) MO Whole (d) PO Crush (e) PO PID (f) PO Whole (g) MP Crush (h) MP PID (i) MP Whole. All maps are thresholded at $p=0.05$ or $-\log_{10}(p)=1.301$.

FIG 5.

Maps of the parameter and variance estimates for the magnitude-only and phase-only models. (a) MO Crush $C\beta$ (b) MO PID $C\beta$ (c) MO Whole $C\beta$ (d) MO Crush σ^2 (e) MO PID σ^2 (f) MO Whole σ^2 (g) PO Crush $D\gamma$ (h) PO PID $D\gamma$ (i) PO Whole $D\gamma$ (j) PO σ^2 Crush (k) PO σ^2 PID (l) PO σ^2 Whole. The colorbar maximum values are 35 AU for (a)-(c), 1250AU² for (d)-(g), 0.5 degree for (h)-(j), and 0.05 degrees² for (k)-(m).

FIG 6.

Power surfaces for $p=0.05$ thresholded (a) MO Power (b) PO Power (c) MP Power. CNR varies from 0 to 0.75 from left to right in 100 equal steps. TRPC varies from 0 to 0.375 in degrees from top to bottom in 100 equal steps.

Table 1

Statistics extracted from the square ROI in the visual cortex. Magnitude coefficient values are in AU, magnitude variance values in AU², phase coefficient values in degrees, and phase variances in degrees².

	Crush			PID			Whole		
	MO	PO	MP	MO	PO	MP	MO	PO	MP
β_0	2470	-	2469	3050	-	3050	3122	-	3122
β_1	17.17	-	-17.18	8.589	-	8.619	-9.731	-	-9.701
β_2	19.84	-	19.80	24.24	-	24.26	41.43	-	41.44
β_3	36.24	-	36.28	38.47	-	38.48	72.51	-	72.51
CB	16.40	-	16.48	14.23	-	14.22	31.15	-	31.08
γ_0	-	158.0	158.1	-	127.0	127.0	-	114.4	114.4
γ_1	-	-0.9969	-1.0078	-	-0.3070	-0.3101	-	-0.1785	-0.1797
γ_2	-	-0.1749	-0.1772	-	-0.1078	-0.1082	-	-0.2155	-0.2158
γ_3	-	-0.0388	-0.0402	-	-0.2070	-0.2074	-	-0.5015	-0.5013
D γ	-	0.1361	0.1371	-	-0.0992	-0.0992	-	-0.2860	-0.2854
σ^2	436.0	3.2483	3234	531.4	0.2542	593.1	532.7	0.2442	593.9
-log ₁₀ (p)	5.919	.2131	.9017	4.276	.8956	3.671	Inf	6.263	Inf