Published in final edited form as:

Magn Reson Imaging. 2008 November ; 26(9): 1201. doi:10.1016/j.mri.2008.03.006.

Phased Array 3D MR Spectroscopic Imaging of the Brain at 7 Tesla

Duan Xu^{1,2}, Charles H Cunningham³, Albert P. Chen^{1,2}, Yan Li^{1,2}, Douglas AC Kelley⁴, Pratik Mukherjee², John M. Pauly⁵, Sarah J. Nelson^{1,2}, and Daniel B. Vigneron^{1,2}

¹Joint UCSF/UC Berkeley Graduate Group in Bioengineering, University of California, San Francisco and Berkeley, CA

²Department of Radiology, University of California, San Francisco, CA

³University of Toronto and Sunnybrook Health Science Center, Toronto, Ontario, Canada

⁴Global Applied Science Laboratory, GE Healthcare Technologies, San Francisco, CA

⁵Department of Electrical Engineering, Stanford University, Stanford, CA

Abstract

Ultrahigh field 7T MR scanners offer the potential for greatly improved MR spectroscopic imaging due to increased sensitivity and spectral resolution. Prior 7T human single-voxel MRS studies have shown significant increases in SNR and spectral resolution as compared to lower magnetic fields, but have not demonstrated the increase in spatial resolution and multivoxel coverage possible with 7T MR spectroscopic imaging. The goal of this study was to develop specialized rf pulses and sequences for 3D MRSI at 7T to address the challenges of increased chemical shift misregistration, B1 power limitations, and increased spectral bandwidth. The new 7T MRSI sequence was tested in volunteer studies and demonstrated the feasibility of obtaining high SNR phased-array 3D MRSI from the human brain.

Keywords

mrsi; brain; phased array; 7T

Introduction

Whole-body 7 Tesla MR systems for research applications have become commercially available over the past few years from the major MRI manufacturers and the imaging benefits provided by these ultra-high field scanners are currently being investigated in several laboratories around the world. The benefits in BOLD contrast at 7T for fMRI have been clearly shown in a number studies ^{1,2}. The improvement in signal to noise ratio (SNR) at 7T has also been shown for volume coils between 7T and 4T ³. Recently, phased array coils at 7T have provided more uniform reception and increased sensitivity to visualize small anatomic structures not previously detected ⁴. Single voxel MRS studies have demonstrated increased SNR and improved spectral separation at 7 T which has provided better detection of more metabolites than at lower fields ⁵⁻⁷. However, 3D MR Spectroscopic Imaging at high fields

Correspondence and Reprint Request: Duan Xu, Byers Hall/QB3 Institute, Suite 102, Department of Radiology, Box 2512, University of California, San Francisco, CA 94158-2512, PH: (415) 514-4455, FAX: (415) 513-4451, email: duan@mrsc.ucsf.edu.

Publisher's Disclaimer: This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our customers we are providing this early version of the manuscript. The manuscript will undergo copyediting, typesetting, and review of the resulting proof before it is published in its final citable form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

with its much larger volume excitation presents additional challenges such as increased B1 inhomogeneity over the selected region, chemical shift misregistration artifacts, pulse profile and peak power concerns. The goal of this study was to develop and apply customized rf pulses and a specialized sequence for acquiring high spatial resolution 3D MRSI from the human brain at 7T.

Methods

Nine healthy volunteers aged 26 to 46 were scanned on a GE EXCITE 7T (General Electric Healthcare Technologies, Waukesha, WI) scanners. All studies were performed with informed consent following a protocol approved by the Committee on Human Research at our institution. A commercial 8-channel array and a volume transmit head coil (NOVA Medical, Wilmington, MA) was used to provide whole brain coverage. Volume transmit was provided by a shielded, highpass birdcage coil; the receive array consists of 8 gapped elements of 12cm by 5cm, tuned to 298MHz with multiple distributed capacitors ⁸.

Imaging

A series of anatomic imaging sequences were performed, followed by a second order automatic shimming routine, and 3D MRSI sequence. A 3 plane localizer with low resolution, followed by a high resolution axial T2/T2* weighted gradient recalled echo (GRE) with TE/TR 11.4ms/250ms with FOV of 18cm and 512×512 matrix at 2mm thickness, and 3D Inversion Recovery Spoiled Gradient Echo (IR-SPGR) with inversion time of 500ms, FOV of 24cm, 256×160 matrix and 3mm thickness were acquired. Axial slices were prescribed through both the supra- and infratentorial brain. A low resolution GRE image set was also collected for summing the spectroscopic data. Second order shimming was performed using the default routine provided by the scanner using the method described by Kim DH et al ⁹.

3D MRSI

Three sets of 3D MRSI acquisitions were performed on volunteers in 3 groups (3 volunteers in each) to investigate the range of the MRSI parameters and RF pulses employed. In group one, the standard long TE (144ms) MRSI was acquired with the custom designed spectral-spatial RF (SSRF) with improved B1 insensitivity and the enlarged chemical shift artifact. In group two, spectra of similar resolution were acquired using routine Shinar-LeRoux (SLR) designed linear-phase pulses and with short TE (35ms) to allow visualization of short T2 metabolites. In group three, regular SLR pulses were employed to achieve short TE (35ms) for acquisition of high spatial resolution spectra. See Table 2 for list of acquisition parameters. Three CHESS water suppression pulses of 120, 105, and 155 degrees with a bandwidth of 200Hz were used. In all acquisitions, the S/I dimensions were selected to be 3 voxels or greater. In the 1cc acquisition, the bottom slice was at the level of the basal ganglia, and the top slice was located above the ventricles. For the high spatial resolution acquisition, the PRESS selection was chosen near the posterior aspect of the brain in the occipital region. For 1cc studies, elliptical k-space sampling was employed to minimize scan time, in which 50% of the corner points in k-space were not collected. For the high resolution 0.35cc studies, all k-space points were collected. The collected spectra were apodized with 4Hz Lorentzian filter and processed using custom routines as previously described¹⁰, in which spectra from individual channels were summed using weighted empirical coil profiles generated by low resolution GRE images. The reported linewidths were calculated before apodization and SNR values were measured after apodization. SNR is given as peak height over the standard deviation from the far right of spectra (approximately -1 to -2ppm) where no resonances were observed. A total of 879 voxels from three volunteers were used.

SSRF—For high field MR systems, the RF power required to obtain the desired flip angle during excitation increases. For the 7T scanner, even with an efficient transmit coil, specially designed RF pulses with low peak power and sufficient bandwidth to include all metabolites of interest are required. In addition, to address the non-uniform B1 excitation at 7 T due to dielectric and wavelength effects, the SSRF were designed with adiabatic characteristics allowing the proper PRESS selection as shown in Figure 1. SSRF pulses with a sweep-width of 712Hz, duration of 30ms and peak power of 0.18G were created using the methodology proposed by Cunningham et al ¹¹. The spectral envelop of the RF was a hyperbolic-secant, and the spatial sub-pulses are of Gaussian shape, which has a time-bandwidth product of 3.5. The spectral window was centered between Cr and NAA, which is wide enough to cover metabolites of current clinical interest from Cho to NAA. The increased power deposition demanded the increase of TR to two seconds in order to stay within FDA limits. Although this caused a lengthening of scan time, it was necessary not only to stay within SAR limit but also to avoid saturation as the T1 of metabolites lengthen with increased field strength.

Very Selective Saturation (VSS) pulses—Due to peak B1 and SAR requirements of the high field system, the very selective suppression pulses ¹² had to be specially designed for this 7 T MRSI study as well as the SSRF pulses. Six, fixed 40mm wide VSS bands with 0.12G peak B1 and bandwidth of 5868Hz were placed around PRESS box for approximately 1000 fold out of volume suppression and to sharpen the edges and reduce chemical shift misregistration for the PRESS-selected volume. It has much higher bandwidth than Le Roux's original design ¹³, thus providing spatial profile with higher selectivity.

Results

Excellent quality 3D MRSI data were acquired from normal volunteers utilizing the specialized 7T PRESS MRSI sequence incorporating the novel rf pulses described above. The MRSI data from a total volume of 168 cm³ in Figure 2 demonstrates the uniformity, high SNR and minimal chemical shift misregistration provided by the new SSRF and VSS pulses designed for this study. Similar to MRSI acquisitions at lower fields, the employment of VSS pulses were important for obtaining PRESS selection with sharp edges. The SNR and linewidth numbers are reported in Table 2 and 3 by axial slice. Although only longer echo times are achievable with the 30 msec SSRF pulses, Figure 3a demonstrates the Glx peak at long TE, in comparison with the short TE acquisition shown in Figure 3b.

Acquisitions of high spatial resolution 3D-MRSI (7mm per size, 0.34cm³) at short TE were also shown to be feasible at 7 T. A selected voxel is demonstrated in Figure 4a. The following SNR and linewidth numbers are for the short TE acquisitions. Choline had a linewidth of 9.6 ±4.5Hz and a SNR of 15.9±6.4; Creatine had a linewidth of 14.3±3.2Hz and a SNR of 34.8 ±12.8; NAA had a linewidth of 17.9±5.3Hz and a SNR of 54.9±19.0.

Discussion

With the recent non-significant risk classification of 7T MRI scanners by the US Food and Drug Administration, ultrahigh field *in vivo* MR images and functional data with improved SNR and susceptibility contrast have been demonstrated at various institutions ¹⁴⁻¹⁶. However, hardware and acquisition changes are necessary to accommodate the increased field strength ^{17,18}. For MR spectroscopic acquisitions at high field, enhanced sensitivity to magnetic susceptibility and uniformed excitation are two major fundamental challenges. Peak and average RF power requirements and coil performance issues also compound the acquisition process. Due to these challenges, prior 7 T MRS studies have focused single voxel acquisitions of relatively large 8 cm³ voxel sizes ^{5-7,19} or 2D acquisitions ^{20,21}. This approach has limited value for many applications due to the small spatial coverage across the brain and large voxel

size. In this paper, we demonstrated the feasibility of acquiring 3D MRSI of the human brain at 7 T using SSRF and regular SLR with respective advantages and disadvantages.

A method employed by previous MRSI studies at lower fields to minimize the chemical shift artifact is to over prescribe the volume (OVERPRESS) and utilize VSS pulses to sharpen the edges of the PRESS-selected volume and saturate the outside¹². With the increased field, power of these VSS is becoming an issue as each surface of the box required at least one pulse and additional ones if graphic saturation were prescribed. This required the redesign and implementation of optimized VSS pulses for 7 T MRSI. Using tools offered from the Electrical Engineering Department at Stanford University, a set of 0.12G peak power VSS were designed and implemented for the use at 7T. Based on visual inspection, it showed excellent spatial profile, yielding sharp edges for the PRESS selection. This is critical as no matter whether SSRF or SLR pulses were employed to select the PRESS volume, the high bandwidth, small transition band VSS pulses were required for out of volume suppression, sharpening the selected region and reducing chemical shift misregistration.

The use of SSRF is important due to the larger bandwidth requirement of the pulses to limit chemical shift artifact, which can be used in conjunction of OVERPRESS. The SSRF employed in this study also has adiabatic properties which allowed the acquisition to be much less sensitive to non-uniform excitation. The symmetric sweep property of the SSRF yielded B1 insensitivity; therefore, spectra remained consistent across the PRESS selection and all voxels were considered in the analyses. The customized low peak power and low SAR VSS pulses were also very important. At lower field, this may not be a concern because the SAR limit would allow multiple VSS to be played. However, at 7T, under a reasonable TR, only one VSS per band can be played within the SAR limit.

The linewidths in this 3D MR spectroscopic imaging study were higher than previously reported for single voxel 7 T MRS studies⁶, due presumably to the much larger spatial coverage and increased Bo inhomogeneities for the MRSI data. The variation in spectral linewidth due to non-uniformity in the field may be improved by future improvements in higher order shimming hardware and software. Increased linewidths were observed near the edge of PRESS volume due to residual Bo inhomogeneities after higher order shimming.

The SSRF pulses provided the major advantage of having large bandwidth and adiabatic properties. This allowed high quality MRSI data to be acquired over a large 3-dimensional volume in the brain. However, a limitation to these pulses is the requirement of longer TE acquisitions due to the pulse length of the SSRF pulses. The two SSRF 180° pulses were each 30ms long compared to the 6.5ms pulse length of the SLR pulses. The SLR pulses allowed a TE of 35ms, in which short T2 metabolites were clearly observed. Due to modulation at higher field and properties of the SSRF, even at long TE, the Glx peak was clearly detected in the volunteer studies, shown in Figure 3a. Simulation has shown this peak is primarily Glu at this TE²². This maybe of clinical interest, as Glu is readily detectable without editing techniques at 7T. Other short TE metabolites may also be visible, but due to the limited sweep width of the SSRF, they are not seen. Further studies are required to determine the modulation and corresponding modifications to the acquisition, which may further improve the visualization.

These initial 3D MRSI investigations of the human brain at 7T highlighted the need of optimal higher order shimming to address Bo variability across the volume. The top slice above the ventricle had the best shim, which can be observed visually and confirmed by the linewidth and SNR computed from the peak heights of the metabolites. The slices near and at the level of the nasal sinus air-tissue interface (moving from slice 2 to slice 3) showed worse SNR and larger linewidth due to high level of magnetic susceptibility. It is especially clear towards the edges of the PRESS box because the higher order shim region was manual selected with an

ellipsoid covering most of the selection. The most inferior axial slice demonstrated the worst shim as it was the closest to the air-tissue interface and the resultant magnetic susceptibility shifts.

This initial investigation incorporated the design and application of new specialized rf pulse designs for obtaining high spatial resolution 3D MRSI at 7T from large selected volumes with increased B1 insensitivity and reduced chemical shift misregistration. The results of this study demonstrated the feasibility of this method to study metabolite distributions at 7T and highlighted the benefit of higher order shimming, low power very selective suppression pulses, and custom designed RF excitation pulses.

Acknowledgments

The authors would like to thank Dr. Julio Carballido-Gamio for Matlab consultations and numerous volunteers who made this work possible.

This work was supported by grants from the National Institutes of Health (RO1 NS40117 and UC Discovery LSIT 01-10107).

References

1. Duong TQ, Yacoub E, Adriany G, et al. High-resolution, spin-echo BOLD, and CBF fMRI at 4 and 7 T. *Magn Reson Med* 2002;48(4):589–593. [PubMed: 12353274]
2. Yacoub E, Duong TQ, Van De Moortele PF, et al. Spin-echo fMRI in humans using high spatial resolutions and high magnetic fields. *Magn Reson Med* 2003;49(4):655–664. [PubMed: 12652536]
3. Vaughan JT, Garwood M, Collins CM, et al. 7T vs. 4T: RF power, homogeneity, and signal-to-noise comparison in head images. *Magn Reson Med* 2001;46(1):24–30. [PubMed: 11443707]
4. Wiggins GC, Potthast A, Triantafyllou C, Wiggins CJ, Wald LL. Eight-channel phased array coil and detunable TEM volume coil for 7 T brain imaging. *Magn Reson Med* 2005;54(1):235–240. [PubMed: 15968650]
5. Tkac I, Gruetter R. Methodology of 1H NMR Spectroscopy of the Human Brain at Very High Magnetic Fields. *Applied Magnetic Resonance* 2005;(29):139–157.
6. Tkac I, Andersen P, Adriany G, Merkle H, Ugurbil K, Gruetter R. In vivo 1H NMR spectroscopy of the human brain at 7 T. *Magn Reson Med* 2001;46(3):451–456. [PubMed: 11550235]
7. Ugurbil K, Adriany G, Andersen P, et al. Ultrahigh field magnetic resonance imaging and spectroscopy. *Magn Reson Imaging* 2003;21(10):1263–1281. [PubMed: 14725934]
8. Ledden, PJ.; Gelderen, PV.; Duyn, JH. Birdcage Volume Transmit, Eight Channel Receive Array System for Brain Imaging at 7T. Miami, FL, USA: May. 2005 p. 322
9. Kim DH, Adalsteinsson E, Glover GH, Spielman DM. Regularized higher-order in vivo shimming. *Magn Reson Med* 2002;48(4):715–722. [PubMed: 12353290]
10. Nelson SJ. Analysis of volume MRI and MR spectroscopic imaging data for the evaluation of patients with brain tumors. *Magn Reson Med* 2001;46(2):228–239. [PubMed: 11477625]
11. Cunningham CH, Vigneron DB, Chen AP, et al. Design of symmetric-sweep spectral-spatial RF pulses for spectral editing. *Magn Reson Med* 2004;52(1):147–153. [PubMed: 15236378]
12. Tran TK, Vigneron DB, Sailasuta N, et al. Very selective suppression pulses for clinical MRSI studies of brain and prostate cancer. *Magn Reson Med* 2000;43(1):23–33. [PubMed: 10642728]
13. Le Roux P, Gilles RJ, McKinnon GC, Carlier PG. Optimized outer volume suppression for single-shot fast spin-echo cardiac imaging. *J Magn Reson Imaging* 1998;8(5):1022–1032. [PubMed: 9786138]
14. Augustinack JC, van der Kouwe AJ, Blackwell ML, et al. Detection of entorhinal layer II using 7Tesla [correction] magnetic resonance imaging. *Ann Neurol* 2005;57(4):489–494. [PubMed: 15786476]
15. Michaeli S, Garwood M, Zhu XH, et al. Proton T2 relaxation study of water, N-acetylaspartate, and creatine in human brain using Hahn and Carr-Purcell spin echoes at 4T and 7T. *Magn Reson Med* 2002;47(4):629–633. [PubMed: 11948722]

16. Lei H, Zhu XH, Zhang XL, Ugurbil K, Chen W. In vivo ³¹P magnetic resonance spectroscopy of human brain at 7 T: an initial experience. *Magn Reson Med* 2003;49(2):199–205. [PubMed: 12541238]
17. Zhang X, Ugurbil K, Chen W. Microstrip RF surface coil design for extremely high-field MRI and spectroscopy. *Magn Reson Med* 2001;46(3):443–450. [PubMed: 11550234]
18. Wang J, Yang QX, Zhang X, et al. Polarization of the RF field in a human head at high field: a study with a quadrature surface coil at 7.0 T. *Magn Reson Med* 2002;48(2):362–369. [PubMed: 12210945]
19. Mangia S, Tkac I, Gruetter R, Van de Moortele PF, Maraviglia B, Ugurbil K. Sustained neuronal activation raises oxidative metabolism to a new steady-state level: evidence from ¹H NMR spectroscopy in the human visual cortex. *J Cereb Blood Flow Metab* 2007;27(5):1055–1063. [PubMed: 17033694]
20. Otazo R, Mueller B, Ugurbil K, Wald L, Posse S. Signal-to-noise ratio and spectral linewidth improvements between 1.5 and 7 Tesla in proton echo-planar spectroscopic imaging. *Magn Reson Med* 2006;56(6):1200–1210. [PubMed: 17094090]
21. Goelman G, Liu S, Fleysher R, Fleysher L, Grossman RI, Gonen O. Chemical-shift artifact reduction in Hadamard-encoded MR spectroscopic imaging at high (3T and 7T) magnetic fields. *Magn Reson Med* 2007;58(1):167–173. [PubMed: 17659608]
22. Xu, D.; Li, Y.; Chen, AP., et al. 2D J-Resolved Spectroscopy at 7T. Berlin, Germany: 2007. p. 1365

Figure 1.

(a) a partially adiabatic SSRF pulse for 7T acquisitions. Blue and green lines are the real and imaginary components, respectively. (b) simulated spectral vs spatial profile (180-degree flip angles) shows an absence of chemical shift mis-registration. (c) spatial profile for RF amplitude of $\pm 33\%$ showing adiabatic behavior showing only slight changes in profile.

Figure 2. 1cc 3D MRSI of the brain using SSRF with VSS at TE of 144ms. It demonstrates good quality spectra but variable linewidth due to residual B_0 inhomogeneity.

Figure 3. Spectra obtained using SSRF with TE 144ms (a) and SLR with TE 35ms (b) at 1cc resolution and TR 2s. $8 \times 8 \times 8$ with elliptical kspace sampling. Total time of 17.5min.

Figure 4. 0.35cc spectra with TE of 35ms and TR 2s. 8×8×8 phase encoding for a total acquisition time of 17.5min. The array (b) demonstrates the gray and white matter metabolite concentration differences, in which choline is higher in white matter.

Table 1**MRSI Parameters**

Parameters	Group 1	Group 2	Group 3
Selection Pulse	SSRF	SLR	SLR
TE	144ms	35ms	35ms
TR	2s	2s	2s
Elliptical k-space sampling	Yes	Yes	No
Total Imaging Time	17.5min	17.5min	17.5min
Spatial Resolution	1cm ³	1cm ³	0.35cm ³
Matrix Size	12×12×8	12×12×8	8×8×8
OVERPRESS	40%	40%	40%

Table 2

Linewidth for 1cc acquisitions in Hz

	Cho		Cr		NAA	
	144ms	35ms	144ms	35ms	144ms	35ms
Slice 1	17.2±3.8	15.3±9.4	17.5±4.7	19.1±8.7	21.0±5.6	23.7±10.3
Slice 2	17.4±4.2	20.1±9.7	17.9±4.4	22.3±6.9	21.4±5.4	27.6±11.3
Slice 3	19.5±7.5	22.8±9.3	21.7±5.6	26.8±7.7	25.5±8.8	32.1±16.0

Table 3

SNR for 1cc acquisitions

	Cho		Cr		NAA	
	144ms	35ms	144ms	35ms	144ms	35ms
Slice 1	54.7±20.5	103.0±27.4	69.7±17.8	167.1±43.2	164.1±30.3	278.5±71.3
Slice 2	57.3±29.2	92.9±30.5	74.5±28.4	141.0±53.2	154.3±50.5	202.1±69.6
Slice 3	34.2±19.1	61.3±26.0	43.9±22.0	94.5±45.0	89.5±39.7	127.8±65.0