Erratum: “Total scatter factors of small beams: A multidetector and Monte Carlo study” [Med. Phys. 35, 504–513 (2008)]

Paolo Francescon,^{a)} Stefania Cora, and Carlo Cavedon

Department of Medical Physics, ULSS 6, Viale Rodolfi, Vicenza, Vicenza 36100, Italy

(Received 9 December 2009; accepted for publication 10 December 2009; published 28 January 2010)

[DOI: [10.1118/1.3284208](https://doi.org/10.1118/1.3284208)]

We would like to draw your attention to an error in our 2008 paper describing a method for estimating the total scatter factors of small beams.¹ Specifically, Table VII, which lists correction factors as a function of detector type and FWHM, includes the value $F_{\text{corr}}=0.940$ for the PTW 60012 diode and FWHM=2.6 mm. This value is a typographical error; the correct value is 0.959. We feel that this correction should be published since the error of approximately 2%

would be reflected directly in the total scatter factor used for clinical dose calculations if this combination of detector and FWHM were found in practice.

^{a)}Electronic mail: paolo.francescon@ulssvicenza.it

¹P. Francescon, S. Cora, and C. Cavedon, “Total scatter factors of small beams: A multidetector and Monte Carlo study,” *Med. Phys.* **35**, 504–513 (2008).