Published in final edited form as:

Bioanalysis. 2010 February 1; 2(2): 311–341. doi:10.4155/bio.09.185.

Stable-isotope dilution LC–MS for quantitative biomarker analysis

Eugene Ciccimaro¹ and Ian A Blair^{2,†}

¹Thermo Fisher Scientific, 265 Davidson Avenue, Somerset, NJ 08873–4120, USA

²Centers of Excellence in Environmental Toxicology and Cancer Pharmacology, Department of Pharmacology, University of Pennsylvania School of Medicine, 421 Curie Blvd, Philadelphia, PA 19104–6160, USA

Abstract

The ability to conduct validated analyses of biomarkers is critically important in order to establish the sensitivity and selectivity of the biomarker in identifying a particular disease. The use of stable-isotope dilution (SID) methodology in combination with LC–MS/MS provides the highest possible analytical specificity for quantitative determinations. This methodology is now widely used in the discovery and validation of putative exposure and disease biomarkers. This review will describe the application of SID LC–MS methodology for the analysis of small-molecule and protein biomarkers. It will also discuss potential future directions for the use of this methodology for rigorous biomarker analysis.

Stable-isotope dilution LC–MS methodology for biomarker analysis

The development of clinically useful biomarkers is a two-stage process. The initial bioanalytical validation stage uses control samples in appropriate biofluids to establish analytical limits of quantification and determine errors introduced by sample handling that are typically used in conventional bioanalytical validation studies [1]. This stage is followed by clinical validation using patient samples to determine the assay's sensitivity for detecting the disease and specificity in discriminating the particular disease. Without rigorous bioanalytical validation, time-consuming and expensive clinical validation studies cannot succeed. Conventional validation strategies require quality control samples to be conducted in the biofluid being examined. Therefore, bioanalytical validation of endogenous bio markers is often very challenging, due to the target analyte's presence in the biofluid of interest. Related to this hurdle, though more difficult to resolve, is the fundamental uncertainty as to whether the endogenous signal is actually the analyte of interest and not an interfering substance with similar physicochemical properties. To best distinguish the target analyte from chemical-background peaks, assays have to be conducted with the maximal specificity and sensitivity

†Author for correspondence Tel.: +1 215 573 9880, Fax: +1 215 573 9889, ianblair@mail.med.upenn.edu.

For reprint orders, please contact reprints@future-science.com

Financial & competing interests disclosure

Support of biomarker research by NIH grants UO1ES16004, RO1CA091016 and RO1130038, held by Ian A Blair and P30ES0130508 held by Trevor Penning at the University of Pennsylvania is gratefully acknowledged. Ian A. Blair is the AN Richards Professor and Vice-Chair of the Department of Pharmacology, Director of the Center for Cancer Pharmacology, Director of the Proteomics and Systems Biology Facility and an employee of the University of Pennsylvania. Eugene Ciccimaro is an LC–MS applications chemist employed by the Scientific Instruments Division of Thermo Fischer Scientific. Description of the studies described in the article was not influenced by the authors' current appointments, although research in many of the areas is currently ongoing. The authors have no other relevant affiliations or financial involvement with any organization or entity with a financial interest in or financial conflict with the subject matter or materials discussed in the manuscript. This includes employment, consultancies, honoraria, stock ownership or options, expert testimony, grants or patents received or pending, or royalties.

No writing assistance was utilized in the production of this manuscript.

that is possible using stable isotope analogs as internal standards. GC- and LC-MS/MS are the two most widely used instrument platforms to employ stable-isotope dilution (SID) methodology. LC-MS/MS is more applicable to the analysis of a wider range of biomarkers than GC-MS/MS and is also inherently easier to use for rigorous validation. Hence, the present review will focus on the use of SID LC-MS/MS for biomarker analysis. However, the underlying principles relating to the specificity of SID methodology are relevant to both techniques. It should also be noted that SID has been used for the elemental analysis of biomarkers, using such techniques as inductively coupled plasma-MS, though this is beyond the scope of this review and will not be discussed.

Quantitative studies often require the most sensitive means of detection possible. The MS platform and analysis mode best suited for a particular analysis needs to be determined empirically and will depend on the molecule and matrix involved. However, in general, a triple quadrupole (TQ) operated in the multiple reaction monitoring (MRM)-MS (also known as selective reactive monitoring [SRM], but referred to herein as MRM) mode will show exceptional levels of sensitivity and selectivity when coupled to LC. In this mode of operation, a precursor ion is preselected and resolved in Q1 of the TQ, fragmented by collision-induced dissociation (CID) in Q2 and the resultant product ion is analyzed in Q3. Under optimal operating conditions, the precursor to product ion 'reaction' is monitored many times per second, resulting in extremely reproducible chromatographic peak shapes and intensity. In this way, a stable (heavy) isotope-labeled standard is used in SID LC-MRM-MS to establish the presence of an endogenous analyte using both the LC retention time and MS/MS mass selection of the TQ platform. This level of specificity cannot be attained with any other bioanalytical technique employed for biomarker analysis.

An authentic stable isotope-labeled analog of a compound is identical to the endogenous molecule except for mass. The term SID most often refers to the use of a stable isotope-labeled internal standard spiked into a sample at a known concentration. The response ratio between the analyte and labeled compound can then be interpolated onto a standard curve to calculate the absolute amount of analyte in the unknown sample. Variations of this technique are also used, most extensively in proteomics, where chemically and metabolically labeled proteins and peptides are often used at unknown concentrations for relative quantification purposes. In either case, the internal standard offers a means to verify the presence of the analyte and normalize experimental variables, such as sample storage and matrix suppression. In our article, the SID technique will be reviewed under this broader understanding of its application in LC-MS experiments. The use of structural analogs as internal standards, rather than authentic isotope-labeled analogs, is undesirable because they will have different retention times and ionization properties compared with the analyte of interest. Therefore, differential ionization can occur between an analyte and a structural analog in the source of the mass spectrometer. This difference arises in part from suppression of ionization by constituents present in the biofluid that is being analyzed and can lead to significant imprecision during quantitative analyses [2]. Unfortunately, suppression effects vary with chromatographic retention time and with biofluid samples from different individuals. It is therefore impossible to standardize the amount of suppression occurring within any particular sample. The ideal control offered by an authentic isotope-labeled internal standard is not always possible because, for many biomarkers, only deuterated and structural analogs are available. Deuterated forms of a compound are not perfect internal standards, since there is a small but significant separation of the deuterium analog internal standards and their corresponding endogenous protium forms during chromatography. This slight difference in chromatography can result in differential suppression or enhancement of ionization and affect the quality of the analytical data. Structural analogs are even less representative of the endogenous compound since, in addition to differences in LC retention time, the structural analog can show different absorptive loss. Selective binding to active sites on glassware or other surfaces can occur during extraction and

chromatography, leading to significant analyte loss. Whereas a structural analog might not account for this loss, an isotope-labeled internal standard has identical physicochemical properties and is therefore lost at exactly the same rate as the endogenous analyte. Due to this feature of stable-isotope analogs, they may act as a carrier, preventing the loss of trace amounts of analyte during extraction and analysis [3]. This is an often-overlooked benefit offered by the isotope-labeled internal standard. Finally, variability introduced during compound isolation can be fully controlled by an authentic isotope-labeled standard. In many cases, compound enrichment is required to improve the analytical performance, as we found with amyloid β -peptides [3] and DNA adducts [4], which required immunoaffinity purification.

Small-molecule biomarkers

Introduction

The search for small-molecule biomarkers that are predictive of cardiovascular disease, neurodegenerative diseases and cancers has been particularly intense [5–8]. By linking biomarker discovery and validation with known pathways of endogenous metabolism it is possible to rationally approach the development of a biomarker for a specific disease. This is exemplified in the case of breast cancer research. Based on figures from 2002 to 2006, 87.3% of breast cancer diagnoses were made in women over the age of 45. Hence, aging is an important risk factor and postmenopausal women are at much higher breast cancer risk than premenopausal women. Paradoxically, postmenopausal women lack the capacity to synthesize ovarian estrogens, yet are at increased risk through exposure to elevated levels of circulating estrogens [9]. Rationally, this increased risk could result from a genotoxic effect of estrogen metabolites [10], from the direct proliferative effect of estrogens [11] or from a combination of both mechanisms. A substantial effort over the last decade to develop methods for analyzing serum estrogen biomarkers that are predictive of breast cancer risk has therefore been underway [7]. Such a biomarker could designate women at a high risk of breast cancer, and chemoprevention strategies could then be implemented [12,13].

Small-molecule biomarkers are derived from modifications to glutathione (GSH), DNA and cellular metabolites, such as lipids and folates (FIGURE 1). The most common biofluids that are used to monitor small-molecule disease biomarkers include, blood, serum, urine, cerebrospinal fluid, red blood cells (RBCs) and nasal and lung lavages [14]. Modifications to major biomolecules can occur through the direct and indirect actions of reactive intermediates and/or reactive oxygen species (ROS), which in turn can arise from the metabolism of both exogenous and endogenous chemicals during inflammation, radiation and the metabolism of hormones, drugs and environmental toxins [15]. Cellular GSH provides a major defense against ROS and reactive intermediates through GSH peroxidase-mediated reduction of ROS and GSH-S-transferase (GST)-mediated conversion of reactive intermediates to GSH adducts (FIGURE 1) [6]. As GSH is consumed there is a concomitant increase in glutathione disulfide (GSSG) concentrations [16]. In addition, increased activity of oxygenases such as cyclooxygenase (COX) and lipoxygenase occurs, which contributes to the oxidative metabolism of xenobiotics and endogenous compounds [17,18]. These processes can ultimately lead to DNA, protein and lipid damage and disrupt cellular metabolite formation (such as homocysteine and folates). Both the ‘damaged’ molecules as well as the aberrant metabolite concentrations may therefore provide potential biomarkers of diseases (FIGURE 1). For example, the direct action of ROS can lead to the formation of 8-oxo-7,8-dihydro-2'-deoxyguanosine (8-oxodGuo) [4] and the F₂-isoprostanes (IsoPs) [5]. Indirectly, ROS-mediated lipid peroxidation can lead to the formation of DNA adducts such as heptanone-etheno-2'-deoxyguanosine (HedGuo) (FIGURE 1) [19]. Since DNA adducts are normally excised from DNA by nucleotide excision repair (NER) as the corresponding 2'-deoxynucleotides or by base excision repair (BER) as the corresponding purine or pyrimidine bases, the resulting DNA adducts excreted in the urine can serve as disease biomarkers. In rare

cases when the DNA adducts are not excised from DNA, this can lead to a mutation, such as a G to T transversion. Such mutations may ultimately result in aberrant mRNA and protein expression (FIGURE 1), which in turn can lead to increased cellular proliferation and tumorigenesis. Following this rationale, the analysis of urinary DNA adducts has been employed to provide biomarkers of cancer risk [19].

The analysis of small-molecule biomarkers using LC–MS/MS-based methodology most often involves the use of reversed-phase chromatography coupled to a TQ mass spectrometer or ion trap, utilizing an atmospheric pressure ionization (API) source such as electrospray ionization (ESI), nanospray or atmospheric pressure chemical ionization (APCI). In some instances, such as our targeted lipidomic profiling of various eicosanoids and other metabolites of lipid peroxidation, chiral LC separations were accomplished using normal phase, rather than reversed-phase chromatography [20]. Normal-phase solvents do not allow efficient ionization so this is performed by electron capture atmospheric pressure chemical ionization (ECAPCI)–MS [21]. Ion traps provide high sensitivity in full scan mode and can be used to acquire data-dependant MS/MS information on multiple ions at extremely fast cycle speeds. Additionally, when coupled to a high-resolving and high-mass-accuracy trap, as in a hybrid instrument, this data can provide outstanding compound identifications during small-molecule biomarker discovery. However, TQs are used most often during biomarker validation due to their improved sensitivity for quantification of targeted analytes. Although modern TQ instruments can conduct MRM scans very quickly, operation under optimal scan conditions for biomarker sensitivity and reproducibility does limit the number of compounds monitored.

Steroids

For brevity, we will focus here on estrogens as a case-in-point of the usefulness and analytical challenges associated with measuring endogenous steroids by SID MS/MS. Estrogen biosynthesis in the breast tissue of post-menopausal women arises from circulating C-19 endogens formed in the adrenal cortex. Estradiol is formed directly from androstenedione or from reduction of estrone by 17 β -hydroxysteroid dehydrogenases (HSDs)-1, -7 and -12 and/or aldo-keto reductase 1C3 (FIGURE 2). Conversely, estradiol can be reduced back to estrone by 17-HSDs 2 and 14. Both estrone and estradiol are converted to catechol estrogens by cytochromes P450 (CYPs) and then to 2- and 4-methoxy-estrogens by catechol-*O*-methyltransferases (COMTs). The ratio of 4-methoxy-estrogens to 2-methoxy-estrogens provides an indirect measurement of the catechol estrogens 4-hydroxy-estradiol/2-hydroxy-estradiol and 2-hydroxy-estrone/4-hydroxy-estrone (FIGURE 2). This is potentially important since the catechols are considered to be genotoxic estrogens while 2-methoxy-estrogens are considered to be antiproliferative and protective against mammary carcinogenesis [22,23]. 16 α -hydroxy-estrone, which is formed by CYP-mediated metabolism of estrone, is also considered to be procarcinogenic (FIGURE 2) [24]. The ability to routinely conduct very high sensitivity analyses of 4-methoxy-estrogens and 2-methoxy-estrogens together with estrone, 16 α -hydroxy-estrone and estradiol in serum will make it possible to develop and evaluate new and improved models of breast cancer risk [13,25]. The development of such models would impact significantly on the implementation of chemoprevention strategies for women newly identified to be in a high breast cancer risk category. Previous studies have shown that this would significantly improve breast cancer prevention [12,13]. Therefore, the ability to routinely analyze plasma and serum estrogens with very high sensitivity could potentially save a large number of women from this devastating disease [25].

For reliable measurements of multiple estrogens in plasma or serum, it is necessary to employ SID methodology in combination with LC–MRM-MS. Unfortunately, endogenous estrogens are not effectively ionized using conventional ESI or APCI methodology. This has restricted the use of these ionization techniques to samples with higher concentrations of plasma and

serum estrogens [26–30]. Fortunately, the ionization characteristics can be readily enhanced by derivatization of the 3-hydroxy-group on the A-ring of the estrogen. Three derivatization approaches for plasma and serum estrogens have been reported. The first approach, which we have pioneered, involves the preparation of an electron capturing pentafluorobenzyl derivative of the estrogen coupled with the use of electron capture atmospheric pressure chemical ionization (ECAPCI)-MS for analysis (FIGURE 3) [21]. Higashi *et al.* have also explored the utility of ECAPCI-MS for estrogen analysis by using different electron capturing derivatives [31]. We showed that it was possible to quantify estrogens in the low picogram per milliliter range in plasma using LC–ECAPCI-MS (TABLE 1) [32]. This second approach uses conventional derivatization coupled with LC–ESI-MS. This approach is exemplified by studies that have employed the dansyl (D) derivative [26,29,33] and others that made use of picolinoyl (P) or pyridyl-3-sulfonyl (PS) derivatives [34,35] (FIGURE 3). The third approach involves the preparation of preionized (quaternized) derivatives, so that ionization is not required in the ESI source of the mass spectrometer. This has involved preparation of *N*-methyl-2-pyridyl (NMP), 1–2,4-dinitro-5-fluorophenyl-4,4,-dimethylpiperazinyl (MPPZ) or *N*-methylnicotinyl (NMN) groups attached to the 3-hydroxy moiety of the estrogen moiety (FIGURE 3) [36–38]. The three derivatization strategies described above make it possible to quantify plasma and serum estrogens by SID LC–MS/MS with limits of quantification (LOQs) in the high femtogram per milliliter to low picogram per milliliter range (TABLE 1) [32–35,38]. Additionally, until recently, only deuterated analogs were available for use as internal estrogen standards. Fortunately, the recent availability of [$^{13}\text{C}_6$]-estrogen analogs from Cambridge Isotope Laboratories (Andover, MA, USA) makes it possible to use internal standards that have identical chromatographic retention times as their corresponding analytes but a mass difference of 6 Da. The availability of methodology based on these novel ionization techniques and [^{13}C]-labeled internal standards will greatly facilitate future studies to rigorously establish the precise levels of individual estrogens that are present in the plasma and serum of postmenopausal women.

The ESI process requires ionization to occur in solution followed by desolvation of the resulting protonated molecules in the source of the mass spectrometer. It is therefore difficult to achieve complete ionization of all analyte molecules. This contrasts with preionized derivatives, which are already completely ionized (FIGURE 3). The NMN preionized derivative for the highly sensitive analysis of plasma and serum estrogens has been employed by Adamec *et al.* (TABLE 1) [38]. Our laboratory has also explored this approach using the Girard T (GT) derivative, which has been previously used to analyze keto steroids [39,40]. Extremely high sensitivity can be obtained with the preionized estrone GT derivative (FIGURE 3) when analyzed using a modern TQ mass spectrometer and nanoflow chromatography. For example, using [$^{13}\text{C}_6$]-estrone as the internal standard, we have demonstrated a limit of detection of 0.6 fg (2.2 amol) of estrone on column as its GT derivative [7]. This suggests that it will be possible to conduct LC–MS/MS assays on multiple estrogen metabolites in serum and plasma an order of magnitude lower than is currently achievable using LC–ECAPCI-MS or LC–ESI-MS methodology (TABLE 1). The ease with which the preionized derivatization strategy can be implemented will make it possible to readily introduce highly sensitive methodology in laboratories that are currently employing LC–MS/MS methodology. We anticipate that the use of preionized estrogen derivatives will also help conserve important plasma and serum samples as it will be possible to conduct high sensitivity analyses using limited sample volumes. This will allow the use of existing banked plasma and serum samples without significantly depleting the amounts that are available. In this way, multiple studies may share the same samples, which may be critical to our understanding of the multivariate factors that cause an increase in breast cancer risk.

Lipids

There are eight classes of lipids: fatty acyls, glycerolipids, glycerophospholipids, sphingolipids, sterol lipids, prenol lipids, saccharolipids and polyketides [41]. Lipid structures are extremely diverse, which results in their wide range of polarity, although most lipids are sparingly soluble in water at neutral pH. Classes 1–4 have similar fatty acid and polyunsaturated fatty acid substituents at their *sn*-1 and *sn*-2 positions. Polyunsaturated fatty acids are susceptible to oxidation through both nonenzymatic actions of ROS and by enzymatic processes such as the action of COXs, lipoxygenases and CYPs (FIGURE 4). Lipid oxidation, the mechanisms by which this process can take place and the numerous potential biomarkers that are formed have generated a substantial amount of biomarker research [19,42,43]. This burgeoning interest is due, at least in part, to the importance of lipids in biological systems and to the negative effects of the lipid oxidation on human health [5]. SID LC–MS/MS-based methodology has had a significant impact on the analysis of five of the lipid classes – fatty acyls, glycerolipids, glycerophospholipids, sphingolipids and sterol lipids [5].

Fatty acyls represent one of the most important classes of bioactive lipids and they have been extensively investigated using MS methodology. In early studies, GC–MS was widely used, but its applicability was limited due to the thermal stability and volatility of the molecules. Therefore, the availability of API-based methodology [44,45] has had a dramatic impact on the field [46]. Fatty acyls form protonated molecules $[M + H]^+$ or negatively charged molecules arising from the loss of a proton $[M - H]^-$ under positive or negative ESI conditions, respectively. Fatty acyls are ionized more efficiently under negative ESI conditions so this mode is usually employed [47,48]. Using a combination of ESI and LC–MS/MS, it is now possible to analyze extremely complex mixtures of fatty acyl biomarkers derived from diverse biological samples [49–51]. LC–ECAPCI-MS analysis of fatty acyls, which have been derivatized as pentafluorobenzyl esters, is two orders of magnitude more sensitive when compared with negative ESI of underivatized fatty acyls [21]. The LC–ECAPCI-MS methodology also allows the chiral lipid peroxidation products such as hydroxyeicosatetraenoic acids to be resolved (FIGURE 5) [52], which makes it possible to determine whether the fatty acyls are derived from nonenzymatic or enzymatic pathways (FIGURE 4) [20,53].

Prostaglandins (PGs) are COX-mediated bioactive metabolites derived from arachidonic acid (FIGURE 4). As with other arachidonic acid metabolites, they form carboxylate anions under ESI and ECAPCI conditions. The MS fragmentation pathways resulting from CID of PGs is complicated, due to the formation of ‘second-generation’ metabolites [54]. However, these fragmentation pathways make it possible to employ specific product ions for selective LC–MRM-MS analyses [20]. PG metabolites have been widely used as biomarkers to study the effects of COX inhibitors. Many of these studies were performed using older SID GC–MS-based methodology [55]. isoPs are a class of arachidonic acid metabolites that are produced by the action of ROS on intact lipids [56,57]. As they are generated nonenzymatically, the isoPs are produced as a mixture of isomers, derived from their multiple chiral centers. There are four classes of isoPs that undergo class-specific fragmentation when subjected to CID, making them good candidates for SID LC–MRM-MS analysis [58]. Using SID GC–MS methodology, they were rigorously validated as biomarkers of oxidative stress and have been widely used as disease biomarkers [59]. Unfortunately, currently many of the isoPs cannot be quantified in biological samples because appropriate heavy isotope-labeled standards are not available. In the near future, however, we anticipate that these standards will be prepared through the autoxidation of $[^{13}C_{20}]$ -arachidonic acid. This will then open up the possibility of conducting the analysis of multiple isoPs using high-sensitivity SID LC–MRM-MS.

Folate/homocysteine

Folate/homocysteine metabolism is required for the methylation of many important intracellular substrates including DNA, proteins and lipids, as well for the generation of thymidylate and purines required for RNA and DNA synthesis (FIGURE 6) [60–63]. These, important cellular processes require different intracellular folate derivatives (FIGURE 6) [64]. Biomarker studies have revealed that low folate status is associated with elevated levels of circulating homocysteine (hyperhomocysteinemia) and a phenotype characterized by low RBC and serum folate, together with high homocysteine [63]. This phenotype has been implicated in diverse diseases ranging from neural-tube defects, such as spina bifida [65,66], to aging-related conditions, such as cardiovascular disease [62] and colorectal cancer [67]. The potentially deleterious effects of hyperhomocysteinemia are a consequence of inadequate levels of the methyl donor 5-methyltetrahydrofolate (5-MTHF) (FIGURE 6) [64,68]. Folate/homocysteine metabolism also modulates GSH biosynthesis through the cystathionine/cysteine pathway, which is in turn is crucial for controlling intracellular redox status (FIGURE 6) [16].

Over the past 10 years, several functional polymorphisms in enzymes involved in folate/homocysteine metabolism have been described [69]. The functional polymorphism with the most readily observed impact on phenotype is the C to T transition at nucleotide 677 (677C→T) of the *MTHFR* gene, which results in a change in amino acid residue from Ala→Val at position 222, located at the bottom of the (β α)8 barrel of the catalytic domain of the enzyme [70]. The 677T allele encodes an enzyme that is ‘thermolabile’ and less efficient at generating the 5-MTHF that is needed for both homocysteine remethylation and the generation of S-adenosyl methionine for methylation reactions (FIGURE 6). It is well established that *MTHFR* 677TT homozygotes with low folate status are at a greatly increased risk of being hyperhomocysteinemic [71]. In RBCs of individuals with this genotype, 5-MTHF comprises only 60% of total RBC folate, whereas this form predominates in the RBCs of their 677CC peers [72]. Furthermore, the *MTHFR* C677T genotype is the primary determinant of nonmethylfolate accumulation in RBCs [73]. The homozygous *MTHFR* 677TT genotype confers a significantly increased risk of many of the conditions with which a low folate, high homocysteine phenotype has been associated. However, the excess individual risk of developing such conditions in relation to their prevalence is insufficient to warrant genetic testing and counseling. Therefore, we established SID LC–MS/MS methodology to determine the degree of variation in the RBC folate phenotypes between and within the three *MTHFR* 677C→T genotypes.

Individual folates are present as folylpoly-glutamates, which prevents their release from the RBCs. Therefore, it was first necessary to convert them to the corresponding monoglutamates prior to analysis. This was accomplished using an individual’s own plasma pteroylpolyglutamate hydrolase by simply lysing the RBCs in a whole blood sample. A SID LC–MRM–MS assay was then developed and used to assess the relationship between different *MTHFR* 677C>T genotypes and RBC folates in thirty genotyped subjects. This showed that there were four different folate biomarker phenotypes that were differentially distributed between the *MTHFR* 677C>T genotype classes [74]. The RBC folate biomarker assay is now being employed for extensive phenotyping studies in human populations with defined genotypes [75,76]. Availability of this new biomarker assay will facilitate the identification subgroups of individuals with genotype/phenotype profiles that confer excess risk of defects known to be associated with dysfunction in folate/homocysteine metabolism. Such genotype/phenotype-based risk estimation may in the future be used in the conduct of clinical studies and to develop predictive and diagnostic screening protocols.

Glutathione & other thiols

Intracellular GSH provides one of the major defenses against oxidative stress in mammalian cells. During oxidative stress, reduced GSH (γ -glutamylcysteinylglycine) is converted to GSSG. GSH is the most abundant small molecule thiol in cells with concentrations in the millimolar range [6]. By contrast, concentrations of GSSG in mammalian cells are usually two orders of magnitude lower [77]. The high abundance of GSH and low abundance of GSSG helps to maintain cells under a reducing environment and prevents oxidative damage to cellular macromolecules. Hydrogen peroxide and lipid hydroperoxides undergo GSH peroxidase-mediated reduction to water and lipid hydroxides, respectively, with GSH providing the reducing equivalents [78]. GSH also readily forms adducts with a great variety of both exogenously and endogenously derived electrophilic reactive intermediates (FIGURE 1) [6, 79]. Formation of the GSH adducts is generally facilitated by GSH *S*-transferases (GSTs) and is normally considered to represent a detoxification of the relevant reactive intermediate. The resulting GSH adducts are then exported from cells by transporters [80,81]. GSH/GSSG homeostasis plays an important role in maintaining cellular redox status [77]. Changes in the half-cell reduction potential of the 2GSH–GSSG couple correlate with the biological status of the cell. Therefore, determinations of the reduction potential can be used to better understand the redox biochemistry that results from oxidative stress. The redox potential of the 2GSH–GSSG couple can be readily obtained from the Nernst Equation:

$$E_h = E_0 + \frac{RT}{nF} \times \ln \left[\frac{(GSSG)}{(GSH)^2} \right]$$

with lower (more negative) redox potential representing more reducing conditions. According to this equation, the cellular redox potential is a second-order function of GSH concentration, which means that a change in concentration of GSH even without a change in GSH/GSSG ratio could alter the cellular redox status.

Determinations of the cellular redox state require a methodology for accurately measuring the intracellular concentrations of both GSSG and GSH. Many different methods have been published for the quantitative analysis of GSH and GSSG. Despite the large variety of methods that are available for GSH and GSSG analysis, many of them still have limitations. It is well known that GSH oxidation is a serious problem, which can lead to overestimating the amount of GSSG, particularly in cell lysates where redox cycling with abundant protein thiols can readily occur (FIGURE 7). Hence, we developed a SID LC–MRM–MS method for the simultaneous quantitation of cytosolic GSH and GSSG. The method utilizes 4-fluoro-7-sulfamoylbenzofuran as a thiol-derivatizing reagent that can rapidly and completely derivatize GSH. This fast and efficient derivatization is essential in preventing oxidation and enabling reliable measurements of GSSG to be made in the presence of a large excess of GSH. The method was validated and shown to give parallel GSH standard curves (FIGURE 8) when GSSG was added in increasing amounts to cell lysates (solid triangles) and buffer samples (solid circles). The intercept on the y-axis reflected the endogenous concentration of GSH in the cell lysate and the modest rise reflected the small amount of GSH present in the GSSG that was added. These data contrast with that obtained using a conventional SID LC–MRM–MS procedure without thiol derivatization. The slope of the GSH standard curve (FIGURE 8) when GSSG was added to the cell lysate (solid squares) deviated considerably from that obtained in buffer alone (solid circles). This parallelism experiment is illustrative of a general method to determine whether a SID LC–MRM–MS method is quantifying the correct endogenous analyte or some interfering substance. The assay was able to detect subtle changes in the redox potentials of two macrophage cell lines with different phenotypes when they were treated with the endogenous reactive electrophile, 4-oxo-2(*E*)-nonenal (ONE) [16]. It is currently being

employed to examine the redox status of epithelial cells from different animal models and circulating lymphocytes from different human disease states. Elaboration of the SID LC–MRM-MS biomarker methodology for GSH and GSSG to other important cellular thiols such as homocysteine, cysteine and coenzyme A, together with their corresponding disulfides, is currently under development. This will then provide a comprehensive method to monitor thiol biomarkers of oxidative stress to complement the isoP methodology.

Glutathione adducts & their metabolites

Intracellular GSH is present in the cytosol (85–90%), with the remainder being found in the mitochondria, nuclear matrix and peroxisomes [82]. This means that intracellular reactions with endogenous electrophiles occur with a large molar excess of GSH. With the exception of bile, which contains up to 10-mM GSH [83], extracellular concentrations of GSH are relatively low. For example, GSH concentrations in plasma are in the range of 2–20 μ M [84]. GSH is involved in the formation of endogenous bioactive eicosanoids [6,85] and is a source of reducing equivalents in a number of biosynthetic reactions [86]. GSH also plays an important role in the detoxification of endogenous reactive intermediates such as ONE, 4-hydroxy-2 (*E*)-nonenal, leukotrienes, oxo-eicosatetraenoic acids (ETEs) and estrogen quinones, as well as reactive metabolites from chemicals and drugs such as acetaminophen, arylamines, clozapine, carbamazepine, 3-methylindole and valproic acid [6,85]. Furthermore, a trend was observed in the association between increased GSH adduct formation *in vitro* and increased drug-induced toxicity *in vivo* [87]. Although some reactive intermediates can form adducts with GSH directly, GST-mediated reactions generally predominate [6]. Therefore, the actual adduct structures are dependent upon the regioselectivity conferred by the particular GST that is involved. When extensive reactive intermediate formation occurs (as in the metabolism of acetaminophen), the resulting GSH adduct formation causes depletion of intracellular GSH, allowing the free sulfhydryl groups on cellular proteins to form adducts through nonenzymatic processes [88,89]. The resulting modified proteins can be immunogenic and stimulate toxic immune reactions [90]. In this manner, GSH adducts are useful biomarkers of potential drug-induced toxicity.

Positive ion LC–MRM-MS has been used extensively to detect GSH adduct biomarkers in cell culture and perfused organs. Analyses are performed either in the full scanning mode or by constant neutral-loss scanning for 129 Da (γ -glutamyl moiety) to detect GSH-derived metabolites, followed by CID and MS/MS analysis of MH^+ [91]. Negative-ion ESI methodology in combination with precursor ion scanning of m/z 272 (M-H-H₂S) can provide a complementary method for detecting GSH adducts in cellular incubations [92]. More recently, SID methodology was used in combination with an isotope pattern-dependent scanning method applied to the data acquisition of GSH adducts from drug-derived reactive metabolites [93]. In these experiments, recorded full-scan MS and MS/MS data sets were further processed with neutral-loss filtering, product-ion filtering and extracted ion chromatographic analysis to search for protonated molecules and MS/MS spectra of GSH adducts. This approach proved very effective in detecting low levels of GSH adducts, regardless of their fragmentation patterns. The characterization of GSH adduct biomarkers from drug candidates during metabolism studies is critical, since it allows potential sites and mechanisms of bioactivation to be determined. Such studies can be coupled with structural modification of the drug candidate to block sites susceptible to metabolic activation and formation of a reactive intermediate. This knowledge can lead to the development of drug candidates that are less able to bind to cellular macromolecules and reducing their potential for causing adverse events if they are ultimately used in clinical studies [79,89,94].

There is increasing interest in the use of SID LC–MRM-MS-based methodology for the analysis of endogenous GSH adducts as biomarkers of oxidative stress [95]. The 4-hydroxy-2

(*E*)-nonenal-GSH adduct can exist as a mixture of eight diastereomers [96], four of which can be separated during LC–MS analysis [97]. Analysis of GSH adduct biomarkers has been found to provide insights into the pathological role of 4-hydroxy-2(*E*)-nonenal [98]. GSH-adducts derived from 4-hydroxy-2(*E*)-nonenal have also been used as biomarkers for monitoring oxidative stress in animal models [99] and Alzheimer's disease [100]. Characterization of the lipid hydroperoxide-derived thiadiazabicyclo-ONE-GSH adduct (TOG) has provided a biomarker that can also be used to monitor intracellular oxidative stress (FIGURE 9) [97]. Using SID LC–MRM-MS analysis of TOG, we recently showed that the lipid hydroperoxide 15(*S*)-hydroperoxyeicosatetraenoic acid (HPETE) is detoxified more rapidly when it is formed in the cytosol, compared with that formed on intact lipids (FIGURE 9) [101]. Other ONE-like bifunctional electrophiles can be formed from the homolytic decomposition of lipid hydroperoxides. Dioxododecenoic acid [102] and dioxooctenoic acid [103] contain the carboxylate terminus from linoleic acid- and arachidonic-derived lipid hydroperoxides, respectively. Both dioxododecenoic acid and dioxooctenoic acid also form TOG-like GSH adducts [102]. In the future it might be possible to use these GSH adduct biomarkers to identify the particular polyunsaturated fatty acid-derived lipid hydroperoxides that are involved in the induction of intracellular oxidative stress. Leukotriene (LT)_{A4}- and oxo-ETE-derived GSH adducts (OEGs), cysteinylglycine adducts (OECs) and FOG-7 (FIGURE 4) can also be analyzed using SID LC–MRM-MS methodology [6,17,85,104]. However, these studies have been restricted primarily to cell culture models owing to the extensive metabolism of the GSH adducts that occurs *in vivo* as described below. Recently, 15-oxo-ETE was identified as a 15-lipoxygenase-derived metabolite formed by the action of 15-PGDH on 15(*S*)-HETE [104]. The 15-oxo-ETE was found to inhibit endothelial cell proliferation *in vitro* and also to rapidly form 15-OEG (FIGURE 10). It is noteworthy that 15-PGDH is downregulated in colon cancer tissues [105]. This implies that it will be possible to use SID LC–MRM-MS analysis of 15-OEG and/or its metabolites such as 15-OEC (FIGURE 4) to determine whether they are useful biomarkers of 15-PGDH deficiency and/or tumor progression.

Glutathione adducts are secreted from cells by various transporters [80,81]. These adducts are then metabolized by GGT, an enzyme that exists both bound to the extracellular surface of the plasma membrane and circulating freely in plasma [106]. The result of GGT metabolism is the formation of a cysteinylglycine adduct, as exemplified in the case of acetaminophen (FIGURE 11). The cysteinylglycine adduct can undergo further dipeptidase-mediated metabolism to a cysteine adduct. Finally, the cysteine adduct can be metabolized by *N*-acetyl transferase (NAT) to an *N*-acetyl cysteine adduct, also known as a mercapturic acid derivative. This pathway of metabolism also occurs for endogenous GSH adducts [6]. For example, the GSH adduct LTC₄ undergoes GGT-mediated metabolism to LTD₄, dipeptidase-mediated metabolism to LTE₄, which is excreted in the urine and so can be analyzed by SID LC–MRM-MS [107]. LTE₄ then undergoes NAT-mediated metabolism to *N*-acetylLTE₄ [6]. Mercapturic acid derivatives (*N*-acetylcysteine adducts) have been used as biomarkers of exposure to many exogenously derived chemicals [108–112] and endogenously derived bioactive substances [6,113–115].

DNA adducts

Stable isotope dilution LC–MS has played a critical role in the establishing DNA adducts as biomarkers for genotoxins derived from environmental chemicals, drugs and endogenous reactive metabolites [116–118]. Availability of these biomarkers has provided insight into both toxic processes and to the multiple steps involved in carcinogenesis. The action of NER and BER enzymes results in the excretion of DNA adducts in the urine (FIGURE 1) [119,120]. Urinary DNA adducts can also be obtained by hMTH1-mediated hydrolysis of oxidized bases that are present in the trinucleotide pool [121]. This makes it possible to use noninvasive LC–MS-based techniques to monitor urinary DNA adducts as both exposure and biological

response biomarkers. The SID LC–MRM–MS approach has proven to be particularly useful for assessing the presence of DNA adducts urine and in tissue DNA samples [122–128].

Over the last decade, we have conducted a series of studies using SID LC–MRM–MS to develop biomarkers of lipid hydroperoxide-induced DNA damage [19]. During these studies we discovered that COX-2 can convert arachidonic acid into 15(*S*)-HPETE, which then undergoes intracellular reduction to 15(*S*)-HETE [129]. We reasoned that, in settings of oxidative stress, where reducing pathways are compromised, 15(*S*)-HPETE might survive long enough to undergo homolytic decomposition to DNA-reactive bifunctional electrophiles that would then generate biomarkers of lipid hydroperoxide-derived DNA damage. In previous studies we had shown that 15(*S*)-HPETE (and other HPETEs) decomposes to 4-hydroperoxy-2(*E*)-nonenal (HPNE) and ONE and that they form ϵ - and H ϵ -DNA adducts, respectively (FIGURE 12). Importantly, COX-2 is localized to the nuclear membrane, which maximizes the potential translocation of any bifunctional electrophiles into the nucleus. Using RIES cells that stably express COX-2, it was possible to show that basal endogenous production of 15(*S*)-HPETE resulted in the formation of H ϵ dGuo adducts [129]. As predicted from *in vitro* studies [130, 131], there was a dose-dependent increase in endogenous H ϵ dGuo adduct formation when vitamin C was added to the RIES cells [129]. H ϵ dGuo-adduct formation and 15(*S*)-HPETE biosynthesis were both inhibited by a selective COX-2 inhibitor and by aspirin, a nonselective inhibitor.

The role of COX-2-induced DNA damage *in vivo* was studied using the Min mouse model of colon carcinogenesis [132]. Min mice provide a useful model to explore the consequences of increased *in vivo* COX-2 expression. The expressed dominant negative 95 kDa-truncated APC protein in Min mice is thought to prevent processing of β -catenin. This results in increased nuclear β -catenin, which, together with T-cell factor, activates expression of COX-2. In our experiments, DNA was isolated from the entire small intestine of C57BL/6J and C57BL/6JAPC^{min} mice and then hydrolyzed in the presence of [¹⁵N₅ or ¹⁵N₃]-internal standards. In separate experiments, it was shown that the H ϵ -DNA adducts were not generated as an artifact during the isolation and hydrolysis by adding [¹³C¹⁵N]-DNA. The use of labeled DNA provides a general method to confirm whether an endogenous DNA adduct can arise artifactually during isolation and hydrolysis of DNA or during LC–MS analysis. SID LC–MRM–MS employed in these studies showed that H ϵ dGuo was increased in the Min mice compared with the wild-type mice, while H ϵ -2'-deoxycytidine (dCyd) was much less abundant. This suggested that DNA repair pathways favored H ϵ dAdo and H ϵ dCyd over H ϵ dGuo in the DNA and that the excised adducts might be excreted in the urine.

DNA is very efficiently repaired and only a limited amount of the total DNA damage results in permanent mutations (7). This means that DNA adducts, which are repaired and appear in the urine, can be used as surrogate biomarkers of DNA damage [119,120]. SID LC–MRM–MS in combination with immunoaffinity purification can provide the specificity and sensitivity for analysis of endogenous and exogenous DNA adducts in mammalian tissue and biofluid samples [133–137]. We anticipate that immunoaffinity purification in combination with highly sensitive SID LC–MRM–MS, will be required for the specific quantification of urinary H ϵ -DNA adducts. Since the finding that H ϵ dCyd is highly mutagenic in human cells [138,139], the quantification of urinary H ϵ -DNA adducts could potentially allow the identification of human populations at risk of cancer through lipid peroxidation-mediated DNA-damage.

Elevated levels of oxidatively damaged DNA bases have been measured as biomarkers to monitor numerous diseases including Alzheimer's disease [140], Parkinson's disease [141], autoimmune diseases [142], cardiovascular diseases [143–145] and cancer [146–148]. It has been suggested that such damage plays an important role in the etiology of these diseases [149,150]. Oxidative DNA damage is also increased through exposure to environmental

chemicals [117,151]. In addition, there is some evidence of a link between ingestion of antioxidants and reduced oxidative damage to DNA [152]. 8-oxo-dGuo is by far the most extensively studied lesion that arises as a consequence of oxidative DNA damage [144,150,153]. Unfortunately, it is extremely easy to induce artifactual formation of 8-oxo-dGuo in cellular DNA during the isolation and hydrolysis procedures that are commonly used [144,154]. Therefore, numerous studies have been conducted to address the problem of quantifying 8-oxo-dGuo in DNA, including the important contributions of the Cadet group [155,156] and the European Standards Committee on Oxidative DNA Damage [157–163]. The comet assay has predicted that basal endogenous concentrations of 8-oxo-dGuo in cellular DNA should be at the level of three to five molecules of 8-oxo-dGuo/10⁷ dGuo [163]. Physicochemical methods developed to analyze 8-oxo-dGuo have consistently obtained values that are almost an order of magnitude higher at 34–37 molecules of 8-oxo-dGuo/10⁷ dGuo [163]. These elevated values were obtained even when highly specific and sensitive methodology based on immunoaffinity purification and stable isotope dilution LC–MS was used [134], which suggests that artifactual formation of 8-oxo-dGuo might occur during DNA isolation and hydrolysis [144,154]. Significant advances have been made to the comet assay by introducing the use of hOGG1 to specifically remove 8-oxo-dGuo from the DNA so that it is now possible to reliably measure basal cellular oxidative DNA damage [164]. hOGG1 is known to excise both 8-oxo-dGuo and 8-oxo-Gua, producing a single -strand break. It will also remove formamidopyrimidine-dGuo, derived to yield a single-strand break. Unfortunately, a linear dose response to KBrO₃-induced oxidative DNA damage was not obtained, even with this significantly improved method [164]. This implies that lesions other than 8-oxo-dGuo can be detected by the comet assay, even when the assay is used in combination with the selective hOGG1 enzyme.

We developed an immunoaffinity purification SID LC–MRM-MS assay for the analysis of 8-oxo-dGuo in cellular DNA. This technique provides the ultimate in analytical sensitivity and specificity that can be attained using physicochemical methodology for analysis of endogenously generated lesions in DNA [165]. Yet, even when this assay was used with conventional DNA isolation procedures, basal levels of 8-oxo-dGuo were similar to those described previously [134]. These experimentally generated adducts were only reduced to acceptable levels when extreme care was taken during sample preparation. First, artifactual generation of 8-oxo-dGuo was reduced when cellular DNA was isolated using commercially available DNAzol (which contains high concentrations of guanidine thiocyanate) [166] instead of other chaotropic- [167,168] or phenol-based [169] extraction methodology. Second, Chelex-treated water was used to reduce transition metal-ion contamination in the buffers that were used [130]. Last, deferoxamine was added to complex any residual transition metal ions that were still present in buffers following the Chelex treatment. These steps prevented Fenton chemistry-mediated generation of ROS and artifactual oxidation of DNA bases [144,154], resulting in the levels of 8-oxo-dGuo obtained using these precautions to agree with levels previously determined using the comet assay (FIGURE 13)[164]. The LC–MS assay was shown to sensitively and specifically measure 8-oxo-dGuo formed in the DNA of human lung bronchoalveolar H358 cells treated with the redox-active benzo[a]pyrene (B[a]P)-7,8-dione, a reaction product of aldoketo reductase-mediated polycyclic aromatic hydrocarbon-*trans*-dihydrodiol activation [170]. The immunoaffinity SID LC–MRM-MS assay has now been elaborated to the analysis of urinary 8-oxo-dGuo where it was able to show a significant increase in the urine of smokers [MANGAL D, MESAROS A, BLAIR IA, UNPUBLISHED DATA]. Therefore, it holds promise as a biomarker to identify subjects at increased risk for oxidative stress.

Exogenously derived DNA adducts, such as those arising from environmental chemicals, can also be employed as biomarkers of exposure [118]. Polycyclic aromatic hydrocarbons, such as B[a]P, are ubiquitous environmental chemicals found in vehicle exhaust emissions and tobacco smoke [171]. A substantial research effort has been conducted on B[a]P-mediated

DNA adduct formation to provide biomarkers of exposure to B[a]P [172]. We developed SID LC–MRM–MS methodology in order to quantify eight anti-7,8-dihydroxy-9,10-epoxy-7,8,9,10-tetrahydro-B[a]P(B[a] PDE)-derived DNA adduct biomarkers in four H358 human bronchoalveolar cell lines with different phenotypes [173]. In CYP1A1/1B1-induced H358 cells exposed to (±)-B[a]P-7,8-dihydrodiol, (+)-anti-trans-B[a]PDE-N²-dGuo was the major DNA adduct [174]. Surprisingly, the greatest amount of (+)-anti-trans-B[a]PDE-N²-dGuo was formed in the control H358 cells. These data raise the intriguing possibility that CYP1A1/1B1 might be protective against (+)-B[a]PDE-mediated DNA damage. Going forward we will assess the relationship between cellular DNA and urinary DNA adducts. Such studies will require the sensitivity offered by new generation LC–MS instrumentation, which will make it possible to conduct analyses in circulating lymphocytes.

Protein biomarkers

Introduction

The need for SID MS quantification of protein and peptide biomarkers arises from both basic and clinical researchers. Driving these assays is a push for increased selectivity and sensitivity during both the discovery and validation phases of biomarker development. Currently, the great majority of validated procedures used to quantify biological macromolecules are based on immunoassay methodologies. Although these antibody-based techniques are exceptionally sensitive (typically picogram per milliliter range), long development time, variability and crossreactivity have stimulated the interest in an equally sensitive, though more selective and controlled, assay platforms. The proven capability of SID MS, considered the ‘gold standard’ for small-molecule quantification, offers an attractive solution to the limitations associated with antibody-based detection. Peptide and protein measurement by SID MS does, however, pose unique challenges and the field is striving to match the lower limits of quantification typical of those seen using ligand-binding assays, such as ELISAs [3,175–180].

There are several key advantages characteristic of SID MS when compared with antibody-based methods alone. As noted by Carr and Anderson, analyte detection by MS can be structurally unambiguous, essentially offering absolute selectivity [176]. Even in instances where selectivity may be compromised in a LC–MS/MS experiment (i.e., a co-eluting analyte has isobaric precursor and product m/z), these interferences can often be identified and the LC–MS method altered to relatively quickly eliminate the problem. The ability to assay multiple peptides in a single LC–MS/MS experiment is another strength not offered by standard immunoassays. Several SID MS experiments have been conducted, showing the ability to quantify peptides correlating to multiple proteins in a single run [176,179,181]. This multiplexing significantly increases assay throughput, which may prove critical to realizing SID MS of macromolecules in the clinical setting, where biomarker panels will most likely be the reality [182]. Last, unlike immunobased detection, which is reliant on the binding interaction of two macromolecules, LC–MS detection is direct and only depends on the chemical nature of the analyte. This characteristic of SID MS often results in excellent assay linearity over several orders of magnitude, whereas ligand-binding assays are often limited in dynamic range and often show a nonlinear responses [183]. The utility of SID MS in all stages of biomarker development, from identification, through validation and ultimately into clinical practice has proven this technology holds a central role in biomarker research [184–187]. The extension of SID MS to protein quantification, however, is an evolving field and its application is not trivial. In this review, we will cover the historical development, critical challenges and current state of the field, highlighting select publications during the discussion.

A typical workflow for the analysis of peptides and proteins by a ‘bottom-up’ approach involves the enzymatic cleavage of the large molecule followed by the analysis of the resultant peptides by LC–MS/MS. Traditionally, the endopeptidase trypsin is used to digest a protein

into its constituent peptides. Trypsin cleaves proteins C-terminal of arginine and lysine, except when blocked by an adjacent C-terminal proline residue [188]. The average tryptic peptide is ideal for LC–MS/MS sequencing, due to its relatively small length and single C-terminally located basic residue. Positive-mode ESI of tryptic peptides will normally produce $M+2H^+$ and $M+3H^+$ species with m/z between 400–1000 and, upon CID, will show a predominant y- and b-fragment ion series [189]. A protein is qualitatively identified in such an experiment when one or several peptides are sequenced that are unique to the protein [190]. The bottom-up approach is similarly used for protein quantification, where the amount of one or several unique peptides corresponds stoichiometrically to the amount of the whole protein. Such peptides, when they also ionize well by positive ESI, are sometimes referred to as ‘prototypic’ or ‘signature’ peptides [191,192]. Protein quantification using a bottom-up workflow involving protease digestion is shown in FIGURE 14. In FIGURE 14, the tryptic peptide AVFPSIVGR, corresponding to human β -actin, was quantified following trypsin digestion of the whole protein. The isotopically labeled peptide AVFPSIVG[$^{13}C_6^{15}N_4$]R was used as an internal standard. The peptide pair in this example was analyzed using a TQ instrument operated in the MRM mode, where the transition masses are the peptides’ monoisotopic doubly charged $M+2H^+$ ions to singly charged y_6 and y_7 ions, respectively. The resultant chromatographic trace for each peptide is then represented by the sum of both transitions (FIGURE 14).

Historical development of SID LC–MS/MS-based approaches

In 1996, Barr *et al.* reported the first attempt to quantify a protein by ‘bottom-up’ isotope dilution-MS methodology [193]. In this seminal study, apolipoprotein (apo)-A-1 was quantified using isotope-labeled reference peptides synthesized using standard *N*-(9-fluorenyl) methoxycarbonyl (F_{moc}) chemistry. The reference peptides incorporated [D_3]-leucine and [$^{13}C_3$]-alanine and corresponded to three apo A-1 tryptic peptides. quantification of the standard reference peptides was accomplished by nitrogen analysis and amino-acid analysis. The experimental design entailed the addition of two reference peptides prior to tryptic digestion. One peptide provided a primary measurement, while the other was used to confirm this calculated amount. A third standard peptide was added postdigestion and was used to determine an absolute recovery value. The results showed the assay to be exceptionally precise, with approximately 4% coefficient of variation (% CV). The goal of this assay was to develop an alternative approach to immuno-based detection for the quantification of reference protein. A noted caveat of the experiment was the requirement that the sample material must be abundant, well characterized and pure. In this analysis, material quantified was purified apo-A1 (~ 1 mg/ml of a certified reference material) that had been extensively characterized and shown to be free from degradation and modification. Since the peptides quantified in the study contained modifiable amino acids and did not control for the rate of trypsin hydrolysis, the ability of the assay to quantify trace amounts of apo-A1 in a complex biological matrix would be questionable. In fact, although the assay was extremely reproducible, 44% of the experimental imprecision was attributable to differences between sample digestions, rather than purely analytical error.

Following the work of Barr *et al.*, in independent studies, the laboratories of Aebersold and Chait used isotope-labeling strategies to quantify proteins expressed in yeast [194]. Using an isotope-coded affinity tag (ICAT), working in the Aebersold group, Gygi *et al.* were able to selectively label the sulfhydryl groups of cysteine-containing proteins with either deuterium-labeled (D_8) or unlabeled reagent. Tagging two experimentally conditioned groups of yeast with the ICAT it was possible to perform relative quantification and determine differential protein expression between the two groups. A limitation to the technique, however, was that only cysteine residues were labeled and, hence, only cysteine-containing peptides served in the analysis [194]. In an alternative approach, Oda *et al.*, in the Chait laboratory, uniformly labeled yeast with ^{15}N . In these experiments, two batches of yeast, one in natural

abundance ^{14}N (99.6%) and the other in ^{15}N (>96%) enriched cell growth media were prepared. The yeast strains used were mutants that expressed or were deficient in, a G1 cyclin protein that regulated G1→S transition in budding yeast. The comparison of the yeast strains, using this labeling strategy, not only revealed alterations in protein expression levels, but showed differences in phosphorylation on important signaling proteins. As the authors' note, the gel fractionation performed would not alter the observed unlabeled to labeled peptide ratios, since the ^{15}N -labeled proteome was mixed with the unlabeled proteome prior to sample workup [195]. The added control offered by the addition of the labeled proteome during early stages of sample preparation is a great strength of the technique.

Protein biomarker discovery & global relative quantification

Today, most LC–MS-based biomarker-discovery experiments using stable isotopes are accomplished by either tagging with chemical reagents or metabolic labeling of cells in culture. The majority of current chemical tagging techniques label peptides on the N-terminus and epsilon-amino of lysine using *N*-hydroxysuccinimide chemistry. Isobaric mass tags are particularly popular for biomarker-discovery experiments since the mass difference between labeled forms occur on fragment ions rather than on the parent analyte [196,197]. In this manner, MS spectra are not complicated and the peptide MS signal is increased by multiplexing. These reagents and their application have been reviewed [198,199]. Notably, demonstrating the multiplexing capabilities and utility of the iTRAQ reagent, Zhang *et al.* monitored the time-resolved tyrosine phosphorylation events downstream of the epidermal growth factor receptor [200]. Following on the promise of this earlier work, the group more recently was able to probe the EGF receptor signaling network in glioblastoma cell lines, which revealed a cross-activation between the epidermal growth factor receptor and c-MET receptor pathways. This result led to the exciting and clinically relevant finding that the dual inhibition of both receptors had greater cytotoxicity than either treatment alone on cultured glioblastoma cells [201]. This and similarly successful experiments showed the utility of the chemical labeling approach in biomarker discovery type experiments [202–206]. A limitation to this application, however, is that the isotope label is introduced postprotein extraction, meaning extreme care must be taken to ensure error is not introduced during sample preparation. Additionally, other factors, such as mixed MS/MS spectra, resulting from the co-isolation of a contaminating ion during MS/MS acquisition, can lead to aberrant quantification [207].

Another powerful technique employed in biomarker discovery experiments is stable isotope labeling by amino acids in cell culture (SILAC). Following the work performed in the Chait laboratory, in which uniformly ^{15}N -labeled yeast were prepared [195], Ong *et al.* extended the concept to mammalian cells and termed the method SILAC [208]. Rather than uniform ^{15}N labeling, SILAC metabolic labeling introduces specific ^{13}C - and ^{15}N -labeled amino acids. In a typical SILAC workflow, an immortalized cell line is grown in media containing unlabeled essential amino acids or media containing labeled counterparts (i.e., [$^{13}\text{C}_6^{15}\text{N}_2$]-lysine and [$^{13}\text{C}_6^{15}\text{N}$]-leucine). Unlabeled and labeled cells are then mixed following a control or experimental treatment and a bottom-up LC–MS/MS analysis is performed [209]. Unlike isobaric labeling, SILAC results in a mass difference at the MS level. Due to this feature, MS spectra do become more complicated upon multiplexing. To improve quantification at the MS level in such cases, high resolution MS instruments are often employed [210,211]. There are numerous examples illustrating the successful identification of candidate biomarkers using SILAC methodology [212–216]. In a recent study, Pan *et al.* in the laboratory of Matthias Mann, described the global evaluation of phosphorylation alterations in response to kinase inhibition [217]. A multiplexed SILAC method using 'light' ([$^{12}\text{C}_6$]-arginine, [$^{12}\text{C}_6$]-lysine), 'medium' ([$^{13}\text{C}_6$]-arginine, [$^2\text{H}_4$]-lysine) and 'heavy' ([$^{13}\text{C}_6^{15}\text{N}_4$]-arginine, [$^{13}\text{C}_6^{15}\text{N}_2$]-lysine) cell growth media was developed. The group was able to relatively measure thousands of phosphorylated peptides from the three SILAC cell populations, which were exposed to

control treatment, EGF stimulus or growth factors in the presence of kinase inhibitors. Following a peptide fractionation and phospho-peptide enrichment step, the fractions were analyzed by LC-MS and the response pattern of the three labeled proteomes was determined. In this manner, phosphorylation events were delineated into nonresponders, growth factor responders and those activated by growth factors but suppressed by inhibitors [217]. As SILAC methodologies become more streamlined and data analysis software is more effectual [218], a greater variety of candidate biomarkers will be uncovered.

A critique often encountered when comparing chemical labeling to metabolic labeling strategies is that only the former is applicable to investigating primary source materials such as biofluids, tissues and biopsies. Following the experimental workflow outlined above, this is indeed true. However, the Blair laboratory has shown the utility of using the SILAC-labeled proteome as a reference standard that can be applied to any source material. Proof of principle for this approach was established in the work of Yu *et al.* [216]. The CAPAN-2 pancreatic cancer cell line was labeled by SILAC and secreted proteins from the conditioned media were collected. This stable isotope-labeled proteome (SILAP) standard was added to pooled pancreatic cancer or control serum (FIGURE 15). Over 100 differentially expressed biomarker candidates were identified, with the proteins intercellular adhesion molecule 1 and basal cell-adhesion molecule independently validated. This approach was extended by Rangiah *et al.* by developing a MRM method combined with a SILAP standard to study biomarkers in the Apcmin mouse, a colon cancer model [219]. Candidate biomarkers were identified in the secreted proteome of the CT26 colon cancer cell line and an MRM method was developed to measure 12 biomarkers of interest. The CT26-derived SILAP standard was added to pooled Apcmin mouse or normal serum samples. Levels of all 12 biomarkers could be quantified with differential expression validated by Western blot analysis for five biomarkers. Demonstrating the versatility of this approach, Shah *et al.* pooled SILAC-labeled human epithelial endocervical-1 and vaginal cell secretomes. This SILAP standard was then used in the characterization and relative quantification of proteins present in patient cervicovaginal fluid in a preterm birth study. The result of the investigation showed that three proteins were significantly elevated in preterm birth cases (desmoplakin isoform 1, stratifin and thrombospondin 1 precursor), providing a foundation for further validation experiments [220]. Limiting the analysis to proteins secreted and overexpressed in the cell line model makes it possible to exclude acute phase proteins and other abundant proteins found in serum or cervicovaginal fluid, while simultaneously focusing on proteins with biological relevance to the disease of interest.

Protein biomarker quantification: targeted approaches

Gygi and colleagues have extended the methodology of protein quantification by stable isotope-labeled peptide standards to cellular proteins resolved on an SDS-PAGE gel and termed the technique AQUA [221]. In their studies, an isotope-labeled phosphopeptide standard was used to quantify endogenous amounts of a phosphorylated protein. This innovation offered accessibility to many important protein biomarkers related to kinase signaling [222]. Additionally, the application of the technique to proteins separated by SDS-PAGE gel provided an attractively simple, established and widely applicable means of reducing sample complexity. AQUA has become a critical technique in the validation of candidate protein biomarkers [186].

An important development of the AQUA technique was reported by Mayya *et al.* in a 2006 study that monitored the cell cycle-dependant phosphorylation of CDKs on multiple residues. These experiments used synthetic peptides representing the CDK regulatory sites T¹⁴ and Y¹⁵ contained in the tryptic peptide IGE^T¹⁴Y¹⁵GVVYK in three phosphorylated and one unphosphorylated form. In this manner, it was possible to quantify not only total CDK, but

also calculate the amount of the phosphorylation forms present [223]. Although the experiments were not able to control for CDK protein immunoprecipitation, results were reproducible and in agreement with quantitative immunoblotting. Many regulatory phosphorylation sites, such as those that occur within the catalytic loop of a kinase, exist on multiple residues within a short stretch of amino acid residues. Therefore, quantification of all phosphorylated and unphosphorylated forms of these peptides can provide valuable information indicating activation state [224,225].

In their initial study, Barr *et al.* highlighted many of the confounding issues associated with the AQUA technique, chief among them being reference material purity/stability and variability of protein hydrolysis. For a peptide standard quantified by SID MS to match its value calculated by amino acid analysis, any resolubilization of the peptide post-amino acid analysis must be complete. Additionally, no modifications to the standard can occur (e.g., oxidation or carbamylation) during peptide storage or early sample preparation. Lastly, degradation and absorptive losses incurred during peptide storage must be minimized [226, 227]. Recent reports have investigated the accuracy and robustness of SID MS of proteins using standard peptides [227,228]. Mirzaei *et al.* compared peptide standards generated synthetically with those expressed from a synthetic gene *in vitro*. In this second approach, termed quantification concatamers (QconCAT), a series of targeted tryptic peptides are concatenated into a single plasmid and expressed using a bacterial transcription/translation system *in vitro*. The expression of the QconCAT is performed in a medium containing isotopically labeled amino acids (e.g., [$^{13}\text{C}_6^{15}\text{N}_2$]-lysine and [$^{13}\text{C}_6^{15}\text{N}_4$]-arginine, used in place of the unlabeled species). A series of 25 tryptic peptides were assembled into QconCATs or synthetically expressed individually. The ability of the two standard methods to report equimolarity between the 25 peptides was then investigated. Multiple resolubilization buffers and digestion parameters were compared, with the result that neither method was proven optimal and both suffered from a combination of the above issues. Interestingly, synthetic peptides appeared to suffer from solubility and modification, while the QconCAT peptide showed variability in digestion and required extensive digestion optimization [229]. This study emphasizes that neither a single nor a concatenated peptide is a true authentic standard for a whole protein and, unless great care is taken, accuracy can suffer.

Protein biomarker validation in complex biofluids

Quantifying trace amounts of protein in complex biofluids, such as tissue, serum and plasma, using SID MS requires sample preparation strategies that enrich the analyte and/or reduce interferences [230,231]. It is estimated that 99% of serum's total protein mass is due to the top 20 most abundant protein species [232]. The dynamic range of protein abundance in serum may cover eleven orders of magnitude with albumin, the most abundant protein, being present at the extraordinarily high concentration of 60 mg/ml [232]. A recent review of the literature identified that direct analysis of plasma is possible for proteins in the microgram per milliliter range [233]. Bonder *et al.* demonstrated the quantification of the candidate prostate cancer biomarker Zn- α 2 glycoprotein directly in serum without fractionation or abundant protein depletion. The results showed a linear dynamic range of 0.32–10 $\mu\text{g/ml}$ and a inter- and intra-assay imprecision of less than 6% C V. The assay was able to quantitatively differentiate between pilot groups of normal men (3.65 $\mu\text{g/ml}$) and those suffering from nonmalignant prostate disease and PCa (6.21 and 7.59 $\mu\text{g/ml}$ respectively). This assay is unique among the many LC–MS assays reported, in that a robust LC configuration utilizing analytical flow rates (250 $\mu\text{l/min}$) was used [234]. Yet, although it is possible to measure a selected protein directly in serum, the level of sensitivity required to quantify many clinically relevant protein biomarkers, such as prostate-specific antigen (PSA) and cardiac troponins, which exist in the low nanogram per milliliter range, require additional sample preparation.

Similar to the benefits seen when quantifying small molecules following solid-phase extraction, peptide fractionated prior to LC–MS can dramatically improve the analytical performance. Coupled to reversed-phase separation, strong cation exchange (SCX) chromatography can serve as an effective and orthogonal form of peptide chromatography. SCX is often utilized in qualitative proteomics experiments to increase proteome coverage [235,236]. Recently, the technique has been used to quantify low level serum biomarkers. In their work, Keshishian *et al.* systematically explored the impact of abundant serum protein depletion followed by SCX fractionation on the ability to quantify a panel of six proteins spiked into human serum. The report indicated a ten- to 40-fold increase in signal-to-noise ratio, comparing fractionated to complex samples. As stated by the authors, the benefits of SCX fractionation was apparently achieved through both a reduction in background noise and an increase in analyte signal attributable to more target analyte being loaded on column per injection [180].

Since the report of Keshishian *et al.* illustrating the ability to quantify serum proteins in the 1–10 ng/ml range using limited fractionation [180], other studies have used the same or similar workflows with success. For example, PSA was quantified over 5–40 ng/ml in serum with excellent accuracy and precision using a combination of abundant serum protein depletion and mixed cation exchange fractionation. A benefit noted by the authors of this assay over existing methods using immunoenrichment of the PSA is that results are not complicated by free and bound forms of PSA. In the later case, some PSA species may have their antibody recognition region buried within a complex resulting in high assay variability. The issue of bound PSA can also cause disagreement between laboratories using antibodies that target different epitopes [237]. In addition to analytical benefits, an antibody-free method may offer an economic advantage due to a reduction in reagent usage and time per assay.

Another means to reduce sample complexity at the peptide level is the technique of stable isotope standards and capture by antipeptide antibodies (SISCAPA). Anderson *et al.* introduced the SISCAPA technique in 2004 [238], and several excellent studies have recently established the approach [177–179,239]. Using the SISCAPA method, Hoofnagle *et al.*, working in the Heinecke laboratory, successfully targeted three thyroglobulin peptides for quantification in patient serum. Thyroglobulin is a clinical tumor marker of thyroid carcinoma; however, commercial ligand-binding assays suffer from endogenous antibody interferences leading to a high false positive rate. One of the three peptides targeted was capable of quantifying thyroglobulin with the necessary level of sensitivity (2.6 ng/ml) and was in good correlation with a commercial immunoassay [177]. Further extending the usefulness of SISCAPA, Kuhn *et al.*, working in the Carr laboratory, verified the ability to multiplex assays. In their work, protein quantification was performed on peptides correlating to troponin I (cTnI) and IL-33, an established and emerging marker of cardiac injury and cardiovascular health. This assay showed satisfactory precision and a limit of LOQs of approximately 3 ng/ml (cTnI) and 5 g/ml (IL-33). Good correlations were found when patient cTnI levels determined by a commercially available immunoassay were compared with the amount calculated using the SISCAPA assay [179].

Although fractionation steps at the peptide level are beginning to show great promise, protein immunoprecipitation prior to SID MS is exceptionally sensitive [240]. N-terminal probrain natri-uretic peptide (NTproBNP), a clinically validated marker of heart failure, was measured following protein immunoprecipitation by Berna *et al.* In these experiments, NTproBNP was immunoprecipitated from serum, digested with trypsin and then quantified using a single isotope-labeled NTproBNP tryptic peptide. Following assay optimization, a lower LOQ was determined to be 0.1 ng/ml with a less than 16% CV. The assay was used to monitor drug-induced cardiac hypertrophy in rats, showing a quantitative correlation between heart weights and circulating NTproBNP concentration. Results indicated that following a 14-day drug

treatment, the heart weight increased by 34%, while the NTproBNP amount increased from 0.107 (± 0.009) to 0.332 (± 0.03) ng/ml [175]. These data are not only remarkably accurate but are an order of magnitude more sensitive than those seen using other LC–MS/MS methodologies.

Highlighting the limited degree of control offered by a standard peptide alone, in their report, Berna *et al.* determined that only approximately 63% of NTproBNP protein was recovered. The authors indicate this recovery value was most likely affected by immunoprecipitation, digestion and ion suppression [175]. Similar levels of protein recovery following immunoprecipitation from serum have been reported [241]. To ultimately control these variables and provide traceability across differing LC–MS platforms and enrichment antibodies, a whole protein standard is required. The production of isotope-labeled whole protein standards was first developed and utilized in NMR analysis [242,243]. The cell-free synthesis of isotopically labeled proteins described by these NMR investigations has been used by our laboratory and others. The production of the cell-free recombinant protein is accomplished using extracted bacterial translation and transcription machinery and, therefore, requires that the target protein be inserted into a suitable expression plasmid. Isotopically labeled amino acids of the experimenter's choice can then be added to the *in vitro* reaction to produce a variety of labeled species. Brun *et al.* investigated the accuracy of using isotope-labeled protein standards (PSAQ), as compared with AQUA or QconCAT, in the quantification of staphylococcal enterotoxins. The group investigated the ability to quantify the toxins in water and urine following a protein capture and fractionation method. It was shown, as expected, that both AQUA and QconCAT quantification severely underestimated the amount of toxin present, while the measurement using the PSAQ calibrant was acceptable [244]. Other groups have since implemented the PSAQ approach in the quantification of therapeutic antibodies and in discerning closely related protein isoforms [245,246].

The Blair laboratory has significantly extended the PSAQ application with the creation of a post-translationally modified PSAQ standard. To facilitate our studies, full-length isotope-labeled and phosphorylated human focal adhesion kinase (FAK) protein was produced *in vitro*. The focus of our investigation was the Src kinase specific phosphorylation within FAK's catalytic loop region. The amounts of total FAK protein, as well as phosphorylated forms, were quantified using standard peptides. Using this characterized reference material, it was then possible to calculate the amount of endogenous FAK and determine the absolute amount of phosphorylation present within the catalytic loop region relative to the amount of total FAK protein following immunoprecipitation [247]. Applying this technique to a cell culture model, we were able to correlate the amount of FAK catalytic loop phosphorylation with Src kinase inhibition, confirming FAK phosphorylation as a biomarker of drug exposure in culture [225]. These experiments are the first use of a post-translationally modified isotope-labeled whole protein standard and its usage provided novel data unattainable via other established methods.

Conclusion & future perspective

Improved sensitivity using lower flow rates and newer generation mass spectrometers should make it possible to use readily accessible circulating lymphocytes as sources of potential *in vivo* biomarkers. This will prove to be particularly useful for DNA adduct analyses as it will be possible to compare DNA adducts in cellular DNA with urinary DNA adduct metabolites (FIGURE 1). These data will ultimately link the activity of DNA-repair enzymes and DNA-adduct formation in the prevention of tumorigenesis [248]. Increasing instrument sensitivity will also revolutionize the approach to folate analysis (FIGURE 6), allowing analyses to be conducted in metabolically competent cells rather than the current use of RBCs [76]. Lastly, the simultaneous quantification of free and oxidized thiols in the lymphocytes will make it

possible to assess interindividual differences in oxidative stress and how they relate to the folate phenotype [74].

In addition to instrumentation, methodologies to produce preionized derivatives will lead to increases in assay sensitivity [7]. This will hopefully prove particularly useful in the analysis of circulating steroids, making it possible to rigorously evaluate the role of CYP-mediated estrogen metabolism [24] and facilitate the development of new models of breast cancer risk [25]. The use of preionized derivatives as described by Hong and Wang could also have applicability for the analysis of urinary DNA adducts [249]. Although LC–ECAPCI–MS still provides the best sensitivity for analysis of lipids and permits chiral analyses to be conducted [5], it is not clear that reduced flow rates and modern TQ instruments will improve ECAPCI sensitivity still further. The requirement for normal phase solvents in order to accomplish chiral separations is currently very restrictive and will most likely stimulate the development of alternative approaches to improving ESI sensitivity, such as using preionized carboxylate derivatives.

The use of appropriate protein standards in SID LC–MRM–MS assays is critically important and is an active area of research within the field of protein biomarkers. This is because the standard will define the scope and accuracy of an assay. While global relative quantification can be accomplished using an isotopically labeled proteome, no measurement in the analysis is absolute. Conversely, the measurement of a specific protein using labeled synthetic peptides can be absolute, but a specific protein or group of proteins is targeted. The choice of standard should be determined based on the fitness for the experiment at hand and no single methodology will likely serve every paradigm. During discovery and the early qualification stages of biomarker development, many candidates need to be screened for their usefulness. Hence, many measurements need to be made with acceptable analytical precision, but no specific degree of accuracy is required. Later stages of biomarker validation will be targeted by nature. As a biomarker assay moves into the clinical setting, the use of a traceable reference standard will become more critical and serve to benchmark interlaboratory comparisons.

Executive summary

- The ability to conduct validated analyses of small molecule (steroids, lipids, folates, glutathione and other thiols, glutathione adducts and their metabolites and DNA adducts) and protein biomarkers is critically important in order to establish their sensitivity and specificity for identifying particular diseases.
- The use of stable isotope dilution LC–MS/MS provides the highest analytical specificity possible for quantitative biomarker determinations.
- Quantitative studies normally require the high sensitivity that can be obtained by the use of multiple reaction monitoring (MRM)–MS.
- Additional increases in sensitivity can also be attained using preionized derivatives. This will prove to be particularly useful in the analysis of circulating steroids, making it possible to rigorously evaluate the role of cytochrome-mediated estrogen metabolism in breast cancer risk.
- The use of appropriate protein standards in SID LC–MRM–MS assays is critically important and is an active area of research within the field of protein biomarkers.
- The choice of a protein standard should be determined based on the fitness for the experiment at hand and no single methodology will likely serve every paradigm.

- As protein biomarker assays move into the clinical setting, the use of an appropriate reference standard will become more critical for rigorous validation and to benchmark interlaboratory comparisons.

Glossary

BIOMARKER	Characteristic that is objectively measured and evaluated as an indicator of normal biologic processes, pathogenic processes or pharmacologic responses to a therapeutic intervention
BIOANALYTICAL VALIDATION	Procedure that demonstrates that a method will successfully meet or exceed the minimum standards recommended in the US FDA guidance for accuracy, precision, selectivity, sensitivity, reproducibility and stability
STABLE ISOTOPE ANALOG	Compound of identical structure in which ideally three or more atoms have been replaced with ^{13}C , ^{15}N or a combination of both
DNA ADDUCT	Covalent attachment of a reactive intermediate to a DNA base
SID METHODOLOGY	Normally involves the use of a stable isotope-labeled internal standard, which is spiked into a sample at a known concentration. The response ratio between the analyte and labeled compound obtained by LC-MRM-MS can then be interpolated onto a standard curve to calculate the absolute amount of analyte in the unknown sample
MRM	In a triple quadrupole mass spectrometer a precursor ion is preselected and resolved in Q1, fragmented by collision induced dissociation in Q2 and the resultant product ion is analyzed in Q3. This is conducted for multiple analytes and their internal standards
STABLE ISOTOPE CARRIER EFFECT	The ability to modulate the effects of nonspecific binding to surfaces during isolation and analysis of an analyte
ESI	Process in which ionized species in the gas phase are produced from a solution via highly charged fine droplets, by means of spraying the solution from a narrow-bore needle tip at atmospheric pressure in the presence of a high electric field (1,000 to 10,000 V potential)
APCI	Chemical ionization that takes place using a nebulized liquid and atmospheric pressure corona discharge in an atmospheric pressure ionization source
ECAPCI	Process in which low energy electrons generated under APCI conditions are captured by the analyte to generate a radical anion. This process often occurs through dissociative electron capture, which results in the loss of a radical and generation of a negative ion
BIOMARKER SENSITIVITY	The ability to detect all subjects with a particular disease or condition (ideally 100%)
LOWER LIMIT OF QUANTIFICATION	The ability to quantify replicate samples (normally $n = 5$) with a precision better than 20% and accuracy of between 80% and 120%

PHENOTYPE

Any observable characteristic or trait of an organism: such as its morphology, development, biochemical or physiological properties or behavior. Phenotypes result from the expression of an organism's genes as well as the influence of environmental factors and interactions between the two

Bibliography

Papers of special note have been highlighted as:

▪ of interest

▪▪ of considerable interest

1. Bansal S, DeStefano A. Key elements of bioanalytical method validation for small molecules. *AAPS. J* 2007;9(1):E109–E114. [PubMed: 17408234] ▪▪ Excellent overview of bioanalytical validation procedures.
2. King R, Bonfiglio R, Fernandez-Metzler C, Miller-Stein C, Olah T. Mechanistic investigation of ionization suppression in electrospray ionization. *J. Amer. Soc. Mass Spectrom* 2000;11(11):942–950. [PubMed: 11073257]
3. Oe T, Ackermann BL, Inoue K, et al. Quantitative analysis of amyloid β peptides in cerebrospinal fluid of Alzheimer's disease patients by immunoaffinity purification and stable isotope dilution LC/negative electrospray ionization tandem MS. *Rapid Commun. Mass Spectrom* 2006;20(24):3723–3735. [PubMed: 17117458]
4. Mangal D, Vudathala DK, Park JH, Lee SH, Penning TM, Blair IA. Analysis of 7,8-dihydro-8-oxo-2'-deoxyguanosine in cellular DNA during oxidative stress. *Chem. Res. Toxicol* 2009;22(5):788–797. [PubMed: 19309085]
5. Mesaros C, Lee SH, Blair IA. Targeted quantitative analysis of eicosanoid lipids in biological samples using LC-tandem MS. *J. Chromatogr. B* 2009;877(26):2736–2745.
6. Blair IA. Endogenous glutathione adducts. *Curr. Drug Metab* 2006;7(8):853–872. [PubMed: 17168687] ▪▪ Comprehensive review of glutathione adduct biomarkers.
7. Blair IA. Analysis of estrogens in serum and plasma from postmenopausal women: past, present, and future. *Steroids*. 2010 (In Press).
8. Lee SH, Blair IA. Targeted chiral lipidomics analysis of bioactive eicosanoid lipids in cellular systems. *BMB. Rep* 2009;42(7):401–410. [PubMed: 19643036]
9. Santen RJ. Assessing individual risk for breast cancer: role of oestrogens and androgens. *Breast Cancer Res* 2008;10:S10. [PubMed: 19128423] ▪▪ Excellent overview of the current use of estrogen biomarkers for assessing breast cancer risk.
10. Liehr JG. Genotoxicity of the steroidal oestrogens oestrone and oestradiol: possible mechanism of uterine and mammary cancer development. *Hum. Reprod. Update* 2001;7(3):273–281. [PubMed: 11392373]
11. Nebert DW. Elevated estrogen 16- α -hydroxylase activity: is this a genotoxic or nongenotoxic biomarker in human breast cancer risk? *J. Natl. Cancer Inst* 1993;85(23):1888–1891. [PubMed: 8230275]
12. Chan K, Morris GJ. Chemoprevention of breast cancer for women at high risk. *Semin. Oncol* 2006;33(6):642–646. [PubMed: 17145342]
13. Santen RJ, Boyd NF, Chlebowski RT, et al. Critical assessment of new risk factors for breast cancer: considerations for development of an improved risk prediction model. *Endocr. Relat Cancer* 2007;14(2):169–187. [PubMed: 17639036]
14. Winnik WM, Kitchin KT. Measurement of oxidative stress parameters using LC-tandem mass spectroscopy (LC-MS/MS). *Toxicol. Appl. Pharmacol* 2008;233(1):100–106. [PubMed: 18547599] ▪▪ Excellent overview of LC-MS methodology for the analysis of oxidative stress biomarkers.

15. Ames BN, Shigenaga MK, Hagen TM. Oxidants, antioxidants, and the degenerative diseases of aging. *Proc. Natl Acad. Sci. USA* 1993;90(17):7915–7922. [PubMed: 8367443]
16. Zhu P, Oe T, Blair IA. Determination of cellular redox status by stable isotope dilution LC/MS analysis of glutathione and glutathione disulfide. *Rapid Commun. Mass Spectrom* 2008;22(4):432–440. [PubMed: 18215009]
17. Lee SH, Rangiah K, Williams MV, Wehr AY, Dubois RN, Blair IA. Cyclooxygenase-2-mediated metabolism of arachidonic acid to 15-oxo-eicosatetraenoic acid by rat intestinal epithelial cells. *Chem. Res. Toxicol* 2007;20(11):665–1675.
18. Jian W, Lee SH, Williams MV, Blair IA. 5-lipoxygenase-mediated endogenous DNA damage. *J. Biol. Chem* 2009;284(25):16799–16807. [PubMed: 19390118]
19. Blair IA. DNA adducts with lipid peroxidation products. *J. Biol. Chem* 2008;283(23):15545–15549. [PubMed: 18285329] •• Comprehensive overview of how lipid peroxidation can cause DNA damage and the potential utility of resulting DNA adducts as disease biomarkers.
20. Lee SH, Blair IA. Targeted chiral lipidomics analysis by LC electron capture atmospheric pressure chemical ionization MS (LC-ECAPCI/MS). *Methods Enzymol* 2007;433:159–174. [PubMed: 17954234] •• Overview of the use of ECAPCI-MS for the analysis of chiral eicosanoids.
21. Singh G, Gutierrez A, Xu K, Blair IA. LC/electron capture atmospheric pressure chemical ionization/MS: analysis of pentafluorobenzyl derivatives of biomolecules and drugs in the attomole range. *Anal. Chem* 2000;72(14):3007–3013. [PubMed: 10939360]
22. Eliassen AH, Missmer SA, Tworoger SS, Hankinson SE. Circulating 2-hydroxy and 16- α -hydroxy estrone levels and risk of breast cancer among postmenopausal women. *Cancer Epidemiol. Biomarkers Prevent* 2008;17(8):2029–2035.
23. Arslan AA, Shore RE, Afanasyeva Y, Koenig KL, Toniolo P, Zeleniuch-Jacquotte A. Circulating estrogen metabolites and risk for breast cancer in premenopausal women. *Cancer Epidemiol. Biomarkers Prevent* 2009;18(8):2273–2279.
24. Kabat GC, O’Leary ES, Gammon MD, et al. Estrogen metabolism and breast cancer. *Epidemiology* 2006;17(1):80–88. [PubMed: 16357599]
25. Santen RJ, Lee JS, Wang S, et al. Potential role of ultra-sensitive estradiol assays in estimating the risk of breast cancer and fractures. *Steroids* 2008;73(13):1318–1321. [PubMed: 18614192]
26. Kushnir MM, Rockwood AL, Bergquist J, et al. High-sensitivity tandem MS assay for serum estrone and estradiol. *Am. J. Clin. Pathol* 2008;129(4):530–539. [PubMed: 18343779]
27. Ceglarek U, Kortz L, Leichtle A, Fiedler GM, Kratzsch J, Thiery J. Rapid quantification of steroid patterns in human serum by on-line solid-phase extraction combined with LC-triple quadrupole linear ion trap MS. *Clin. Chim. Acta* 2009;401(1–2):114–118. [PubMed: 19071101]
28. Hosogi J, Tanaka H, Fujita K, et al. LC-MS/MS coupled with immunoaffinity extraction for determination of estrone, 17 β -estradiol and estrone 3-sulfate in human plasma. *J. Chromatogr. B Analyt. Technol. Biomed. Life. Sci* 2009;878(2):222–227.
29. Tai SS, Welch MJ. Development and evaluation of a reference measurement procedure for the determination of estradiol-17 β in human serum using isotope-dilution LC-tandem MS. *Anal. Chem* 2005;77(19):6359–6363. [PubMed: 16194100]
30. Hsing AW, Stanczyk FZ, Belanger A, et al. Reproducibility of serum sex steroid assays in men by RIA and MS. *Cancer Epidemiol. Biomarkers Prevent* 2007;16(5):1004–1008.
31. Higashi T, Takayama N, Nishio T, Taniguchi E, Shimada K. Procedure for increasing the detection responses of estrogens in LC-MS based on introduction of a nitrobenzene moiety followed by electron capture atmospheric pressure chemical ionization. *Anal. Bioanal. Chem* 2006;386(3):658–665. [PubMed: 16609843]
32. Penning TM, Lee SH, Jin Y, Gutierrez A, Blair IA. Liquid-chromatography MS (LC–MS) of steroid hormone metabolites and its applications. *J. Steroid Biochem. Mol. Biol.* 2010 (In Press).
33. Xu X, Roman JM, Issaq HJ, Keefer LK, Veenstra TD, Ziegler RG. Quantitative measurement of endogenous estrogens and estrogen metabolites in human serum by LC-tandem MS. *Anal. Chem* 2007;79(20):7813–7821. [PubMed: 17848096]
34. Yamashita K, Okuyama M, Watanabe Y, Honma S, Kobayashi S, Numazawa M. Highly sensitive determination of estrone and estradiol in human serum by LC-electrospray ionization tandem MS. *Steroids* 2007;72(11–12):819–827. [PubMed: 17716700]

35. Xu L, Spink DC. Analysis of steroidal estrogens as pyridine-3-sulfonyl derivatives by LC electrospray tandem MS. *Anal. Biochem* 2008;375(1):105–114. [PubMed: 18162162]
36. Lin YH, Chen CY, Wang GS. Analysis of steroid estrogens in water using LC/tandem MS with chemical derivatizations. *Rapid Commun. Mass Spectrom* 2007;21(13):1973–1983. [PubMed: 17526068]
37. Nishio T, Higashi T, Funaishi A, Tanaka J, Shimada K. Development and application of electrospray-active derivatization reagents for hydroxysteroids. *J. Pharm. Biomed. Anal* 2007;44(3):786–795. [PubMed: 17420107]
38. Yang WC, Regnier FE, Sliva D, Adamec J. Stable isotope-coded quaternization for comparative quantification of estrogen metabolites by high-performance LC-electrospray ionization MS. *J. Chromatogr. B Analyt. Technol. Biomed. Life. Sci* 2008;870(2):233–240.
39. Griffiths WJ, Liu S, Alvelius G, Sjövall J. Derivatisation for the characterisation of neutral oxosteroids by electrospray and matrix-assisted laser desorption/ionisation tandem MS: the Girard P derivative. *Rapid Commun. Mass Spectrom* 2003;17(9):924–935. [PubMed: 12717765]
40. Johnson DW. Ketosteroid profiling using Girard T derivatives and electrospray ionization tandem MS: direct plasma analysis of androstenedione, 17-hydroxyprogesterone and cortisol. *Rapid Commun. Mass Spectrom* 2005;19(2):193–200. [PubMed: 15593251]
41. Fahy E, Subramaniam S, Brown HA, et al. A comprehensive classification system for lipids. *J. Lipid Res* 2005;46(5):839–861. [PubMed: 15722563] •• Nice primer on different lipid classes and the nomenclature associated with them.
42. Praticó D, Rokach J, Lawson J, FitzGerald GA. F₂-isoprostanes as indices of lipid peroxidation in inflammatory diseases. *Chem. Phys. Lipids* 2004;128(1–2):165–171. [PubMed: 15037161]
43. Milne GL, Yin H, Morrow JD. Human biochemistry of the isoprostane pathway. *J. Biol. Chem* 2008;283(23):15533–15537. [PubMed: 18285331] •• Excellent overview of how isoprostanes are formed and their utility as biomarkers of oxidative stress.
44. Fenn JB, Mann M, Meng CK, Wong SF, Whitehouse CM. Electrospray ionization for mass-spectrometry of large biomolecules. *Science* 1989;246(4926):64–71. [PubMed: 2675315]
45. Siuzdak G. The emergence of mass-spectrometry in biochemical-research. *Proc. Natl Acad. Sci. USA* 1994;91(24):11290–11297. [PubMed: 7972052]
46. Han X, Gross RW. Global analyses of cellular lipidomes directly from crude extracts of biological samples by ESI MS: a bridge to lipidomics. *J. Lipid Res* 2003;44(6):1071–1079. [PubMed: 12671038]
47. Han X, Gross RW. Shotgun lipidomics: multidimensional MS analysis of cellular lipidomes. *Expert Rev. Proteomics* 2005;2(2):253–264. [PubMed: 15892569]
48. Han X, Gross RW. Shotgun lipidomics: electrospray ionization mass spectrometric analysis and quantitation of cellular lipidomes directly from crude extracts of biological samples. *Mass Spectrom. Rev* 2005;24(3):367–412. [PubMed: 15389848] •• Excellent review on the use of MS for conducting lipidomics analysis.
49. Kempen EC, Yang P, Felix E, Madden T, Newman RA. Simultaneous quantification of arachidonic acid metabolites in cultured tumor cells using high-performance LC/electrospray ionization tandem MS. *Anal. Biochem* 2001;297(2):183–190. [PubMed: 11673886]
50. Pettinella C, Lee SH, Cipollone F, Blair IA. Targeted quantitative analysis of fatty acids in atherosclerotic plaques by high sensitivity LC/tandem MS. *J. Chromatogr. B: Analyt. Technol. Biomed. Life. Sci* 2007;850(1–2):168–176.
51. Zhang JH, Pearson T, Matharoo-Ball B, et al. Quantitative profiling of epoxyeicosatrienoic, hydroxyeicosatetraenoic, and dihydroxyeicosatetraenoic acids in human intrauterine tissues using LC/electrospray ionization tandem MS. *Anal. Biochem* 2007;365(1):40–51. [PubMed: 17418798]
52. Lee SH, Williams MV, Dubois RN, Blair IA. Targeted lipidomics using electron capture atmospheric pressure chemical ionization MS. *Rapid Commun. Mass Spectrom* 2003;17(19):2168–2176. [PubMed: 14515314]
53. Lee SH, Williams MV, Blair IA. Targeted chiral lipidomics analysis. *Prostaglandins Other Lipid Mediat* 2005;77(1–4):141–157. [PubMed: 16099399]
54. Murphy RC, Barkley RM, Zemski BK, et al. Electrospray ionization and tandem MS of eicosanoids. *Anal. Biochem* 2005;346(1):1–42. [PubMed: 15961057]

55. Meagher EA, FitzGerald GA. Indices of lipid peroxidation *in vivo*: strengths and limitations. *Free Radic. Biol. Med* 2000;28(12):1745–1750. [PubMed: 10946216]
56. Morrow JD. The isoprostanes - unique products of arachidonate peroxidation: their role as mediators of oxidant stress. *Curr. Pharm. Des* 2006;12(8):895–902. [PubMed: 16533158]
57. Montuschi P, Barnes P, Roberts LJ. Insights into oxidative stress: the isoprostanes. *Curr. Med. Chem* 2007;14(6):703–717. [PubMed: 17346157]
58. Musiek ES, Yin HY, Milne GL, Morrow JD. Recent advances in the biochemistry and clinical relevance of the isoprostane pathway. *Lipids* 2005;40(10):987–994. [PubMed: 16382569]
59. Kadiiska MB, Gladen BC, Baird DD, et al. Biomarkers of oxidative stress study II: are oxidation products of lipids, proteins, and DNA markers of CCl₄ poisoning? *Free Radic. Biol. Med* 2005;38(6):698–710. [PubMed: 15721980]
60. Selhub J. Folate, vitamin B12 and vitamin B6 and one carbon metabolism. *J. Nutr. Health Aging* 2002;6(1):39–42. [PubMed: 11813080]
61. Stover PJ. Physiology of folate and vitamin B12 in health and disease. *Nutr. Rev* 2004;62(62):S3–S12. [PubMed: 15298442]
62. Smulders YM, Stehouwer CD. Folate metabolism and cardiovascular disease. *Semin. Vasc. Med* 2005;5(2):87–97. [PubMed: 16047262]
63. Huang Y, Lu ZY, Brown KS, Whitehead AS, Blair IA. Quantification of intracellular homocysteine by stable isotope dilution LC/tandem MS. *Biomed. Chromatogr* 2007;21(1):107–112. [PubMed: 17121413]
64. Brown KS, Huang Y, Lu ZY, Jian W, Blair IA, Whitehead AS. Mild folate deficiency induces a proatherosclerotic phenotype in endothelial cells. *Atherosclerosis* 2006;189(1):133–141. [PubMed: 16469322]
65. Jensen LE, Etheredge AJ, Brown KS, Mitchell LE, Whitehead AS. Maternal genotype for the monocyte chemoattractant protein 1 A(-2518)G promoter polymorphism is associated with the risk of spina bifida in offspring. *Am. J. Med. Genet. A* 2006;140(10):1114–1118. [PubMed: 16596675]
66. Pitkin RM. Folate and neural tube defects. *Am. J. Clin. Nutr* 2007;85(1):285S–288S. [PubMed: 17209211]
67. Sanderson P, Stone E, Kim YI, et al. Folate and colo-rectal cancer risk. *Br. J. Nutr* 2007;98(6):1299–1304. [PubMed: 17764601]
68. Blount BC, Mack MM, Wehr CM, et al. Folate deficiency causes uracil misincorporation into human DNA and chromosome breakage: implications for cancer and neuronal damage. *Proc. Natl Acad. Sci. USA* 1997;94(7):3290–3295. [PubMed: 9096386]
69. Schwahn B, Rozen R. Polymorphisms in the methylenetetrahydrofolate reductase gene: clinical consequences. *Am. J. Pharmacogenomics* 2001;1(3):189–201. [PubMed: 12083967]
70. Pejchal R, Campbell E, Guenther BD, Lennon BW, Matthews RG, Ludwig ML. Structural perturbations in the Ala→Val polymorphism of methylenetetrahydrofolate reductase: how binding of folates may protect against inactivation. *Biochem* 2006;45(15):4808–4818. [PubMed: 16605249]
71. Strain JJ, Dowey L, Ward M, Pentieva K, McNulty H. B-vitamins, homocysteine metabolism and CVD. *Proc. Nutr. Soc* 2004;63(4):597–603. [PubMed: 15831132]
72. Bagley PJ, Selhub J. A common mutation in the methylenetetrahydrofolate reductase gene is associated with an accumulation of formylated tetrahydrofolates in red blood cells. *Proc. Natl Acad. Sci. USA* 1998;95(22):13217–13220. [PubMed: 9789068]
73. Botto LD, Yang Q. 5,10-methylene-tetrahydrofolate reductase gene variants and congenital anomalies: a HuGE review. *Am. J. Epidemiol* 2000;151(9):862–877. [PubMed: 10791559]
74. Huang Y, Khartulyari S, Morales ME, et al. Quantification of key red blood cell folates from subjects with defined MTHFR 677C>T genotypes using stable isotope dilution LC/MS. *Rapid Commun. Mass Spectrom* 2008;22(16):2403–2412. [PubMed: 18634122]
75. Lu ZY, Morales M, Khartulyari S, et al. Genetic and biochemical determinants of serum concentrations of monocyte chemoattractant protein-1, a potential neural tube defect risk factor. *Birth Defects Res. A Clin. Mol. Teratol* 2008;82(10):736–741. [PubMed: 18937353]
76. Mitchell LE, Morales M, Khartulyari S, et al. Folate and homocysteine phenotypes: comparative findings using research and clinical laboratory data. *Clin. Biochem* 2009;42(12):1275–1281. [PubMed: 19427846]

77. Schafer FQ, Buettner GR. Redox environment of the cell as viewed through the redox state of the glutathione disulfide/ glutathione couple. *Free Radic. Biol. Med* 2001;30(11):1191–1212. [PubMed: 11368918] •• In-depth discussion of cellular oxidative stress and how it can be monitored.
78. Arthur JR. The glutathione peroxidases. *Cell. Mol. Life Sci* 2000;57(13–14):1825–1835. [PubMed: 11215509]
79. Doss GA, Baillie TA. Addressing metabolic activation as an integral component of drug design. *Drug Metab. Rev* 2006;38(4):641–649. [PubMed: 17145693]
80. Deeley RG, Westlake C, Cole SP. Transmembrane transport of endo- and xenobiotics by mammalian ATP-binding cassette multidrug resistance proteins. *Physiol. Rev* 2006;86(3):849–899. [PubMed: 16816140]
81. Awasthi YC, Sharma R, Yadav S, Dwivedi S, Sharma A, Awasthi S. The non-ABC drug transporter RLIP76 (RALBP-1) plays a major role in the mechanisms of drug resistance. *Curr. Drug Metab* 2007;8(4):315–323. [PubMed: 17504221]
82. Lu SC. Regulation of glutathione synthesis. *Curr. Topics Cell. Regulat* 2000;36:95–116.
83. Griffith OW. Biologic and pharmacologic regulation of mammalian glutathione synthesis. *Free Radic. Biol. Med* 1999;27(9–10):922–935. [PubMed: 10569625]
84. Jones DP. Redox potential of GSH/GSSG couple: assay and biological significance. *Methods Enzymol* 2002;348:93–112. [PubMed: 11885298]
85. Murphy RC, Zarini S. Glutathione adducts of oxyeicosanoids. *Prostaglandins Lipid Mediat* 2002;68–69:471–482.
86. Dickinson DA, Forman HJ. Cellular glutathione and thiols metabolism. *Biochem. Pharm* 2002;64(5–6):1019–1026. [PubMed: 12213601]
87. Gan J, Ruan Q, He B, Zhu M, Shyu WC, Humphreys WG. *In vitro* screening of 50 highly prescribed drugs for thiol adduct formation – comparison of potential for drug-induced toxicity and extent of adduct formation. *Chem. Res. Toxicol* 2009;22(4):690–698. [PubMed: 19253935]
88. Baillie TA. Metabolic activation and drug design: challenges and opportunities in chemical toxicology. *Chem. Res. Toxicol* 2006;19(7):889–893. [PubMed: 16841955]
89. Baillie TA. Metabolism and toxicity of drugs. Two decades if progress in industrial drug metabolism. *Chem. Res. Toxicol* 2008;21(1):129–137. [PubMed: 18052111] •• Comprehensive review of drug-induced toxicity and potential biomarkers that can be employed to monitor the toxicity.
90. Rinaldi R, Eliasson E, Swedmark S, Morgenstern R. Reactive intermediates and the dynamics of glutathione transferases. *Drug Metab. Dispos* 2002;30(10):1053–1058. [PubMed: 12228179]
91. Levsen K, Schiebel HM, Behnke B, et al. Structure elucidation of phase II metabolites by tandem MS: an overview. *J. Chromatogr. A* 2005;1067(1–2):55–72. [PubMed: 15844510]
92. Dieckhaus CM, Fernandez-Metzler CL, King R, Krolikowski PH, Baillie TA. Negative ion tandem MS for the detection of glutathione conjugates. *Chem. Res. Toxicol* 2005;18(4):630–638. [PubMed: 15833023]
93. Ma L, Wen B, Ruan Q, Zhu M. Rapid screening of glutathione-trapped reactive metabolites by linear ion trap MS with isotope pattern-dependent scanning and postacquisition data mining. *Chem. Res. Toxicol* 2008;21(7):1477–1483. [PubMed: 18549250]
94. Baillie TA. Future of toxicology-metabolic activation and drug design: challenges and opportunities in chemical toxicology. *Chem. Res. Toxicol* 2006;19(7):889–893. [PubMed: 16841955]
95. Blair IA. Analysis of endogenous glutathione adducts and their metabolites. *Biomed. Chromatogr.* 2009 DOI: 10.1002/bmc 1374. (In Press).
96. Uchida K. 4-hydroxy-2-nonenal: a product and mediator of oxidative stress. *Progress Lipid Res* 2003;42(4):318–343.
97. Jian W, Lee SH, Mesaros C, Oe T, Silva Elipse MV, Blair IA. A novel 4-oxo-2(E)-nonenal-derived endogenous thiadiazabicyclo glutathione adduct formed during cellular oxidative stress. *Chem. Res. Toxicol* 2007;20(7):1008–1018. [PubMed: 17550273]
98. Carini M, Aldini G, Facino RM. MS for detection of 4-hydroxy-trans-2-nonenal (HNE) adducts with peptides and proteins. *Mass Spectrom. Rev* 2004;23(4):281–305. [PubMed: 15133838]

99. Völkel W, Varez-Sanchez R, Weick I, Mally A, Dekant W, Pahler A. Glutathione conjugates of 4-hydroxy-2(E)-nonenal as biomarkers of hepatic oxidative stress-induced lipid peroxidation in rats. *Free Radic. Biol. Med* 2005;38(11):1526–1536. [PubMed: 15890627]
100. Völkel W, Sicilia T, Pahler A, et al. Increased brain levels of 4-hydroxy-2-nonenal glutathione conjugates in severe Alzheimer's disease. *Neurochem. Int* 2006;48(8):679–686. [PubMed: 16483694]
101. Zhu P, Jian W, Blair IA. A 4-oxo-2(E)-nonenal-derived glutathione adduct from 15-lipoxygenase-1-mediated oxidation of cytosolic and esterified arachidonic acid. *Free Radic. Biol. Med* 2009;47(7):953–961. [PubMed: 19576981]
102. Jian W, Arora JS, Oe T, Shuvaev VV, Blair IA. Induction of endothelial cell apoptosis by lipid hydroperoxide-derived bifunctional electrophiles. *Free Radic. Biol. Med* 2005;39(9):1162–1176. [PubMed: 16214032]
103. Jian W, Lee SH, Arora JS, Silva Elipe MV, Blair IA. Unexpected formation of etheno-2'-deoxyguanosine adducts from 5(S)-hydroperoxyeicosatetraenoic acid: evidence for a bis-hydroperoxide intermediate. *Chem. Res. Toxicol* 2005;18(3):599–610. [PubMed: 15777099]
104. Wei C, Zhu P, Shah SJ, Blair IA. 15-oxo-eicosatetraenoic acid, a metabolite of macrophage 15-hydroxyprostaglandin dehydrogenase that inhibits endothelial cell proliferation. *Mol. Pharm* 2009;76(3):516–529.
105. Backlund MG, Mann JR, Holla VR, et al. 15-hydroxyprostaglandin dehydrogenase is downregulated in colorectal cancer. *J. Biol. Chem* 2005;280(5):3217–3223. [PubMed: 15542609]
106. Lee DH Jr, Jacobs DR. Is serum γ -glutamyltransferase a marker of exposure to various environmental pollutants? *Free Radic. Res* 2009;43(6):533–537. [PubMed: 19370474]
107. Armstrong M, Liu AH, Harbeck R, Reisdorph R, Rabinovitch N, Reisdorph N. Leukotriene-E4 in human urine: comparison of on-line purification and LC-tandem MS to affinity purification followed by enzyme immunoassay. *J. Chromatogr. B Analyt. Technol. Biomed. Live. Sci* 2009;877(27):3169–3174.
108. Hecht SS. Human urinary carcinogen metabolites: biomarkers for investigating tobacco and cancer. *Carcinogenesis* 2002;23(6):907–922. [PubMed: 12082012]
109. Tang W. The metabolism of diclofenac – enzymology and toxicology perspectives. *Curr. Drug Metab* 2003;4(4):319–329. [PubMed: 12871048]
110. Keum YS, Jeong WS, Kong AN. Chemopreventive functions of isothiocyanates. *Drug News Perspect* 2005;18(7):445–451. [PubMed: 16362084]
111. Haufroid V, Lison D. Mercapturic acids revisited as biomarkers of exposure to reactive chemicals in occupational toxicology: a minireview. *Int. Arch. Occup. Environ. Health* 2005;78(5):343–354. [PubMed: 15883816]
112. Anders MW. Chemical toxicology of reactive intermediates formed by the glutathione-dependent bioactivation of halogen-containing compounds. *Chem. Res. Toxicol* 2008;21(1):145–159. [PubMed: 17696489]
113. Sidell KR, Amamath V, Montine TJ. Dopamine thioethers in neurodegeneration. *Curr. Topics Med. Chem* 2001;1(6):519–527.
114. Pombrio JM, Giangreco A, Li L, et al. Mercapturic acids (N-acetylcysteine S-conjugates) as endogenous substrates for the renal organic anion transporter-1. *Mol. Pharm* 2001;60(5):1091–1099.
115. Stevens JF, Maier CS. Acrolein: sources, metabolism, and biomolecular interactions relevant to human health and disease. *Mol. Nutrit. Food Res* 2008;52(1):7–25. [PubMed: 18203133]
116. Farmer PB. DNA and protein adducts as markers of genotoxicity. *Toxicol. Lett* 2004;149(1–3):3–9. [PubMed: 15093242]
117. Gabelova A, Valovicova Z, Labaj J, Bacova G, Binkova B, Farmer PB. Assessment of oxidative DNA damage formation by organic complex mixtures from airborne particles PM(10). *Mutat. Res* 2007;620(1–2):135–144. [PubMed: 17403525]
118. Phillips DH, Arlt VM. Genotoxicity: damage to DNA and its consequences. *EXS* 2009;99:87–110. [PubMed: 19157059]
119. Shuker DE, Farmer PB. Relevance of urinary DNA adducts as markers of carcinogen exposure. *Chem. Res. Toxicol* 1992;5(4):450–460. [PubMed: 1391611]

120. Sharma RA, Farmer PB. Biological relevance of adduct detection to the chemoprevention of cancer. *Clin. Cancer Res* 2004;10(15):4901–4912. [PubMed: 15297390] •• Outstanding review on the potential use of DNA adducts as biomarkers of DNA damage.
121. Cooke MS, Olinski R, Loft S. Measurement and meaning of oxidatively modified DNA lesions in urine. *Cancer Epidemiol. Biomarkers Prevent* 2008;17(1):3–14. •• Excellent overview on the use of urinary DNA adducts as biomarkers
122. Lao Y, Villalta PW, Sturla SJ, Wang M, Hecht SS. Quantitation of pyridyloxobutyl DNA adducts of tobacco-specific nitrosamines in rat tissue DNA by high-performance LC-electrospray ionization-tandem MS. *Chem. Res. Toxicol* 2006;19(5):674–682. [PubMed: 16696570]
123. De Kok TM, Moonen HJ, van DJ, Van Schooten FJ. Methodologies for bulky DNA adduct analysis and biomonitoring of environmental and occupational exposures. *J. Chromatogr. B Analyt. Technol. Biomed. Life. Sci* 2002;778(1–2):345–355.
124. Doerge DR, Churchwell MI, Beland FA. Analysis of DNA adducts from chemical carcinogens and lipid peroxidation using LC and electrospray MS. *J. Environ. Sci. Health C. Environ. Carcinog. Ecotoxicol. Rev* 2002;20(1):1–20. [PubMed: 12734050]
125. Turesky RJ, Vouros P. Formation and analysis of heterocyclic aromatic amine-DNA adducts in vitro and in vivo. *J. Chromatogr. B Analyt. Technol. Biomed. Life. Sci* 2004;802(1):155–166.
126. Zhang S, Villalta PW, Wang M, Hecht SS. Detection and quantitation of acrolein-derived 1,N2-propanodeoxyguanosine adducts in human lung by LC-electrospray ionization-tandem MS. *Chem. Res. Toxicol* 2007;20(4):565–571. [PubMed: 17385896]
127. Farmer PB, Singh R. Use of DNA adducts to identify human health risk from exposure to hazardous environmental pollutants: the increasing role of MS in assessing biologically effective doses of genotoxic carcinogens. *Mutat. Res* 2008;659(1–2):68–76. [PubMed: 18468947]
128. Neale JR, Smith NB, Pierce WM, Hein DW. Methods for aromatic and heterocyclic amine carcinogen-DNA adduct analysis by LC-tandem MS. *Polycycl. Aromat. Compd* 2008;28(4–5):402–417. [PubMed: 19122802]
129. Lee SH, Williams MV, Dubois RN, Blair IA. Cyclooxygenase-2-mediated DNA damage. *J. Biol. Chem* 2005;280(31):28337–28346. [PubMed: 15964853]
130. Lee SH, Oe T, Blair IA. Vitamin C-induced decomposition of lipid hydroperoxides to endogenous genotoxins. *Science* 2001;292(5524):2083–2086. [PubMed: 11408659]
131. Williams MV, Lee SH, Blair IA. LC/MS analysis of bifunctional electrophiles and DNA adducts from vitamin C mediated decomposition of 15-hydroperoxy-eicosatetraenoic acid. *Rapid Commun. Mass Spectrom* 2005;19(6):849–858. [PubMed: 15723435]
132. Williams MV, Lee SH, Pollack M, Blair IA. Endogenous lipid hydroperoxide-mediated DNA adduct formation in Min mice. *J. Biol. Chem* 2006;281(15):10127–10133. [PubMed: 16449227]
133. Roberts DW, Churchwell MI, Beland FA, Fang JL, Doerge DR. Quantitative analysis of etheno-2'-deoxycytidine DNA adducts using on-line immunoaffinity chromatography coupled with LC/ES-MS/MS detection. *Anal. Chem* 2001;73(2):303–309. [PubMed: 11199982]
134. Singh R, McEwan M, Lamb JH, Santella RM, Farmer PB. An improved LC/tandem MS method for the determination of 8-oxo-7,8-dihydro-2'-deoxyguanosine in DNA samples using immunoaffinity column purification. *Rapid Commun. Mass Spectrom* 2003;17(2):126–134. [PubMed: 12512091]
135. Bhattacharya S, Barbacci DC, Shen M, Liu JN, Casale GP. Extraction and purification of depurinated benzo[a]pyrene adducted DNA bases from human urine by immunoaffinity chromatography coupled with HPLC and analysis by LC/quadrupole ion-trap MS. *Chem. Res. Toxicol* 2003;16(4):479–486. [PubMed: 12703964]
136. Ham AJ, Engelward BP, Koc H, et al. New immunoaffinity-LC-MS/MS methodology reveals that Aag null mice are deficient in their ability to clear 1,N6-etheno-deoxyadenosine DNA lesions from lung and liver *in vivo*. *DNA Repair (Amst)* 2004;3(3):257–265. [PubMed: 15177041]
137. Hoberg AM, Otteneider M, Marnett LJ, Poulsen HE. Measurement of the malondialdehyde-2'-deoxyguanosine adduct in human urine by immuno-extraction and LC/atmospheric pressure chemical ionization tandem MS. *J. Mass Spectrom* 2004;39(1):38–42. [PubMed: 14760611]
138. Pollack M, Yang IY, Kim HY, Blair IA, Moriya M. Translesion DNA Synthesis across the heptanone - etheno-2'-deoxycytidine adduct in cells. *Chem. Res. Toxicol* 2006;19(8):1074–1079. [PubMed: 16918247]

139. Yang IY, Hashimoto K, de WN, Blair IA, Moriya M. Two distinct translesion synthesis pathways across a lipid peroxidation-derived DNA adduct in mammalian cells. *J. Biol. Chem* 2009;284(1):191–198. [PubMed: 18981176]
140. Gackowski D, Rozalski R, Siomek A, et al. Oxidative stress and oxidative DNA damage is characteristic for mixed Alzheimer disease/vascular dementia. *J. Neurol. Sci* 2008;266(1–2):57–62. [PubMed: 17888453]
141. Nakabeppu Y, Tsuchimoto D, Yamaguchi H, Sakumi K. Oxidative damage in nucleic acids and Parkinson's disease. *J. Neurosci. Res* 2007;85(5):919–934. [PubMed: 17279544]
142. Bashir S, Harris G, Denman MA, Blake DR, Winyard PG. Oxidative DNA damage and cellular sensitivity to oxidative stress in human autoimmune diseases. *Ann. Rheum. Dis* 1993;52(9):659–666. [PubMed: 8239761]
143. Collins AR, Gedik CM, Olmedilla B, Southon S, Bellizzi M. Oxidative DNA damage measured in human lymphocytes: large differences between sexes and between countries, and correlations with heart disease mortality rates. *FASEB J* 1998;12(13):1397–1400. [PubMed: 9761783]
144. Lee SH, Blair IA. Oxidative DNA damage and cardiovascular disease. *Trends Cardiovasc. Med* 2001;11(3–4):148–155. [PubMed: 11686005]
145. Ohtsubo T, Ohya Y, Nakamura Y, et al. Accumulation of 8-oxo-deoxyguanosine in cardiovascular tissues with the development of hypertension. *DNA Repair (Amst.)* 2007;6(6):760–769. [PubMed: 17280880]
146. Olinski R, Gackowski D, Rozalski R, Foksinski M, Bialkowski K. Oxidative DNA damage in cancer patients: a cause or a consequence of the disease development? *Mutat. Res* 2003;531(1–2):177–190. [PubMed: 14637254]
147. Caporaso N. The molecular epidemiology of oxidative damage to DNA and cancer. *J. Natl Cancer Inst* 2003;95(17):1263–1265. [PubMed: 12953074]
148. Loft S, Moller P. Oxidative DNA damage and human cancer: need for cohort studies. *Antioxid. Redox Signal* 2006;8(5–6):1021–1031. [PubMed: 16771692]
149. Cooke MS, Evans MD, Dizdaroglu M, Lunec J. Oxidative DNA damage: mechanisms, mutation and disease. *FASEB J* 2003;17(10):1195–1214. [PubMed: 12832285]
150. Cooke MS, Olinski R, Evans MD. Does measurement of oxidative damage to DNA have clinical significance? *Clin. Chim. Acta* 2006;365(1–2):30–49. [PubMed: 16214123]
151. Singh R, Kaur B, Kalina I, et al. Effects of environmental air pollution on endogenous oxidative DNA damage in humans. *Mutat. Res* 2007;620(1–2):71–82. [PubMed: 17434188]
152. Loft S, Moller P, Cooke MS, Rozalski R, Olinski R. Antioxidant vitamins and cancer risk: is oxidative damage to DNA a relevant biomarker? *Eur. J. Nutr* 2008;47:19–28. [PubMed: 18458832]
153. Marnett LJ. Oxyradicals and DNA damage. *Carcinogenesis* 2000;21(3):361–370. [PubMed: 10688856]
154. Cadet J, D'Ham C, Douki T, Pouget JP, Ravanat JL, Sauvaigo S. Facts and artifacts in the measurement of oxidative base damage to DNA. *Free Radic. Res* 1998;29(6):541–550. [PubMed: 10098458]
155. Pouget JP, Douki T, Richard MJ, Cadet J. DNA damage induced in cells by γ and UVA radiation as measured by HPLC/GC-MS and HPLC-EC and Comet assay. *Chem. Res. Toxicol* 2000;13(7):541–549. [PubMed: 10898585]
156. Ravanat JL, Douki T, Duez P, et al. Cellular background level of 8-oxo-7,8-dihydro-2'-deoxyguanosine: an isotope based method to evaluate artefactual oxidation of DNA during its extraction and subsequent work-up. *Carcinogenesis* 2002;23(11):1911–1918. [PubMed: 12419840]
157. European Standards Committee on Oxidative DNA damage. Comparative analysis of baseline 8-oxo-7,8-dihydroguanine in mammalian cell DNA, by different methods in different laboratories: an approach to consensus. *Carcinogenesis* 2002;23(12):2129–2133. [PubMed: 12507938]
158. Lunec J. ESCODD: European Standards Committee on Oxidative DNA Damage. *Free Radic. Res* 1998;29(6):601–608. [PubMed: 10098465]
159. Riis B. Comparison of results from different laboratories in measuring 8-oxo-2'-deoxyguanosine in synthetic oligonucleotides. *Free Radic. Res* 2002;36(6):649–659. [PubMed: 12180190]

160. ESCODD (European Standards Committee on Oxidative DNA Damage). Inter-laboratory validation of procedures for measuring 8-oxo-7,8-dihydroguanine/8-oxo-7,8-dihydro-2'-deoxyguanosine in DNA. *Free Radic. Res* 2002;36(3):239–245. [PubMed: 12071341]
161. European Standards Committee on Oxidative DNA damage (ESCODD). Measurement of DNA oxidation in human cells by chromatographic and enzymic methods. *Free Radic. Biol. Med* 2003;34(8):1089–1099. [PubMed: 12684094]
162. Collins AR, Cadet J, Moller L, Vina J. Are we sure we know how to measure 8-oxo-7,8-dihydroguanine in DNA from human cells? *Arch. Biochem. Biophys* 2004;423(1):57–65. [PubMed: 14989265]
163. Gedik CM, Collins A. Establishing the background level of base oxidation in human lymphocyte DNA: results of an interlaboratory validation study. *FASEB J* 2005;19(1):82–84. [PubMed: 15533950]
164. Smith CC, O'Donovan MR, Martin EA. hOGG1 recognizes oxidative damage using the comet assay with greater specificity than FPG or ENDOIII. *Mutagenesis* 2006;21(3):185–190. [PubMed: 16597659]
165. Blair, IA. MS approaches to elucidate the role of oxidative stress in cancer and toxicology. In: Caprioli, RM.; Gross, ML., editors. *Encyclopedia of MS*. Elsevier, UK: 2005. p. 283–307.
166. Chomczynski P, Mackey K, Drews R, Wilfinger W. DNAzol: a reagent for the rapid isolation of genomic DNA. *Biotechniques* 1997;22(3):550–553. [PubMed: 9067036]
167. Helbock HJ, Beckman KB, Shigenaga MK, et al. DNA oxidation matters: the HPLC-electrochemical detection assay of 8-oxo-deoxyguanosine and 8-oxo-guanine. *Proc. Natl Acad. Sci. USA* 1998;95(1):288–293. [PubMed: 9419368]
168. Hamilton ML, Guo Z, Fuller CD, et al. A reliable assessment of 8-oxo-2-deoxyguanosine levels in nuclear and mitochondrial DNA using the sodium iodide method to isolate DNA. *Nucleic Acids Res* 2001;29(10):2117–2126. [PubMed: 11353081]
169. Sai K, Takagi A, Umemura T, Hasegawa R, Kurokawa Y. Changes of 8-hydroxy-deoxyguanosine levels in rat organ DNA during the aging process. *J. Environ. Pathol. Toxicol. Oncol* 1992;11(3):139–143. [PubMed: 1625183]
170. Smithgall TE, Harvey RG, Penning TM. Spectroscopic identification of ortho-quinones as the products of polycyclic aromatic *trans*-dihydrodiol oxidation catalyzed by dihydrodiol dehydrogenase. A potential route of proximate carcinogen metabolism. *J. Biol. Chem* 1988;263(4):1814–1820. [PubMed: 3276678]
171. Smith CJ, Perfetti TA, Rumble MA, Rodgman A, Doolittle DJ. "IARC Group 2B carcinogens" reported in cigarette mainstream smoke. *Food Chem. Toxicol* 2001;39(2):183–205. [PubMed: 11267712]
172. Baird WM, Hooven LA, Mahadevan B. Carcinogenic polycyclic aromatic hydrocarbon-DNA adducts and mechanism of action. *Environ. Mol. Mutagen* 2005;45(2–3):106–114. [PubMed: 15688365]
173. Ruan Q, Kim HY, Jiang H, Penning TM, Harvey RG, Blair IA. quantification of benzo[a]pyrene diol epoxide DNA adducts by stable isotope dilution LC/tandem MS. *Rapid Commun. Mass Spectrom* 2006;20(8):1369–1380. [PubMed: 16557497]
174. Ruan Q, Gelhaus SL, Penning TM, Harvey RG, Blair IA. Aldo-keto reductase-and cytochrome P450-dependent formation of benzo[a]pyrene-derived DNA adducts in human bronchoalveolar cells. *Chem. Res. Toxicol* 2007;20(3):424–431. [PubMed: 17295519]
175. Berna M, Ott L, Engle S, Watson D, Solter P, Ackermann B. Quantification of NTproBNP in rat serum using immunoprecipitation and LC/MS/MS: a biomarker of drug-induced cardiac hypertrophy. *Anal. Chem* 2008;80(3):561–566. [PubMed: 18179251]
176. Carr SA, Anderson L. Protein quantitation through targeted MS: the way out of biomarker purgatory? *Clin. Chem* 2008;54(11):1749–1752. [PubMed: 18957555]
177. Hoofnagle AN, Becker JO, Wener MH, Heinecke JW. Quantification of thyroglobulin, a low-abundance serum protein, by immunoaffinity peptide enrichment and tandem MS. *Clin. Chem* 2008;54(11):1796–1804. [PubMed: 18801935]

178. Nicol GR, Han M, Kim J, et al. Use of an immunoaffinity-MS-based approach for the quantification of protein biomarkers from serum samples of lung cancer patients. *Mol. Cell. Proteomics* 2008;7(10):1974–1982. [PubMed: 18388126]
179. Kuhn E, Addona T, Keshishian H, et al. Developing multiplexed assays for troponin I and interleukin-33 in plasma by peptide immunoaffinity enrichment and targeted MS. *Clin. Chem* 2009;55(6):1108–1117. [PubMed: 19372185]
180. Keshishian H, Addona T, Burgess M, et al. Quantification of cardiovascular biomarkers in patient plasma by targeted MS and stable isotope dilution. *Mol. Cell. Proteomics* 2009;8(10):2339–2349. [PubMed: 19596694] •• Excellent report showing the current state-of-the-art and highlighting the challenges associated with quantifying protein biomarkers in a complex matrix using a bottom-up approach.
181. Anderson L, Hunter CL. Quantitative mass spectrometric multiple reaction monitoring assays for major plasma proteins. *Mol. Cell Proteomics* 2006;5(4):573–588. [PubMed: 16332733]
182. Lee JW, Hall M. Method validation of protein biomarkers in support of drug development or clinical diagnosis/prognosis. *J. Chromatogr. B Analyt. Technol. Biomed. Life Sci* 2009;877(13):1259–1271. •• Comprehensive review detailing analytical method optimization and validation for protein quantification SID MS assays.
183. Viswanathan CT, Bansal S, Booth B, et al. Quantitative bioanalytical methods validation and implementation: best practices for chromatographic and ligand binding assays. *Pharm. Res* 2007;24(10):1962–1973. [PubMed: 17458684]
184. Barnidge DR, Goodmanson MK, Klee GG, Muddiman DC. Absolute quantification of the model biomarker prostate-specific antigen in serum by LC-MS/MS using protein cleavage and isotope dilution MS. *J. Proteome Res* 2004;3(3):644–652. [PubMed: 15253448]
185. Kuhn E, Wu J, Karl J, Liao H, Zolg W, Guild B. Quantification of C-reactive protein in the serum of patients with rheumatoid arthritis using multiple reaction monitoring MS and ¹³C-labeled peptide standards. *Proteomics* 2004;4(4):1175–1186. [PubMed: 15048997]
186. Rifai N, Gillette MA, Carr SA. Protein biomarker discovery and validation: the long and uncertain path to clinical utility. *Nat. Biotechnol* 2006;24(8):971–983. [PubMed: 16900146] •• Offers the reader a realistic perspective on the difficult steps necessary to move an assay from discovery to validation and, ultimately, into a clinical setting.
187. Kulasingam V, Diamandis EP. Strategies for discovering novel cancer biomarkers through utilization of emerging technologies. *Nat. Clin. Pract. Oncol* 2008;5(10):588–599. [PubMed: 18695711]
188. Olsen JV, Ong SE, Mann M. Trypsin cleaves exclusively C-terminal to arginine and lysine residues. *Mol. Cell. Proteomics* 2004;3(6):608–614. [PubMed: 15034119]
189. Coon JJ. Collisions or electrons? Protein sequence analysis in the 21st century. *Anal. Chem* 2009;81(9):3208–3215. [PubMed: 19364119]
190. Aebersold R, Mann M. MS-based proteomics. *Nature* 2003;422(6928):198–207. [PubMed: 12634793] •• Excellent basic primer in proteomics methodology.
191. Anderson NL, Anderson NG, Pearson TW, et al. A human proteome detection and quantitation project. *Mol. Cell Proteomics* 2009;8(5):883–886. [PubMed: 19131327]
192. Mallick P, Schirle M, Chen SS, et al. Computational prediction of proteotypic peptides for quantitative proteomics. *Nat. Biotechnol* 2007;25(1):125–131. [PubMed: 17195840]
193. Barr JR, Maggio VL, Patterson DG Jr, et al. Isotope dilution–mass spectrometric quantification of specific proteins: model application with apolipoprotein A–I. *Clin. Chem* 1996;42(10):1676–1682. [PubMed: 8855153]
194. Gygi SP, Rist B, Gerber SA, Turecek F, Gelb MH, Aebersold R. Quantitative analysis of complex protein mixtures using isotope-coded affinity tags. *Nat. Biotechnol* 1999;17(10):994–999. [PubMed: 10504701]
195. Oda Y, Huang K, Cross FR, Cowburn D, Chait BT. Accurate quantitation of protein expression and site-specific phosphorylation. *Proc. Natl Acad. Sci. USA* 1999;96(12):6591–6596. [PubMed: 10359756]

196. Ross PL, Huang YN, Marchese JN, et al. Multiplexed protein quantitation in *Saccharomyces cerevisiae* using amine-reactive isobaric tagging reagents. *Mol. Cell Proteomics* 2004;3(12):1154–1169. [PubMed: 15385600]
197. Thompson A, Schafer J, Kuhn K, et al. Tandem mass tags: a novel quantification strategy for comparative analysis of complex protein mixtures by MS/MS. *Anal. Chem* 2003;75(8):1895–1904. [PubMed: 12713048]
198. Nita-Lazar A, Saito-Benz H, White FM. Quantitative phosphoproteomics by MS: past, present, and future. *Proteomics* 2008;8(21):4433–4443. [PubMed: 18846511]
199. Bantscheff M, Schirle M, Sweetman G, Rick J, Kuster B. Quantitative MS in proteomics: a critical review. *Anal. Bioanal. Chem* 2007;389(4):1017–1031. [PubMed: 17668192]
200. Zhang Y, Wolf-Yadlin A, Ross PL, et al. Time-resolved MS of tyrosine phosphorylation sites in the epidermal growth factor receptor signaling network reveals dynamic modules. *Mol. Cell. Proteomics* 2005;4(9):1240–1250. [PubMed: 15951569]
201. Huang PH, Mukasa A, Bonavia R, et al. Quantitative analysis of EGFRvIII cellular signaling networks reveals a combinatorial therapeutic strategy for glioblastoma. *Proc. Natl Acad. Sci. USA* 2007;104(31):12867–12872. [PubMed: 17646646]
202. Schmelzle K, Kane S, Gridley S, Lienhard GE, White FM. Temporal dynamics of tyrosine phosphorylation in insulin signaling. *Diabetes* 2006;55(8):2171–2179. [PubMed: 16873679]
203. Kim JE, White FM. Quantitative analysis of phosphotyrosine signaling networks triggered by CD3 and CD28 costimulation in Jurkat cells. *J. Immunol* 2006;176(5):2833–2843. [PubMed: 16493040]
204. Wolf-Yadlin A, Kumar N, Zhang Y, et al. Effects of HER2 overexpression on cell signaling networks governing proliferation and migration. *Mol. Syst. Biol* 2006;2:54. [PubMed: 17016520]
205. Choe L, D'Ascenzo M, Relkin NR, et al. 8-plex quantitation of changes in cerebrospinal fluid protein expression in subjects undergoing intravenous immunoglobulin treatment for Alzheimer's disease. *Proteomics* 2007;7(20):3651–3660. [PubMed: 17880003]
206. Rajcevic U, Petersen K, Knol JC, et al. iTRAQ based proteomic profiling reveals increased metabolic activity and cellular crosstalk in angiogenic compared to invasive Glioblastoma phenotype. *Mol. Cell. Proteomics*. 2009 DOI: 10.1074/mcp.M900124-MCP200 (Epub ahead of print).
207. Ow SY, Salim M, Noirel J, Evans C, Rehman I, Wright PC. iTRAQ underestimation in simple and complex mixtures: the good, the bad and the ugly. *J. Proteome Res*. 2009 DOI: 10.1021/pr900634c. (Epub ahead of print).
208. Ong SE, Blagoev B, Kratchmarova I, et al. Stable isotope labeling by amino acids in cell culture, SILAC, as a simple and accurate approach to expression proteomics. *Mol. Cell Proteomics* 2002;1(5):376–386. [PubMed: 12118079]
209. Ong SE, Mann M. Stable isotope labeling by amino acids in cell culture for quantitative proteomics. *Methods Mol. Biol* 2007;359:37–52. [PubMed: 17484109]
210. Graumann J, Hubner NC, Kim JB, et al. Stable isotope labeling by amino acids in cell culture (SILAC) and proteome quantitation of mouse embryonic stem cells to a depth of 5,111 proteins. *Mol. Cell. Proteomics* 2008;7(4):672–683. [PubMed: 18045802]
211. Venable JD, Wohlschlegel J, McClatchy DB, Park SK, Yates JR III. Relative quantification of stable isotope labeled peptides using a linear ion trap-Orbitrap hybrid mass spectrometer. *Anal. Chem* 2007;79(8):3056–3064. [PubMed: 17367114]
212. Gronborg M, Kristiansen TZ, Iwahori A, et al. Biomarker discovery from pancreatic cancer secretome using a differential proteomic approach. *Mol. Cell Proteomics* 2006;5(1):157–171. [PubMed: 16215274]
213. Yocum AK, Busch CM, Felix CA, Blair IA. Proteomics-based strategy to identify biomarkers and pharmacological targets in leukemias with t(4;11) translocations. *J. Proteome Res* 2006;5(10):2743–2753. [PubMed: 17022645]
214. Luo W, Slebos RJ, Hill S, et al. Global impact of oncogenic Src on a phosphotyrosine proteome. *J. Proteome Res* 2008;7(8):3447–3460. [PubMed: 18563927]
215. Guo A, Villen J, Kornhauser J, et al. Signaling networks assembled by oncogenic EGFR and c-Met. *Proc. Natl Acad. Sci. USA* 2008;105(2):692–697. [PubMed: 18180459]

216. Yu KH, Barry CG, Austin D, et al. Stable isotope dilution multidimensional LC-tandem MS for pancreatic cancer serum biomarker discovery. *J. Proteome Res* 2009;8(3):1565–1576. [PubMed: 19199705]
217. Pan C, Olsen JV, Daub H, Mann M. Global effects of kinase inhibitors on signaling networks revealed by quantitative phosphoproteomics. *Mol. Cell. Proteomics*. 2009 DOI: 10.1074/mcp.M900285-MCP200. (Epub ahead of print).
218. Cox J, Mann M. MaxQuant enables high peptide identification rates, individualized p.p.b.-range mass accuracies and proteome-wide protein quantification. *Nat. Biotechnol* 2008;26(12):1367–1372. [PubMed: 19029910]
219. Rangiah K, Tippornwong M, Sangar V, et al. Differential secreted proteome approach in murine model for candidate biomarker discovery in colon cancer. *J. Proteome Res*. 2009 DOI: 10.1021/pr900518v. (Epub ahead of print).
220. Shah SJ, Yu KH, Sangar V, Parry SI, Blair IA. Identification and quantification of preterm birth biomarkers in human cervicovaginal fluid by LC/tandem MS. *J. Proteome Res* 2009;8(5):2407–2417. [PubMed: 19271751]
221. Gerber SA, Rush J, Stemman O, Kirschner MW, Gygi SP. Absolute quantification of proteins and phosphoproteins from cell lysates by tandem MS. *Proc. Natl Acad. Sci. USA* 2003;100(12):6940–6945. [PubMed: 12771378]
222. Krueger KE, Srivastava S. Posttranslational protein modifications: current implications for cancer detection, prevention, and therapeutics. *Mol. Cell Proteomics* 2006;5(10):1799–1810. [PubMed: 16844681]
223. Mayya V, Rezual K, Wu L, Fong MB, Han DK. Absolute quantification of multisite phosphorylation by selective reaction monitoring MS: determination of inhibitory phosphorylation status of cyclin-dependent kinases. *Mol. Cell Proteomics* 2006;5(6):1146–1157. [PubMed: 16546994]
224. Singh S, Springer M, Steen J, Kirschner MW, Steen H. FLEXIQuant: a novel tool for the absolute quantification of proteins, and the simultaneous identification and quantification of potentially modified peptides. *J. Proteome Res* 2009;8(5):2201–2210. [PubMed: 19344176]
225. Ciccimaro E, Hanks SK, Blair IA. Quantification of focal adhesion kinase activation loop phosphorylation as a biomarker of Src activity. *Mol. Pharmacol* 2009;75(3):658–666. [PubMed: 19098120]
226. Kraut A, Marcellin M, Adrait A, et al. Peptide storage: are you getting the best return on your investment? Defining optimal storage conditions for proteomics samples. *J. Proteome Res* 2009;8(7):3778–3785. [PubMed: 19397304]
227. Arsene CG, Ohlendorf R, Burkitt W, et al. Protein quantification by isotope dilution MS of proteolytic fragments: cleavage rate and accuracy. *Anal. Chem* 2008;80(11):4154–4160. [PubMed: 18447320]
228. Addona TA, Abbatiello SE, Schilling B, et al. Multi-site assessment of the precision and reproducibility of multiple reaction monitoring-based measurements of proteins in plasma. *Nat. Biotechnol* 2009;27(7):633–641. [PubMed: 19561596]
229. Mirzaei H, McBee JK, Watts J, Aebersold R. Comparative evaluation of current peptide production platforms used in absolute quantification in proteomics. *Mol. Cell Proteomics* 2008;7(4):813–823. [PubMed: 18089551]
230. Immler D, Greven S, Reinemer P. Targeted proteomics in biomarker validation: detection and quantification of proteins using a multi-dimensional peptide separation strategy. *Proteomics* 2006;6(10):2947–2958. [PubMed: 16619308]
231. Keshishian H, Addona T, Burgess M, Kuhn E, Carr SA. Quantitative, multiplexed assays for low abundance proteins in plasma by targeted MS and stable isotope dilution. *Mol. Cell Proteomics* 2007;6(12):2212–2229. [PubMed: 17939991]
232. Anderson NL, Anderson NG. The human plasma proteome: history, character, and diagnostic prospects. *Mol. Cell Proteomics* 2002;1(11):845–867. [PubMed: 12488461]
233. Pan S, Aebersold R, Chen R, Rush J, Goodlett DR, McIntosh MW, et al. MS based targeted protein quantification: methods and applications. *J. Proteome Res* 2009;8(2):787–797. [PubMed: 19105742]

234. Bondar OP, Barnidge DR, Klee EW, Davis BJ, Klee GG. LC-MS/MS quantification of Zn- α 2 glycoprotein: a potential serum biomarker for prostate cancer. *Clin. Chem* 2007;53(4):673–678. [PubMed: 17317883]
235. Wolters DA, Washburn MP, Yates JR III. An automated multidimensional protein identification technology for shotgun proteomics. *Anal. Chem* 2001;73(23):5683–5690. [PubMed: 11774908]
236. Slebos RJ, Brock JW, Winters NF, et al. Evaluation of strong cation exchange versus isoelectric focusing of peptides for multidimensional LC-tandem MS. *J. Proteome Res* 2008;7(12):5286–5294. [PubMed: 18939861]
237. Fortin T, Salvador A, Charrier JP, Lenz C, Lacoux X, Morla A, et al. Clinical quantitation of prostate-specific antigen biomarker in the low nanogram/milliliter range by conventional bore LC-tandem MS (multiple reaction monitoring) coupling and correlation with ELISA tests. *Mol. Cell Proteomics* 2009;8(5):1006–1015. [PubMed: 19068476]
238. Anderson NL, Anderson NG, Haines LR, Hardie DB, Olafson RW, Pearson TW. Mass spectrometric quantitation of peptides and proteins using stable isotope standards and capture by anti-peptide antibodies (SISCAPA). *J. Proteome Res* 2004;3(2):235–244. [PubMed: 15113099]
239. Ahn YH, Lee JY, Lee JY, Kim YS, Ko JH, Yoo JS. Quantitative analysis of an aberrant glycoform of TIMP1 from colon cancer serum by I-PHA-enrichment and SISCAPA with MRM MS. *J. Proteome Res* 2009;8(9):4216–4224. [PubMed: 19645485]
240. Ackermann BL, Berna MJ. Coupling immunoaffinity techniques with MS for quantitative analysis of low-abundance protein biomarkers. *Expert. Rev. Proteomics* 2007;4(2):175–186. [PubMed: 17425454]
241. Mayr BM, Kohlbacher O, Reinert K, et al. Absolute myoglobin quantitation in serum by combining two-dimensional LC-electrospray ionization MS and novel data analysis algorithms. *J. Proteome. Res* 2006;5(2):414–421. [PubMed: 16457608]
242. Kigawa T, Muto Y, Yokoyama S. Cell-free synthesis and amino acid-selective stable isotope labeling of proteins for NMR analysis. *J. Biomol. NMR* 1995;6(2):129–134. [PubMed: 8589601]
243. Muchmore DC, McIntosh LP, Russell CB, Anderson DE, Dahlquist FW. Expression and nitrogen-15 labeling of proteins for proton and nitrogen-15 nuclear magnetic resonance. *Methods Enzymol* 1989;177:44–73. [PubMed: 2691846]
244. Brun V, Dupuis A, Adrait A, et al. Isotope-labeled protein standards: toward absolute quantitative proteomics. *Mol. Cell Proteomics* 2007;6(12):2139–2149. [PubMed: 17848587]
245. Heudi O, Barteau S, Zimmer D, et al. Towards absolute quantification of therapeutic monoclonal antibody in serum by LC-MS/MS using isotope-labeled antibody standard and protein cleavage isotope dilution MS. *Anal. Chem* 2008;80(11):4200–4207. [PubMed: 18465883]
246. Janecki DJ, Bemis KG, Tegeler TJ, Sanghani PC, Zhai L, Hurley TD, et al. A multiple reaction monitoring method for absolute quantification of the human liver alcohol dehydrogenase ADH1C1 isoenzyme. *Anal. Biochem* 2007;369(1):18–26. [PubMed: 17692277]
247. Ciccimaro E, Hanks SK, Yu KH, Blair IA. Absolute quantification of phosphorylation on the kinase activation loop of cellular focal adhesion kinase by stable isotope dilution LC/MS. *Anal. Chem* 2009;81(9):3304–3313. [PubMed: 19354260]
248. Paz-Elizur T, Brenner DE, Livneh Z. Interrogating DNA repair in cancer risk assessment. *Cancer Epidemiol. Biomarkers Prevent* 2005;14(7):1585–1587.
249. Hong H, Wang Y. Derivatization with Girard reagent T combined with LC-MS/MS for the sensitive detection of 5-formyl-2'-deoxyuridine in cellular DNA. *Anal. Chem* 2007;79(1):322–326. [PubMed: 17194156]

Figure 1. Formation of small molecule and protein biomarkers

BER: Base excision repair; GSH: Glutathione; NER: Nucleotide excision repair; ROS: Reactive oxygen species.

Figure 2. Metabolism of estrogens in menopausal women

AKR: Adlo-keto reductase; COMT: Catechol-O-methyltransferase; CYP: Cytochrome.

Figure 3. Derivatives for improving electrospray ionization characteristics of estrogens

Figure 4. Formation of oxidized lipids

5-HEDH: 5-hydroxyeicosanoid dehydrogenase; EH: Epoxide hydrolase; FLAP: 5-lipoxygenase-activating protein; FOG-7: 5-oxoeicosatetraenoic acid GSH adduct; LO: Lipoxygenase; LTAH: Leukotriene A₄ hydrolase; LTAS: Leukotriene A₄ synthase; PGS: Prostaglandin synthase; ROS: Reactive oxygen species; TX: Thromboxane; TXS: Thromboxane synthase.

Figure 5. Stable isotope dilution chiral LC–multiple reaction monitoring-MS analysis of HETEs
Reproduced with permission from [20].

Figure 6. Formation and metabolism of folate derivatives

AIICARTF: Aminoimidazolecarboxamide ribotide transformylase; B6: Vitamin B6 [pyridoxine, 3-hydroxy-4,5-bis(hydroxymethyl)-2-methylpyridine]; B12: Vitamin B12 (cyanocobalamin); CysRS: Cysteine-t-RNA synthase; DHFR: Dihydrofolate reductase; FTHFL: Formate-tetrahydrofolate ligase; GR: Glutathione reductase; GS: glutathione synthase; γ GCL: γ -glutamylcysteine ligase; GCT: γ -cystathionase; HR: Homocysteine reductase; IMPCH: Inosine monophosphate cyclohydrolase; MAT: Methionine adenosylmethyltransferase; MetRS: Methionine-t-RNA synthase; MT: Methyl transferase; MTHFC: 5,10-methenylcyclohydrolase; MTHFD: 5,10-methylene tetrahydrofolate reductase; MTHFS: 5,10-methenyltetrahydrofolate synthase; MTR: 5-methyltetrahydrofolate-homocysteine S-methyltransferase; MTRR: Methionine synthase reductase.

Figure 7. Redox cycling of protein sulfhydryl (ProSH) and disulfides (ProS-SPro) with glutathione and glutathione disulfide

GSH: Glutathione; GS-SG: Glutathione disulfide; Pro: Protein; SG: Glutathionyl.

Figure 8. Parallel standard curves for GSH derivative in macrophage cell line lysates
ABD-F: 4-aminosulfonyl-7-fluoro-2,1,3-benzoxdiazole GSH: Glutathione. Adapted from [16].

Figure 9. Formation of thiadiazabicyclo-ONE-GSH adduct from a lipid hydroperoxide (15 (S) hydroperoxyeicosatetraenoic acid)

Figure 10. Formation of 15-oxo-eicosatetraenoic acid-glutathione adduct in macrophage/monocytes

GST: GSH-S-transferase; OEC: Oxo-EETE-derived cysteinylglycine adducts; OEG: Oxo-EETE-derived GSH adduct.

Figure 11. Metabolism of acetaminophen to a mercapturic acid derivative

Figure 12. Formation of DNA adducts from 15 (*S*)-HPETE and other lipid hydroperoxides
 HPETE: Hydroperoxyeicosatetraenoic acid.

Figure 13. Analysis of 8-oxo-dGuo in human lung cells under different conditions
Reprinted with permission from [4].

>gil4501885|reflnp_001092.1|β actin [*Homo sapiens*]

MDDDI AALVVDNGSGMCKAGFAGDDAPR **AVFPSIVGR** PRHQGVMVGMGQKDS
 YVGDEAQSKRGILTLKYP I EHGIVTNWDDMEKIWHHTFYNELRVAP E EHPVLLTEAPLN
 PKANREKMTQIMFETFNTPAMYVAIQAVLSLYASGR TTGIVMDSGDGVTH TVPIYEGYA
 LPHAILRLDLAGRDLTDYLMKILTERGYSFTTTAEREIVRDIKEKLCYVALDFEQEMATAA
 SSSSLEKSYELPDGQVITIGNERFRCPEALFQPSFLGMESCGIHETTFNSIMKCDVDIRK
 DLYANTVLSGGTTMYPGIADRMQKEITALAPSTMKIKI IAPPERKYSVWIGGSILASLSTF
 QQMWISKQEYDESGPSIVHRKCF

Figure 14. Protein quantification using the 'bottom-up' approach

In this example, significant trypsin-mediated cleavage occurred at RP.

Figure 15. Protocol for analysis of serum pancreatic cancer biomarkers

SID: Stable-isotope dilution; SILAC: Stable isotope labeling by amino acids in cell culture;

SILAP: Stable isotope-labeled proteome.

Adapted with permission from [216].

Table 1

Limits of quantitation for analysis of serum and plasma estradiol and estrone by LC-MS/MS after derivatization to improve ionization efficiency.

Ionization	Derivative	LOQ (pg/ml)		LOQ (p mol/l)		Ref.
		<i>Estradiol</i>	<i>Estrone</i>	<i>Estradiol</i>	<i>Estrone</i>	
ECAPCI	PFB	5.00	5.0	18.4	18.5	[32]
ESI	D	8.00	8.0	29.4	29.6	[33]
ESI	P	0.50	1.0	1.8	3.6	[34]
ESI	PS	10.00	ND	36.7	ND	[35]
ESI	NMN	0.44	0.36	1.6	1.3	[38]

D: Dansyl; ESI: Electrospray ionization; ECAPCI: Electron capture atmospheric pressure chemical ionization; ND: Not detected; NMN: *N*-methyl-nicotinyl; P: Picolinoyl; PFB: Pentafluorobenzyl; PS: Pyridyl-3-sulfonyl.