Published in final edited form as:

Semin Roentgenol. 2010 October ; 45(4): 238–249. doi:10.1053/j.ro.2009.12.007.

MRI of the Knee: Optimizing 3T Imaging

Lauren Shapiro, B.A.¹, Ernesto Staroswiecki, M.S.^{1,4}, and Garry Gold, M.D.^{1,2,3}

¹Department of Radiology, Stanford University

²Department of Bioengineering, Stanford University

³Department of Orthopaedic Surgery, Stanford University

⁴Department of Electrical Engineering, Stanford University

Introduction

Magnetic resonance imaging (MRI), with its multiplanar capabilities and excellent soft-tissue contrast, has established itself as the leading modality for noninvasive evaluation of the musculoskeletal system (1-5). It is regarded as the top imaging and diagnostic tool for the knee joint as a result of its ability to evaluate a wide range of anatomy and pathology varying from ligamentous injuries to articular cartilage lesions. Imaging of the knee requires excellent contrast, high resolution and the ability to visualize very small structures, all of which can be provided by MR imaging. The development of advanced diagnostic MR imaging tools for the joints is of increased clinical importance as it has been recently shown that musculoskeletal imaging is the most rapidly growing field in MR imaging, second only to neuroradiology applications (6).

Currently, most clinical evaluation of the musculoskeletal system is performed at intermediate field strengths of 1.5 T or lower. High field systems, like 3.0 T, are now becoming increasingly available for clinical use. Although at first used primarily for neurological imaging, an increasing number of studies have demonstrated the abilities and advantages of 3.0 T systems in musculoskeletal imaging (7-10). The most notable advantage includes an increased signal-to-noise ratio (SNR) which can lead to a shorter imaging time or improved image resolution. However, with the increase to a 3.0 T field strength comes a various number of considerations that must be dealt with in order to optimize its intrinsically superior imaging capabilities.

Advantages of using 3.0 T

3.0 T imaging is of special interest to the musculoskeletal system due to the increased MR signal and higher SNR. SNR is a function of the main magnetic field strength, the volume of tissue being imaged and the radiofrequency coil used. Therefore if the tissue volume imaged and coil used remain the same, the transition from 1.5 T to 3.0 T should result in twice the intrinsic SNR. This increase in SNR then allows for up to four times faster image acquisition on multiple-average scans or double the resolution in one direction. Positive clinical applications abound. The increase in scan speed has the ability to provide for increased patient

© 2010 Elsevier Inc. All rights reserved.

Corresponding author: Garry Gold, MD, Department of Radiology, Grant Building Room SO68B, Stanford, CA 94305, gold@stanford.edu.

Publisher's Disclaimer: This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our customers we are providing this early version of the manuscript. The manuscript will undergo copyediting, typesetting, and review of the resulting proof before it is published in its final citable form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

comfort and throughput while the increase in resolution may prove invaluable for the visualization of small structures. The increase in resolution and visualization of anatomy and pathology also provide the advantage of improved preoperative planning.

Shortly after the introduction of 3.0 T imaging capabilities several researchers began studies in an attempt to investigate the anatomical and pathological accuracy of the new high field imaging systems compared to 1.5 T and lower field imaging systems. Numerous studies have demonstrated the high accuracy and sensitivity and specificity of tissue anatomy and pathology in the knee joint. The ligamentous structures of the knee joint have been shown to be better visualized at 3.0 T when compared with lower field imaging (11-14). 3.0 T imaging also offers the possibility of delineation of fine detail which was not previously offered at lower field imaging strengths (14) (Fig. 1). Meniscal anatomy acquired at 3.0 T has been demonstrated to be displayed with enhanced visibility while meniscal pathology obtained at 3.0 T has been shown to allow for better clinical assessment as demonstrated by the superior sensitivity and specificity measures when compared with lower field imaging systems (11,13-16) (Fig. 2). Bone marrow edema has also been demonstrated to be seen with greater resolution and detail providing for increasing diagnostic accuracy of knee joint pathology (8,17) (Fig. 3).

SNR and Relaxation Time Considerations

While it appears that 3.0 T imaging should provide double the intrinsic SNR of imaging at 1.5 T, changes in both T1 and T2 relaxation times along with the lack of optimized coils results in an SNR improvement of slightly less than double. Much work has been done to measure relaxation times in order to optimize imaging protocols for the musculoskeletal system at 3.0 T. (7) These studies have concluded that T1 relaxation times must be increased 14 - 20% at 3.0 T from 1.5 T. It was also noted that this lengthening of T1 time is much less pronounced in cartilage while it is much more pronounced in fluid and fatty bone marrow. In contrast, the T2 relaxation time has been demonstrated to be much less dependent upon magnetic field strength requiring approximately only a 10% decrease from 1.5 T to 3.0 T. These changes in T1 and T2 relaxation times affect the appropriate TR and TE for 3.0 T and ultimately affect the SNR and contrast of the images acquired (Fig. 4).

Tissue contrast in MRI is determined by several different variables, including the chosen TR and TE, the T1 and T2 relaxation times of the tissues and the use of fat saturation. Manipulation of TR and TE should reflect the tissues being imaged as well as the contrast desired. As a result of the increase in T1 relaxation times at 3.0 T, the TR must be increased in order to achieve the tissue contrast seen at 1.5 T. Along the same lines, the TE should be decreased slightly in order to account of the T2 relaxation time decrease with 3.0 T.

Technical Considerations

Technical considerations must be accounted for in order to optimize 3.0 T imaging. The most apparent of these include: chemical shift, fat saturation, and radiofrequency power deposition. Since the resonant frequency of fat and water protons increases linearly with the magnetic field strength, chemical shift artifact in the frequency encoding direction will be double with 3.0 T as compared to 1.5 T, if imaging bandwidth is kept the same (Fig. 5). One way of correcting for the chemical shift artifact includes doubling the receiver bandwidth. Increasing the receiver bandwidth from (+/-) 32 kHz at 1.5 T to (+/-) 64 kHz at 3.0 T will result in the same amount of chemical shift artifact experienced at 1.5 T. In addition to correcting for the chemical shift artifact increase, doubling the bandwidth also allows for a greater number of slices, less metal artifact and shorter TEs and echo spacing. However, doubling the bandwidth also results in an SNR decrease of β^2 since the overall readout window length at a higher bandwidth is shorter (Fig 6).

The doubling of the chemical shift distance between the fat and water resonance at 3.0 T and 1.5 T, makes fat saturation much easier. With a chemical shift of 440 Hz (18) the peaks are much further apart meaning that the fat saturation pulse lengths can be shortened to 8 msec instead of 16 msec. Resulting from this is the advantage of acquiring more slices at a given TR, slice thickness and bandwidth.

A third technical consideration is that of the radiofrequency power deposition. The radiofrequency power for excitation at 3.0 T is four times that at 1.5 T (19,20) due to the fact that the resonant frequency at 3.0 T is double that at 1.5 T. Many sequences used in musculoskeletal imaging, including fast spin-echo, have the potential for high radiofrequency power. The overall deposition depends on the amplitude and number of radiofrequency pulses. The use of rapid imaging sequences may reduce the radiofrequency power deposition. This complication should be minimized in small volume areas such as the knee since the radiofrequency power deposited is a function of tissue volume excited (19) (Fig. 7).

The primary disadvantage of imaging with short TE fast spin-echo sequences is blurring due to decreased signal echoes at the edges of k-space. This image blurring can partly be reduced by using a short echo-train length and higher receiver bandwidth on short TE fast spin-echo images. These technical considerations and other issues that are apparent with 3.0 T high field imaging are laid out in Table 1. It is important to note that it is much better to use a transmit/receive RF coil than a body coil transmit. However, if a body coil transmit is used, lowering the refocusing pulses or shortening the scan time to limit SAR would be appropriate. FDA limitations must also be taken into account. The FDA limits are 4 W/kg for the whole body for a 15 minute period for all patients, and the local SAR limit is 8 W/kg for extremities over a time period of 5 minutes.

Protocols

Protocols have been developed at 3.0 T MRI that show promising results for assessment of the knee joint (Tables 2 and 3). Imaging with the 3.0 T high resolution knee protocol keeps imaging time between 30 - 45 minutes while maintaining excellent image quality (Fig. 8). Another option, the 3.0 T rapid knee imaging protocol, allows for a scan time of 10 minutes while still producing outstanding quality images (Fig. 9). 3.0 T MRI provides a significant increase in SNR which can be used to either decrease exam time, thereby decreasing chances of motion artifact and increasing patient throughput and comfort or to enhance image quality by increasing resolution. Several studies have confirmed the previous statement and displayed remarkable contrast between fat, muscle, hyaline cartilage, fibrocartilage and fluid (10). Clinical 3.0 T high field imaging is becoming increasingly available and has shown promise for evaluation of the anatomy and pathology in the knee joint.

Future Directions

Isotropic Imaging

Isotropic or three dimensional (3D) imaging techniques allow for the acquisition of isotropic voxels as opposed to the typically acquired anisotropic voxels with two dimensional (2D) imaging. With isotropic voxels comes the ability to retrospectively reformat images into numerous planes allowing for better visualization of oblique structures, for example the anterior and posterior cruciate ligaments. A significant reduction in scan time also results from the ability to reformat images as only one acquisition is needed thereby avoiding multiple image plane acquisitions as occurs in 2D MRI. 2D imaging results in relatively thick slices which have gaps between them leading to partial-volume artifact. Isotropic imaging corrects this limitation by acquiring thin continuous slices thereby eliminating slice gaps and reducing

partial-volume artifact (21,22). While isotropic imaging is possible at 1.5 T, the increased SNR advantage at 3.0 T allows for better visualization of reformatted images (Fig. 10).

uTE imaging

Human tissues contain several components that have a wide range of T2 values. In tissues such as the liver and white matter, the spins typically have long T2 values. However, in the knee, ligaments, menisci, tendons, cortical bone and periosteum have short T2 values that range from hundreds of microseconds to tens of milliseconds (23). In conventional T2-weighted clinical imaging techniques, changes in the signal from only long T2 spins are highlighted while little or no signal is produced from tissues with short T2 values. Ultrashort TE (uTE) sequences are able to detect signal from tissues with short T2s by utilizing TEs that are 20-50 times shorter than those used in conventional MR sequences. Studies have shown that uTE imaging is capable of obtaining high signal from typically low signal tissues which allows for defects and layers of articular cartilage to be identified, zones of this menisci to be differentiated between and ligamentous scar tissue to be enhanced (24,25) (Fig. 11). The greater SNR at higher field imaging strengths improves uTE imaging.

T2 Mapping

Another recent advancement in the field of MR imaging is that of T2 mapping. While T2 relaxation times for a certain tissue are typically constant, tissue pathology can result in changes in these relaxation times. Even before symptoms arise, physiologic changes in the cartilage matrix begin taking place. The earliest detectable change in cartilage degeneration is an increased permeability throughout the matrix which allows for increased content and motion of water. A greater stress is generated in the cartilage matrix as the increased hydrodynamic fluid pressure is unable to sustain load support. This undue stress causes degeneration of the proteoglycan-collagen matrix as well as a loss of cartilage tissue. Since T2 relaxation time is a function of both the proteoglycan-collagen matrix and water content of the articular cartilage, quantitative T2 relaxation measures show promise as a valuable, noninvasive measure of articular cartilage integrity. Initial findings in several studies have demonstrated an increase in cartilage T2 relaxation times with asymptomatic degenerative changes (26-28).

Attention must be taken in selecting the appropriate MR imaging technique in attempting to accurately measure T2 relaxation time (29). A multiecho spin-echo technique is most often used and signal levels are matched to one or more decaying exponentials, contingent upon whether two or more T2 distributions are thought to be contained within the sample (30). However, a single exponential fit is acceptable for TEs that are used in conventional MR imaging. An image can be constructed with a color or grey-scale map that depicts the T2 relaxation times (28) (Fig. 12). T2 maps can also be made at 1.5 T, however, the increased SNR provided at 3.0 T imaging allows for better depiction of T2 relaxation times and early cartilage degeneration.

T1rho Imaging

T1rho imaging, or spin lattice relaxation in the rotating frame, is possible when the magnetization is “spin-locked” by a constant RF field after being tipped into the transverse plane. It is a method of examining the slow-moving interactions that occur between the static water molecules and the extracellular environment in which they live. Proteoglycan loss, an early biomarker of osteoarthritis (OA), results in changes to the macromolecular environment which can be indicated in T1rho measurements. This technique is able to acquire valuable biomedical information in low frequency systems and initial studies have shown it to be a promising tool the study of early OA development (28,31,32) (Fig. 12). T1rho imaging techniques can be utilized at both 1.5 T and 3.0 T field strengths, however depiction of proteoglycan loss is better optimized at 3.0 T due to the SNR increase.

Sodium

Sodium imaging, like T1rho imaging, has shown promise in measuring proteoglycan content as a marker of early, asymptomatic OA. This technique is made possible by the fact that the atom sodium-23 (^{23}Na), like the ^1H atom, has an odd number of protons or neutrons and therefore possesses a net nuclear spin allowing it to exhibit the MR phenomenon. ^{23}Na is much less prevalent in the body than ^1H , however it can be found in normal human cartilage at concentrations of up to approximately 320 mM. As a result of the lower concentration, lower resonant frequency and shorter T2 relaxation times than ^1H , ^{23}Na imaging presents new challenges and requires that special transmit-and-receive coils and long imaging times be used.

These accommodations prove worthwhile as ^{23}Na has demonstrated a promising ability to image early stages of OA because it has the capability of depicting regions of proteoglycan depletion (33). Since proteoglycans have a fixed negative charge that attracts the positive sodium atoms, as proteoglycan depletion occurs with OA, measurements of sodium levels within cartilage can give an accurate illustration of the level of pathology. While some spatial variation of sodium concentration is notable in healthy cartilage (34), sodium imaging has been demonstrated in studies to be sensitive to relatively small proteoglycan concentration changes (28,35,36) (Fig. 13).

Conclusion

MRI is accepted as one of the most accurate imaging modalities for assessment and evaluation of the musculoskeletal system while its advancement to 3.0 T high field imaging is becoming more refined and established in the clinical realm. 3.0 T imaging systems offer either superior image resolution or shortened imaging times, both resulting in several aforementioned advantages. Much promising research surrounding technical issues is being done in order to allow 3.0 T imaging to reach its full clinical potential. As research on optimization of 3.0 T high field imaging systems continues, the clinical world follows suit allowing for exquisite evaluation of the musculoskeletal system. The future of 3.0 T imaging is bright as research never ceases to push the boundaries of this already outstanding imaging modality.

Acknowledgments

Supported in part by:

NIH EB002524

NIH EB 005790

GE Healthcare

SCBT/MR

References

1. Standaert CJ, Herring SA. Expert opinion and controversies in musculoskeletal and sports medicine: stingers. *Arch Phys Med Rehabil* 2009;90:402–406. [PubMed: 19254603]
2. Khoury NJ, El Khoury GY. Imaging of musculoskeletal diseases. *J Med Liban* 2009;57:26–46. [PubMed: 19459576]
3. Kan JH. Major pitfalls in musculoskeletal imaging-MRI. *Pediatr Radiol* 2008;38(Suppl 2):S251–255. [PubMed: 18401621]
4. Cimmino MA, Grassi W, Cutolo M. Modern imaging techniques: a revolution for rheumatology practice. *Best Pract Res Clin Rheumatol* 2008;22:951–959. [PubMed: 19041071]
5. Bolog N, Nanz D, Weishaupt D. Musculoskeletal MR imaging at 3.0 T: current status and future perspectives. *Eur Radiol* 2006;16:1298–1307. [PubMed: 16541224]

6. Livstone BJ, Parker L, Levin DC. Trends in the utilization of MR angiography and body MR imaging in the US Medicare population: 1993-1998. *Radiology* 2002;222:615-618. [PubMed: 11867774]
7. Gold GE, Han E, Stainsby J, et al. Musculoskeletal MRI at 3.0 T: relaxation times and image contrast. *AJR Am J Roentgenol* 2004;183:343-351. [PubMed: 15269023]
8. Gold GE, Suh B, Sawyer-Glover A, et al. Musculoskeletal MRI at 3.0 T: initial clinical experience. *AJR Am J Roentgenol* 2004;183:1479-1486. [PubMed: 15505324]
9. Uematsu H, Takahashi M, Dougherty L, et al. High field body MR imaging: preliminary experiences. *Clin Imaging* 2004;28:159-162. [PubMed: 15158217]
10. Craig JG, Go L, Blechinger J, et al. Three-tesla imaging of the knee: initial experience. *Skeletal Radiol* 2005;34:453-461. [PubMed: 15968554]
11. Wong S, Steinbach L, Zhao J, et al. Comparative study of imaging at 3.0 T versus 1.5 T of the knee. *Skeletal Radiol* 2009;38:761-769. [PubMed: 19350234]
12. Niitsu M, Nakai T, Ikeda K, et al. High-resolution MR imaging of the knee at 3 T. *Acta Radiol* 2000;41:84-88. [PubMed: 10665878]
13. Bauer JS, Barr C, Henning TD, et al. Magnetic resonance imaging of the ankle at 3.0 Tesla and 1.5 Tesla in human cadaver specimens with artificially created lesions of cartilage and ligaments. *Invest Radiol* 2008;43:604-611. [PubMed: 18708853]
14. Masi JN, Sell CA, Phan C, et al. Cartilage MR imaging at 3.0 versus that at 1.5 T: preliminary results in a porcine model. *Radiology* 2005;236:140-150. [PubMed: 15987970]
15. Magee T, Williams D. 3.0-T MRI of meniscal tears. *AJR Am J Roentgenol* 2006;187:371-375. [PubMed: 16861540]
16. Schoth F, Kraemer N, Niendorf T, et al. Comparison of image quality in magnetic resonance imaging of the knee at 1.5 and 3.0 Tesla using 32-channel receiver coils. *Eur Radiol* 2008;18:2258-2264. [PubMed: 18463874]
17. Kuo R, Panchal M, Tanenbaum L, et al. 3.0 Tesla imaging of the musculoskeletal system. *J Magn Reson Imaging* 2007;25:245-261. [PubMed: 17260407]
18. Collins CM, Smith MB. Signal-to-noise ratio and absorbed power as functions of main magnetic field strength, and definition of "90 degrees" RF pulse for the head in the birdcage coil. *Magn Reson Med* 2001;45:684-691. [PubMed: 11283997]
19. Brix G, Seebass M, Hellwig G, et al. Estimation of heat transfer and temperature rise in partial-body regions during MR procedures: an analytical approach with respect to safety considerations. *Magn Reson Imaging* 2002;20:65-76. [PubMed: 11973031]
20. Shellock FG. Radiofrequency energy-induced heating during MR procedures: a review. *J Magn Reson Imaging* 2000;12:30-36. [PubMed: 10931562]
21. Kijowski R, Davis KW, Woods MA, et al. Knee joint: comprehensive assessment with 3D isotropic resolution fast spin-echo MR imaging--diagnostic performance compared with that of conventional MR imaging at 3.0 T. *Radiology* 2009;252:486-495. [PubMed: 19703886]
22. Gold GE, Busse RF, Beehler C, et al. Isotropic MRI of the knee with 3D fast spin-echo extended echo-train acquisition (XETA): initial experience. *AJR Am J Roentgenol* 2007;188:1287-1293. [PubMed: 17449772]
23. Robson MD, Gatehouse PD, Bydder M, et al. Magnetic resonance: an introduction to ultrashort TE (UTE) imaging. *J Comput Assist Tomogr* 2003;27:825-846. [PubMed: 14600447]
24. Gatehouse PD, Thomas RW, Robson MD, et al. Magnetic resonance imaging of the knee with ultrashort TE pulse sequences. *Magn Reson Imaging* 2004;22:1061-1067. [PubMed: 15527992]
25. Gold GE, Thedens DR, Pauly JM, et al. Imaging of Articular Cartilage of the Knee: New Methods Using Ultrashort TEs. *AJR Am J Roentgenol* 1998;170:1223-1226. [PubMed: 9574589]
26. Mosher TJ, Dardzinski BJ, Smith MB. Human articular cartilage: influence of aging and early symptomatic degeneration on the spatial variation of T2--preliminary findings at 3 T. *Radiology* 2000;214:259-266. [PubMed: 10644134]
27. Dardzinski BJ, Mosher TJ, Li S, et al. Spatial variation of T2 in human articular cartilage. *Radiology* 1997;205:546-550. [PubMed: 9356643]
28. Gold GE, Chen CA, Koo S, et al. Recent advances in MRI of articular cartilage. *AJR Am J Roentgenol* 2009;193:628-638. [PubMed: 19696274]

29. Poon CS, Henkelman RM. Practical T2 quantitation for clinical applications. *J Magn Reson Imaging* 1992;2:541–553. [PubMed: 1392247]
30. Smith HE, Mosher TJ, Dardzinski BJ, et al. Spatial variation in cartilage T2 of the knee. *J Magn Reson Imaging* 2001;14:50–55. [PubMed: 11436214]
31. Akella SV, Regatte RR, Gougoutas AJ, et al. Proteoglycan-induced changes in T1rho-relaxation of articular cartilage at 4T. *Magn Reson Med* 2001;46:419–423. [PubMed: 11550230]
32. Li X, Han ET, Ma CB, et al. In vivo 3T spiral imaging based multi-slice T(1rho) mapping of knee cartilage in osteoarthritis. *Magn Reson Med* 2005;54:929–936. [PubMed: 16155867]
33. Reddy R, Insko EK, Noyszewski EA, et al. Sodium MRI of human articular cartilage in vivo. *Magn Reson Med* 1998;39:697–701. [PubMed: 9581599]
34. Shapiro EM, Borthakur A, Gougoutas A, et al. ²³Na MRI accurately measures fixed charge density in articular cartilage. *Magn Reson Med* 2002;47:284–291. [PubMed: 11810671]
35. Borthakur A, Shapiro EM, Beers J, et al. Sensitivity of MRI to proteoglycan depletion in cartilage: comparison of sodium and proton MRI. *Osteoarthritis Cartilage* 2000;8:288–293. [PubMed: 10903883]
36. Borthakur A, Hancu I, Boada FE, et al. In vivo triple quantum filtered twisted projection sodium MRI of human articular cartilage. *J Magn Reson* 1999;141:286–290. [PubMed: 10579951]

Fig. 1. Proton density-weighted images displaying a healthy anterior cruciate ligament at 1.5 and 3.0 T

A and **B**, sagittal images at 1.5 T (**A**) and at 3.0 T (**B**). 3.0 T provides better visualization and delineation of the anterior cruciate ligament when compared with 1.5 T (*arrows*, **A**, **B**).

Fig. 2.
Images displaying meniscal pathology at 1.5 and 3.0 T

A and **B**, sagittal images at 1.5 T (**A**) and at 3.0 T (**B**). High field, 3.0 T imaging, allows for better visualization of a meniscal tear when compared to lower field imaging at 1.5 T (*arrows*, **A**, **B**).

Fig. 3.
T2-weighted images showing bone marrow edema at 1.5 and 3.0 T
A and **B**, sagittal images at 1.5 T (**A**) and at 3.0 T (**B**). Bone marrow edema is visualized in much greater detail with 3.0 T than with 1.5 T (*arrows, A, B*).

Fig. 4.
Images of a healthy knee obtained at 1.5 and 3.0 T
A and **B**, axial proton density-weighted images at 1.5 T (**A**) and 3.0 T (**B**). Increase in signal-to-noise ratio is evident in 3.0 T. Signal-to-noise ratio is 1.8.

C and **D**, sagittal proton density-weighted images at 1.5 T (**C**) and 3.0 T (**D**). Increased signal-to-noise ratio of 3.0 T is noticeable. Signal-to-noise ratio is 1.8.

Fig. 5.

Images of a healthy knee obtained at 1.5 and 3.0 T

A and **B**, sagittal images at 1.5 T (**A**) and at 3.0 T (**B**) with a bandwidth of 20 kHz. Pronounced chemical shift is visualized in 3.0 T images since bandwidth is the same at both field strengths (*arrows, A, B*).

Fig. 6. Proton density-weighted images of the knee at 3.0 T
A and **B**, sagittal images at bandwidths of 15 kHz(**A**) and at 42 kHz(**B**). Chemical shift is minimized with an almost three-fold increase in bandwidth which can be seen as a significantly sharper anatomy and much thinner subchondral bone thickness (*arrows, A, B*).

Fig. 7.
Coronal T1-weighted 3.0 T images of the healthy knee.

Radiofrequency power deposition complications with 3.0 T use can be reduced with limited use of fast spin-echo imaging for decreased TR sequences.

A, T1-weighted spin-echo image at 3.0 T (TR = 800) caused power monitor to reach 66% the limit of the average radiofrequency power.

B, T1-weighted fast spin-echo image at 3.0 T (TR/TE = 800/2) caused power monitor to reach 33% of the average radiofrequency power limit shows slight blurring resulting from the use of a short TE and two echoes.

Fig. 8. High resolution knee imaging protocol at 3.0 T obtained with a fast spin echo sequence. **A**, axial proton density-weighted, **B**, coronal T1-weighted, **C**, coronal T2-weighted, **D**, sagittal proton density-weighted, **E**, sagittal T2-weighted, and **F**, coronal 3D proton density-weighted images obtained using the high resolution knee imaging protocol at 3.0 T.

Fig. 9.
Rapid knee imaging protocol at 3.0 T obtained with a fast spin echo sequence.

A, axial proton density-weighted, **B**, coronal T1-weighted, **C**, coronal T2-weighted, **D**, sagittal proton density-weighted, and **E**, sagittal T2-weighted images obtained using the rapid knee imaging protocol at 3.0 T.

Fig. 10.

Isotropic imaging with 3D Fast Spin Echo.

A, coronal acquisition, **B**, sagittal reformat, **C**, axial reformat. **B** and **C** display high quality images that can be obtained by acquiring and reformatting only **A**.

Fig. 11.

uTE images of patellar articular cartilage.

A, acquired with TR of 500 milliseconds and a TE of 13.9 microseconds, **B**, acquired with a TR of 300 milliseconds and a TE of 8 microseconds. By allowing for direct visualization of short T2 components, signal alteration in the superficial cartilage at the median patella ridge shows subchondral bone disease (*arrows, A, B*). (Images courtesy of Christine Chung, UCSD Medical Center, San Diego, CA)

Fig. 12.

Images of a 6-month post synthetic biphasic co-polymer plug repair.

A, proton image **B**, color T2 map depicting relaxation times. Color scale displays the range of T2 relaxation times with larger relaxation times displayed in red and shorter relaxation times displayed in blue. The red and orange regions show the beginning of degenerative changes in the cartilage matrix. **C**, color T1rho map. Color scale exhibits the range of T1rho measurements. Regions of proteoglycan loss can be seen in orange and red (*arrows, B, C*). (Images courtesy of Hollis Potter, Hospital for Special Surgery, New York, NY)

Fig. 13.

3.0 T images of a 3 year after anterior cruciate ligament repair.

A, proton image, **B**, sodium heatmap overlay on a proton image, **C**, registered sodium 3D cones image. **B** and **C** depict the capabilities of sodium imaging to display proteoglycan content.

Note focal area of sodium reduction indicating a decreased proteoglycan content (*arrows*, **B**, **C**).

Table 1

Summary of technical considerations and modifications that occur while imaging at 3.0 T. Solutions and disadvantages of each modification are also listed

Protocol Optimization at 3.0 T Summary			
3.0 T Consideration	Solution	Benefits of Solution	Disadvantages of Solution
Increased T1	Lengthen TR	Increasea SNR	Increase in scan time
Decreased T2	Shorten TE	Increases SNR	Blurring due to decreased signal echoes at the edges of k-space
Chemical Shift Artifact	Double receiver bandwidth on non fat sat sequences or use fat sat	More slices, less metal artifact and shorter TEs and echo spacing	SNR decrease of $\sqrt{2}$
RF Power Deposition	Use of more rapid sequence of decreased flip angle refocusing pulses on fast spin echo	Non-issue with small volume transmit-receive coils	More rapid imaging or decreased flip angle refocusing pulses lowers image SNR

Table 2

Sample protocol design for high resolution imaging at 3.0 T

High Resolution Knee Protocol Design at 3.0 T						
Fast Spin Echo Sequence						
Imaging Parameter	Axial PD *	Coronal T1	Coronal T2 *	Sagittal PD	Sagittal T2	Coronal 3D
Repetition Time (msec)	5000	1000	5000	5000	5000	3000
Echo Time (msec)	20	15	60	15	54	35
Matrix Size	416 × 320	416 × 320	416 × 320	512 × 320	384 × 320	288 × 288
Field of View (cm)	14	14	14	16	16	17
Number of Slices	26	18	22	30	30	200
Bandwidth (kHz)	32	41	32	41	32	62
Echo Train Length	8	3	8	8	8	60
Section Thickness (mm)	2.5	2.5	2.5	2.5	2.5	.6
Number of Averages	2	2	2	15	1.5	1.5
Imaging Time (min)	6:00	3:30	6:00	5:00	5:00	5:00

* Fat Suppressed Imaging

Table 3

Sample rapid protocol design for 3.0 T imaging

Rapid Knee Protocol Design at 3.0 T					
Fast Spin Echo Sequence					
Imaging Parameter	Axial PD *	Coronal T1	Coronal T2 *	Sagittal PD	Sagittal T2
Repetition Time (msec)	5000	1000	4000	5000	6400
Echo Time (msec)	35	20	54	35	60
Matrix Size	320 × 224	384 × 224	320 × 224	384 × 224	320 × 224
Field of View (cm)	14	16	16	14	14
Number of Slices	26	18	22	30	30
Bandwidth (kHz)	32	32	32	32	32
Echo Train Length	8	4	8	8	10
Section Thickness (mm)	4	4	4	3	3
Imaging Time (min)	1:25	1:43	2:24	2:30	2:40

* Fat Suppressed Imaging