CORRESPONDENCE

Point-of-Care Testing in Hospitals and Primary Care

by Prof. Dr. med. Ralf Junker, Prof. Dr. med. Harald Schlebusch,
Prof. Dr. med. Peter B. Luppä in volume 33/2010

Practice Laboratory

The point of care testing mentioned in the article is by no means a new idea. I have practiced as a general practitioner since 1967, and up to the 1990s I used to conduct comprehensive absorption photometry (wet chemistry) and dry chemistry (analysis with carrier-bound reagents) in my own practice laboratory. This enabled me to tell patients their blood counts in next to no time and initiate treatment accordingly. Today, a patient visiting their doctor on a Friday will have to wait until the following Monday.

The extremely low reimbursement paid for laboratory services by the statutory sickness funds, combined with the high costs of reagents and test strips (which exceeded the reimbursement amount by multiples) and the disproportionately large package sizes (which were only partly used up by the use-by date) on the part of the delivering pharmaceutical companies and the expensive costs of purchasing a suitable photometer (which did not pay for itself) prompted me to give up my own practice laboratory. It had become a financial burden that I could not justify any more, an unprofitable hobby.

If the thinking is about to change then the mentioned negative measures would need to be cancelled out and corrected. In the interest of patients, a renaissance of previous diagnostic options would be welcome and desirable, even if it is probably not really to be expected.

DOI: 10.3238/arztebl.2010.0129a

REFERENCES

1. Junker R, Schlebusch H, Luppä PB: Point-of-care testing in hospitals and primary care. *Dtsch Arztebl Int* 2010; 107(33): 561–7.

Dr. med. Gerhard Schnorrenberg
Stettiner Str. 1a, 83088 Kiefersfelden, Germany

Unsatisfactory Quality

I wish to add to the review article by Junker et al that selecting a suitable test according to the medical requirements presents a great challenge: both the quality of the testing procedure as well as the entire (pre-)analytical process have to be considered. At the point of care setting, comparative methods are often not available, and users are likely unaware of the unsatisfactory quality of their testing results.

One example: cholesterol testing in the pharmacy constitutes a big hazard. This testing is performed by pharmacists, therefore outside medicine, but it is requested by patients to monitor the effects of

dietary or pharmacological therapies. POCT tests are unsuitable for this purpose as the highly problematic preanalytics —variable squeezing of the finger tip—makes measuring cholesterol from capillary blood an invalid method and the (necessary) testing of more analytes such as total cholesterol and LDL or HDL and LDL is impossible due to the limited sample volume available.

My second remark concerns the point of care testing of Hb_{A1c}. A reduction of Hb_{A1c} by 10% lowers the risk of diabetic retinopathy by ~44% (1), and exact monitoring is obligatory, also because of the possible fatal side effects of adjusting diabetes treatment too stringently. In order to comply with the guidelines of the American Diabetes Association (2) the Hb_{A1c} method used should have a sufficiently low deviation (<5%) from the “true” value (3). Some point of care tests have shown an extremely poor performance and, in addition, large lot-to-lot differences (3). Because of the very poor performance of numerous test systems in the last German external quality schemes (“Ringversuche”), the allowable deviation was increased to 30%!

For the patients' wellbeing, physicians should only use appropriate test systems.

DOI: 10.3238/arztebl.2010.0129b

REFERENCES

1. The effect of intensive treatment of diabetes on the development and progression of long-term complications in insulin-dependent diabetes mellitus. The Diabetes Control and Complications Trial Research Group. *N Engl J Med* 1993; 329: 977–86.
2. Executive Summary: Standards of medical care in diabetes—2010. *Diabetes Care* 2010; 33: 4–10.
3. Leters-Westra E, Slingerland RJ: Six of eight Hemoglobin A1c Point-of-Care instruments do not meet the general accepted analytical performance criteria. *Clinical Chemistry* 2010; 56: 44–52.
4. Junker R, Schlebusch H, Luppä PB: Point-of-care testing in hospitals and primary care. *Dtsch Arztebl Int* 2010; 107(33): 561–7.

PD Dr. med. Matthias Orth
Institut für Laboratoriumsmedizin
Vinzenz von Paul Kliniken gGmbH
Postfach 103161, 70027 Stuttgart, Germany
orth@vinzenz.de

In Reply:

In some extra-European studies, an allegedly positive influence of point of care measuring of a patient's lipid status and of monitoring the metabolic situation in pharmacies by means of point of care testing have been described, which manifested as a lowering of individual parameters of the lipid metabolism (1, 2). By contrast to the situation in Germany, however, treatment was directed immediately by pharmacists. Such an approach leaves out medical advice with regard to the type and extent of the diagnostic evaluation (including the critical preanalytics) as well as regarding different necessary and possible therapeutic

approaches. The problem is less one of measuring techniques or the idea of point of care testing per se as, rather, one of the way in which laypersons deal with medical questions.

Orth rightly points out the problems with faulty Hb_{A1c} measurements, whose negative effects have been investigated epidemiologically. Systematic methodological comparisons between point of care methods and conventional laboratory diagnostics have shown a high performance for point of care testing for many parameters. Hb_{A1c}, however, presents many analytical problems for many measuring devices. The deluge of new systems entering the market and of tests that are available from internet providers, but also from pharmacies or other distributors, thus poses a challenge in distinguishing appropriate from less appropriate methods.

Schnorrenberg describes the difficulties in running a practice laboratory in an economical fashion. In the past this has resulted in laboratory communities of physicians in private practice, which helped to provide laboratory services centrally and cost effectively, and allowed for these to be accounted for by the practices that requested them. Today's point of care systems differ from the practice laboratory in that, according to a definition from the German Society of Clinical Chemistry and Laboratory Medicine, such tests can be performed by persons without laboratory training (3). The German Medical Association therefore requires for any point of care laboratory test the complete quality assurance that is required for a conventional labora-

tory—unless so called unit-use reagents are used, which are based on real once-only measurements with single-use portions of reagents (4). Such tests are usually much more expensive than conventional laboratory tests and do not really present a modern alternative to the practice laboratory, as long as the reimbursement from the statutory sickness funds remains at its current levels.

DOI: 10.3238/arztebl.2010.0129c

REFERENCES

1. Mahtabjafari M, Masih M, Emerson AE: The value of pharmacist involvement in a point-of-care service, walk-in lipid screening program. *Pharmacotherapy* 2001; 21: 1403–6.
2. Nola KM, Gourley DR, Portner TS, et al.: Clinical and humanistic outcomes of a lipid management program in the community pharmacy setting. *J Am Pharm Assoc* 2000; 40: 166–73.
3. Briedigkeit L, Müller-Plathe O, Schlebusch H, Zierass J: Patientennahe Laboratoriumsdiagnostik (Point-of-Care-Testing). *J Lab Med* 1998; 22: 414–20.
4. Bundesärztekammer. Richtlinie der Bundesärztekammer zur Qualitätssicherung laboratoriumsmedizinischer Untersuchungen. *Dtsch Arztebl* 2008; 105(7): A 341–55.
5. Junker R, Schlebusch H, Lupp PB: Point-of-care testing in hospitals and primary care. *Dtsch Arztebl Int* 2010; 107(33): 561–7.

Prof. Dr. med. Ralf Junker

Universitätsklinikum Schleswig-Holstein
Institut für Klinische Chemie, Campus Kiel
Arnold-Heller-Str. 3
Haus 17, 24105 Kiel, Germany

Conflict of interest statement

The authors of all contributions declare that no conflict of interest exists according to the guidelines of the International Committee of Medical Journal Editors.