Published in final edited form as:

Methods Mol Biol. 2010 ; 656: 465–479. doi:10.1007/978-1-60761-746-4_27.

Mass Spectrometry Imaging Using the Stretched Sample Approach

Tyler A. Zimmerman, Stanislav S. Rubakhin, and Jonathan V. Sweedler

Department of Chemistry and Beckman Institute, University of Illinois at Urbana-Champaign, Urbana, IL

Abstract

Matrix-assisted laser desorption/ionization (MALDI) mass spectrometry imaging (MSI) can determine tissue localization for a variety of analytes with high sensitivity, chemical specificity, and spatial resolution. MS image quality typically depends on the MALDI matrix application method used, particularly when the matrix solution or powder is applied directly to the tissue surface. Improper matrix application results in spatial redistribution of analytes and reduced MS signal quality. Here we present a stretched sample imaging protocol that removes the dependence of MS image quality on the matrix application process and improves analyte extraction and sample desalting. First, the tissue sample is placed on a monolayer of solid support beads that are embedded in a hydrophobic membrane. Stretching the membrane fragments the tissue into thousands of nearly single-cell sized islands, with the pieces physically isolated from each other by the membrane. This spatial isolation prevents analyte transfer between beads, allowing for longer exposure of the tissue fragments to the MALDI matrix, thereby improving detectability of small analyte quantities without sacrificing spatial resolution. When using this method to reconstruct chemical images, complications result from non-uniform stretching of the supporting membrane. Addressing this concern, several computational tools enable automated data acquisition at individual bead locations and allow reconstruction of ion images corresponding to the original spatial conformation of the tissue section. Using mouse pituitary, we demonstrate the utility of this stretched imaging technique for characterizing peptide distributions in heterogeneous tissues at nearly single-cell resolution.

Keywords

mass spectrometry imaging; matrix-assisted laser desorption/ionization; nervous tissue; pituitary; mouse; stretched sample; image reconstruction; automated data acquisition

1. Introduction

Both invertebrate and mammalian nervous systems exhibit high levels of biochemical and morphological heterogeneity. Neighboring neurons often possess different sets of intercellular signaling peptides, with several prohormones encoding multiple peptides that are expressed differently among individual neurons. Investigation of the mammalian nervous system, where neurons number in the billions, presents a significant challenge when using classical labeling approaches to examine one or more cell-to-cell signaling molecules at a time. In contrast to other bioimaging techniques, mass spectrometry imaging (MSI) can uncover the distribution of a variety of analytes within tissues while simultaneously

*Corresponding author Address: Department of Chemistry, University of Illinois, 600 South Mathews Ave. 63-5, Urbana, IL 61801
Voice: 217-244-7359 Fax: 217-244-8068 sweedler@scs.uiuc.edu.

determining their chemical identities, without the need for specific labeling or immunostaining (1–5). MSI has broad applications in academic, clinical, and industrial research, having had significant impact on cancer studies (6–9), the search for new pharmaceuticals (10), and investigations of the nervous system (11). A variety of MSI approaches targeting different types of analytes have been developed over the years. MALDI MSI has become one of most successful technologies for investigation of peptide and protein distributions in fixed and freshly prepared tissues. The analyte desorption and ionization processes occurring during exposure of the MALDI matrix/analyte layer to UV or IR laser light allows detection of intact/unfragmented analytes. Not only is the entire sequence of ion desorption, formation, separation, and detection fast, the laser beam can be focused to sub-micrometer diameters. However, because the amount of proteins and peptides present decrease concomitantly with the size of the area being probed, the smallest laser diameters are not commonly used. Obviously, larger laser spot sizes allow desorption of increased amounts of analyte. Typical chemical images are generated with 25 – 50 μm spatial resolution.

In MALDI MSI, a liquid or powder matrix is deposited on top of the sample, incorporating the analyte into the matrix crystals. When illuminated with the laser, the matrix and analyte are vaporized and ionized. Although longer exposure of the sample to matrix facilitates extraction of the analyte of interest from the tissue, it can also delocalize the analyte. Shorter extraction times ameliorate this problem, but result in poorer signal. This issue is particularly problematic when investigating small hydrophilic substances that diffuse during matrix application. Recent advances in MALDI matrix solution application approaches have helped to create uniform MALDI matrix layers; these include spray coating (12), electrospray deposition (13), and automated acoustic deposition (14). Using these techniques, the imaging of fine analyte spatial distributions has been achieved. However, each type of sample and even class of analytes requires individualized optimization of the MALDI matrix exposure duration and drying time.

The stretched sample protocol resolves these issues by eliminating redistribution of analytes during the matrix application stage. A tissue slice is placed on top of a monolayer comprised of ~40- μm diameter glass beads, which has been partially embedded into a layer of Parafilm M (15). As the Parafilm M layer is manually stretched to a ~16-fold increase in area, the beads separate from each other and the tissue, which adheres to the beads, is fragmented into thousands of islands. Because each bead contains only one or a few cells, chemically and spatially separated by areas of hydrophobic membrane, the sample can be exposed to the MALDI matrix solution for a longer period of time without sacrificing spatial resolution. The spatial isolation of tissue fragments allows rare signals from small cellular clusters, single cells, and even subcellular regions to be better detected and spatially distinguished. Furthermore, multiple MALDI matrix wetting/recrystallization cycles can be accomplished via temperature-dependent condensation of solvents onto the stretched sample, which allows for increased incorporation of the analyte into the matrix, contributing to further signal enhancement.

Effective MS imaging of stretched samples demands new methods of data acquisition, along with image reconstruction protocols, to register the spectral data with the corresponding conformation of the tissue before stretching. Although classical MSI experiments collect data in a regular raster pattern over the sample (16), the small tissue/cell islands in the stretched sample occur in irregular spatial patterns on the Parafilm M membrane. Incorporating a step to identify bead positions from optical images of the sample via light thresholding allows automated MS data acquisition from the individual bead positions. Image reconstruction is done *in silico* with a free transform process that mimics the actual stretching process (17). During stretching, the beads tear sizable craters partially through the

layer of Parafilm M, visible in an optical image of the stretched sample. Image reconstruction is performed by aligning an image of the initial positions of the embedded beads with an image of the tissue sample after stretching. The spectral data taken from the stretched sample are assigned to the nearest corresponding initial bead positions to reconstruct an ion image of the tissue in its original conformation. This novel stretched imaging method shows increased potential for identifying rare signals from heterogeneous tissue samples (18).

2. Materials

2.1 Preparation of Parafilm M Substrate

1. Parafilm M (Pechiney, Neenah, WI).
2. Glass slides, $25 \times 75 \times 1.1$ mm, (Delta Technologies, Stillwater MN).
3. At least 100 mg of ~40-um diameter clear glass beads (Mo-Sci Corp., Rolla, MO). Blue beads are optional, and are used as markers to aid image reconstruction. Beads of other types and sizes including liquid chromatography solid phase materials can be also used.
4. A heated aluminum block.

2.2 Tissue Preparation and Sample Stretching

1. Four month-old C57BL/6 mice obtained from an in-house colony bred by the Greenough group, University of Illinois at Urbana-Champaign, were used. Animals of similar strains can be purchased for research purposes from the Jackson (<http://www.jax.org>) or Harlan laboratories (<http://www.harlan.com>). A variety of tissues from different animals can be investigated using the protocol presented here.
2. SPECTRA-SONIC (or similar) solution, pH 7 (Spectrum Surgical Instruments Corp., Stow, OH) for surgical instrument clean up.
3. Modified Gey's balanced salt solution (mGBSS): 1.5 mM CaCl_2 , 4.9 mM KCl, 0.2 mM KH_2PO_4 , 11 mM MgCl_2 , 0.3 mM MgSO_4 , 138 mM NaCl, 27.7 mM NaHCO_3 , 0.8 mM Na_2HPO_4 , 25 mM HEPES and 10 mM glucose, pH 7.2 adjusted with NaOH.
4. Dissection tools including forceps, scissors (available online from Fine Science Tools <http://www.finescience.com> or World Precision Instruments - <http://www.wpiinc.com>), and a properly sharpened guillotine.
5. Cryostat capable of keeping the specimen temperature at -15 to -20°C and of cutting 10- μm sections, e.g., Microm HM550 (Thermo Scientific, Waltham, MA).
6. Indium tin oxide (ITO)-coated glass slides, $25 \times 75 \times 1.1$ mm, (Delta Technologies, Stillwater MN).
7. A piece of firm paper.
8. Liquid nitrogen and dry ice.
9. Vials for specimen storage.
10. Protective lab coat, gloves and goggles.

2.3 MALDI Matrix Application

1. MALDI matrix solution: 300 mg of 2,5-dihydroxybenzoic acid (Sigma-Aldrich, St. Louis, MO) in 10 mL of 75:25 acetone:water.
2. Artist's spray brush (Badger, Franklin Park, IL).
3. In-house built condensation chamber consisting of a Peltier device (Melcor, Trenton, NJ) connected to a cooling basin of water, and a thermocouple connected to a CN77000 temperature controller (Omega, Stamford, CT).
4. Acetone.

2.5 Mass Spectrometry and Automated Imaging

1. Optical stereomicroscope.
2. Inverted transmission light microscope with 2.5 – 10× magnification and equipped with a digital camera (e.g., AxioCam MRc camera controlled by the AxioVision digital image processing software package, Carl Zeiss, Bernreid, Germany; AxioVision LE is free and a sample version of the full package is available at <http://www.zeiss.com/>).
3. Ultraflex II MALDI-TOF mass spectrometer (Bruker Daltonics, Billerica, MA) with a solid-state UV laser.
4. MTP slide adapter (Bruker Daltonics) for insertion of slides into the MS instrument.
5. ImageJ, version 1.38 (National Institutes of Health, <http://rsb.info.nih.gov/ij/>).
6. Java SDK, version 1.6.0 (Sun Microsystems, <http://java.sun.com>).
7. FlexControl 3.0 (Bruker Daltonics).
8. Bead geometry application (free at <http://neuroproteomics.scs.uiuc.edu/imaging.html>).

2.6 Data Conversion

1. Software tool: CompassXport (Bruker Daltonics, free at <http://www.brukerdaltonics.com>; for more information see: http://www.ionsource.com/functional_reviews/CompassXport/CompassXport.htm).
2. Software package: MATLAB R2006a, version 7.2, and the Bioinformatics Toolbox 3.0 (The MathWorks, Natick, MA).
3. Batch conversion MATLAB code (free at <http://neuroproteomics.scs.uiuc.edu/imaging.html>).

2.7 Image Reconstruction

1. Photoshop CS, version 8.0 (Adobe Systems).
2. Java-based code to create image of dots at the initial bead positions (free at <http://neuroproteomics.scs.uiuc.edu/imaging.html>).
3. MSIRestructor application (free at <http://neuroproteomics.scs.uiuc.edu/imaging.html>).

3. Methods

3.1. Preparation of Parafilm M Substrate

1. Parafilm M is cut into a square measuring approximately 5 cm × 5 cm and placed on top of a glass slide (*see* Note 1). In this step, the slide is used as a clean solid support and so does not require a conductive ITO-coated slide.
2. Approximately 100 mg of beads are transferred to the Parafilm M surface. Another glass slide is placed on top of the beads and vertical pressure is manually applied to partially embed the beads into the Parafilm M layer.
3. Application of a nitrogen stream to the substrate removes loose beads, ensuring an even monolayer is attached to the Parafilm M surface.
4. Placing the substrate between two glass slides and heating it on top of an aluminum block at ~60 °C for 10 – 15 s with downward manual pressure allows the beads to become more strongly and uniformly attached to the Parafilm M. Care must be taken to ensure that peripheral parts of the Parafilm M section, which might touch the metal block, do not melt onto it. A small separate piece of Parafilm M can be used to test if the temperature is such that the Parafilm M might be melted by the metal block.
5. An optical image of the initial bead/Parafilm M substrate is taken in transmission mode (*see* Note 2).

3.2. Tissue Preparation and Sample Stretching

1. Experimental animals are selected and euthanized by decapitation. Importantly, work performed on animals should comply with local and federal rules and regulations for the humane care and treatment of animals.
2. Surgical/dissection instruments are cleaned and sterilized by ultrasonic treatment in SPECTRA-SONIC (or similar) solution for 5 – 10 min, followed by autoclaving according to the manufacturer's manual.
3. Vials and paper are prepared; protective lab coat, gloves, and goggles are worn.
4. The animal is decapitated using a sharp guillotine.
5. The cranium is exposed by pushing the skin in a rostral direction using a piece of firm paper.
6. The cranial bones in the frontal plane are cut using long, thin scissors.
7. The brain is carefully lifted and discarded after removal of the previously cut dorsal part of the cranium. The pituitary typically remains in the skull, held in place by connective tissue.
8. The connective tissue surrounding the pituitary should be removed first; the pituitary is quickly removed using fine forceps.

¹Optionally, Parafilm M may be soaked for 1 h in either acetic acid (100%) or ammonium hydroxide solution (28.8 %) to soften the film (19). After drying, this treatment allows the film to be stretched by a greater degree into a ~7-fold increase in each dimension. Use of the more elastic film results in formation of small, concentrated droplets of solution upon matrix application with less spatial spreading. In addition, soaking can reduce polymer signals resulting from the Parafilm M.

²If only performing mass spectral profiling without imaging on stretched samples, as described in reference (15), this step and the steps related to image reconstruction and geometry file creation can be omitted.

9. The pituitary is briefly washed in ice cold mGBSS. The tissue is quickly frozen in liquid nitrogen and stored in a vial over dry ice for transport to the cryostat environment.
10. The pituitary is placed on a cooled (to -20°C) sample stage inside of the cryostat, without addition of embedding solution (see Note 3).
11. Tissue sections are made (10- μm thick).
12. Within the cryostat, the room temperature bead substrate is positioned onto the tissue section and briefly pressed using an index finger or an artist's brush handle. This ensures transfer of the tissue from the cryostat surface onto the substrate surface. Using a magic marker, the orientation and perimeter of the tissue section within the bead substrate are marked on the back of the Parafilm M substrate.
13. An ITO-coated glass slide is mounted onto a tall, thin vertical support (the slide box cover works well) with tape, conductive side facing upwards (Fig. 1A). A digital multimeter can be used to determine which side is conductive. The glass slide box cover is the appropriate shape to support the slide as the stretched substrate is pushed onto the slide. This vertical support enables stretching without having to also hold the glass slide and can be stabilized by placing it between two holders such as large books, as illustrated in Fig. 1A.
14. The sample is stretched by hand and attached to the ITO-coated slide (Fig. 1B – E), and the excess Parafilm M is manually torn off of the sides of the slide (see Note 4). To ease the subsequent process of image reconstruction, the sample should be stretched with the maximum directional uniformity possible. The magic marker label along the perimeter (described above) helps when visually inspecting the sample to ensure it retains gross shape after the stretching process (see Fig. 1F).

3.3 MALDI Matrix Application

1. MALDI matrix is applied using the artist's spray brush at an $\sim 25\text{ cm}$ distance from the sample (see Note 5). The spray brush is washed with pure acetone after use.
2. Water is condensed onto the sample with the condensation chamber at 14°C for 60 s and the sample evaporated at 28°C for 90 s. This cycle is repeated three times for increased analyte extraction, followed by returning the sample to room temperature (see Note 6).
3. The specimen is loaded into the mass spectrometer. Mass spectral profiling (15) is used to assess the quality of peptide signal received from the specimen before MS imaging of the sample.

³Most embedding media interferes with obtaining good signals in mass spectrometry investigations. One exception to this is embedding of tissues in low melting point agarose gel blocks or gelatin. It was found that a block of solidified saturated agarose solution that is freeze-mounted onto the dissection stage, followed by sectioning through the top layers of the gel, creates a flat surface for better freeze-mounting and orienting of the tissue.

⁴For mass spectrometry imaging, it is important to prevent the tissue from completely drying as this will cause bead clumping and may reduce incorporation of analyte into the MALDI matrix crystals. Excess sample drying is prevented by immediately applying MALDI matrix after stretching while the tissue is still partially wet. Thus, if more than one section is to be taken from the tissue, these sections are sectioned with the microtome after immediately applying MALDI matrix to the preceding stretched section.

⁵A larger sprayer-to-sample distance helps in not over-wetting the sample, as larger ($\sim 0.5\text{ mm}$) droplets can cause spreading, even in a stretched sample. A light microscope can be used to monitor the drying process, so that the sample is completely dry before the next spray application. The light microscope also helps to visually monitor the amount of matrix applied. Generally, several spraying-drying cycles over 10 – 20 min is sufficient. Alternatively, matrix application can be done by capillary deposition to control the size of the matrix spots and prevent the spatial redistribution of analytes (19).

⁶The solvent condensation/MALDI matrix recrystallization procedure has shown the ability to improve mass spectra quality by reducing the number and intensity of inorganic salt ion adducts typical for traditional MS imaging sample preparations. This reduction in potassium and sodium salt adducts creates less complex mass spectra (20).

3.4 Mass Spectrometry and Automated Imaging

1. The glass slide with the stretched sample is loaded into the mass spectrometer. Although a calibration bar is typically used (18), we determined that regularly-spaced laser-melted holes in the Parafilm M serve as more accurate spatial calibration markers. Provided that the sample is stretched to sufficient thinness, the mass spectrometer's UV laser beam is used to melt several ~100-micrometer diameter holes through the Parafilm M surface at several of the ordered positions found in the "MTP Slide Adapter II" geometry file included in the Bruker FlexControl software. The location of these points should be chosen so that they span the area of the tissue sample; depending on the size of the sample, 3 – 4 points is sufficient.
2. The specimen is unloaded from the mass spectrometer and a transmission mode optical image is taken with a digital camera coupled to an optical microscope. If several optical images are needed to cover the entire area of the sample, the Photomerge function in Photoshop can be used to stitch multiple images together.
3. ImageJ, along with the color threshold plugin, is used to automatically report the pixel coordinates of the beads (*see* Note 7). The computational steps for bead identification by thresholding are summarized in Fig. 2A. The Analyze-->Set Scale function is used to specify the units of the coordinates as pixels. The results of the threshold are viewed by selecting Analyze-->Analyze Particles-->Show Outlines. The circularity and size parameters can be adjusted and the process repeated until the outputted outlines image appears not to be highlighting non-bead regions and irregular shapes.
4. In ImageJ, the pixel coordinates of the center of the melted calibration regions are recorded and the distances between them are calculated using the Cartesian distance formula (*see* Note 8). If several equivalent distances can be calculated between the various calibration points, the variations in these distances are averaged (*see* Note 9).
5. The coordinates of the calibration points and stretched sample bead positions are entered into an in-house written Java-based application available on the web (*see* Note 10). The steps for geometry file creation are summarized in Fig. 2B.
6. The resulting geometry file is placed in the FlexControl software's geometry files root folder and can be easily found and automatically loaded by the software.
7. The sample is loaded into the mass spectrometer and mass calibration is performed using peptide standards.
8. An AutoXecute sequence is created using the geometry file, specifying an appropriate maximum laser intensity, a value of 100 laser shots per spot, and a 50 Hz repetition rate before starting the MS imaging run.

⁷The success rate of the bead position identification depends on the quality of the optical images. Transmission-mode images are easier to threshold for bead positions as they appear brighter than the background Parafilm M.

⁸As the FlexControl "MTP Slide Adapter II" geometry file uses a fractional distance coordinate system where the distance between each point in the regular array is separated by exactly 0.086957 units, a new geometry file can be created in this coordinate system to acquire data at the bead positions.

⁹Variations in the distances between calibration points that are larger than bead diameter signify inaccuracies in the optical image of the stretched sample such that the resulting geometry file may not accurately represent the bead positions. Inaccuracies sometimes occur because of errors in stitching of images by Photoshop and can be prevented by taking well-focused images with sufficient spatial overlap.

¹⁰All in-house written Java software is available online, along with an example dataset with step-by-step instructions, at <http://neuroproteomics.scs.uiuc.edu/imaging.html>.

9. The region of interest is imaged using MALDI-TOF MS.

3.5 Data Conversion

1. The data must be converted from the Bruker fid file format to the more general mzXML format. The CompassXport software is run along with the `-multiName` tag at an MS-DOS prompt to create a file called `new.mzXML` within each spectrum directory.
2. MATLAB is used along with the bioinformatics toolbox and the batch conversion wrapper code available online (*see* Note 10) to convert mzXML files into spectra-containing text files. During this step, the data may be processed by baseline subtraction and smoothing with the Bioinformatics Toolbox functions in MATLAB to eliminate noise and create more uniform ion images.

3.6 Image Reconstruction

1. The coordinates of the initial bead positions are found in the same manner as for the stretched sample image, as outlined above in step 3.4.3 (*see* Note 11).
2. A simple code (*see* Note 10) is compiled in Java and used to create a separate image that places small dots at the initial positions. These dots are easier to see and aid the free transform process. The computational steps for image reconstruction are summarized in Fig. 2C.
3. The small dots image and the stretched sample optical images are opened into Photoshop. The free transform command (Ctrl + T) is used to report the centroid coordinates of each of the two images from the options bar. A duplicate background layer is created, and the black background is removed from the small dots image to create a transparent image using the Magic Wand tool in Photoshop. A new blank Photoshop image file is created, large enough (in the range of 5000 pixels $\times$ 5000 pixels) to hold both images when placed side-by-side in separate layers, allowing adequate work space to manipulate the images when aligning on top of each other. The small dots image must be placed in a layer above the stretched image layer. The new centroid positions of each image in the blank image file are recorded from the options bar. The free transform command is used to translate, rotate, and resize the small dots image until it is appropriately aligned on top of the stretched sample image. The small dots should each align within one of the bead-torn regions of the stretched sample image.
4. Before applying the transformation to the transformed images in Photoshop, the final width, height, and rotation angle are recorded into a text file from the Info palette, along with the final centroid position of each image.
5. The text files of the initial positions and the image reconstruction parameters, recorded both before and after the free transformation, are inputted into another in-house Java program (*see* Note 10) to create reconstructed ion images at select m/z .

¹¹Alternatively, image reconstruction can be done before geometry file creation so that some time is saved in the rare event that image reconstruction is not successful, upon which the sample is discarded. Any difficulties with image reconstruction using the free transform approach arise from highly non-uniform stretching of the Parafilm M that can be prevented by visually adjusting for the shape of the marked perimeter of the sample while stretching. Overall, image reconstruction is fairly reproducible, as a set of six samples resulted in a classification rate of 84.1% for bead position matching between the stretched and initial samples, with the remaining portion being only near-neighbor mismatches (18).

ratios of interest as seen in Fig. 3 (*see* Note 12). The example shown in Fig. 3 is with a thin tissue section from mouse pituitary.

Acknowledgments

We thank Georgina M. Aldridge, University of Illinois at Urbana-Champaign, for providing the animals. The project described was supported by Award Nos. P30 DA018310 and 5RO1DA017940 from the National Institute On Drug Abuse and Award No. 5RO1DE018866 from the National Institute of Dental and Craniofacial Research (NIDCR) and the Office of Director (OD), National Institutes of Health (NIH). The content is solely the responsibility of the authors and does not necessarily represent the official views of the NIDA, NIDCR, or NIH.

References

1. Seeley EH, Caprioli RM. Molecular imaging of proteins in tissues by mass spectrometry. *Proc Natl Acad Sci U S A*. 2008; 105:18126–18131. [PubMed: 18776051]
2. McDonnell LA, Heeren RM. Imaging mass spectrometry. *Mass Spectrom Rev*. 2007; 26:606–643. [PubMed: 17471576]
3. Pacholski ML, Winograd N. Imaging with mass spectrometry. *Chem Rev*. 1999; 99:2977–3006. [PubMed: 11749508]
4. Rubakhin SS, Jurchen JC, Monroe EB, Sweedler JV. Imaging mass spectrometry: fundamentals and applications to drug discovery. *Drug Discov Today*. 2005; 10:823–837. [PubMed: 15970265]
5. Becker, JS. Inorganic mass spectrometry: principles and applications. John Wiley & Sons; Hoboken, NJ: 2007.
6. Lemaire R, Menguellet SA, Stauber J, Marchaudon V, Lucot J-P, Collinet P, Farine M-O, Vinatier D, Day R, Ducoroy P, et al. Specific MALDI imaging and profiling for biomarker hunting and validation: fragment of the 11S proteasome activator complex, reg alpha fragment, is a new potential ovary cancer biomarker. *J. Proteome Res*. 2007; 6:4127–4134. [PubMed: 17939699]
7. Chandra S, Tjarks W, Lorey DR, Barth RF. Quantitative subcellular imaging of boron compounds in individual mitotic and interphase human glioblastoma cells with imaging secondary ion mass spectrometry (SIMS). *J Microsc*. 2008; 229:92–103. [PubMed: 18173648]
8. Chaurand P, Norris JL, Cornett DS, Mobley JA, Caprioli RM. New developments in profiling and imaging of proteins from tissue sections by MALDI mass spectrometry. *J Proteome Res*. 2006; 5:2889–2900. [PubMed: 17081040]
9. Chaurand P, Rahman MA, Hunt T, Mobley JA, Gu G, Latham JC, Caprioli RM, Kasper S. Monitoring mouse prostate development by profiling and imaging mass spectrometry. *Mol Cell Proteomics*. 2008; 7:411–423. [PubMed: 17991918]
10. Hsieh Y, Chen J, Korfmacher WA. Mapping pharmaceuticals in tissues using MALDI imaging mass spectrometry. *Pharmacol Toxicol Methods*. 2007; 55:193–200.
11. Rubakhin SS, Hatcher NG, Monroe EB, Heien ML, Sweedler JV. Mass spectrometric imaging of the nervous system. *Curr Pharma Design*. 2007; 13:3325–3334.
12. Chaurand P, Norris JL, Cornett DS, Mobley JA, Caprioli RM. New developments in profiling and imaging of proteins from tissue sections by MALDI mass spectrometry. *J Proteome Res*. 2006; 5:2889–2900. [PubMed: 17081040]
13. Kruse R, Sweedler JV. Spatial profiling invertebrate ganglia using MALDI MS. *Am Soc Mass Spectrom*. 2003; 14:752–759.
14. Aerni HR, Cornett DS, Caprioli RM. Automated acoustic matrix deposition for MALDI sample preparation. *Anal Chem*. 2006; 78:827–834. [PubMed: 16448057]
15. Monroe EB, Jurchen JC, Koszczuk BA, Losh JL, Rubakhin SS, Sweedler JV. Massively parallel sample preparation for the MALDI MS analyses of tissues. *Anal Chem*. 2006; 78:6826–6832. [PubMed: 17007502]

¹²The success of image reconstruction can be verified using an in-house written code (<http://neuroproteomics.scs.uiuc.edu/imaging.html>) that plots an image of the calculated transformed initial positions. This image can be checked against the transformed image in Photoshop to verify for any calculation or positional errors.

16. Clerens S, Ceuppens R, Arckens L. Createtarget and analyze this!: new software assisting imaging mass spectrometry on Bruker Reflex IV and Ultraflex II instruments. *Rapid Commun Mass Spectrom*. 2006; 20:3061–3066. [PubMed: 16969765]
17. Decker JD. Image editing. *Am J Orthod Dentofacial Orthop*. 2004; 125:215–219. [PubMed: 14765061]
18. Zimmerman TA, Monroe EB, Sweedler JV. Adapting the stretched sample method from tissue profiling to imaging. *Proteomics*. 2008; 8:3809–3815. [PubMed: 18712762]
19. Wang J, Chen R, Ma M, Li L. MALDI MS sample preparation by using paraffin wax film: systematic study and application for peptide analysis. *Anal Chem*. 2008; 80:491–500. [PubMed: 18189446]
20. Monroe EB, Koszczuk BA, Losh JL, Jurchen JC, Sweedler JV. Measuring salty samples without adducts with MALDI MS. *Int J Mass Spectrom*. 2007; 260:237–242.

Figure 1.

Manual stretching of the sample. (A) An apparatus holds the sample, so that both hands can be used during stretching; here this device consists of the glass slide (arrow) taped to the thin side of the slide box cover, laterally stabilized by heavy objects. (B) The bead substrate, with the location and orientation of the tissue marked, is manually stretched along one axis, (C) rotated 90 degrees and stretched, and (D) rotated and stretched again. (E) The stretched membrane is placed so that the marked sample area is on the glass slide. Use of thumbs provides a last bit of stretching before the sample is applied to the surface. (F) The arrow marking the orientation of the sample is clearly visible after stretching and can be marked again on the area of the glass slide if necessary. (G) The excess Parafilm M is removed from the edges of the glass. (H) A finished slide is ready for MALDI matrix coating.

Figure 2.

Flow chart for the computational steps in the stretched imaging method, including the process of (A) bead identification by light thresholding, (B) geometry file creation, and (C) image reconstruction. The custom software routines can be found online at <http://neuroproteomics.scs.uiuc.edu/imaging.html>.

Figure 3.

Reconstructed MALDI-MSI ion images from a 10- μ m section of mouse pituitary prepared with the stretched sample method. (A) Optical photomicrograph of the pituitary section showing (I) the posterior lobe, (II) a darker band corresponding to the intermediate lobe, and (III) the anterior lobe. Only the outlined region tissue was imaged. Reconstructed ion images correspond to the outlined tissue area showing: (B) oxytocin, 1007 m/z ; (C) di-Ac- α -MSH, 1707 m/z ; (D) vasopressin, 1083 m/z ; (E) POMC J-peptide, 1883 m/z ; and (F) Arg-CLIP [1–22], 2505 m/z . The intermediate lobe is a small band that is highlighted by signals from the di-acetylated- α -MSH and J-peptide ion images.