Published in final edited form as:

J Am Chem Soc. 2010 December 15; 132(49): 17408–17410. doi:10.1021/ja109042w.

A DUAL PLATFORM FOR SELECTIVE ANALYTE ENRICHMENT AND IONIZATION IN MASS SPECTROMETRY USING APTAMER-CONJUGATED GRAPHENE OXIDE

Basri Gulbakan¹, Emir Yasun¹, M. Ibrahim Shukoor¹, Zhi Zhu¹, Mingxu You¹, Xiaohong Tan¹, Hernan Sanchez³, David H. Powell¹, Hongjie Dai³, and Weihong Tan^{*,1}

¹Center for Research at Bio/nano Interface, Department of Chemistry and Department of Physiology and Functional Genomics, Shands Cancer Center, University of Florida, Gainesville, Florida 32611-7200 (USA)

²State Key Laboratory of Chemo/Biosensing and Chemometrics, College of Biology, College of Chemistry and Chemical Engineering, Hunan University, Changsha 410082 (P.R. China)

³Department of Chemistry and Laboratory for Advanced Materials, Stanford University, Stanford, CA, USA

Abstract

This study demonstrates the use of aptamer-conjugated graphene oxide as an affinity extraction and detection platform for analytes from complex biological media. We have shown that cocaine and adenosine can be selectively enriched from plasma samples and that direct mass spectrometric readout can be obtained without a matrix and with greatly improved signal-to-noise ratios. The aptamer conjugated graphene oxide has clear advantages in target enrichment and in generating highly efficient ionization of target molecules for mass spectrometry. These results demonstrate the utility of the approach for analysis of small molecules in real biological samples.

Analysis of biologically important small molecules and their subsequent detection by mass spectrometry (MS) in complex biological matrices (*e.g.*, serum, plasma) is an area of great interest. Despite the great potential of MS based on its high sensitivity, speed, reproducibility and label-free readout, signal suppression effects and the need for careful sample preparation still limits its overall use.^{1,2} To address these problems, chromatographic extraction and fractionation methods are often applied before MS is performed.³ However, since these techniques lack specificity, the next best option utilizes affinity reagents tethered on a substrate from which direct MS readout can be obtained. To achieve this goal, different kinds of affinity reagents have so far been applied in different formats, such as mass spectrometric immunoassay (MSIA), nanoprobe affinity mass spectrometry and surface-enhanced laser desorption/ionization (SELDI).⁴

As a novel affinity reagent, aptamers could provide the specificity lacking in many extraction matrixes. Aptamers are single-stranded oligonucleotides that bind target molecules with very high affinity in a manner similar to antibodies. However, as a capturing reagent, the affinity of aptamers can be adjusted, depending on the application. Aptamers can be generated against a variety of targets, including metal ions, metabolites; proteins and even whole cells.⁵ Aptamers have very distinct advantages as capturing reagents, such as

tan@chem.ufl.edu.

Supporting Information Available: Experimental details, additional MS analyses, FTIR and TEM images. This information is available free of charge via the Internet at <http://pubs.acs.org/>.

small size, non-toxicity, easy modification and easy surface immobilization. Moreover, aptamers can be produced without the need of an animal source, and they can be chemically modified with various functional groups. All of these unique properties increase the likelihood that aptamers will outperform other affinity reagents. 6

To date, very few studies have been carried out on developing a single platform whose function is based on simultaneous capture and ionization.⁷ This particular area still requires new high-efficiency techniques and new materials. Recently, graphene oxide (GO) has attracted interest as a substrate for analyte detection based on its unique electronic, thermal and mechanical properties.⁸ Herein, we report aptamer-modified GO as a selective enrichment and matrix-free detection platform for MS detection of cocaine and adenosine from complex biological systems.

Synthesis of graphene oxide has been reported elsewhere.⁹ Briefly, GO was obtained by oxidizing graphite using Hummer's method, which results in water soluble GO having a carboxyl-rich structure. It has been previously demonstrated that graphene (G) can be used as a matrix for laser desorption ionization.^{10, 11} To demonstrate that GO has very similar properties, a series of small molecules were analyzed using GO as an efficient energy-absorbing molecule (Figs. S2–S5). Next, we immobilized thiol-functionalized cocaine and adenosine aptamers onto GO by activating carboxyl (-COOH)-rich groups with the use of EDC(1-Ethyl-3-(3-dimethylaminopropyl)carbodiimide)/NHS (N-hydroxysuccinimide) / NHS chemistry and by introducing a bifunctional PEG(SH-PEG-NH₂) as a spacer molecule, which provides stability in physiological media, such as serum or blood.¹¹ Since aptamers intrinsically possess a secondary structure, the PEG linker enables them to attain their 3-D conformation, thus permitting target recognition. The bifunctional PEG linker carries amine (-NH₂) groups on one side to bind to the carboxyl on the GO surface and thiol (SH) groups to help anchor the SH-functionalized aptamers through disulfide bond formation (Fig.1). Aptamer-modified GO was characterized using FTIR (Fig. S6), and then aptamer-conjugated GO was applied for selective enrichment and detection of cocaine spiked in human plasma, along with respective control experiments.

Figure 2–A shows the MS analysis of cocaine-spiked plasma samples analyzed directly on GO without any enrichment step (no aptamer modification) where GO served only as a matrix. The cocaine peaks can be detected among huge background ions with an average S/N ratio of 15. Figure 2B shows the analysis of cocaine-spiked plasma extracted and ionized with unmodified GO. Even in the presence of unmodified GO, enrichment of the sample was, to some extent, observed, along with a reduction in background ions. This result can be explained by the structure of cocaine bearing an aromatic ring which aids in π - π interactions with GO. In another control experiment, GO with physically adsorbed cocaine aptamer was used for target extraction. In this case, following washing steps, no, or scant, extraction of cocaine with an S/N ratio of 5 was obtained. On the contrary, GO was shown to physically adsorb DNA in high yields.⁹ To understand this phenomenon, we assume that (i) surface coverage of GO by physically adsorbed aptamer prevents the loading of cocaine by means of nonspecific interactions (as is the case with π - π interactions, Fig. 2B) and (ii) upon target capture, aptamer is released from the GO surface following washing steps by the absence of any chemical functionalization. This result is very consistent with the previous results of GO-based fluorescence sensors. In these biosensor experiments, DNA (either cDNA or aptamer) leaves the surface upon target addition, and this leads to restoration of fluorescence, which is previously quenched by GO.¹² In our experiments, we use MS for detection, and the signal is generated from the GO surface. Consequently, we also see a decrease in signal upon washing steps. However, when chemical conjugation was applied to covalently modify GO with cocaine aptamer, an efficient capture of target analyte with an average S/N ratio of 52 was achieved (Figure 2 D). Since we see an increase in signal to

noise ratios, this result further proves our previous experiments, and indicates that chemical conjugation of the aptamer is necessary for analyte capturing. (Schematic representation of the whole series of experiments is given in Figure S1.) To further demonstrate the analytical utility of our approach, we applied the same concept to an adenosine aptamer.

As shown in Figure 3 top, a trend similar to that of cocaine can be clearly observed. Even if some extraction can be achieved with pristine GO, this result proves that aptamer conjugation yields significantly improved extraction efficiency and significant S/N increase as shown in Figure 3 bottom.

In conclusion, we have prepared aptamer-conjugated GO for selective enrichment and detection of cocaine and adenosine from plasma samples. This method has several advantages: first, it represents a conjugation chemistry that is very easy to perform; second, the use of GO eliminates 1) the need for any additional energy-absorbing matrix for ionization and 2) background interference, which is the biggest problem when conventional MALDI matrices are employed. When combined with aptamer-based affinity capture, our results show that GO provides an efficient platform for selective enrichment of target analytes and the attainment of direct mass spectrometric readout, even from very complex media, which would not be possible by using either GO or aptamers alone. Since a huge repertoire of aptamers exists for different targets, including proteins, metabolites, and cell-surface markers, we believe, based on our proof of concept study that graphene based aptamer enhanced extraction mass spectrometry can be extended to a more complex system and be used when analyzing different biological samples.

Supplementary Material

Refer to Web version on PubMed Central for supplementary material.

Acknowledgments

This work was supported by NIH grants.

References

1. a) Want EJ, Nordstrom A, Morita H, Siuzdak G. J. Proteome Res. 2007; 6:459–468. [PubMed: 17269703] Nordstrom A, Want E, Northen T, Lehtio J, Siuzdak G. Anal. Chem. 2008; 80:421–429. [PubMed: 18085752]
2. a) Chipuk JE, Gelb MH, Brodbelt JS. Anal. Chem. 2010; 82:4130–4139. [PubMed: 20402469] b) Northen TR, Lee J, Hoang L, Raymond J, Hwang D, Yannone SM, Wong C, Siuzdak G. Proc. Natl. Acad. Sci. U.S.A. 2008; 105:3678–3683. [PubMed: 18319341]
3. a) Zhang W, Zhou G, Zhao Y, White MA, Zhao Y. Electrophoresis. 2003; 24:2855–2863. [PubMed: 12929181] b) Bauer A, Kuster B. Eur. Jour. of Biochem. 2003; 270:570–578. [PubMed: 12581197] c) Pernemalm M, Lewensohn R, Lehtio J. Proteomics. 2009; 9:1420–1427. [PubMed: 19235168]
4. a) Hutchens TW, Yip T. Rapid Commun. Mass Spectrom. 1993; 7:576–580. b) Chou P, Chen S, Liao H, Lin P, Her G, Lai AC, Chen J, Lin C, Chen Y. Anal. Chem. 2005; 77:5990–5997. [PubMed: 16159132] c) Lopez MF, Rezai T, Sarracino DA, Prakash A, Krastins B, Athanas M, Singh RJ, Barnidge DR, Oran P, Borges C, Nelson RW. Clin. Chem. 2010; 56:281–290. [PubMed: 20022981]
5. a) Famulok M, Hartig J, Mayer G. Chem. Rev. 2007; 107:3715–3743. [PubMed: 17715981] b) Fang X, Tan W. Acc. Chem. Res. 2010; 43:48–57. [PubMed: 19751057] c) Shangguan D, Li Y, Tang Z, Cao ZC, Chen HW, Mallikaratchy P, Sefah K, Yang CJ, Tan W. Proc. Natl. Acad. Sci. U.S.A. 2006; 103:11838–11843. [PubMed: 16873550]
6. a) Sefah K, Phillips JA, Xiong X, Meng L, Simaey DV, Chen H, Martin J, Tan W. Analyst. 2009; 134:1765–1775. [PubMed: 19684896] b) Meir, A.; Marks, RS.; Stojanovic, MN. John Wiley & Sons, Ltd.;

7. a) Huang Y, Chang H. *Anal. Chem.* 2007; 79:4852–4859. [PubMed: 17523592] b) Lowe RD, Szili EJ, Kirkbride P, Thissen H, Siuzdak G, Voelcker NH. *Anal. Chem.* 2010; 82:4201–4208. [PubMed: 20411919]
8. a) Allen MJ, Tung VC, Kaner RB. *Chem. Rev.* 2010; 110:132–145. [PubMed: 19610631] b) Rao CNR, Sood AK, Voggu R, Subrahmanyam KS. *J. Phys. Chem. Let.* 2010; 1:572–580. (c) He QY, et al. *ACS Nano.* 2010; 4:3201–3208. [PubMed: 20441213] (d) Wang ZJ, et al. *J. Phys. Chem. C.* 2010; 113:14071–14075.
9. a) Sun X, Liu Z, Welsher K, Robinson J, Goodwin A, Zaric S, Dai H. *Nano Res.* 2008; 1:203–212. [PubMed: 20216934] b) Wang H, Wang X, Li X, Dai H. *Nano Res.* 2009; 2:336–342. (c) Zhou XZ, et al. *J. Phys. Chem. C.* 2009; 113:10842–10846.
10. a) Jung J, Cheon D, Liu F, Lee K, Seo T. *Angew. Chem. Int. Ed.* 2010; 49:5708–5711. b) a) Hsu W, Lin W, Hwu W, Lai C, Tsai F. *Anal. Chem.* 2010; 82:6814–6820. [PubMed: 20704371]
11. a) Zhou X, Wei Y, He Q, Boey F, Zhang Q, Zhang H. *Chem. Commun.* 2010; 46:6974–6976. b) Tang LAL, Wang J, Loh KP. *J. Am. Chem. Soc.* 2010; 132:10976–10977. [PubMed: 20698647]
12. a) Lu C, Yang H, Zhu C, Chen X, Chen G. *Angew. Chem., Int. Ed.* 2009; 48:4785–4787. b) Wang Y, Li Z, Hu D, Lin C, Li J, Lin Y. *J. Am. Chem. Soc.* 2010; 132:9274–9276. [PubMed: 20565095]

Figure 1.
Scheme for aptamer modification and GO-assisted target capture and analysis

Figure 2.

Analysis of cocaine a) spiked into plasma, b) extracted with GO, c) extracted with noncovalently aptamer-modified GO (cocaine is washed away without chemical modification before MS) and d) extracted with covalently aptamer-modified GO.

Figure 3.

Top. Analysis of adenosine a) spiked into plasma, b) extracted with GO, c) extracted with noncovalently aptamer-modified GO (adenosine is washed away without chemical modification before MS), and d) extracted with covalently aptamer-modified GO.

Bottom: S/N ratios of adenosine 1) spiked into plasma, 2) extracted with GO, 3) extracted with noncovalently aptamer-modified GO, and 4) extracted with covalently aptamer-modified GO.