Conventional diathermy in arthroscopic anterior cruciate ligament reconstruction – a safe and cost-effective alternative to commercial thermal devices

CP CHARALAMBOUS, F ALVI, P SUTTON
Northern General Hospital, Sheffield, UK

CORRESPONDENCE TO

Farhan Alvi, Department of Orthopaedics and Trauma,
Manchester Royal Infirmary, Oxford Rd, Manchester M13
9WL, UK. E: alvi@btinternet.com

During arthroscopic anterior cruciate ligament (ACL) reconstruction, removal of soft tissue is essential in clearing the remnant of the ruptured ACL and defining the posterolateral wall of the femoral notch. Removing soft tissue using diathermy devices as compared to a motorised shaver minimises bleeding. There are several commercially available, single-use, monopolar diathermy systems, the price of which varies from tens to hundreds of pounds.

Figure 1 The 5-mm ball-end monopolar diathermy probe

We use a 5-mm ball-end monopolar diathermy probe (Trewais Surgical Instruments Pty Ltd, Victoria, Australia) for soft tissue removal (Fig. 1). This is run on a conventional diathermy source, using the coagulation mode set at maximum. This probe is reusable, costing about £40. Due to its curvature and low profile, this probe allows precise soft tissue removal. When using this device, sterile water rather than normal saline is used for irrigation. We have used the above probe in hundreds of ACL reconstructions without any adverse effects.

A simple technique for exchanging the liner in ankle replacements

C TAYLOR, T BALL, J DAVIS
Department of Trauma and Orthopaedic Surgery, Plymouth
Hospitals NHS Trust, Plymouth, UK

CORRESPONDENCE TO

Clare Taylor, Department of Trauma and Orthopaedic Surgery,
Plymouth Hospitals NHS Trust, Derriford Road, Crownhill,
Plymouth PL6 8DH, UK. E: cft199@doctors.org.uk

Exchange of the liner in Mobility Ankle replacements (Depuy Orthopedics Inc., Warsaw, IN 46582, USA) for infection, wear, or

to change the size can be troublesome. Getting purchase on the liner is difficult, it is under soft tissue tension, and semi-constrained due to the contouring of the liner. The use of metal instruments to lever it out can damage the articular surfaces of the implant. We insert a 3.5 mm small fragment screw into the anterior aspect of the liner (Fig. 1). The screw head is gripped with an artery clip or pliers to facilitate removal of the liner without trauma to the joint surfaces.

Figure 1 Representation of the technique.

Safely dealing with the right hepatic vein during a right hepectomy

ATUL BAGUL, GREG MCMAHON, FAHREYAR ALAM,
ASHLEY R DENNISON
Hepato-biliary Unit, University Hospitals of Leicester,
Leicester, UK

CORRESPONDENCE TO

Atul Bagul, Department of General Surgery, Leicester General
Hospital, Leicester LE5 4PW, UK. E: atulbagul@hotmail.com

Figure 1 'T'-tube placed around the vein and fixing the stapler