Published in final edited form as:

Opt Lett. 2012 February 15; 37(4): 659–661.

Double-illumination photoacoustic microscopy

Junjie Yao^a, Konstantin I. Maslov^a, Ernest R. Puckett^a, Kathryn J. Rowland^b, Brad W. Warner^b, and Lihong V. Wang^{a,*}

^aOptical Imaging Laboratory, Department of Biomedical Engineering, Washington University in St. Louis, St. Louis, MO 63130, USA

^bDivision of Pediatric Surgery, St. Louis Children's Hospital, Department of Surgery, Washington University School of Medicine, St. Louis, MO 63110, USA

Abstract

Recent developments of optical-resolution photoacoustic microscopy (OR-PAM) have improved its spatial resolution and imaging speed. However, the penetration depth of OR-PAM is still limited to ~1 mm in tissue due to the strong tissue scattering. Here, we have developed double-illumination PAM (DI-PAM), which illuminates the sample from both top and bottom sides simultaneously. Through phantom and *in vivo* experiments, we have demonstrated for thin targets that DI-PAM has a penetration depth of ~2 mm in tissue at 532 nm and a focal zone of 260 μ m, both significant improvements over traditional reflection or transmission mode OR-PAM.

After several years of rapid development, photoacoustic microscopy (PAM) is now playing an increasingly important role in biomedical research and preclinical studies [1–6]. Taking advantage of rich optical absorption contrast and weak ultrasonic scattering in tissue, PAM has proved to be capable of anatomical, functional and molecular imaging, with highly scalable spatial resolution, penetration depth and imaging speed [3, 7–9]. Based on the confocal configuration of the optical illumination and acoustic detection, PAM can be classified into two types: acoustic-resolution PAM (AR-PAM) and optical-resolution PAM (OR-PAM). While AR-PAM has achieved so far tens of micrometers resolution with a penetration depth up to 3 mm in tissue [10], OR-PAM focuses on superficial imaging (up to ~1 mm in tissue) with at least capillary level resolution [2]. Recently, the imaging performance of OR-PAM has been significantly improved in terms of spatial resolution and imaging speed. By using a water-immersion optical objective with a 1.23 numerical aperture (NA), a 220 nm lateral resolution at 532 nm wavelength has been achieved [11]. Meanwhile, by using a fast voice-coil scanner, the sectional imaging speed of OR-PAM has been pushed to 40 Hz over a 1 mm scanning range [12]. Despite the above improvements, the penetration depth of OR-PAM is still restricted to one transport mean free path due to the strong optical scattering in tissue. Moreover, the strong absorption of hemoglobin in the visible spectral range further reduces the 1/e penetration depth to less than 50 μ m in blood vessels, which may cause shallower vessels to shadow deeper vessels. This penetration limitation applies to both reflection-mode and transmission-mode OR-PAM. Only a layer of tissue on the optical illumination can be imaged [3–5, 11]. By illuminating light from both the top and bottom sides of the sample, we have developed a double-illumination PAM (DI-PAM) to improve the penetration depth to ~2 mm in thin biological tissue. At the same time, the optical focal zone is expanded to ~260 μ m.

In DI-PAM (Figure 1), a Nd:YVO₄ laser (Elforlight, SPOT) generates 1.5 ns pulses at a 532 nm wavelength. The pulses are reshaped by an iris (ID25SS, Thorlabs) and attenuated by a

*Corresponding author: lhwang@biomed.wustl.edu.

neutral density filter (NDC-50C-2M, Thorlabs). The attenuated beam is then split into two sub-beams (top and bottom) by a 50/50 beam splitter (BSW04, Thorlabs). The top beam is focused by a condenser lens (LA1131, Thorlabs) before passing through a 50 μm pinhole (P50C, Thorlabs) for further spatial filtering. The filtered beam is then focused by an optical objective (AC127-050-A, Thorlabs, NA: 0.1 in air) into the sample from the top. A beam combiner composed of a thin layer of silicone oil sandwiched by a right-angle prism (NT32-545, EdmundOptics) and a rhomboid prism (NT49-419, EdmundOptics) provides acoustic-optical coaxial alignment. The resultant photoacoustic waves are detected by an ultrasonic transducer (V214-BB-RM, Olympus-NDT) with a central frequency of 50 MHz. An acoustic lens with an NA of 0.5 is ground into the bottom of the rhomboid prism to provide an acoustic focal diameter of 30 μm . An optical correction lens is attached to the top of the beam combiner to correct the aberration. The acoustic lens is submerged in a water-tank for ultrasound coupling. The bottom beam duplicates the same path as the top beam, except that it is directly focused into the sample from the bottom without going through the beam combiner and water tank. As shown in the inset of Figure 1, by carefully adjusting the positions of two objectives, we achieved a co-axial configuration of the two optical foci and the acoustic focus. To best expand the optical focal zone, the top and bottom optical foci are both approximately one Rayleigh range apart from the acoustic focus.

Volumetric imaging is acquired by two-dimensional raster scanning of the sample. While the lateral resolution is defined by the optical focus, the depth resolution is determined by the temporal resolution of the acoustic detection. An eyepiece is added to the bottom beam path to view the imaging region through the reverse path of the illumination. A photodiode is also added to monitor the fluctuations of the laser pulse intensity.

To measure the lateral resolution of the DI-PAM system, a sharp ink edge was imaged in water at varied depths while the optical focus was fixed. From the edge spread function, the full width half maximum of the line spread function was computed to determine the lateral resolution. As a comparison, the lateral resolution under either top or bottom illumination was also quantified using the same protocol. The measured results were fitted using the

theoretical model for a Gaussian beam $w(z) = \sqrt{w_0^2 + \left[\frac{\lambda(z - z_0)}{\pi w_0} \right]^2}$, where $w(z)$ is the lateral resolution at depth z , w_0 is the focal diameter, λ is the wavelength and z_0 is the focal depth. While λ is known, w_0 and z_0 are the unknown parameters to fit for. As shown in Figure 2(a), the top illumination gives a focal diameter of 2.3 μm with a focal zone of 100 μm . Here, focal zone is defined as the range of depths over which the system maintains a lateral resolution no worse than $\sqrt{2}w_0$. Similarly, the bottom illumination gives a focal diameter of 2.9 μm , with a focal zone of 170 μm . The two foci are 160 μm apart. The top illumination has a tighter focusing due to the addition of the correction lens which gives an effective NA of 0.133 in water, while the bottom illumination has an NA of 0.1 in air. As can be seen in Figure 2(b), under double illumination, the lateral resolution has a 'W' shape distribution along the depth, which is determined by the overall light intensity of the two beams. The two valleys of the 'W' distribution are 2.3 μm and 2.9 μm in lateral resolution, which equal to the top and bottom focal diameters, respectively. The effective focal zone is expanded to 260 μm , which is the sum of the focal zones of the two valleys (100 μm and 160 μm , respectively). Thus, by increasing the focal zone by 160 μm over the top illumination or 90 μm over the bottom illumination, DI-PAM maintains the lateral resolution over a longer depth range.

To measure the penetration depth of the DI-PAM system, a black human hair was obliquely embedded in a 2-mm-thick piece of fresh chicken breast tissue. The two ends of the hair tightly pressed against the top and bottom surfaces of the tissue, and served as landmarks of

the tissue boundaries. The top and bottom illuminations both had a pulse energy of 80 nJ. As shown in Figure 2(c), while either single illumination can image only the top (1.16 mm) or bottom (1.25 mm) part of the hair, DI-PAM can clearly image the whole hair [Figure 2(d)], demonstrating a penetration depth of at least ~2 mm in biological tissue.

The improved system performance enables DI-PAM to image deeper vasculature while maintaining capillary level lateral resolution. The mouse ear was chosen for a first demonstration. The ear of a 12-week-old nude mouse (Hsd:Athymic Nude-Foxn1NU, Harlan) is composed of a layer of auricular cartilage sandwiched by two layers of skin tissue [Figure 3(d)]. The ear of our experimental mouse became much thicker seven days after the injection of $\sim 10^6$ U87 glioblastoma cells [Figure 3(c)]. A 10×15 mm² area including the tumor region was sequentially imaged under top, bottom and double illuminations. In Figure 3(a), the depth-encoded maximum amplitude projection (MAP) images show that only the top layer vessels (shown in green) or bottom layer vessels (shown in red) can be well imaged under top or bottom illumination, respectively. Optical shadowing is the key reason. Although some big vessels from the other layer are discernable, they are very much blurred due to the degraded lateral resolution. In contrast, DI-PAM can clearly image both layers of vasculature without shadowing or blurring. The side-view MAP images in Figure 3(b) further demonstrate the improvements of DI-PAM in penetration depth and focal zone, imaging through the whole ear with a thickness more than 1 mm.

In a second demonstration of DI-PAM, we imaged the vascular network surrounding the small intestine of a living black mouse (C57BL/6Hsd, Harlan), which is of interest in embryonic development and sepsis studies [13, 14]. After surgical exposure, part of the small intestine was gently pulled out and placed on the sample holder for DI-PAM imaging. The high vessel density on one side of the intestine would prevent imaging the other side by single-illumination PAM systems [Figures 4(e–f)]. In contrast, DI-PAM can obtain complete cross-sectional images of the intestine that is close to 2 mm in diameter [Figures 4(a–b)]. The depth-encoded MAP image shows a dense mesenteric vascular network, where the vessels from the top side were colored in green and those from bottom side in red [Figure 4(a)]. The side-view MAP image of the section indicated by the dashed box in Figure 4(a) appears to be a closed loop [Figure 4(b)]. The villi on the lumen of the intestine are also clearly resolved [Figure 4(c)].

Simply doubling the incident light intensity for single illumination OR-PAM can also increase the penetration depth. However, the resultant improvement is not equivalent to that provided by DI-PAM. First, doubling incident light will not result in doubling of the penetration depth, since the attenuation of light by tissue is exponential rather than linear. Second, in DI-PAM, the light intensity on the tissue surface was 22 mJ/cm² for both top and bottom illuminations, which is still close to the American National Standards Institute (ANSI) safety limit (20 mJ/cm² in the visible spectral region). However, simply doubling the light intensity for single illumination (44 mJ/cm²) will break the ANSI limit and thus increase the risk of damaging tissue.

The focal zone of single illumination OR-PAM can also be expanded by depth scanning and stacking the data sets together. However, this is not equivalent to the improvement due to DI-PAM. First, strong tissue scattering would degrade the lateral resolution as the focal position deepens. Second, the imaging time would be increased by depth scanning. In contrast, DI-PAM expands the focal zone without sacrificing lateral resolution and imaging speed.

In conclusion, we have developed double-illumination PAM, which offers improved penetration depth and expanded focal zone. Compared with traditional reflection or

transmission mode OR-PAM, the penetration depth has been improved to at least 2 mm in biological tissue. In addition, the focal zone has been improved to 260 μm , with a lateral resolution no worse than 4 μm . Deep microscopy of microvascular anatomy in a mouse ear and small intestine was performed. Although not demonstrated here, DI-PAM is intrinsically ready for label-free measurements of oxygen saturation, blood flow, and oxygen metabolism.

Acknowledgments

The authors appreciate Prof. James Ballard's close reading of the manuscript, and thank Yan Liu and Joon-Mo Yang for useful discussion and technical assistance. This work was sponsored by National Institutes of Health (NIH) grants R01 EB000712, R01 EB008085, R01 CA134539, U54 CA136398, R01 CA157277, R01 CA159959 and 5P60 DK02057933. Lihong Wang has a financial interest in Microphotoacoustics, Inc. and Endra, Inc., which, however, did not support this work.

References

1. Wang LV. Multiscale photoacoustic microscopy and computed tomography. *Nature Photonics*. 2009; 3(9):503–509. [PubMed: 20161535]
2. Yao J, Wang LV. Photoacoustic tomography: fundamentals, advances and prospects. *Contrast Media Mol Imaging*. 2011; 6(5):332–345. [PubMed: 22025335]
3. Maslov K, et al. Optical-resolution photoacoustic microscopy for in vivo imaging of single capillaries. *Optics Letters*. 2008; 33(9):929–931. [PubMed: 18451942]
4. Hu S, Maslov K, Wang LV. Second-generation optical-resolution photoacoustic microscopy with improved sensitivity and speed. *Optics Letters*. 2011; 36(7):1134–1136. [PubMed: 21479007]
5. Xie ZX, et al. Laser-scanning optical-resolution photoacoustic microscopy. *Optics Letters*. 2009; 34(12):1771–1773. [PubMed: 19529698]
6. Hajireza P, Shi W, Zemp RJ. Label-free in vivo fiber-based optical-resolution photoacoustic microscopy. *Opt Lett*. 2011; 36(20):4107–4109. [PubMed: 22002401]
7. Yao J, et al. Evans blue dye-enhanced capillary-resolution photoacoustic microscopy in vivo. *Journal of Biomedical Optics*. 2009; 14(5):054049. [PubMed: 19895150]
8. Yao J, et al. In vivo photoacoustic imaging of transverse blood flow by using Doppler broadening of bandwidth. *Opt. Lett*. 2010; 35(9):1419–1421. [PubMed: 20436589]
9. Yao J, et al. Label-free oxygen-metabolic photoacoustic microscopy in vivo. *Journal of Biomedical Optics*. 2011; 16(7):076003. [PubMed: 21806264]
10. Zhang HF, et al. Functional photoacoustic microscopy for high-resolution and noninvasive in vivo imaging. *Nature Biotechnology*. 2006; 24(7):848–851.
11. Zhang K, Maslov K, Wang LV. Subwavelength-resolution label-free photoacoustic microscopy of optical absorption in vivo. *Opt Lett*. 2010; 35(19):3195–3197. [PubMed: 20890331]
12. Wang L, et al. Fast voice-coil scanning optical-resolution photoacoustic microscopy. *Opt Lett*. 2011; 36(2):139–141. [PubMed: 21263479]
13. Abtahian F, et al. Regulation of blood and lymphatic vascular separation by signaling proteins SLP-76 and Syk. *Science*. 2003; 299(5604):247–251. [PubMed: 12522250]
14. Whitworth PW, et al. Hypoperfusion of the Intestinal Microcirculation without Decreased Cardiac-Output during Live Escherichia-Coli Sepsis in Rats. *Circulatory Shock*. 1989; 27(2):111–122. [PubMed: 2650913]

Figure 1. Schematic of DI-PAM

The inset shows the coaxial configuration of two optical foci and the acoustic focus. The two optical foci are one Rayleigh range ($d_{1,2}$) apart from the acoustic focus. ND, neutral density filter; BS, beam-splitter; ConL, condenser lens; PH, pinhole; M, mirror; OL, objective lens; CorL, correction lens; RAP, right-angle prism; SOL, silicone oil layer; RhP, rhomboid prism; UT, ultrasonic transducer; WT, water tank; SH, sample holder; MS, motor scanner; EP, eyepiece; PD, photodiode.

Figure 2. System characterization of DI-PAM

(a) Lateral resolution as a function of the axial position relative to the starting depth of data-acquisition, under top and bottom illuminations, respectively. Solid curves in (a–b): theoretical fitting. (b) Lateral resolution as a function of the relative axial position under double illumination. (c) Side-view DI-PAM images of the human hair inside the chicken tissue after the compensation for the acoustic attenuation, under top, bottom and double illuminations, respectively. The two dashed lines indicate the top and bottom surfaces of the chicken tissue. Scale bar: 200 μm . (d) Volumetric rendering of the human hair imaged by DI-PAM.

Figure 3. DI-PAM of mouse ear *in vivo*

(a) Top-view depth-encoded DI-PAM images of the right ear of a nude mouse bearing a U87 glioblastoma tumor, under top, bottom and double illuminations, respectively. z is coded from green (superficial) to red (deep). TI, top illumination; BI, bottom illumination; DI, double-illumination. Scale bar: 500 μ m. (b) Side-view DI-PAM images under top, bottom and double illuminations, respectively. Only double-illumination can imaging through the whole ear. C_{Hb} , total hemoglobin concentration. Scale bar: 500 μ m. (c) Photograph of the mouse ear. The imaged area is indicated by the blue dashed box. Scale bar: 2 mm. (d) Histology image of H&E stained tissue slice across the mouse ear, showing an auricular cartilage (AC) layer sandwiched by two layers of skin tissue. Some of the blood vessels in the skin tissue are indicated by arrowheads. Scale bar: 100 μ m.

Figure 4. DI-PAM of the small intestine of a C57BL/6 mouse *in vivo*

(a) Top-view depth-encoded DI-PAM image of the small intestine under double illumination. z is coded from green (top side) to red (bottom side). Scale bar: 500 μm . (b) Side-view DI-PAM image of the region indicated by the dashed box in (a). C_{Hb} , total hemoglobin concentration. Scale bar: 250 μm . (c) Close-up of the small region indicated by the dashed box in (b). The intestinal villi are indicated by the arrowheads. Scale bar: 100 μm . (d) Volumetric rendering of the small intestine imaged by DI-PAM. (e) Photograph of the small intestine after surgical exposure. Scale bar: 1 mm. (f) Histology image of H&E stained cross-section of the small intestine, showing villi on the lumen surface. Some of the blood vessels in the submucosa layer are indicated by red arrowheads. Scale bar: 200 μm .