ANDROLOGY

Centrifugal Pulsing Increases the Recovery of Viable Spermatozoa

PETER BIELFELD,¹ VIDA ACOSTA,² and R. S. JEYENDRAN^{2,3}

Submitted: May 19, 1998

Accepted: June 29, 1998

Purpose: Our purpose was to compare the number and quality of spermatozoa recover following the standard centrifugation method with those after pulsing the relative centrifugal field during centrifugation.

Methods: A prospectively controlled in vitro study using ejaculates from 10 healthy sperm donors was undertaken. Number of sperm and sperm motility at 0, 3, and 24 hr post-incubation at 5°C in TEST-yolk results determined the efficacy of the centrifugation techniques.

Results: The number of sperm and sperm motility at 0, 3, and 24 hr post-incubation at 5°C in TEST-yolk were significantly higher for the sperm recovered following pulse centrifugation compared with those after standard centrifugation.

Conclusions: Pulsing the relative centrifugal field during centrifugation yields a higher number of viable spermatozoa.

KEY WORDS: pulse; centrifugation; viable; spermatozoa.

INTRODUCTION

Centrifugation of semen is essential to remove decapacitation factors, prostaglandin, and other deleterious substances present in the seminal plasma (1,2). Although the centrifugation of semen is a routine procedure, variations in the number and quality of sperm recovered following centrifugation are poorly understood. Recovery is influenced by the initial quality of

ejaculate and extent of dilution, in addition to the intensity of the relative centrifugal field (RCF) and length of centrifugation (3–6). Centrifugation may induce sublethal intracellular and ultrastructural damage to spermatozoa (6–8) and can also induce superoxide radical production (9). This leads to peroxidation of lipids in the sperm plasma membrane, further compromising spermatozoan quality (9,10).

Ideally, a centrifuge should increase the rate of sedimentation without increasing sublethal damage to spermatozoa. This report describes a new centrifugation method that involves pulsing the RCF. By using the pulsing technology during centrifugation for short periods, one may gradually and discontinuously exert an increased gravitational force without inducing latent injury to the spermatozoa.

MATERIALS AND METHODS

Ejaculates obtained by masturbation from 10 men were analyzed after liquefaction. The ejaculates were then diluted with 4 parts of Tyrodes's salt solution (Sigma Chemical Co., St. Louis, MO) and divided into two equal portions. One portion was centrifuged at 1000 RCF for 5 min in a Spinnet centrifuge (Damon/IEC Division, Needham Heights, MA). The other portion was centrifuged in the same centrifuge but the RCF was pulsed as follows: the sample was centrifuged at 1000 RCF for 2 min, followed by 1 min of rest. The process was repeated three more times, while centrifugation and the rest period lasted for 1 min each.

The supernatant from both portions was carefully removed and the sperm pellet resuspended in 1 ml of Tyrode's salt solution. After determining the sperm

¹ Department of Obstetrics and Gynecology, Heinrich Heine University, Dusseldorf, Germany.

² Andrology Laboratory Services, Chicago, Illinois.

³ To whom correspondence should be addressed at 680 North Lake Shore Drive, Suite 807, Chicago, Illinois 60611.

concentration and motility, an equal volume of TEST-yolk (Refrigeration Media; Irvine Scientific, Santa Ana, CA) was added to both portions and cooled to 5°C in a water jacket to stress the spermatozoa slowly. The sperm motility was again determined at room temperature following 1- and 24-hr incubations at 5°C.

Statistical Analyses

The mean $\pm$ standard error of the mean (SE) for all data was calculated. The log-transformed sperm concentration and arcsin-transformed percentage sperm motility data were subjected to paired *t*-test using the InStat program (Graph Pad, San Diego, CA) to assess the significant differences in means between the two centrifugation methods.

RESULTS

The ejaculate mean values (95% confidence interval) for sperm concentration and motile sperm concentration were $71.3 \times 10^6/\text{ml}$ (48.8–93.8) and $56.6 \times 10^6/\text{ml}$ (36.9–76.2), respectively. The number and quality of sperm recovered following the pulsing method of centrifugation were significantly higher ($P < 0.002$) than those of sperm recovered from standard centrifugation (Table I). Similar results were obtained following 1- and 24-hr post-incubation at 5°C in TEST-yolk.

DISCUSSION

Centrifugation is designed to increase the rate of sedimentation of particles in a medium. Owing to sensitivity, the sedimentation of spermatozoa requires a minimum RCF during centrifugation. The sperm pellet

must be firm enough to prevent liquid vortex disturbance but soft enough to minimize cell damage and loss of function. The detrimental influences on sperm quality mount when the RCF and the length of time of centrifugation are increased to boost recovery. In contrast, relatively low RCF results in poor recovery. Therefore, a compromise is continually sought to maximize the recovery of viable spermatozoa (5,6).

Also, centrifugation is the only practical technique for the concentration of spermatozoa from an ejaculate, and it is essential for any selection process, such as density gradient centrifugation, Sephadex, or glass-wool column separation. However, centrifugation of intact cells over some density gradient results in poor total recovery (11) but is the recommended procedure for sperm preparation (12), because no other comparable methods are available.

Compared with standard centrifugation, pulsing the RCF during centrifugation increased spermatozoan recovery (Table I). From the data presented, incubation of spermatozoa at 5°C did indeed reveal the deleterious effects of standard centrifugation, confirming previous reports (7,8,10). Therefore, stressing the sperm by incubation at 5°C may augment any subtle alteration or latent injury to sperm caused by centrifugation. A significant decrease in sperm motility was observed following standard centrifugation compared with pulsing centrifugation method.

Although an RCF of 600 or more is considered deleterious to spermatozoa (5,6), by pulsing the RCF for short periods, as demonstrated in the present study, even a higher RCF (1000 rcf) can be used. Because the pulsing centrifugation method yielded a significantly higher recovery of viable spermatozoa compared with standard centrifugation under identical conditions, pulsing the RCF during centrifugation will greatly facilitate any processes requiring sedimentation of viable spermatozoa.

Table I. Mean $\pm$ SEM Number and Quality of Sperm Recovered from 10 Ejaculates Following Standard and Pulsing Methods of Centrifugation and Following Incubation at 5°C in TEST-yolk

Sperm parameter	Centrifugation method		P value
	Standard	Pulsing	
Concentration $\times 10^6/\text{ml}$	35.3 ± 7.2	44.6 ± 8.2	0.002
Motility (%)	71.0 ± 3.7	80.7 ± 2.8	0.002
Motile concentration $\times 10^6/\text{ml}$	25.7 ± 5.8	36.9 ± 7.1	0.002
1 hr post-incubation at 5°C in TEST-yolk			
Motility (%)	57.4 ± 5.3	71.5 ± 5.0	0.002
24 hr post-incubation at 5°C in TEST-yolk			
Motility (%)	32.8 ± 3.1	47.7 ± 5.0	0.004

REFERENCES

1. Van der Ven H, Bhattacharyya A, Binor Z, *et al.*: Inhibition of human sperm capacitation by a high-molecular-weight factor from human seminal plasma. *Fertil Steril* 1982;38:753–755
2. Gutmann JN, Corson SL: Artificial insemination. *In* Gynecology and Obstetrics, JJ Sciarra (ed). Philadelphia, J. P. Lippincott, 1995; Vol 5, Chap 67
3. Makler A, Jakobi P: Effects of shaking and centrifugation on human sperm motility. *Arch Androl* 1981;7:21–26
4. Makler A, Murillo O, Huszar G, *et al.*: Improved technique for separating motile spermatozoa from human semen. II. An atraumatic centrifugation method. *Int J Androl* 1984;7:71–78
5. Jeulin C, Serres C, Jouannet P: The effects of centrifugation, various synthetic media and temperature on the motility and vitality of human spermatozoa. *Reprod Nutr Dev* 1982;22:81–91
6. Mack SR, Zaneveld LJD: Acrosomal enzymes and ultrastructure of unfrozen and cryotreated human spermatozoa. *Gamete Res* 1987;18:375–383
7. Overstreet JW, Drobnis EZ: Sperm transport in the female tract. *In* Advances in Donor Insemination, CLR Barratt, ID Cooke (eds). Cambridge, Cambridge University Press, 1993, pp 33–49
8. Alvarez JG, Lasso JL, Blasco L, *et al.*: Centrifugation of human spermatozoa induces sublethal damage; Separation of human spermatozoa from seminal plasma by a dextran swim-up procedure without centrifugation extends their motile lifetime. *Hum Reprod* 1993;8:1087–1092
9. Aitken RJ, Clarkson JS: Significance of reactive oxygen species and antioxidants in defining the efficacy of sperm preparation techniques. *J Androl* 1988;9:367–376
10. Shekariz M, DeWire DM, Thomas AJ, *et al.*: A method of human semen centrifugation to minimize the iatrogenic sperm injuries caused by reactive oxygen species. *Eur Urol* 1995;28:31–35
11. Ford TC, Graham JM (eds): An Introduction to Centrifugation. Oxford, Bios Scientific, 1991.
12. World Health Organization: WHO Laboratory Manual for the Examination of Human Semen and Semen-Cervical Mucus Interaction. Cambridge, Cambridge University Press, 1992