Published in final edited form as:

Opt Commun. 2012 October 1; 285(21-22): 4632–4637. doi:10.1016/j.optcom.2012.07.043.

Morphological characterization of cells in concentrated suspensions using multispectral diffuse optical tomography

Mohammad Reza Hajihashemi, Xiaoqi Li, and Huabei Jiang*

Department of Biomedical Engineering, University of Florida, Gainesville, FL

Abstract

Based on a non-spherical model of particle scattering, we investigate the capabilities and limitations of a T-matrix based inverse algorithm to morphologically characterize cells in concentrated suspensions. Here the cells are modeled as randomly orientated spheroidal particles with homogenous dielectric properties and suspended in turbid media. The inverse algorithm retrieves the geometrical parameters and the concentration of cells simultaneously by inverting the reduced scattering coefficient spectra obtained from multispectral diffuse optical tomography (MS-DOT). Both round and spheroidal cells are tested and the role of multiple and higher order scattering of particles on the performance of the algorithm is evaluated using different concentrations of cells.

1. Introduction

Morphological characterization of cells plays an important role in clinical diagnosis [1–6]. For example, tumorous cells are crowded and enlarged compared to normal cells [1]. Microscopy is routinely used to examine the cellular morphology, but it is not applicable to *in vivo* tissue. It has been shown that multispectral diffuse optical tomography (MS-DOT) is a promising modality for imaging of cellular morphology since it is non-invasive and can be used to *in vivo* characterize cells located deep in tissue [7]. The basic idea of MS-DOT is to illuminate the tissue by near-infrared light and collect the scattered light along the tissue boundary. By analyzing the spectra of scattered light, one can retrieve information regarding the morphology and composition of cells.

It is known that the MS-DOT or other optical based particle sizing methods are dealing with an ill-posed inverse problem that require some a priori information about the cells in order to obtain stable and unique solutions [7, 8]. For example, the spherical model of Mie scattering theory is widely used so far. However, cells generally have complex irregular shapes; thus the effect of non-sphericity cannot be adequately addressed by the Mie based techniques [9]. This necessitates a non-spherical model of particle scattering for accurate cellular sizing. The T-matrix or extended boundary conditions method is one of the most powerful and widely used tools for rigorous computation of electromagnetic scattering by single or an aggregate of non-spherical particles [10–11]. In the T-matrix approach, the incident and scattered waves are expanded in terms of incoming and outgoing vector spherical wave functions. The relationship between the coefficient of the incident and scattered waves is

© 2012 Elsevier B.V. All rights reserved.

*Corresponding author: hjiang@bme.ufl.edu.

Publisher's Disclaimer: This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our customers we are providing this early version of the manuscript. The manuscript will undergo copyediting, typesetting, and review of the resulting proof before it is published in its final citable form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

established through a transmission matrix (i.e., the T-matrix). The main advantage of T-matrix representation is that it is independent of the incident/scattered wave directions and that it depends only on the shape, size, orientation and refractive index of the particles [10].

The application of the T-matrix method to assess the morphology of spheroidal particles has been reported in several publications [12–14]. However, these studies did not demonstrate the ability to retrieve the concentration of particles or the methods used were not applicable to characterization of cell aggregates in deep tissue. The purpose of our current work is to demonstrate and evaluate our MS-DOT-T-matrix based inverse algorithm for characterization of cells in concentrated suspensions. In our method, the scattering spectra of cell suspensions provided by MS-DOT are used to retrieve the statistical parameters of cells including mean size, volume fraction and aspect ratio through the T-matrix based algorithm. Both round and spheroidal cells embedded in turbid media are studied. This work is the first demonstrating the application of MS-DOT for morphological characterization of cells suspended in a culture medium, by analyzing diffusely scattered light collected along the phantom boundary.

2. Methodology

2.1 Theoretical summary and the inverse algorithm

The first step in our technique is to shine laser beams into a mixture, containing the cells to be characterized. The light sources are placed sufficiently far from the suspensions. The presence of cells influences the light scattering and propagation within the suspensions. Afterwards, photon densities along the medium boundary are recorded by some detectors for further analysis. We suppose that cells have spheroidal shapes and they are dispersed in suspensions with random orientations. The photon propagation in suspensions satisfies the following photon diffusion equation, which can be rigorously solved by the finite element method [7].

$$\nabla \cdot (D(r)\Phi(r, \omega)) - [\mu_a(r) - i\omega/c] \Phi(r, \omega) = -S(r, \omega) \quad (1)$$

where $\Phi(r, \omega)$ is the photon density, ω is the angular frequency, $D(r) = 1/3 (\mu_a(r) + \mu'_s(r))$ is the diffusion coefficient, $\mu_a(r)$ is the absorption coefficient, $\mu'_s(r)$ is the reduced scattering coefficient, c is the speed of light in the medium, $S(r, \omega)$ is the source term, and an $\exp(-i\omega t)$ time variation is assumed. Once the reduced scattering spectra are recovered using the MS-DOT algorithm, as elaborated in Ref. [7], they have to be further processed to recover the morphological parameters of the cells.

At each computational node, the measured reduced scattering coefficient is related to the particle morphology parameters through the following expression:

$$\mu'_s(\lambda) = \varphi \int_{r=r_{\min}}^{r=r_{\max}} \int_{\varepsilon=\varepsilon_{\min}}^{\varepsilon=\varepsilon_{\max}} \frac{e^{-\frac{1}{2} \left(\frac{(r-r_m)^2}{\delta_m^2} + \frac{(\varepsilon-\varepsilon_m)^2}{\delta_\varepsilon^2} \right)}}{2\pi\delta_r\delta_\varepsilon} \left(1 - \frac{\alpha_1^1(r, \varepsilon)}{3} \right) \langle C_{sca}(r, \varepsilon) \rangle dr d\varepsilon \quad (2)$$

The reduced scattering coefficient, at a wavelength λ , is denoted by the symbol $\mu'_s(\lambda)$. $r \in [r_{\min} r_{\max}]$ is the equal-volume radius of the scattering particle. $\varepsilon \in [\varepsilon_{\min} \varepsilon_{\max}]$ is the aspect ratio, which is the fraction of the equatorial radius to the polar radius of the spheroidal particle. We have utilized a bivariate normal distribution function to describe the random variations in the cell size and aspect ratio. The mean values of the particle radius and aspect

ratio are represented by the symbols r_m and ε_m , respectively. The concentration of cells is represented by the symbol φ , which represents the number of cells per cubic millimeter. The following relationship holds between the cell volume fraction and its concentration:

$$VF(\%) = 0.1\varphi \text{ (in million per cubic millimeter)} \times \frac{4\pi}{3} r_m^3 \text{ (in } \mu\text{m)}^3 \quad (3)$$

δ_r and δ_e are the standard deviations of the cell size and aspect ratio, respectively. In this work we assumed that the standard deviations equal to 10% of their corresponding mean values. $\langle C_{sca}(r, \varepsilon) \rangle$ denotes the averaged scattering cross section which can be calculated using the T-matrix method. $\alpha_1^1(r, \varepsilon)$ is a coefficient obtained by expanding the phase function [15]. The values of $\langle C_{sca}(r, \varepsilon) \rangle$ and $\alpha_1^1(r, \varepsilon)$ are *a priori* calculated and stored in a T-matrix database.

In order to recover the cell size ($2r_m$), aspect ratio (ε_m) and the concentration (φ), we first need to find the best estimates for these parameters, in a range of their possible values, by

minimizing the error function $\chi^2 = \sum_{\lambda_j} (\mu'_s(\lambda_j)^o - \mu'_s(\lambda_j)^c)^2$, where $\mu'_s(\lambda_j)^o$ is the observed reduced scattering coefficient and $\mu'_s(\lambda_j)^c$ is the calculated one from Eq. 2 and $\lambda_j, j=1, 2, \dots$ are the measurement wavelengths.

The updates for unknown parameters are determined by solving the following equation

$$(J^T J + \gamma I) \Delta \zeta = J^T \Delta \chi \quad (4)$$

where γ is an appropriately chosen regularization parameter, I is the identity matrix, the vector $\Delta \chi = [\mu'_s(\lambda_1)^o - \mu'_s(\lambda_1)^c, \dots]^T$ represents the error between the observed and calculated reduced scattering coefficients and $\Delta \zeta = [\Delta \varphi \ \Delta \delta_r \ \Delta r_m \ \Delta \delta_e \ \Delta \varepsilon_m]^T$ is the vector which updates the parameters of interest. We can obtain explicit expressions for the elements of the Jacobian matrix J by direct differentiation of Eq. (2) with respect to the model parameters.

2.2 Experimental setup and the data collection

Briefly, light from one of the five continue-wave laser modules was transmitted to an optical switch, which sequentially passes it to 16 pre-selected points at the surface of the phantom for 2D imaging experiments. 16×16 measured data was then input into our MS-DOT reconstruction algorithm to generate a 2D cross-sectional image of the phantom [7]. Measurement data at five wavelengths of 733, 775, 808, 840, and 915 nm were used for the image reconstructions.

For cellular sizing, depending on the cell type and its refractive index, an *a priori* stored T-matrix database with different combinations of cell size and aspect ratio was used to calculate the reduced scattering coefficient [16] and recover the unknown morphological parameters using the inverse algorithm [15]. A separate T-matrix database was used for characterization of the scattering Intralipid particles in the background [17–18]. Cells to be characterized were suspended in a glass tube which was embedded off-center in the cylindrical background phantom consisting of 1% Intralipid and India Ink (Fig. 1).

3. Results and Discussion

3.1 Chinese hamster ovary (CHO) cells

CHO cells are one of the most commonly used cell cultures in biological and medical researches. CHO cells have epithelial morphology in culture and grow attached to the surface of a flask. While attached to the surface, CHO cells have polygonal and elongated shapes as shown in Fig. 2a. The size of CHO cells is about $\sim 14\text{--}15\text{ }\mu\text{m}$ with refractive index of ~ 1.38 [19–21]. Optical scattering from CHO cells cannot be accurately measured while they are attached to the plastic surface. Hence, we utilized a serum free medium to suspend CHO cells following the procedure described in Ref. [22]. While in suspensions, the irregular-shaped CHO as anchorage-dependent cells change their morphology and become rounded. As shown in Fig. 2b, CHO cells have almost spherical shapes in serum free medium.

The T-matrix database was generated for cells with size ranging from 3 to 9 μm in equivalent radius, and 0.8–1.2 in aspect ratio [16]. The background was a 70mm-diameter cylindrical solid phantom made of Intralipid and India ink, and had an absorption coefficient of 0.005mm^{-1} and a reduced scattering coefficient of $\sim 0.2\text{mm}^{-1}$. A 9mm-diameter glass tube, containing CHO cells, was embedded 20 mm away from the center of the background phantom. Table 1 gives the averaged reduced scattering coefficient of CHO cells/target and the background phantom at five wavelengths of interest. The presented data show that the background phantom had higher scattering coefficient compared with the target due to relatively low concentration of CHO cells used. Figs. 3a, 3c and 3e, respectively, show the recovered images of particle size, concentration and aspect ratio using our T-matrix based algorithm. For better quantitative assessment, each of the recovered parameters is also plotted along a line passing through the centers of the background and target (Figs. 3b, 3d, and 3f).

According to Figs. 3a and 3b, the recovered size of CHO cells is about 14 μm , in good agreement with that measured by microscopy (Fig. 2b). The aspect ratio images (Figs. 3e and 3f) demonstrate that the cells in the target and the Intralipid particles in the background are spherical in shape. Again, this agrees well with that measured by microscopy for both the cells (Fig. 2b) and Intralipid [17]. The concentration of CHO cells used in our experiments could not be accurately measured using microscopy, but we estimate from the reconstructed value of volume fraction (Figs. 3c and 3d) that the recovered concentration of CHO cells is $\sim 60,000\text{ cells/mm}^3$.

3.2 Red Blood Cells

Erythrocytes or human red blood cells (RBCs) as target with different concentrations were also examined. Unlike the mammalian cells such as CHO, RBCs retain their non-spherical shape in isotonic solutions (with the same salt concentration of blood cells). Normal saline solution (0.9% NaCl) is considered isotonic with blood cells. Therefore, RBCs are a good candidate to demonstrate the capability of our algorithm in retrieving the morphological parameters of non-spherical cells.

The optical properties of individual RBCs and compounded blood samples were studied previously [23–26]. The size of RBCs is in the range of 6–7 μm with a relative refractive index of ~ 1.05 compared to the surrounding saline medium. In optical study of RBCs, RBCs are typically represented by homogenous volume-equivalent oblate spheroids [25]. In our study the T-matrix database for RBCs was generated with particle size of 2–8 μm and aspect ratio of 0.4–4. Fig. 4 presents the microscopic images of RBCs suspended in saline solution with three different volume fractions of 9%, 16% and 33%. As shown in Fig. 4, when the concentration of RBCs increases, the average distance among cells decreases subsequently,

and some cells are tightly attached to each other or significantly deformed (Fig. 4c) compared to the lower concentration cases (Fig. 4a and 4b).

MS-DOT experiments were conducted when RBCs suspensions with different concentration were embedded in the Intralipid/ink background phantom. Table 2 lists the averaged reduced scatterings coefficients of the RBCs/target with three different concentrations/volume fractions at three wavelengths of interest.

It is worth to notice that when the concentration of RBCs is increased from 9% to 16%, the corresponding scattering coefficient increases almost proportionally. However, when the concentration of RBCs is increased from 16% to 33%, this linearity breaks down. Fig. 5 shows the retrieved images of particle size, concentration and aspect ratio for the three cases of RBCs. Here the particle size is defined as the diameter of sphere with equal volume of oblate RBC.

From the results shown in Fig. 5, the recovered equivalent size and aspect ratio of RBCs are $\sim 4.5 - 5 \mu\text{m}$ and ~ 2.3 , respectively, resulting in an equatorial diameter of $6 - 6.5 \mu\text{m}$ for the RBCs. This is quantitatively consistent with that measured by microscopy (see Fig. 4) and that available from the literature [25]. The concentration images shown in Figs. 5b and 5e demonstrate that the recovered volume fractions of RBCs are almost same as the actual values in saline solution (i.e., 9 and 16%). However, when the actual concentration of RBCs is increased to 33%, the retrieved volume fraction of the RBCs is only 18% (see Fig. 5h). This large error in volume fraction (15%) is mainly caused by the interactions between cells in highly concentrated suspensions that are not included in our current T-matrix based method [15].

In low to moderately concentrated suspensions, the distances among particles are relatively large and the interactions between neighboring particles can be ignored. In such cases, the reduced scattering coefficient is linearly proportional to the concentration of particles. As the concentration of particles within the suspensions is increased, the particle-particle interactions become significant and the reduced scattering coefficient is no longer a linear function of particle concentration. Compared with low to moderately concentrated suspensions, the scattering efficiency is substantially reduced for highly concentrated suspensions [27].

As shown in previous studies [27–28], when the volume fraction of particles reached $\sim 10\%$, the corresponding scattering coefficient saturated and further increment of volume fraction did not increase the scattering coefficient. In order to account for this effect, the static and angle-integrated structure factors were introduced and incorporated into their Mie-based scattering model. Consequently, improvements were obtained in calculation of reduced scattering coefficient in highly concentrated suspensions of particles (up to 40% in volume fraction) [28].

The absorption coefficient is not directly used in this work since it is a measure of the percent of energy not scattered/reflected. However, the absorption should be small since a greater amount of photons are scattered and detected by the light detectors, rather than absorbed in the medium. Therefore, we only process the reduced scattering spectra in order to extract the morphological parameters of cells in suspensions.

In all examples presented in this work, the absorption coefficient is much smaller than the reduced scattering coefficient. Compared with CHO cells, red blood cells (RBC) absorb the diffused light more strongly. However, the absorption coefficients for RBCs are still several orders of magnitude smaller than the scattering coefficients in the near infrared range [26].

Our algorithm has relatively robust performance against noisy data. Before testing our algorithm using experimental data, we conducted extensive simulations and the obtained results were in good agreement with the exact values, for up to 50% noise added to the “measured” data [15].

4. Conclusions

We have demonstrated the applicability of our T-matrix based algorithm for simultaneous recovery of shape parameters and concentration of cell suspensions embedded in scattering media using multispectral diffusing light measurements. Both spherical and non-spherical cells were examined. The recovered values of cellular morphology were validated by microscopy. For highly concentrated colloidal suspensions, the current particle scattering model needs some modifications to account for particle-particle interactions. This study is currently underway in our lab and the results will be reported in the near future.

Acknowledgments

We would like to thank Prof. M. Mishchenko for providing the FORTRAN code used for preparing the T-matrix database. This research was supported in part by an NIH grant (R01 NS069848).

References

1. Li C, Grobmyer SR, Massol N, Liang X, Zhang Q, Chen L, Fajardo LL, Jiang H. Noninvasive in vivo tomographic optical imaging of cellular morphology in the breast: possible convergence of microscopic pathology and macroscopic radiology. *Med Phys.* Jun.2008 35:2493–2501. [PubMed: 18649482]
2. Giacomelli MG, Chalut KJ, Ostrander JH, Wax AH. Review of the Application of T-Matrix Calculations for Determining the Structure of Cell Nuclei With Angle-Resolved Light Scattering Measurements. *Selected Topics in Quantum Electronics, IEEE Journal of.* Jul-Aug;2010 16(4):900–908.
3. Backman V, Gurjar R, Badizadegan K, Itzkan I, Dasari RR, Perelman LT, Feld MS. Polarized light scattering spectroscopy for quantitative measurement of epithelial cellular structures in situ. *Selected Topics in Quantum Electronics, IEEE Journal of.* Jul-Aug;1999 5(4):1019–1026.
4. Bartlett, Matthew; Huang, George; Larcom, Lyndon; Jiang, Huabei. Measurement of Particle Size Distribution in Mammalian Cells *In Vitro* by Use of Polarized Light Spectroscopy. *Appl Opt.* 2004; 43:1296–1307. [PubMed: 15008533]
5. Gurjar RS, Backman V, Perelman LT, Georgakoudi I, Badizadegan K, Itzkan I, Dasari RR, Feld MS. Imaging human epithelial properties with polarized light-scattering spectroscopy. *Nat Med.* 2001; 7:1245–1248. [PubMed: 11689891]
6. Wax, A.; Backman, V. *Biomedical Applications of Light Scattering.* McGraw-Hill; 2009.
7. Jiang, H. *Diffuse Optical Tomography: Principles and Applications.* CRC Press; 2010.
8. Riefler N, Wriedt T. Intercomparison of Inversion Algorithms for Particle-Sizing Using Mie Scattering. *Particle & Particle Systems Characterization.* Sep; 2008 25(3):216–230.
9. Tropea C. Optical Particle Characterization in Flows. *Annual Review of Fluid Mechanics.* 2010; 43(1):399–426.
10. Mishchenko MI, Travis LD, Mackowski DW. T-matrix method and its applications to electromagnetic scattering by particles: A current perspective. *J of Quantitative Spectroscopy and Radiative Transfer.* Jul; 2010 111(11):1700–1703. 50 Years of QRSRT.
11. Mishchenko MI, Zakharova NT, Videen G, Khlebtsov NG, Wriedt T. Comprehensive T-matrix reference database: A 2007–2009 update. *Journal of Quantitative Spectroscopy and Radiative Transfer.* Mar; 2010 111(4):650–658.
12. Duncan, Donald D.; Thomas, Michael E. Particle shape as revealed by spectral depolarization. *Appl Opt.* 2007; 46:6185–6191. [PubMed: 17712384]

13. Veselovskii I, Dubovik O, Kolgotin A, Lapyonok T, Summa D, Whiteman DN, Mishchenko M, Tanré D. Application of randomly oriented spheroids for retrieval of dust particle parameters from multi-wavelength lidar measurements. *Journal of Geophysical Research*. 2010; 115:D21203, 16.
14. Giacomelli, Michael G.; Chalut, Kevin J.; Ostrander, Julie H.; Wax, Adam. Application of the T-matrix method to determine the structure of spheroidal cell nuclei with angle-resolved light scattering. *Opt Lett*. 2008; 33:2452–2454. [PubMed: 18978884]
15. Reza Hajihashemi M, Jiang Huabei. T-matrix-based inverse algorithm for morphologic characterization of nonspherical particles using multispectral diffuse optical tomography. *Appl Opt*. 2011; 50:3896–3906. [PubMed: 21772373]
16. Mishchenko M. Capabilities and limitations of a current FORTRAN implementation of the T-matrix method for randomly oriented, rotationally symmetric scatterers. *Journal of Quantitative Spectroscopy and Radiative Transfer*. Sep; 1998 60(3):309–324.
17. Michels, Rene; Foschum, Florian; Kienle, Alwin. Optical properties of fat emulsions. *Opt Express*. 2008; 16:5907–5925. [PubMed: 18542702]
18. Li C, Jiang H. Imaging of particle size and concentration in heterogeneous turbid media with multispectral diffuse optical tomography. *Optics Express*. Dec; 2004 12(25):6313–6318. [PubMed: 19488277]
19. Han Y, Liu XM, Liu H, Li SC, Wu BC, Ye LL, Wang QW, Chen ZL. Cultivation of Recombinant Chinese Hamster Ovary Cells Grown as Suspended Aggregates in Stirred Vessels. *Journal of Bioscience and Bioengineering*. 2006; 102(5):430–5. [PubMed: 17189170]
20. Takagi M, Kitabayashi T, Ito S, Fujiwara M. Noninvasive measurement of three-dimensional morphology of adhered animal cells employing phase-shifting laser microscope. *J Biomed Opt*. 2007; 12:054010. [PubMed: 17994898]
21. Chang, Yi-Ren; Hsu, Long; Chi, Sien. Optical trapping of a spherically symmetric sphere in the ray-optics regime: a model for optical tweezers upon cells. *Appl Opt*. 2006; 45:3885–3892. [PubMed: 16724154]
22. Sinacore MS, Drapeau D, Adamson SR. Adaptation of mammalian cells to growth in serum-free media. *Molecular Biotechnology*. 2000; 15(3):249–257. [PubMed: 10986701]
23. Nilsson AMK. T-matrix computations of light scattering by red blood cells. *Appl Opt*. 1998; 37:2735–2748. [PubMed: 18273219]
24. Karlsson A, He J, Swartling J, Andersson-Engels S. Numerical Simulations of Light Scattering by Red Blood Cells. *IEEE Trans on Biomedical Engineering*. Jan; 2005 52(1):13–18.
25. Tsinopoulos, Stephanos V.; Polyzos, Demosthenes. Scattering of He-Ne Laser Light by an Average-Sized Red Blood Cell. *Appl Opt*. 1999; 38:5499–5510. [PubMed: 18324059]
26. Meinke, Martina; Müller, Gerhard; Helfmann, Jürgen; Friebe, Moritz. Empirical model functions to calculate hematocrit-dependent optical properties of human blood. *Appl Opt*. 2007; 46:1742–1753. [PubMed: 17356617]
27. Sun Z, Tomlin CD, Sevick-Muraca EM. Approach for Particle Sizing in Dense Polydisperse Colloidal Suspension Using Multiple Scattered Light. *Langmuir*. 2001; 17(20):6142–6147.
28. Sun Z, Sevick-Muraca EM. Investigation of structure factors in dense colloidal suspensions using frequency domain photon migration: polydisperse systems. *Journal of Colloid and Interface Science*. 2004; 270:329–336. [PubMed: 14697698]

Figure 1.
Diagram of the experimental setup for MS-DOT measurements

Figure 2.
(a) CHO cells attached to the plastic surface. (b) CHO cells suspended in serum free medium.

Figure 3. Reconstructed particle size (a, b), concentration (c, d), and aspect ratio (e, f) of CHO cells.

Figure 4.
(a) RBC in saline with 9% VF. (b) RBC in saline with 16% VF. (c) RBC in saline with 33% VF.

Figure 5. Reconstructed particle size (a, d and g), concentration (b, e and h), and aspect ratio (c, f and i) of red blood cells at three different concentrations (Top row: 9%, middle row: 16%, and bottom row: 33%).

Table 1

Averaged reduced scattering coefficient at five wavelengths of interest (mm^{-1})

	733 nm	775 nm	808 nm	840 nm	915 nm
CHO cells/Target	0.145	0.146	0.142	0.142	0.149
Background	0.202	0.204	0.205	0.205	0.209

Table 2Averaged reduced scattering coefficient at three wavelengths of interest (mm^{-1})

	808 nm	840 nm	915 nm
9% VF	0.6942	0.6280	0.6279
16% VF	1.7146	0.7163	1.0417
33% VF	1.7884	1.2616	1.1515