POSTER PRESENTATION

Open Access

Automatic motion probe setting assist system for cardiac magnetic resonance imaging

Shigehide Kuhara^{1*}, Shuhei Nitta², Taichiro Shiodera², Tomoyuki Takeguchi², Kenichi Yokoyama³, Rieko Ishimura³, Toshiaki Nitatori³

From 16th Annual SCMR Scientific Sessions
San Francisco, CA, USA. 31 January - 3 February 2013

Background

Planning assist systems for cardiac MR examinations are necessary for easier operation and shorter examination times. This is also true for motion probe setting in whole-heart MR imaging. In previous reports, we proposed an automatic planning assist system for couch adjustment, local shimming, and axial multislice imaging using single scout volume data of the chest. This system employs an atlas-based segmentation technique, so it can detect not only the heart region but also various anatomical structures in the chest. In the present study, this technique was employed to detect the position of the top of the right hemidiaphragm for motion probe setting. The results were also compared against the degree of interobserver error in manual annotation.

Methods

An ECG-non-gated 3D fast field echo (FFE) single volume covering the entire chest area was acquired using a 1.5-T MRI scanner (Excelart Vantage™ powered by Atlas, Toshiba Medical Systems) during a single breath-hold with TR/TE = 3.7/1.3, FOV = 500x350x350 mm³ (coronal slab), and readout/phase/slice encode steps = 256/64/35 in an acquisition time of approximately 9 seconds. The acquired volume was then transformed to match a prepared model volume with manual annotation of the heart region and the position of the top of the right hemidiaphragm, permitting the motion probe region of the input data to be located. Accuracy was assessed by measuring the Euclidean distance between the position of the top of the right hemidiaphragm obtained by our

Figure 1 Automatic motion probe setting assist system for MRI exams

¹MRI Systems Development Department, Toshiba Medical Systems Corporation, Otawara-shi, Tochigi, Japan
Full list of author information is available at the end of the article

method and that obtained by manual annotation, and evaluation was performed by comparison with the differences between two manual annotations as a measure of interobserver error.

Results

The proposed method successfully segmented the heart region and detected the position of the top of the right hemidiaphragm for motion probe setting in 48 datasets from 15 healthy volunteers. The processing time was approximately 1.6 seconds (2.5 GHz CPU, single-thread processing). The interobserver error was also measured for 15 datasets from 15 healthy volunteers manually annotated by two operators. The average Euclidean distance error and the interobserver error were 11.04 ± 4.73 and 14.76 ± 19.13 mm, respectively.

Conclusions

Our previously proposed automatic planning assist system has been extended to motion probe setting. The results of the present study showed that the position of the top of the right hemidiaphragm could be detected by our method almost as accurately as by manual annotation. It is expected that this method should be useful in the clinical setting.

Funding

No funding was received for this research.

Author details

¹MRI Systems Development Department, Toshiba Medical Systems Corporation, Otawara-shi, Tochigi, Japan. ²Corporate Research & Development Center, Toshiba Corporation, Kawasaki-shi, Japan. ³Department of Radiology, Faculty of Medicine, Kyorin University, Mitaka-shi, Japan.

Published: 30 January 2013

doi:10.1186/1532-429X-15-S1-E21

Cite this article as: Kuhara et al.: Automatic motion probe setting assist system for cardiac magnetic resonance imaging. *Journal of Cardiovascular Magnetic Resonance* 2013 15(Suppl 1):E21.

Submit your next manuscript to BioMed Central and take full advantage of:

- Convenient online submission
- Thorough peer review
- No space constraints or color figure charges
- Immediate publication on acceptance
- Inclusion in PubMed, CAS, Scopus and Google Scholar
- Research which is freely available for redistribution

Submit your manuscript at
www.biomedcentral.com/submit