Published in final edited form as:

Surf Interface Anal. 2013 January 1; 45(1): 302–304. doi:10.1002/sia.4882.

C60-ToF SIMS imaging of frozen hydrated HeLa cells

Alan M. Piwowar¹, Selda Keskin², Melissa Ortiz Delgado, Kan Shen², Jonathan J. Hue², Ingela Lanekoff³, Andrew G. Ewing⁴, and Nicholas Winograd²

¹The Dow Chemical Company, 727 Norristown Road, Spring House, Pennsylvania 19002, United States

²Department of Chemistry, The Pennsylvania State University, University Park, Pennsylvania 16802, United States

³Pacific Northwest National Laboratory, 902 Battelle Boulevard, Richland, Washington 99352

⁴Department of Chemistry, University of Gothenburg, Göteborg, Sweden SE-412 96

Abstract

Sample preparation continues to be a major challenge for secondary ion mass spectrometry studies of biological materials. Maintaining the native hydrated state of the material is important for preserving both chemical and spatial information. Here, we discuss a method which combines a sample wash and dry protocol discussed by Berman *et al*¹ followed by plunge freezing in liquid ethane for a frozen-hydrated analysis of mammalian cells (HeLa). This method allows for the removal of the growth media and maintains the hydrated state of the cells so that they can be prepared frozen-hydrated without the need for a freeze-fracture device. The cells, which were grown on silicon, have been successfully re-grown after the cleaning procedure, confirming that a significant portion of the cells remain undamaged during the wash and dry. Results from preliminary SIMS measurements show that it is possible to detect a large variety of bio-molecular signals, including intact lipids from the plasma membrane in the mass range of 700–900 Da from single cells, with little external water interference at the surface.

Introduction

The study of lipids and lipid metabolism is an important part of understanding both the function of cells and the role lipids play in disease progression.² Identifying the chemical signatures of lipids and other cellular species spatially on the dimensions of single cells (10–30 μm) is a challenging prospect. ToF-SIMS has the potential to provide this information, with several studies already showing the application for lipid identification at the surface of a material as well as for molecular identification of cellular materials into the bulk of the specimen.^{2–4}

Perhaps the largest obstacle for routine characterization of cells with ToF-SIMS is sample preparation. The SIMS technique obviously requires a vacuum environment. This reduced pressure causes several stresses to cells including molecule migration and morphological changes due to the rapid removal of volatile compounds such as water. Frozen-hydrated sample preparation has been shown to be an excellent method to closely maintain the native hydrated state and distribution of cellular materials in a high vacuum environment.⁵ The frozen-hydrated environment has also been shown to increase the ion yields for lipids and other cellular species.^{6,7} However, almost all frozen-hydrated cellular studies utilize a freeze-fracture device which often causes a fracture plane through the cell. Although extremely useful for sub-surface characterization, it can complicate identification for an intact cell including characterization of the outer cell membrane.

In this study, we discuss a preparation routine to characterize mammalian cells in a frozen-hydrated state after a sample washing procedure. The process does not require freeze-fracturing, which allows us to probe the lipid-rich region of the outer plasma membrane without interfering water-ice accumulation.

Experimental

ToF-SIMS studies were performed using a Bio-TOF SIMS instrument described previously.⁸ Briefly, data was collected using a 40 keV C₆₀⁺ primary ion source (Ionoptika Ltd., Chandler's Ford, United Kingdom) with an 80 ns pulse width and a beam current of approximately 30 pA rastered over a 300 × 300 μm² area while in delayed extraction mode. For the depth profiling, data were collected within an etched area that was three times less than the raster area to avoid edge effects. Experiments were performed at ~90K based upon thermocouple measurements. To achieve cryogenic temperatures, liquid nitrogen-cooled nitrogen gas was pumped from an external cylinder into cooling lines connected to the analysis stage.

The procedure for preparing frozen-hydrated cells follows closely the protocol detailed by Berman *et al.*¹ Specifically for our experiments, cells are pipetted into a petri dish that also contained several silicon shards (smooth side up) and growth media (DMEM, Cellgro, Manassas, VA). This dish was placed in a cell incubator for 24 hours. Longer or shorter incubation periods and the number of cells initially deposited affect surface coverage of cells on the silicon shards. Ideally, only 30–40% of the silicon surface should be covered with cells. Fewer numbers create difficulty in identifying the location of cells with imaging SIMS while too many cells exhibit large patches of overgrown areas, making isolation of single cells not possible. Once the surface coverage is appropriate, the silicon is removed from the media and immediately dipped into a 0.15 M ammonium formate cleaning solution, where it is gently swished around for ~5 seconds before repeating the process with a second clean solution. After the second wash, the edge of the silicon is contacted with an absorbent paper, which removes much of the excess washing solution. Following this procedure, a gentle stream of nitrogen gas is flowed over the silicon surface, again forcing any residual washing solution to a corner, where it is removed by contact with paper. Within less than 30 seconds the surface visually appears to be dry. To examine the state of the cells after the washing procedure, the samples are re-introduced into a new petri dish and given growth media. After 24 hours there is clear cell growth, supporting the idea that the cells are not damaged via the washing routine (see figure 1). Results indicating the health of the cells after similar cleaning protocols have also been reported.^{1,9}

Immediately following the drying process, the cells are plunged into liquid nitrogen cooled liquid ethane (99.0%, Scott Specialty Gases, Plumsteadville, PA). Although ethane and propane have similar cooling rates, ethane is used because it has been observed in several of our experiments to produce less chemical interference in collected SIMS data. The cells are held in liquid ethane for ~3–4 seconds and then are rapidly transferred to liquid nitrogen, where the silicon is attached to the SIMS sample holder. The cells are maintained in liquid nitrogen until their transfer into the instrument.

It is important to maintain adequate sample temperature before and during the experiment. The most serious concern for this type of experiment is allowing the sample temperature to increase above ~165 K in vacuum, at which point rapid sublimation of water from the surface begins.¹⁰ For our procedure, the sample stub is stored in liquid nitrogen before being transferred to a sample transfer plunger which is also cooled by emersion in liquid nitrogen.⁸ The transfer device is used to place the sample in the instrument chamber while minimizing the amount of atmospheric water which can adsorb to the surface of the sample. Once inside

of the instrument, the sample is quickly transferred to a pre-cooled analysis stage which is held at ~90 K. During the transfer, the sample stub temperature is not observed to increase by more than 20 K, well below the sublimation threshold temperature of 165 K. Therefore, cells properly prepared are maintained in their frozen-hydrated state.

Water deposition onto the sample surface during transfer is a major concern for this type of preparation. Although we have occasionally observed water cluster ions from some samples, the majority of the sample surface area, including the cellular regions, is generally free from a thick interfering ice layer. This effect may be a consequence of the presence of a residual amount of cryogen initially retained on the sample surface. We have observed that when a sample is characterized immediately after introduction to the sample analysis stage, SIMS peaks from ethane (the cryogen) can be observed. However, under normal conditions this ethane over layer is not spectrally observed after a few minutes in vacuum. It may be that the sublimation of the ethane over layer prevents thick water-ice from adsorbing to the surface.

Characterization of frozen-hydrated cells

The goal of this type of sample preparation is to maintain the structural integrity of the sample and to allow for proper identification of lipid membrane as well as other features at the surface. The SIMS data collected from HeLa cells grown on silicon, rinsed with 0.15 M ammonium formate, plunge frozen in liquid ethane and interrogated in their frozen-hydrated state are shown in Figure 2. This spectrum was collected using 40 keV C_{60}^+ bombardment *without* sputter removal of any surface material and therefore describes the outermost surface of the frozen-hydrated cell (i.e. the plasma membrane). The examination of some common lipids species observed for this membrane include intact lipid molecules in the range of approximately m/z 700–900, cholesterol (m/z 369) and various fragment ions for lipids such as m/z 184 and 86. An individual lipid species located at m/z 760, which is likely 1-palmitoyl-2-oleoyl-sn-glycero-3-phosphocholine, is also observed. As detailed in the image, these species all localize to regions localized on the HeLa cells. The size of these chemical features is slightly larger than that determined for suspended HeLa (15–20 μm). However these cells are grown on a surface and are expected to elongate as they adhere to the silicon. Hence, lipid species from single cell regions are preserved and show ion yields useful for imaging.

Spectral data collected from the same cells after receiving an etch dose of 2.0×10^{14} ions/ cm^2 are presented in Figure 3. As observed in the figure the key regions where the cells likely reside are still populated by common cellular ions. Perhaps most striking is the retention of a small but significant amount of information regarding the presence of cholesterol (m/z 369) and the high mass lipids (m/z 650–900). It is important to note that although there is no intense ion formation between m/z 650–900, integrating this area for imaging still shows localization to the cells. The intact generation of species in this region is significant because it indicates that even after sustained ion bombardment the individual cells still yield intact lipid species, albeit with very low intensity. With a modest increase in sensitivity we may be able to image these intact lipid species throughout a depth profile of the cell. This concept will be explored further on optimized instrumentation (11).

Conclusion

Sample preparation remains a key concern for biological SIMS investigations. We describe here a method to prepare cells in a frozen-hydrated state after sample washing with ammonium formate without the need for a fracturing device. The data show that cellular structure is maintained and that lipids species can be detected with useful intensities at the

surface of the cells. Sub-surface characterization is also possible and intact lipid species are likely detected, albeit at very low intensities.

Acknowledgments

The authors acknowledge the National Institutes of Health LIPID MAPS consortium (GM 069338-07) and grant 2R01 EB002016-18 for partial financial support and extend a special thanks to Irene Lee of Case Western Reserve University for her help with HeLa cell preparation.

References

1. Berman ESF, Fortson SL, Checchi KD, Wu L, Felton JS, Wu KJ, Kulp KS. JASMS. 2008; 19:1230.
2. Passarelli MK, Winograd N. Biochim. Biophys. Acta. 2011 **In Press**.
3. Fletcher JS, Lockyer NP, Vaidyanathan S, Vickerman JC. Anal. Chem. 2007; 79:2199. [PubMed: 17302385]
4. Nygren H, Hagenhoff B, Malmberg P, Nilsson M, Richter K. Microscop. Res. Tech. 2007; 70:696.
5. Chandra S, Bernius MT, Morrison GH. Anal. Chem. 1986; 58:493. [PubMed: 3963401]
6. Roddy TP, Cannon DM Jr, Ostrowski SG, Ewing AG, Winograd N. Anal. Chem. 2003; 75:4087. [PubMed: 14632121]
7. Piwowar AM, Fletcher JS, Kordys J, Lockyer NP, Winograd N, Vickerman JC. Anal. Chem. 2010; 82:8291. [PubMed: 20836508]
8. Braun RM, Blenkinsopp P, Mullock SJ, Corlett C, Willey KF, Vickerman JC, Winograd N. Rapid Commun. Mass Spectrom. 1998; 12:1246. [PubMed: 9772767]
9. Malm J, Giannaras D, Riehle MO, Gadegaard N, Sjoval P. Anal. Chem. 2009; 81:7197. [PubMed: 19639962]
10. Cannon DM Jr, Pacholski ML, Winograd N, Ewing AG. J. Am. Chem. Soc. 2000; 122:603.
11. Fletcher JS, Rabbani S, Henderson A, Blenkinsopp P, Thompson SP, Lockyer NP, Vickerman JC. Anal. Chem. 2008; 80:9059.

Figure 1.

Optical image of a silicon shard 24 hours after a wash and dry procedure reported in the experimental section. Note the presence of new cells which have grown over the edge of the silicon, indicating cell viability after sample cleaning.

Figure 2. Mass spectral image of HeLa cells prepared frozen-hydrated for different chemical species as identified in the figure (total refers to a the total counts for the spectrum and HMR corresponds to high-mass range and is the cumulative image for ions between m/z 700–900. Analysis field of view is $298 \mu\text{m} \times 298 \mu\text{m}$ with a spectral dose of 3.5×10^{11} ions/cm².

Figure 3.

Mass spectral image of HeLa cells prepared frozen-hydrated for different chemical species as identified in the figure (total refers to a the total counts for the spectrum and HMR corresponds to high-mass range and is the cumulative image for ions between m/z 700–900. Analysis field of view is $298\ \mu\text{m} \times 298\ \mu\text{m}$ with a spectral dose of 3.5×10^{11} ions/ cm^2 following an etch dose of 2.0×10^{14} ions/ cm^2 .