Published in final edited form as:

Methods Mol Biol. 2013 ; 931: 213–257. doi:10.1007/978-1-62703-056-4_12.

Laser Capture Microdissection for Protein and NanoString RNA analysis

Yelena Golubeva¹, Rosalba Salcedo¹, Claudius Mueller², Lance A. Liotta², and Virginia Espina^{2,*}

¹National Cancer Institute-Frederick/SAIC, Frederick, MD

²George Mason University, Center for Applied Proteomics and Molecular Medicine, Manassas, VA 20110

Abstract

Laser capture microdissection (LCM) allows the precise procurement of enriched cell populations from a heterogeneous tissue, or live cell culture, under direct microscopic visualization. Histologically enriched cell populations can be procured by harvesting cells of interest directly, or isolating specific cells by ablating unwanted cells. The basic components of laser microdissection technology are a) visualization of cells via light microscopy, b) transfer of laser energy to a thermolabile polymer with either the formation of a polymer-cell composite (capture method) or transfer of laser energy via an ultraviolet laser to photovolatilize a region of tissue (cutting method), and c) removal of cells of interest from the heterogeneous tissue section. The capture and cutting methods (instruments) for laser microdissection differ in the manner by which cells of interest are removed from the heterogeneous sample. Laser energy in the capture method is infrared (810nm), while in the cutting mode the laser is ultraviolet (355nm). Infrared lasers melt a thermolabile polymer that adheres to the cells of interest, whereas ultraviolet lasers ablate cells for either removal of unwanted cells or excision of a defined area of cells. LCM technology is applicable to an array of applications including mass spectrometry, DNA genotyping and loss-of-heterozygosity analysis, RNA transcript profiling, cDNA library generation, proteomics discovery, and signal kinase pathway profiling. This chapter describes laser capture microdissection using an Arcturus^{XT} instrument for protein LCM sample analysis, and using a mmi CellCut Plus® instrument for RNA analysis via NanoString technology.

Keywords

DNA; infrared laser; laser capture microdissection; molecular profiling; NanoString; phosphoprotein; pre-analytical variability; protein; RNA; tissue; tissue heterogeneity; UV laser

1. Introduction

Tissue heterogeneity is a normal biological attribute of multicellular organisms. A variety of cell types constitute tumors as well as healthy tissue. The ability to analyze a specific cell population, within a heterogeneous tissue sample, poses difficulties for both genomic and proteomic researchers during tissue analysis. Molecular analysis of heterogeneous tissue is currently ambiguous because it is impossible to discern which cells contribute which cellular constituents to a given tissue lysate (1). Molecular profiling of pure cell populations, which

*Corresponding author: George Mason University Center for Applied Proteomics and Molecular Medicine 10900 University Blvd MS1A9 Manassas, VA 20110 USA Phone: 703-993-8062 Fax: 703-993-8606 vespina@gmu.edu.

is reflective of the cell population's *in vivo* genomic and proteomic state, is essential for correlating molecular signatures in normal and diseased tissue (2-8). Laser capture microdissection is a technique that allows the identification, selection and isolation of pure cell populations from a heterogeneous tissue section, cytological preparation, or from live cell culture via direct microscopic visualization of the cells (2, 3). LCM enables researchers to isolate normal and diseased cells, as well as different stages of cells such as pre-malignant and malignant cells without contamination from surrounding cells (1, 4-7, 9-14). Xenograft tissues and host tissue may be isolated via LCM (15). Downstream analysis of the microdissected cells may be performed with any method that has adequate sensitivity. Recently a method for *in situ* proteomic analysis of microdissected cells has been described, permitting potential biomarker discovery in small numbers of breast cells (16).

LCM technology encompasses two general classes, infrared (IR) laser capture systems (2, 3) and ultraviolet (UV) laser cutting systems (17-21). The basic components of LCM technology are (a) visualization of the cells of interest via microscopy, (b) transfer of laser energy to a thermolabile polymer with formation of a polymer-cell composite (IR system) or photo volatilization of cells surrounding a selected area (UV system), and (c) removal of the cells of interest from the heterogeneous tissue section. The Arcturus^{XTM} system (Applied Biosystems/Life Technologies) incorporates both laser types in one instrument providing options as to the type of microdissection to be performed. The mmi CellCut Plus® instrument utilizes a UV laser for “cutting” selected cells from a tissue section.

The uses a stationary near-infrared laser mounted in the optical axis of the microscope stage to melt a thermolabile polymer film (*see* Note 1). The polymer film is manufactured on the bottom surface of an optical-quality plastic support cap. The cap acts as an optic for focusing the laser in the same plane as the tissue section. The polymer melts only in the vicinity of the laser pulse, forming a polymer-cell composite. A dye integrated into the polymer serves two purposes: 1) it absorbs laser energy, preventing damage to the cellular constituents, and 2) it aids in visualizing areas of melted polymer (Figure 1). Microdissection occurs when the polymer is removed from the tissue surface, at which point the embedded cells of interest are sheared away from the heterogeneous tissue section. The exact cellular morphology, as well as the DNA, RNA and proteins of the procured cells, remain intact and bound to the polymer (Figure 2). Ultraviolet lasers in both the Arcturus^{XTM} and the mmi CellCut Plus® instruments photovolatilize cells and/or the mounting medium, typically a polyethylene naphthalate (PEN) membrane. In UV cutting mode, a microdissection cap is placed on the surface of the tissue, encompassing the area of photovolatilization, and the infrared laser melts the thermolabile polymer in several defined areas around the edge of the cut area, thus holding the cut area of tissue/membrane in place. The mmi CellCut Plus® instrument uses an adhesive membrane mounted on the lid of a collection tube to secure the microdissected cells/membrane during microdissection. Following microdissection, extraction buffer may be applied directly to the polymer film for solubilizing the cells, allowing the collection of nucleic acids or protein for downstream analysis.

Tissue preservation methods should be taken into consideration when planning laser capture microdissection and downstream analyses. Although LCM is compatible with frozen, fixed, or paraffin embedded specimens, tissue preservation methods vary greatly in their ability to preserve nucleic acids, proteins, and post-translational modifications of proteins (22-24). Pre-analytical variability originating from tissue procurement and preservation may cause

¹ Significant amounts of heat are not deposited at the tissue surface during capture mode (IR) microdissection. Heat deposition is limited by the following engineering safeguards: a) the short laser pulse durations used, b) the low laser power levels required (the near-IR laser diode has a maximum output of 100 mW), and c) absorption of the laser pulse by the dye-impregnated polymer.

significant variability and bias in downstream molecular analysis. Depending on the *ex vivo* delay time in tissue processing, and the manner of tissue handling, nucleic acids, proteins, and phosphoproteins will be elevated or suppressed in a manner that does not represent the molecule at the time of excision. Consequently, molecular profiling requires stabilization, or preservation, of the molecules of interest immediately post tissue procurement. Proteins can be extracted with variable yield from formalin fixed tissue (25). The yield depends on the time, chemistry of formalin fixation, and the tissue geometry and density. Formalin penetrates tissue at a variable rate, reported to be within the range of mm/hr (26-28). During this time the portion of the living tissue deeper than several millimeters would be expected to undergo significant fluctuations with regard to phosphoprotein analytes. The preferred specimen for protein and RNA analysis is frozen tissue because formaldehyde becomes hydrated in aqueous solutions, forming methylene glycol (26, 28). Methylene glycol readily penetrates tissue, yet it is the small percentage of carbonyl formaldehyde that covalently cross-links proteins and nucleic acids, resulting in tissue fixation (26, 28). Formalin cross-linking, the formation of methylene bridges between amide groups of protein, blocks analyte epitopes thereby decreasing the yield of proteins extracted from the tissue.

Desirable aspects of tissue preservation are a) preservation of nucleic acids, proteins and protein post-translational modifications such as phosphorylation state, b) maintenance of nuclear size and morphology, c) functional at room temperature, d) one-step system, and e) compatible with standard paraffin embedding and processing. We have developed a fixative for protein and phosphoprotein preservation that meets these criteria (Figure 3) (22, 23, 29).

This chapter presents microdissection of frozen tissue sections for proteomic analysis as an illustration of the operation of the Arcturus^{XT}™ in both capture and cutting modes, and microdissection of formalin fixed paraffin embedded (FFPE) tissue section on a mmi CellCut Plus® instrument for RNA analysis via NanoString technology (30).

2. Materials

Tissue for microdissection is usually prepared as a cryosection, or a paraffin embedded section, on a glass or PEN (polyethylene naphthalate) membrane, or PET (polyethylene terephthalate) slide. The tissue section is stained for histomorphologic identification of cells. The stained section is used immediately for laser capture microdissection, or stored in a desiccator, depending on the tissue and molecular analysis type.

2.1. Preparation of Tissue Sections for Arcturus^{XT} Infrared Capture Mode

1. Uncharged, pre-cleaned glass microscope slides, 25 × 75 mm
2. Optimal cutting temperature cryopreservation solution (Tissue-Tek® O.C.T.™) (Sakura Finetek) for embedding frozen tissue sections.)
3. Cryomolds
4. Specimen for protein analysis: frozen tissue or ethanol fixed tissue sections cut at 2-15 μm (thickness of 5-8 μm is optimal for capture methods) (see Note 2)

²Optimal tissue thickness for capture mode microdissection is 5-8 μm. Tissue sections cut less than 5 μm may not provide a full cell thickness, requiring microdissection of more cells for a particular downstream assay. Tissue sections thicker than 8 μm may not microdissect completely, leaving essential cellular components adhering to the slide. Recently, a modified LCM technique was described by Iyer and Cox to microdissect neurons from *Drosophila* embryos that were more than 8 μm thick (38). UV mode microdissection is capable of cutting tissue sections up to 200 μm. Specimen types for DNA or RNA analysis are frozen-tissue sections, ethanol fixed, or formalin fixed paraffin embedded tissue sections cut at 2-15 μm.

5. Dry ice; Inhalation and contact hazard. Use with appropriate ventilation and personal protective equipment.

2.1.2. Preparation of Frozen Tissue Sections for UV Cutting Mode (protein analysis)—

1. PEN membrane or PEN frame slides, 25 × 75 mm
2. Optimal cutting temperature cryopreservation solution (Tissue-Tek® O.C.T.™) (Sakura Finetek) for embedding frozen tissue sections.)
3. Cryomolds
4. Specimen for protein analysis: frozen of ethanol fixed tissue sections cut at 2-200 μm (*see* Note 2)
5. Dry ice

2.1.3. Preparation of FFPE tissue sections for UV Cutting Mode (NanoString analysis)—

6. Automated microtome
7. PET PALM® metal framed membrane slide (polyethylene terephthalate) (P.A.L.M. Microlaser Technologies)
8. RNase-AWAY™ (Molecular BioProducts)
9. RNase-free water (Hydro Picopure® 2UV Plus)
10. FFPE tissue paraffin block for RNA analysis
11. Ice pan with solid ice (kept at −20°C)
12. Water bath with RNase-free water at +43°C
13. 8% 3-aminopropyltriethoxysilane (APES) (Sigma) in dry acetone

2.2. Hematoxylin and Eosin (H&E) staining for frozen/FFPE sections for protein analysis

1. Mayer's Hematoxylin Solution (Sigma). Inhalation and contact hazard; wear gloves when handling.
2. Eosin Y Solution, alcoholic (Sigma). Highly flammable; store away from heat, sparks, and open flames. Contact hazard; wear gloves when handling.
3. Scott's Tap Water Substitute (Fisher), also known as Blueing Solution (*see* Note 3).
4. Ethanol (ethyl alcohol, absolute, 200 proof molecular biology grade). Flammable; store away from heat, sparks, and open flames.
5. Ethanol gradient: 70% (v/v in dH₂O), 95% and 100% ethanol (*see* Note 4).
6. Type 1 reagent-grade water (dH₂O).
7. Xylene. Xylene vapor is harmful and can be fatal if inhaled. Use in a well-ventilated area and discard in appropriate hazardous waste container. Flammable; Contact hazard; wear gloves when handling.

³Scott's Tap Water Substitute is an alkaline solution that is used to develop the blue color of the hematoxylin stain (39). Scott's Tap Water Substitute can be prepared by dissolving 3.5g sodium carbonate and 20.0g magnesium sulfate in 1 L of water.

⁴Prepare fresh solutions after staining more than 20-25 slides at a time or if the ambient humidity is greater than 50%. Solutions can be stored at 4°C for one week.

8. Protease inhibitors (Complete Protease Inhibitor Cocktail tablets; Roche). Use one inhibitor tablet for every 10ml of solution (*see* Note 5).

2.2.1. One-step Cresyl Violet Acetate/Eosin Y staining of FFPE sections for NanoString analysis—

1. Cresyl violet acetate (Sigma)
2. Eosin Y (VWR)
3. Ethyl alcohol USP Absolute-200 proof (AAPER Alcohol and Chemical Co.)
4. RNase-free water
5. ProtectRNA™ RNase inhibitor (Sigma)
6. Xylene
7. 50 ml conical (Falcon) tubes

2.3. Laser Capture Microdissection Infrared Mode

1. 1. Arcturus^{XT} laser capture microdissection system (Applied Biosystems)
2. CapSure® Macro LCM Caps (Applied Biosystems) are recommended for protein analysis. CapSure® HS LCM Caps (Applied Biosystems) are recommended for DNA or RNA downstream analysis (*see* Note 6).
3. SealRight tubes 0.5mL (USA Scientific) or GeneAmp® 0.5mL Thin-walled Reaction Tubes with Domed Cap (cat. No. N801-0611; Applied Biosystems) (*see* Note 7)

2.4. Laser Capture Microdissection UV mode

1. Arcturus^{XT} laser capture microdissection system (Applied Biosystems/Life Technologies)
2. mmi CellCut Plus® microdissection system (MMI)
3. mmi IsolationCap® (MMI)
4. PCR tube (200 µL)
5. CapSure® Macro LCM Caps (Applied Biosystems) are recommended for protein analysis. CapSure® HS LCM Caps (Applied Biosystems) are recommended for DNA or RNA downstream analysis (*see* Note 6)
6. SealRight tubes 0.5mL (USA Scientific) or GeneAmp® 0.5mL Thin-walled Reaction Tubes with Domed Cap (cat. No. N801-0611; Applied Biosystems) (*see* Note 7)

⁵To limit protein degradation, add protease inhibitors to the 70% ethanol, water, hematoxylin, and Scott's Tap Water Substitute staining solutions. Complete protease inhibitor tablets are soluble in aqueous solutions. Dissolve the tablets in Type 1 reagent grade water (dH₂O) and use this solution to prepare 70% ethanol.

⁶CapSure® HS caps can be used successfully for RNA, DNA, or protein extraction. HS caps are designed with a 12 µm rail on the polymer surface, which prevents the polymer from directly touching the tissue except in the vicinity of the laser pulse. An extraction device is designed for use with the CapSure® HS caps, allowing extraction buffer to contact the polymer within a centrally designated area. These features limit any potential RNA contamination from surrounding cells. The area of the CapSure® HS cap outside the rail should be cleaned of any accumulated material because of the risk of accidental lysis of this material during centrifugation after extraction (35). In contrast, CapSure® Macro caps are placed in direct contact with the tissue, are not equipped with an extraction device, and any cellular material on the polymer surface will be available for extraction.

⁷Other brands/types of 0.5mL microcentrifuge tubes may not form adequate seals with the LCM cap. Leakage may occur if the cap/tube assembly is placed cap-down during incubation with buffer inside the tube.

2.5. Protein and RNA Extraction Buffers

1. Protein extraction buffer: 450 μ L T-PER Tissue Protein extraction reagent (Pierce), 450 μ L Novex® Tris-glycine 2X SDS loading buffer (Invitrogen), and 100 μ L TCEP (Tris (2-carboxyethyl)phosphine) Bond Breaker® 10% v/v (Pierce).
2. RNA lysis buffer for NanoString analysis: 2 μ L of buffer RLT with β -mercaptoethanol from AllPrep RNA/DNA extraction kit (Qiagen), 3.8 μ L of RNase-free water (Hydro Picopure® 2UV Plus) 1.0 μ L Proteinase K from FFPE RNA extraction kit (Qiagen), 0.3 μ L of RNaseOut™ RNase inhibitor (Invitrogen).

2.6. Adaptation of LCM workflow for NanoString platform

1. Aperio ScanScope® XT scanner
2. ImageScope™ software
3. Agilent Bioanalyzer (Agilent Technologies)
4. Agilent RNA 6000 Pico LabChip kit
5. AllPrep RNA/DNA extraction kit (Qiagen)

2.7. NanoString analysis of LCM samples

1. NanoString technology platform
2. LCM sample lysate
3. Reporter CodeSet
4. Capture ProbeSet

3. Methods

The protocols described below illustrate a) sample preparation of frozen tissue sections, b) hematoxylin and eosin (H&E) tissue staining for frozen sections and formalin fixed paraffin embedded sections (FFPE) or ethanol fixed paraffin embedded section, c) infrared capture mode microdissection, d) ultraviolet cutting mode microdissection using the Arcturus^{XT} and *mmiCellCutPlus*, e) cell lysis and protein extraction, and e) NanoString analysis (30).

3.1. Frozen Tissue Sectioning for Protein Analysis

Tissue biopsy samples should be stabilized promptly (within 15 minutes) after procurement. Stabilization methods for tissue that will be frozen include embedding in a cryopreservative solution or snap freezing as soon as the specimen is excised from the patient/animal. Prompt preservation of the sample reduces protein and RNA degradation as a result of protease and RNase activity respectively and limits reactive changes in phosphorylated kinases (22, 24, 31-33).

1. Embed tissue directly in a cryomold, covering the tissue with cryopreservative solution (OCT). Place OCT-embedded tissue on dry ice, or at -80°C , to freeze the tissue. Store the frozen tissue at -80°C (*see Note 8*).
2. Cut frozen sections at 2-15 μm thickness (5-8 μm is optimal for IR laser capture) and place sections on labeled, uncharged, pre-cleaned glass microscope slides (*see Note 9*). Position the tissue section near the center of the slide, avoiding the top and

⁸In our experience, it is best to store the block of tissue rather than storing the cut tissue sections. We have had successful protein recovery from frozen rhabdomyosarcoma blocks up to 12 years old when stored at -80°C (7).

bottom thirds of the slide (*see* Note 10). Place the slide directly on dry ice or keep it in the cryostat at -20°C or colder, or place the slide in a pre-chilled slide box on dry ice until the slides can be stored at -80°C . Alternatively the frozen section can be immediately stained and microdissected (*see* Note 11).

3.1.1. FFPE Tissue Sectioning for NanoString analysis—Maintain nuclease free conditions throughout the procedure: Tools should be wiped with RNase-AWAY™, rinsed with RNase-free water and wiped dry with a new kimwipe, A new disposable blade should be installed prior to cutting each sample. Use RNase-free water in the water bath and use individual containers for soaking the tissue blocks (*see* Note 12). Discard the first 20-25 μm of tissue before cutting an LCM section (or use these first sections for reference H&E). Sections should be protected from dust while drying.

Coat PET slides with 8% 3-aminopropyltriethoxysilane (APES) in dry acetone.

1. Expose PET slides to UV at 352 nm for 30 minutes.
2. Immerse the slides in APES coating solution for 10 minutes.
3. Wash slides in two changes of acetone for 1 minute.
4. Rinse slides in two changes of RNase-free water for 10 minutes.
5. Dry slides overnight at 37°C (*see* Note 13).

Design a block trimming template and trim the block

6. Use an Aperio annotated digital image of the H&E reference slide (Figure 4A) to design a block trimming template (Figure 4B).
7. Trim the block to maximize the number of sections mounted on the PET slide (Figure 4C, D), and set a water bath at $+43^{\circ}\text{C}$.
8. Cut serial 7 μm sections.
9. On a water bath, separate the number of sections that will fit inside the dissection area of the PET framed slide.
10. Position the tissue sections on the flat side of the PET slide, opposite the slide window. Center the sections such that the annotated area can be optimally dissected from around the slide midline (Figure 4D).

⁹Tissue with a thin open architecture, such as lung tissue, can be sectioned and collected on charged or silanized slides to prevent the tissue from nonspecifically adhering to the polymer during microdissection. Coated slides are generally not used for microdissection due to increased adhesive forces between the tissue and the slide. Effective microdissection is a balance between three adhesive forces: (1) maximizing downward adhesive forces between the polymer and the tissue, (2) minimizing lateral adhesive forces between the cells, and (3) minimizing upward adhesive forces between the slide and the tissue. Working with a new LCM tissue, conduct a pick-up-efficiency test to select an optimal type of slide for infrared mode LCM (35).

¹⁰The frame on membrane style slides limits the usable region of the microscope slide. In order to prevent potential damage to the instrument, the area available for microdissection on any slide type is limited by the width of the frame. Therefore it is best to place the tissue sections on the slide near the middle third of the slide. Tissue sections placed at extreme edges of the slide will not be within the usable microdissection region.

¹¹Do not allow the tissue section to dry on the slide at room temperature. Repeated fluctuations in temperature may cause the tissue to strongly adhere to the slide, preventing procurement of the cells of interest. Paraffin-embedded sections for DNA analysis can be stored at room temperature indefinitely prior to microdissection. Ethanol-fixed sections for RNA or protein analysis can be stored at room temperature for up to 3 months. Frozen sections for RNA analysis can be stored at -80°C for 1 month. Frozen sections for protein analysis can be stored at -80°C for up to 3 months.

¹²Wipe down the interior of the water bath with a kimwipe soaked in RNase-AWAY and rinse with RNase-free water. FFPE block should be well soaked for better adherence of the section to the membrane.

¹³The adherence of sections is tissue dependent. The majority of tissues adhere firmly to the membrane subjected to UV for 30 minutes before sectioning. Tissue containing collagen, muscle, bone marrow and bone always require adhesive treatment. We recommend testing adherence of sections to the membrane slide with the staining protocol of choice before serial sectioning.

11. Air-dry slides for 1 hour in a vertical position. Place a filter paper under the slide so that water which may have accumulated under the bottom section will drain out. Incubate the slides at 37°C overnight, followed with a 1 h incubation at 58°C.
12. Cool slide to room temperature and place each sample set in an individual slide box with a desiccant pouch or Drierite. Store slides at -20°C.

3.2. Hematoxylin & Eosin (H&E) Staining

Selection of tissue staining protocols should be based on compatibility with the downstream analysis (*see* Note 14). Staining protocols allow visualization and identification of the tissue or cells of interest with a standard inverted light microscope. Incorporation of protease inhibitors in the staining reagents, along with a microdissection session limited to 1 hr, minimizes protein degradation during the staining process. RNase inhibitors are recommended for tissues sensitive to RNA degradation. Skin tissue, cartilage, and samples prepared on charged slides might be difficult to microdissect and may require additional slide or tissue treatments (*see* Note 15) (34). The staining procedure employed depends on whether the tissue is frozen or paraffin embedded.

3.2.1. H&E staining procedure for frozen tissue sections—

1. Remove the frozen section slide from freezer/dry ice and proceed immediately with the staining protocol. Do not allow the slide to thaw.
2. Dip the slide in each of the following solutions, for the time indicated. Blot the slide on absorbent paper in between the different solutions to prevent carryover from the previous solution.
 - a. 70% Ethanol fixative, 3-10 s
 - b. dH₂O, 10 s
 - c. Mayer's hematoxylin, 15-20 s
 - d. dH₂O, 10 s
 - e. Scott's Tap Water Substitute, 10 s
 - f. 70% Ethanol, 10s
 - g. Eosin Y (optional), 3-5 s (*see* Note 16)

¹⁴Stains compatible with LCM include H&E, Methylene blue, Wright-Giemsa or Toluidine blue, Cresyl Violet acetate, and Methyl green. Staining of the cytoplasm with Eosin is not necessary for visualization of cells during microdissection, unless it is needed to enhance the contrast between target and adjacent non-specific tissue on the LCM dissecting screen. Fluorescence stains are compatible with fluorescence-equipped systems. Arcturus^{XT} instruments include metal halide lamps with blue (455-495 nm excitation, 510 nm emission), green (503-548 nm excitation, 565 nm emission), and red filter cubes (570-630 nm excitation, 655 nm emission) that can be used for immuno-LCM protocols (40).

¹⁵Tissue with strong intracellular adhesion might be difficult to microdissect and may require a modified staining protocol. The following protocol incorporates glycerol in the staining procedure for frozen sections as adopted from Agar, NS et al (41):

- a. Mayer's hematoxylin, 30 s
- b. dH₂O, 15 s
- c. 70% Ethanol fixative, 10 s
- d. 95% Ethanol, 10 s
- e. dH₂O, 10 s
- f. Scott's Tap Water Substitute, 15 s
- g. 70% Ethanol, 2 min
- h. 3% glycerol in PBS, 5-10 min
- i. 100% Ethanol, 10 s
- j. 100% Ethanol, 1 min
- k. Xylene, 30-60 s
- l. Xylene, 30-60 s
- m. Air dry slide as quickly as possible

- h. 95% Ethanol, 10 s
 - i. 95% Ethanol, 10 s
 - j. 100% Ethanol, 30-60 s
 - k. 100% Ethanol, 30-60 s
 - l. Xylene, 30-60 s (*see* Note 17)
 - m. Xylene, 30-60 s
3. Allow the stained slide to air dry as quickly as possible (*see* Note 18).
 4. Proceed immediately with microdissection. Do not coverslip the slide.

3.2.2. H&E staining procedure for formalin-fixed or ethanol-fixed paraffin-embedded tissue sections—

1. Paraffin-embedded tissue sections must be de-paraffinized and rehydrated to allow staining of the tissue elements. Dip the slide in each of the following solutions, for the time indicated. Blot the slide on absorbent paper in between the different solutions to prevent carryover from the previous solution.
 - a. Xylene, 5 - 15 min (*see* Note 19)
 - b. Xylene, 5 - 15 min
 - c. 100% Ethanol, 30 s
 - d. 95% Ethanol, 30 s
 - e. 70% Ethanol, 30 s
 - f. dH₂O, 10 s
 - g. Mayer's hematoxylin, 15 s
 - h. dH₂O, 10 s
 - i. Scott's Tap Water Substitute, 10 s
 - j. 70% Ethanol, 10 s
 - k. Eosin Y (optional), 3-5 s (*see* Note 16)
 - l. 95% Ethanol, 10 s
 - m. 95% Ethanol, 10 s
 - n. 100% Ethanol, 30-60 s
 - o. 100% Ethanol, 30-60 s
 - p. Xylene, 30-60 s
 - q. Xylene, 30-60 s

¹⁶Staining solutions can be prepared in 50mL conical tubes. Use forceps or tweezers to hold the slide and dip the slide in and out of the solutions for the times indicated. Dilute Eosin Y 1:1 with 100% ethanol to prevent over-staining of cytoplasmic proteins. Complete dehydration of the tissue is necessary for minimizing the upward adhesive forces between the tissue section and the slide. Increasing incubation time to 2 min for the 100% ethanol and xylene rinses may enhance dehydration, resulting in maximized microdissection efficiency.

¹⁷If absolutely necessary, the slide may be left in xylene for a maximum of 5 min before proceeding with microdissection.

¹⁸Xylene dissolves the polymer on CapSure® LCM caps. It is crucial that the tissue-slide be completely dry before cap placement for microdissection. When dry, the slide will appear as a grayscale (non-refractive index-matched) image.

¹⁹Allow formalin fixed or ethanol fixed paraffin embedded slides to soak in Xylene for 5- 15 minutes to dissolve the paraffin.

2. Air dry slide as quickly as possible.
3. Proceed immediately with microdissection. Do not coverslip the slide (*see Note 18*).

3.2.3. One-step Cresyl Violet Acetate/Eosin Y staining procedure for FFPE sections (NanoString analysis)—Our previously described protocol (35) was modified for a) improved contrast on the dissecting screen of the mmi CellCut Plus® instrument, B) gentle treatment of sections to minimize adherence to the membrane, and c) better preservation of tissue RNA. RNase free conditions and reagents should be used for the whole procedure. Staining is performed in 50 ml Falcon tubes filled with 45 ml of required reagent (*see Note 20*).

1. Prepare fresh staining mixture for 4 slides: 300µl of cresyl violet stock (dissolve 250 mg of cresyl violet in 25 ml of 100% ethanol, mix on a shaker overnight, filter for RNase-free conditions, store at +4°C for at least 6 months for the stain to reach full strength), 100µL Eosin Y, 400 µL RNase-free water, and 400 µL 100% ethanol. Vortex vigorously for 30 seconds, centrifuge at 6000rpm for 1 minute, and pipette from the surface of the stain.
2. Dissect one sample at a time. After –20°C storage, equilibrate slide to +4°C for 15 minutes in a closed box, and then to room temperature for 20 minutes.
3. Move slide to a desiccator for at least 10 minutes before staining.
4. Place the slide in xylene #1 and incubate for 5 minutes.
5. Transfer the slide to xylene #2 and incubate for 5 minutes.
6. Transfer the slide to 100% ethanol #1, gently invert the tube and incubate for 1 minute.
7. Transfer the slide to 100% ethanol #2 and incubate for 1 minute.
8. By pipette, apply a maximum of 200µL of stain to each section with ProtectRNA RNase inhibitor (1:500) for 20 seconds, and immediately drain the slide by touching a kimwipe to the edge of the slide.
9. Dip the slide in 100% ethanol for 5 seconds to rinse the stain off.
10. Transfer the slide to 100% ethanol, and incubate for 30 seconds.
11. Transfer the slide to xylene #1, gently invert the tube and incubate for 2 minutes.
12. Transfer the slide to xylene #2, and incubate for 3 minutes.
13. Air-dry the slide for 5 minutes in a fume hood in a vertical position (*see Note 21*).
14. Transfer the slide to a desiccator at least for 10 minutes prior to LCM.

3.3. Laser Capture Microdissection: Infrared (IR) Capture Mode

The Arcturus^{XT} system combines IR capture microdissection (LCM) and ultraviolet (UV) laser cutting in one instrument. UV laser cutting microdissection allows the “cut and capture” of cells of interest by first ablating unwanted cells, thus preventing contamination during cell capturing. This system is particularly useful for microdissection of tissue sections

²⁰The size of metal framed slides as well as Falcon tubes is not standard. 45 ml of reagent will completely cover the top sections on the slide even if a slide can't reach the bottom of the tube.

²¹Xylene accumulates at the bottom of the slide window. When xylene evaporates from the section, drain it from the bottom of the slide window with a kimwipe and continue to dry.

up to 200 μm thick, such as plant tissue sections (5, 9). The Arcturus^{XT} instrument features imaging software for creating composite images of the tissue, allowing the user to identify differences in tissue architecture during cell selection. The Graphical User Interface permits the control of all operations in the system, including stage movement, slide selection, focus and light intensity, laser parameters, objective selection, cap transfers, and camera settings (Figure 1). Images may be captured, annotated, and saved as JPEG or TIFF files; live video can also be taken at any point during the microdissection process. This automated system is equipped with a trackball activated stage and a mouse for navigation across a slide. An interactive touch screen (stylus) monitor permits selection of single cells or groups of cells (36). A variety of slide types can be used with this automated system including glass, glass membrane (PEN), and/or framed membrane (PET) slides. This flexibility, along with the IR laser and UV laser capability of the system, allows the user to technically prepare any specimen for automated laser microdissection (*see* Note 22).

Microdissection is performed without cover slips or immersion oils. Lack of immersion fluids on any of the optics prevents refraction of light from the tissue image. Thus, the color and detail of a given tissue stain is lost as the stained slide dries. Manual LCM methods capitalize on the index refraction of a wet tissue slide for visualizing and reviewing an index-matched image of the tissue (11, 12). The Arcturus^{XT} allows an index-matched image or images to be digitally saved which can be used to directly mark the cells of interest for microdissection.

The microdissection process consists of six steps: 1) loading the slides and LCM caps, 2) locating the cells of interest, 3) LCM cap placement and laser location, 4) marking the cells of interest, 5) capturing the cells, and 6) unloading of the samples and caps containing the captured tissue.

3.3.1. Arcturus^{XT} instrument set-up—

1. Turn on the PC, then the Arcturus^{XT} instrument, and touch screen monitor.
2. Open the Arcturus^{XT} software program by double clicking on the Arcturus^{XT} software icon.
3. Click on “Present Stage” in the SetUp tool pane. The microscope stage moves forward to allow loading of CapSure® caps and slides.
4. Load the CapSure® caps by sliding a CapSure® cap cartridge into the stage slot. Each cartridge holds 4 individual LCM caps.
5. Load up to 3 slides on the stage.
6. Remove any existing caps in the QC/Unload area.
7. Click the “i” options button to open the “Load Options” dialog box. Enter information about your slides and caps.
 - a. Check each slide that is loaded.
 - b. Check “Load with Overview” to create a full slide overview images.
 - c. Select the type of slide: glass, membrane or frame.
 - d. Enter a slide name/ID in the “SlideName” field.

²²Any specimen preparation may be used with the Arcturus^{XT} system: thin or thick sections, frozen or formalin-fixed tissues, stained, fluorescently stained, or unstained sections, hydrated or dehydrated specimens, fine needle aspirates, forensic smears, live plant specimens, and cell cultures.

- e. Enter any comments for each slide in the “SlideNotes” field.
- f. Click on the “Caps” tab to enter information about the caps.
- g. Select the type of cap: Macro or HS.
- h. Check each cap that is loaded.
- i. Click on “File Paths” tab to enter information regarding the location of saved images. Enter your desired file path/name information.
- j. Click “OK”.

3.3.2. Inspect image and locate cells of interest—

1. Use the trackball or mouse to move the stage to an area for microdissection. The live video image is displayed on the monitor. Alternatively, you can tap the stylus on the overview image at the location of interest.
2. To view a different slide, tap the slide button for the slide of interest (*see* Note 23).
3. In the “Inspect” tool pane, select the desired objective, brightness and focus by tapping on the corresponding button. The selected objective is red. Autobrightness and Autofocus settings are the middle icon between the up and down arrows (*see* Note 24).
4. Locate the cells of interest using the trackball, mouse, or stylus to move the stage to the desired location (*see* Note 25).

3.3.3. LCM cap placement and laser location—

1. Place a cap on the slide by clicking “Place Cap” icon in the “Microdissect” tool pane. The instrument places a cap at the center of the red box in the slide overview image. The apparent size of this box will change depending on the microscope objective. The cap location is outline in green.
2. Use the trackball to move to an area without tissue but still under the cap. This can be at the side of the tissue, a luminal area, or any other area without cells.
3. Tap the “i” options button in the “Select” tool pane. Tap the “IR Spot Sizes” tab.
4. Click on the desired spot size button. Click OK.
5. Double-click the mouse in the area without tissue to fire a test IR laser shot. This test spot allows you to verify the polymer wetting, spot size and IR laser location.
6. Place the mouse cursor directly in the center of the test spot. Right-click in the center of the spot and select “Located IR laser” (*see* Note 26).

²³If the slide overview is blank or does not update, right click in the slide overview and select “Reacquire Overview Image”.

²⁴Older manual LCM systems such as the PixCell® II/IIe instruments did not allow the user to microdissect directly from an index-matched image of the tissue. As a work around, map images were saved while the tissue was wet (or rewetted with a drop of xylene prior to microdissection), providing a guide for microdissection (11, 14). The Arcturus^{XT} instrument illumination system greatly enhances the image properties during microdissection. For optimization of cell visualization adjustments to the illumination system can be made by selecting the “i” options button under the Inspect tools pane. Click on the “Illumination” tab to adjust the brightfield lamp settings, white balance, camera gain and diffuser settings.

²⁵To capture a static image of the main image window, tap the “Camera” button in the “Inspect” tools pane. The images will be saved in the folder specified in the “File Paths” tab of the “Load Options” dialog box in the “Present Stage” tool pane. The 3 general types of image are: Before image - tissue before microdissection; After image - tissue after microdissection; Cap image - microdissected tissue only.

²⁶The laser must be located each time the objective is changed for microdissection. It is recommended to fire the test pulse using the objective you intend to use for microdissection. Microdissection may be performed with any suitable laser spot size and a 2x, 10x, 20x, 40x, 60x or 100x (no oil) objective.

7. Move to another area free of cells and fire another test IR laser shot. Assess the quality of the spot for the following parameters (Figure 1) (*see Note 27*):

- a. Clear center
- b. Dark ring around spot
- c. Appropriate spot size for cell type to be microdissected

Spot adjustments can be made using the options button in the “Select” tool pane. If the spot diameter is larger than desired, reduce the power and/or duration. If the spot diameter is too small, increase the power and/or duration (*see Note 28*). Fire additional test spots until the desired spot is achieved (*see Note 29*).

8. Measure the diameter of the test spot. Click on the ruler icon in the “Select” tool pane. Using the mouse or stylus, click and drag from one inner edge of the spot to the opposite inner edge. A line and label will be displayed showing the diameter of the spot. Click on the Move Stage (hand) icon to deactivate the ruler tool.
9. Tap the “i” options button in the “Select” tool pane to edit the spot diameter, laser power/duration, and IR laser location.
10. Tap the “IR Spot Sizes” tab. Enter the measured spot diameter for the selected spot size. Click OK.

3.3.4. Mark the cells for microdissection—

1. Locate the cells of interest using the trackball, mouse, or stylus to move the stage to the desired location.
2. Click on any tool in the “Select” tool pane. The pencil icon is for freehand drawing. Freehand drawing objects must be a closed figure, i.e. the ends of the drawing must touch to enclose the desired area. The Defined circle tool allows the selection of defined areas of known diameter. The Single IR spot tool is for single cell or single spot microdissection. The IR Spot Line tool permits microdissection of a line, either straight or curved. A drawing tool is active if the icon is gray. Deactivate each drawing tool by clicking again on the icon.
3. Using the stylus or mouse, mark cells on the live image that are to be microdissected.

²⁷The dark ring produced by pulsing the laser is caused by a combination of migration of the dye and changes in the thickness of the polymer wall at the site of the laser pulse, permitting visualization of the melted polymer. The black ring should be sharp in appearance with a clear center (Figure 1). This pattern indicates proper laser focusing, adequate laser operation and acceptable performance of the polymer. A “fuzzy” ring could indicate improper focusing of the laser, uneven placement of the cap on the tissue, or inadequate power and/or duration of the laser pulse. The following steps can be followed when troubleshooting a poorly wetted polymer spot:

- a. Reposition the cap on the tissue; the cap may be crooked or uneven in relation to the tissue.
- b. Relocate the laser; the laser may be out of alignment.
- c. Adjust the power and duration of the laser pulse. Increase the laser power in increments of approximately 10 mW and the duration by 2.0 ms and fire additional laser test pulses after each adjustment. Observe the wetted polymer for the appropriate appearance.
- d. If the above steps fail to resolve the problem, discard the cap and repeat the process with a new cap.

²⁸A phenomenon termed “polymer depletion” occurs when microdissecting large, polygon-shaped areas from the perimeter toward the center. As the laser melts the polymer downward onto the cells, the polymer is depleted on the edges of the laser fire area. As more and more polymer is melted onto the cells in a localized area, this depletion effect becomes more apparent. This can be prevented by microdissecting large, enclosed areas from the center of the area toward the perimeter, or using the freehand drawing microdissection tool.

²⁹Single-cell microdissection is possible by adjusting the power and duration settings such that a very narrow area of the polymer is melted with each laser pulse. Select the smallest spot size setting and manually adjust the laser power and duration. Suggested settings for single-cell microdissection are power 45 mW and duration 650 μ s using Macro caps.

4. Each drawing object is numbered and appears in the Drawing Items list, located to the right of the Capture Groups List (Figure 1-2).
5. Selected areas may be deleted or erased if they were erroneously selected. Right-click on the area to be deleted and select “delete” or use the eraser icon to delete partial areas. Alternatively, the drawing object may be deleted by right clicking on the item in the Drawing Items List and selecting “Delete Object”. Selected areas may also be copied and pasted to another location or dragged as needed to duplicate or correct the drawing item placement.
6. Different cell types can be marked for microdissection and separated out by utilizing Capture Groups. This feature allows cells to be marked from one image but microdissected separately. A Capture Group is a group of drawing items that will be microdissected during one microdissection session, on one cap (*see* Note 30).
 - a. Prior to marking cells for microdissection, click on the desired Capture Group (A, B, C, D) in the “Select” tool pane.
 - b. Mark the desired cell population for microdissection.
 - c. Select a new Capture Group. Mark the second desired cell population for microdissection.
 - d. Tap the “i” options button in the “Select” tool pane and click on Capture Group. Click the desired group and enter the attributes for that group (name, IR spot color, UV cut color). Click OK.
7. To view the area of microdissection or the number of laser spots in each area, tap the “i” options button in the “Select” pane. Click on the “Drawing Items” tab to determine the area microdissected (μm^2) or number of laser spots (*see* Note 31).

3.3.5. Capturing the cells—

1. Tap the Capture Group to be microdissected if capture groups were assigned.
2. Click on the IR capture icon in the “Microdissect” tool pane.
3. If necessary, move the cap to new areas of the tissue/slide to microdissect additional cells (of the same cell type) that were not included under the original cap diameter. To move the cap, use the trackball, mouse or stylus to move the stage to the new location. Right click on the slide overview image and select “Place cap at region center”.
4. Fire a test IR pulse in an area lacking cells to check the laser spot morphology (*see* section 3.3.5 steps 4-8). Adjust the power and/or duration as needed to achieve the desired spot size and quality.
5. Click on the IR capture icon in the “Microdissect” tool pane.

³⁰Capture groups are used to mark different types of cells from the same slide such that the cell populations will be microdissected separately, using two different caps. As an example, areas of tumor and stroma on the same slide can be marked for microdissection. Tumor cells can be assigned to Capture Group A, and will be color coded to indicate Group A, while stromal cells could be assigned to Capture Group B, and color coded to indicate Group B.

³¹It is possible to estimate the number of captured cells based upon the number of laser pulses counted during microdissection (which is automatically counted on the toolbar), the spot size, and the efficiency of microdissection. The percent efficiency of microdissection can be estimated by observing the polymer for cellular material within the diameter of the melted laser spot. 30 μm laser spot size: Number of pulses $\times 5 \times \% \text{ efficiency} = \text{total cells captured}$ 15 μm laser spot size: Number of pulses $\times 3 \times \% \text{ efficiency} = \text{total cells captured}$ 7.5 μm laser spot size: Number of pulses $\times 1 \times \% \text{ efficiency} = \text{total cells captured}$

3.3.6. Unload the samples and microdissected tissue—

1. After the desired number of cells, or a capture group, has been collected on a cap, move the cap to the QC station. Click on the “Move Cap to QC” icon in the “Microdissect” tool pane. The cap will be moved to the QC position on the stage. The stage will be aligned so the QC station holding the cap is in the optical path of the microscope. The cap image will be displayed in the live image window.
2. Inspect the tissue on the cap for efficiency of microdissection. Adjust the magnification (maximum of 20× in the QC station), focus and brightness.
 - a. Observe the cap for microdissection of the desired cells and for debris and/or adhesion of non-specific tissue to the polymer surface.
 - b. Estimate the percentage efficiency of microdissection by observing the polymer for cellular material within the diameter of the melted laser spot. Efficiency of microdissection is a critical factor for estimating the number of cells procured by LCM.
 - c. Debris or non-specific tissue adhering to the cap may be removed by gently blotting the polymer surface with the tacky side of an adhesive note. Do not use “super sticky” style adhesive notes.
3. The cap can either be removed for storage/cell lysis or placed back on the slide to continue microdissection. To place the cap back on the slide, right click on the appropriate cap icon in the “QC Caps” pane (to the right of the slide overview) (Figure 2), and select “Replace Cap on Slide”.
4. If no further microdissection is to be performed with the cap, remove the cap from the stage by carefully lifting the cap straight up from the QC position. Place the cap in a labeled microcentrifuge tube, label the outer edge of the cap, and place the cap/tube assembly in dry ice or at −80°C (for protein/RNA analysis) or at room temperature (for DNA analysis).
5. After all dissections are completed, remove all slides and caps from the stage. Click “Present Stage” for access to the slides and caps.
6. Close the Arcturus^{XT} software.
7. Turn-off the Arcturus^{XT} instrument.
8. Turn off the PC and the monitor.

3.4. UV Cutting Mode Microdissection Using the Arcturus^{XT}

By default, the instrument will perform IR capture first followed by UV cutting. Depending on the type of slide loaded in the instrument, details of the UV cutting vary (36). In UV cutting mode, glass slides will have a moat cut around the region of interest. For PEN membrane and PET frame slides, the instrument cuts around the region of interest, leaving tabs, or sections of uncut tissue. Tabs prevent the tissue from curling-up or detaching from the surface of the membrane slide. Initial slide preparation and instrument set-up are similar to the IR capture mode (sections 3.3.1. to 3.3.3. #2).

3.4.1. UV Laser Location—

1. To locate the UV laser for each objective, tap the “i” options button in the “Microdissection” tool pane. Select the “UV Locate” tab. Follow the steps in the dialog box.
 - a. Tap the locate UV button. The UV laser will be fired.

- b. From the live image, tap the location of the UV laser spot (designated by a circle) then tap the OK button in the dialog box next to the UV button.
 - c. If the UV laser is not visible, tap “Push On” to manually turn on the UV cutting laser.
 - d. Click OK to close the dialog box.
2. Click OK.

3.4.2. Mark the cells for UV cutting—

1. Locate the cells of interest using the trackball, mouse, or stylus to move the stage to the desired location.
2. Click on the freehand drawing tool, or the circle tool, in the “Select” tool pane.
3. Using the stylus or mouse, mark the cells on the live image that are to be cut or ablated with the UV laser. Assign capture groups if desired (see section 3.3.4. #6).
4. To determine the area (μm^2) of tissue to be microdissected tap the “i” options button in the “Select” pane. Click on the “Drawing Items” tab to view the microdissection area (μm^2).

3.4.3. Setting cut and capture properties—

1. Tap the “i” options button in the “Microdissect” tool pane. Select the “Select Settings” tab.
2. Tap “IRSpotSpacing” and enter a value of 100%. This allows the spots to touch but not overlap.
3. Tap “UV Settings” and enter values for the following settings. Values will depend on the area to be cut. Smaller areas require fewer tabs and fewer IR spots compared to larger areas.
 - a. IR SpotsPerCutLength – number of IR spots per cut length
 - b. Tab Length – distance (microns) between each UV cut
 - c. UV CutLength – length of UV cut before a tab
 - d. UV CuttingSpeed – speed of the UV laser. Increase the speed to cut more quickly. Alternatively, the UV laser cutting speed can be adjusted using the slide bar in the “Microdissect” tool pane, allowing adjustment in real time during UV cutting if necessary.

3.4.4. Cutting the cells—

1. Tap the Capture Group to be microdissected if capture groups were assigned.
2. Click on either the UV cut and capture or UV cut icon in the “Microdissect” tool pane.
3. Move the cap to the QC area and unload caps/slides.

3.4.5. UV Cutting Mode Microdissection for NanoString analysis using the mmi CellCut Plus®—The mmi CellCut Plus® instrument is designed to dissect sections on PET (polyethylene terephthalate) metal framed slides. A plain glass histological slide is used as a cover for the tissue section and a support for the membrane during microdissection. This two-slide ‘sandwich’ is inserted in the slide holder of the microscope

dissection stage. Dissections are software navigated. Microdissected tissue is collected with the inert adhesive lid on the collection tube. Cell lysis is performed by adding lysis buffer to the collection tube, closing the lid and placing the tube upside down (*see* Note 32). A slide can be scanned, and the scan used as a navigation reference during dissection. We recommend setting up the laser and camera configuration on a stained test slide of the tissue of interest prior to microdissecting experimental samples.

3.4.6. mmi CellCut Plus® instrument set-up—

1. Turn on PC and wait for Windows to boot.
2. Turn on the microscope power supply and the laser key.
3. Start the mmi CellTools® software, allow it to load fully, and press laser button on the control box.
4. Load a glass support slide on the microscope dissection stage (*see* Note 33).
5. Load a test membrane slide on the glass slide (slide window facing up) and make sure that the slide ‘sandwich’ is clipped tightly in the holder.
6. Select the 4x-objective on the instrument and in the software tool panel.
7. Set-up slide limits in the software tool panel.
8. Define the scanning area and press scan button in the overview section of the tool panel for a reference slide view.

3.4.7. Laser set-up—

1. On the reference scan of the LCM slide, move the red square to the clear membrane area.
2. Set a 10x-objective on the instrument and in the software tool panel, and disable ‘AutoShape’ function.
3. Bring membrane in focus, press on a drawing tool button and draw a long line across the dissecting screen.
4. Set laser speed at 10%, and laser power at 80%.
5. Press the Cut button in the tool panel and adjust the focus slider in a continuous slow motion to observe a cut line. Stop the focus slider when the optimal cut line appears.
6. Repeat step 5 with the speed of choice for serial dissections.
7. With the Laser Position button, set the laser position to match the cut line to the drawing.
8. On the reference scan of the LCM slide, move a red square to the target area and bring it in focus.
9. In groups 1, 2, and 3, mark one target per group with the drawing tool.
10. Cut the target in each group with decreasing laser power (group 1--80%, group 2--70%. Group 3--60%).

³²The cap is not necessary for dissected targets 1000µm in diameter. They can be collected with the forceps directly from the slide on a dissecting stage.

³³For a metal-framed slide to be positioned tightly in the slide holder, the glass support slide should be the same width or narrower than the metal-framed slide. The flat side of the glass side (without frosting) should face up.

11. Load the collection tube in the tube holder and attach the tube holder to the cap lift (the cap is in up position by default).
12. With the Cap Lift button, set the cap on the slide over the dissected target.
13. Observe the target image on the dissecting screen for color and brightness. If the image is discolored, use a Camera Configuration dialog box to set white balance (WB) and adjust brightness.
14. Lift the cap and observe target for a complete pick-up.
15. Repeat step 12 and 14 for each of the dissected targets.
16. The speed and focus of choice (from step 6), and the lowest power providing a complete target pick up (from step 12, 14 and 15) will be the parameters of choice for serial dissections (*see* Note 34).

3.4.8. Dissection of serial slides and target collection—The dissection and collection approach described below allows acquisition of an LCM sample for tissues such as mouse skin and skin papilloma that contain abundant collagen and have sub-optimal adherence of the tissue sections to the PET membrane slide. A low volume of lysis buffer is combined with multiple dissections from serial tissue sections to provide adequate material for downstream analysis. The method is applicable to target areas larger than 200 μm (*see* Note 35).

1. Remove the cap holder from the cap lift and remove the test slide from the dissection stage together with the glass slide.
2. Set the 4x-objective on the instrument and in the software tool panel.
3. Load an RNase-free glass support slide (*see* Note 36) on the microscope dissection stage, and insert the first serial slide.
4. Define the scanning area and press scan button in the overview section of the tool panel for a reference slide view.
5. Set the 10x-objective on the instrument and in the software tool panel.
6. Check the laser position.
7. Load the collection tube in the tube holder and attach the tube holder to the cap lift.
8. Disable ‘Collect with Cap Up’ function and select ‘Auto New Cap’ (*see* Note 37).

³⁴Despite a standard sample preparation it is impossible to have standard serial sections due to changing tissue architecture through the sample, and laser parameters are very sensitive to such change. The same is true with the slide characteristics. So, here and there, some targets won't be completely cut out. Make subtle adjustments to the setting, changing the following parameters one at a time: increase power, or decrease speed, and adjust focus by two notches up or down.

³⁵Dissectates equal or larger than 200 μm can be comfortably picked up from the cap with the fine tip forceps under the dissecting microscope. Since the optimum input in NanoString reaction is 100 ng, with the smaller targets number of dissected slides and labor hours diminish the feasibility of the project. However, small targets from multiple caps can be combined into one tube by the following procedure: Put 3–5 μl of 100% ethanol over the targets on the collection tube lid (ethanol preserves RNA and can be effectively removed by drying in a desiccator before sample lysis). With the pipette tip carefully release the targets from the cap into alcohol. Press the tube over the lid to seal the tube, and centrifuge it at 16,000 rpm for 15 seconds to collect the dissectates on the tube bottom. Place the tube in the desiccator for 15 minutes (dissectates will stick to the tube walls while ethanol evaporates). Cut the lid with dissectates off the next tube and repeat the procedure using the tube with the dissectates from the first cap.

³⁶RNase-free slides were prepared as follows: place glass slides in the staining rack and incubate in RNase-AWAY away for 2 minutes, blot the rack with slides on a new paper towel and transfer into the container with RNase-free water, incubate for 5 minutes, discard water and rinse slides inside the container under running RNase-free water for 5 minutes, blot the rack and dry slides overnight at 58°C. Store slides in a new slide box at RT.

³⁷‘Auto New Cap’ allows collecting subsequent dissectates overlaying each other in the center of the cap. It suppresses static and allows easy detachment of the dissectates as one unit from the cap.

9. With the cap in 'up' position navigate to the first section on the review scan and draw the target on the dissecting screen.
10. Press the Cut button, then collect the microdissected tissue by putting the cap in the 'down' and then 'up' position.
11. Navigate to the next section on the review scan, and draw and cut the target (*see* Note 38).
12. Repeat step 11 until all the dissected targets from the slide are collected on the cap (*see* Note 39) (Figure 5).
13. Remove the cap holder from the cap lift and carefully transfer the collection tube under a dissecting microscope set on 1x- magnification.
14. Dispense 3 μ L of 100% ethanol on the bottom of 200 μ L PCR tube. With fine tip forceps, detach the microdissected tissue from the lid and transfer them into the PCR tube by touching the tissue piece to the ethanol (*see* Note 40).
15. Spin the tube at 16,000 rpm for 15 seconds to collect microdissected cells on the tube bottom.
16. Place the tube in the desiccator until the next serial slide will be dissected (*see* Note 41).
17. Remove the dissected slide from the slide holder leaving the glass slide on the stage (*see* Note 42), and set a 4x-objective on the instrument and in the software tool panel. Insert the next serial slide.
18. Continue with the dissections following steps 4-7 and 9-18.
19. Upon sample completion, open the lid of PCR tube with microdissected tissue and place it in a desiccator for 15 minutes to evaporate the ethanol. Close the lid and place the tube with dry microdissected tissue/cells in 15 mL Falcon tube with a dessicant pouch. Store at -20°C before lysis.

3.5.Lysis and Protein Extraction of Microdissected Material for Downstream Analysis

The LCM cap, containing microdissected cells for protein analysis, can be stored at -80°C for extraction at a later date (*see* Note 43). Extraction of proteins from microdissected cells should be performed just prior to the downstream analysis to prevent aggregation of proteins, degradation of proteins, or binding of protein to the walls of the microcentrifuge

³⁸Silylation of slides with 8% APES makes the membrane sticky for the collection cap. In this case, simultaneous annotation and automated collection of dissectates is not feasible due to the membrane damage by the collection cap during the move from one dissectate to the next. The damaged membrane makes cutting incomplete, and also, the cutting line may not match the drawing. Drawing and cutting one target at a time eliminates this problem. If cutting is incomplete due to the poor tissue adherence, the dissectate can be detached from the membrane under dissecting microscope with forceps by the following procedure: place an RNase-free glass slide on top of the membrane slide and move a three-slide assembly under the dissecting microscope. Carefully remove the top slide and gently detach the dissectates from the membrane.

³⁹Each dissectate will be positioned very close to each other on the collection cap. When the overlap is substantial, the cap won't be able to pick up the next dissectate. In this case, move the slide with a Move tool from the initial position to a distance of a dissectate's width; the cap will pick it up.

⁴⁰Membrane dissectates are static. Ethanol allows easy detachment of membrane cutouts from the forceps tips.

⁴¹From ten to fifteen slides (3-6 sections per slide) can be dissected in 4 hours. Four to six hours of desiccator storage of stained FFPE sections doesn't affect RNA quality (42). The majority of the samples (depending on a tissue type by RNA content and a target size) will satisfy the 100 ng of total RNA requirement for NanoString analysis with 20-70 sections per sample.

⁴²Re-use the glass slide and the collection tube for all the serial slides of the sample unless they get contaminated with loose tissue. Use a new glass slide and collection tube for the next sample.

⁴³Microdissected cells for DNA analysis can be stored desiccated at room temperature up to 1 week prior to extraction. Samples for RNA analysis should be extracted immediately after microdissection because condensation in the microcentrifuge tube during storage may be a potential source of RNase contamination.

tube during prolonged storage. Microdissected samples for western blotting and/or reverse phase protein array analysis, can be prepared with the following denaturing cell lysis/protein extraction buffer:

1. Protein extraction buffer: 450 μ L T-PER Tissue Protein extraction reagent, 450 μ L Novex® Tris-glycine 2X SDS loading buffer, and 100 μ L TCEP Bond Breaker®. The final extraction buffer is a 10% (v/v) solution of TCEP in T-PER/Tris-glycine 2x SDS buffer (*see* Note 44).
2. Thaw each LCM cap at room temperature and remove all traces of condensation from the edges and rim of the cap. Place the CapSure® cap containing microdissected cells on a flat-clean surface, film side up. Using a pipette, dispense the desired quantity of extraction buffer directly on the cap film and incubate for 1 min. The maximum volume of extraction buffer that can be used to cover the surface of a CapSure® cap is 15 μ L.
3. Pipette the extraction buffer up and down on the surface of the cap to solubilize the cells. Be careful not to scrape the cap polymer.
4. Collect the extraction buffer containing the solubilized cells in a 0.5mL microcentrifuge tube. If more than one CapSure® cap was used to microdissect the cells of interest, solubilized cells from these caps can be collected in the same microcentrifuge tube.
5. Denature the proteins by heating the closed microcentrifuge tube at 100°C for 5-8 minutes prior to downstream proteomic analysis.
6. Briefly spin the microcentrifuge tubes at 14,000rpm to pellet any condensation in the tube. Store the samples at –80°C.

3.5.1. Lysis of Microdissected Material for NanoString Analysis—

1. Transfer tubes from –20°C storage to a +4°C refrigerator for 30 minutes.
2. Prepare lysis buffer (*see* section 2.5.) adding one extra reaction to the mix (*see* Note 45).
3. Centrifuge the tube at 16,000 rpm for 30 seconds to collect the microdissected tissue on the bottom of the tube.
4. Add 7.5 μ L of lysis buffer to each tube, vortex vigorously for 15 seconds and centrifuge at 16,000 rpm for 30 seconds.
5. Make sure that all the microdissected tissue is covered with buffer and not stuck to each other.
6. Seal the lid of the tube with Parafilm and place the tube in a metal dry heat block. Incubate at 56 °C for 68 hours.
7. Transfer the block with tubes to an oven preheated at +80°C, and incubate for 15 minutes.
8. Vortex the tube vigorously for 30 seconds and centrifuge at 16,000 rpm for 30 seconds.

⁴⁴The optimal protein extraction buffer for electrophoresis or microarray analysis consists of a detergent, a denaturing agent, and a buffer. 10% v/v TCEP in T-PER/Tris-glycine 2X SDS buffer is considered a mild denaturing extraction buffer for the solubilization of cellular proteins. If the microdissected cells are to be analyzed via mass spectrometry, a urea-based buffer is recommended. An example mass spectrometry compatible denaturing buffer is: 8M urea in TrisHCl pH 7.0-7.5 (34).

⁴⁵NanoString chip is design for the analysis of 12 samples. We prepared a lysate mix for 14 samples.

9. Cut off the tip of the RNase-free long-tip 20 μ L barrier tip, leaving a 3 mm stump, and insert it in the labeled PCR tube as a filter for the lysate.
10. Pipette in the lysate and transfer it to the filter by touching the barrier with the pipette tip.
11. Gather microdissected tissue in a lump with the pipette tip and transfer them to the bottom of the filter (*see* Note 46).
12. Centrifuge the filter-tube assembly at 16,000 rpm for 30 seconds, discard the filter and store the sample in -80°C prior to NanoString analysis.

3.6. Adaptation of LCM workflow for NanoString platform requirements

Considering that a large amount of dissected material for one sample should be collected in a timely manner we found it advantageous to use an automated microtome for serial sectioning, a modified target collection approach, and the use of an Aperio slide scanner to facilitate the LCM sample collection process (*see* Note 47).

3.6.1. Aperio digital image of H&E sections—The Aperio ScanScope®XT instrument allows scanning of the whole slide area at 20x and 40x magnification (we use 20x scanning mode for LCM projects). The resulting digital image, annotated by the pathologist, serves as a reference during target annotation on the dissecting screen of mmi CellCut instrument (Figure 6). A digital image can be accessed through the database and viewed from any computer (we use a wall mounted monitor to display a digital image during LCM). The magnification of the digital image can be matched to the magnification of the LCM objective of choice. The rotation tool allows placing the Aperio reference image into the same orientation as an mmi CellCut® reference view of the membrane slide before laser dissection.

1. During serial sectioning, mount each 10th section on a positively charged (plus) glass slide for an H&E stain and label slides sequentially.
2. Scan H&E slides in Aperio ScanScope®XT, review digital images for scanning quality, and assign them to the study folder for annotation by the pathologist.
3. Create a list of samples with Aperio image IDs for prompt image retrieval from the Spectrum™ database during LCM. With the search function of the ImageScope™ digital slide viewer, locate the image in the database.
4. When stained slides are drying in a desiccator, retrieve the first annotated image to the display monitor (Figure 7A).
5. Insert a test slide in the holder of the dissection stage, create a reference view of an LCM slide and compare the target image on the dissecting screen (Figure 7A, B) with the Aperio image on the display monitor.
6. With the rotation tool, rotate the annotated image (if required) for the same orientation as the LCM target image (Figure 7B, C).

⁴⁶The dissectates should be filtered out of the lysate due to the tendency of plugging the tip during pipetting, thus jeopardizing the input into NanoString hybridization reaction. The filtered volume of lysate varies across the samples. Prolong digestion and lysate transfer to the filter contributes to the loss of lysate volume (up to 3 μ L). For NanoString input samples should be adjusted to 5 μ L volume with buffer RLT as needed.

⁴⁷An automated microtome allows cutting sections of a standard thickness that improves laser focusing and, as a result, makes dissections more efficient. Dissection and collection approach, and a use of a digital slide scanner dramatically increased a speed of dissection.

7. With the zoom slider, match the Aperio image magnification to the magnification of LCM objective (Figure 7B, D), substitute the test slide with the first serial slide, and proceed to LCM.

3.6.2. Pilot studies: Amount of microdissected material for NanoString input—

We conduct pilot studies in order to determine a lysis procedure suitable for NanoString sample requirements, and to estimate the number of LCM slides per sample. The pilot experiment is recommended for each tissue type/project to ensure adequate cell number and RNA yield/

1. Use ALL Prep RNA/DNA extraction kit (Qiagen) for RNA extraction from pilot samples.
2. Calculate the area of each annotated target by Aperio ImageScope™ software, based on annotated H&E image.
3. Microdissect areas of various size containing the cells of interest and extract the RNA from microdissected samples. Example microdissected areas are: 2.7 mm², 3.3 mm², 6 mm², 12 mm², 12 mm², 24 mm². Calculate an average RNA yield (ng) for one mm² of target tissue (see Note 48).
4. Calculate the *Total Collected Area* (mm²) for NanoString input by the formula: quotient of division of *100 ng of Total RNA* by *Average RNA Yield* (ng/mm²) from step 4 (see Note 49).
5. Calculate the *Total Number of Sections* (per sample) by the formula: quotient of division of *Total Collected Area* (mm²) by *Annotated Target Area* (mm²).
6. Calculate the *Number of Slides* (per sample) by the formula: quotient of division of *Total Number of Sections* by *Number of Mounted Sections* (per slide).

3.6.3. Optimize LCM sample lysis—We recommend preparing a test block to optimize cell lysis conditions. We used twenty five dissectates (3 mm²), corresponding by area to the largest target across a sample-set, were incubated for 72 h in 5 µL of buffer RLT (see Note 50), and the final lysate volume was measured. On average, 2 µL of lysis buffer were lost during lysis procedure. Increasing the initial lysis volume to 7 µL resulted in 3.5-5 µL of lysate for NanoString input (see Note 51).

3.7. NanoString analysis of LCM samples

An example application of NanoString analysis with LCM samples is described below using microdissected skin papillomas from C57BL/6 mice. Skin papillomas were generated in C57BL/6 mice using a skin painting protocol (34). Briefly, 400 nmoles of DMBA (9, 10-Dimethyl-1,2-Benzanthracene) in 200 µL of acetone were applied to the skin, and ten days later, 40 nmoles of TPA (12-O tetradecanocylphorbol-13 acetate) in 200 µL of acetone were applied twice weekly for a total of 20 weeks. Developed papillomas were fixed in 10% neutral buffered formalin for 24 hours and paraffin embedded.

⁴⁸Recovery of RNA by column base methods is variable depending on nature and size of the target, temperature of the column, binding efficiency, etc. To estimate average RNA content in papillomas, we extracted RNA from single (2.7 and 3.3 mm²) and combined targets (6, 12 and 24 mm²). Compared to Trizol® (Invitrogen) extraction with Phase Lock Gel Tube (Eppendorf) (31), the loss of RNA in column based extraction is around 20%-30% (data not shown). Accordingly, an estimated average RNA yield in LCM lysate (see 3.6.2. step 4) was adjusted upward by 1.30%. Since RNA loss is inevitable with any RNA extraction method, the RNA content in LCM lysate, without extraction, will be higher than estimated.

⁴⁹The nCounter™ gene expression hybridization procedure for cell lysate has been optimized by NanoString Technologies, Inc. for ~10,000 mammalian cells per reaction, or the equivalent of approximately 100 ng of total RNA. The protocol requirements are: 4-5 µL at a concentration of 25-33 ng/µL (37).

⁵⁰Samples with a volume below 5 µL were adjusted for NanoString input by addition of buffer RLT.

1. Laser microdissected papillomas are used to prepare total RNA lysates for NanoString analysis. Twenty genes involved in inflammation and five housekeeping genes, used as an internal control (Table 1), are selected for a custom NanoString probe set (Reporter Code set and Capture Probe set).
2. The nCounter™ Gene Expression Assay is performed using two specific probes (capture and reporter) for each gene of interest. 5 µL of LCM lysate from each laser dissected papilloma is hybridized with the designed Reporter CodeSet and Capture ProbeSet (according to the manufacturer's instructions (NanoString Technologies, Seattle-USA), for direct labeling of mRNAs of interest with molecular barcodes, without the use of reverse transcription or amplification.
3. The hybridized samples are recovered with the NanoString Prep Station and the mRNA molecules counted with the NanoString nCounter™.
4. The resulting counts are corrected by subtracting the average value of the negative control (alien probes from the CodeSet, lacking spiked transcript) from the raw counts obtained for each RNA. Values less than zero are considered equal to 1.
5. The corrected raw data are finally normalized using values for the housekeeping genes.

3.7.1. Assay validity and sample quality assessment—The quality control of the NanoString analysis platform is based on the value of FOV (fields of view per sample) counted, a binding density, seven negative and six positive controls per sample, and presence or absence of messages under 'Lane Attributes' as displayed by the NanoString nCounter™. A sensitivity level of 600 FOV counts was used for the papilloma analysis (Table 2).

1. For the analysis to be qualified as 'successful', the FOV counted (corresponding to the counted number of reporters and showing the level of sensitivity in the system) should be close to 600 across the samples set (twelve samples).
2. The values for Binding Density (a measure of sample saturation) should be between 0.05 and 2.25 across the sample set. Positive control values should show

⁵¹Our transfer protocol allows transfer of intact sections from glass slides to PET metal framed membrane slides with subsequent laser dissection of target areas for molecular analysis:

- a) Heat coverslipped and unstained slides at +65-70°C for 5 minutes (proceed to step d with unstained slides).
- b) Incubate in xylene for 30 minutes to 1 hour, and remove coverslips.
- c) Rinse in xylene 2×1 minute (up and down movement) and proceed to step e.
- d) Deparaffinize sections in xylene (4× 5 minutes).
- e) Hydrate to water (100%--2×2 minutes, 95%--2 minutes, 70%--2 minutes, 30%--2 minutes, distilled water--2×2 minutes)
- f) Dehydrate to xylene (30%--2 minutes, 70%--2 minutes, 95%--2 minutes, 100%-- 2×2 minutes, xylene- 4×2 minutes)
- g) Cover section with thin layer of Krystalon™ (Harleco) (place the large drop in the middle of the section and let it spread by itself to avoid bubbles later on). The media should cover the section and 2-3 mm of the slide around the section (to avoid section damaging during peeling).
- h) Heat at +65-70°C for 2.5 hours (media will be hard).
- i) Remove slide from heat and cool down.
- j) Incubate the slide in distilled water at +65°C for 1.5 hours to soften the Krystalon.
- k) Gently lift the section with the forceps and peel it off the slide (at this step tissue can be cut for optimal mounting on PET slides).
- l) Cover the PET slide with +65°C distilled water.
- m) Put the tissue section embedded in Krystalon the same side down as it was on the original slide and cover the section with +65°C water to flatten the section.
- n) Dip slide in +65°C water until the section is visibly flat.
- o) Heat slides at +65-70°C for 1 hour and cool down for 5 minutes.
- p) Carefully drop xylene on the edges of the section and incubate the slide in xylene (4×2 minutes) to remove Krystalon (not longer than 20 minutes to avoid membrane detachment from the metal frame).
- q) Dry and laser dissect, or hydrate back to water for additional staining.

linearity with corresponding dilutions in a descending order, and the first (largest) values should be close across a sample set.

3. The negative control values should be low and comparable across the sample set in a range from 0 to 10. There should be no messages displayed in the 'Lane Attributes'. An example from our papilloma data set that satisfied all quality control conditions of the NanoString analysis platform is shown in Table 2.
4. The results are analyzed using the least variable gene (*GAPDH*) for normalization. Alternatively, the average of the three least variable genes (*GAPDH*, *RPI9* and *RPI30*) may be used for comparison. The calculations for the background subtraction, correction factor and normalization are performed as per standard NanoString analysis protocols (NanoString Technologies, Seattle-USA) (37) (Figure 8).

The mRNA counts pattern, similar using both analytical approaches (least variable gene or 3 least variable genes), demonstrated the suitability of our LCM sample preparation method for NanoString input, and the reliability of NanoString analysis of LCM samples for gene expression profiling. Our LCM procedure for a NanoString sample will suit profiling studies on archival clinical cases stored as FFPE sections mounted on glass slides, providing that sections are transferred from glass slides to PET membrane slides (*see* Note 51).

Acknowledgments

Thanks to the staff of Pathology-Histotechnology Laboratory, SAIC-NCI-Frederick. The work was funded in part by NCI Contract HHSN261200800001E.

5. References

1. Wulfkuhle JD, Speer R, Pierobon M, et al. Multiplexed cell signaling analysis of human breast cancer applications for personalized therapy. *J Proteome Res.* 2008; 7:1508–1517. [PubMed: 18257519]
2. Bonner RF, Emmert-Buck M, Cole K, et al. Laser capture microdissection: molecular analysis of tissue. *Science.* 1997; 278:1481–1483. [PubMed: 9411767]
3. Emmert-Buck MR, Bonner RF, Smith PD, et al. Laser capture microdissection. *Science.* 1996; 274:998–1001. [PubMed: 8875945]
4. Ma XJ, Dahiya S, Richardson EA, et al. Gene expression profiling of tumor microenvironment during breast cancer progression. *Breast Cancer Res.* 2009; 11:R7. [PubMed: 19187537]
5. Nakazono M, Qiu F, Borsuk LA, et al. Laser-capture microdissection, a tool for the global analysis of gene expression in specific plant cell types: identification of genes expressed differentially in epidermal cells or vascular tissues of maize. *Plant Cell.* 2003; 15:583–596. [PubMed: 12615934]
6. Petricoin EF 3rd, Bichsel VE, Calvert VS, et al. Mapping molecular networks using proteomics: a vision for patient-tailored combination therapy. *J Clin Oncol.* 2005; 23:3614–3621. [PubMed: 15908672]
7. Petricoin EF 3rd, Espina V, Araujo RP, et al. Phosphoprotein pathway mapping: Akt/mammalian target of rapamycin activation is negatively associated with childhood rhabdomyosarcoma survival. *Cancer Res.* 2007; 67:3431–3440. [PubMed: 17409454]
8. Wulfkuhle JD, Sgroi DC, Krutzsch H, et al. Proteomics of human breast ductal carcinoma in situ. *Cancer Res.* 2002; 62:6740–6749. [PubMed: 12438275]
9. Angeles G, Berrio-Sierra J, Joseleau JP, et al. Preparative laser capture microdissection and single-pot cell wall material preparation: a novel method for tissue-specific analysis. *Planta.* 2006; 224:228–232. [PubMed: 16721624]
10. Gallup JM, Kawashima K, Lucero G, et al. New quick method for isolating RNA from laser captured cells stained by immunofluorescent immunohistochemistry; RNA suitable for direct use

- in fluorogenic TaqMan one-step real-time RT-PCR. *Biol Proced Online*. 2005; 7:70–92. [PubMed: 16136226]
11. Mouldous L, Hunt S, Harcourt R, et al. Navigated laser capture microdissection as an alternative to direct histological staining for proteomic analysis of brain samples. *Proteomics*. 2003; 3:610–615. [PubMed: 12748941]
 12. Nakamura N, Ruebel K, Jin L, et al. Laser capture microdissection for analysis of single cells. *Methods Mol Med*. 2007; 132:11–18. [PubMed: 17876072]
 13. VanMeter AJ, Rodriguez AS, Bowman ED, et al. Laser capture microdissection and protein microarray analysis of human non-small cell lung cancer: differential epidermal growth factor receptor (EGFR) phosphorylation events associated with mutated EGFR compared with wild type. *Mol Cell Proteomics*. 2008; 7:1902–1924. [PubMed: 18687633]
 14. Wong MH, Saam JR, Stappenbeck TS, et al. Genetic mosaic analysis based on Cre recombinase and navigated laser capture microdissection. *Proc Natl Acad Sci U S A*. 2000; 97:12601–12606. [PubMed: 11050178]
 15. Kennedy J, Katsuta H, Jung MH, et al. Protective unfolded protein response in human pancreatic beta cells transplanted into mice. *PLoS One*. 2010; 5:e11211. [PubMed: 20585452]
 16. Cha S, Imielinski MB, Rejtar T, et al. In situ proteomic analysis of human breast cancer epithelial cells using laser capture microdissection: annotation by protein set enrichment analysis and gene ontology. *Mol Cell Proteomics*. 2010; 9:2529–2544. [PubMed: 20739354]
 17. Kolble K. The LEICA microdissection system: design and applications. *J Mol Med*. 2000; 78:B24–25. [PubMed: 11043387]
 18. Micke P, Ostman A, Lundeberg J, et al. Laser-assisted cell microdissection using the PALM system. *Methods Mol Biol*. 2005; 293:151–166. [PubMed: 16028418]
 19. Palinauskas V, Dolnik OV, Valkiunas G, et al. Laser microdissection microscopy and single cell PCR of avian hemosporidians. *J Parasitol*. 2009; 96:420–424. [PubMed: 20001098]
 20. Schermelleh L, Thalhammer S, Heckl W, et al. Laser microdissection and laser pressure catapulting for the generation of chromosome-specific paint probes. *Biotechniques*. 1999; 27:362–367. [PubMed: 10457845]
 21. Schutze K, Posl H, Lahr G. Laser micromanipulation systems as universal tools in cellular and molecular biology and in medicine. *Cell Mol Biol (Noisy-le-grand)*. 1998; 44:735–746. [PubMed: 9764744]
 22. Espina V, Edmiston KH, Heiby M, et al. A portrait of tissue phosphoprotein stability in the clinical tissue procurement process. *Mol Cell Proteomics*. 2008; 7:1998–2018. [PubMed: 18667411]
 23. Espina V, Mueller C, Edmiston K, et al. Tissue is alive: New technologies are needed to address the problems of protein biomarker pre-analytical variability. *Proteomics Clin Appl*. 2009; 3:874–882. [PubMed: 20871745]
 24. Lim MD, Dickherber A, Compton CC. Before you analyze a human specimen, think quality, variability, and bias. *Anal Chem*. 2011; 83:8–13. [PubMed: 21114268]
 25. Becker KF, Schott C, Hipp S, et al. Quantitative protein analysis from formalin-fixed tissues: implications for translational clinical research and nanoscale molecular diagnosis. *J Pathol*. 2007; 211:370–378. [PubMed: 17133373]
 26. Fox CH, Johnson FB, Whiting J, et al. Formaldehyde fixation. *J Histochem Cytochem*. 1985; 33:845–853. [PubMed: 3894502]
 27. Helander KG. Kinetic studies of formaldehyde binding in tissue. *Biotech Histochem*. 1994; 69:177–179. [PubMed: 8068812]
 28. Srinivasan M, Sedmak D, Jewell S. Effect of fixatives and tissue processing on the content and integrity of nucleic acids. *Am J Pathol*. 2002; 161:1961–1971. [PubMed: 12466110]
 29. Mueller C, Liotta LA, Espina V. Reverse phase protein microarrays advance to use in clinical trials. *Mol Oncol*. 2010; 4:461–481. [PubMed: 20974554]
 30. Geiss GK, Bumgarner RE, Birditt B, et al. Direct multiplexed measurement of gene expression with color-coded probe pairs. *Nat Biotechnol*. 2008; 26:317–325. [PubMed: 18278033]
 31. Botling J, Edlund K, Segersten U, et al. Impact of Thawing on RNA Integrity and Gene Expression Analysis in Fresh Frozen Tissue. *Diagn Mol Pathol*. 2009

32. Micke P, Ohshima M, Tahmasebpour S, et al. Biobanking of fresh frozen tissue: RNA is stable in nonfixed surgical specimens. *Lab Invest.* 2006; 86:202–211. [PubMed: 16402036]
33. Xiang CC, Mezey E, Chen M, et al. Using DSP, a reversible cross-linker, to fix tissue sections for immunostaining, microdissection and expression profiling. *Nucleic Acids Res.* 2004; 32:e185. [PubMed: 15604454]
34. Espina V, Wulfschlegel JD, Calvert VS, et al. Laser-capture microdissection. *Nat Protoc.* 2006; 1:586–603. [PubMed: 17406286]
35. Golubeva Y, Rogers K. Collection and preparation of rodent tissue samples for histopathological and molecular studies in carcinogenesis. *Methods Mol Biol.* 2009; 511:3–60. [PubMed: 19347291]
36. Arcturus^{XT} user guide: Flexible and Modular Laser Capture Microdissection, Rev. A. Applied Biosystems/Life Technologies; Foster City, CA: 2007.
37. v.20091113. NanoString Technologies, Inc; Seattle, WA: 2009. NanoString Technical note: Sample Flexibility for Gene Expression Analysis with the nCounterTM Analysis System.
38. Iyer EP, Cox DN. Laser capture microdissection of *Drosophila* peripheral neurons. *J Vis Exp.* 2010
39. Kiernan, J. *Histological and Histochemical Methods.* 4th ed. Scion; Oxfordshire: 2008. Histological staining in one or two colours; p. 146-147.
40. Buckanovich RJ, Sasaroli D, O'Brien-Jenkins A, et al. Use of immuno-LCM to identify the in situ expression profile of cellular constituents of the tumor microenvironment. *Cancer Biol Ther.* 2006; 5:635–642. [PubMed: 16627987]
41. Agar NS, Halliday GM, Barnetson RS, et al. A novel technique for the examination of skin biopsies by laser capture microdissection. *J Cutan Pathol.* 2003; 30:265–270. [PubMed: 12680959]
42. Luzzi VI, Holtschlag V, Watson MA. Gene expression profiling of primary tumor cell populations using laser capture microdissection, RNA transcript amplification, and GeneChip microarrays. *Methods Mol Biol.* 2005; 293:187–207. [PubMed: 16028420]

Figure 1. Arcturus^{XT} software for microdissection

A. The Arcturus^{XT} instrument operates via a series of software tool panes and option buttons. Each step of the microdissection process can be accessed from the main software screen. Mouse bony ear tissue is shown at 2x magnification in the live image, and an overview of the tissue slide is shown below. The overview image allows the operator to navigate across the slide and/or tissue section. B. Properly melted spots have a dark outer ring and a clear center, indicating that the polymer has melted and is in direct contact with the slide. Inadequate power and/or duration settings create spots with a hazy appearance, lacking a distinct black ring (live image shown as 60x magnification). The area to be microdissected can be measured to match the microdissection area to the cell size.

Figure 2. Arcturus^{XT} Infrared Laser Capture Microdissection process

A. A section of mouse inner ear tissue is mounted and stained on a glass slide and viewed under high magnification (40x) to locate the ciliated epithelium. The ciliated epithelium are identified visually and designated for microdissection using software annotation tools at 2x, 10x, 40x, or 60x magnification. The infrared laser locally expands a thermoplastic polymer that captures the cell(s) only in the vicinity of the laser pulse. B. Microdissection occurs when the film is lifted from the tissue section, shearing the selected cells from the tissue section. The presence of stained material inside the melted margins of the polymer indicates effective microdissection (10x).

Figure 3. Mouse bony ear tissue adequately preserved for laser capture microdissection
Bony inner ear tissue, preserved in a novel protein/phosphoprotein preservative (22, 23) and paraffin embedded, shows retention of morphology adequate for microdissection of inner ear ciliated epithelium.

Figure 4. Sectioning approach for LCM slide preparation (large scale target collection for NanoString analysis)

A. Aperio digital image of H&E slide with annotated papilloma (Aperio ‘fit’ magnification).
B. FFPE block trimming template (Aperio 4x magnification). C. View of FFPE block trimmed according to the template. D. Serial sections mounted on PET membrane slide. Target area in A, B and D is indicated by a rectangular.

Figure 5. LCM target dissection and collection for NanoString analysis with ‘Auto New Cap’ set-up (Aperio 4x magnification)

A. View of the PET membrane after collection of the first serial target. B. mmi IsolationCap® with the first collected target (a) positioned over the second target (b) with laser cutting path (open line). C. mmi IsolationCap® with the first (a) and second (b) collected targets is positioned over the third (c) target with laser cutting path (open line). D. View of mmi IsolationCap® with the first (a), the second (b) and the third (c) overlapped targets.

Figure 6. Mouse skin papilloma dissection based on pathology annotation of Aperio digital image of the reference H&E slide

A. Digital image of H&E reference slide with pathology annotation (black line) (Aperio 4x magnification). B. View of LCM target on the dissecting screen with the reflected drawing of pathology annotation (white line) (mmi CellCut® objective 4x, mmi IsolationCap® is in 'down' position). C. View of the PET membrane after target collection (mmi CellCut® objective 4x, mmi IsolationCap® is in a down position).

Figure 7. Digital image manipulation with ImageScope™ digital slide view prior to LCM session
 A. Digital image of annotated H&E section in ImageScope™ Main Window (Aperio 0.4x magnification) (a) zoom slider (b) slide label window (c) rotation tool (d) thumbnail window (e) magnifier window with LCM target. B. Orientation of LCM target (dissected) on LCM dissecting screen (mmi CellCut® objective 4x, mmi IsolationCap® is in 'up' position). C. Rotated digital image matches the orientation of LCM section on dissecting screen during laser cutting. D. Enlarged H&E image of the target (Aperio 4x magnification) matches the magnification of LCM target on dissecting screen.

Figure 8. NanoString expression results for selected interleukin genes in murine skin papillomas

The expression of several interleukins associated with inflammatory response was investigated in murine papillomas developed post DMBA/TPA treatment. The data was normalized using *GAPDH* as housekeeping gene (A) or using the mean of the three least variable housekeeping genes (*GAPDH*, *RPI30*, *RPI9*) B. The results of each normalization approach are very similar. Each circle corresponds to gene expression in a single mouse. Mean is indicated by a solid line.

The targeted region and corresponding sequence of interleukin genes selected for NanoString Reporter Code and Capture Probe set.

(*RP19*, *RP130*, *RP19*, *GAPDH*, and *HPRT* were used as housekeeping genes).

Gene	Accession	Targeted Region	Target Sequence
Il10	NM_010548.1	985-1085	GGGCCCTTTGGCTATGGTGTCTCTTTCAAATTGCTCTCATCCCTGAGTTCAGAGCTCTAAGAGAGATTGTGAAGAAACTCATGGGTCTTTGGGAAGAGAAACCA
Il11	NM_008350.2	285-385	GCGCTGGGACATTGGGATCTTTGCAGCTTCTCTGGTGTCTGACAAGGCTTCGAGTAGACTTGATGTCTCTACCTCCGGCATGTACAAATGGCTGCGCGTGC
Il12a	NM_008351.1	355-455	TCATGAAGACATCACACGGGACCAAAACCAGCACATTGAAGACCTGTTTACCAGTGGAACACAAAGAACGAGAGTTGCCTGGCTACTAGAGAGACTTCT
Il12b	NM_008352.1	1045-1145	TCGTAGAGAAAGACATCTACCGAAGTCCAATGC AAAAGCGGGAATGTCGTGTCAAGCTCAGGATCGCTATTACAAATTCCTCTGTGCAGCAAGTGGGCATG
Il13	NM_008355.2	425-525	AGCTACACAAAGCAACTGTTTCGCCACGGCCCTTCTAATGAGGAGAGACCATCCCTGGGCATCTCAGCTGTGGACTCATTTTCTCTTCTCACATCAGAC
Il15	NM_008357.1	205-305	GTGTTTGGAAAGGCTGAGTTCACATCTAACAGCTCAGAGAGGTCAGGAAAGAATCCACCTTGACACATGGCCCTCTGGCTCTTCAAAGCACTGCCTCTTC
Il17a	NM_010552.3	205-305	ACCTCAAAGTCTTAACTCCCTTGGCGCAAAAAGTAGCTCCAGAAAGCCCTCAGACTACCTCAAACCGTTCCACGTCACCCCTGGACTCTCCACCCGCAATGA
il17e	NM_080729.2	649-749	TTGGGGAGAAACTCTGACTTTTGC ACTTTTGGAAAGCACTTTTGGGAAAGGAGCAGGTTCCGCTTGTGCTCTAGAGGATGCTGTGTGGCATTTTCTACTC
Il18	NM_008360.1	270-370	ACTTTGGCCGACTTCACCTGTACAAACCGCAGTAATACGGAAATATAATGACC AAGTTCTCTTCGTTGACAAAAGACAGCCTGTGTCGAGGATATGACTGA
Il18bp	NM_010531.1	585-685	TTTGTTTGTGGATCCTCGGACAAAGTGGCCCAAGTATCACATCATTTCTGGCCCAAGTCTTGGGATGCGTTGAAAGACAGCTCCGTCCTTCTCAAAGAAAACCCCTC
Il1a	NM_010554.4	225-325	ACCTCTGAAAACGTCAAAAGATGTCCAACCTTCACCTTCAAAGGAGAGCGCGGTGACAGTATCAGCAACGTCAAGCAACGGGAAGATTCTGAAGAAAGAGACGGC
Il1b	NM_008361.3 NM_00102560	1120-1220	GTTGATTCAAAGGGGACATTAGGCAGCCTCTCTAGAACAGAACCTAGCTGTCACGTTGGGGGATGAATTGGTCATAGCCCGCACTGAGGTCTTTTCATT
Il1rl1	2.2	815-915	ATAGGAAAACAGCAAGTATTGCCCTGTTCAGCTTGTCTTGGCAAAAGCTCTCACCTTCTGGCTGATGTCCTGTGGCAGATTAAACAAACAGTAGTTGGAA
Il1rm	NM_031167.5	224-324	CAACGAGCTCATTTGCTGGGTACTTACAAGGACCAAAATACAAACTAGAAGAAAAGATAGACATGGTGCCTATTGACCTTCATAGTGTGTTCTTGGGCATC
Il21	NM_021782.2	1762-1862	ATGGCTGGGGATGGTTTTGATCTAAAGGAAAAGGTGTCTGTACCTCACAGTGCCTTTAAACAAGCAGAGATCCCGCTTACCGCCCTAAGATAGCACAA
Il22	NM_016971.1	477-577	AGAAGAAATGTCAGAAAGGCTGAAGGAGACAGTGA AAAAGCTTGGAGAGAGTGAGAGATCAAAGCGCATGGGGAACCTGGACCTGCTGTTTATGTCTGAG
Il23a	NM_031252.1	360-460	CAAGGACAACAGCCAGTTCTGCTTGGCAAGGATCCGCCAAGGTCTGGCTTTTATAAGCACCTGTTGACTCTGACATCTTCAAAAGGGGAGCCTGCTCTA
Il27 ebi3	NM_015766.2	1015-1115	ATGTACTGGGCTGCCGAAAGCACTGGGATAATTCACTTGACTTCTTTCAGACCTCAAATTTCCAAACCCCTGTGGGATGATCTTTCTTCTCTGCCGGTGC
Il27p28	NM_145636.1	175-275	GCTATGTCCACAGCTTTGCTGAATCTCGATTGCCAGGAGTGAACCTGGACCTCCTGCCCTGGGATACCATCTTCCCAATGTTTCCCTGACTTTC
il33	NM_133775.1	1011-1111	ATGACTTACGGCGTTGGTAAAGAAAAGTGAAGGAGATTCAGCCTTGCTTTCCTTTCTCTGCTTGAGTCCCTGTATGAAATCACACTCACGGACTTCAG
Il4	NM_021283.1	345-445	TGCTTGAAGAAAGAACTCTAGTGTTCATGGAAGCTGCAAGACTCTTTTCGGGCTTTTCGATGCCCTGGATTTCATCGATAAGCTGCACCATGAAATGAGTCCA
Il6	NM_031168.1	40-140	CTCTCTGCAAGAGACTTCCATCCAGTTGCCCTTCTTGGGACTGATGCTGGTGACAACCAACGGCTTCCCTACTTCAACAGTCCGGAGAGGAGACTTTCACAG
Rpl27	NM_011289.3	2-102	GCTGCTGCTGCTGCTAATAAAGCTTGGTTCAAAAACCGAGTGGCCGACCCCTCCCTGAGGAGCTCTTCCGCCAACAACTTTGACCAACCCGTTGGGCA
Rpl30	NM_009083.4	14-114	CCCCGGCCAATCTTGGCGGCTGTGTGTGAGTGAGCTTGTGGGGGTAACCGATTAGGGGGCTCGGGGAGCTCCGCTAGCTGTGTGTTTGACGCTCTGGAT
Rpl9	NM_011292.2 NM_00100130	240-340	GGTAAACAGAAAAGGAACTGGCCACCGTCAGGACCATCTGCAAGTATGTTTCAAGATGATCAAGGTGTCAACGCTGGGCTTCCGATACAAAGATGCGGTCTG
Gapdh	3.1	890-990	AGGTTGTCTCTGCGACTTCACAAGCAACTCCCACTTTCACCTTCGATGCCGGGGCTGGCATTTGCTCTCAATGACAACACTTTGTCAAAGCTCATTTTCCTG
Hpri	NM_013556.2	30-130	TGCTGAGCGCGCGGAGAGCGTTGGGCTTACCTCACTGCTTTCGAGCGGTAGCACCTCTCCGCGGCTTCCCTCCAGACCGCTTTTTCGCCGCGA

Table 2

NanoString nCounter™ Report for LCM papilloma samples.

File Attributes						
File name	20110224					
Sample ID	Pap1330	Pap1332	Pap338-1	Pap338-2	Pap339-2	
Owner						
Sample Date	2/24/2011	2/24/2011	2/24/2011	2/24/2011	2/24/2011	
File Version	1.6	1.6	1.6	1.6	1.6	
GeneRLF						
Comments						
Lane Attributes						
Lane ID	1	2	3	4	5	
FOV Count	600	600	600	600	600	
FOV Counted	600	600	598	600	597	
Scanner ID	DA36	DA36	DA36	DA36	DA36	
StagePosition	2	2	2	2	2	
Binding Density	0.15	0.16	0.23	0.17	0.2	
Messages						
Reporter Counts						
Code Class	Name	Accession				
Positive	POS_A(128)	ERCC_00117.1	6196	6755	6607	7530 7786
Positive	POS_B(32)	ERCC_00112.1	1619	1897	1787	1980 2029
Positive	POS_C(8)	ERCC_00002.1	331	460	370	420 444
Positive	POS_D(2)	ERCC_00092.1	109	136	97	143 135
Positive	POS_E(0.5)	ERCC_00035.1	16	27	25	28 20
Positive	POS_F(0.125)	ERCC_00034.1	10	15	18	20 18
Negative	NEG_A(0)	ERCC_00096.1	3	1	1	1 0
Negative	NEG_B(0)	ERCC_00041.1	2	1	4	1 2
Negative	NEG_C(0)	ERCC_00019.1	3	5	2	1 3
Negative	NEG_D(0)	ERCC_00076.1	2	1	2	0 2
Negative	NEG_E(0)	ERCC_00098.1	2	2	1	0 3
Negative	NEG_F(0)	ERCC_00126.1	2	4	6	4 5
Negative	NEG_G(0)	ERCC_00144.1	0	0	2	2 0
Negative	NEG_H(0)	ERCC_00154.1	4	2	2	0 1