Research Article

Threshold Levels of Dual Electrode Stimulation in Cochlear Implants

JORIEN SNEL-BONGERS,¹ JEROEN J. BRIAIRE,¹ ERIKA H. VAN DER VEEN,¹ RANDY K. KALKMAN,¹
AND JOHAN H. M. FRIJNS¹

¹ENT-department, Leiden University Medical Centre, PO box 9600, 2300 RC Leiden, The Netherlands

Received: 10 April 2012; Accepted: 29 April 2013; Online publication: 22 May 2013

ABSTRACT

Simultaneous stimulation on two contacts (current steering) creates intermediate pitches between the physical contacts in cochlear implants. All recent studies on current steering have focused on Most Comfortable Loudness levels and not at low stimulation levels. This study investigates the efficacy of dual electrode stimulation at lower levels, thereby focusing on the requirements to correct for threshold variations. With a current steered signal, threshold levels were determined on 4 different electrode pairs for 7 different current steering coefficients (α). This was done psychophysically in twelve postlingually deafened cochlear implant (HiRes90K, HiFocus1J) users and, in a computer model, which made use of three different neural morphologies. The analysis on the psychophysical data taking all subjects into account showed that in all conditions there was no significant difference between the threshold level of the physical contacts and the intermediate created percepts, eliminating the need for current corrections at these very low levels. The model data showed unexpected drops in threshold in the middle of the two physical contacts (both contacts equal current). Results consistent with this prediction were obtained for a subset of 5 subjects for the apical pair with wider spacing (2.2 mm). Further analysis showed that this decrease was only observed in subjects with a long duration of deafness. For current steering on adjacent contacts, the results from the psychophysical experiments were

in line with the results from computational modelling. However, the dip in the threshold profile could only be replicated in the computational model with surviving peripheral processes without an unmyelinated terminal. On the basis of this result, we put forward that the majority of the surviving spiral ganglion cells in the cochlea in humans with a long duration of deafness still retain peripheral processes, but have lost their unmyelinated terminals.

Keywords: current steering, computational model, peripheral processes, unmyelinated terminal

INTRODUCTION

Cochlear implants (CIs) are widely used as treatment for profoundly hearing impaired children and adults. The main goal of a CI is to recover a perception of sound and, in particular, speech information. However, implanted adults and children are still experiencing problems in challenging listening situations, such as background noise or listening to music. One reason for this might be reduced spectral resolution caused by the limited number of physical electrodes in the array (Brendel et al. 2009; Hughes and Goulson 2011). Dual electrode stimulation (DES), either simultaneous (Donaldson et al. 2005; Townshend et al. 1987) or sequential (Kwon and van den Honert 2006; McDermott and McKay 1994), has been proposed as a method to increase the number of perceivable pitches with CIs. In simultaneous DES (also called current steering), the summation of two electrical fields produces excitation of nerve fibers intermediate to the stimulated contacts. Therefore it

Correspondence to: Jeroen J. Briaire • ENT-department • Leiden University Medical Centre • PO box 9600, 2300 RC Leiden, The Netherlands. Telephone: +31-715262819; fax: +31-715248201; email: j.j.briaire@lumc.nl

can lead to generation of additional pitch percepts beyond those generated by stimulation of a single contact. The percept is controlled by a current steering coefficient α . This coefficient is defined as the fraction of the total current delivered through the more basal contact of the pair. Therefore, $\alpha=0$ denotes stimulation of the apical contact only and $\alpha=1$ pure stimulation of the basal one. The number of intermediate percepts thus created along the whole array varies among CII/HiRes 90K implant users (ranging from 8 to 466 percepts) and between studies (the average ranging from 20 to 93 percepts; Firszt et al. 2007; Koch et al. 2007; Snel-Bongers et al. 2012).

The perceived pitch depends on the neural excitation pattern induced by the DES. Frijns et al. (2009b) examined the neural excitation pattern for both simultaneous and sequential DES, both psychophysically and using computational modeling. For current levels close to Most Comfortable Loudness level (MCL), they found that the excitation area moves smoothly between the contacts as a function of α producing an almost constant loudness (Figure 1C, D). However, the computational model also showed that, in some situations, there was discontinuous stimulation for the different α values, especially at the lowest stimulation levels (Figure 1A, B). Similar effects were found while evaluating DES in animals using neural recordings (Bonham and Litvak 2008). Consequently, the centre of excitation jumped from one electrode contact region to the other. Both our model and the physiological studies suggest that at a constant stimulation level the number of activated neurons changes for different α values. To stimulate at threshold level (TL) for various α values, then this requires a current level correction during DES.

The human cochlea model developed at Leiden University Medical Center has been used to evaluate effects of intra-cochlear position (Frijns et al. 2001), neural degeneration (Briaire and Frijns 2006) and effects of multipolar stimulation (Frijns et al. 2011) on thresholds and spread of excitation. The realistic tapering structure of the cochlear model allows for

the evaluation of excitation differences between the base and the apex of the cochlea. In the model, TL has been shown to depend on variables such as the position of the electrode contacts in the scala tympani (i.e. lateral or peri-modiolar positions) and the presence or degeneration of the peripheral axonal processes. The model consists of two parts, a volume conduction part simulating the current flow through an implanted cochlea and a neural part simulating the response of the nerve fibers (Briaire and Frijns 2000, 2005, 2006; Frijns et al. 2001, 2009a, b). In the present study, this model will be used to evaluate the behavior of the threshold under different current steering conditions.

All recent studies on current steering have focused on MCL level and not on TL. When current steering is used in speech coding strategies, it is of interest to know what happens at TL. Firstly, because sounds are presented at levels ranging from threshold to MCL level and not only at MCL level. If large corrections were needed then if, for example, one bit of the frequency spectrum were dominated by a component whose frequency caused it to be sent entirely to one electrode, and another region had a frequency content that led to current sharing, this could distort the representation of the frequency spectrum. In addition, TL is part of the clinical fitting procedure. At MCL, a current correction is needed to maintain constant loudness for non-simultaneous DES on adjacent electrodes, while it is not for most simultaneous DES (Donaldson et al. 2005; Frijns et al. 2009b; Snel-Bongers et al. 2011; Snel-Bongers et al. 2012). Based on the findings with our computational model described above (Figure 1A, B), it was hypothesized that such a current correction is also needed when current steering is employed close to TL. This was evaluated in the present study by computational modeling and with a psychophysical experiment. Experiments were conducted at two locations along the cochlea (apical and basal), for two adjacent contacts and for two non-adjacent contacts with one electrode in between (Snel-Bongers et al. 2011).

FIG. 1. Current steering plots for AS contacts in a HiFocus 1 electrode (outer wall position) calculated by Frijns et al. (2009b). The nerve fibers exhibit a peripheral process without an unmyelinated terminal (UT). The ordinate axis denotes the associated place pitch of

the fibers according to the Greenwood (GW) map. The excitation area is shaded grey and its center is indicated by a dashed line. Panes from left to right show results for stimulation levels from near threshold (-9 dB re MCL) to MCL.

METHOD

Subjects

Twelve postlingually deafened adults implanted at LUMC in 2007 with a HiRes 90K HiFocus-1JTM cochlear implant (Advanced Bionics, Sylmar, CA) took part. All subjects had at least 1 year experience with their implant and a phoneme score of at least 70 %, measured at 65 dB SPL with the standard Dutch monosyllabic word speech test (Bosman and Smoorenburg 1995). Specific subject information is provided in Table 1. The study was approved by the Leiden University Medical Center Ethics Committee under number P02.106.L. Written informed consent was obtained from the participants.

Assessment of electrode position

The position of the electrode array and, thereby, the electrode contacts, was determined from post-operative CT scans which are a routine part of the clinical CI program. To measure the exact position of the electrode contacts, a multiplanar reconstruction (MPR) was made (Verbist et al. 2005). A system of coordinates, according to an international consensus (Verbist et al. 2010), was entered in the postoperative MPR, using a custom Matlab application (MathWorks, Natick, MA) as described in a previous paper (Snel-Bongers et al. 2011). All electrode contacts were marked by an experienced physician. In line with the consensus, the angles used in this study were calculated with the round window as the 0° reference.

For each subject the electrode contact closest to 360° (apical site, AS) and the one closest to 180° (basal site, BS) were selected on the electrode array (Table 1). Current steering was applied between the

contact at the reference angles e (360° and 180°), and the two contacts in the basal direction ($e+1$ and $e+2$), creating two pairs with a spacing of one and two contacts (1.1 and 2.2 mm), respectively. Throughout this paper, the pairs consisting of electrode contacts (e and $e+1$) and (e and $e+2$) will be referred to as ‘pair $e+1$ ’ and ‘pair $e+2$ ’, respectively.

The CT scan was further used to determine the distance from the electrode contacts to the medial wall of the cochlea. After locating each electrode contact individually, a line was generated from the contact to the center of the modiolus. After identifying the location of the medial wall, the distance was calculated along this line (Snel-Bongers et al. 2012).

Computer model

For the modeling part of this study, a computational model of the implanted human cochlea, which has been developed over the years at Leiden University Medical Center, was used (Briaire and Frijns 2005; Frijns et al. 2001; Frijns et al. 2011; Frijns et al. 2009a). The first part of the model consists of a volume-conduction model that calculates electrical potentials at the auditory nerve fibers located in a realistic three-dimensional representation of a human cochlea implanted with a geometrically accurate representation of the HiFocus electrode array (Advanced Bionics, Sylmar, CA, USA). Secondly, an active nerve fiber model simulates neural responses to specific stimulation patterns produced by the cochlear implant, using the calculated potentials at the location of the nodes of Ranvier of the primary auditory nerve fibers. Neural excitation profiles were generated at the same electrode locations and with the same stimulus configurations as investigated in the subject

TABLE 1
Subject demographics

	Gender	Age (years)	Duration of deafness (years)	CI side	CI usage (months)	CVC Ph (%)	Electrode used		Aetiology
							AS	BS	
S1	Male	64	23	Right	23	89	4	11	Otosclerosis
S2	Female	69	36	Right	52	81	5	11	Meningitis
S3	Male	60	4	Left	52	92	6	13	Medication
S4	Female	57	27	Right	44	83	3	11	Unknown
S5	Female	58	42	Right	33	87	2	9	Unknown
S6	Female	77	12	Right	52	72	5	11	Ménière's disease
S7	Male	55	4	Right	61	84	6	13	Loudness
S8	Female	44	36	Right	20	92	5	11	Unknown
S9	Male	71	26	Left	29	84	4	11	Familial Progressive
S10	Male	50	6	Left	38	96	4	12	Otosclerosis
S11	Female	72	30	Right	37	86	4	11	Familial Progressive
S12	Male	53	43	Left	32	89	2	9	Unknown
Mean		61	24		39	86			

Speech perception scores are given as percentage phonemes correct (Ph %) in phonetically balanced monosyllabic (CVC) words

group, except that the current steering parameter (α) ranged from 0 to 1 in increments of 0.1. Three sets of 320 nerve fibers with different stages of degeneration were created, based on the morphology with an unmyelinated cell body as described in Briaire and Frijns (2005). The first set contained intact neurons with an unmyelinated terminal (UT) added at the tip of the peripheral process (Figure 2). Based on earlier research the length of this UT was set to 10 μm (Liberman and Oliver 1984; Parkins and Colombo 1987). Previous research (Kujawa and Liberman 2009; Lin et al. 2011) demonstrated that the UT is the first part that degenerates after acoustic overexposure with moderate hearing loss as a consequence. The second set therefore, simulates an initial stage of degeneration with peripheral processes, but now without an UT, replacing it by a 1 μm node. The third set consisted of neurons in a further stage of degeneration, without peripheral processes but with their cell bodies and central axons still intact (Briaire and Frijns 2006). The undegenerated peripheral processes followed non-radial trajectories based on data from Stakhovskaya et al. (2007) as described in Frijns et al. (2009b).

From the excitation profiles, it was possible to determine how the width of the excitation area depends on the stimulus level. This excitation width is expressed as the length spanned by the tips of the peripheral processes, in other words, the equivalent length along the basilar membrane that corresponds to the neural activation. Based on this loudness definition equivalent loudness was considered to be the equivalent width of excitation in mm along the basilar membrane. Current levels were determined for total basilar membrane excitation lengths ranging from 0 (i.e. absolute threshold) to 4 mm. The value of 4 mm excitation was considered to correspond to maximum comfortable loudness (Briaire and Frijns 2006). Throughout the manuscript two definitions for TL were defined: the minimum amount of current needed to excite at least one fiber ('0 mm width'; Briaire and Frijns 2006) called 1st fiber threshold ($\text{TL}_{0\text{mm}}$) and a model equivalent of the perceptual

threshold ($\text{TL}_{1\text{mm}}$) where 1 mm of the length along the basilar membrane is excited.

Psychophysical experiment

The main outcome for this experiment was the TL_{Φ} , when the signal was just heard, at various values of the current steering coefficient α . The experiment was performed using the Bionic Ear Data Collection System (BEDCS, Advanced Bionics, USA) for the stimulation configuration and using the PsychoAcoustic Test Suite (PACTS, Advanced Bionics, Belgium) for the psychoacoustic tests. Stimuli were 300 ms bursts of symmetric biphasic pulses with phase duration of 32 μs and 1,400 pulses per second. The inter-stimulus interval was 500 ms. Dual electrode stimuli were always simultaneous.

Rough TLs and most comfortable loudness levels were first determined for each of the six pre-selected electrode contacts individually. The subject was asked to indicate when the signal was just heard (rough TL) and also when the signal sounded most comfortably loud (MCL).

A single run of a 2-Alternative-Forced-Choice, 1-up/2-down adaptive procedure was used to determine TL_{Φ} (Levitt 1971) for each $\alpha=0, 0.17, 0.33, 0.50, 0.67, 0.83$ or 1 for all four electrode pairs. The target stimulus consisted of a dual-electrode stimulus at a current level between MCL and TL. The reference stimulus was 0 mA. Within each trial, the two stimuli were presented in a random order. The subject was required to select the stimulus that generated a sound percept. A response was considered to be correct when the subject chose the target stimulus. The current level of the target stimulus was decreased until the subject chose the wrong interval because no sound was heard for either stimulus. The procedure covered ten reversals (i.e. changes in the direction of the signal level), where the test outcome was calculated over the last six reversals. The current level was altered in steps of 15 % increments or decrements for the first four reversals and in steps of 7 % for the remaining six. If a downward or upward trend was detected over the last six reversal points by the

FIG. 2. Representation of the nerve fiber morphology. The peripheral process consists of six scalable segments to adjust for the variable length from the organ of Corti to the cell body, with the unmyelinated terminal (UT) at the end.

program (PACTS), determined using an algebraic algorithm (website 2009), then the test was extended, assuming that either the adaptive procedure had not yet converged to the subject's discrimination limit or that the measurement was unreliable due to the behavioral status of the subject, for example loss of attention (Snel-Bongers et al. 2011).

To correct for the variation in the absolute thresholds between all subjects and the four stimulated positions, the data were normalized. TL for each α was normalized on both ends ($\alpha=0$ (TL_0) and $\alpha=1$ (TL_1)) with linear interpolation in between, using the following formula:

$$TL_{\text{normalized}} = TL_{\alpha} / (TL_0 + \alpha(TL_1 - TL_0)) \quad (1)$$

where TL_{α} is the TL for that specific α ($0 \leq \alpha \leq 1$).

As a consequence, all threshold corrections are calculated relative to $TL_{\text{normalized}}$, and do not reflect the absolute current level.

Statistical analysis

For the threshold detection experiments a linear mixed model analysis (Fitzmaurice et al. 2004) was performed on the normalized data. A linear mixed model can take several data points from one subject into account, and gives an overview of the interaction of all these data with the main values, the comparison of all data with one data point and corrects for the individual subject (Snel-Bongers et al. 2012). Further, the linear mixed model corrects for missing data. A Bonferroni correction was applied to all statistical tests, to correct for the fact that (up to five) different tests were performed on the same data, resulting in a significance criterion of $p < 0.01$.

All data were analyzed using the SPSS 17 (Statistical Package for the Social Sciences, SPSS inc., Chicago, IL) statistics software package.

RESULTS

Electrode position

The mean location of the apical and basal electrode sites, measured from the round window, was at 364° ($\pm 11^\circ$) and 180° ($\pm 7^\circ$), respectively. The apical electrode contacts were significantly closer ($p=0.026$) to the medial wall (1.19 mm $\pm$ 0.11) than the basal contacts (1.33 mm $\pm$ 0.12).

Computer model

Figure 3 shows normalized current levels needed to achieve 0 ($TL_{0\text{mm}}$) and 1 mm ($TL_{1\text{mm}}$) excitation

along the basilar membrane in a cochlear model with either intact neurons (SG cells with a peripheral process) with an UT (3A), without an UT (3B, No UT) and a degenerated peripheral process (3C, DPP), for current steered and spanning pairs ($e+1$: open vs. $e+2$: filled symbols) at AS and BS (AS: triangles vs. BS: circles). Generally speaking, the $e+1$ pair needs less current correction than the $e+2$ pair and the apical pair needs more current correction than the basal pair, regardless of the threshold criterion. Further, more current correction is needed in the fiber model without UT in comparison with the other two fiber degeneration states.

In Figure 3B1, it is visible that apically more current correction is needed to reach $TL_{0\text{mm}}$ than basally. The curve for $e+1$ at BS is almost flat, while the others peak more or less around $\alpha=0.5$. The basal electrodes have the same pattern for the $TL_{1\text{mm}}$ condition (Figure 3B2). However, for the apical sites current levels for 1.0 mm show a clear dip of 60 % relative to the $TL_{0\text{mm}}$ condition. Interestingly, in the case of the dip, the calculated excitation profiles showed that initial excitation takes place at the peripheral processes, while it occurs in the central axons in most other cases.

This reduced TL around $\alpha=0.5$ can also be seen in Figure 1, which was generated with nerve fibers with peripheral processes, but without an UT. From Figure 1A (for stimuli 9 dB below MCL), it is clear there is a dependence of the number of excited fibers on α , even resulting in an absence of excited fibers at the center ($\alpha=0.5$). This indicates that the current level needs to be varied (increased) as a function of α if a constant number of excited fibers has to be maintained regardless of α . In Figure 1B, the situation for a 3 dB higher current level is shown. The center of the excitation now consists of two regions (one around each individual contact). For $\alpha=0.5$, both regions co-exist and lead to a higher number of excited nerve fibers (i.e., increased loudness with the same amount of current). As a result, the current to maintain constant loudness over the whole range of α from 0 to 1 shows a minimum at $\alpha=0.5$.

Figure 3A, C shows the plots for the other two neural conditions (with UT and without peripheral process, respectively). These plots are not essentially different for the basal pairs. For the apical pairs however for $TL_{0\text{mm}}$, particularly pair $e+1$, less current correction is needed to reach TL (Figure 3A1, C1) in comparison with the graphs in Figure 3B1. Contrary to Figure 3B2, A2, C2 do not show a dip for apical electrode pairs.

For both cases (Figure 3A, C), the excitation region will be a single area that shifts gradually between e and $e+x$. However, the excited region does decrease in size as α approaches 0.5 for low current levels. In line with this, the required stimulus levels are elevated for pair $e+1$ by approximate 10 % around $\alpha=0.5$ for both TL

FIG. 3. Computational modeling results of the loudness balancing experiment for the model with a peripheral process with an UT (A), a peripheral process without UT (B) and without peripheral processes (C). The circles represent the data at the AS of the array and the triangles data at the BS of the array, where the open symbols are pair

$e+1$ and the filled symbols pair $e+2$. The data were normalized on both ends ($\alpha=0$ and $\alpha=1$; Eq. 1) and the other currents (I_{TL}) are expressed in percentages of this values. Panes on the left show results for excitation lengths of 0 and of 1 mm on the right.

criteria (first fiber, or on the basis of a distinct width of the excited region). With elevation up to 65 %, this effect is larger for pair $e+2$ than for adjacent electrode contacts.

Psychophysical results

In Figure 4, the individual psychophysical results, normalized on both the apical and basal electrode of the pair (Eq. 1), are shown for the two different positions in the cochlea (AS: circles vs. BS: triangles) and both measured electrode pairs ($e+1$: open vs. $e+2$: filled symbols). The average results are shown in Figure 5. Linear mixed modeling performed on the data excluding $\alpha=0.0$ and $\alpha=1.0$ showed that there is no difference between the curves for the four electrode pairs ($F_{(221)}=1,822$, $p=0.145$). The average curves suggest (Figure 5) that the pairs $e+2$ need on average more current, correction like in the computer model (Figure 3), but the $e+1$ (101.29 %) and $e+2$ (104.26 %) pairs were not significantly different ($F_{(223)}=1,969$, $p=0.144$) from each other. Also the

position in the cochlea had no significant ($F_{(223)}=1,969$, $p=0.162$) effect on TLs (AS, 101.35 % and BS: 104.20 %), which is in line with the model predictions, which use 1 mm excitation as threshold criterion. No relation was found between TL_{Φ} and the distance to the medial wall ($r^2_{(24)}=-0.072$, $p=0.737$).

Contrary to some of the model predictions, the TL_{Φ} for intermediate values of α are not significantly different ($F_{(265)}=0.774$, $p=0.569$) from the flanking electrodes ($\alpha=0$ and $\alpha=1$), as demonstrated with the linear mixed model run on all data together. In line with this result, TL_{Φ} found with $\alpha=0.5$ on both adjacent (pair $e+1$) (103.55 % on average) and nonadjacent electrode (pair $e+2$) (100.88 % on average) pairs were not significantly different from those of the apical contact ($t_{(22)}=-1,380$, $p=0.182$ and $t_{(23)}=-1,132$, $p=0.269$, respectively).

The model outcomes, shown in Figure 3B2, exhibit a dip for $\alpha=0.5$ in TL_{1mm} at AS, which is also visible in the psychophysical results for pair $e+2$ at 360° (Figure 5). A standard Student's T test showed a decrease of the TL_{Φ} of $\alpha=0.5$ (94.33 %) compared

FIG. 4. Individual results for the twelve subjects (S1–12) on the loudness balancing experiment. The *circles* represent the data of the AS pairs and the *triangles* the BS pairs, where *open symbols* are for pair e+1 and *filled symbols* are for pair e+2. The data were normalized and are expressed in percentages. The different values of α are denoted on the horizontal axis.

FIG. 5. The average data of the loudness balancing experiment on AS (*circles*) and BS (*triangles*) for pair e+1 (*open*) and pair e+2 (*filled*). The current was normalized on both ends, where the other percentages are calculated from Eq. 1. The symbols are plotted at staggered α values along the x axis in order to avoid overlapping of the error bars.

with $\alpha=0.33$ (104.03 %), but no difference with $\alpha=0.67$ (102.87 %) ($F_{(66)}=1.505$, $p=0.042$, and 0.072, respectively). After application of the Bonferroni correction, these differences did not meet the significance criterion.

From Figure 4, it is clear that such a dip in AS pair e+2 (at least 10 % lower threshold values at $\alpha=0.5$ compared to $\alpha=0.33$ and $\alpha=0.67$) is visible in the individual results of subjects S1, S2, S5, S8 and S12. To find out, whether these subjects share a common factor, the subjects were divided into two groups, one group with the dip (5 subjects) and a group lacking it (7 subjects). A post-hoc analysis with a standard Student's *T* test was performed, using $p=0.01$ as the level of significance after correcting according to Bonferroni. The average duration of deafness for the group without a dip (15.5 years) was significantly lower ($t_{(10)}=3.599$, $p=0.005$) than that for the group with a dip (36.0 years), while the parameters age at implantation, electrode distance to the medial wall and the absolute value of TL_{Φ} showed no significant difference between these groups ($t_{(10)}=-0.956$, $p=$

0.364, $t_{(10)} = -0.778$, $p = 0.456$ and $t_{(23)} = 2.010$, $p = 0.056$, respectively). This suggests that the duration of deafness is a possible predictor for the dip.

DISCUSSION

Based on previous computational and physiological results (Bonham and Litvak 2008; Frijns et al. 2009b) it was hypothesized that current steering at TL requires a current correction in order to maintain equal loudness for all current weighting combinations between a pair of electrodes. However, TLs are not significantly influenced by the current steering coefficient, at least for adjacent and nonadjacent electrode pairs up to 2.2 mm wide. For the $e+1$ electrode pairs model outcomes, using TL_{1mm} as threshold criterion, were comparable to the patient data for the same condition. In the model larger distances between the contacts give rise to the need of current correction, which was not observed in the subjects.

The model showed a dip in the amount of current correction at $\alpha = 0.5$, specifically for the apical contacts, and only in the condition where the UT was degenerated but the rest of the fiber was still intact (Figure 3B2). This same effect was observed in some of the subject specific outcomes at the same location, although the effect was not large enough to reach a significant level in the whole group. With post-hoc analysis with the dip as group factor, the duration of deafness was the only parameter which showed a significant correlation with the presence of this decrease in threshold at $\alpha = 0.5$. The group with the dip had the longest mean duration of deafness (15.5 years vs. 36.0 years).

This observation can shed some light on the possible time course of neural degeneration in humans. The UT could be the first structure that degenerates after several years of deafness. This would imply that the majority of SG cells in subjects with a long duration of deafness still have their peripheral processes, although the UT is degenerated. In line with this hypothesis, and contrary to earlier reports (Nadol and Eddington 2006; Sly et al. 2007), Rask-Andersen et al. (2010) found that both peripheral processes and SG cells are well preserved in the apical region of the cochlea even after 28 years duration of deafness. At the same time the evidence is growing that the first part that degenerates is the UT, even before the hair cells and the nerve fiber (Kujawa and Liberman 2009; Lin et al. 2011).

As stated above, the model showed only a dip when using the fibers with intact peripheral processes without an UT, while simulations either without peripheral processes or with intact peripheral processes with UT did not exhibit such a dip. The

physiological mechanism underlying this fact can be understood as follows: The UT is the unmyelinated connection between the nerve fiber and the hair cell and therefore is likely to behave as a large node of Ranvier. As a consequence, the UT is expected to possess a large electrical capacitance, which must be loaded with the intra neural potential before the fiber can fire in the peripheral process. In this case, the action potential threshold is more easily initiated in the central axon than at the peripheral process, as it must also be in cases of degeneration of the entire peripheral process. Without a UT the action potentials will be initiated in the peripheral process, as a consequence of its reduced thresholds. Additional simulations (data not shown) demonstrated that even partial preservation of peripheral processes suffices to produce this dip. The differences with respect to the location within the cochlea can be understood from the excitation site along the nerve fiber. Contacts in the more apical turns of the cochlea excite the fibers in the peripheral process, while contacts in the base directly stimulate the central axon and thus the influence of the UT is minimized (Briaire and Frijns 2006).

It is likely, that there are other factors, like the cause of deafness, which influence the course of the degenerative process of the auditory nerve. For instance, Teufert et al. (2006) found, that subjects with idiopathic deafness had the highest residual SG cell count, while subjects with bacterial labyrinthitis had the lowest count. Unfortunately, the present group size and the known large variation in aetiology, preclude an analysis of our data along these lines.

Contrary to the model predictions, only two of the five subjects with a dip in pair AS $e+2$ also exhibited a dip in pair AS $e+1$. An explanation for this could be that the simulation conditions are extreme situations, where all the fibers exhibit the same morphological situation. The subjects probably have all three degenerative fiber conditions mixed along the cochlear partition, which may mask effects only present in one specific condition.

It is very likely that the threshold detection criterion is a patient specific parameter, which is not only determined by the situation in the cochlea (besides the number and condition of the residual nerve fibers, also tissue growth, electrode position etc.), but also by central effects. Also in the model the width and depth of the dip is influenced by the threshold criterion. As shown in Figure 3B, with TL_{0mm} the dip is completely absent, while the model predicts a much shallower dip in $e+2$ and an almost absent dip for $e+1$ if a value of 0.7 mm (rather than 1 mm) along the basilar membrane is used as threshold criterion (data not shown). It should be noted that the absolute value of the threshold

criterion depends on the spread of excitation in the model, which, in turn, depends on various model parameters.

As stated above, both the model and the psychophysics did not show evidence for the need of current correction to maintain equal loudness for pair $e+1$. However, for pair $e+2$ the model showed a clear need for current correction at $\alpha=0.5$, which was not seen in the psychophysical data. This discrepancy might well be caused by an underestimation of the spread of excitation (SOE) in the model. The narrower the SOE, the higher the need for current correction (Frijns et al. 2009b). The SOE in the model is not only influenced by the TL criterion (Figure 3) but also by the (absence of) stochastic neural behaviour and the spatial distribution of the cell bodies in the SG.

Unfortunately, published literature cannot provide much additional evidence, as this is the first clinical study on current steering at threshold, while all other ones have been performed at MCL. Donaldson et al. (2005) showed that for most of the subjects no current correction was needed for creating the same loudness for e (monopolar stimulation) and $e+1$ with $\alpha=0.5$. This finding was confirmed by a previous study from our group (Snel-Bongers et al. 2011), but it was demonstrated that higher current levels were required to reach MCL levels for $e+x$ up to an excitation width of 4.4 mm ($e+4$).

We conclude that with present electrode arrays simultaneous dual electrode stimulation is possible at low current levels and that no current adjustment is necessary to compensate for loudness variations in most cases. The observations for apical contacts in patients with a long duration of deafness are consistent with the model predictions regarding loss of UTs. Therefore, we hypothesize that the majority of the surviving spiral ganglion cells in the cochlea in humans with a long duration of deafness still retain peripheral processes, but have lost their UTs.

In addition, these outcomes are consistent with the notion that psychophysical thresholds are not reached at excitation of the first fiber, but of a region of nerve fibers of approximately 1 mm along the basilar membrane.

ACKNOWLEDGEMENTS

The authors thank Filiep Vanpoucke for his technical support, Ron Wolterbeek for his advise on the statistical analysis, David Dekker for his support during the writing process and the twelve subjects who participated in this study. This research was financially supported by the Heinsius Houbolt Fund and Advanced Bionics. The sponsors had no influence on the method and findings of this study.

Sources of support

Heinsius Houbolt Fund
Advanced Bionics

REFERENCE

- BONHAM BH, LITVAK LM (2008) Current focusing and steering: modeling, physiology, and psychophysics. *Hear Res* 242:141–153
- BOSMAN AJ, SMOORENBURG GF (1995) Intelligibility of Dutch CVC syllables and sentences for listeners with normal hearing and with three types of hearing impairment. *Audiology* 34:260–284
- BRENDEL M, FROHNE-BUECHNER C, STOEVE T, LENARZ T, BUECHNER A (2009) Investigation of pitch discrimination and the effect of learning for virtual channels realized by current steering. *Acta Otolaryngol* 129:1425–1433
- BRIAIRE JJ, FRIJNS JH (2000) Field patterns in a 3D tapered spiral model of the electrically stimulated cochlea. *Hear Res* 148:18–30
- BRIAIRE JJ, FRIJNS JH (2005) Unraveling the electrically evoked compound action potential. *Hear Res* 205:143–156
- BRIAIRE JJ, FRIJNS JH (2006) The consequences of neural degeneration regarding optimal cochlear implant position in scala tympani: a model approach. *Hear Res* 214:17–27
- DONALDSON GS, KREFT HA, LITVAK L (2005) Place-pitch discrimination of single- versus dual-electrode stimuli by cochlear implant users (L). *J Acoust Soc Am* 118:623–626
- FIRSIT JB, KOCH DB, DOWNING M, LITVAK L (2007) Current steering creates additional pitch percepts in adult cochlear implant recipients. *Otol Neurotol* 28:629–636
- FITZMAURICE GM, LAIRD NM, WARE JH (2004) Linear mixed effects model. In G.M.Fitzmaurice, N. M. Laird, & J. H. Ware (Eds.). *Applied Longitudinal Analysis* (pp. 187–236). John Wiley & Sons
- FRIJNS JH, BRIAIRE JJ, GROTE JJ (2001) The importance of human cochlear anatomy for the results of modiolus-hugging multichannel cochlear implants. *Otol Neurotol* 22:340–349
- FRIJNS JH, KALKMAN RK, BRIAIRE JJ (2009A) Stimulation of the facial nerve by intracochlear electrodes in otosclerosis: a computer modeling study. *Otol Neurotol* 30:1168–1174
- FRIJNS JH, KALKMAN RK, VANPOUCKE FJ, BONGERS JS, BRIAIRE JJ (2009B) Simultaneous and non-simultaneous dual electrode stimulation in cochlear implants: evidence for two neural response modalities. *Acta Otolaryngol* 129:433–439
- FRIJNS JH, DEKKER DM, BRIAIRE JJ (2011) Neural excitation patterns induced by phased-array stimulation in the implanted human cochlea. *Acta Otolaryngol* 131:362–370
- [HTTP://WWW.CURVEFIT.COM/LINEAR_REGRESSION.HTM](http://www.curvefit.com/linear_regression.htm) (2009)
- HUGHES ML, GOULSON AM (2011) Electrically evoked compound action potential measures for virtual channels versus physical electrodes. *Ear Hear* 32:323–330
- KOCH DB, DOWNING M, OSBERGER MJ, LITVAK L (2007) Using current steering to increase spectral resolution in CII and HiRes 90K users. *Ear Hear* 28:38S–41S
- KUJAWA SG, LIBERMAN MC (2009) Adding insult to injury: cochlear nerve degeneration after "temporary" noise-induced hearing loss. *J Neurosci* 29:14077–14085
- KWON BJ, VAN DEN HONERT C (2006) Dual-electrode pitch discrimination with sequential interleaved stimulation by cochlear implant users. *J Acoust Soc Am* 120:EL1–EL6
- LEVITT H (1971) Transformed up-down methods in psychoacoustics. *J Acoust Soc Am* 49:Suppl
- LIBERMAN MC, OLIVER ME (1984) Morphometry of intracellularly labeled neurons of the auditory nerve: correlations with functional properties. *J Comp Neurol* 223:163–176

- LIN HW, FURMAN AC, KUJAWA SG, LIBERMAN MC (2011) Primary neural degeneration in the Guinea pig cochlea after reversible noise-induced threshold shift. *J Assoc Res Otolaryngol* 12:605–616
- McDERMOTT HJ, McKAY CM (1994) Pitch ranking with nonsimultaneous dual-electrode electrical stimulation of the cochlea. *J Acoust Soc Am* 96:155–162
- NADOL JB JR, EDDINGTON DK (2006) Histopathology of the inner ear relevant to cochlear implantation. *Adv Otorhinolaryngol* 64:31–49
- PARKINS CW, COLOMBO J (1987) Auditory-nerve single-neuron thresholds to electrical stimulation from scala tympani electrodes. *Hear Res* 31:267–285
- RASK-ANDERSEN H, LIU W, LINTHICUM F (2010) Ganglion cell and 'dendrite' populations in electric acoustic stimulation ears. *Adv Otorhinolaryngol* 67:14–27
- SLY DJ, HEFFER LF, WHITE MW ET AL (2007) Deafness alters auditory nerve fibre responses to cochlear implant stimulation. *Eur J Neurosci* 26:510–522
- SNEL-BONGERS J, BRIAIRE JJ, VANPOUCKE FJ, FRIJNS JH (2011) Influence of widening electrode separation on current steering performance. *Ear Hear* 32:221–229
- SNEL-BONGERS J, BRIAIRE JJ, VANPOUCKE FJ, FRIJNS JH (2012) Spread of excitation and channel interaction in single- and dual-electrode cochlear implant stimulation. *Ear Hear* 33:367–376
- STAKHOVSKAYA O, SRIDHAR D, BONHAM BH, LEAKE PA (2007) Frequency map for the human cochlear spiral ganglion: implications for cochlear implants. *J Assoc Res Otolaryngol* 8:220–233
- TEUFERT KB, LINTHICUM FH JR, CONNELL SS (2006) The effect of organ of corti loss on ganglion cell survival in humans. *Otol Neurotol* 27:1146–1151
- TOWNSHEND B, COTTER N, VAN CD, WHITE RL (1987) Pitch perception by cochlear implant subjects. *J Acoust Soc Am* 82:106–115
- VERBIST BM, FRIJNS JH, GELEIJNS J, VAN BUCHEM MA (2005) Multisection CT as a valuable tool in the postoperative assessment of cochlear implant patients. *AJNR Am J Neuroradiol* 26:424–429
- VERBIST BM, SKINNER MW, COHEN LT, ET AL. (2010) Consensus Panel on a Cochlear Coordinate System Applicable in Histologic, Physiologic, and Radiologic Studies of the Human Cochlea. *Otol Neurotol*