Detection of changes of high-frequency activity by statistical time-frequency analysis in epileptic spikes

Katsuhiko Kobayashi^{a,*}, Julia Jacobs^b, and Jean Gotman^b

^aDepartment of Child Neurology, Okayama University Graduate School of Medicine, Dentistry and Pharmaceutical Sciences, Okayama University Hospital, Shikatacho 2-chome 5-1, Okayama 700-8558, Japan

^bMontreal Neurological Institute and Hospital, McGill University, Montreal, Que., Canada

Abstract

Objective—A novel type of statistical time-frequency analysis was developed to elucidate changes of high-frequency EEG activity associated with epileptic spikes.

Methods—The method uses the Gabor Transform and detects changes of power in comparison to background activity using t-statistics that are controlled by the false discovery rate (FDR) to correct type I error of multiple testing. The analysis was applied to EEGs recorded at 2000 Hz from three patients with mesial temporal lobe epilepsy.

Results—Spike-related increase of high-frequency oscillations (HFOs) was clearly shown in the FDR-controlled *t*-spectra: it was most dramatic in spikes recorded from the hippocampus when the hippocampus was the seizure onset zone (SOZ). Depression of fast activity was observed immediately after the spikes, especially consistently in the discharges from the hippocampal SOZ. It corresponded to the slow wave part in case of spike-and-slow-wave complexes, but it was noted even in spikes without apparent slow waves. In one patient, a gradual increase of power above 200 Hz preceded spikes.

Conclusions—FDR-controlled *t*-spectra clearly detected the spike-related changes of HFOs that were unclear in standard power spectra.

Significance—We developed a promising tool to study the HFOs that may be closely linked to the pathophysiology of epileptogenesis.

Keywords

High-frequency; Time-frequency analysis; False discovery rate; Hippocampus; Mesial temporal lobe epilepsy; Spike

*Corresponding author. Tel.: +81 86 235 7372; fax: +81 86 235 7377. k_koba@md.okayama-u.ac.jp (K. Kobayashi).

Conflict of interest

None of the authors has any conflict of interest to disclose.

1. Introduction

Recent development of digital EEG recording with high sampling frequency rates (1000–2000 Hz) has facilitated observation of high-frequency activity of up to 600 Hz (Le Van Quyen et al., 2006). High-frequency oscillations (HFOs) ranging from 80 to 250 Hz (ripples) can be recorded from the hippocampus and entorhinal cortex of normal rodents (Buzsáki et al., 1992; Chrobak and Buzsáki, 1996) and also from human hippocampus (Bragin et al., 1999a). In kainic acid-treated rats and patients with mesial temporal lobe epilepsy, even faster HFOs of 250–500 Hz (fast ripples) have been detected using microelectrodes or microwires from regions close to the epileptogenic lesion, and are suggested to reflect pathological hypersynchronous events crucially associated with seizure generation (Bragin et al., 1999a,b; Staba et al., 2002; Rampp and Stefan, 2006). Jirsch et al. (2006) demonstrated for the first time that HFOs in the range of 100–500 Hz can be recorded from human epileptic patients during focal seizures using depth macroelectrodes: they made an important advance for the study of HFOs that had been limited by the mandatory usage of microelectrodes or microwires to record fast ripples.

The interictal epileptic spikes on EEG are traditional indications of epileptic irritability of the brain. An epileptic spike sometimes occurs together with a slow wave immediately following the spike and then is referred to as a spike-and-slow-wave complex. Urrestarazu et al. (2006) found using depth macroelectrodes that the power of high-frequency activity above 100 Hz is decreased during the period immediately following spikes, and that this decrease is prominent in the hippocampus but less consistent in amygdala and neocortex. It was indicated that the reduction in fast activity could reflect a depression in neuronal activity during post-spike slow waves. Experimental studies have shown an inhibitory action of neuronal hyperpolarization caused by both intrinsic and synaptic mechanisms in the slow wave phase following the paroxysmal depolarization shift (PDS) that corresponds to spikes (Ayala et al., 1973; Neckelmann et al., 2000). Given a hypothesis that reduction of the inhibitory action, especially altered balance between excitation and inhibition, is involved in epileptogenesis (Ayala et al., 1973), further analysis with respect to the changes of high-frequency activity associated with epileptic spikes might lead to a better understanding of the mechanisms of seizure generation.

We have developed a novel type of statistical time-frequency analysis in order to elucidate such spike-related changes of high-frequency activity in more detail regarding their frequency characteristics and temporal sequence. We performed a preliminary evaluation of this analysis with clinical data in the present study.

2. Materials and methods

2.1. Patients and EEG recording

Between 2005 and 2007, 47 patients with medically intractable epilepsy were recorded with intracranial electrodes and a high sampling rate at the Montreal Neurological Institute. The first three patients with bi-temporal implantations and mesial temporal lobe epilepsy were selected for this study. No other selection criteria were used. The Montreal Neurological

Institute and Hospital Research Ethics Committee approved this study and informed consent was obtained from each patient.

Stereoelectroencephalography (SEEG) was performed by implanting depth electrodes orthogonally through the middle temporal gyrus in anterior, mid and posterior locations so that the deepest contacts were situated in the amygdala, hippocampus and parahippocampus. The implantation method is described by Olivier et al. (1994). Electrode bundles were implanted stereotactically using an image-guidance system (SSN Neuronavigation System, Mississauga, Ontario, Canada) through percutaneous holes drilled in the skull.

Intracranial depth electrodes (electrode bundles) were manufactured on site by wrapping 3/1000 in. (0.076 mm) stainless steel wire around a 10/1000 in. (0.254 mm) stainless steel central core. These wires were coated with Teflon except for regions where the insulation was stripped to form electrode contacts. In total, there were nine contacts on each electrode bundle that were spaced along the length of the core wire at 5-mm intervals. The deepest contact (contact 1) was made from the tip of the core wire and had an uninsulated length of 1 mm, while more superficial contacts (contacts 2–9) were formed from stripped sections of the marginal wire that was tightly wound to make 0.5-mm long coils. The effective surface area for each of the eight superficial contacts was 0.80 mm², and was 0.85 mm² for the single deep contact. Impedances were measured once immediately after implantation; they varied between 1 and 20 k Ω (Urrestarazu et al., 2006).

The EEG telemetry signal was digitally recorded with a 128-channel Harmonie system for long-term monitoring (Stellate, Montreal, Canada) with filter settings of 0.1 and 500 Hz and a sampling frequency of 2000 Hz in conformity to Jirsch et al. (2006) and Urrestarazu et al. (2006). The analysis was performed using a bipolar montage. Each channel compared two adjacent contacts of the same electrode bundle. (Table 1) shows the information about the implanted electrodes and the finding of the seizure onset zone (SOZ) of each patient. The SOZ was defined as the set of contacts that showed the earliest ictal activity during the intracranial investigation.

2.2. Spectral analysis

SEEG recordings were visually reviewed to identify the interictal epileptiform discharges. The discharges were categorized according to spatial distribution and morphology. The polarity of spike peak was negative in some discharges and positive in others depending on the relation between the generator and the electrode locations, but the identified spike types had a consistent polarity. In each patient, four different types of discharges were classified and marked in the interictal SEEG data: (1) a spike-and-slow-wave complex recorded from the SOZ (SW_{SOZ}), (2) a spike without slow wave recorded from the SOZ (S_{SOZ}), (3) a spike-and-slow-wave complex recorded from outside the SOZ (NSOZ)(SW_{NSOZ}), and (4) a spike without slow wave recorded from outside the SOZ (S_{NSOZ}).

With respect to each spike type, 50 EEG segments lasting 2000 ms (comprising 4000 data-points) and non-overlapping were sampled for analysis, each segment including a marked spike and surrounding background activity. The spike segments were selected not to include any other spike or artifact in the channel of the marked spikes, and the minimal interval

between the selected spikes was 2.5 s. A time-frequency power spectrum was built for each spike segment of bipolar EEG data by using the Gabor (Windowed Fourier) Transform (Kobayashi et al., 2004) with a sliding Gaussian window of 50 ms FWHM (full width half maximum) (the spike-related spectral data). The frequency range was 50–500 Hz. The Fourier Transform was performed on 512 data-points (256 ms; frequency resolution 3.9 Hz) of the window at each time-step, and the step of the sliding window was 2 ms. The signal power was converted to logarithmic scale to obtain a more Gaussian distribution (Gasser et al., 1982). The average spike-related spectra were obtained by averaging the 50 data segments regarding each spike type.

For statistical comparison of the power of fast activity between the spikes and the background, 300 EEG segments lasting 256 ms and non-overlapping were selected in the background, and the Fourier Transform was similarly applied to each background segment to obtain the control spectral data. The background segments were at least 1 s away from any spike, and they were obtained from wakefulness, as were the spikes. Only the wakeful EEG recordings were used, because it was difficult to select a sufficient number of spike-free background segments from the sleep EEGs, which showed very frequent spikes. We used a large number of background segments (76.8 s of EEG data in a total) for the control data because there are generally temporal fluctuations in the EEG background activity. The unpaired *t*-test was performed between the spike-related and control spectral data to obtain the *t* value and the corresponding *p* value at each pixel of the time-frequency spectrum. Because there is an enormous number of pixels (115,000) in the time-frequency spectrum, a test with an ordinary significance level α (say, 0.05 or a 5% error rate) but without any correction procedure for multiple testing would result in declaring too many pixels as active (type I errors). Standard methods for controlling the false-positive rate in case of multiple testing are, however, too conservative: in the Bonferroni correction which is commonly used, the nominal significance level α is replaced with α/V where *V* is the number of tests performed (the number of pixels in this situation) (Genovese et al., 2002).

Therefore we adopted a statistical procedure for controlling the false discovery rate (FDR) (Genovese et al., 2002), as frequently used for neuroimaging studies. The FDR is defined as the ratio of the number of false positive pixels to the number of pixels declared active. At first, one should select the FDR bound *q* that is the maximum tolerable FDR on average. The *p* values obtained by *t*-test with respect to *V* pixels are ordered from smallest to largest, and are indexed by *j* ranging from 1 to *V* as $p_{(j)}$. To define the pixels of significantly small *p* values, the largest *j* for which $p_{(j)} \leq \frac{j \cdot q}{c \cdot V}$ is determined and denoted as *r*, where *c* is

$\sum_{j=1}^V 1/j$ when statistics at different pixels are not independent. The spectrum of *t* values (*t*-spectrum) controlled by the FDR was built with the threshold corresponding to the *p* value $p_{(r)}$.

In the present study, the two-tailed test was used with $q = 0.025$. To determine the pixels that had significantly smaller power than the background control data, the above procedure was performed. Conversely, to determine the pixels that had significantly greater power than the control, the above procedure was performed with the *p* value replaced by the value of $1 - p$. In the FDR-controlled *t*-spectra, pixels with significantly small power indicated by *t* values

below the lower limit are shown in blue, and pixels with significantly great power indicated by t values above the upper limit are indicated in red; pixels with t values that did not reach either limit are in green (Fig. 1 for example).

We also applied the discrete and subsequent inverse Fourier Transform to the SEEG data of epileptiform discharges to filter out the EEG activity lower than 100 Hz. The high-pass filtered SEEG traces of 50 spikes were piled to directly illustrate the changes of fast activity associated with spikes.

The computation was done with an in-house program written in MATLAB (version 6.5.1; MathWorks, USA).

2.3. Simulation study

We were concerned with the possibility that the transient changes of fast activity detected before and after the spikes could be artifacts produced by the application of spectral analysis to peaky signals such as spikes. To investigate this possibility, 50 transients resembling spike-and-slow-wave complexes were generated by combining parts of Gaussian pulses with a small computer-generated random jitter and were added to spike-free background data recorded from the left hippocampus in patient 1. An analysis similar to the above description was applied to these simulated spike data.

3. Results

3.1. Real SEEG data

In patient 1 whose seizures originated in the left hippocampus, the interictal SW_{SOZ} and S_{SOZ} were recorded from the left hippocampus at the SOZ, and the SW_{NSOZ} and S_{NSOZ} were from the left temporal neocortical region and the right temporal neocortical region, respectively. In patient 2, who had seizures from the right hippocampus, the SW_{SOZ} and S_{SOZ} were recorded from the right hippocampus at the SOZ, and the SW_{NSOZ} and S_{NSOZ} were from the left amygdala. In patient 3 with seizures from the right hippocampus, the SW_{SOZ} and S_{SOZ} were recorded from the right hippocampus at the SOZ, and SW_{NSOZ} and S_{NSOZ} were from the left hippocampus. Results for these three patients are shown in Figs. 1, 2 and 3 respectively.

The epileptiform discharges recorded from the SOZ showed increase of fast activity associated with the spikes in the average time-frequency spectra in all three patients. This change could be observed in t -spectra without FDR-control, but it was not nearly as clear because there were so many significant values everywhere in the time-frequency plane. The spike-related increase of power was unmistakably illustrated in the FDR-controlled t -spectra, and it largely involved the whole frequency range of 50–500 Hz in SW_{SOZ} in the three patients (Figs. 1, 2, and 3A) and in S_{SOZ} in patient 1 (Fig. 1B). It was mostly limited to the frequencies below 200 Hz in S_{SOZ} of patients 2 and 3 (Figs. 2 and 3B). Depression of fast activity was observed immediately after the spikes in the t -spectra, especially with FDR-control, and it involved the whole frequency range with a tendency of predominance or longer durations in the frequencies below 200 Hz in SW_{SOZ} and S_{SOZ} in patients 1 and 2 (Figs. 1 and 2A and B). It was noted in the frequencies below about 100 Hz in SW_{SOZ} in

patient 3 (Fig. 3A). The power depression temporally corresponded with the slow wave part of the discharges in SW_{SOZ} , but it was observed even in S_{SOZ} without apparent slow waves. In patient 1, the t -spectra of SW_{SOZ} and S_{SOZ} showed gradual increase of power in frequencies above 200 Hz that *preceded* the spikes by 100–200 ms (Fig. 1A and B arrows).

The epileptiform discharges recorded from outside the SOZ (SW_{NSOZ} and S_{NSOZ}) showed moderate increase of power below 300 Hz or no change in association with spikes in the time-frequency t -spectra. In contrast to the SOZ, post-spike depression of fast activity was not found in either SW_{NSOZ} or S_{NSOZ} (Figs. 1, 2, and 3CD).

The SEEG traces of the epileptiform discharges that were high-pass filtered at 100 Hz showed spike-related increase and post-spike depression of fast activity in SW_{SOZ} and S_{SOZ} of patients 1 and 2 (Fig. 4A and B) and in SW_{SOZ} of patient 3, confirming the FDR-controlled time-frequency t -spectra. As seen in the t -spectra, power in frequencies higher than 200 Hz started to increase 100–200 ms before the spike peaks in SW_{SOZ} of patient 1 (Fig. 4B arrow). In contrast, the high-pass filtered traces of SW_{NSOZ} and S_{NSOZ} showed, if any, minimal increase of fast activity in association with spikes and no noticeable decrease of fast activity (Fig. 4C and D).

3.2. Simulation study

In the analysis of the simulated spike-and-slow-wave complexes, increase of high-frequency signals was shown by both the FDR-controlled time-frequency t -spectrum and filtered SEEG traces at the time of the spike, but we did not find any leading increase of fast activity before the spikes or decrease of fast activity after the spikes (Fig. 5). This confirmed that the analysis itself did not result in changes in high frequencies outside of the time in which the spikes were present.

4. Discussion

The changes of high-frequency activity associated with the interictal epileptiform discharges were best visualized by the FDR-controlled time-frequency t -spectra in the present study. The time-frequency power spectra showed spike-related changes in power, but because fast activity generally has much smaller power than slow activity, the raw power spectra without statistical manipulation were not so informative with respect to the behavior of high-frequency activity. The time-frequency t -spectra indicated increase and decrease of high-frequency activity associated with spikes in comparison to the background activity, but they were still difficult to interpret without an appropriate threshold of t or p values because of many noisy peaks and troughs. The objective significance level that was obtained by the control of FDR for the time-frequency t -spectra was demonstrated to be useful for the study of spike-related changes of high-frequency activity, as illustrated in Fig. 1.

The FDR-controlled time-frequency t -spectra clearly showed post-spike depression of high-frequency activity with respect to the epileptiform discharges recorded from the hippocampus in the SOZ. The high-pass filtered SEEG traces of the epileptiform discharges also illustrated the depression of high-frequency activity after spikes, in accordance with the earlier demonstration of a post-spike depression in fast activity using a statistical comparison

of power in various frequency bands (Urrestarazu et al., 2006). The analysis of the simulated spike-and-slow-wave complexes demonstrated that the post-spike changes of fast activity were not artifacts produced by the spectral analysis of peaky signals. The post-spike depression was more striking in spike-and-slow-wave complexes than in spikes without slow waves, but it was still detectable even without apparent slow waves in two patients out of three. This post-spike depression was not obvious in the discharges recorded from the amygdala or neocortex, in agreement with the findings of Urrestarazu et al. (2006).

Regarding the epileptiform discharges recorded from the non-ictogenic left hippocampus in patient 3, the post-spike depression was not observed, in contrast to the discharges from the right hippocampus in the SOZ. The behavior of post-spike high-frequency activity might therefore depend on whether spikes are in the SOZ or outside. The number of patients was small in the present preliminary clinical application of the statistical time-frequency analysis, and these questions will be addressed in future studies with more patients.

Increase of high-frequency activity was observed in association with SEEG spikes, but it was also found in simulated spikes that had a very sharp edge at the peak. Differentiation between the true HFOs associated with epileptiform discharges and high-frequency components of fast transients is difficult with the present methodology. Independently of its origin, the increase of high-frequency activity was much more prominent in discharges from the SOZ than in those from outside, indicating its relation to epileptogenesis (Jacobs et al., 2008). We used the Gabor Transform, based on the Fourier Transform with a good frequency resolution, and we made a compromise regarding the temporal resolution. We could have a better temporal resolution with a narrower Gaussian window, compromising the frequency resolution to some extent. Development of another method with a good temporal resolution, such as a method based on the wavelet transform, would also reach the above aim. Using the present method, however, the t-spectra showed gradual increase of power in frequencies above 200 Hz that preceded the spikes recorded from the SOZ in patient 1 (Fig. 1A and B arrows): this increase was shown in narrow spectral streaks heralding the spikes, and the frequency resolution of the present method may be appropriate to detect this sort of activity. This increase of high-frequency activity preceding the spikes was observed only in this patient, and the problems of its actual incidence and relation to epileptogenesis require more investigation.

We used the FDR for type I error control in multiple testing in the present study according to Genovese et al. (2002). There are other versions of FDR and other methods for type I error control (Strimmer, 2008; Hemmelmann et al., 2005). It is beyond the scope of the present study to compare these various methods. Application of FDR to EEG data was reported with respect to event-related electrocorticographic gamma activity (Ray et al., 2008). There were no statistical time-frequency studies with type I error control regarding the changes associated with EEG spikes.

HFOs are divided into ripples and fast ripples: ripples are detected not only from epileptogenic regions but also from non-epileptogenic regions, and they may include physiological activity. Fast ripples of 250–500 Hz have a close relation to epileptogenesis (Le Van Quyen et al., 2006). Urrestarazu et al. (2006) found that post-spike depression of

activity in the 250–500 Hz frequency band coincides with the localization of largest spike amplitude, and that it was spatially more restricted than that of lower frequency bands. In the present study, post-spike depression of fast activity tended to be more intense and longer in the frequencies below 200 Hz than in the higher frequencies. The post-spike changes of high-frequency activity appeared to depend on frequency with a separation around 200 or 250 Hz, although depressed post-spike fast activity cannot be equated with ripples or fast ripples that are short-lasting oscillatory events, as pointed out by Urrestarazu et al. (2006)

The changes of high-frequency activity before and after EEG spikes were visualized in a statistical time-frequency spectra in this study. This method can be modified depending on the type of target activity, as mentioned above. The present method based on the Gabor Transform with a favorable frequency resolution appears to be a promising tool to explore the detailed frequency characteristics of changes in high-frequency activity around spikes. This method may be also useful for a detailed study on the changes of HFOs associated with seizures, and we hope that it will be helpful to explore the most epileptogenic brain region.

Acknowledgments

This study was supported in part by Grant MT-10189 of the Canadian Institutes of Health Research.

References

- Ayala GF, Dichter M, Gumnit RJ, Matsumoto H, Spencer WA. Genesis of epileptic interictal spikes. New knowledge of cortical feedback systems suggests a neurophysiological explanation of brief paroxysms. *Brain Res.* 1973; 52:1–17. [PubMed: 4573428]
- Bragin A, Engel J Jr, Wilson CL, Fried I, Buzsáki G. High-frequency oscillations in human brain. *Hippocampus.* 1999a; 9:137–42. [PubMed: 10226774]
- Bragin A, Engel J Jr, Wilson CL, Fried I, Mathern GW. Hippocampal and entorhinal cortex high-frequency oscillations (100–500 Hz) in human epileptic brain and in kainic acid-treated rats with chronic seizures. *Epilepsia.* 1999b; 40:127–37. [PubMed: 9952257]
- Buzsáki G, Horváth Z, Urioste R, Hetke J, Wise K. High-frequency network oscillation in the hippocampus. *Science.* 1992; 256:1025–7. [PubMed: 1589772]
- Chrobak JJ, Buzsáki G. High-frequency oscillations in the output networks of the hippocampal–entorhinal axis of the freely behaving rat. *J Neurosci.* 1996; 16:3056–66. [PubMed: 8622135]
- Gasser T, Bächer P, Möcks J. Transformations towards the normal distribution of broad band spectral parameters of the EEG. *Electroencephalogr Clin Neurophysiol.* 1982; 53:119–24. [PubMed: 6173196]
- Genovese CR, Lazar NA, Nichols T. Thresholding of statistical maps in functional neuroimaging using the false discovery rate. *Neuroimage.* 2002; 15:870–8. [PubMed: 11906227]
- Hemmelmann C, Horn M, Süsse T, Vollandt R, Weiss S. New concepts of multiple tests and their use for evaluating high-dimensional EEG data. *J Neurosci Methods.* 2005; 142:209–17. [PubMed: 15698661]
- Jacobs J, Levan P, Chander R, Hall J, Dubeau F, Gotman J. Interictal high-frequency oscillations (80–500 Hz) are an indicator of seizure onset areas independent of spikes in the human epileptic brain. *Epilepsia.* 2008; 49:1893–907. [PubMed: 18479382]
- Jirsch JD, Urrestarazu E, LeVan P, Oliver A, Dubeau F, Gotman J. High-frequency oscillations during human focal seizures. *Brain.* 2006; 129:1593–608. [PubMed: 16632553]
- Kobayashi K, Oka M, Akiyama T, Inoue T, Abiru K, Ogino T, et al. Very fast rhythmic activity on scalp EEG associated with epileptic spasms. *Epilepsia.* 2004; 45:488–96. [PubMed: 15101830]
- Le Van Quyen M, Khalilov I, Ben-Ari Y. The dark side of high-frequency oscillations in the developing brain. *Trends Neurosci.* 2006; 29:419–27. [PubMed: 16793147]

- Neckelmann D, Amzica F, Steriade M. Changes in neuronal conductance during different components of cortically generated spike-wave seizures. *Neuroscience*. 2000; 96:475–85. [PubMed: 10717428]
- Olivier A, Germano IM, Cukiert A, Peters T. Frameless stereotaxy for surgery of the epilepsies: preliminary experience. Technical note. *J Neurosurg*. 1994; 81:629–33. [PubMed: 7931603]
- Rampp S, Stefan H. Fast activity as a surrogate marker of epileptic network function? *Clin Neurophysiol*. 2006; 117:2111–7. [PubMed: 16843722]
- Ray S, Niebur E, Hsiao SS, Sinai A, Crone NE. High-frequency gamma activity (80–150 Hz) is increased in human cortex during selective attention. *Clin Neurophysiol*. 2008; 119:116–33. [PubMed: 18037343]
- Staba RJ, Wilson CL, Bragin A, Fried I, Engel J Jr. Quantitative analysis of high-frequency oscillations (80–500 Hz) recorded in human epileptic hippocampus and entorhinal cortex. *J Neurophysiol*. 2002; 88:1743–52. [PubMed: 12364503]
- Strimmer K. A unified approach to false discovery rate estimation. *BMC Bioinformatics*. 2008; 9:303. [PubMed: 18613966]
- Urrestarazu E, Jirsch JD, LeVan P, Hall J, Avoli M, Dubeau F, et al. High-frequency intracerebral EEG activity (100–500 Hz) following interictal spikes. *Epilepsia*. 2006; 47:1465–76. [PubMed: 16981862]

Fig. 1.

Time-frequency power spectra and FDR-controlled t -spectra of SEEG spikes in patient 1. (A) Spike-and-slow-wave complexes from the SOZ in the left hippocampus (SW_{SOZ}). (B) Spikes without following slow wave from the left hippocampus (S_{SOZ}). (C) Spike-and-slow-wave complexes from the left temporal neocortical region that was outside of the SOZ (SW_{NSOZ}). (D) Spikes without slow waves from the right temporal neocortical region that was outside of the SOZ (S_{NSOZ}). For each spike type, ordered from top to bottom: overlaid SEEG traces including spikes, average power spectrum, t -spectrum without control, and t -spectrum controlled by FDR with $q = 0.025$ (two-tailed test). Power of fast activity significantly increased in association with spikes and transiently decreased after the spikes in the epileptiform discharges recorded from the left hippocampus, especially in SW_{SOZ} . Increase of activity above 200 Hz appears to precede the spikes in SW_{SOZ} and S_{SOZ} (white arrows). Activity up to about 100 Hz increased in SW_{NSOZ} , but not in S_{NSOZ} . Post-spike depression of fast activity was not observed in either SW_{NSOZ} or S_{NSOZ} .

Fig. 2.

Time-frequency power spectra and FDR-controlled t -spectra of S EEG spikes in patient 2. (A) Spike-and-slow-wave complexes from the SOZ in the right hippocampus (SW_{SOZ}). (B) Spikes without slow waves from the right hippocampus (S_{SOZ}). (C) Spike-and-slow-wave complexes from the left amygdala that was outside of the SOZ (SW_{NSOZ}). (D) Spikes without slow waves from the left amygdala (S_{NSOZ}). For each spike type, from top to bottom: overlaid S EEG traces including spikes, average power spectrum, and FDR-controlled t -spectrum. Power of fast activity significantly increased with spikes and transiently decreased after the spikes in SW_{SOZ} , but these changes were less profound in S_{SOZ} . Fast activity moderately increased with spikes in SW_{NSOZ} and S_{NSOZ} , but post-spike depression was not observed.

Fig. 3. Time-frequency power spectra and FDR-controlled *t*-spectra of SEEG spikes in patient 3. (A) Spike-and-slow-wave complexes from the SOZ in the right hippocampus (SW_{SOZ}). (B) Spikes without succeeding slow waves from the right hippocampus (S_{SOZ}). (C) Spike-and-slow-wave complexes from the left hippocampus that was outside of the SOZ (SW_{NSOZ}). (D) Spikes without slow waves from the left hippocampus (S_{NSOZ}). In SW_{SOZ}, power of fast activity significantly increased with spikes, but post-spike depression was limited below about 100 Hz. Some increase of fast activity was noted in S_{SOZ}, SW_{NSOZ} and S_{NSOZ}, but post-spike depression was not seen.

Fig. 4.

High-pass filtered SEEG traces including spikes in patient 1. SEEG traces including spike-and-slow-wave complexes recorded from the left hippocampal electrodes in the SOZ (SW_{SOZ}) are shown with high-pass filters at 0.5 Hz (A) and 100 Hz (B). Fast activity higher than 100 Hz started to increase around 100 ms before the spike peaks (arrow) and were transiently suppressed after the spikes in (B). Traces including spike-and-slow-wave complexes from the left temporal neocortical region outside of the SOZ (SW_{NSOZ}) are similarly shown with a filter at 0.5 Hz (C) and at 100 Hz (D). Changes of fast activity were not observed in (D).

Fig. 5.

Spectral analysis in simulated spike-and-slow-wave complexes. (A) Spike-and-slow-wave complexes were simulated by adding transients like spike-and-slow-wave complexes generated by the computer to the background SEEG data recorded from the left hippocampus in the SOZ in patient 1. Overlaid traces including simulated discharges, the average power spectrum and the FDR-controlled t -spectrum are ordered from top to bottom. High-frequency signals were detected from these simulated discharges, but increase of fast activity did not start before the spikes. Post-spike depression of fast activity was not found. The traces including the simulated discharges are shown with a high-pass filter at 0.5 Hz (B) and at 100 Hz (C): this demonstrates that the analysis method does not introduce artificial changes before and after the transients.

Table 1

Summary of the results of analysis.

Patient	Spike type	Electrode location	Relation to seizure origin	Fast activity in spike	Post-spike depression of fast activity
1	SW	Left hippocampus	SOZ	+++	+++
	S	Left hippocampus	SOZ	+++	++
	SW	Left temporal neocortical region	NSOZ	+	No
	S	Right temporal neocortical region	NSOZ	No	No
2	SW	Right hippocampus	SOZ	+++	+++
	S	Right hippocampus	SOZ	+	++
	SW	Left amygdala	NSOZ	++	No
	S	Left amygdala	NSOZ	++	No
3	SW	Right hippocampus	SOZ	+++	+
	S	Right hippocampus	SOZ	+	No
	SW	Left hippocampus	NSOZ	+	no
	S	Left hippocampus	NSOZ	+	no

SW, spike slow wave; S, spike without slow wave.

SOZ, seizure onset zone; NSOZ, non-seizure onset zone.

Fast activity in spike and the post-spike depression of fast activity were rated as +++ when they involved the whole frequency band of 50–500 Hz, as ++ when they extended beyond 200 Hz but did not involve the whole frequency band of 50–500 Hz, and as + when they did not extend beyond 200 Hz.