Published in final edited form as:

Health Phys. 2013 November ; 105(5 0 2): S199–S208. doi:10.1097/HP.0b013e31828e56b6.

Interceptor and Phantom Trials of EDNS at UPMC

Ryan Dickson, Jong Oh Kim, M. Saiful Huq, Greg Bednarz, Joe Suyama, Donald M. Yealy, Mark Izadbakhsh, and Joel S. Greenberger

Abstract

University of Pittsburgh Medical Center (UPMC) installed an Emergency Department Notification System (EDNS) in one of its hospitals. The system, manufactured by Thermo Fisher Scientific (Thermo Fisher Scientific, Inc., 81 Wyman Street, Waltham, MA 02454), consists of four NaI(Tl) scintillation detectors, a 2.5 L PVT gamma counter, a 512 channels multi-channel analyzer, a system controller, and a database-monitoring server. We evaluated a portable *Interceptor*TM hand-held detector as part of the system for potential ambulance-based early detection and warning unit. We present the minimum detectable activity, distance, and isotope identification success rates along with the change in detector response to various radioisotope sources placed in a Rando[®] humanoid phantom. The present paper reports these results.

Introduction

The nuclear accident at the Fukushima Daiichi Power Plant resulting from damage the plant suffered from the 9.0 earthquake and subsequent tsunami that hit Japan on March 10, 2011 has shown the impact a radiological disaster can have on a region without direct serious injuries or deaths from radiation exposure. The number of people killed by the earthquake and tsunami dwarf the 0.1% increase in the incidence of cancer, the long-term psychological effects related to the fear of the radiation may affect a large population in the contaminated areas (Von Hippel 2011). The terrorist attacks on September 11th 2001 demonstrated that unconventional methods can deliver powerful results. Many authors and organizations have suggested the possible use of radiological weapons in future attacks (McDonald 2007) (Stather 2007) (Morgan, et al. 2008). While the amount of material released and area affected would be smaller than for a nuclear weapon or nuclear power plant accident, the number of people exposed and resultant panic would be large in a major metropolitan target. In 1987, 51 TBq of ¹³⁷Cs were stolen from an abandoned radiotherapy institute in Goiânia Brazil. The source was handled by many people unaware of the danger, resulting in four deaths. About 130,000 people overwhelmed hospitals, of which 250 people were found to be contaminated and 20 people showed signs of radiation sickness and required treatment (IAEA 1988).

The above incidents have shown the need for detection and dosimetric measurements in response to a radiological emergency, accidental or deliberate (terrorism). Comprehensive measures must be taken to determine the level of radioactivity, specific radionuclides, and the resultant hazard posed by the contamination. These measurements will very likely take place in conjunction with the medical response, either by first responders at the scene or in the emergency department of a nearby hospital.

The Emergency Department Notification System (EDNS) system has been designed specifically to detect and identify radioactive isotopes with the ability to discriminate non-medical isotopes from sources used in nuclear medicine and radiation brachytherapy. A series of experimental studies using the EDNS system were carried out at UPMC hospitals. The purpose of these tests was to measure the detector response and identification accuracy of the portable *Interceptor*TM detector. The identification capabilities of which are crucial in

distinguishing between hazardous contaminations and “false positive” alarms from clinically indicated medical isotopes. Deployment of the portable detector in an ambulance could allow for earlier detection and notification, reducing the number of personnel exposed to a minimum.

EDNS System

Description

Developed in conjunction with Thermo Fisher Scientific the Emergency Department Notification System (EDNS) is a multi-detector radiation detector system customized for the hospital environment. The configuration currently installed at one of the UPMC hospitals in Pittsburgh consists of four large volume (10.2 cm × 5.1 cm × 40.6 cm) sodium iodide [NaI(Tl)] scintillation detectors mounted in the ceiling, three in the emergency department and the fourth in a high traffic area on the hospital's third floor. The four NaI detectors are networked to a control unit possessing a multi-channel analyzer for spectrum analysis for isotope identification. A large polyvinyltoluene (PVT) scintillator is also installed on the third floor for gross gamma counting and as a possible trigger for future camera surveillance use. A full description of the system was previously published (Kim, et al. 2011).

The system also supports Wi-Fi hand-held isotope identification devices to extend monitoring beyond the portal locations. The Thermo Fisher Scientifics *Interceptor*TM detector is a portable multi-channel analyzer as well as a personal radiation monitor, providing emergency personnel the capability to locate, identify and quantify dispersed radioactive contamination on patient following or independent of the initial notification of EDNS. The portability of the device allows it to be brought to the patient or source for early notification and identification.

The various detectors, hardware controllers and computers are networked over the UPMC local area network as shown in Figure 1. The monitoring and control of the various detectors is handled by the *Viewpoint* server. The system provides real-time event display, alarm handling, and logging events and data of interest to a SQL database for further analysis. The data can be viewed from a number of software clients.

Prior Tests

The performance of the NaI detectors has previously been reported (Kim, et al. 2011). The system had excellent capability of detecting low activity non-medical sources. Initial issues with false positives, from CT scanners and other X-ray sources were largely eliminated after improvements to the isotope library. The system identified combinations of several isotopes with excellent sensitivity, with a minimum detectable activity of 0.13 MBq (3.5 µCi). Subsequent testing after a recent software upgrade verified the integrity of the system yielded compatible measurements of detectable activity, minimum distances, response times and identification of isotopes.

Setup

We sought to test the potential utility extending the EDNS system outside the hospital environment. To achieve these goals, we tested the *Interceptor*TM detector's portability, sensitivity and identification capabilities against a variety of sources, distances, and presence of phantom material. The primary requirements for its role in the EDNS system are sufficient sensitivity and identification accuracy to detect radionuclide contamination on a patient being transported by ambulance in time to alert the emergency personnel before arrival at the hospital.

Throughout our measurements, distances were chosen to be representative of dimensions encountered in an ambulance, with a working distance from likely mounting points and the center of the stretcher between 0.5–2.5 meters.

Sources

We tested against a wide array of low activity check sources and several common medical sources, listed in Table 1. Initial phantom tests were performed. We tested initially with 0.037–0.92 MBq (1–25 μ Ci) of ^{137}Cs sealed button sources using both the NaI detectors and the Interceptor. Next, we tested in phantom studies and *Interceptor*TM testing with higher activity ^{137}Cs , ^{133}Ba , ^{60}Co sources.

Interceptor

The *Interceptor*TM is a hand-held isotope Spectroscopic Personal Radiation Detector (SPRD) manufactured by Thermo Scientific. The device measures 68mm (W) $\times$ 122mm (L) $\times$ 30mm (T), with a 2.2" backlit LCD display in rugged case. It is equipped three Cadmium-Zinc-Telluride (CdZnTe) detectors, two for monitoring gamma dose rate and one high resolution detector for nuclide identification. An optional ^3He proportional counter is used for detecting thermal neutrons. The two dose-rate detectors are each (15 $\times$ 13 $\times$ 5) mm with a sensitivity (^{137}Cs) of 1.5 cps per $\mu\text{R hr}^{-1}$ and a dose-rate range response of 5 $\mu\text{R hr}^{-1}$ to 100 mR hr^{-1} ($\pm$ 30 %). The high resolution CdZnTe detector volume is 7 $\times$ 7 $\times$ 3.5 mm, attached to a 1024 channel analyzer with an energy resolution (^{137}Cs) 2.4% – 3.5% FWHM (crystal dependent) and range 30keV–1.5 MeV. Best identification was obtained by placing source on the top edge of the device, nearest the internal detector elements. The detector can operate in either Unattended Surveillance Mode (USM) or User Attended Mode (UAM). Figure 2 shows the device measuring a 0.037 MBq (1 μ Ci) ^{137}Cs sealed source.

In UAM, the device can act as finder reporting count rate, monitor dose rate and total dose measured, be placed in identification mode and manually acquire spectra. The identification routine can acquire data from one to thirty minutes as needed to acquire sufficient statistics. The isotope library used by the device covers 28 medical, industrial, NORM, and SNM isotopes with additional 23 available in the extended library. Voice recording and pictures may be taken with the device to document the source measured. In either mode the device can be set with user-defined warning and alarm levels between 250 nSv hr^{-1} and 50 mSv hr^{-1} and 3.0 to 9.9 times the local background rate. The device alerts the user through visual, audio and vibration notification of warnings and alarms. Audio and vibration alarms can be disabled in the settings, allowing for measurements to be taken without alarming members of the public if needed. Figure 3 shows an example ^{137}Cs spectrum. Full width half maximum of the ^{137}Cs 662 keV peak was found to be 21.2 keV (3.2 %) for our device, within the manufacturer specification (Thermo Fisher Scientific Inc. 2009).

Phantom

The attenuation of the gamma spectra of radionuclides is important for both identification and dose calculations. Estimation of the activity and resultant dose from an internal contamination from external measurements is important for applying proper treatment plan. The *Viewpoint* and *Interceptor* systems both use template-based identification software, which can be susceptible to energy-dependent shielding effects. To measure this effect, we used a humanoid phantom to simulate human tissue shielding an internal source of radiation. The Rando[®] humanoid is designed for radiotherapy dose planning and mapping. It is constructed with a natural human skeleton cast inside material that is radiologically equivalent to soft tissue and is sliced into 2.5cm sections (The Phantom Laboratory n.d.). By placing a number of layers between the radioactive source and detector, the attenuation of gamma rays could be measured simulating internal contamination at different tissue depths.

Protocol

We first placed 0.37–0.92 MBq (10–25 μCi) of ^{137}Cs source(s) on a hand cart and varying the number phantom layers placed on top the source. Then, we pushed the cart at a walking pace underneath one of the NaI detectors. This configuration roughly simulates detection of contamination in (on) a patient being transported within the hospital and was very similar to the method used previously in commissioning studies. The trial was repeated at each of the NaI detector locations. The system reliably detected and identified the ^{137}Cs sources except for the lowest activity (0.37 MBq) at the emergency department main entrance, which is the location with the highest ceiling and therefore greatest distance from source to detector. Peak gamma counts did decrease with increasing phantom thickness, but detector response times and the confidence level reported by the identification software were similar both with and without the phantom.

We then developed a fixed position setup by suspending the source from the ceiling one meter from the floor and securing the *Interceptor* on a tripod. The active volume of the device was placed forward of the tripod mount for unobstructed solid-angular coverage and to minimize scattering. We measured at distances from 0.5 to 2.5 meters with and without phantom layers present in front of source. We recorded the count rate, dose (as reported by the device), time-to-identification (TTI), and confidence level (CL) of the isotope assignment at each distance with multiple trials using ^{60}Co (1.09 MBq), ^{137}Cs (3.01 MBq), and ^{133}Ba (4.96 MBq) sources. We altered the number of chest layers of the Rando[®] phantom by clamping them in a vertical position using a cart just beneath the source. Distances were measured from the front of the *Interceptor*[™] to the center of the source to the source externally.

Finally, the device was tested in an ambulance for false alarms and interference due to radio equipment present. The distances from several possible mounting locations to the center of the stretcher were measured.

Results

Interceptor

The *Interceptor* identified a range of medical and industrial sources. Table 1 shows the time required and isotopes identified with associated confidence levels. With the exception of palladium and iridium, the interceptor correctly identified the isotope tested, though occasionally a secondary nuclide was also reported by the device.

Figure 5 shows the count rate measured by the *Interceptor* versus distance from 0.37 MBq (10 μCi) ^{137}Cs , 3.01 MBq (81.3 μCi) ^{137}Cs and 4.96 MBq (134 μCi) ^{133}Ba sources. The average background rate of 6 cps has been subtracted from the data. The data were fit to the function

$$I = A [(x + d_{\text{offset}})^{\alpha} + c]$$

where A , c and α were free parameters in the fit. A is the source term constant and the background c term allows for any increase in background due to backscattering in the room. d_{offset} accounts for the stance to the center of the active volume of the *Interceptor*, from the *Interceptor* manual this is quoted to be 19.75 mm. Ideally α would be two according to the inverse-square-law, but deviation due to scattering is expected. The slope α was found to be 1.73 ± 0.05 for 3.01 MBq (81.3 μCi) ^{137}Cs and 1.76 ± 0.05 for 4.96 MBq (134 μCi) ^{133}Ba . The background offset c was practically consistent with zero for ^{133}Ba , at 0.017 ± 0.016

compared with 0.099 ± 0.023 for ^{137}Cs . There were too few data points for 0.37 MBq (10 μCi), for a meaningful fit but an α of 1.75 gives a qualitatively reasonable description of the data. Using these results one can roughly estimate the measured count-rate for a given distance and source activity.

Phantom

Figure 6 shows the dose rate measured using the *Interceptor* versus increasing thickness of phantom. The dose decreases with increasing thickness of phantom material for a given distance from source to detector. The dose rate falls off more rapidly than the count rate, indicating the detection of scattered photons having less energy. For ^{137}Cs and ^{60}Co , the count rate actually increases slightly for the first few layers of phantom. The deviation between separate measurements of dose rate was usually less than 1% and never greater than 3%.

The acquisition time required by the *Interceptor*'s identification algorithm remained nearly constant for ^{60}Co and ^{137}Cs , but increased by 40–80% for ^{133}Ba . Figure 7 shows time-to-identification (TTI) for the isotopes tested. TTI increased linearly with distance for a given isotope and phantom arrangement. The average deviation between separate measurements of a given setup was 2.6%. The accuracy of isotope identification was isotope dependent. ^{137}Cs was identified with good confidence level with the default acquisition time for all distances and thicknesses of phantom. The confidence level did decrease with distance and thickness but was nearly always greater than 90% and no secondary hypothesis had greater than 55% probability.

The ^{133}Ba and ^{60}Co results varied more greatly, with confidence levels decreasing to 60–70%. Occasionally in ^{133}Ba measurements a secondary isotope was assigned a greater probability, even with as little as 5 cm of phantom at one meter. ^{214}Pb was the most common misidentification for ^{133}Ba , occurring 10 times out of 40 measurements, with greater occurrence at larger distances with increased thickness of phantom material.

To determine whether identification accuracy improved with greater statistics were recorded via a manual device mode to obtain longer acquisition times for ^{60}Co and ^{133}Ba . Figures 8a and 8b show the confidence level of the *Interceptor*'s results with respect to different acquisition times. The ^{133}Ba and ^{60}Co sources were placed one meter from the detector and the phantom layers placed as usual. First data were acquired using the built-in identification-routine's acquisition time. Then the device was placed in its advanced spectrum mode and counts were acquired for approximately 30 minutes, then the spectra identified as usual. The typical acquisition times were 18 and 34 minutes for ^{60}Co for the automatic (short) and manual (long) methods respectively, averaging over different thicknesses of phantom material. For ^{133}Ba the automatic time averaged 4.5 minutes and the manual time averaged 33 minutes.

An examination of the spectra obtained using longer identification periods show the magnitude of the attenuation. Figure 9 shows the ^{133}Ba gamma spectra in air, with phantom and compared to background. The attenuation shows up quite strongly, especially visible in the 81 keV line and the ^{137}Cs x-ray line below it. A closer look at the peaks in the 200–400 keV range, shown in Figure 10, shows the peak bins are in line with their expected values (Heath 1997) for all thicknesses of phantom. From gaussian fits of the most prominent 356 keV gamma the FWHM was found to not vary significantly with phantom thickness from its average value (in mm:ss) of $4:4 \pm 0:3\%$.

Finally, the device was tested in an ambulance for false alarms or interference due to radio equipment present. We observed no change in background count-rate or device operation

from either of the radios present in the vehicle. The size and weight of the device were comparable to other electronics used with several mounting locations possible within 1–1.5 meters of the center of the primary stretcher location.

Discussion

Interceptor

The *Interceptor* detector has the features needed to be a powerful device for detection and identification of gamma-emitting radioisotopes. The small physical form factor of the device can easily be accommodated in a typical ambulance. If placed using a removable mount, the device can be moved closer to the source to obtain greater statistics for better identification and to pinpoint location as needed to discriminate between inhaled, ingested, surface or clothing contamination on the patient through observation of count rate.

The primary limiting factor is the small detector volume. The user can set warning and alarms levels as low as 250 nSv hr⁻¹ with an optional 3.0–9.9 sigma alarm (relative to monitored background), allowing for detection of relatively low activity sources. However, sufficient statistics for accurate identification of the isotopes present require either bringing the device quite near the source (< 20 cm) or a long dwell time (10–30 minutes) to acquire enough counts to correctly identify the radionuclides present.

The industrial check sources tested were of low activity, 3.7–370 kBq (0.01–1 µCi), such that statistics were low even with the sealed source in direct contact with the *Interceptor* unit. ¹⁰³Pd decays primarily by electron capture to ¹⁰³Rh, emitting gamma-rays with 21 keV of energy. 40 keV and 357 keV gamma peaks are also produced at rates of 68 and 22 per 100,000 disintegrations respectively. The 357 keV gamma peak is quite near the 352 keV gamma of ²¹⁴Pb decay and thus is likely confusing the identification algorithm. Attempts to identify a HDR ¹⁹²Ir brachytherapy unit from outside the treatment room were unsuccessful, likely due to shielding of the lower energy gammas of iridium's complex decay scheme.

From our results we can estimate the minimum activity needed for identification within a given time frame. A time-to-identification goal of five minutes at one meter from source with no phantom corresponds to activities of 3.1, 4.0, and 7.6 MBq for ¹³³Ba, ⁶⁰Co, and ¹³⁷Cs respectively. Adding five centimeters of phantom material increases the estimated required activity of ¹³³Ba to 3.9 MBq, while slightly decreasing the activities needed for ⁶⁰Co and ¹³⁷Cs to 3.8 and 7.3 respectively. For 10 cm of phantom the activity needed for ¹³³Ba further increases to 5.0 MBq with ⁶⁰Co unchanged, and 7.4 MBq needed for ¹³⁷Cs. The identification algorithm seems to converge based on accumulated count, thus decay schemes that produce multiple gamma-rays per disintegration such as ⁶⁰Co and ¹³³Ba terminate earlier than for single gamma-ray spectra such as ¹³⁷Cs.

From Table 1 one can see the accuracy of identification is not solely a function of statistics, as several isotopes were identified with high confidence level with relatively low statistics, while other isotopes with more complex decay schemes were more difficult to identify despite larger statistics. This was also present in other tests where ¹³⁷Cs (single gamma peak) was easily identified at large distances, while the higher activity ¹³³Ba source (multiple gamma peaks) had lower reported confidence levels and more frequent misidentifications.

The neutron detection capabilities of the device were not studied in this report.

Phantom

The results of the studies using the Rando[®] phantom show that attenuation of gammas is not negligible in identification of radionuclides from internal contamination. The accuracy of isotope identification decreases with phantom thickness for all isotopes tested. The effect is especially pronounced for ¹³³Ba, for which even a single 2.5 cm layer of phantom can produce inconclusive results or misidentification at one meter from the source. Improving statistics using longer measurement times did little to improve accuracy for deep contamination (large phantom thickness). Thus, it is important to deploy the detector in a position that reduces the amount of tissue between likely source location and the detector by. For inhalation and ingestion mounting the device above the patient on the wall or ceiling of the ambulance would be good choices.

False positives remain a concern, as we did not measure any medical isotopes with the phantom setup. The next stage of testing will focus on source identification and location in ambulance testing. The goal is to distinguish between ingested, inhaled, skin and clothing contamination and correctly identify the isotope(s) present in a reasonable timeframe for emergency response.

Conclusion

The *Interceptor* hand-held isotope spectroscopic personal radiation detector (SPRD) has many features which make it well suited for early radiation detection and identification in a small environment, such as an emergency department exam room or ambulance compartment. The device is sensitive enough to detect low-moderate activity radiation sources. The accuracy of identification depends greatly upon the spectral complexity of the source isotope and attenuation from intervening human tissue in the case of internal contamination. A fixed detector with a larger active volume would allow for faster detection and identification compared to an SPRD, but would not have the flexibility to pinpoint localized contamination without multiple detector elements. A portable device can also be used outside of the ambulance if contamination is suspected beforehand. Such a detector could also be used to complement a more sensitive fixed detector to pinpoint location and identity after an alarm is raised.

References

- Heath, RL. Scintillation Spectrometry, Gamma-ray Spectrum Catalogue. USAEC; 1997.
- IAEA. The Radiological Accident in Goiânia. Vienna: IAEA; 1988.
- Kim, Jong Oh; Huq, MS.; Novotny, J., Jr; Bednarz, G.; Palatine, R.; Reilly, M.; Izadbakhsh, M.; Paris, P.; Greenberger, JS. Performance Characteristics of a Novel Radioactive Isotope Detection and Notification System Designed for Use in Hospitals. Operational Radiation Safety. 2011;71–79.10.1097/HP.0b013e3181fa370f
- McDonald JC. Radiation protection and international intrigue. Radiation Protection Dosimetry. 2007; 123:1–2.10.1093/rpd/ncl569 [PubMed: 17276996]
- Morgan OW, Page L, Forrester S, Maguire H. Polonium-210 poisoning in London: Hypochondriasis and public health. Prehospital and Disaster Medicine. 2008; 23:96–97.10.1017/S1049023X00005665 [PubMed: 18491669]
- Stather JW. The polonium-210 poisoning in London. Journal of Radiological Protection. 2007; 27:1–3.10.1088/0952-4746/27/1/E02 [PubMed: 17341800]
- The Phantom Laboratory. The RANDO[®] Phantom, RAN100 and RAN110. n.d.
www.phantomlab.com
- Thermo Fisher Scientific Inc. Interceptor[™] Spectroscopic Personal Radiation Detector. Dec. 2009
- Von Hippel FM. The radiological and psychological consequences of the Fukushima Daiichi accident. Bulletin of the Atomic Scientists. 2011; 67:27–36.10.1177/0096340211421588

Biography

Ryan Dickson is a postdoctoral research associate at the University of Pittsburgh. He received his PhD degree in physics from Carnegie Mellon University in Pittsburgh, PA where he studied meson spectroscopy as part of the medium energy experimental research group. His primary research experience has been performing analysis of data with significant background noise. His email address is dicksonrd@upmc.edu.

Fig. 1.
UPMC EDNS System

Fig. 2. Thermo Fisher Scientific Interceptor™ hand-held Spectroscopic Personal Radiation Detector (SPRD).

Fig. 3.
Sample spectra obtained during identification of ^{137}Cs as displayed on InterceptorTM screen.

Fig. 4.
The Rando[®] Phantom

Fig. 5. Count rate measured by the Interceptor™ versus distance from 0.37MBq(10μCi) ^{137}Cs , 3.01MBq(81.3μCi) ^{137}Cs and 4.96MBq(134μCi) ^{133}Ba sources. Curve fitted to the data described in the text.

Fig. 6. Dose rate measured by InterceptorTM versus different thicknesses of Rando[®] phantom material using 4.96MBq(134 μ Ci) ^{133}Ba , 3.01MBq(81.3 μ Ci) ^{137}Cs and 1.09MBq(29.4 μ Ci) ^{60}Co sources. Data were averaged over two measurements per setup.

Fig. 7. Time-to-identification measured by InterceptorTM versus different thicknesses of Rando[®] phantom material using 4.96MBq(134 μ Ci) ^{133}Ba , 3.01MBq(81.3 μ Ci) ^{137}Cs and 1.09MBq(29.4 μ Ci) ^{60}Co sources. Data were averaged over two measurements per setup.

Fig. 8. Comparison of confidence level for short (default) versus long (manually acquired) identification dwell times. Distance from source to detector was one meter, with typical

times quoted in the text. Note at 12.5cm of phantom the detector did not identify ^{60}Co as present.

Fig. 9. Comparison of spectra from 4.96MBq ^{133}Ba source in air and with different thicknesses of phantom material. Background taken without source present also shown.

Fig. 10. Inspection of ^{133}Ba gamma ray peaks in the 200–400keV range for different thicknesses of phantom material. The expected energy values are shown as vertical lines.

Table 1

Interceptor response to tested medical and industrial check sources.

Isotope	Counts	TTI	Identification CL
¹⁰³ Pd	151560	60 s	75% ²¹⁴ Pb, 51% ¹⁰³ Pd
¹²⁵ I	493140	60 s	98% ¹²⁵ I
¹⁹² Ir	20000	60 s	Various, no ¹⁹² Ir
²² Na	5100	289 s	91% ²² Na, 44% ¹⁰³ Ru
U(nat.)	252260	240 s	88% U, 41% ⁶⁷ Ga
¹⁵² Eu	48600	60 s	99% ¹⁵² Eu
¹³³ Ba	9480	82 s	95% ¹³³ Ba
¹⁰⁹ Cd	10080	60 s	96% ¹⁰⁹ Cd, 50% ¹³³ Xe
⁵⁷ Co	2580	389 s	99% ⁵⁷ Co
⁶⁰ Co	3900	520 s	90% ⁶⁰ Co, 40% ⁸⁸ Y
²³² Th	12060	148 s	73% ²³² Th, 49% ²⁰⁷ Pb

The default minimum time-to-identification (TTI) for spectrum acquisition for the device is 60 seconds.