Published in final edited form as:

J Am Coll Radiol. 2014 February ; 11(2): 161–168.e3. doi:10.1016/j.jacr.2013.07.004.

Dose reduction efforts for pediatric head CT imaging in Washington State trauma centers: follow-up survey results

Janessa M. Graves, PhD, MPH,

Harborview Injury Prevention and Research Center (HIPRC), Department of Pediatrics, School of Medicine, Box 359960, University of Washington, Seattle, WA 98104, Tel. 206-744-9861 Fax. 206-744-9962, janessa@uw.edu

Kalpana M. Kanal, PhD,

Harborview Injury Prevention and Research Center (HIPRC), Department of Radiology, School of Medicine, University of Washington, Seattle, WA

Frederick P. Rivara, MD, MPH,

Harborview Injury Prevention and Research Center (HIPRC), Department of Pediatrics, School of Medicine, Department of Epidemiology, School of Public Health, University of Washington, Seattle, WA

Jeffrey G. Jarvik, MD, MPH, and

Comparative Effectiveness, Cost and Outcomes Research Center (CECORC), Departments of Radiology and Neurological Surgery, School of Medicine, Department of Health Services, School of Public Health, Department of Pharmacy, School of Pharmacy, University of Washington, Seattle, WA

Monica S. Vavilala, MD

Harborview Injury Prevention and Research Center (HIPRC), Departments of Pediatrics and Anesthesiology & Pain Medicine, School of Medicine, University of Washington, Seattle, WA

Abstract

Purpose—To examine variation in pediatric trauma head CT imaging protocols in Washington State trauma centers (TCs) in 2012 and compare to a previous survey conducted in 2008/09.

Methods—A mixed-mode (online and paper) survey was sent to all adult and pediatric Washington State TCs (levels 1–5). Respondents provided information about the CT scanner used for pediatric head scans and technical information about pediatric dose reduction protocols. Mean head effective dose and organ dose for a female baby were estimated. Results were compared to previous data.

Results—Sixty-one of 76 TCs responded to the 2012 survey (response rate 80.3%; response rate in 2008/09 survey was 76%). In 2012, 91.7% reported having a dedicated pediatric protocol (87.7% in 2008/09). Protective shielding use ranged from 80 to 100% across both survey years. In 2012, 2.5 times more TCs provided sufficient information to conduct dose calculations than in 2008/09. Estimated mean CTDI_{vol} was 23.1 mGy in 2012, compared to 34.8 mGy in 2008/09 ($p=0.01$). Estimated mean DLP was also significantly lower in 2012 than 2009/08 (307.6 mGy*cm

© 2013 American College of Radiology. Published by Elsevier Inc. All rights reserved.

Correspondence to: Janessa M. Graves.

Publisher's Disclaimer: This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our customers we are providing this early version of the manuscript. The manuscript will undergo copyediting, typesetting, and review of the resulting proof before it is published in its final citable form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

vs. 430.1 mGy*cm, respectively; $p=0.04$). Wide variation in mean effective dose was observed for level 3 and 4 TCs in 2012, similar to variation observed in 2008/09 among level 4 TCs. Mean organ dose was significantly lower in 2012 for eye lens and brain, but higher for thyroid than 2008/09 ($p<0.05$).

Conclusions—While most Washington State TCs employ dose reduction protocols for pediatric head CTs, and some measures were lower in 2012, variation in protocols use and estimated dose continues to exist. More complete responses in 2012 suggest improved understanding of the importance of pediatric dose reduction efforts. Education and institutional protocols are necessary to reduce pediatric radiation dose from head CTs.

Keywords

CT imaging protocols; pediatric head CT; trauma center; radiation dose; Washington State

INTRODUCTION

In recent years, numerous studies have reported the rising use of computed tomography (CT) among pediatric patients. [1–4] In the emergency department (ED), where timely diagnosis of traumatic injury is critical, there has been a 5-fold increase in the use of CTs for children from 1995–2008. [4] However, given the recent work showing that clinical decision rules can accurately identify the presence or absence of serious intracranial bleed in the more common mild traumatic brain injury (TBI), [5] researchers and clinical practice guidelines have called for providers and institutions to more carefully weigh the benefit of diagnostic needs vs. the risks posed by imaging. [6–10] Each year, approximately 435,000 children aged 14 years and younger seek ED care for a TBI. [11]

Computed tomography imaging poses a potential risk of radiation exposure, especially among children. [9, 12, 13] Until recently, this association was largely informed by the precautionary principle and research from nonmedical sources (e.g. Chernobyl or Hiroshima), [14, 15]; however, recent studies have since provided evidence for increased cancer risk associated with CTs, specifically of the head region. [13] Pearce and colleagues found significant positive associations between cumulative ionizing radiation dose from childhood CT scans and subsequent leukemia and brain cancer. [13]

To avoid unnecessary risks, it is therefore important to pursue the lowest dose of radiation possible for pediatric patients in need of medical imaging. This can be achieved through education and increased awareness, promotion of alternative nonionizing imaging modalities, utilization review or controls, decision support software, and the use of pediatric protocols when operating CT imaging equipment. [16–18] In order to optimize pediatric dose reduction efforts, it is important to understand how often and where appropriate pediatric CT protocols are used and explore variations in the possible radiation risks from CT imaging. Previous research in Washington State showed substantial variation in radiation dose from pediatric head CT in a statewide survey of trauma centers (TCs) conducted from July 2008 to August 2009. Within level 4 trauma centers, the estimated median effective dose for a baby ranged from 0.60 to 9.60 mSv, a 10-fold variation (median 3.5 ± 0.84 mSv). [19] Since that time, national campaigns, such as Image Gently, have been launched to educate and encourage adoption of child-appropriate protocols when CTs are used. [20, 21] Possible outcomes of education and awareness efforts include increased use of pediatric dose reduction protocols and protective shielding and lower radiation doses from pediatric CT. This study sought to investigate the changes over time by conducting a follow-up survey in Washington State, 4 years after the previous survey. The objective of this study

was to examine variation in pediatric trauma head CT imaging protocols in Washington State trauma centers (TCs) in 2012 and compare to a previous survey conducted in 2008/09.

METHODS

Survey design and contents

We based the current survey instrument on one used in the previous study in Washington State, incorporating changes based on that experience. [19] The current survey comprised of questions asking the following information: 1) Institutional information, 2) CT scanner information, 3) Protocol scan parameters for trauma head CT scan for an infant patient and 4) Protective shielding (survey instrument available online as supplemental material).

Survey sample and administration

We administered the current survey to all Washington State TCs. The state recognizes five levels of trauma services, ranging from level 1 (most comprehensive) to level 5 (basic). All institutions, levels 1–5, that served children (including pediatric-only trauma centers) were considered eligible for this study. There were slight changes in TC designations between 2008/09 and 2012, shown in Table 1.

The current survey was administered by email and mail. First, we obtained email addresses for radiology contacts (chiefs of radiology, lead CT technologists) through professional contacts, websites, and by contacting TCs directly. For TCs with email addresses, an email describing the study was sent, along with a link to the survey and an educational module about radiation dose reduction for pediatric patients. [22] In order to access the educational module, respondents were required to complete the online survey. A reminder email was sent to non-respondents 3 and 6 weeks after the initial contact. Non-respondent institutions were then contacted by telephone and a hardcopy of the survey was sent by mail three times thereafter. Incentives of \$5 gift cards for a popular coffee shop were offered to the first respondent from each TC. If more than one survey was received from an institution, responses were compared and the most complete response (with the fewest missing items) was used. The survey was administered from April to December 2012.

Descriptive data about TCs were collected from multiple sources. We used existing information from the Washington State Department of Health to categorize institutions by TC designation (levels 1–5) and population served (adult and pediatric, or pediatric only). [23] To categorize TCs by American College of Radiology (ACR) accreditation for pediatric CT imaging, we cross-referenced the list of TCs with the list on the ACR's Web site. [24] Finally, we used data from the Washington State Department of Health Community Hospital Quarterly Files to describe TCs bed size and number of emergency visits for 2011. [25] In cases where these data were not available, the American Hospital Directory was used to ascertain bed size information (N=4). [26]

Because survey questions inquired about institutional and not individual practices, this study was deemed to be exempt from human subjects review by the University of Washington institutional review board.

Measures

Based on questions in the survey instrument, three main outcome measures were developed: use of a dedicated pediatric protocol, use of shielding, and estimated radiation dose. In order to compare the results of this survey to the 2008/09 data, survey questions focused on CT scans for head trauma among pediatric patients. Other scanning modalities or body parts were not covered.

Dedicated pediatric protocol—Each institution was asked whether it had a dedicated protocol for a pediatric trauma patient receiving a head CT scan, and if so, what year this protocol was implemented. Having a dedicated protocol is indicative of institutional awareness and action to address pediatric radiation dose exposure. This question was asked in the previous version of the survey as well.

Protective shielding—Use of shielding was ascertained in the survey questionnaire for each TC in the previous and current survey. Respondents were asked if their hospital uses external shielding to protect any organs during head CT scans, and if so, which organs are shielded and the type of shielding used.

Estimated radiation dose—Respondents were asked to report the manufacturer and model for the CT scanner and protocols used most often used for pediatric patients. We calculated the estimated effective dose (ED) and estimated absorbed dose for organs - brain, eye lens, and thyroid for a hypothetical patient with a head trauma. We also calculated CT Dose Index (CTDI_{vol}) and Dose Length Product (DLP), additional measures of estimated dose, for each TC. In order to compare with the previous survey, estimated dose was calculated for a female infant (length, 57 cm; anteroposterior size, 12.2 cm; weight, 4.2 kg) for all facilities that reported technical data. Rationale for selecting these hypothetical patient characteristics was explained previously. [19] In short, a female infant represents a patient with the highest potential risk for exposure and sensitivity to medical radiation. We calculated estimated doses using the CT-Expo Dosimetry Spreadsheet version 1.7 (Medizinische Hochschule, Hannover, Germany), a tool used previously for calculating patient organ and ED from CT scans. [27] ED is measured in milliSieverts (mSv), CTDI_{vol} and absorbed organ dose are measured in milliGray (mGy), and DLP is measured in milliGray * centimeter (mGy*cm). If information about more than one CT scanner was reported (e.g., institutions use scanners in equal frequency), we reported the mean dose estimates for that institution.

Analysis

Data merging, cleaning, and statistical analysis were conducted using Stata/SE 11.2 statistical software. [28] We calculated descriptive statistics for respondents and non-respondents (i.e. TC level, population served, ACR accreditation, bed size, emergency visits) and compared them using χ^2 tests (substituted in pertinent situations with Fisher's exact test). We described the overall proportion of TCs that used a dedicated pediatric protocol and protective shielding. ED, CTDI_{vol}, DLP and organ dose estimates were summarized using means, standard deviations, medians, and inter-quartile ranges. We used χ^2 tests to compare the proportion of TCs that reported using a dedicated pediatric protocol and protective shielding in 2012 to the proportion reported in 2008/09. Student t-tests were used to compare dose estimates between baseline and follow-up.

RESULTS

Of the 79 designated TCs in 2012, 3 reported not having a CT scanner on site. Of the remaining 76, we received 65 replies, yielding 61 usable responses for analysis: a completion rate of 80.3% (Figure 1). (Four responses had incomplete information that prevented their inclusion in analyses.) Sixty general TCs (levels 1–5) and one pediatric-only TC responded to the survey (Table 2). The distribution of respondents across TC levels in 2012 was not significantly different from 2008/09 (Fisher's exact test, $p=0.87$). The respondent group closely resembled non-respondents in terms ACR CT accreditation, bed size, and ED visits (no significant differences, Table 2).

Use of dedicated pediatric protocols

Most responding TCs reported used a dedicated pediatric dose reduction protocol in 2012 (91.7%) (Table 3). Eighty-four percent (N=46) of TCs that reported use of a protocol provided the year of adoption; in 39%, this occurred prior to 2007, 24% between 2007 and 2010, and 37% in or after 2010. There was no significant difference in year of adoption (grouped in categories above) across TC level ($p=0.73$).

In 2008/09, fewer reported using a dedicated pediatric dose reduction protocol (87.8%) compared to 2012, though this difference was not statically significant ($p=0.54$). The proportion of responding TCs that reported using pediatric protocols were not significantly different in 2012 than 2008/09 ($p=0.49$) (Figure 2).

Use of protective shielding

In 2012, 87.9% of responding TCs reported using protective shielding when conducting head CT scans on pediatric patients, compared to 84.3% in 2008/09 ($p=0.59$) (Table 3). Most centers that reporting using shielding in 2012 indicated use of bismuth shielding (53.6%, N=30); the remainder used either lead (N=19) or combination of lead and bismuth shielding (N=1) (one center did not answer this question).

Across trauma level designations, a non-significantly higher proportion of respondents reported using shielding in 2012 than in 2008/09 (with the exception of one level 1 TC, which reported having a CT scanner that cannot use shielding due to its automatic tube current modulation technology; in this case use of shielding would increase the radiation dose and is therefore not employed) (Figure 2).

Estimated Radiation Dose

For all responding TCs in 2012, 48 (78.7%) provided sufficient information to conduct dose calculations, between 2 to 3 times as many as provided data in 2008/09 (N=19). [19] (For these analyses, we were able to access 17 responses from 2008/09 to conduct comparisons.) In 2008/09, significantly fewer hospitals (N=3) reported using helical mode for scanning, compared to in 2012 (N=24; $p<0.01$). Estimated mean CTDI_{vol} was 23.1 mGy in 2012, compared to 34.8 mGy in 2008/09 ($p=0.01$). Estimated mean DLP was also significantly lower in 2012 than 2009/08 (307.6 mGy*cm vs. 430.1 mGy*cm, respectively; $p=0.04$). No significant difference was observed for estimated effective dose (Table 3). For eyes, the estimated mean dose in 2012 was significantly lower than 2008/09 (30.3 mGy vs. 45.2 mGy, respectively) ($p=0.01$). The estimated organ dose for thyroid increased significantly from 6.8 mGy in 2008/09 to 14.7 mGy in 2012 ($p=0.02$). There was not a significant change in brain organ dose between the two time-points.

The 2008/09 survey results reported a 10-fold variation in estimated ED across level 4 TCs. [19] The 2012 survey shows similar variation among level 4 TCs (range: 0.9–9.2 mSv), and also 10-fold variation among level 3 TCs (range 0.9–9.1 mSv) that was not observed in 2008/09 (Figure 3).

DISCUSSION

While CT imaging is an extremely valuable tool, it brings with it risks to the patient. Adoption of appropriate dose reduction protocols helps reduce radiation risk while maintaining diagnostic imaging quality. [29, 30] This study serves as a follow-up to a previous survey of Washington State TCs to understand their use of dose reduction protocols among pediatric patients receiving head CT scans.

Most Washington State TCs reported use of dedicated pediatric protocols to reduce radiation dose among children receiving head CT scans at their institution in 2012. In the 2008/09 survey, estimated effective dose showed substantial variation among level 4 TCs, which persisted in 2012. Increased reporting among level 3 TCs in 2012 allowed for dose calculations among those institutions, which also showed considerable variation in estimated effective dose. Overall, item response rates were higher in 2012 than in 2008/09, which allowed for a more complete analysis of results. Improved item response in 2012 may be associated with greater understanding of the importance of dose reduction among pediatric patients and increased familiarity of dose reduction efforts, such as Image Gently, which was implemented in 2008 and has gained exposure since. [20]

Estimated radiation dose from head CT scans decreased from 2008/09 to 2012 for some measures, whereas for others, estimated mean dose did not change (e.g., effective dose and brain organ dose) or increased (e.g. thyroid organ dose). This variation may be due to the greater proportion of helical scan mode used in the current survey compared to axial scan mode previously. [31]

Our study showed that while use of protective shielding was widely reported among Washington State TCs in 2012, a smaller proportion of respondents indicated thyroid shielding compared to the previous survey. It is possible that respondents did not provide complete information when answering this open-ended question in the survey. While open-ended questions may provide greater depth of information, closed-ended (fixed-response) questions generally place a lesser burden on the respondent. In the future, a closed-ended question with check-boxes for organs could be considered. Shielding is a simple and effective method of dose reduction, [32] but its use is not universally recommended as use of shields during scanning may introduce noise or increase beam-hardening artifacts. [33] Also, proper use of shielding depends on the design of an individual scanner – a universal approach to shielding is not appropriate for all CT scanners. In fact, inappropriately used shielding may lead to increased radiation dose. [16]

Online educational modules inform radiologists and CT technologists about appropriate CT protocols and strategies to optimize image quality at the lowest dose possible. [34] In Washington State, the Department of Health supported the creation of a free, instructive web-based module that covers the following topics: CT dose monitoring, CT dosimetry, and factors that influence CT dose. Participants are also provided a list of educational resources for future reference. As part of the 2012 survey, TCs were provided with a link to access this module. (It is also accessible to the public at <http://depts.washington.edu/hiprc/Education%20and%20Training/CTModule.html>.)

The lack of significant differences between 2008/09 and 2012 in terms of the proportion of TCs using dose reduction protocols or the estimated effective dose may be due to a ceiling effect. Even in 2008/09, most responding trauma centers reported using pediatric dose reduction protocols, leaving limited room for improvement in 2012. Currently, Washington State does not have any state regulations for users of CT scanners to monitor or report dose. Therefore, the variation we observed in estimated dose may be due to the lack of monitoring or enforcement. Ideally, all TCs should have pediatric dose reduction protocols in place, and substantial variation in dose should not be present. Variation in dose may also be attributed to the wide range of machines present in TCs, including older models. Modern CT scanners have dose reduction options such as automatic tube current modulation, which automatically decreases or increases the tube current depending on anatomy being scanned. Over time, one would expect institutions to upgrade to more modern machines that provide more dose reduction technology than old scanners. Our survey did not explicitly ask about these options, which might have resulted in an overestimation of estimated dose for institutions

with such modern scanners. Finally, the unit of analysis for this study was the institution and individual impacts were not captured. While we did not observe significant changes from 2008/09 to 2012, if one institution shows improvement of CT practices (either adoption of a pediatric protocol or reduction in estimated effective dose), many patients will be affected.

The wide variation observed among level 3 and 4 TCs provides evidence for additional efforts to reduce exposure to CT radiation among pediatric patients. Educational campaigns, such as Image Gently and the local web-module, are crucial to support ongoing learning and skill development among CT radiologists and technologists. Appropriate, pediatric dose reduction protocols should be employed at each institution. Educational efforts and institutional protocols are necessary to reduce pediatric radiation dose from head CTs. Involvement on the part of the radiologist, technologist and physicist if available is critical for dose reduction efforts. Forming a CT protocol review committee and reviewing protocols periodically aids in dose reduction. Studies have shown that systematic use of CT dose reduction protocols can result in significant dose reduction while maintaining image quality. [35, 36]

This study has several limitations. First, the outcome measures from this study are based solely on self-report. We did not confirm or evaluate the compliance with the use of a dedicated pediatric protocol or shielding at each hospital. Dose calculations are general estimates based on information provided by respondents. If a respondent inaccurately reported one of the dose inputs, the estimated dose may be erroneously high or low. We were not able to verify the results of the surveys with institutions. In addition, information about image quality at each institution is not known. We are therefore not able to make conclusions about the relationship between image quality and dose for individual institutions. Second, it is possible that institutions may collaborate as radiology groups, thus sharing decisions about protocols. We were not able to measure this aspect in this study. Finally, while we were only to access raw data for 17 responses from the 2008/09 survey to conduct comparisons, our data were also compared to the published 2008/09 results.

In summary, we conducted a state-wide survey of Washington TCs to evaluate pediatric head CT practices and estimated radiation dose in 2012 and describe changes since 2008/09. The variation in pediatric head CT imaging protocols and radiation dose continues among Washington State TCs. For level 3 and 4 TCs, there was more than 10-fold variation in mean effective dose in 2012, similar to wide variation observed in 2008/09 among level 4 TCs.

Acknowledgments

We gratefully acknowledge Harriet Saxe and Leah Thompson for their assistance in the data collection process and Mary Rotert at the Washington State Department of Health for providing information on trauma center designations. We would like to acknowledge the Washington State trauma centers that participated in this study and the Washington State Department of Health and the Emergency Medical Services for Children (EMSC) Program for its support of this work. Dr. Graves receives fellowship support from the National Institute of Child Health and Human Development (T32HD057822). This work was also supported by the Harborview Injury Prevention & Research Center, University of Washington.

REFERENCES

1. Blackwell CD, Gorelick M, Holmes JF, Bandyopadhyay S, Kuppermann N. Pediatric head trauma: changes in use of computed tomography in emergency departments in the United States over time. *Annals of emergency medicine*. 2007; 49(3):320–324. [PubMed: 17145113]
2. Broder J, Fordham LA, Warshauer DM. Increasing utilization of computed tomography in the pediatric emergency department, 2000–2006. *Emergency radiology*. 2007; 14(4):227–232. [PubMed: 17505849]

3. Hryhorczuk AL, Mannix RC, Taylor GA. Pediatric abdominal pain: use of imaging in the emergency department in the United States from 1999 to 2007. *Radiology*. 2012; 263(3):778–785. [PubMed: 22535565]
4. Larson DB, Johnson LW, Schnell BM, et al. Rising Use of CT in Child Visits to the Emergency Department in the United States, 1995–2008. *Radiology*. 2011
5. Kuppermann N, Holmes JF, Dayan PS, et al. Identification of children at very low risk of clinically-important brain injuries after head trauma: a prospective cohort study. *Lancet*. 2009; 374(9696): 1160–1170. [PubMed: 19758692]
6. Smith-Bindman R, Miglioretti DL, Larson EB. Rising use of diagnostic medical imaging in a large integrated health system. *Health affairs (Project Hope)*. 2008; 27(6):1491–1502. [PubMed: 18997204]
7. Brenner DJ, Hall EJ. Computed Tomography — An Increasing Source of Radiation Exposure. *New England Journal of Medicine*. 2007; 357(22):2277–2284. [PubMed: 18046031]
8. American College of Radiology and Society for Pediatric Radiology. ACR-SPR practice guideline for the performance of pediatric computed tomography (CT). 2008 Available from: http://www.acr.org/SecondaryMainMenuCategories/quality_safety/guidelines/pediatric/PediatricCT.
9. Brody AS, Frush DP, Huda W, Brent RL. a.t.S.o. Radiology. Radiation Risk to Children From Computed Tomography. *Pediatrics*. 2007; 120(3):677–682. [PubMed: 17766543]
10. Dorfman AL, Fazel R, Einstein AJ, et al. Use of Medical Imaging Procedures With Ionizing Radiation in Children: A Population-Based Study. *Arch Pediatr Adolesc Med*. 2011; 165(5):458–464. [PubMed: 21199972]
11. Langlois, JA.; Rutland-Brown, W.; Thomas, KE. Traumatic Brain Injury in the United States: Emergency Department Visits, Hospitalizations, and Deaths. Atlanta (GA): Centers for Disease Control and Prevention, National Center for Injury Prevention and Control; 2004.
12. Huda W, Vance A. Patient radiation doses from adult and pediatric CT. *AJR. American journal of roentgenology*. 2007; 188(2):540–546. [PubMed: 17242266]
13. Pearce MS, Salotti JA, Little MP, et al. Radiation exposure from CT scans in childhood and subsequent risk of leukaemia and brain tumours: a retrospective cohort study. *Lancet*. 2012; 380(9840):499–505. [PubMed: 22681860]
14. Pierce DA, Preston DL. Radiation-related cancer risks at low doses among atomic bomb survivors. *Radiation research*. 2000; 154(2):178–186. [PubMed: 10931690]
15. Preston DL, Ron E, Tokuoka S, et al. Solid cancer incidence in atomic bomb survivors, 1958–1998. *Radiation research*. 2007; 168(1):1–64. [PubMed: 17722996]
16. Strauss KJ, Goske MJ, Kaste SC, et al. Image Gently: Ten Steps You Can Take to Optimize Image Quality and Lower CT Dose for Pediatric Patients. *American Journal of Roentgenology*. 2010; 194(4):868–873. [PubMed: 20308484]
17. Coakley FV, Gould R, Yeh BM, Arenson RL. CT Radiation Dose: What Can You Do Right Now in Your Practice? *American Journal of Roentgenology*. 2011; 196(3):619–625. [PubMed: 21343506]
18. Brenner DJ. Slowing the increase in the population dose resulting from CT scans. *Radiation research*. 2010; 174(6):809–815. [PubMed: 20731591]
19. Kanal KM, Vavilala MS, Raelson C, et al. Variation in Pediatric Head CT Imaging Protocols in Washington State. *Journal of the American College of Radiology*. 2011; 8(4):242–250. [PubMed: 21458762]
20. Goske M, Applegate K, Boylan J, et al. The ‘Image Gently’ campaign: increasing CT radiation dose awareness through a national education and awareness program. *Pediatric Radiology*. 2008; 38(3):265–269. [PubMed: 18202842]
21. Goske MJ, Applegate KE, Boylan J, et al. The Image Gently Campaign: Working Together to Change Practice. *American Journal of Roentgenology*. 2008; 190(2):273–274. [PubMed: 18212208]
22. Kanal KM, Vavilala MS. Pediatric Dose Reduction in Pediatric Head CT Imaging Training Module. 2012 Available from: <http://depts.washington.edu/hiprc/Education%20and%20Training/CTModule.html>.

23. Office of Community Health Systems. Washington State Trauma Services, DOH Publication Number 530-101. In: Washington State Department of Health. , editor. 2011.
24. American College of Radiology. Accredited Facility Search. Available from: <http://www.acr.org/Quality-Safety/Accreditation/Accredited-Facility-Search>.
25. Washington State Department of Health. [cited 2013 February 4] Community Hospital Quarterly Files, 2011. 2012. Available from: <http://www.doh.wa.gov/DataandStatisticalReports/HealthcareinWashington/HospitalandPatientData/HospitalFinancialData/QuarterlyReports.aspx>.
26. American Hospital Directory. Free Hospital Profiles. 2012 Available from: <http://www.ahd.com/freesearch.php>.
27. Stamm G, Nagel HD. CT-expo--a novel program for dose evaluation in CT. RoFo : Fortschritte auf dem Gebiete der Rontgenstrahlen und der Nuklearmedizin. 2002; 174(12):1570–1576. [PubMed: 12471531]
28. LP, S. Stata/SE 11.2 for Windows. StataCorp LP: College Station, TX; 2012.
29. Siegel MJ, Schmidt B, Bradley D, Suess C, Hildebolt C. Radiation dose and image quality in pediatric CT: effect of technical factors and phantom size and shape. Radiology. 2004; 233(2): 515–522. [PubMed: 15358847]
30. Yu L, Bruesewitz MR, Thomas KB, et al. Optimal tube potential for radiation dose reduction in pediatric CT: principles, clinical implementations, and pitfalls. Radiographics : a review publication of the Radiological Society of North America, Inc. 2011; 31(3):835–848.
31. Bushberg, JT.; Seibert, JA.; Leidholdt, EM.; Boone, JM. The Essential Physics of Medical Imaging. 3rd ed.. Philadelphia: Pennsylvania: Lippincott Williams & Wilkins; 2011.
32. Curtis JR. Computed tomography shielding methods: a literature review. Radiol Technol. 2010; 81:428–436. (1943–5657 (Electronic)). [PubMed: 20445137]
33. McCollough CH, Primak AN, Braun N, et al. Strategies for reducing radiation dose in CT. Radiologic clinics of North America. 2009; 47(1):27–40. [PubMed: 19195532]
34. Goske MJ, Applegate KE, Bell C, et al. Image Gently: providing practical educational tools and advocacy to accelerate radiation protection for children worldwide. Seminars in ultrasound, CT, and MR. 2010; 31(1):57–63.
35. Smith AB, Dillon WP, Lau BC, et al. Radiation dose reduction strategy for CT protocols: successful implementation in neuroradiology section. Radiology. 2008; 247(2):499–506. [PubMed: 18372456]
36. Zacharia TT, Kanekar SG, Nguyen DT, Moser K. Optimization of patient dose and image quality with z-axis dose modulation for computed tomography (CT) head in acute head trauma and stroke. Emergency radiology. 2011; 18(2):103–107. [PubMed: 20857165]

TAKE HOME POINTS

- Most Washington State trauma centers employ pediatric dose reduction protocols for head CTs.
- There is wide variation in estimated dose among level 3 and 4 trauma centers in 2012.
- Survey responses were much more complete in 2012 than 2008/09, suggesting improved understanding of the importance of pediatric dose reduction efforts.
- Educational efforts and institutional protocols are necessary to reduce pediatric radiation dose from head CTs.

Figure 1. Response rate calculation for 2012 survey of Washington State trauma centers (using Council of American Survey Research Organizations formula)

The response rate was the number of hospitals who completed surveys (61) divided by the sum of the number who were eligible and initially consented (75) plus the estimated number who were eligible among those who refused (1): $61/(75 + 1) = 80.3\%$.

Figure 2. Dose reduction efforts across trauma level designations in 2008/09 and 2012

Data from 2008/09 were obtained from previous publication (Kanal et al, 2011). No significant differences comparing dose reduction efforts across trauma level designations and survey years ($p>0.05$). One level 1 TC reported having a CT scanner that cannot use shielding due to its automatic tube current modulation technology; in this case use of shielding would increase the radiation dose and is therefore not employed.

Figure 3. Estimated effective dose (mSv) across TC levels in 2008/09 and 2012

Dose estimates based on calculations for a female infant (aged 1 year). Data from 2008/09 reprinted from the Journal of the American College of Radiology, Vol 8/Issue 4, Kanal, K. M., M. S. Vavilala, C. Raelson, A. Mohan, W. Cohen, J. Jarvik, F. P. Rivara, and B. K. Stewart. Variation in Pediatric Head CT Imaging Protocols in Washington State, Pages 242–250, Copyright (2011), with permission from Elsevier.

Table 1

Washington State trauma center designations.

General trauma level designation	2008/09 [*]		2012 [†]	
	Adult	Pediatric	Adult	Pediatric
1	1	1	1	1
2	4	0	4	1
3	23	5	25	6
4	33	0	33	0
5	15	0	16	0
Pediatric-only [‡]	-	2	-	1
All Levels	76	8	79	9

Values indicate number of trauma centers (by level of designation) in Washington State at the time when the surveys were administered. Unless otherwise indicated, pediatric trauma centers also serve as adult trauma centers. Data provided by the Washington State Department of Health, Office of Emergency Medical Services and Trauma System.

^{*} Current as of 11/31/2008.

[†] Current as of 08/31/2011

[‡] Level 2 pediatric-only trauma center

Table 2

Characteristics of trauma centers in 2012 survey, respondents and nonrespondents.

	Respondents (N=61)		Nonrespondents (N=18)		All Washington Trauma Centers (N=79)	
	Number	Percent	Number	Percent	Number	Percent
General trauma center level						
1	1	2%	0	0%	1	1%
2	2	3%	2	11%	4	5%
3	18	30%	6	33%	24	30%
4	28	46%	5	28%	33	42%
5	11	18%	2	28%	13	20%
Pediatric-only*	1	2%	0	0%	1	1%
Population served						
Pediatric only	1	2%	0	0%	1	1%
Adult only	54	88%	17	94%	71	90%
Both	6	10%	1	6%	7	9%
American College of Radiology CT accreditation[†]						
Yes	19	31%	7	39%	26	33%
No	42	69%	11	61%	53	67%
Bed size[‡]						
0–50	24	39%	7	39%	31	39%
51–150	19	31%	3	17%	22	28%
150+	18	30%	6	33%	24	30%
Emergency department visits[‡]						
5,000	17	28%	5	28%	22	28%
5,001–15,000	13	21%	2	11%	15	19%
15,001–50,000	23	38%	4	22%	27	34%
> 50,001	8	13%	5	28%	13	16%

Values in table indicate the number and percentage of trauma centers that responded or did not respond to the 2012 survey. Percentages may not sum to 100% due to rounding or missing responses. No significant differences were observed between respondents and non-respondents (χ^2 and Fisher's exact tests). Missing values were excluded from significance testing.

* Level 2 pediatric-only trauma center

[†]From American College of Radiology accredited facility search (<http://www.acr.org/Quality-Safety/Accreditation>) (Accessed 08/31/2012)

[‡]From Washington State Department of Health, Community Hospital Quarterly Files, 2011. (<http://www.doh.wa.gov/DataandStatisticalReports/HealthcareinWashington/HospitalandPatientData/HospitalFinancialData/QuarterlyReports.aspx>). Accessed 03/03/2013. Bed size data for hospitals that did not submit data to the State Department of Health were obtained from the American Hospital Directory (www.ahd.com) profile search (N=4).

Table 3

Survey results from general trauma centers in 2008/09 and 2012.

	2008/09 Survey	2012 Survey	Sig
Use of a dedicated pediatric protocol, N(%)	N=49	N=60	0.54
Yes	43 (87.8%)	55 (91.7%)	
No	6 (12.2%)	5 (8.3%)	
Use of protective shielding, N(%)	N=51	N=58	0.59
Yes	43 (84.3%)	51 (87.9%)	
No	8 (15.7%)	7 (12.1%)	
	2008/09 Survey	2012 Survey	
Estimated dose	N=17	N=48	Sig
Effective dose, (mSv)			
Mean (SD)	3.1 (2.2)	3.0 (1.8)	0.89
Median (IQR)	2.6 (1.7)	2.6 (2.0)	
CTDI_{vol}, (mGy)			
Mean (SD)	34.8 (25.6)	23.1 (10.9)	0.01
Median (IQR)	30.9 (15.5)	20.7 (15.0)	
DLP, (mGy*cm)			
Mean (SD)	430.1 (315.7)	307.6 (161.4)	0.04
Median (IQR)	366.9 (233.9)	273.0 (198.5)	
	2008/09 Survey	2012 Survey	
Estimated organ dose	N=17	N=48	Sig
Brain, (mGy)			
Mean (SD)	39.0 (28.8)	25.4 (12.0)	0.08
Median (IQR)	34.7 (17.3)	22.4 (15.5)	
Eye lens, (mGy)			
Mean (SD)	45.2 (33.7)	30.3 (14.2)	0.01
Median (IQR)	40.1 (20.0)	26.6 (18.3)	
Thyroid, (mGy)			
Mean (SD)	6.8 (5.7)	14.7 (12.8)	0.02
Median (IQR)	4.6 (3.3)	9.4 (11.6)	

Values in table indicate responses to 2008/09 and 2012 surveys. Values represent the number (percentage), mean (SD), or median (IQR) as indicated. Significance values represent results from χ^2 test (p-values) comparing proportions between the 2008/09 and 2012 surveys or unpaired Student t-tests for dose comparisons. For comparison purposes, pediatric-only trauma centers are excluded from these analyses.

Abbreviations: IQR, inter-quartile range; SD, standard deviation; DLP, dose-length product; mSv, millisievert; mGy, megagray.