Published in final edited form as:

Annu Rev Biomed Eng. 2013 ; 15: 383–405. doi:10.1146/annurev-bioeng-071910-124640.

Implants and Decoding for Intracortical Brain Computer Interfaces

Mark L. Homer, M.S.¹, Arto V. Nurmikko, PhD², John P. Donoghue, Ph.D.^{4,3}, and Leigh R. Hochberg, M.D., Ph.D.^{4,2,5}

Mark L. Homer: Mark_Homer@brown.edu; Arto V. Nurmikko: Arto_Nurmikko@brown.edu; John P. Donoghue: John_Donoghue@brown.edu; Leigh R. Hochberg: Leigh_Hochberg@brown.edu

¹Biomedical Engineering, Brown University, Providence, RI

²School of Engineering, Brown University, Providence, RI

³Department of Neuroscience, Brown University, Providence, RI

⁴Center for Neurorestoration and Neurotechnology, Veterans Affairs Medical Center, Providence, RI

⁵Neurology, Massachusetts General Hospital, Harvard Medical School, Boston, MA

Abstract

Intracortical brain computer interfaces (iBCIs) are being developed to enable a person to drive an output device, such as a computer cursor, directly from their neural activity. One goal of the technology is to help people with severe paralysis or limb loss. Key elements of an iBCI are the implanted sensor that records the neural signals and the software which decodes the user's intended movement from those signals. Here, we focus on recent advances in these two areas, with special attention being placed on contributions that are or may soon be adopted by the iBCI research community. We discuss how these innovations increase the technology's capability, accuracy, and longevity, all important steps that are expanding the range of possible future clinical applications.

Keywords

brain machine interfaces; multielectrode arrays; signal processing; neural engineering

Introduction

Voluntary limb movement in primates, including able-bodied people, is achieved in part by the conveyance of high level motor commands from neurons in the motor cortex to neurons within the spinal cord. In response to input from these so-called “upper motor neurons”, as well as inputs from other descending systems and cells local to the spinal cord, alpha (“lower”) motor neurons then activate muscle fibers generating motion. Injury or disease, such as brainstem stroke or loss of a limb, can disconnect the cortex from its effector target. Ongoing research is focused on building technological bridges between the motor cortex and external devices in order to restore motor function. When these systems work as designed, the intent to move again becomes translated into overt action.

Various labeled as neural interfaces, neural prostheses, brain machine interfaces (BMI) or brain computer interfaces (BCI), this technology promises to be useful to a large number of people. In the United States alone, it is estimated that 1.6 million Americans live with limb loss (1) and 5.5 million people experience some form of paralysis (2) including 50,000 individuals who have complete tetraplegia (severe functional impairment or paralysis of both arms and legs) (3). Commonly, people with tetraplegia rely on others to perform routine, everyday actions. Responding to the need for better assistive devices, BCI technology could provide a control signal for any range of devices: a computer, robotic or prosthetic arm, or a powered wheelchair. For a true restorative function, people suffering from paralysis of an arm might, in the future, be able to recover control by re-animating the affected muscles via brain-controlled functional electrical stimulation (FES); others with limb loss could employ a robotic prosthesis that emulates arm and hand actions.

This review focuses on intracortical brain computer interfaces (iBCIs) where, for example, microscale probes having 96 fine tipped microelectrodes in a compact 4×4 mm array are inserted into the cortex (reviews of scalp-based electroencephalography and brain surface-based electrocorticography BCIs can be found in (4) and (5) respectively). The sensors are sensitive enough to pick up the discrete all-or-none output of single neurons, the action potential, commonly referred to as a “spike”, as well as the summed voltage fluctuations from small to large numbers of neurons, called field potentials. Each electrode provides spiking from up to a few neurons (typically 1 or 2), yielding the population's time evolving output pattern. These represent but a small sample of the entire set of neurons in this limited region, as spiking can only be detected by microelectrodes closely approximated to a neuron.

Spikes are desired because they are the information-rich signal nearly universally employed by neurons for communication in the brain. Beginning approximately 50 years ago, experiments by Evarts and others helped to reveal how individual cortical neurons, recorded one at a time, control voluntary movement, helping to lay the groundwork for current iBCI work, e.g., (6–11). Multielectrode approaches helped to provide (12, 13) early examples of real-time prediction of an animal's wrist position from the decoded motor cortex output (14, 15), and enabled animals to perform simple tasks via an external device (16). The first clinical realization of these technologies occurred in 1998, when a person with amyotrophic lateral sclerosis (ALS) paralysis controlled an on-off switch using two electrodes implanted in the motor cortex (17).

Based upon the many years of publicly funded basic science, research beginning in 2000 demonstrated that monkeys could use motor cortex spiking patterns to control a computer cursor in two or three dimensions (18–22). Soon afterwards, in 2004, the first demonstrations of intracortically-directed 2D cursor movements and simple robotic control were accomplished by people with tetraplegia using an iBCI (23). This was followed by iBCI-enabled 3-D control of a prosthetic arm by non-human primates (24). Most recently, multi-dimensional robotic arm control was successfully achieved for reaching and grasping actions by two people with tetraplegia (25). The future promise of brain-controlled FES devices was further demonstrated by iBCI-driven muscle control in monkeys with temporary arm paralysis (26–28), further underscoring the potential highlighted in an initial demonstration by a human with paralysis using an iBCI to control a simulated, dynamic FES system and arm (29).

Modern iBCIs have several major components, illustrated in Figure 1. The user imagines or attempts a movement, resulting in motor cortical neuronal activity. An implanted multiprobe sensor of electrodes, commonly referred to as a microelectrode array (MEA), records neural activity in the form of extracellular voltage signals. Data acquisition (DAQ) hardware

transports, amplifies, and digitally samples these signals. The data is processed on a dedicated, real time hardware and software platform, which hosts various decoding routines. In the typical real-time data processing stream, neural signals first are processed to extract informative features related to intended movement, after which estimation techniques translate those features into device commands. The device then carries out the desired action as well as provides feedback to the user.

The core elements of an iBCI (Figure 2), from neural signal collection to actuation of the effector, each require the iBCI neuroengineering team to make important decisions. Each of these elements (which could also be broken into several smaller elements) also presents opportunities for focused research and development. We take the engineering perspective in this review, where the domain of decoding encompasses the combination of the feature extraction, estimation, and calibration elements (30); this viewpoint is used to organize subsequent sections.

Prior to use, the BCI must be calibrated. Plainly stated, a map must be built between the observed neural activity associated with intended movement and the actual movement of the effector. First, the user (with, for example, tetraplegia due to stroke or ALS) engages in a series of instructed behavioral tasks where various real world actions, such as the preprogrammed motion of a computer cursor, are observed while imagined or attempted by the user (who is unable to make them). Next, with both the neural signals and the attempted (or, at least, instructed) movements being known quantities, calibration techniques tune the decoder. It is important to note that during this process only the decoder has been calibrated; no real training or learning has (yet) occurred from the user's standpoint. While reasonable to believe that users presented with a new and reasonably useful tool (an iBCI) will become more proficient in its use over time (31), important questions regarding recorded signal stability as well as the potential for simultaneous co-adaptation (32) of both the decoder and the user yield the clear demonstration of such learning to future research.

A thorough review of the several active research areas highlighted in Figure 2 would extend far beyond this review (see (33) for several excellent chapters in this regard). Here we focus on recent advances using the implanted sensor as well as the state-of-the-art in decoding algorithms. In each case, the emphasis is on results that have been or we believe are likely to be widely adopted by the iBCI research community.

Implant

Microelectrode arrays

Overview—From an electrical and electrophysiological point of view, microelectrode arrays (MEAs) for iBCIs need to be designed considering details of both the electrode-brain tissue interface and the macroscale (i.e. lumped) circuit equivalents. Quantitative modeling for optimizing capture of the neuronal electrical signals is complex, influenced by the role of the electrolytic bilayer at the interface between the electrode recording surface and brain tissue, electrode tip shapes, potential tissue reactions, etc. (34). Metals such as Pt, W, and Ir are chemically stable in the electrolyte-rich and proteinaceous environment of brain parenchyma and have reasonable work functions to match the Na⁺, K⁺, Cl⁻, Ca⁺⁺, and Mg⁺⁺ concentrations that dominate. To capture both spikes and lower frequency field potentials, the equivalent circuit of the single microelectrode/tissue recording interface should typically cover a bandwidth of ≈ 0.5 Hz – 10 KHz. A single action potential might generate recorded amplitudes from few tens up to >100 μ V, registered by an electrode of ~ 500 k Ω in impedance (i.e. sub-nA currents), thereby framing the design requirements for subsequent analog preamplifiers (gain, noise figure etc.), digitizing circuits, and the downstream signal processing and data management electronics. Note that apart from

fundamental sources of noise common to all electrical systems (Johnson noise, etc.), the brain operates within its own inherently “noisy” background (since its circuits are continuously ‘alive’), thereby adding to the challenges of MEAs acquiring consistent neural signals that correlate with specific cognitive process or behavior. Another set of important contemporary questions is how reaction by the body tissue on implantable electrodes and related device structures affects their truly long term chronic performance ($\gg 1$ year).

Early intracortical microelectrode arrays composed of bundles of wires have been now largely replaced by ‘monolithic’ arrays for work in primates. Of these, we mention the Si-based arrays where lithographic and electrochemical techniques are combined to fabricate tapered microscale ‘beds of needles’ (35–37). For example, in the so-called “Utah” MEA (inset of Figure 3), each approximately conically-shaped 1–2 mm long electrode, with its p-Si shank insulated by biocompatible parylene, is coated at its tip by Pt or Pt/Ir. The heavily doped p-Si provides a low-loss conducting path onto the silicone insulated, planar support substrate wherefrom a wirebonded bundle (say, 100 insulated 1 mil Au wires) transports the recorded neural signals through the skull and the skin to exterior electronics (Figure 3). Operationally, specialized neurosurgical techniques are employed to access cortical areas of interest, with MEAs inserted into the brain typically using calibrated pneumatic single-impact pulsed drives. While these MEA implants have shown their mettle as chronically viable implants as described in this review, further improvements in their reliability and further control of their critical dimensions on the micrometer scale precision are expected both from improvements to current manufacturing approaches as well as innovations in both materials and advanced microfabrication (such as wafer scale processing (38)).

Based on silicon substrate shanks, the “Michigan array” can run multiple connective lines down to pads of iridium along a shank to achieve multi-depth electrodes (37). One shank can efficiently support as many as 16 electrode sites at different depths, and can also horizontally cluster electrodes in placing several at the same depth along a single shank (39). Conducting polymer nanotubes have also been used to dramatically decrease impedance when attached to the surface of the electrode sites (40), potentially enabling the creation of thinner but still functional recording electrodes (41) which can be more densely placed in cortex without overly disturbing the surrounding tissue. Yet another technology, the “neurotrophic” electrode (17) aims to provide for long-term intracortical recordings by attracting the growth of neurites into a wire electrode-embedded glass enclosure.

Wireless implants—Several successful clinical technologies have benefitted from having percutaneous components during their initial development and testing (e.g., cardiac pacemakers, deep brain stimulators). Nevertheless, there are a number of aesthetic and safety benefits to developing systems which dispense with an externalized “tether”, and which ultimately become fully implanted. One current research trend focuses on efforts to integrate the intracortical microelectrode recording arrays with wireless electronics so that the physical cabled connection between a subject and external electronics and external assistive devices can be removed. Wireless approaches will greatly improve the mobility for a subject, be that in human neuroprosthetics applications or in freely moving non-human primates in fundamental neuroscience research. Many groups worldwide are now working towards wireless active microelectronic neural interfaces, especially focusing on very low power, application-specific system-on-chip integrated circuits (ASIC/SoC) which consolidate the analog, digital, and telemetry components to a wearable system, respectively. Several sophisticated designs have been demonstrated at the benchtop level. Yet, much of this integrated ASIC-based engineering work still awaits transition to in-vivo use for primates. One successful strategy has been to retain the percutaneous wiring from the MEA to the skull-mounted connector pedestal and affix battery-powered head-mounted

external modules on these pedestals in non-human primates, to demonstrate transmission of multichannel neuronal data as a broadband radiofrequency digital stream (42–45).

Transcutaneous systems—The emergent wireless approaches to iBCIs have the goal of eventual implementation as fully implanted electronic microsystems whereby neural signals from MEAs are amplified and transmitted from hermetically sealed modules which are embedded entirely within the body (head, or head and chest), without any skin-penetrating wiring. The ability to employ fully wireless, and eventually bidirectional, iBCI interfaces that are enveloped by the body's most potent protection against infections etc., namely the skin, opens up potentially entirely new vistas for next generations of neurotechnologies. As a state-of-the-art example (ca. 2012), we show in Figure 4 a 100-channel wireless neural broadband implant which has been successfully tested in mobile pig and non-human primate models (D. Borton, M. Yin, J. Aceros, W.R. Patterson, and A.V. Nurmikko, in progress) (46). The device shares a commonality with cochlear and cardiac implants in that the active microelectronics, battery power, and telemetry circuits are encased within a hermetically sealed titanium enclosure. By contrast to these low data rate neural stimulation devices, however, the amount of information which is being acquired wirelessly from the brain and transmitted through the skin to nearby receivers for decoding is huge (on the order of 50 MBits/sec at this writing, and expected to increase in the future). Such demands on data rates not only impact the required sophistication in designing and fabricating the wireless implants, but will require new strategies for handling and storing user-specific neural information in the future.

Decoding

Neural signal feature extraction

Overview—Once the recorded signals have gone through data acquisition steps, the decoder first performs the step of neural signal feature extraction (Figure 5). During each operational cycle of the decoder (aka time step), a vector of values is computed from the multivariate time series sampled from the many voltage waveforms streaming in from the sensor. One classical approach depends upon the visual inspection of the action potentials (spikes) emanating from individual neurons, e.g. (23), which are a complex time derivative of the extracellular voltage gradients shaped by filter settings. The typically biphasic, ms-long voltage waveforms of each channel are visually inspected to manually set time and amplitude windows, converting them to discrete events in time, permitting detection of spikes and their assignment to a single neuron or a cluster of them. A unit's spikes are then time binned to compute the spike counts of each unit in each of the time bins, creating a point process time series for each channel.

Threshold crossings—Sometimes, the shape and size of a unit's waveform can change significantly over time. A systematic analysis of instabilities found several cases where the combination of time-amplitude bins traditionally used in spike sorting failed to detect spikes due to drifts in their amplitude which result from even small movements (on the order of a few tens of microns of the electrode recording site with respect to a particular neuron) (J.A. Perge, M.L. Homer, W.Q. Malik, G. Friebs, J.P. Donoghue, and L.R. Hochberg, submitted) (47). A simpler and more robust approach may be to count the number of voltage threshold crossings within a given channel and time bin (48). Termed threshold crossings, the technique enabled no appreciable drop in decoding performance despite significant reductions in spike amplitudes (47). Use of threshold crossings is gaining adoption with their use a key element in recent pilot clinical studies (25), but success may be sensitive to signal to noise ratio, electrode properties, the size and shape of neurons and other unknown factors.

Multisunit activity—An alternative or addition to detecting spikes analyzes signal fluctuations available in the high frequency band, e.g., between 300 – 6,000 Hz. The “multisunit activity” (MUA) can also be used for iBCI control after passing through signal processing such as a nonlinear filter (49). More recently, in offline decoding, non-linear filtering of MUA was reported to yield better results than either spike detection and sorting or threshold crossing (50). 3D reach and grasp in monkeys was decoded offline by extracting power information in the 200–400 Hz frequency bands (which extends into the MUA range) (51). In preliminary closed-loop, head-to-head comparisons by a person with tetraplegia operating an iBCI, MUA performance was found comparable to spike counts (52). Perhaps most attractive is the evidence given of MUA’s non-stationary statistics being less pronounced than other high frequency signal options (50). Again, caution is needed because these advantages may be modified by amplifier or electrode properties.

Local field potentials—Other frequency bands within the range of signals generated by neural activity also contain information about movement intent. Unlike spikes, the primary sources for a given local field potential (LFP) signal measured in the motor cortex remains unclear. Although it has been postulated that LFPs are the result of synaptic currents, many factors such as spike afterhyperpolarizations can play a role (53). Prominent oscillations, one example of which is the commonly referred to “beta rhythm”, have been linked to attention (54) and found to noticeably decrease in amplitude between movement preparation and onset (11). In the last decade, reach direction was decoded from LFPs in monkeys in offline studies (55–57). The features extracted varied, for example, a Gaussian kernel smoother (55) or a multitaper spectral analysis (57). The frequency bands studied may contain different types of information about intended movement; for example, the average amplitude within the 63–200 Hz band increased at movement onset relative to baseline, whereas it was observed to decrease between 16–42 Hz (56). Three dimensional reach and grasp in monkeys can be decoded, offline, from intracortically-recorded low, midrange and higher frequency bands, with the quality approaching, but not surpassing, that of spikes (58). Figure 6 provides an example from a person with tetraplegia with a sensor placed in the arm area of motor cortex. In the presented spectrogram, there is a decrease in amplitude during intended arm movement in the 12–40 Hz range, with both increases and decreases in power seen at other frequency bands throughout the intended movement.

Using two arrays and frequency bands, 0.3–4 Hz and 48–200 Hz, classification of eight target directions was achieved to an accuracy comparable with spikes (59). Other reports decoded natural arm reaches to predict arm endpoint motion and grip aperture (58, 60). Despite these advances, comparisons between decoding with LFPs vs. spikes do not point to LFPs as a clear winner (55, 58). The real value then of the signal is thought to lie in providing more robust features, particularly over the course of time. For example, the ability to detect a spike might be susceptible to array micromotions whereas the aggregate nature of LFPs may mitigate signal changes due to tiny array movements because of their volume conduction properties. This advantage might explain one study’s report of channels that had lost their spikes still retaining most of their predictive power in the LFP features (61).

An alternative use for LFPs lies in identifying discrete states linked to movement. In electrocortical studies, a person’s visual attention (versus eye movement) can be classified above chance levels into one of eight directions (62). In the posterior parietal cortex (PPC), 0–10 and 20–40 Hz bands have revealed the onset of motion as compared to the preparation phase (63). The accomplishment is an especially significant step because the decoding was done online where the monkey used the iBCI to perform the task. Thus, LFPs could play a part in a hierarchical decoding strategy where identification of the discrete motor state dictates how to approach estimating the continuous variables.

Also, the precise features extracted from these broadband neural signals can make a substantive difference in their decoding utility. Instead of LFP amplitude, recent work is starting to reveal the role of LFP phase. Investigations into beta oscillations, for example, have found a statistical dependence between the oscillation's phase and target reach direction (64). These observations are supported by the revealed coherence between beta oscillations and EMG recordings (65). Spiking activity has also been linked to the phase of beta oscillation, where directional information of the former was a function of the latter (66). These new scientific insights suggest that extracting phase information from LFPs might improve iBCI decoding by considering the influence of LFP waveform phase on modulations in spiking activity.

Decoder calibration

Overview—In order to customize the decoder to the user, initial filter calibration exercises engage the iBCI user. In animals, the subject is trained to perform the behavioral task, using overt arm actions. Meanwhile, recordings of the brain activity are collected (16, 18–22). In contrast, people with paralysis are asked to attempt arm and hand motions that would, for example, generate planar ‘mouse’ trajectories and clicks matching the computer controlled cursor on the screen (23, 67). In these calibration sessions, the recorded neural signals and any information processed from them are not presented to the user and thus are called open-loop. Similar approaches can be used for initial training for the control of multi-dimensional reach and grasp movements of a robotic or prosthetic arm (25).

The recorded neural signals and the corresponding movement information are then fed into calibration routines which set parameters within the decoder. For example, when a person with paralysis used an iBCI to control a computer cursor's motion, linear regression processed the calibration data to arrive at the needed matrices for the linear filter (23). In the case of a Bayesian classifier, once the neural signal samples are separated into their respective classes, class conditioned distributions can be fit from the data, e.g., calculating means for independent Poisson distributions (22). The calibration procedures usually extend to processing the features themselves. For instance, from the set of spiking units collected, only those well modulating with reach direction as shown through ANOVA were selected for inclusion into the decoder (21). Such steps help to best fit the decoder's parameters given the calibration data while protecting against over fitting.

Closed-loop calibration—The concept behind closed-loop calibration is to adjust the decoder's calibration after a closed-loop block of trials, one where the iBCI is used to carry out the task. The process has an iterative nature where the previous calibration block adjusts the decoder used in the next calibration block (19). Supporting the approach, differences were found in neuronal activity between actual and observed movement (68, 69). Furthermore, while observation-tuned neurons were found to be predictive of movement direction, strictly action modulating neurons did a better job. Even more recently, such tuning differences, depending upon cognitive context, have also been reported in preliminary studies in clinical trial participants with tetraplegia (70).

Closed-loop decoder calibration has been used successfully to enable control in able-bodied monkeys (24, 71) and in people with tetraplegia (see Figure 7) (25). An innovation is to run an automated software routine to assist epochs of “brain control” (that is, iBCI-controlled movements, as opposed to hand-controlled movements), thereby increasing trial success. In early iterations, considerable assistance is provided, but with each iteration it lessens until the task is completely governed by brain-control. Not only is the technique thought to improve the decoder, but the user's neural activity appears to change to accommodate the iBCI. In several studies, the response characteristics of units changed as the monkey learned

a task (72). When a tuned decoder was intentionally altered while a monkey was operating an iBCI, units' firing rates changed during continued use of the iBCI (32). In another closed-loop iBCI experiment, rats improved their task success rate without changing their overt behavior (73). Thus, use of closed-loop calibration appears to allow both the user and the decoder to update their internal calculations in an effort to converge on stable configurations for useful closed-loop iBCI operation.

Sometimes, long after a stable mapping has been reached, it changes, making the initial decoding model inaccurate. Fortunately, there are mathematical frameworks to formalize possible paths of deviation and ways to fix this problem. On the efficiency side, a recursive least squares formulation can prevent one from having to reconsider all prior calibration blocks (74). Similarly, the problem can be cast in a Bayesian context (75). These innovations get closer to how best to utilize and efficiently process the closed-loop calibration data to adapt the decoder.

Dimensionality reduction & regularization—Several methods seek to remove or discount one or more neural activity features collected during filter calibration in order to make the decoding more accurate. One direct way to go about this is to focus on well-isolated single unit activity (SUA). By comparing changes in waveform characteristics, spike intervals, and directional tuning, one study chose a small subset of 10–15 units from the 75–100 recorded (31). The calibrated decoder performed well over many days (9 days in one monkey, 19 in another).

Others have employed more automated means, such as cross validation, whereby the calibration data is divided into three sets: training, validation, and assessment. First, the decoder training data is used to build models, i.e. the mathematical mapping between features and motor states. These models are then checked with the validation data set to see which performs best and finally the true test of performance occurs with the assessment data. If models under comparison differ by the features chosen, then a greedy method, akin to forward stepwise regression, can find promising, though typically not optimal, subsets (76). This strategy yielded reductions in the number of features chosen, in some cases by a factor of four, in 3D reach and grasp in monkeys (51, 77).

Alternatively, the decoder can treat the neural signal features as being driven by a small number of latent variables, which then represent motor intent – a technique called dimensionality reduction. Gaussian Process Factor Analysis (GPFA), for example, both smooths in time as well as finds a smaller set of latent features (78). Using data from neural ensembles, GPFA was found to improve trajectory reconstruction accuracy over standard methods. Use of dimensionality reduction in real-time decoding reduced the number of variables driving the decoder by approximately 90% in one case and 60% in another (79). Furthermore, error rates in directional classification dropped dramatically from roughly 20 to 5%. Interestingly, such latent features may generate new theories about fundamental, computational principles of the motor cortex. In a reduced space, characteristic, repeatable oscillations emerged, suggesting that cell populations in M1 might not encode movement variables per se, but rather serve as a pattern generator to drive downstream computations and eventually the movement itself (80).

These methods become particularly important if more advanced feature extraction methods are used. With LFP histories, the number of features can grow easily into the thousands. Given such a large data set, many of the features display a redundancy which can be taken advantage of to reject a variety of noise sources such as those of a biological nature and recording artifacts. Dimensionality reduction and regularization is one way to be robust to such disturbances and indeed has been successfully wielded to handle feature deluges of as

much as 6,400 features in order to achieve stable performance in ECoG work over many days (81).

Motor state estimation

Overview—Given a set of features extracted from the neural signals, computations estimate the intended motor state. Decoded motor states have focused largely on those surrounding arm reaching and grasping, particularly as these functions are listed as most important to restore by people with high cervical spinal cord injury (82). Specifically, many earlier studies investigated decoding endpoint velocity and/or position in order to control the movement of a computer cursor on a screen (18–20, 23). For discrete selection, monkeys have chosen a reach target position/direction from a finite list of options (21, 22). One end goal of the work is to realize the possibility of enabling someone with paralysis or loss of a limb to control a robotic aid or FES-driven limb for reaching and grasping. The relatively high level motor commands of selecting a target or controlling the endpoint trajectory could then be converted to control commands needed to drive the effector.

A standard estimation approach is the linear filter (83) and it has been used for continuous decoding, e.g., (9, 18, 19, 23, 24). Each feature value specifies the length of a vector of fixed orientation in the motor state space. The vectors are then summed to yield the state estimate. For discrete states, Bayesian classifiers have also proven useful (21, 22). They typically model the feature distributions conditioned on the discrete states as either Poisson or Gaussian. Both methods have the advantage of being calibrated by exact methods, fast online computation times, and relative transparency in the behavior of the tuned estimator.

More accurate estimation models—In a head-to-head comparison, however, between linear and Kalman filtering, the latter was shown better using several standard performance metrics of cursor control (84). Interestingly, as Figure 8 shows, the Kalman filter did not always yield better cursor trajectories. The algorithm performs the estimation based upon linear, multivariate Gaussian models linking the features to the motor state as well as describing the motor state dynamics. The decoder can run on the same time scale as a linear filter since the results from the linear algebra calculations quickly reach steady state (85). Of especial note is the fact that the evaluation was done by people with tetraplegia operating an iBCI to control the motion of a computer cursor. Since then, the method has provided a reliable estimation approach, even in trials 1000 days after implantation of the sensor (67). A variant of the technique recently enabled two people with severe paralysis to perform 3D reach and grasp tasks by way of a robotic arm aid (25), in one case >5 years after sensor implantation.

Careful evaluation of the true functional mapping showed a large fraction (35%) of the unit spike counts to be nonlinearly related to intended motor control states (86) and for this reason, some earlier attempts with iBCIs worked with nonlinear decoders (14). Another line of research added an auto-regressive element to the feature histories (87, 88). By combining the two types of model refinements, animals were able to operate iBCIs within an unscented Kalman filter framework, yielding improved performance (89). Very general mappings between the features and states can be handled by particle filters (48, 90–92) or neural networks (14, 93).

Discrete motor states can also be introduced into the model. The resulting hybrid state space model switches between a series of discrete modes (94). Computational machinery infers the state transitions during the model fitting process via methods such as expectation maximization (EM). The technique was applied to augment the Kalman filter (95) as well as a point process filter (96). The augmented set of hidden motor states, like the nonlinear models, free the estimator from purely linear assumptions should the calibration data show a

clear nonlinear relationship. In terms of metrics such as mean squared error, use of hybrid state models have achieved reductions as high as 22% (96). However, the results should be treated with caution as the real test lies in closed-loop, head-to-head comparisons (97).

Improving the noise modeling, another line of inquiry makes the more accurate assumption that the spike counts follow a Poisson process (rather than a Gaussian). This is the territory of point process models and they account for the exact timing of the spiking sequences (94, 98). Unlike most linear filter methods which process spike count features within, for example, 30 millisecond bins (24), point processes desire to isolate each unit firing event in time for a typical time bin of 1 millisecond (99). In offline studies, decoders running point process models have been demonstrated in animals (96, 100, 101) as well as in people (99). The cost for the improved modeling fidelity, however, is that decoding becomes more computationally intensive.

Choice of motor control states—Increasing the degrees of freedom decoded from motor cortical activity would enable more tasks to be performed with an iBCI and/or with a greater degree of control. Reports of 3D reach and grasp with a robotic arm in able-bodied monkeys (24) and in people with tetraplegia (25) provide direct evidence of this and relied on expanding the state dimension from two to four (one extra spatial dimension and a variable for grasp). For more precise control, decoders could first determine whether or not the person is attempting a move and in what general direction. Initial creation of such a first stage classifier was reported in monkeys (102). Even finer stages of motor preparation and execution can be detected with iBCIs, such as shown in macaques in which pre-trial, pre-cue, memory (remembering target location during a delay), and post-saccade phases could be readily decoded (103).

Alternatively, the measured neural activity may better associate with a different set of motor states. For example, a unit could be more tuned to wrist extension rather than general upward motion (104). Responding to the need, researchers decoded 25 joint angles under natural reach and grasp conditions in monkeys (77) under open-loop conditions. The strategy relied on statistical dependencies between the joint angles, allowing the kinematics states to be compressed down to 10. As the 10 variables were estimated at each time step, a linear transform expanded the estimate to a vector representing all 25 joint angles. Thus, even with a relatively small sample of neurons in motor cortex, iBCIs based on motor cortical recordings should enable complex reach and grasp movement. Forces are also encoded in the stream of neural activity measurements (8, 105). In planar manipulandum experiments, software simultaneously decoded force and velocity in offline analyses (106). Estimating intended force (grip force, for example) in people with paralysis will be an important venture, and is likely to benefit from feedback from the effector.

Decoding muscle contractions had led to temporarily paralyzed monkeys regaining control of their muscles through functional electric stimulation (FES) driven by an iBCI. In 2008, monkeys were reported to control wrist torques in closed-loop experiments (26). More recently, five electrodes stimulated three muscles that drove wrist flexion and grip (28). Effectively building an artificial bridge from the central nervous system to the muscles, the animals were able to pick up balls and place them into a receptacle. For people, control of a virtual arm was shown possible by a person with tetraplegia, using a simulation that incorporates the complex, dynamic response behavior of an FES system (29). These advancements mark key milestones towards restoring limb control to those with paralysis.

While the majority of studies in non-human primates have focused on reaching and grasping, research has also been proceeding in other types of motor control. Going beyond hand grip, individual finger movements (i.e., which finger moved) has been correctly

classified better than 80% of the time (107, 108) in offline studies. Leg kinematic offline decoding studies in monkeys have also been performed (109), although full closed-loop locomotion presents interlimb coordination and balance challenges. Rats, with electrodes in the hind limb/trunk region of M1, successfully lightened a load applied to their hips, in time with their gait (110). While neither study demonstrated closed-loop control over the entire combination of movements needed to enable walking, they do suggest a route to the more modest goals of controlling walking direction and speed as well as making any major load adjustments.

Adaptive estimation—The relationship between the intended movement of a person with tetraplegia and the relatively tiny ensemble of neural signals used to deduce that intended movement is by no means expected to remain stationary. Although this relationship can be stable over time (31), both the technical factors (e.g., micromotion of the electrode with respect to the tissue, changing characteristics of recording devices) and biological ones (e.g., cortical plasticity) discussed earlier often change the mapping (111–113). Thus, adaptive decoders have been employed to refine continually the relationship between the recorded neural signals and the expected output. For example, using Bayesian regression methods, decoding performance was maintained in the face of model drift on the time scale of days (114).

Ultimately, the cause of the nonstationarities may arise from a range of cognitive or physiological state variables. For example, spikes rates may modulate by levels of attention. The amplitude of beta oscillations reduces upon initiation of movement (11) and the phasing of the rhythm has been shown to be predictive of spike activity (66). Thus, directly accounting for movement onset in the estimator may enable it to account for some of the nonstationarities. Similarly, incorporating eye position into the estimation process can raise iBCI performance (115). In this view, the observed nonstationarities more reflect the incomplete nature of existing models underlying estimators than changes to the random noise corrupting the neural signal.

Conclusion

Research progress in implants and decoders for iBCIs is advancing rapidly. 2006 saw the reporting of people with paralysis controlling a computer cursor and simple robotic devices. Only six years later, complex 3D reach and grasp movements of assistive robotic devices was achieved (see Figure 9) (25). Wireless, transcutaneous sensors have reached the milestone of initial successful animal testing (46). Furthermore, as has been the case in the neural prosthesis field for 50+ years, publicly funded preclinical research continues to provide vital insights and new directions for clinical investigation focused on developing restorative neurotechnologies for people with paralysis.

At the same time, the number of decodable states is expanding (77) and other areas besides the arm are proving to be amenable to brain-control (110). The push for additional motor states is being supported by an expansion of the list of features extracted such as those seen with LFPs (61, 81). The developments promise to increase the capability of robotic aids or prostheses.

Just as important as the degrees of freedom is the quality of control. With the introduction of algorithms such as dimensionality reduction (79) and improved implant circuitry, the neural signal features are becoming less contaminated by noise. Noise free signals however will not cure errors in the model, but fortunately model sophistication is increasing as well (84, 99) and so are ways to avoid poor calibration (24, 25). These efforts fuel the hope of seeing finer control tasks performed in the coming years.

Finally, for the technology to be clinically viable, it would ideally be reliable for very long time periods, with a span of more than a decade being a plausible target. Preliminary reports, with one person successfully using an investigational iBCI (BrainGate) after 5 years, suggest great potential, however additional advances in sensor manufacturing will be essential (67). For example, the current work on a wireless, transcutaneous implant eliminates some connectors and tethering that can wear. When sensor characteristics change with time or the physiological mapping between intended movement and neural activity evolves, adaptive features built into the decoder can mitigate such deviations (114). These innovations all underscore the expectation that iBCI users will soon see implants functioning well for 10 years, and frequent cases where iBCIs are used for days at a time without any recalibration.

The iBCI field is tasked simultaneously with creating better devices for fundamental neuroscience research and with creating viable clinical neurotechnologies. This latter goal requires continued advances in neuroengineering, as well as rigorous clinical evaluation. With iBCI's becoming more capable, precise, and reliable, future research will need to then bridge the gap between demonstrating a capability in a research trial and incorporating the technology into the daily lives of its users. For people who are unable to move or speak due to brainstem stroke or motor neuron disease, the impact of even modest increases in command signal output will be immense. While this and other goals for iBCIs will likely require years of additional research to become reality, recent strides make clear that such technologies are moving steadily from the realm of science fiction toward the domain of devices that support the communication, mobility, and independence of people with neurologic disease or injury.

Acknowledgments

This work is supported in part by the Rehabilitation Research and Development Service, Office of Research and Development, Department of Veterans Affairs. Additional support is provided by NIH: NIDCD (R01DC009899) and NIBIB (R01EB007401-01); DARPA REPAIR (N66001-10-C-2010); Doris Duke Charitable Foundation; MGH-Deane Institute of Integrated Research on Atrial Fibrillation and Stroke; Katie Samson Foundation.

The contents do not represent the views of the Department of Veterans Affairs or the United States Government.

References

1. Ziegler-Graham K, MacKenzie EJ, Ephraim PL, Travison TG, Brookmeyer R. Estimating the prevalence of limb loss in the United States: 2005 to 2050. *Arch Phys Med Rehabil.* 2008; 89:422–429. [PubMed: 18295618]
2. Foundation CDR. One Degree of Separation: Paralysis and Spinal Cord Injury in the United State. Short Hills, NJ: 2009.
3. Jackson AB, Dijkers M, Devivo MJ, Poczatek RB. A demographic profile of new traumatic spinal cord injuries: change and stability over 30 years. *Arch Phys Med Rehabil.* 2004; 85:1740–1748. [PubMed: 15520968]
4. Mak JN, Wolpaw JR. Clinical Applications of Brain-Computer Interfaces: Current State and Future Prospects. *IEEE Rev Biomed Eng.* 2009; 2:187–199. [PubMed: 20442804]
5. Schalk G, Leuthardt EC. Brain-computer interfaces using electrocorticographic signals. *IEEE Rev Biomed Eng.* 2011; 4:140–154. [PubMed: 22273796]
6. Evarts EV. Relation of pyramidal tract activity to force exerted during voluntary movement. *J Neurophysiol.* 1968; 31:14–27. [PubMed: 4966614]
7. Fetz EE. Operant conditioning of cortical unit activity. *Science.* 1969; 163:955–958. [PubMed: 4974291]
8. Humphrey DR, Schmidt EM, Thompson WD. Predicting measures of motor performance from multiple cortical spike trains. *Science.* 1970; 170:758–762. [PubMed: 4991377]

9. Georgopoulos AP, Schwartz AB, Kettner RE. Neuronal population coding of movement direction. *Science*. 1986; 233:1416–1419. [PubMed: 3749885]
10. Kalaska JF, Crammond DJ. Cerebral cortical mechanisms of reaching movements. *Science*. 1992; 255:1517–1523. [PubMed: 1549781]
11. Sanes JN, Donoghue JP. Oscillations in local field potentials of the primate motor cortex during voluntary movement. *Proc Natl Acad Sci U S A*. 1993; 90:4470–4474. [PubMed: 8506287]
12. BeMent SL, Wise KD, Anderson DJ, Najafi K, Drake KL. Solid-state electrodes for multichannel multiplexed intracortical neuronal recording. *IEEE Trans Biomed Eng*. 1986; 33:230–241. [PubMed: 3957372]
13. Maynard EM, Nordhausen CT, Normann RA. The Utah intracortical Electrode Array: a recording structure for potential brain-computer interfaces. *Electroencephalogr Clin Neurophysiol*. 1997; 102:228–239. [PubMed: 9129578]
14. Burrow, MDJ.; Humphrey, D. Cortical control of a robot using a time-delay neural network; Presented at International Conference on Rehabilitation Robotics; 1997.
15. Humphrey DRD, Hochberg LR, Burrow M, Dugger J. Cortical Control of Neural Prosthetic Devices. Rep. NIH/NINDS Contract N01-NS-1-2308. 1997
16. Chapin JK, Moxon KA, Markowitz RS, Nicolelis MA. Real-time control of a robot arm using simultaneously recorded neurons in the motor cortex. *Nat Neurosci*. 1999; 2:664–670. [PubMed: 10404201]
17. Kennedy PR, Bakay RA. Restoration of neural output from a paralyzed patient by a direct brain connection. *Neuroreport*. 1998; 9:1707–1711. [PubMed: 9665587]
18. Serruya MD, Hatsopoulos NG, Paninski L, Fellows MR, Donoghue JP. Instant neural control of a movement signal. *Nature*. 2002; 416:141–142. [PubMed: 11894084]
19. Taylor DM, Tillery SI, Schwartz AB. Direct cortical control of 3D neuroprosthetic devices. *Science*. 2002; 296:1829–1832. [PubMed: 12052948]
20. Carmena JM, Lebedev MA, Crist RE, O'Doherty JE, Santucci DM, et al. Learning to control a brain-machine interface for reaching and grasping by primates. *PLoS Biol*. 2003; 1:E42. [PubMed: 14624244]
21. Musallam S, Corneil BD, Greger B, Scherberger H, Andersen RA. Cognitive control signals for neural prosthetics. *Science*. 2004; 305:258–262. [PubMed: 15247483]
22. Santhanam G, Ryu SI, Yu BM, Afshar A, Shenoy KV. A high-performance brain-computer interface. *Nature*. 2006; 442:195–198. [PubMed: 16838020]
23. Hochberg LR, Serruya MD, Friehs GM, Mukand JA, Saleh M, et al. Neuronal ensemble control of prosthetic devices by a human with tetraplegia. *Nature*. 2006; 442:164–171. [PubMed: 16838014]
24. Velliste M, Perel S, Spalding MC, Whitford AS, Schwartz AB. Cortical control of a prosthetic arm for self-feeding. *Nature*. 2008; 453:1098–1101. [PubMed: 18509337]
25. Hochberg LR, Bacher D, Jarosiewicz B, Masse NY, Simeral JD, et al. Reach and grasp by people with tetraplegia using a neurally controlled robotic arm. *Nature*. 2012; 485:372–375. [PubMed: 22596161]
26. Moritz CT, Perlmutter SI, Fetz EE. Direct control of paralysed muscles by cortical neurons. *Nature*. 2008; 456:639–642. [PubMed: 18923392]
27. Pohlmeier EA, Oby ER, Perreault EJ, Solla SA, Kilgore KL, et al. Toward the restoration of hand use to a paralyzed monkey: brain-controlled functional electrical stimulation of forearm muscles. *PLoS One*. 2009; 4:e5924. [PubMed: 19526055]
28. Ethier C, Oby ER, Bauman MJ, Miller LE. Restoration of grasp following paralysis through brain-controlled stimulation of muscles. *Nature*. 2012; 485:368–371. [PubMed: 22522928]
29. Chadwick EK, Blana D, Simeral JD, Lambrecht J, Kim SP, et al. Continuous neuronal ensemble control of simulated arm reaching by a human with tetraplegia. *J Neural Eng*. 2011; 8:034003. [PubMed: 21543840]
30. Linderman MD, Santhanam G, Kemere CT, Gilja V, O'Driscoll S, et al. Signal processing challenges for neural prostheses. *Ieee Signal Processing Magazine*. 2008; 25:18–28.

31. Ganguly K, Secundo L, Ranade G, Orsborn A, Chang EF, et al. Cortical representation of ipsilateral arm movements in monkey and man. *J Neurosci*. 2009; 29:12948–12956. [PubMed: 19828809]
32. Jarosiewicz B, Chase SM, Fraser GW, Velliste M, Kass RE, Schwartz AB. Functional network reorganization during learning in a brain-computer interface paradigm. *Proc Natl Acad Sci U S A*. 2008; 105:19486–19491. [PubMed: 19047633]
33. Wolpaw, JR.; Wolpaw, EW. *Brain-computer interfaces : principles and practice*. Oxford; New York: Oxford University Press; 2012. p. xviii400 p. pp.
34. Robinson DA. Electrical Properties of Metal Microelectrodes. *Proceedings of the Institute of Electrical and Electronics Engineers*. 1968; 56 1065-&.
35. Jones KE, Campbell PK, Normann RA. A Glass Silicon Composite Intracortical Electrode Array. *Annals of Biomedical Engineering*. 1992; 20:423–437. [PubMed: 1510294]
36. Bai Q, Wise KD. Single-unit neural recording with active microelectrode arrays. *Ieee Transactions on Biomedical Engineering*. 2001; 48:911–920. [PubMed: 11499528]
37. Kipke DR, Vetter RJ, Williams JC, Hetke JF. Silicon-substrate intracortical microelectrode arrays for long-term recording of neuronal spike activity in cerebral cortex. *Ieee Transactions on Neural Systems and Rehabilitation Engineering*. 2003; 11:151–155. [PubMed: 12899260]
38. Bhandari R, Negi S, Rieth L, Normann RA, Solzbacher F. A novel masking method for high aspect ratio penetrating microelectrode arrays. *Journal of Micromechanics and Microengineering*. 2009;19.
39. Seymour JP, Langhals NB, Anderson DJ, Kipke DR. Novel multi-sided, microelectrode arrays for implantable neural applications. *Biomed Microdevices*. 2011; 13:441–451. [PubMed: 21301965]
40. Abidian MR, Ludwig KA, Marzullo TC, Martin DC, Kipke DR. Interfacing Conducting Polymer Nanotubes with the Central Nervous System: Chronic Neural Recording using Poly (3,4-ethylenedioxythiophene) Nanotubes. *Advanced Materials*. 2009; 21:3764–3770.
41. Ludwig KA, Langhals NB, Joseph MD, Richardson-Burns SM, Hendricks JL, Kipke DR. Poly(3,4-ethylenedioxythiophene) (PEDOT) polymer coatings facilitate smaller neural recording electrodes. *J Neural Eng*. 2011;8.
42. Santhanam G, Linderman MD, Gilja V, Afshar A, Ryu SI, et al. HermesB: a continuous neural recording system for freely behaving primates. *IEEE Trans Biomed Eng*. 2007; 54:2037–2050. [PubMed: 18018699]
43. Chestek CA, Gilja V, Nuyujukian P, Kier RJ, Solzbacher F, et al. HermesC: low-power wireless neural recording system for freely moving primates. *IEEE Trans Neural Syst Rehabil Eng*. 2009; 17:330–338. [PubMed: 19497829]
44. Harrison RR, Kier RJ, Chestek CA, Gilja V, Nuyujukian P, et al. Wireless neural recording with single low-power integrated circuit. *IEEE Trans Neural Syst Rehabil Eng*. 2009; 17:322–329. [PubMed: 19497825]
45. Miranda, H.; Gilja, V.; Chestek, C.; Shenoy, KV.; Meng, TH. A high-rate long-range wireless transmission system for multichannel neural recording applications; 2009. Presented at Circuits and Systems, 2009. ISCAS, 2009. IEEE International Symposium on
46. Yin, M.; Borton, DA.; Aceros, J.; Patterson, WR.; Nurmikko, AV. A 100-channel hermetically sealed implantable device for wireless neurosensing applications; 2012. Presented at Circuits and Systems (ISCAS), 2012 IEEE International Symposium on
47. Chestek CA, Gilja V, Nuyujukian P, Foster JD, Fan JM, et al. Long-term stability of neural prosthetic control signals from silicon cortical arrays in rhesus macaque motor cortex. *J Neural Eng*. 2011; 8:045005. [PubMed: 21775782]
48. Fraser GW, Chase SM, Whitford A, Schwartz AB. Control of a brain-computer interface without spike sorting. *J Neural Eng*. 2009; 6:055004. [PubMed: 19721186]
49. Humphrey DR. Systems, methods, and devices for controlling external devices by signals derived directly from the nervous system. Google Patents. 2001
50. Stark E, Abeles M. Predicting movement from multiunit activity. *J Neurosci*. 2007; 27:8387–8394. [PubMed: 17670985]

51. Zhuang J, Truccolo W, Vargas-Irwin C, Donoghue JP. Decoding 3-D reach and grasp kinematics from high-frequency local field potentials in primate primary motor cortex. *IEEE Trans Biomed Eng.* 2010; 57:1774–1784. [PubMed: 20403782]
52. Malik, WQ.; Stavisky, SD.; Bacher, D.; Simeral, JD.; Truccolo, W., et al. Society for Neuroscience Meeting Planner. San Diego, CA: 2010. Decoding multiunit activity in neural interfaces for individuals with tetraplegia.
53. Buzsaki G, Anastassiou CA, Koch C. The origin of extracellular fields and currents--EEG, ECoG, LFP and spikes. *Nat Rev Neurosci.* 2012; 13:407–420. [PubMed: 22595786]
54. Bouyer JJ, Montaron MF, Rougeul A. Fast fronto-parietal rhythms during combined focused attentive behaviour and immobility in cat: cortical and thalamic localizations. *Electroencephalogr Clin Neurophysiol.* 1981; 51:244–252. [PubMed: 6163613]
55. Mehring C, Rickert J, Vaadia E, Cardosa de Oliveira S, Aertsen A, Rotter S. Inference of hand movements from local field potentials in monkey motor cortex. *Nat Neurosci.* 2003; 6:1253–1254. [PubMed: 14634657]
56. Rickert J, Oliveira SC, Vaadia E, Aertsen A, Rotter S, Mehring C. Encoding of movement direction in different frequency ranges of motor cortical local field potentials. *J Neurosci.* 2005; 25:8815–8824. [PubMed: 16192371]
57. Scherberger H, Jarvis MR, Andersen RA. Cortical local field potential encodes movement intentions in the posterior parietal cortex. *Neuron.* 2005; 46:347–354. [PubMed: 15848811]
58. Bansal AK, Truccolo W, Vargas-Irwin CE, Donoghue JP. Decoding 3D reach and grasp from hybrid signals in motor and premotor cortices: spikes, multiunit activity, and local field potentials. *J Neurophysiol.* 2012; 107:1337–1355. [PubMed: 22157115]
59. Ince NF, Gupta R, Arica S, Tewfik AH, Ashe J, Pellizzer G. High accuracy decoding of movement target direction in non-human primates based on common spatial patterns of local field potentials. *PLoS One.* 2010; 5:e14384. [PubMed: 21200434]
60. Bansal AK, Vargas-Irwin CE, Truccolo W, Donoghue JP. Relationships among low-frequency local field potentials, spiking activity, and three-dimensional reach and grasp kinematics in primary motor and ventral premotor cortices. *J Neurophysiol.* 2011; 105:1603–1619. [PubMed: 21273313]
61. Flint RD, Lindberg EW, Jordan LR, Miller LE, Slutzky MW. Accurate decoding of reaching movements from field potentials in the absence of spikes. *J Neural Eng.* 2012; 9:046006. [PubMed: 22733013]
62. Gunduz A, Brunner P, Daitch A, Leuthardt EC, Ritaccio AL, et al. Decoding covert spatial attention using electrocorticographic (ECoG) signals in humans. *Neuroimage.* 2012; 60:2285–2293. [PubMed: 22366333]
63. Hwang EJ, Andersen RA. Brain control of movement execution onset using local field potentials in posterior parietal cortex. *J Neurosci.* 2009; 29:14363–14370. [PubMed: 19906983]
64. Rubino D, Robbins KA, Hatsopoulos NG. Propagating waves mediate information transfer in the motor cortex. *Nat Neurosci.* 2006; 9:1549–1557. [PubMed: 17115042]
65. Baker SN. Oscillatory interactions between sensorimotor cortex and the periphery. *Curr Opin Neurobiol.* 2007; 17:649–655. [PubMed: 18339546]
66. Reimer J, Hatsopoulos NG. Periodicity and evoked responses in motor cortex. *J Neurosci.* 2010; 30:11506–11515. [PubMed: 20739573]
67. Simeral JD, Kim SP, Black MJ, Donoghue JP, Hochberg LR. Neural control of cursor trajectory and click by a human with tetraplegia 1000 days after implant of an intracortical microelectrode array. *J Neural Eng.* 2011; 8:025027. [PubMed: 21436513]
68. Tkach D, Reimer J, Hatsopoulos NG. Observation-based learning for brain-machine interfaces. *Curr Opin Neurobiol.* 2008; 18:589–594. [PubMed: 18838120]
69. Dushanova J, Donoghue J. Neurons in primary motor cortex engaged during action observation. *Eur J Neurosci.* 2010; 31:386–398. [PubMed: 20074212]
70. Feldman, JM.; King, B.; Truccolo, W.; Hochberg, LR.; Donoghue, JD. Society for Neuroscience Meeting Planner. San Diego, CA: 2011. Decoding neural representations of action from motor cortex ensembles during action observation in humans with tetraplegia.

71. Orsborn AL, Dangi S, Moorman HG, Carmena JM. Closed-loop decoder adaptation on intermediate time-scales facilitates rapid BMI performance improvements independent of decoder initialization conditions. *IEEE Trans Neural Syst Rehabil Eng.* 2012; 20:468–477. [PubMed: 22772374]
72. Green AM, Kalaska JF. Learning to move machines with the mind. *Trends Neurosci.* 2011; 34:61–75. [PubMed: 21176975]
73. Koralek AC, Jin X, Long JD 2nd, Costa RM, Carmena JM. Corticostriatal plasticity is necessary for learning intentional neuroprosthetic skills. *Nature.* 2012; 483:331–335. [PubMed: 22388818]
74. Wu W, Hatsopoulos NG. Real-time decoding of nonstationary neural activity in motor cortex. *IEEE Trans Neural Syst Rehabil Eng.* 2008; 16:213–222. [PubMed: 18586600]
75. Li Z, O'Doherty JE, Lebedev MA, Nicolelis MA. Adaptive decoding for brain-machine interfaces through Bayesian parameter updates. *Neural Comput.* 2011; 23:3162–3204. [PubMed: 21919788]
76. Hastie, T.; Tibshirani, R.; Friedman, JH. *Springer series in statistics.* New York: Springer; 2009. The elements of statistical learning data mining, inference, and prediction; p. 1online resource (xxii, 745 p.)
77. Vargas-Irwin CE, Shakhnarovich G, Yadollahpour P, Mislow JM, Black MJ, Donoghue JP. Decoding complete reach and grasp actions from local primary motor cortex populations. *J Neurosci.* 2010; 30:9659–9669. [PubMed: 20660249]
78. Yu BM, Cunningham JP, Santhanam G, Ryu SI, Shenoy KV, Sahani M. Gaussian-process factor analysis for low-dimensional single-trial analysis of neural population activity. *J Neurophysiol.* 2009; 102:614–635. [PubMed: 19357332]
79. Santhanam G, Yu BM, Gilja V, Ryu SI, Afshar A, et al. Factor-analysis methods for higher-performance neural prostheses. *J Neurophysiol.* 2009; 102:1315–1330. [PubMed: 19297518]
80. Churchland MM, Cunningham JP, Kaufman MT, Foster JD, Nuyujukian P, et al. Neural population dynamics during reaching. *Nature.* 2012; 487:51–56. [PubMed: 22722855]
81. Chao ZC, Nagasaka Y, Fujii N. Long-term asynchronous decoding of arm motion using electrocorticographic signals in monkeys. *Front Neuroeng.* 2010; 3:3. [PubMed: 20407639]
82. Anderson KD. Targeting recovery: priorities of the spinal cord-injured population. *J Neurotrauma.* 2004; 21:1371–1383. [PubMed: 15672628]
83. Oppenheim, AV.; Schaffer, RW. *Discrete-time signal processing.* Pearson: Upper Saddle River; 2010. p. xxviii+1108 p. pp.
84. Kim SP, Simeral JD, Hochberg LR, Donoghue JP, Black MJ. Neural control of computer cursor velocity by decoding motor cortical spiking activity in humans with tetraplegia. *J Neural Eng.* 2008; 5:455–476. [PubMed: 19015583]
85. Malik WQ, Truccolo W, Brown EN, Hochberg LR. Efficient decoding with steady-state Kalman filter in neural interface systems. *IEEE Trans Neural Syst Rehabil Eng.* 2011; 19:25–34. [PubMed: 21078582]
86. Paninski L, Shoham S, Fellows MR, Hatsopoulos NG, Donoghue JP. Superlinear population encoding of dynamic hand trajectory in primary motor cortex. *J Neurosci.* 2004; 24:8551–8561. [PubMed: 15456829]
87. Fisher J, Black M. Motor cortical decoding using an autoregressive moving average model. *Conf Proc IEEE Eng Med Biol Soc.* 2005; 2:2130–2133. [PubMed: 17282650]
88. Shpigelman L, Lalazar H, Vaadia E. Kernel-ARMA for hand tracking and brain-machine interfacing during 3D motor control. *Advances in neural information processing systems.* 2009; 21:1489–1496.
89. Li Z, O'Doherty JE, Hanson TL, Lebedev MA, Henriquez CS, Nicolelis MA. Unscented Kalman filter for brain-machine interfaces. *PLoS One.* 2009; 4:e6243. [PubMed: 19603074]
90. Gao Y, Black M, Bienenstock E, Shoham S, Donoghue J. Probabilistic inference of hand motion from neural activity in motor cortex. *Advances in neural information processing systems.* 2002; 1:213–220.
91. Brockwell AE, Rojas AL, Kass RE. Recursive bayesian decoding of motor cortical signals by particle filtering. *J Neurophysiol.* 2004; 91:1899–1907. [PubMed: 15010499]
92. Wood F, Prabhat, Donoghue J, Black M. Inferring attentional state and kinematics from motor cortical firing rates. *Conf Proc IEEE Eng Med Biol Soc.* 2005; 1:149–152. [PubMed: 17282133]

93. Sussillo D, Nuyujukian P, Fan JM, Kao JC, Stavisky SD, et al. A recurrent neural network for closed-loop intracortical brain-machine interface decoders. *J Neural Eng.* 2012; 9:026027. [PubMed: 22427488]
94. Srinivasan L, Eden UT, Mitter SK, Brown EN. General-purpose filter design for neural prosthetic devices. *J Neurophysiol.* 2007; 98:2456–2475. [PubMed: 17522167]
95. Wu, W.; Black, MJ.; Mumford, D.; Gao, Y.; Bienenstock, E.; Donoghue, JP. A switching Kalman filter model for the motor cortical coding of hand motion; Proceedings of the 25th Annual International Conference of the IEEE; 2003. Presented at Engineering in Medicine and Biology Society, 2003
96. Lawhern V, Wu W, Hatsopoulos N, Paninski L. Population decoding of motor cortical activity using a generalized linear model with hidden states. *J Neurosci Methods.* 2010; 189:267–280. [PubMed: 20359500]
97. Chase SM, Schwartz AB, Kass RE. Bias, optimal linear estimation, and the differences between open-loop simulation and closed-loop performance of spiking-based brain-computer interface algorithms. *Neural Netw.* 2009; 22:1203–1213. [PubMed: 19502004]
98. Brown EN, Frank LM, Tang D, Quirk MC, Wilson MA. A statistical paradigm for neural spike train decoding applied to position prediction from ensemble firing patterns of rat hippocampal place cells. *J Neurosci.* 1998; 18:7411–7425. [PubMed: 9736661]
99. Truccolo W, Friehs GM, Donoghue JP, Hochberg LR. Primary motor cortex tuning to intended movement kinematics in humans with tetraplegia. *J Neurosci.* 2008; 28:1163–1178. [PubMed: 18234894]
100. Truccolo W, Eden UT, Fellows MR, Donoghue JP, Brown EN. A point process framework for relating neural spiking activity to spiking history, neural ensemble, and extrinsic covariate effects. *J Neurophysiol.* 2005; 93:1074–1089. [PubMed: 15356183]
101. Nazarpour K, Ethier C, Paninski L, Rebesco JM, Miall RC, Miller LE. EMG prediction from motor cortical recordings via a nonnegative point-process filter. *IEEE Trans Biomed Eng.* 2012; 59:1829–1838. [PubMed: 21659018]
102. Kemere C, Santhanam G, Yu BM, Afshar A, Ryu SI, et al. Detecting neural-state transitions using hidden Markov models for motor cortical prostheses. *J Neurophysiol.* 2008; 100:2441–2452. [PubMed: 18614757]
103. Campos M, Breznen B, Andersen RA. A neural representation of sequential states within an instructed task. *J Neurophysiol.* 2010; 104:2831–2849. [PubMed: 20739594]
104. Kakei S, Hoffman DS, Strick PL. Sensorimotor transformations in cortical motor areas. *Neurosci Res.* 2003; 46:1–10. [PubMed: 12725907]
105. Fagg AH, Ojakangas GW, Miller LE, Hatsopoulos NG. Kinetic trajectory decoding using motor cortical ensembles. *IEEE Trans Neural Syst Rehabil Eng.* 2009; 17:487–496. [PubMed: 19666343]
106. Gupta R, Ashe J. Offline decoding of end-point forces using neural ensembles: application to a brain-machine interface. *IEEE Trans Neural Syst Rehabil Eng.* 2009; 17:254–262. [PubMed: 19497832]
107. Acharya S, Tenore F, Aggarwal V, Etienne-Cummings R, Schieber MH, Thakor NV. Decoding individuated finger movements using volume-constrained neuronal ensembles in the M1 hand area. *IEEE Trans Neural Syst Rehabil Eng.* 2008; 16:15–23. [PubMed: 18303801]
108. Baker J, Bishop W, Kellis S, Levy T, House P, Greger B. Multi-scale recordings for neuroprosthetic control of finger movements. *Conf Proc IEEE Eng Med Biol Soc.* 2009; 2009:4573–4577. [PubMed: 19963841]
109. Fitzsimmons NA, Lebedev MA, Peikon ID, Nicolelis MA. Extracting kinematic parameters for monkey bipedal walking from cortical neuronal ensemble activity. *Front Integr Neurosci.* 2009; 3:3. [PubMed: 19404411]
110. Song W, Giszter SF. Adaptation to a cortex-controlled robot attached at the pelvis and engaged during locomotion in rats. *J Neurosci.* 2011; 31:3110–3128. [PubMed: 21414932]
111. Kim, SP.; Wood, F.; Fellows, M.; Donoghue, JP.; Black, MJ. Statistical Analysis of the Non-stationarity of Neural Population Codes; 2006. Presented at Biomedical Robotics and Biomechanics, 2006. BioRob 2006 The First IEEE/RAS-EMBS International Conference on

112. Perge, JA.; Donoghue, JP.; Hochberg, LR. Society for Neuroscience Meeting Planner. San Diego, CA: 2010. Firing rate nonstationarity can contribute to performance variations in neural cursor control: A BrainGate2 study.
113. Gilja V, Chestek C, Diester I, Henderson J, Deisseroth K, Shenoy K. Challenges and Opportunities for Next-Generation Intra-Cortically Based Neural Prostheses. IEEE Trans Biomed Eng. 2011
114. Li, Z.; O'Doherty, E.; Lebedev, MA.; Nicolelis, MA. Closed-loop adaptive decoding using Bayesian regression self-tuning. San Diego, CA: Presented at Society for Neuroscience Meeting Planner; 2010.
115. Batista AP, Yu BM, Santhanam G, Ryu SI, Afshar A, Shenoy KV. Cortical neural prosthesis performance improves when eye position is monitored. IEEE Trans Neural Syst Rehabil Eng. 2008; 16:24–31. [PubMed: 18303802]

Figure 1.
Illustration of an iBCI setup.

Figure 2. Block diagram of iBCI architecture. Arrows indicate the major flows of information whereas the red-shaded box contains elements associated with decoding.

Figure 3.
A silicon-based cortical MEA (inset); implanted for intracortical neural recording via a percutaneous connection to a skull mounted pedestal connector.

Figure 4.

A wireless neural interface, hermetically sealed, for untethered neuroscience and potential clinical use. (a) An image of the device after hermetic closure and ready for implant (device placed in special holder for sterilization and transport to the operating room). The standard 10×10 element microelectrode array (Blackrock Microsystems, UT) can be seen exiting the enclosure in a 100-wire bundle. (b) In a view through the single crystal sapphire window, used for electromagnetic transparency, the receiving power coil (red) and the wideband data telemetry chip antenna (blue) can be seen. The sapphire is braised to the titanium, maintaining the hermetic seal. (c) The intracortical neural sensor, manufactured at Blackrock Microsystems, and a reference wire can be seen in relation to the supracranial/subcutaneous enclosure. (d) All 100 wires individually enter the enclosure through gold and ceramic feedthroughs, after fanout onto a flexible polymer (Kapton) interconnect board sealed to the (outside) bottom of the titanium enclosure. Here, we show the 100 wire bonding site (grid, right) and the interconnect circuit board soldered to the feedthrough assembly. The

interconnect board is packaged in biocompatible silicone rubber to maintain robust electrical isolation of the input electrodes.

Figure 5. Illustration of a decoder's architecture and the flow of information passing through the various components.

Figure 6.

Representative spectrogram of the LFP signal in one channel, averaged over 70 repeated trials. Time 0 denotes the instructed onset of movement (dashed line). Colors indicate amplitude, normalized by dividing by average baseline activity prior to movement onset (blue greatest reduction in activity, red greatest increase in activity). The data was recorded from a person with tetraplegia intending to move her own hand during repeated cursor control calibration trials. Note the appearance of three differing bands (approximately 0–3, 3–40, and 40–100 Hz). (courtesy of J. A. Perge, unpublished)

Figure 7.

Illustration of closed-loop calibration technique. (a) White shows cursor trajectory; the blue and red dots show the starting point and goal respectively. At each sample time, the direction between the cursor and goal is calculated (red arrows). This direction is then compared to the current cursor heading in order to generate an error signal with which to adjust the decoding tuning parameters. (b) A similar technique can be applied in the 3D case (where the white coloring is switched to black for contrast). ((a) courtesy of B. Jarosiewicz, unpublished) ((b) courtesy of (25), adapted)

Figure 8.

A comparison of the paths under brain-control using the Kalman filter vs. linear filter. The paths were performed by a person with tetraplegia using velocity-based decoders during the eight-target center-out task. Performance results from two recording sessions are shown where both the linear filter and the Kalman filter were sequentially tested. Eight targets were located at (0, 45, 90, 135, 180, 225, 270 and 315 degrees). Units indicate the number of features employed by the respective decoding filter. (courtesy of (84))

Figure 9. Successful iBCI-based neural control of a robotic arm to reach and grasp a thermos of coffee by a person with tetraplegia. The MEA recording the neural data had been implanted for more than five years at the time of this achievement. (courtesy of (25))