Published in final edited form as:

Magn Reson Med. 2014 April ; 71(4): 1349–1357. doi:10.1002/mrm.25168.

A Subspace Approach to High-Resolution Spectroscopic Imaging

Fan Lam and Zhi-Pei Liang

Department of Electrical and Computer Engineering, and Beckman Institute for Advanced Science and Technology University of Illinois at Urbana-Champaign

Abstract

Purpose—To accelerate spectroscopic imaging using sparse sampling of (k, t) -space and subspace (or low-rank) modeling to enable high-resolution metabolic imaging with good signal-to-noise ratio (SNR).

Methods—The proposed method, called SPICE (SPectroscopic Imaging by exploiting spatioSpectral CorRElation), exploits a unique property known as partial separability of spectroscopic signals. This property indicates that high-dimensional spectroscopic signals reside in a very low-dimensional subspace and enables special data acquisition and image reconstruction strategies to be used to obtain high-resolution spatioSpectral distributions with good SNR. More specifically, a hybrid CSI/EPSI pulse sequence is proposed for sparse sampling of (k, t) -space, and a low-rank model-based algorithm is proposed for subspace estimation and image reconstruction from sparse data with the capability to incorporate prior information and field inhomogeneity correction.

Results—The performance of SPICE has been evaluated using both computer simulations and phantom studies, which produced very encouraging results. For 2D spectroscopic imaging experiments on a metabolite phantom, a factor of 10 acceleration was achieved with a minimal loss in SNR compared to the long CSI experiments and with a significant gain in SNR compared to the accelerated EPSI experiments.

Conclusion—SPICE is able to significantly accelerate spectroscopic imaging experiments, making high-resolution metabolic imaging possible.

Keywords

Spectroscopic imaging; chemical shift imaging; echo-planar spectroscopic imaging; partial separability; low-rank model; subspace modeling

INTRODUCTION

MR spectroscopic imaging (MRSI or spatially-resolved MR spectroscopy) has been recognized as a powerful tool for non-invasive metabolic studies (1–4), but clinical and

Correspondence to: Zhi-Pei Liang, Ph.D., Franklin W. Woeltge Professor of Electrical and Computer Engineering, Beckman Institute for Advanced Science and Technology, University of Illinois at Urbana-Champaign, 405 N. Mathews Ave, Urbana, IL 61801 USA, z-liang@illinois.edu.

research applications of this technology have been developing more slowly than was expected. One reason for this situation is the problems of long data acquisition time, poor spatial resolution, and low signal-to-noise (SNR) that have limited the use of the existing spectroscopic imaging techniques for in vivo applications.

Over the past few decades, significant efforts have been made to address the above problems, resulting in a large number of new data acquisition and reconstruction methods for spectroscopic imaging. For fast data acquisition, one key approach is to incorporate echo-planar-type of data acquisition schemes with spectroscopic imaging. Many methods (and pulse sequences) have been proposed to implement this data acquisition strategy (see (5) for a comprehensive review; also see (6–14) for a more detailed discussion of some of the EPSI methods). Although EPSI methods can significantly reduce the data acquisition time for spatio-spectral encoding, it is at the expense of SNR, which is a significant limitation for in vivo spectroscopic imaging experiments. Another more recent approach to accelerated spectroscopic imaging is to use parallel imaging (15, 16). It is expected that parallel data acquisition using phased array coils will be an integral part of any method to accelerate spectroscopic imaging experiments.

Advanced image reconstruction for MRSI has been focusing on using prior information to compensate for the lack of sufficient measurements or SNR. To this end, a number of reconstruction models have also been proposed (17–25), but reconstruction methods alone have failed to provide the level of improvements in spatial resolution, data acquisition speed, and SNR needed to have a major impact on in vivo spectroscopic imaging.

In this paper, we propose a new approach to both data acquisition and image reconstruction to accelerate spectroscopic imaging. The proposed approach, called SPICE (SPectroscopic Imaging by exploiting spatio-spectral CorrElation), exploits a unique property known as partial separability (PS) of spectroscopic signals (26,27). This property indicates that high-dimensional spectroscopic signals reside in a very low-dimensional subspace and enables special data acquisition and image reconstruction strategies to be used to obtain high-resolution spatio-spectral distributions with good SNR. More specifically, a hybrid CSI/EPsi pulse sequence is proposed for rapid and sparse sampling of (k, t) -space (or spatio-spectral encoding), and a PS/low-rank model-based algorithm is proposed for subspace estimation and image reconstruction from sparse data with the capability to incorporate prior information and field inhomogeneity correction. The performance of SPICE has been evaluated using both computer simulations and phantom studies. For 2D spectroscopic imaging experiments on a brain metabolite phantom, SPICE achieved effectively 100×100 spatial encodings in 6 minutes. The SNR of the resulting SPICE reconstructions is significantly better than that of the EPSI reconstructions with the same spatio-spectral resolution and data acquisition time, and is comparable to that of the CSI reconstructions from 60×60 spatial encodings acquired in one hour. The proposed method and some representative results are described in the subsequent sections.

THEORY

Subspace Modeling

In spectroscopic imaging, the measured data can be expressed as

$$\begin{aligned} s(\mathbf{k}, t) &= \int_{V\Omega_f} \rho(\mathbf{r}, f) e^{-i2\pi f t} e^{-i\gamma \Delta B(\mathbf{r}) t} e^{-i2\pi \mathbf{k} \mathbf{r}} d\mathbf{r} + \xi(\mathbf{k}, t) \\ &= \int_V \tilde{\rho}(\mathbf{r}, t) e^{-i\gamma \Delta B(\mathbf{r}) t} e^{-i2\pi \mathbf{k} \mathbf{r}} d\mathbf{r} + \xi(\mathbf{k}, t), \end{aligned} \quad [1]$$

where $\rho(\mathbf{r}, f)$ is the desired spatio-spectral function, V the excited volume of interest, Ω_f the desired spectral bandwidth, γ the gyromagnetic ratio, $B(\mathbf{r})$ the B_0 field inhomogeneity map and $\xi(\mathbf{k}, t)$ the measurement noise (often assumed to be white Gaussian).

SPICE exploits an important property of $\tilde{\rho}(\mathbf{r}, t)$, known as spatiotemporal partial separability (PS) (26). More specifically, we express $\tilde{\rho}(\mathbf{r}, t)$ as (26, 27)

$$\tilde{\rho}(\mathbf{r}, t) = \sum_{l=1}^L c_l(\mathbf{r}) \phi_l(t), \quad [2]$$

where $\{\phi_l(t)\}_{l=1}^L$ can be viewed as a set of temporal basis functions, $\{c_l(\mathbf{r})\}_{l=1}^L$ are the corresponding spatial coefficients (can also be viewed as spatial basis), and L is the model order (also called the order of separability (26)). In spectroscopic imaging, the PS model can be justified by the fact that the object imaged has only a finite (L) number of tissue types, each of which has a distinct spectral structure, or that $\rho(\mathbf{r}, f)$ has a finite number (L) of spectral components. It has been shown that partial separability implies low-rankness (26).

More specifically, the Casorati matrix formed from $\tilde{\rho}(\mathbf{r}, t)$ over any point set $\{(\mathbf{r}_p, t_q)\}_{p,q=1}^{P,Q}$

$$\mathcal{C} \left(\{\tilde{\rho}(\mathbf{r}_p, t_q)\}_{p,q=1}^{P,Q} \right) = \begin{bmatrix} \tilde{\rho}(\mathbf{r}_1, t_1) & \tilde{\rho}(\mathbf{r}_1, t_2) & \cdots & \tilde{\rho}(\mathbf{r}_1, t_Q) \\ \tilde{\rho}(\mathbf{r}_2, t_1) & \tilde{\rho}(\mathbf{r}_2, t_2) & \cdots & \tilde{\rho}(\mathbf{r}_2, t_Q) \\ \vdots & \vdots & \ddots & \vdots \\ \tilde{\rho}(\mathbf{r}_P, t_1) & \tilde{\rho}(\mathbf{r}_P, t_2) & \cdots & \tilde{\rho}(\mathbf{r}_P, t_Q) \end{bmatrix}$$

has a rank upper-bounded by L . This implies that $\tilde{\rho}(\mathbf{r}, t)$ resides in a low-dimensional space (or has a small number of degrees of freedom). More specifically, assume that $\tilde{\rho}(\mathbf{r}, t)$ can be represented by $\{\tilde{\rho}(\mathbf{r}_m, t_n)\}_{m,n=1}^{M,N}$ in the conventional pixel representation. Then, when viewed as a vector (formed from $\tilde{\rho}(\mathbf{r}_m, t_n)$), it resides in a MN -dimensional space. However, the PS property implies that $\{\tilde{\rho}(\mathbf{r}_m, t_n)\}_{m,n=1}^{M,N}$ actually resides in a very low-dimensional subspace (spanned by $\{\phi_l(t)\}_{l=1}^L$, for example). This property enables special data acquisition and image reconstruction strategies to be used to obtain high-resolution spatio-spectral distributions with good SNR, which are described next.

Data Acquisition: Extended (k, t) Coverage with Sparse Sampling

To exploit the PS property for high-resolution spectroscopic imaging, SPICE uses a special data acquisition (or spatio-spectral encoding) strategy to achieve extended (k, t) coverage with sparse sampling. The data acquisition scheme (illustrated in Fig. 1) has two key features: (a) acquisition of a data set, $\mathcal{D}_1$ (with limited k -space coverage but in high temporal resolution), for determination of the temporal basis (in high temporal resolution), and (b) acquisition of a data set, $\mathcal{D}_2$ (with extended k -space coverage but limited temporal sampling), for determination of the spatial coefficients (in high spatial resolution). This data acquisition scheme can be implemented using a hybrid CSI/EPSI pulse sequence, an example of which is shown in Fig. 2. The CSI component is used to generate $\mathcal{D}_1$ (with good SNR) while the EPSI component is used to generate $\mathcal{D}_2$ (with high data acquisition speed). With the PS model, SPICE offers a range of flexibility for implementing the EPSI component to collect $\mathcal{D}_2$. For example, in the sequence in Fig. 2, SPICE encodes two spatial dimensions in the FID signals (in contrast to traditional EPSI sequences that usually encodes one spatial dimension) and uses echo shifting (of different excitations) for spectral encoding. As a result, the SPICE sequence samples (k, t)-space only sparsely. This sparse sampling scheme is enabled by the PS model with temporal basis (capturing the full spectral information). It improves the efficiency of spatio-spectral encoding, thus reducing data acquisition time. More alternative data acquisition strategies for collecting $\mathcal{D}_2$ are discussed in the Discussion section.

Reconstruction from Sparse Data

For notation convenience, we assume that $\mathcal{D}_1 = \{s_1(\mathbf{k}_m, t_n)\}_{m,n=1}^{M_1, N_1}$ and $\mathcal{D}_2 = \{s_1(\hat{\mathbf{k}}_m, \hat{t}_n)\}_{m,n=1}^{M_2, N_2}$. It is understood that: (a) $\{t_n\}_{n=1}^{N_1}$ sample the time interval of interest in high resolution while $\{\hat{t}_n\}_{n=1}^{N_2}$ sample the interval sparsely, and (b) $\{\mathbf{k}_m\}_{m=1}^{M_1}$ cover limited k -space locations while $\{\hat{\mathbf{k}}_m\}_{m=1}^{M_2}$ cover the entire desired region of k -space (to provide the necessary spatial resolution). SPICE reconstructs $\rho(\mathbf{r}, t)$ from the data in $\mathcal{D}_1$ and $\mathcal{D}_2$ using a two-step procedure: (a) determination of the temporal basis, and (b) determination of the spatial coefficients.

Determination of the temporal basis (or temporal subspace), $\{\phi_l(t_n)\}_{l=1}^L$ for $n = 1, 2, \dots, N_1$, from $\mathcal{D}_1$ is fairly easy if the effect of field inhomogeneity $B(\mathbf{r})$ on $s_1(\mathbf{k}_m, t_n)$ is negligible. In practice, $B(\mathbf{r})$ can be significant, and we need to first remove/reduce its effect on $s_1(\mathbf{k}_m, t_n)$. Assuming that $B(\mathbf{r})$ is available (which can be acquired during the MRSI experiment), field inhomogeneity correction on limited k -space spectroscopic data can be done using the method in (28), although more advanced methods can also be developed and used for solving this problem. The corrected data, denoted as $s_1(\tilde{\mathbf{k}}_m, t_n)$, is then used to form an $M_1 \times N_1$ Casorati matrix $\mathcal{C} \left(\{s_1(\tilde{\mathbf{k}}_m, t_n)\}_{m,n=1}^{M_1, N_1} \right)$. The SVD is then applied to this Casorati matrix and its L principal right singular vectors are chosen as $\{\phi_l(t_n)\}_{l=1}^L$.

After $\{\phi_l(t_n)\}$ are determined, the spatial coefficients, $\{c_l(\mathbf{r}_m)\}_{l=1}^L$, can be determined from the sparse data in $\mathcal{D}_2$ by solving a least-squares problem. For notation convenience, we

rewrite $\{\tilde{\rho}(\mathbf{r}_m, t_n)\}_{m,n=1}^{M,N} = \{\sum_{l=1}^L c_l(\mathbf{r}_m) \phi_l(t_n)\}_{m,n=1}^{M,N}$ as $\mathbf{C}\Phi$ (27, 29) (with $\mathbf{C} \in \mathbb{C}^{M \times L}$ and $\Phi \in \mathbb{C}^{L \times N}$ such that $\mathbf{C}_{ml} = c_l(\mathbf{r}_m)$ and $\Phi_{ln} = \phi_l(t_n)$, where $M = M_2$ and $N = N_1$ for the proposed data acquisition scheme), and use $\mathbf{s}_2 \in \mathbb{C}^{M_2 N_2 \times 1}$ to denote the vector containing all the data in $\mathcal{D}_2$, $\mathcal{F}_B$ to represent the Fourier transform operator including the B_0 field inhomogeneity effect described in Eq. (1), and $\square$ to represent the (k, t) -space sampling operation used for collecting the data in $\mathcal{D}_2$. Then, the spatial coefficients can be determined as

$$\hat{\mathbf{C}} = \arg \min_{\mathbf{C}} \|\mathbf{s}_2 - \square \mathcal{F}_B \{\mathbf{C}\Phi\}\|_2^2 + \lambda \Psi(\mathbf{C}, \Phi), \quad [3]$$

where the $\|\cdot\|_2^2$ term measures the data consistency of a reconstruction, and $\Psi(\cdot)$ is a regularization function with regularization parameter λ . There are many choices for $\Psi(\cdot)$ to incorporate prior information about $\rho(\mathbf{r}, t)$ or $\rho(\mathbf{r}, f)$ (including both quadratic and sparsity-promoting penalties (20–22,25)). In this paper, we focus on demonstrating the concept and potential of SPICE and use $\Psi(\mathbf{C}, \Phi) = \|\mathbf{WDC}\Phi\|_F^2$, where $\mathbf{D}$ is a finite difference operator and $\mathbf{W}$ contains edge weights derived from a high-resolution anatomical image (20). Note that if $\{\phi_l(t_n)\}$ are orthonormal basis, $\|\mathbf{WDC}\Phi\|_F^2 = \|\mathbf{WDC}\|_F^2$ and an additional diagonal weighting matrix can be included as $\|\mathbf{WDC}\mathbf{A}\|_F^2$ to penalize noisy basis. The weightings in **3** can be determined from the singular value distribution of the Casorati matrix. Integrating this regularization function into Eq. [3] yields a weighted- ℓ_2 regularized least-squares problem that can be solved efficiently (20).

METHODS

We have evaluated SPICE using computer simulations and experimental studies on phantoms.

Simulation Study

We used the high-resolution spatio-spectral numerical phantom described in (30) for our simulation study. The phantom has 128×128 matrix size and 512 time points sampled at a bandwidth of 1200 Hz, and contains spatial distributions of commonly observed ^1H metabolites (see (30) for a more detailed description). We generated (k, t) -space measurements from this phantom with or without field inhomogeneity effects introduced through a coregistered in vivo B_0 map according to Eq. [1]. A simulated full data set without the field inhomogeneity effects was used as the gold standard to compare three acceleration methods: (i) SPICE, (ii) low-resolution CSI, and (iii) EPSI (7). The effect of field inhomogeneity was included for all three schemes, which were set up to have the same acquisition time. Additional noise was added into the simulated EPSI data to account for the SNR difference due to different readout bandwidths. The field inhomogeneity corrected conjugate phase (CP) algorithm (31) was used to obtain the CSI and EPSI reconstructions, which are compared with the corresponding SPICE reconstructions.

Phantom Study

Two phantoms were constructed for our study: a water-oil phantom and a metabolite phantom. The water-oil phantom is a cylindrical jar containing NaCl-doped water and three rows of vials filled with vegetable oil, corn oil and peanut oil, respectively. The metabolite phantom is a cylindrical jar containing NaCl-doped water and five vials with different sizes (Fig. 5a), which were filled with metabolite solutions of physiologically relevant concentrations. More specifically, vial 1 contains 20 mmol/L NAA, 15 mmol/L creatine (Cr) and 10 mmol/L choline-chloride (Cho); Vial 2 contains 20 mmol/L NAA, 10 mmol/L Cr and 5 mmol/L Cho; Vial 3 contains NAA, Cr, Cho and myo-inositol (mI) all at 10 mmol/L; Vial 4 contains 15 mmol/L NAA, 8 mmol/L Cr and 5 mmol/L Cho; and Vial 5 contains the same solution as vial 3.

Experiments on the phantoms were conducted on a 3T Siemens Trio scanner (Siemens Health-care USA) equipped with a 12-channel receiver headcoil. The experiments on the water-oil phantom were focused on evaluating the feasibility of SPICE under practical conditions. So the proposed (k , t)-space sparse sampling scheme and the corresponding reconstruction algorithm were implemented as described. A set of experimental results is shown in Fig. 4, with the following imaging parameters: TR/TE = 1000/30 ms, 220×220 mm² FOV, 5 mm slice thickness, 128×128 matrix size and 167 kHz readout bandwidth (i.e., 6 μ s dwell time). A total of 64 excitations without water suppression was used for acquiring the SPICE data set (16 for $\mathcal{D}_1$ with eight spatial encodings and two temporal interleaves and 48 for $\mathcal{D}_2$ each with 128×128 spatial encodings). The SPICE reconstruction was performed coil by coil with $L = 16$ and followed by SVD-based coil combination (32). For comparison, an EPSI data set (with the same k -space coverage and 256 echoes) was also acquired using 256 excitations. The total acquisition time for the SPICE experiment is 1/4 of that for the EPSI experiment. In addition, a T_1 -weighted anatomical image was acquired using a gradient-echo sequence for defining $\mathbf{W}$, and a B_0 map was acquired using a double gradient-echo field mapping sequence for field inhomogeneity correction.

For the experiments on the metabolite phantom, water suppression is necessary to reveal the spatospectral distributions of the metabolites. Therefore, water-suppressed CSI data were acquired using a modified version of the commercial CSI sequence with: TR/TE = 1000/30 ms, FOV = 250×250 mm², slice thickness = 10 mm, matrix size = 60×60 , FID length = 512, and spectral bandwidth = 2000 Hz. The total data acquisition time was one hour. Water-suppressed EPSI data were acquired using a customized 2D EPSI sequence that encodes chemical shifts plus one spatial dimension for each FID using oscillating gradients. The other spatial dimension was phase encoded. The data acquisition parameters for the EPSI experiment were: TR/TE = 1000/30 ms, FOV = 250×250 mm², slice thickness = 10 mm, matrix size = 100×100 , number of echoes = 256, readout bandwidth = 167 kHz and echo spacing = 830 μ s. The total data acquisition time for two temporal interleaves and five averages was 16.7 minutes. For both the CSI and EPSI sequences, the WET pulses were used for water suppression (33). The HSVD algorithm was also used to further remove any residual water signals in both the CSI and EPSI data (34). As in the water-oil phantom experiments, a T_1 -weighted image (Fig. 5a) and a B_0 map (Fig. 5b) were also acquired.

Similarly to the simulation study, we used a field inhomogeneity corrected reconstruction obtained from the 60×60 CSI data as a reference to compare three acceleration methods: (i) SPICE, (ii) low-resolution CSI, and (iii) EPSI, with the same data acquisition time (6 minutes). The SPICE data set contains 12×12 CSI encodings (for $\mathcal{D}_1$) and 45 echo shifts each with 100×100 spatial encodings averaged five times (for $\mathcal{D}_2$). Note in this experiment, $\mathcal{D}_2$ was retrospectively sampled from the full EPSI data to avoid potential water removal issues with sparse data and only data acquired on the positive gradients were used to avoid echo inconsistency in bipolar acquisition. With the same data acquisition time (6 minutes), the CSI acquisition generated 19×19 spatial encodings while the EPSI acquisition generated 100×100 spatial encodings averaged twice. The SPICE reconstruction was performed coil by coil with $L = 12$ before coil combination. The field inhomogeneity corrected CP algorithm was used for both the CSI and EPSI reconstructions.

RESULTS

Figure 3 shows one set of representative simulation results. The SPICE reconstruction was obtained with $\mathcal{D}_1$ containing 8×8 CSI encodings each with 512 FID samples, $\mathcal{D}_2$ containing 48 echo shifts each with 128×128 spatial encodings (averaged four times) and $L = 8$. The CSI reconstruction was obtained from 16×16 spatial encodings while the EPSI reconstruction was obtained from 128×128 spatial encodings (averaged twice) both with 512 FID samples (under the assumption that the SPICE, CSI, and EPSI have the same acquisition time). As can be seen, the CSI reconstruction shows significant blurring and ringing artifacts, which were reduced in the EPSI reconstruction but at the expense of a significant loss of SNR, as expected. The SPICE reconstruction shows very good resolution and SNR.

Figure 4 presents one set of representative experimental results from the water-oil phantom, which demonstrates that SPICE works with its data acquisition and reconstruction scheme under practical spectroscopic imaging conditions. In this experiment, SPICE offers a factor of four acceleration over the EPSI experiment with comparable results.

Figure 5 shows a set of experimental results from the metabolite phantom. As can be seen, the low-resolution CSI reconstruction has high SNR but suffers from serious truncation artifacts. The EPSI reconstruction has high spatial resolution but very low SNR due to the small voxel size and the high readout bandwidth. SPICE achieves both high spatial resolution and reasonable SNR with the same acquisition time. Note also that the SPICE reconstruction has almost comparable SNR to the high-resolution CSI reconstruction (see the spectra in the right column of Fig. 5), but with a factor of 10 acceleration in data acquisition. Figure 6 compares SPICE reconstructions with different amounts of data. As expected, the reconstructions improve as the number of echo shifts used increases. Note that if a full (instead of sparsely sampled) (k, t) -space EPSI data set is used for determination of the spatial coefficients, SPICE is equivalent to subspace denoising (30) but with a “known” subspace structure determined from a complementary high SNR data set ($\mathcal{D}_1$). The subspace constraint makes the denoising more effective especially for data of very low SNRs.

DISCUSSION

We have demonstrated the feasibility of SPICE to achieve high-resolution spectroscopic imaging with good SNR. Both data acquisition and image reconstruction can be further optimized for practical applications.

First, SPICE data acquisition involves selecting several key parameters including M_1 (the number of encodings in $\mathcal{D}_1$) and N_2 (the number of echo shifts in $\mathcal{D}_2$). Theoretically, M_1 and N_2 need to be not smaller than L , which is easy to satisfy since L is typically a small number. In practice, however, larger M_1 and N_2 are needed to ensure accurate estimation of the temporal basis in the presence of field inhomogeneity and stable solution of the spatial coefficients from Eq. [3]. Based on our experience, $M_1 = 10 \times 10$ for $\mathcal{D}_1$ and $N_2 = 3L$ for $\mathcal{D}_2$ work reasonably well. Nevertheless, M_1 and N_2 should be optimized for a specific MRSI experiment under the experimental time constraints. In addition, SPICE offers a lot of flexibility in sampling (k, t) -space for collecting $\mathcal{D}_1$ and $\mathcal{D}_2$. In our current implementation, $\mathcal{D}_1$ is collected using a CSI sequence, which can be replaced by a traditional EPSI scheme if further improvement in data acquisition speed is desired. $\mathcal{D}_2$ is collected using a modified EPSI sequence with extended k -space coverage and using echo-shifting for spectral encoding. SPICE can also collect $\mathcal{D}_2$ in traditional EPSI trajectories (Fig. 7a) or spiral EPSI trajectories (Fig. 7b). The advantages and disadvantages of these data acquisition schemes should be further studied for a specific MRSI application.

Second, SPICE image reconstruction requires specifying L (the PS model order) and λ (regularization parameter). In principle, L is upper-bounded by the number of spectral components and the number of distinct tissue types in $\rho(\mathbf{r}, f)$ (26, 30). The choice of L can be further modified by examining the singular value distribution of the Casorati matrix formed from the data in $\mathcal{D}_1$ (as is done in this paper), or based on some rank selection criteria (e.g., (30, 35, 36)). In practice, it is desirable to use an L larger than needed or estimated to avoid any potential model truncation errors. This would not create a problem associated with using noisy temporal basis because of the weightings (Λ) on the corresponding spatial coefficients. The regularization parameter, λ , was chosen based on the discrepancy principle (37) in this work. More advanced schemes can also be used (37). Furthermore, SPICE allows for the use of advanced regularization schemes. The current weighted- ℓ_2 regularization is advantageous in terms of computational efficiency and performance characterization. Additional constraints along the spectral dimension can be included. Non-quadratic or sparsity-promoting penalties (21, 25) can also be used to further improve reconstruction quality.

Several other factors should also be considered in optimizing SPICE for practical MRSI applications, including correction of B_0 inhomogeneity, suppression/removal of water and fat signals for ^1H MRSI experiments, and parallel data acquisition. When the field inhomogeneity map is not accurate, the residual field inhomogeneity can increase the PS model order and broaden the spectral lineshape. Based on our experience, with a proper choice of L (discussed above), SPICE is rather tolerant to residual field inhomogeneity. But this issue should be further investigated. For ^1H spectroscopic imaging, suppression/removal of water and fat signals is necessary, which has been well studied for CSI and EPSI

acquisitions. SPICE samples (k, t) -space sparsely for efficient spatio-spectral encoding, thus specialized water and fat suppression/removal methods need to be developed. This issue is under investigation and will be addressed in a subsequent paper. Integrating parallel imaging with SPICE can further accelerate high-resolution spectroscopic imaging experiments, especially for 3D experiments. In the current paper, the parallel imaging capability was not fully exploited and SPICE reconstructions were performed coil by coil and then combined together. Further investigation is necessary to optimize both data acquisition and image reconstruction for parallel SPICE spectroscopic imaging experiments.

It is also worth noting that partial separability, which SPICE is based on, is a general property that exists in various kinds of high-dimensional imaging data. The PS model has been successfully used in several imaging applications, including cardiac imaging (38), dynamic speech imaging (39), parameter mapping (40) and diffusion imaging (41). It has also been utilized for processing CSI data sets (25,30), although SPICE is the first attempt to use the subspace model to dictate both data acquisition and image reconstruction for high-resolution spectroscopic imaging.

CONCLUSIONS

We have presented a new approach called SPICE to accelerate spectroscopic imaging through subspace modeling and sparse sampling of (k, t) -space. Very encouraging results have been obtained from computer simulations and phantom studies, which demonstrate an unprecedented capability of SPICE for high-resolution spectroscopic imaging. With further development to optimize SPICE data acquisition and image reconstruction, SPICE may provide a powerful tool for high-resolution metabolic imaging with many potential applications.

Acknowledgments

The work was supported in part by the following research grants: NIH-P41-EB015904, NIH-P41-EB001977, and NIH-1R01-EB013695, and a Beckman Graduate Fellowship (Fan Lam).

References

1. Lauterbur PC, Kramer DM, House WV, Chen CN. Zeugmatographic high resolution nuclear magnetic resonance spectroscopy: Images of chemical inhomogeneity within macroscopic objects. *J Amer Chem Soc.* 1975; 97:6866–6868.
2. Brown TR, Kincaid BM, Ugurbil K. NMR chemical shift imaging in three dimensions. *Proc Natl Acad Sci.* 1982; 79:3523–3526. [PubMed: 6954498]
3. Maudsley AA, Hilal SK, Perman WH, Simon HE. Spatially resolved high resolution spectroscopy by “four-dimensional” NMR. *J Magn Reson.* 1983; 51:147–152.
4. Posse S, Otazo R, Dager SR, Alger J. MR spectroscopic imaging: Principles and recent advances. *J Magn Reson Imag.* 2013; 37:1301–1325.
5. Pohmann R, von Kienlin M, Haase A. Theoretical evaluation and comparison of fast chemical shift imaging methods. *J Magn Reson.* 1997; 129:145–160. [PubMed: 9441879]
6. Mansfield P. Spatial mapping of the chemical shift in NMR. *Magn Reson Med.* 1984; 1:370–386. [PubMed: 6571566]
7. Posse S, Tedeschi G, Risinger R, Ogg R, Bihan DL. High speed ^1H spectroscopic imaging in human brain by echo planar spatial-spectral encoding. *Magn Reson Med.* 1995; 33:34–40. [PubMed: 7891533]

8. Mulkern RV, Panych LP. Echo planar spectroscopic imaging. *Concepts Magn Reson*. 2001; 13:213–237.
9. Posse S, Otazo R, Caprihan A, Bustillo J, Chen H, Henry PG, Marjanska M, Gasparovic C, Zuo C, Magnotta V, Mueller B, Mullins P, Renshaw P, Ugurbil K, Lim KO, Alger JR. Proton echo-planar spectroscopic imaging of J-coupled resonances in human brain at 3 and 4 Tesla. *Magn Reson Med*. 2007; 58:236–244. [PubMed: 17610279]
10. Ebel A, Maudsley AA, Weiner MW, Schuff N. Achieving sufficient spectral bandwidth for volumetric 1H echo-planar spectroscopic imaging at 4 Tesla. *Magn Reson Med*. 2005; 54:697–701. [PubMed: 16086316]
11. Adalsteinsson E, Irarrazabal P, Topp S, Meyer C, Macovski A, Spielman DM. Volumetric spectroscopic imaging with spiral-based k-space trajectories. *Magn Reson Med*. 1998; 39:889–898. [PubMed: 9621912]
12. Schirda CV, Tanase C, Boada FE. Rosette spectroscopic imaging: Optimal parameters for alias-free, high sensitivity spectroscopic imaging. *J Magn Reson Imag*. 2009; 29:1375–1385.
13. Guimaraes AR, Baker JR, Jenkins BG, Lee PL, Weisskoff RM, Rosen BR, Gonzalez RG. Echoplanar chemical shift imaging. *Magn Reson Med*. 1999; 41:877–882. [PubMed: 10332868]
14. Maudsley AA, Domenig C, Govind V, Darkazanli A, Studholme C, Arheart K, Bloomer C. Mapping of brain metabolite distributions by volumetric proton MR spectroscopic imaging (MRSI). *Magn Reson Med*. 2009; 61:548–559. [PubMed: 19111009]
15. Lin FH, Tsai SY, Otazo R, Caprihan A, Wald LL, Belliveau JW, Posse S. Sensitivity-encoded (SENSE) proton echo-planar spectroscopic imaging (PEPSI) in the human brain. *Magn Reson Med*. 2007; 57:249–257. [PubMed: 17260356]
16. Tsai SY, Otazo R, Posse S, Lin YR, Chung HW, Wald LL, Wiggins GC, Lin FH. Accelerated proton echo planar spectroscopic imaging (PEPSI) using GRAPPA with a 32-channel phased-array coil. *Magn Reson Med*. 2008; 59:989–998. [PubMed: 18429025]
17. Hu X, Levin DN, Lauterbur PC, Spraggins T. SLIM: Spectral localization by imaging. *Magn Reson Med*. 1988; 8:314–322. [PubMed: 3205158]
18. Liang ZP, Lauterbur PC. A generalized series approach to MR spectroscopic imaging. *IEEE Trans Med Imag*. 1991; 10:132–137.
19. Jacob M, Zhu X, Ebel A, Schuff N, Liang ZP. Improved model-based magnetic resonance spectroscopic imaging. *IEEE Trans Med Imag*. 2007; 26:1305–1318.
20. Haldar JP, Hernando D, Song SK, Liang ZP. Anatomically constrained reconstruction from noisy data. *Magn Reson Med*. 2008; 59:810–818. [PubMed: 18383297]
21. Eslami R, Jacob M. Robust reconstruction of MRSI data using a sparse spectral model and high resolution MRI priors. *IEEE Trans Med Imag*. 2010; 29:1297–1309.
22. Kornak J, Young K, Soher BJ, Maudsley AA. Bayesian k-space-time reconstruction of MR spectroscopic imaging for enhanced resolution. *IEEE Trans Med Imag*. 2010; 29:1333–1350.
23. Zhang Y, Gabr RE, Schar M, Weiss RG, Bottomley PA. Magnetic resonance spectroscopy with linear algebraic modeling (SLAM) for higher speed and sensitivity. *J Magn Reson*. 2012; 218:66–76. [PubMed: 22578557]
24. Chatnuntawech, I.; Bilgic, B.; Adalsteinsson, E. Undersampled spectroscopic imaging with model-based reconstruction. *Proc Intl Symp Magn Reson Med*; Salt Lake City, UT, USA. 2013. p. 3960
25. Kasten J, Lazeyras F, Van De Ville D. Data-driven MRSI spectral localization via low-rank component analysis. *IEEE Trans Med Imag*. 2013; 32:2266–2279. [PubMed: 2266259]
26. Liang, ZP. Spatiotemporal imaging with partially separable functions. *Proc IEEE Int Symp on Biomed Imag*; Arlington, VA, USA. 2007. p. 988–991.
27. Haldar, JP.; Liang, ZP. Spatiotemporal imaging with partially separable functions: A matrix recovery approach. *Proc IEEE Int Symp on Biomed Imag*; Rotterdam, Netherlands. 2010. p. 716–719.
28. Peng X, Nguyen H, Haldar JP, Hernando D, Wang XP, Liang ZP. Correction of field inhomogeneity effects on limited k-space MRSI data using anatomical constraints. *Proc IEEE Eng Med Biol Soc*. 2010:883–886.
29. Haldar JP, Hernando D. Rank-constrained solutions to linear matrix equations using power-factorization. *IEEE Signal Process Lett*. 2009; 16:584–587. [PubMed: 22389578]

30. Nguyen HM, Peng X, Do MN, Liang ZP. Denoising MR spectroscopic imaging data with low-rank approximations. *IEEE Trans Biomed Eng.* 2013; 60:78–89. [PubMed: 23070291]
31. Noll DC, Fessler JA, Sutton BP. Conjugate phase MRI reconstruction with spatially variant sample density correction. *IEEE Trans Med Imag.* 2005; 24:325–336.
32. Bydder M, Hamilton G, Yokoo T, Sirlin CB. Optimal phased-array combination for spectroscopy. *Magn Reson Imag.* 2008; 26:847–850.
33. Ogg RJ, Kingsley RB, Taylor JS. WET, a T1- and B1-insensitive water-suppression method for in vivo localized 1H NMR spectroscopy. *J Magn Reson.* 1994; 104:1–10.
34. Barkhuysen H, de Beer R, van Ormondt D. Improved algorithm for noniterative time-domain model fitting to exponentially damped magnetic resonance signals. *J Magn Reson.* 1987; 73:553–557.
35. Ulfarsson M, Solo V. Dimension estimation in noisy PCA with SURE and random matrix theory. *IEEE Trans Signal Process.* 2008; 56:5804–5816.
36. Minka T. Automatic choice of dimensionality for PCA. *Adv Neural Inform Process Syst.* 2001; 13:598–604.
37. Vogel, CR. *Computational Methods for Inverse Problems.* Philadelphia, PA: SIAM; 2002.
38. Zhao B, Haldar JP, Christodoulou AG, Liang ZP. Image reconstruction from highly undersampled (k, t)-space data with joint partial separability and sparsity constraints. *IEEE Trans Med Imag.* 2012; 31:1809–1820.
39. Fu, M.; Christodoulou, AG.; Naber, AT.; Kuehn, DP.; Liang, ZP.; Sutton, BP. High-frame-rate multi-slice speech imaging with sparse sampling of (k,t)-space. *Proc Intl Symp Magn Reson Med; Melbourne, Australia.* 2012. p. 12
40. Zhao, B.; Lu, W.; Liang, Z. Highly accelerated parameter mapping with joint partial separability and sparsity constraints. *Proc Intl Symp Magn Reson Med; Melbourne, Australia.* 2012. p. 2233
41. Lam F, Babacan SD, Haldar JP, Weiner MW, Schuff N, Liang ZP. Denoising diffusion-weighted magnitude MR images using rank and edge constraints. *Magn Reson Med.* 2013;10.1002/mrm.24728.

Figure 1.

An example of SPICE (k , t)-space sampling for 2D spectroscopic imaging (with k_x pointing into the page): (a) (k , t)-space sampling for data in $\mathcal{D}_1$ for subspace estimation. $\mathcal{D}_1$ covers only a limited region of central k -space (based on SNR consideration), and sample the free precession period ($T_{A;1}$) fully (to capture the spectral information); (b) (k , t)-space sampling for data in $\mathcal{D}_2$ for determination of the spatial coefficients. Note that a set of FIDs with different echo shifts ($t_1, t_2, \dots, t_q$) is collected, each of which traverses entire k -space (also see Fig. 2) but with limited spectral encoding ($T_{A;2} < T_{A;1}$).

Figure 2.

A prototypical hybrid CSI/EPSI sequence for SPICE: (a) the CSI component used to collect the data in $\mathcal{D}_1$ with limited k -space coverage but full spectral encoding, and (b) an EPSI-like component used to acquire the data in $\mathcal{D}_2$ with extended k -space coverage but limited spectral encoding. Note that there is only one k_y reversal in each TR due to SNR consideration, although more k_y reversals can be included in principle. Note also that the proposed EPSI component supports bipolar acquisition but requires additional correction if data acquired on both positive and negative Gx are used.

Figure 3.

Simulation results: (a) the gold standard, (b) CSI reconstruction from 16×16 spatial encodings, (c) EPSI reconstruction from 128×128 spatial encodings averaged twice, and (d) SPICE reconstruction from 8×8 CSI encodings in $\mathcal{D}_1$, 48 echo-shifts in $\mathcal{D}_2$ averaged four times and $L = 8$. The left column shows the spatial distributions of a frequency component at 345 Hz and the right column shows the spectra corresponding to the voxel identified by the red dot for each case.

Figure 4.

Experimental results from a water-oil phantom: (a) field inhomogeneity corrected CP reconstruction from an EPSI data with 128×128 spatial encodings, and (b) SPICE reconstruction with eight conventional EPSI encodings in $\mathcal{D}_1$ (two temporal interleaves), 48 echo shifts in $\mathcal{D}_2$ (a total of 64 excitations) and $L = 16$. Note that the reconstructions are almost comparable but the data acquisition time for the SPICE experiment is only 1/4 of that for the EPSI experiment.

Figure 5.

Experimental results from a metabolite phantom shown in (a) with spectroscopic imaging data acquired in the presence of B_0 inhomogeneity shown in (b): (c) CSI reconstruction from 60×60 spatial encodings, (d) CSI reconstruction from 19×19 spatial encodings, (e) EPSI reconstruction from 100×100 spatial encodings with two averages, and (f) SPICE reconstruction with 12×12 CSI encodings in $\mathcal{D}_1$, 45 echo shifts in $\mathcal{D}_2$ with five averages, and $L = 12$. The left column shows the spatial distributions of NAA and the right column shows the spectra from the voxel identified by the red dot in (a). Field inhomogeneity correction was included for all the cases as described in the text. The results in (d)–(f) correspond to a factor of 10 reduction in data acquisition time (6 min) compared to the high-resolution CSI acquisition in (c).

Figure 6.

Experimental results from the metabolite phantom in Fig. 5a: (a) the high-resolution CSI reconstruction in Fig. 5c and SPICE reconstructions with 12×12 CSI encodings in $\mathcal{D}_1$, $L = 12$ and different numbers of echo shifts in $\mathcal{D}_2$: (b) 45 echo shifts, (c) 80 echo shifts, and (d) the entire EPSI data set.

Figure 7.

Two alternative (k, t) -space sampling trajectories that can be used to generate $\mathcal{D}_2$: (a) EPSI trajectories, and (b) spiral EPSI trajectories. Note that in (a), the echo spacing is not constrained by the spectral Nyquist criterion as in traditional EPSI schemes. In both cases, different colors represent trajectories for different excitations.