Published in final edited form as:

Diagn Microbiol Infect Dis. 2013 January ; 75(1): 9–15. doi:10.1016/j.diagmicrobio.2012.09.003.

Single-Tier Testing with the C6 Peptide ELISA Kit Compared with Two-Tier Testing for Lyme Disease

Gary P. Wormser, M.D.^a, Martin Schriefer, Ph.D.^b, Maria E. Aguero-Rosenfeld, M.D.^c, Andrew Levin, Ph.D.^d, Allen C. Steere, M.D.^e, Robert B. Nadelman, M.D.^a, John Nowakowski, M.D.^a, Adriana Marques, M.D.^f, Barbara J. B. Johnson, Ph.D.^b, and J. Stephen Dumler, M.D.^g

^aDivision of Infectious Diseases, New York Medical College, Valhalla, NY

^bNational Center for Emerging and Zoonotic Infectious Diseases, Centers for Disease Control and Prevention, Fort Collins, CO

^cDepartment of Pathology, New York University Langone Medical Center, New York, NY

^dImmunetics, Inc., Boston, MA

^eDepartment of Medicine, Harvard Medical School, Boston, MA

^fNational Institute of Allergy and Infectious Diseases, National Institutes of Health, Bethesda, MD

^gDepartment of Pathology, Johns Hopkins University School of Medicine

Abstract

Background—The two-tier serologic testing protocol for Lyme disease has a number of shortcomings including low sensitivity in early disease; increased cost, time and labor; and subjectivity in the interpretation of immunoblots.

Methods—The diagnostic accuracy of a single-tier commercial C6 ELISA kit was compared with two-tier testing.

Results—The C6 ELISA was significantly more sensitive than two-tier testing with sensitivities of 66.5% (95% C.I.:61.7-71.1) and 35.2% (95% C.I.:30.6-40.1), respectively ($p<0.001$) in 403 sera from patients with erythema migrans. The C6 ELISA had sensitivity statistically comparable to two-tier testing in sera from Lyme disease patients with early neurological manifestations (88.6% vs. 77.3%, $p=0.13$) or arthritis (98.3% vs. 95.6%, $p=0.38$).

© 2012 Elsevier Inc. All rights reserved.

Corresponding Author: Gary P. Wormser, M.D., Division of Infectious Diseases, Munger Pavilion, Rm. 245, New York Medical College, Valhalla, NY 10595, 914-493-1560 (Tel), 914-493-7289 (Fax), gary_wormser@nymc.edu.

Publisher's Disclaimer: This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our customers we are providing this early version of the manuscript. The manuscript will undergo copyediting, typesetting, and review of the resulting proof before it is published in its final citable form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

Competing interests

AL is President and CEO of Immunetics, Inc.; GPW has received research support under the above-referenced NIAID grant to Immunetics, Inc. to fund the study reported herein.

The specificities of C6 ELISA and two-tier testing in over 2200 blood donors, patients with other conditions, and Lyme disease vaccine recipients were found to be 98.9% and 99.5%, respectively ($p < 0.05$, 95% C.I. surrounding the 0.6 percentage point difference of 0.04 to 1.15).

Conclusions—Using a reference standard of two-tier testing, the C6 ELISA as a single step serodiagnostic test provided increased sensitivity in early Lyme disease with comparable sensitivity in later manifestations of Lyme disease. The C6 ELISA had slightly decreased specificity. Future studies should evaluate the performance of the C6 ELISA compared with two-tier testing in routine clinical practice.

Keywords

Lyme disease; *Borrelia burgdorferi*; serology; C6 ELISA

Lyme disease, a tick-borne spirochetosis, has been recognized as the most frequent vector-borne disease in the United States since its identification over two decades ago (Bacon et al., 2008; Steere et al., 1983). The mainstay of laboratory diagnosis is detection of antibody to *Borrelia burgdorferi*, the etiologic agent (Aguero-Rosenfeld et al., 1993; Aguero-Rosenfeld et al., 2005; Craft et al., 1986; GrodzickiSteere, 1988; Wormser et al., 2006). To improve the specificity of serologic testing for Lyme disease, a two-tier approach was recommended in 1995 (Centers for Disease Control and Prevention, 1995). Accordingly, a serum sample yielding a positive or indeterminate result in the first-tier assay (typically an ELISA) is re-tested by separate IgM and IgG immunoblots usually based on *B. burgdorferi* whole cell sonicate (WCS). The immunoblot is the second tier assay and must be positive for the serum to be considered seropositive.

The public health service recommendations in support of two-tier testing (Centers for Disease Control and Prevention, 1995) provide for the development of alternatives to one or both steps provided that equal or better performance is demonstrated by such alternative methods ("Proceedings of the Second National Conference on Serologic Diagnosis of Lyme Disease," 1995). In principle, an ELISA of sufficiently high specificity and sensitivity would provide diagnostic information similar to existing two-tier testing. A variety of recombinant and synthetic antigens have been evaluated in previous studies for use in serodiagnosis of Lyme disease (Bacon, et al., 2008; Branda et al., 2010; CincoMurgia, 2006; Embers et al., 2007a; Embers et al., 2007b; Gomes-Solecki et al., 2007; Gottner et al., 2004; HauserWilske, 1997; Heikkila et al., 2003; Heikkila et al., 2002; Jobe et al., 2008; Lawrenz et al., 1999; Liang et al., 1999a; Liang et al., 1999b; Magnarelli et al., 2000; Magnarelli et al., 2002; Mathiesen et al., 1998; Mogilyansky et al., 2004; Padula et al., 1994; Panelius et al., 2001; Peltomaa et al., 2004; Rasiah et al., 1994; Sillanpaa et al., 2007; Skarpaas et al., 2007; Smismans et al., 2006; Steere et al., 2008; Tjernberg et al., 2007; Tjernberg et al., 2008; Wormser et al., 2008a). In particular, the VlsE protein and the highly conserved 25-amino acid peptide ("C6 peptide") derived from the sixth invariant region of this protein have been shown to be both sensitive and specific antigens in the ELISA (Bacon et al., 2003; CincoMurgia, 2006; Embers, et al., 2007a; Embers, et al., 2007b; Gomes-Solecki, et al., 2007; Gottner, et al., 2004; Heikkila, et al., 2003; Lawrenz, et al., 1999; Liang, et al., 1999a; Liang, et al., 1999b; Magnarelli, et al., 2002; Mogilyansky, et al., 2004; Peltomaa, et al., 2004; Sillanpaa, et al., 2007; Skarpaas, et al., 2007; Smismans, et al., 2006; Steere, et al.,

2008; Tjernberg, et al., 2007; Tjernberg, et al., 2008; Wormser, et al., 2008a) and immunoblot formats (Branda, et al., 2010).

The aim of the present study was to determine if the C6 ELISA by itself is a suitable alternative to two-tier testing by evaluating the comparative diagnostic accuracy in over 550 sera from well-characterized Lyme disease patients and in more than 2200 control sera.

Methods

Serological assays. The C6 Lyme ELISA kit (Immunetics, Inc.), a test kit approved by the Food and Drug Administration (FDA) for use as a first-tier assay, was modified for this study. Kits provided by the manufacturer for this study incorporated a simplified and comparable cut-off formula that was based on a negative calibrator serum. This modification yielded sensitivity and specificity statistically equivalent to what had been demonstrated with the original cut-off formula. Reproducibility testing of the C6 ELISA kit using the simplified cut-off yielded intra-assay and inter-assay coefficients of variation (CV) of 10% and 11.6%, respectively. Intra-assay and inter-assay CV's for the kit with the original cut-off were 9.5% and 14.3%, respectively. In all other respects the modified and original kits were the same.

The C6 peptide used as antigen in the kit is derived from the *B. burgdorferi* B31 strain sequence, which differs from the originally described IP90 sequence (Liang, et al., 1999b) by 4 amino acids. The kit is formatted as an indirect ELISA in which both IgG and IgM antibodies to C6 peptide are detected by an enzyme conjugate. The C6 ELISA testing was performed at New York Medical College, the Centers for Disease Control and Prevention (CDC) and at Immunetics, Inc. Two-tier serology was performed using IgG/IgM ELISA kits from Wampole Laboratories (at New York Medical College and Immunetics, Inc.) or IgG/IgM Vidas II Lyme Screening kits from bioMerieux Corp. (at the CDC), followed by Lyme IgG and IgM immunoblot kits from MarDx/Trinity Biotech. The two-tier ELISA and immunoblot kits were based on a *B. burgdorferi* WCS antigen and were also FDA-approved for in vitro diagnostic use. An alternative two-tier result was calculated based on the C6 ELISA result as the first step, combined with the same IgG and IgM immunoblot results as above. For the majority of the Lyme disease sera tested, the WCS ELISA-two-tier serologic testing was performed as part of the current study. However, for 156 sera from one of the panels for which the volume of serum was limited, we instead used the WCS ELISA-two-tier serologic test results that were obtained in 2002 as part of another study (Bacon, et al., 2003); these tests had also been performed using the Vidas II Lyme Screening kit and the MarDx/Trinity Biotech kits.

Lyme Patient Sera. The patient sera selected for this study were chosen to represent a broad range of clinical manifestations of Lyme disease (Table 1) from well-characterized patients. The sera comprised multiple reference panels from different sources that had been previously collected and frozen at -80°C . The 569 Lyme disease sera studied were obtained from 528 patients. Erythema migrans was defined based on clinical diagnosis alone for 172 sera and on clinical diagnosis in conjunction with microbiologic confirmation of *B. burgdorferi* infection by either culture or polymerase chain reaction (PCR) for 231 sera. The

test results for some of these sera were previously reported (Wormser, et al., 2008a; Wormser et al., 2008b).

Early neurologic Lyme disease was defined based on the presence of objective clinical (e.g., facial nerve palsy) and/or laboratory findings (e.g., cerebrospinal fluid [CSF] lymphocytic pleocytosis) in association with concomitant/recent erythema migrans for 39 sera; for the other 5 sera evidence for *B. burgdorferi* infection was based on prior seropositivity by at least a positive WCS ELISA on an acute or convalescent phase serum specimen. Late neurologic Lyme disease was defined based on the presence of a compatible objective clinical finding (e.g., encephalopathy, polyneuropathy or encephalomyelitis) in association with serologic evidence of borrelial infection demonstrated by at least a positive WCS ELISA. Acute EM or neurologic Lyme disease sera were collected prior to or on the date of first antibiotic treatment; convalescent sera were collected after the completion of antibiotic treatment.

Lyme arthritis was defined as the presence of joint swelling that was clinically compatible with Lyme arthritis in conjunction with serologic evidence of borrelial infection demonstrated by at least a positive WCS ELISA.

Serum samples which had been previously found to be seronegative by an ELISA and which had been obtained from patients with erythema migrans in whom the diagnosis was confirmed by culture or PCR (n= 58) were purposely included in this study to challenge the sensitivity of the methods under evaluation.

Controls. Serum samples were obtained from healthy blood donors from endemic (U.S. Northeast) and non-endemic regions for Lyme disease (U.S. Southeast and Southwest). In addition, other control sera were tested from patients with other diseases or conditions and from recipients of the LymeRx® vaccine

The protocol was approved by the Institutional Review Board of New York Medical College. Serum panels were either obtained from volunteers who gave written informed consent to have their serum used to improve diagnostic tests for Lyme disease, or came from anonymous, pre-existing reference samples that had previously been unlinked from the serum donor.

Data and Statistical Analysis

Sensitivity was calculated as percent positive in the Lyme disease populations, while specificity was calculated as per cent negative in the control sera. Comparisons were made using the Fisher exact test. Where appropriate, McNemar's test was used to account for the paired measurements on samples. P values were two-tailed and a value of <0.05 was considered significant. All analyses were conducted using Stata (version 11.2, StataCorp, TX).

Results

Specificity

Serum samples were obtained from healthy blood donors from endemic (n=1329) (U.S. Northeast) and non-endemic (n=513) regions for Lyme disease (U.S. Southeast and Southwest) (Table 2). In addition, 399 other control sera were tested including sera from individuals with the conditions listed in Table 3 and sera from recipients of the LymeRx® vaccine (n=33). Specificities of the C6 ELISA and WCS ELISA-based two-tier testing in a population of healthy blood donors from regions of the U.S. non-endemic for Lyme disease were both above 99%, but the specificity of two-tier testing exceeded that of the C6 ELISA by 0.6 percentage points (95% C.I. surrounding the 0.6 percentage point difference of 0 to 2.0, $p=0.38$) (Table 2). Among donors from regions endemic for Lyme disease, the specificities of both the C6 ELISA and two-tier testing were approximately 0.5 percentage points lower than among donors from non-endemic areas (98.6% vs. 99.4%, respectively, $p=0.006$). Among the entire blood donor population of 1,842 individuals, two-tier testing was associated with a significantly higher specificity of 99.5% compared with 98.8% for the C6 ELISA ($p=0.002$, 95% C.I. surrounding the 0.7 percentage point difference of 0.11 to 1.36).

Testing 366 sera from individuals with 14 other disease conditions, or with several common potential interferences, yielded a specificity for the C6 ELISA of 99.5% (95% C.I.: 98.0-99.9) compared with 99.2% (95% C.I.: 97.6-99.8) for WCS ELISA-based two-tier testing (Table 3) ($p=1.0$). The two sera that were positive by C6 ELISA in this group were also positive by two-tier testing. Of the 33 LymeRx® vaccinee sera, 1 was C6 ELISA positive; none was two-tier positive, although 29 were positive and 1 was in the indeterminate range on WCS ELISA.

Overall, the specificities of C6 ELISA and WCS ELISA-based two-tier testing in the 2241 control sera evaluated were 98.9% and 99.5%, respectively ($p<0.05$; 95% C.I. surrounding the 0.6 percentage point difference of 0.04 to 1.15). Among the 25 sera that were reactive by the C6 ELISA, 9 (36.0%) tested positive by the two-tier method. Conversely, among the 12 sera that were reactive by two-tier testing, 9 (75%) were positive by the C6 ELISA. By comparison, the specificity of the first-tier WCS ELISAs was significantly less than that of the C6 ELISA for the 4 groups of controls (95.9% for WCS ELISA vs. 99.2% for C6 ELISA in non-endemic blood donors, $p<0.001$; 96.5% for WCS ELISA vs. 98.6% for C6 ELISA in endemic blood donors, $p<0.001$; 89.3% for WCS ELISA vs. 99.5% for C6 ELISA in other diseases or interfering conditions, $p<0.001$; and 9.1% for WCS ELISA vs. 97.0% for C6 ELISA in LymeRx® vaccine recipients, $p<0.001$) (Tables 2 and 3).

Substitution of the C6 ELISA for the WCS ELISA in the two-tier testing yielded essentially identical results for specificity in each of the control groups (Tables 2 and 3).

Sensitivity

Sensitivity vs. disease manifestation—Sample demographics are shown in Tables 1 and 4. Patients who had both erythema migrans and an extracutaneous manifestation were tabulated only under the category of the extracutaneous manifestation in this analysis. The

overall sensitivities of the C6 ELISA and of WCS ELISA-based two-tier testing were 75.0% (95% C.I.: 71.3-78.5) and 51.5% (95% C.I.: 47.3-55.7), respectively, in the group of 569 Lyme disease sera previously collected from 528 patients (Table 4) ($p < 0.001$). Sera were selected for this study to represent a broad range of clinical manifestations of Lyme disease. The elimination of the 157 sera from patients who had previously been treated with antimicrobials did not significantly ($p = 0.5$) alter these values (overall sensitivity for C6 ELISA = 73.1%, 95% C.I.: 68.5-77.3%; for WCS ELISA-based two tier testing = 53.9%, 95% C.I.: 48.9-58.8%). The higher sensitivity of the C6 ELISA was entirely due to the insensitivity of the second-tier immunoblot test, primarily during early disease; indeed, the sensitivity of the first-tier WCS ELISA exceeded that of the C6 ELISA (81.7% versus 75.0% for all Lyme sera combined, $p < 0.001$).

On acute phase samples, the difference in sensitivity between the C6 ELISA and WCS ELISA-based two-tier testing was most evident in the group of sera from patients presenting with a single erythema migrans skin lesion, in which the sensitivity of the C6 ELISA was more than twice that of two-tier testing (57.9% vs. 27.1%, $p < 0.001$) (Table 4). In sera from patients with multiple erythema migrans skin lesions, the sensitivity of both methods rose substantially, but the C6 ELISA was 41% more sensitive than two-tier testing in this subgroup ($p < 0.001$) (Table 4). Seropositivity on acute phase serum samples of culture or PCR-confirmed cases of erythema migrans was similar to that of clinically-defined cases for both the C6 ELISA (135/209 [64.6%, 95% C.I.: 57.7-71.1%]) versus 56/89 [62.9%, 95% C.I.: 52.0-72.9%]) and for two-tier testing (80/209 [38.3%, 95% C.I.: 31.7-45.2%]) versus 34/89 [38.2%, 95% C.I.: 28.1-49.1%]).

In patients with early neurologic Lyme disease including facial palsy, meningitis and radicular neuropathy, a total of 20 sera obtained during the acute phase before antibiotic treatment and 24 obtained during the convalescent phase after antibiotic therapy was initiated were evaluated in this study. The convalescent phase samples were obtained from patients separate from those who provided the acute phase serum samples. The sensitivity of the C6 ELISA was comparable to that of WCS ELISA-based two-tier testing in these groups (90.0% vs. 80.0%, $p = 0.50$, in the acute phase serum samples and 87.5% vs. 75.0%, $p = 0.38$, in the convalescent phase group).

In the group of 114 sera from patients diagnosed with Lyme arthritis, sensitivities of the C6 ELISA and WCS ELISA-based two-tier testing were 98.2% and 95.6%, respectively ($p = 0.38$). Similarly, in the 8 patients with late neuroborreliosis, the sensitivities of the C6 ELISA and of two-tier testing were both high (100%, $p > 0.99$).

WCS ELISA-based two-tier testing was positive on sera from only 7 patients who failed to react by the C6 ELISA, of whom 5 had erythema migrans, 1 had Lyme arthritis and 1 had early neurologic Lyme disease. Among the 200 serum samples found to be two-tier positive by at least a positive IgG immunoblot, two were negative by C6 ELISA. In contrast, among the 92 sera found to be two-tier positive based on a positive WCS ELISA and only a positive IgM immunoblot, 5 were negative by C6 ELISA testing (2/200 vs. 5/92, $p = 0.03$).

The sensitivity of the C6 ELISA-based two-tier testing was indistinguishable from that of WCS ELISA-based two-tier testing overall (50.6% vs. 51.5%, $p=0.81$) and in each of the individual patient groups (Table 4).

COMMENT

In this evaluation of the performance of serologic methods for detection of antibodies to *B. burgdorferi*, the C6 ELISA as a stand-alone test had significantly greater sensitivity than either the WCS ELISA-based two-tier testing method (66.5% vs. 35.2%, $p<0.001$) or the C6 ELISA-based two-tier testing method (66.5% vs. 34.5%, $p<0.001$) for patients with erythema migrans (Table 4). In contrast, on sera from the groups of patients with neurologic or rheumatologic manifestations of Lyme disease each of the three testing methods had high sensitivity without a statistically significant difference ($p>0.05$) (Table 4). The specificity of either of the two-tier testing methods on more than 2200 control sera, however, exceeded that of the C6 ELISA alone by 0.6 percentage points (99.5% versus 98.9%, 95% C.I. surrounding the difference: 0.04 to 1.15), and this difference was significant ($p<0.05$). The 0.4-0.7 percentage point lower specificity values observed for these serologic methods in the endemic compared with the non-endemic blood donor group might be explained by the inclusion of individuals with prior exposure to *B. burgdorferi* in the endemic group, although this hypothesis cannot be evaluated due to the lack of information on donor histories.

Our overall results for specificity are quite comparable to that for the C6 ELISA in another published study that involved a sizeable but smaller number of control serum samples (present study 98.9% [2216/2241] vs. 98.4% [1279/1300], $p=0.22$) (Branda et al., 2011). The authors of that report (Branda, et al., 2011) observed that a novel two-tier strategy using a WCS ELISA followed by the C6 ELISA would have increased the specificity to 99.5%, a figure that was identical to the conventional WCS ELISA-immunoblot based two-tier testing. A re-analysis of our study data with such an algorithm yielded the same 99.5% specificity value. Interestingly, however, in the group of individuals with other disease conditions as well as the group of LymeRx® vaccine recipients, which collectively might be more representative of the target population for testing than blood donors, the specificity of this two-ELISA algorithm was identical to that of the C6 ELISA alone. Serial two-step algorithms generically offer improved specificity at the cost of some loss of sensitivity; in the present study, the serial ELISA algorithm would result in a loss of about 6% in sensitivity for EM patients (Table 4), and a similar decrease was observed by Branda et al (Branda, et al., 2011). A cost-benefit analysis of this tradeoff may indicate its utility relative to other algorithms.

Five C6 ELISA-positive blood donor sera out of 8 which were retested two years later were reproducibly C6 ELISA-positive, suggesting that random assay error was not responsible for the positive results among control sera in most cases. In parallel, a subset of 188 endemic blood donor sera that were negative when first assayed - about 14% of the total in this group - were retested, and yielded an almost identical distribution of absorbance readings (none positive, none indeterminate, median Lyme index value = 0.182 vs. 0.155 when originally

tested [Lyme index value of cut-off = 1 for reference]). These results indicate the relatively consistent performance of the C6 ELISA in testing control populations.

The low sensitivity of WCS ELISA-based two-tier testing in detection of patients with early Lyme disease associated with erythema migrans reported in previous studies (Bacon, et al., 2003; Grodzicki and Steere, 1988; Nowakowski et al., 2001) was confirmed in the present study. Sensitivity was under 30% in patients with a single erythema migrans skin lesion. By comparison, sensitivity was over twice this value when the same sera were tested by the C6 ELISA. The maximum sensitivity which can be achieved at this stage of the infection may be limited by the lag time intrinsic to the primary immune response, rather than the specific assay methodology. In patients with multiple erythema migrans, sensitivity rose substantially with both methods, to 63.8% and 89.7% for two-tier testing and C6 ELISA, respectively. Our results suggest that an immune response to the C6 peptide appears at an early point in time, often developing more rapidly than either the IgM or IgG response to the combination of antigens on which the two-tier immunoblot criteria are based (Bacon, et al., 2003). The *vlsE* gene encoding the C6 peptide is expressed early in the infection of mammals, but is poorly expressed in cultured borreliae, the source of antigens for WCS ELISAs and immunoblots (Norris et al., 1992).

Although some of the sera obtained from patients with erythema migrans that were evaluated in this study were specifically selected based on negative archival results with a WCS ELISA, negative results on the first-tier WCS ELISAs used in this study were not the explanation for the lower sensitivity of two-tier testing (Table 4). The selection of the MarDx/Trinity Biotech immunoblot kits as the second tier test, however, may partially explain the reduced sensitivity, as suggested in at least one study of IgM immunoblot kit performance (Mogilyansky, et al., 2004). An alternative explanation for the lower sensitivity of the IgM immunoblot is a reduction of IgM activity due to prolonged sample storage or a prior freeze-thaw cycle (Jobe, et al., 2008; Petrakis, 1985). Against the latter hypothesis are the results of a recent report of 76 culture-positive patients with erythema migrans for whom two-tier testing was performed prospectively using the same immunoblot kit as was used in our study (Steere, et al., 2008). In that report the two-tier seroreactivity rate on acute phase sera was 30.3% (23/76), which was even lower than what we observed (38.3%).

In the current study not all of the sera from patients with Lyme arthritis were seropositive by two-tier testing; this unexpected result is likely attributable either to misclassification of some samples or to sample degradation as many of these samples were originally collected more than a decade earlier.

The sensitivity of the C6 ELISA was comparable to that of WCS ELISA-based two-tier testing at 90.0% and 80.0%, respectively ($p=0.50$), in the group of 20 acute phase sera from patients with early neurologic Lyme disease. The sensitivity of two-tier testing would not have changed in this group if IgM seropositivity were not restricted to an illness of 30 days duration (Centers for Disease Control and Prevention, 1995). Sera from patients with facial palsy accounted for the majority of both the acute (13 of 20) and convalescent (13 of 24) phase samples in the neurologic Lyme disease group. In selecting specimens from patients with extracutaneous manifestations of Lyme disease, sample selection bias may have

inadvertently been introduced through the prior use of serology to characterize such specimens by the sources. Although the sensitivity of an in-house C6 ELISA in early neurologic Lyme disease patients was 60% in one study from the United States (Bacon, et al., 2003), rates of at least 80% have been generally found in studies from both the United States (Liang, et al., 1999b; Peltomaa, et al., 2004; Steere, et al., 2008) and Europe (Skarpaas, et al., 2007; Tjernberg, et al., 2008) (see below).

Several recent reports have described performance evaluations of independently developed in-house C6 ELISA tests (Bacon, et al., 2003; Branda, et al., 2010; Gottner, et al., 2004). The sensitivities and specificities observed in these assays have varied from those determined in the present study for the commercial C6 ELISA kit. Sensitivities and/or specificities lower than those demonstrated with the commercial kit currently marketed in the United States may be due to several factors, including different assay chemistry, reagents or procedures, and in some cases, use of C6 peptide sequences other than the B31 sequence used in the kit. The present study also made use of significantly larger patient and control populations than those evaluated using in-house kits in previous reports, providing a more robust statistical basis for the findings reported herein. Conclusions relating to the performance of in-house C6 ELISA assays cannot be reliably extended to that of the commercial C6 ELISA kit evaluated in this study and vice-versa.

Substitution of the C6 ELISA result for the WCS ELISA in the first tier of the two-tier procedure yielded virtually identical results for both sensitivity and specificity, consistent with the observation that sensitivity of two-tier testing is limited by the immunoblot step. Use of the C6 ELISA as the first-tier test, however, was associated with 82 fewer false positive results among the 2208 (non-vaccinee) control sera, which translates to a 77.4% reduction in unnecessary immunoblot testing. These results should stimulate further cost effectiveness studies comparing WCS ELISA-based two-tier testing with that of C6 ELISA-based two-tier testing.

Overall, our results indicate that the C6 ELISA should be preferred over two-tier testing in patients with early infection, for example in patients with a skin lesion(s) for whom a clinical diagnosis is uncertain. Given the ease of performance, objectivity and greater simplicity of single-tier testing using an ELISA, the C6 ELISA may be preferred over two-tier testing in patients with early neurologic Lyme disease and other manifestations of early disseminated Lyme disease. An advantage of two-tier testing over the C6 ELISA is its higher specificity. The high specificity of two-tier testing found in this study, however, is not consistent with observations in clinical practice where over-reading of weak bands on the IgM immunoblot in particular has served to reduce specificity and stimulate interest in alternative testing strategies to avoid the IgM immunoblot entirely (Branda, et al., 2010; Seriburi et al., 2011; Weinstein, 2008). Another potential advantage of two-tier testing over C6 ELISA is that the former provides information on the presence of an expanded IgG response specifically, a serologic prerequisite for the diagnosis of late Lyme disease (Centers for Disease Control and Prevention, 1995; Wormser, et al., 2006). Thus, for the diagnosis of late Lyme disease, it may be advantageous to perform an IgG immunoblot to supplement the C6 or another ELISA.

Acknowledgments

This work was supported by Public Health Service Grant R43 AI051926 from the National Institute of Allergy and Infectious Diseases. It was also supported in part by the Intramural Research Program of the National Institute of Health (NIH). The authors thank Paul Visintainer, Diane Holmgren, Susan Bittker, Denise Cooper, Carol Carbonaro, Lois Zentmaier and Cecil Encarnacion for assistance.

The findings and conclusions in this article are those of the authors and do not necessarily represent the views of the Centers for Disease Control and Prevention or the NIH.

REFERENCES

- Aguero-Rosenfeld ME, Nowakowski J, McKenna DF, Carbonaro CA, Wormser GP. Serodiagnosis in early Lyme disease. *J Clin Microbiol.* 1993; 31:3090–5. [PubMed: 8308100]
- Aguero-Rosenfeld ME, Wang G, Schwartz I, Wormser GP. Diagnosis of Lyme borreliosis. *Clin Microbiol Rev.* 2005; 18:484–509. [PubMed: 16020686]
- Bacon RM, Biggerstaff BJ, Schrieffer ME, Gilmore RD Jr, Philipp MT, Steere AC, Wormser GP, Marques AR, Johnson BJ. Serodiagnosis of Lyme disease by kinetic enzyme-linked immunosorbent assay using recombinant VlsE1 or peptide antigens of *Borrelia burgdorferi* compared with 2-tiered testing using whole-cell lysates. *J Infect Dis.* 2003; 187:1187–99. [PubMed: 12695997]
- Bacon RM, Kugeler KJ, Mead PS. Surveillance for Lyme disease--United States, 1992–2006. *MMWR Surveill Summ.* 2008; 57:1–9. [PubMed: 18830214]
- Branda JA, Aguero-Rosenfeld ME, Ferraro MJ, Johnson BJ, Wormser GP, Steere AC. 2-tiered antibody testing for early and late Lyme disease using only an immunoglobulin G blot with the addition of a VlsE band as the second-tier test. *Clin Infect Dis.* 2010; 50:20–6. [PubMed: 19947857]
- Branda JA, Linskey K, Kim YA, Steere AC, Ferraro MJ. Two-tiered antibody testing for Lyme disease with use of 2 enzyme immunoassays, a whole-cell sonicate enzyme immunoassay followed by a VlsE C6 peptide enzyme immunoassay. *Clin Infect Dis.* 2011; 53:541–7. [PubMed: 21865190]
- Centers for Disease Control and Prevention. Recommendations for test performance and interpretation from the Second National Conference on Serologic Diagnosis of Lyme Disease. *MMWR Morb Mortal Wkly Rep.* 1995; 44:590–1. [PubMed: 7623762]
- Cinco M, Murgia R. Evaluation of the C6 enzyme-linked immunoadsorbent assay for the serodiagnosis of Lyme borreliosis in north-eastern Italy. *New Microbiol.* 2006; 29:139–41. [PubMed: 16841555]
- Craft JE, Fischer DK, Shimamoto GT, Steere AC. Antigens of *Borrelia burgdorferi* recognized during Lyme disease. Appearance of a new immunoglobulin M response and expansion of the immunoglobulin G response late in the illness. *J Clin Invest.* 1986; 78:934–9. [PubMed: 3531237]
- Embers ME, Jacobs MB, Johnson BJ, Philipp MT. Dominant epitopes of the C6 diagnostic peptide of *Borrelia burgdorferi* are largely inaccessible to antibody on the parent VlsE molecule. *Clin Vaccine Immunol.* 2007a; 14:931–6. [PubMed: 17567769]
- Embers ME, Wormser GP, Schwartz I, Martin DS, Philipp MT. *Borrelia burgdorferi* spirochetes that harbor only a portion of the lp28-1 plasmid elicit antibody responses detectable with the C6 test for Lyme disease. *Clin Vaccine Immunol.* 2007b; 14:90–3. [PubMed: 17108288]
- Gomes-Solecki MJ, Meirelles L, Glass J, Dattwyler RJ. Epitope length, genospecies dependency, and serum panel effect in the IR6 enzyme-linked immunosorbent assay for detection of antibodies to *Borrelia burgdorferi*. *Clin Vaccine Immunol.* 2007; 14:875–9. [PubMed: 17538122]
- Gottner G, Schulte-Spechtel U, Wilske B. Heterogeneity of the immunodominant surface protein VlsE among the three genospecies of *Borrelia burgdorferi* pathogenic for humans. *Int J Med Microbiol.* 2004; 293(Suppl 37):172–3. [PubMed: 15147003]
- Grodzicki RL, Steere AC. Comparison of immunoblotting and indirect enzyme-linked immunosorbent assay using different antigen preparations for diagnosing early Lyme disease. *J Infect Dis.* 1988; 157:790–7. [PubMed: 3279140]
- Hauser U, Wilske B. Enzyme-linked immunosorbent assays with recombinant internal flagellin fragments derived from different species of *Borrelia burgdorferi* sensu lato for the serodiagnosis of Lyme neuroborreliosis. *Med Microbiol Immunol.* 1997; 186:145–51. [PubMed: 9403843]

- Heikkilä T, Huppertz HI, Seppälä I, Sillanpää H, Saxen H, Lahdenne P. Recombinant or peptide antigens in the serology of Lyme arthritis in children. *J Infect Dis*. 2003; 187:1888–94. [PubMed: 12792865]
- Heikkilä T, Seppälä I, Saxen H, Panelius J, Peltomaa M, Julin T, Carlsson SA, Lahdenne P. Recombinant BBK32 protein in serodiagnosis of early and late Lyme borreliosis. *J Clin Microbiol*. 2002; 40:1174–80. [PubMed: 11923327]
- Jobe DA, Lovrich SD, Asp KE, Mathiason MA, Albrecht SE, Schell RF, Callister SM. Significantly improved accuracy of diagnosis of early Lyme disease by peptide enzyme-linked immunosorbent assay based on the borreliacidal antibody epitope of *Borrelia burgdorferi* OspC. *Clin Vaccine Immunol*. 2008; 15:981–5. [PubMed: 18417666]
- Lawrenz MB, Hardham JM, Owens RT, Nowakowski J, Steere AC, Wormser GP, Norris SJ. Human antibody responses to VlsE antigenic variation protein of *Borrelia burgdorferi*. *J Clin Microbiol*. 1999; 37:3997–4004. [PubMed: 10565921]
- Liang FT, Alvarez AL, Gu Y, Nowling JM, Ramamoorthy R, Philipp MT. An immunodominant conserved region within the variable domain of VlsE, the variable surface antigen of *Borrelia burgdorferi*. *J Immunol*. 1999a; 163:5566–73. [PubMed: 10553085]
- Liang FT, Steere AC, Marques AR, Johnson BJ, Miller JN, Philipp MT. Sensitive and specific serodiagnosis of Lyme disease by enzyme-linked immunosorbent assay with a peptide based on an immunodominant conserved region of *Borrelia burgdorferi* vlsE. *J Clin Microbiol*. 1999b; 37:3990–6. [PubMed: 10565920]
- Magnarelli LA, Ijdo JW, Padula SJ, Flavell RA, Fikrig E. Serologic diagnosis of Lyme borreliosis by using enzyme-linked immunosorbent assays with recombinant antigens. *J Clin Microbiol*. 2000; 38:1735–9. [PubMed: 10790090]
- Magnarelli LA, Lawrenz M, Norris SJ, Fikrig E. Comparative reactivity of human sera to recombinant VlsE and other *Borrelia burgdorferi* antigens in class-specific enzyme-linked immunosorbent assays for Lyme borreliosis. *J Med Microbiol*. 2002; 51:649–55. [PubMed: 12171295]
- Mathiesen MJ, Christiansen M, Hansen K, Holm A, Asbrink E, Theisen M. Peptide-based OspC enzyme-linked immunosorbent assay for serodiagnosis of Lyme borreliosis. *J Clin Microbiol*. 1998; 36:3474–9. [PubMed: 9817857]
- Mogilyansky E, Loa CC, Adelson ME, Mordechai E, Tilton RC. Comparison of Western immunoblotting and the C6 Lyme antibody test for laboratory detection of Lyme disease. *Clin Diagn Lab Immunol*. 2004; 11:924–9. [PubMed: 15358654]
- Norris SJ, Carter CJ, Howell JK, Barbour AG. Low-passage-associated proteins of *Borrelia burgdorferi* B31: characterization and molecular cloning of OspD, a surface-exposed, plasmid-encoded lipoprotein. *Infect Immun*. 1992; 60:4662–72. [PubMed: 1398980]
- Nowakowski J, Schwartz I, Liveris D, Wang G, Aguero-Rosenfeld ME, Girao G, McKenna D, Nadelman RB, Cavaliere LF, Wormser GP. Laboratory diagnostic techniques for patients with early Lyme disease associated with erythema migrans: a comparison of different techniques. *Clin Infect Dis*. 2001; 33:2023–7. [PubMed: 11700579]
- Padula SJ, Dias F, Sampieri A, Craven RB, Ryan RW. Use of recombinant OspC from *Borrelia burgdorferi* for serodiagnosis of early Lyme disease. *J Clin Microbiol*. 1994; 32:1733–8. [PubMed: 7929767]
- Panelius J, Lahdenne P, Saxen H, Heikkilä T, Seppälä I. Recombinant flagellin A proteins from *Borrelia burgdorferi* sensu stricto, *B. afzelii*, and *B. garinii* in serodiagnosis of Lyme borreliosis. *J Clin Microbiol*. 2001; 39:4013–9. [PubMed: 11682523]
- Peltomaa M, McHugh G, Steere AC. The VlsE (IR6) peptide ELISA in the serodiagnosis of Lyme facial paralysis. *Otol Neurotol*. 2004; 25:838–41. [PubMed: 15354020]
- Petrakis NL. Biologic banking in cohort studies, with special reference to blood. *Natl Cancer Inst Monogr*. 1985; 67:193–8. [PubMed: 4047147]
- Association of State and Territorial Public Health Laboratory Directors; Proceedings of the Second National Conference on Serologic Diagnosis of Lyme Disease; Dearborn, Michigan. 1995; p. 1-111.

- Rasiah C, Rauer S, Gassmann GS, Vogt A. Use of a hybrid protein consisting of the variable region of the *Borrelia burgdorferi* flagellin and part of the 83-kDa protein as antigen for serodiagnosis of Lyme disease. *J Clin Microbiol.* 1994; 32:1011–7. [PubMed: 8027303]
- Seriburi V, Ndukwe N, Chang Z, Cox ME, Wormser GP. High frequency of false positive IgM immunoblots for *Borrelia burgdorferi* in clinical practice. *Clin Microbiol Infect.* Dec 7.2011 2011. doi: 10.1111/j.1469-0691.2011.03749.x. [Epub ahead of print].
- Sillanpaa H, Lahdenne P, Sarvas H, Arnez M, Steere A, Peltomaa M, Seppala I. Immune responses to borrelial VlsE IR6 peptide variants. *Int J Med Microbiol.* 2007; 297:45–52. [PubMed: 17234451]
- Skarpaas T, Ljostad U, Soby M, Mygland A. Sensitivity and specificity of a commercial C6 peptide enzyme immunoassay in diagnosis of acute Lyme neuroborreliosis. *Eur J Clin Microbiol Infect Dis.* 2007; 26:675–7. [PubMed: 17605055]
- Smismans A, Goossens VJ, Nulens E, Bruggeman CA. Comparison of five different immunoassays for the detection of *Borrelia burgdorferi* IgM and IgG antibodies. *Clin Microbiol Infect.* 2006; 12:648–55. [PubMed: 16774561]
- Steere AC, Grodzicki RL, Kornblatt AN, Craft JE, Barbour AG, Burgdorfer W, Schmid GP, Johnson E, Malawista SE. The spirochetal etiology of Lyme disease. *N Engl J Med.* 1983; 308:733–40. [PubMed: 6828118]
- Steere AC, McHugh G, Damle N, Sikand VK. Prospective study of serologic tests for Lyme disease. *Clin Infect Dis.* 2008; 47:188–95. [PubMed: 18532885]
- Tjernberg I, Kruger G, Eliasson I. C6 peptide ELISA test in the serodiagnosis of Lyme borreliosis in Sweden. *Eur J Clin Microbiol Infect Dis.* 2007; 26:37–42. [PubMed: 17180348]
- Tjernberg I, Schon T, Ernerudh J, Wistedt AC, Forsberg P, Eliasson I. C6-peptide serology as diagnostic tool in neuroborreliosis. *Apmis.* 2008; 116:393–9. [PubMed: 18452429]
- Weinstein A. Editorial commentary: laboratory testing for Lyme disease: time for a change? *Clin Infect Dis.* 2008; 47:196–7. [PubMed: 18532894]
- Wormser GP, Dattwyler RJ, Shapiro ED, Halperin JJ, Steere AC, Klempner MS, Krause PJ, Bakken JS, Strle F, Stanek G, Bockenstedt L, Fish D, Dumler JS, Nadelman RB. The clinical assessment, treatment, and prevention of Lyme disease, human granulocytic anaplasmosis, and babesiosis: clinical practice guidelines by the Infectious Diseases Society of America. *Clin Infect Dis.* 2006; 43:1089–134. [PubMed: 17029130]
- Wormser GP, Liveris D, Hanincova K, Brisson D, Ludin S, Stracuzzi VJ, Embers ME, Philipp MT, Levin A, Aguero-Rosenfeld M, Schwartz I. Effect of *Borrelia burgdorferi* genotype on the sensitivity of C6 and 2-tier testing in North American patients with culture-confirmed Lyme disease. *Clin Infect Dis.* 2008a; 47:910–4. [PubMed: 18724824]
- Wormser GP, Nowakowski J, Nadelman RB, Visintainer P, Levin A, Aguero-Rosenfeld ME. Impact of clinical variables on *Borrelia burgdorferi*-specific antibody seropositivity in acute-phase sera from patients in North America with culture-confirmed early Lyme disease. *Clin Vaccine Immunol.* 2008b; 15:1519–22. [PubMed: 18716009]

Table 1

Sera selected for evaluation

Panel	n	Culture or PCR Positive ²	Days Post Onset of Symptoms							
			0-7	8- 14	15- 21	22- 30	31- 60	61- 90	> 90	NA ³
All EM	403	231	160	65	42	34	76	13	10	3
Single EM ¹	133	131	65	30	22	7	5	1	3	0
Multiple EM ¹	58	57	27	18	7	3	2	1	0	0
Acute EM	298	209	159	65	37	19	14	1	0	3
Convalescent EM	105	22	1	0	5	15	62	12	10	0
Early/Acute NL	20	4	3	2	3	2	1	5	4	0
Early/Convalescent NL	24	5	4	3	3	2	4	2	6	0
Late NL	8	0	0	0	0	0	0	0	6	2
Arthritis	114	10	0	0	1	0	6	4	50	53
Total	569	250	167	70	49	38	87	24	76	58

Abbreviations: EM, erythema migrans; NL, neurologic Lyme disease

¹ EM not characterized as single or multiple for 212 sera² Not all sera were from patients whose blood or skin lesion was tested by culture or PCR³ Time after onset data not available

Table 2
Reactivity in Blood Donor Populations

Method	Number [Percent] Negative ¹ (95% C.I.) ²		
	Non-Endemic	Endemic	Total
C6 ELISA	509 [99.2%] (98.0, 99.8%)	1311 [98.6%] ³ (97.9, 99.2%)	1820 [98.8%] (98.2, 99.3%)
Two-Tier (WCS ELISA/WB) ^a	512 [99.8%] (98.9, 100%)	1321 [99.4%] ³ (98.8, 99.7%)	1833 [99.5%] (99.1, 99.8%)
Two-Tier (C6 ELISA/WB) ^b	513 [100%] (99.3, 100%)	1320 [99.3%] (98.7, 99.7%)	1833 [99.5%] (99.1, 99.8%)
WCS ELISA	492 [95.9%] (93.8, 97.4%)	1283 [96.5%] (95.4, 97.5%)	1775 [96.4%] (95.4, 97.2%)
N (total # donors) =	513	1,329	1,842
P*	0.38	0.006	0.002
P[†]	0.13	0.004	<0.001

Abbreviations: WCS, whole cell sonicate

¹ % Negative calculated as (Total # samples – Total # Reactive Samples)/Total # Samples

² calculated by the Exact Method

³ one donor serum which was positive by both C6 ELISA and IgG immunoblot, but negative in WCS ELISA, was categorized as two-tier negative.

^a Two-tier testing with WCS ELISA as 1st step and IgG + IgM Western Blot as 2nd step

^b Two-tier testing with C6 ELISA as 1st step and IgG + IgM Western Blot as 2nd step

* For the comparison of C6 ELISA with WCS ELISA-based two-tier testing

[†] For the comparison of C6 ELISA with C6 ELISA-based two-tier testing

Table 3
Reactivity with Sera from Patients with Other Diseases or that may Contain Interfering Antibodies or Substances

Disease Condition	n	No. of Positive Sera			
		C6 ELISA Positive or Indeterminate	Two-Tier (WCS ELISA) ^a Positive	Two-Tier (C6 ELISA) ^b Positive	WCS ELISA Positive or Indeterminate
<i>Helicobacter pylori</i>	20	0	1	0	3
<i>Mycoplasma pneumoniae</i>	38	0	0	0	6
Cytomegalovirus	20	0	0	0	4
Epstein-Barr Virus	20	0	0	0	3
Human Immunodeficiency Virus	20	0	0	0	2
Hepatitis A	25	0	0	0	1
Hepatitis B	23	0	0	0	4
Hepatitis C	15	0	0	0	5
Influenza Vaccinated	25	0	0	0	0
Antinuclear Antibody +	20	0	0	0	1
Rapid Plasma Reagin + ¹	20	1	1	1	3
Lipemic	20	0	0	0	4
Icteric	20	0	0	0	0
Hemolyzed ¹	20	1	1	1	1
Systemic Lupus Erythematosus	20	0	0	0	0
Rheumatoid Arthritis	20	0	0	0	1
Rheumatoid Factor +	20	0	0	0	1
Total	366	2	3	2	39
Overall # [%] Negative² (95% C.I.)	-	364 [99.5%] (98.0, 99.9%)	363 [99.2%] (97.6, 99.8%)	364 [99.5%] (98.0, 99.9%)	327 [89.3%] (85.7, 92.3%)

Abbreviations: WCS, whole cell sonicate

¹The same serum sample was positive in all assays

²% Negative calculated as (Total # samples – Total # Reactive Samples)/Total # Samples

^aTwo-tier testing with WCS ELISA as 1st step and IgG + IgM Western Blot as 2nd step

^bTwo-tier testing with C6 ELISA as 1st step and IgG + IgM Western Blot as 2nd step

Table 4
Sensitivity by Symptom Category in Patients with Lyme Disease

Symptom Category	Number [Percent] Positive/Indeterminate [‡] (95% C.I.)					P [*]	P [†]
	C6 ELISA	WCS ELISA	Two-Tier WCS ELISA ^a	Two-Tier C6 ELISA ^b	n		
All Lyme patient sera	427 [75.0%] (71.3, 78.5)	465 [81.7%] (78.3, 84.8)	293 [51.5%] (47.3, 55.7)	288 [50.6%] (46.4, 55.7)	569	<0.001	<0.001
All Erythema Migrans	268 [66.5%] (61.7, 71.1)	302 [74.9%] (70.4, 79.1)	142 [35.2%] (30.6, 40.1)	139 [34.5%] (29.9, 39.4)	403	<0.001	<0.001
Single Erythema Migrans	77 [57.9%] (49.0, 66.4)	91 [68.4%] (59.8, 76.2)	36 [27.1%] (19.7, 35.5)	36 [27.1%] (19.7, 35.5)	133	<0.001	<0.001
Multiple Erythema Migrans	52 [89.7%] (78.8, 96.1)	55 [94.8%] (85.6, 98.9)	37 [63.8%] (50.1, 76.0)	36 [62.1%] (48.4, 74.5)	58	<0.001	<0.001
Acute Erythema Migrans	191 [64.1%] (58.4, 69.5)	214 [71.8%] (66.3, 76.8)	114 [38.3%] (32.7, 44.0)	112 [37.6%] (32.1, 43.4)	298	<0.001	<0.001
Convalescent Erythema Migrans	77 [73.3%] (63.8, 81.5)	88 [83.8%] (75.3, 90.3)	28 [26.7%] (18.5, 36.2)	27 [25.7%] (17.7, 35.2)	105	<0.001	<0.001
All Early Neurologic Lyme Disease	39 [88.6%] (75.4, 96.2)	43 [97.7%] (88.0, 99.9)	34 [77.3%] (62.2, 88.5)	33 [75.0%] (59.7, 86.8)	44	0.13	0.03
Early Neurologic Lyme Disease-Acute	18 [90.0%] (68.3, 98.8)	19 [95.0%] (75.1, 99.9)	16 [80.0%] (56.3, 94.3)	16 [80.0%] (56.3, 94.3)	20	0.5	0.5
Early Neurologic Lyme Disease-Convalescent	21 [87.5%] (67.6, 97.3)	24 [100%] (85.8, 100)	18 [75.0%] (53.3, 90.2)	17 [70.8%] (48.9, 87.4)	24	0.38	0.13
Late Neurologic Lyme Disease	8 [100%] (63.1, 100)	8 [100%] (63.1, 100)	8 [100%] (63.1, 100)	8 [100%] (63.1, 100)	8	>0.99	>0.99
Lyme Arthritis	112 [98.2%] (93.8, 99.8)	112 [98.2%] (93.8, 99.8)	109 [95.6%] (90.1, 98.6)	108 [94.7%] (88.9, 98.0)	114	0.38	0.13

Abbreviations: EM, erythema migrans; WCS, whole cell sonicate.

[‡] Indeterminate category refers only to ELISA results, which can be reported as positive, negative or indeterminate; two-tier testing is reported as either positive or negative. For purposes of calculation, the % positive values in the table in the ELISA categories sum both positive and indeterminate results. % Positive values were calculated as Total # Positive (or Positive + Indeterminate) Samples/Total # Samples. C6 ELISA and WCS ELISA had totals of 5 and 17 indeterminate results, respectively.

^a Two-tier testing with WCS ELISA as 1st step and IgG + IgM Western Blot as 2nd step; For two-tier testing, 319 serum samples were tested by Wampole WCS ELISA and 250 samples by Vidas WCS ELISA.

^b Two-tier testing with C6 ELISA as 1st step and IgG + IgM Western Blot as 2nd step

* p calculated for C6 ELISA vs. WCS ELISA-based two-tier testing

[†] p calculated for C6 ELISA vs. C6 ELISA-based two-tier testing