Meta-Analysis of Prevention of Surgical Site Infections following Incision Closure with Triclosan-Coated Sutures: Robustness to New Evidence

Frederic C. Daoud,¹ Charles E. Edmiston, Jr.,² and David Leaper³

Abstract

Background: A systematic literature review (SLR) and meta-analysis of surgical site infections (SSIs) after surgical incision closure with triclosan-coated sutures (TS) compared with non-antibacterial coated sutures (NTS) published previously by the authors suggested that fewer SSIs occurred in the TS study arm. However, the results were vulnerable to the removal of one key randomized controlled trial (RCT) because of insufficient data. Furthermore, recently published RCTs highlighted the need for an update of the SLR to challenge the robustness of results.

Methods: The protocol for the new SLR included more stringent tests of robustness than used initially and the meta-analysis was updated with the results of two new RCTs as well as the count of patients and SSIs by U.S. Centers for Disease Control and Prevention (CDC) incision class.

Results: The updated SLR included 15 RCTs with 4,800 patients. No publication bias was suggested in the analysis. The predominant effect estimated a relative risk of 0.67 (95% CI: 0.54–0.84, $p=0.00053$) with an overall lower frequency of SSI in the TS arm than in the NTS arm. Results were robust to sensitivity analysis.

Conclusions: The two additional peer-reviewed double-blind RCTs of this update confirmed the predominant effect found in the authors' previous meta-analysis and established the robustness of conclusions that were lacking previously. This SLR and meta-analysis showed that the use of triclosan antimicrobial sutures reduced the incidence of SSI after clean, clean-contaminated, and contaminated surgery. The two additional peer-reviewed double blind RCTs reinforced the evidence level of this SLR (CEBM level 1a).

TRICLOSAN (POLYCHLORO PHENOXY PHENOL) is an antimicrobial biocide that exhibits broad-spectrum activity against gram-positive and gram-negative bacteria.[1,2] In-vitro studies have shown that microbial pathogens associated commonly with surgical site infections (SSIs) are inhibited from adhering to the surface of triclosan-coated polyglactin 910 braided sutures.[3] Animal studies have also documented the efficacy of polyglactin 910 with triclosan in-vivo, and further in vitro and in vivo studies have shown the efficacy of triclosan when incorporated in polydioxanone sutures.[4,5] The triclosan dosage producing the intended biocidal effect in these absorbable sutures is a maximum of 2,360 mcg/m in both polydioxanone and poliglecaprone 25; and 472 mcg/m in polyglactin 910.[6–8]

Tests and clinical trials have also shown that surgeons cannot differentiate the presence or absence of triclosan in

braided or monofilament sutures, making it possible to design randomized double-blind trials for clinical comparison.[9,10] Several randomized, controlled clinical trials (RCTs) have compared the frequency of SSIs after closure of surgical incisions with triclosan-coated sutures (TS) compared with non-triclosan sutures (NTS) in different clinical settings, and surgical procedures, to ensure the comparability of study arms, diagnostic criteria, suture materials, patient demographics, background diseases, and surgical operations. These clinical trials are the subject of this updated systematic literature review (SLR) and meta-analysis.

An earlier SLR and meta-analysis, published by the authors of this current article, was conducted according to a predefined written protocol, and identified 13 peer-reviewed eligible RCTs that compared TS vs. NTS in different clinical settings (different surgical operations, suture materials,

¹Unit of Biostatistics and Bioinformatics, Medextens, Paris, France.

²Department of Surgery, Medical College of Wisconsin, Milwaukee, Wisconsin.

³School of Applied Sciences, Huddersfield University, Queensgate, Huddersfield, United Kingdom.

underlying diseases, concentration of incision contamination, and methods to assess SSI occurrence) [9,11–23]. The result was a lower risk of SSI in the TS arm with a point estimate relative risk (RR) of 0.69, and a 95% confidence interval (CI) of 0.52–0.92 ($p < 0.011$). Sensitivity analysis showed that this result was vulnerable to the removal of one trial, resulting in a borderline p -value, suggesting that the conclusions of the meta-analysis depended on that trial alone [16]. This article reports an update of the prior SLR and meta-analysis with inclusion of new clinical trials as well as additional information about the study methods and patient characteristics communicated by the authors of the previously reported eligible trials. The primary objective was to determine if incision closure presented the same risk of SSI when triclosan was present or absent on the surface of the sutures. The secondary objectives were to assess potential bias or confounding factors that could invalidate the primary conclusion and to determine the extent to which the primary conclusion could be generalized to the various types of patients included in the pooled RCTs.

Materials and Methods

The SLR protocol of the first meta-analysis reported previously was developed according to PRISMA recommendations and is repeated in the current communication with a number of differences described below.[24]

Study selection

Embase/Medline, the Cochrane database (Central Register of Controlled Trials [CENTRAL], Cochrane Database of Systematic Reviews [CDRS], Health Economic Evaluations Database [HEED], and Database of Health Technology Assessments [HTA]) as well as www.clinicaltrials.gov were searched on July 30, 2013 using their own search engines with a syntax that combined the following keywords: Triclosan AND suture* AND (random* OR RCT). Clinicaltrials.gov was searched to identify potentially completed relevant RCTs. The “*” character at the end of keyword radicals indicated to search engines to include any character sequence that ended the keyword. This syntax differed from the previous meta-analysis by not searching broader keyword combinations that had proved to retrieve only non-randomized studies or irrelevant publications.

Study eligibility

Study eligibility criteria were the same as reported previously [11]. The results of the clinicaltrials.gov search and the reference lists of all relevant publications were meticulously searched to “catch back” any potentially eligible RCTs that might have been missed by the electronic study extraction and filtering process.

Data extraction

Data extraction was performed according to the same rules as previously described [11]. The four data items extracted from each study were the number of patients in the TS group (N_{TS}) with the number of patients presenting an SSI in the TS arm (n_{TS}), and their counterparts in the NTS arm (N_{NTS} , n_{NTS}).

Whenever a study reported outcomes on both an intention-to-treat (ITT) basis and a per-protocol (PP) basis, ITT results

were used. Whenever a study reported outcomes PP only, but indicated the number of randomized patients together with demographics on an ITT basis, the PP sample size was replaced by ITT sample size.

In the case of RCTs reporting outcomes of the closure of several incisions per patient, only outcomes of the incision defined with the primary endpoint were included. For example, when the primary endpoint was the occurrence of SSI in the closure of a leg incision, after venous graft harvesting for coronary artery bypass surgery (CABG), outcomes of chest closure in the same patient were reported as a secondary endpoint. Trials where the same patient was systematically used in the active arm and the control arm were excluded as these trials were not “head-to-head” comparisons and did not provide independence of groups.

Compared with the previous protocol, the extraction also recorded the blinding method (double-blind, single-blind, open-label), the SSI diagnostic method used (U.S. Center for Disease Control and Prevention [CDC] criteria or other) as well as the number of patients with and without SSI per treatment arm broken down by class of incision contamination (class I/clean, class II/clean-contaminated, class III/contaminated, class IV/dirty-infected) [25,26]. Whenever these data were not reported in the publications and could not be deducted from the study context, the reviewers contacted the corresponding authors of the publications to obtain the missing information.

Quality of evidence

As performed with the RCTs included in the previous SLR, newly published studies were tested against the eligibility criteria defined in this SLR, against the concentration of evidence criteria proposed by the Centre for Evidence-Based Medicine (CEBM) at the University of Oxford, and the Cochrane criteria for quality and low risk of bias [27,28]. Full publications of all new RCTs were acquired and reviewed.

Statistical analysis

The data extracted from eligible RCTs was analyzed using meta-analytic techniques as previously, and the same CMA software (Comprehensive Meta-Analysis v2.2.027, Englewood, NJ) and reproducibility of calculations was checked by running an auditable script in STATA 12 (StataCorp LP, College Station, TX, USA). A rigorous 13-step analytical strategy was formulated (Fig. 1) and implemented in order to include the additional questions to be compared to the meta-analysis. Subgroup meta-analysis was considered to be powered sufficiently if it pooled a minimum of five estimates. Similarly, the comparison of subgroups was considered powered sufficiently if all compared subgroups pooled a minimum of five estimates. Subgroups that pooled four estimates were considered only to describe the available evidence and to assess its internal validity, but not for generalization of conclusions.

Step 1. Calculated the risk ratio (RR) as the measure of effect between treatment arms in each RCT. $RR = (n_{TS}/N_{TS}) / (n_{NTS}/N_{NTS})$, where in each treatment arm, n is the number of patients with SSI and N is the number patients treated.

Step 2. Tested the null hypothesis (H_0) of absence of publication bias. H_0 was rejected if the Egger intercept test

FIG. 1. Flowchart of the meta-analysis.

(significant if $p < 0.05$) detected an asymmetric inverted funnel shape in the funnel plot [29–31].

Step 3. Tested the null hypothesis (H_0) of no heterogeneity between trials, meaning that the RRs of all RCTs were drawn from a common population with the same true mean. H_0 was rejected if the Cochran Q-test was significant (if $p < 0.05$) [32]. The I^2 estimated the percentage of variability of study RRs that could not be explained by random error only [33,34]. Given differences between trials in terms of methods and clinical settings (blinding, diagnostic criteria, compared suture materials, types of operations, CDC incision classes, and site/organ location), the algorithm temporarily maintained the assumption of heterogeneity between study populations until all these factors would be ruled out further in the analysis.

Step 4. Tested the effect between treatment arms across studies by calculating the random effects pooled RR (weighted

average) [35,36]. This estimated the predominant effect among the populations from which the RCTs were drawn but not the true mean RR of any study population in particular. Under the null hypothesis (H_0), the pooled RR is one, meaning the same frequency of SSIs between treatment arms. H_0 would be rejected in favor of the alternative that the predominant effect was significant if the pooled RR's 95% CI included one.

Step 5. Sensitivity analysis of the robustness of meta-analysis was performed on all eligible RCTs. This consisted of repeating steps two, three, and four iteratively after removing one RCT at a time, then two RCTs at a time (any pair of RCTs), then three RCTs at a time, and so on, until 95% CI of the pooled RR reached one, thus non-significance. This process enabled identification of the RCTs upon which depended significance.

Step 6. Assessed the potential bias in the predominant effect potentially caused by differences in blinding method

between RCTs (double-blind, single-blind, open-label). The pooled random effects RR and 95% CIs of each blinding subgroup with more than four individual RRs were calculated. The null hypothesis (H_0), that subgroups shared the same population mean RR, was rejected if $p < 0.05$, in favor of the alternative where subgroups had heterogeneous population means. Sensitivity analysis of robustness was planned similarly to step five in the double-blinded subgroup if the subgroup RR was significant and pooled more than four individual RRs.

Step 7. Assessed the potential bias in the predominant effect potentially caused by differences in SSI diagnostic method between RCTs (CDC criteria, other methods). The process was similar to step six but no sensitivity analysis was planned.

Step 8. Assessed potential bias in the predominant effect caused by differences in compared suture materials between RCTs (polyglactin 910, polydioxanone, poliglecaprone, combinations of various materials, silk sutures) as performed [25]. The process was similar to step six with sensitivity analysis.

Step 9. Assessed potential bias in the predominant effect caused by differences in CDC incision class reported either between RCTs or class of contamination. The process was similar to step six with sensitivity analysis.

Step 10. Assessment of the potential bias in the predominant effect caused by differences in incision depth or site, between RCTs or within RCTs, was planned if reviewers were able to extract that information from publications. Incisional and organ/space surgical site definitions were those proposed by the CDC [37]. The process was similar to step six with sensitivity analysis.

Step 11. Assessed potential bias in the predominant effect caused by differences in operation type as reported by RCT authors. The process was similar to step six with sensitivity analysis.

Step 12. Assessed the potentially confounding relationship between the RR and the SSI incidence rate (IR) in each RCT. Incidence rate is an outcome computed with the same inputs as RR but examines the frequency of SSI as an average, ignoring differences between treatment arms: $IR = (n_{TS} + n_{NTS}) / (N_{TS} + N_{NTS})$. In order to enable a linear comparison of RR and IR, the $\text{Log}(RR)$ of each RCT was plotted against the corresponding RCT's IR in a meta-regression. The null hypothesis (H_0) that RR was independent of IR was tested by testing the slope of $\text{Log}(RR)$ as a function of IR. The alternative hypothesis that RR did vary with IR was accepted if that slope was (i.e. $p < 0.05$) different from 0.

Step 13. Conclusion: If previous steps showed that the overall random-effects pooled RR was robust, and that RR did not dependent on the IR, four conclusion options were possible:

- i. If the pooled RR of the double-blind RCTs subgroup was robust AND if no heterogeneity was found among all RRs, AND all tested potential bias/heterogeneity factors were ruled out: The conclusion was

a similar treatment effect across study populations probably caused by triclosan.

- ii. If the pooled RR of the double-blind RCTs subgroup was robust BUT heterogeneity was found among all RRs, OR any tested potential bias/heterogeneity factors was not ruled out: The conclusion was a significant but parameter treatment effect across study populations probably caused by triclosan.
- iii. If the pooled RR of the double-blind RCTs subgroup was not robust BUT no heterogeneity was found among all RRs, AND all tested potential bias/heterogeneity factors were ruled out: The conclusion was a similar treatment effect across study populations but an uncertain role of triclosan in the effect.
- iv. If the pooled RR of the double-blind RCTs subgroup was robust AND if heterogeneity was found among all RRs, OR if any tested potential bias/heterogeneity factors was not ruled out: The conclusion was a parameter treatment effect across study populations and an uncertain role of triclosan in the effect.

Results

Seventy-six references were identified and one additional reference was caught back [40]. Forty-three duplicate references were removed, 34 references were screened and 15 were confirmed eligible RCTs (Fig. 2). The eligible trials included the 13 RCTs from the previous meta-analysis and two double-blind prospective RCTs published later [11,38,39]. The new RCTs met all eligibility criteria of this SLR protocol as well as CEBM and Cochrane criteria for quality and low risk of bias [27,28]. The 19 ineligible references included four abstracts of non-peer reviewed RCTs [40–43]. The eligibility of those four abstracts could not be determined because of insufficient methods descriptions or incomplete reporting of results reporting.

The SSI diagnostic method used and CDC incision class were reported in four RCTs and CDC incision class could be deducted as clean from the type of surgery in two others. Ten authors were contacted to obtain or confirm the count of patients with and without SSI by CDC class and the confirmation of the SSI diagnostic method when these were not specified in the publication. Seven authors responded, two of whom also added that SSI had been diagnosed using CDC criteria. The lead author of one trial on colorectal surgery, who did not use the CDC criteria, responded that all cases were clean or clean-contaminated so the reviewers assumed that all cases were clean-contaminated, as in other colorectal trials [17]. Three authors did not respond. Altogether outcomes were available by class of incision contamination in 12 RCTs. Table 1 summarizes the characteristics of the eligible RCTs obtained from publications and authors directly. Overall, the 15 RCTs enabled the extraction of outcomes in 4,800 patients; (TS: $n = 2,323$ and NTS: $n = 2,477$), with head-to-head comparisons of a single incision per patient. There were 453 patients who had an SSI (TS: $n = 180$ and NTS: $n = 273$). The 13-step analysis revealed the following:

The 15 trial RRs and their 95% CIs were calculated with data reported (Table 1) on an intention to treat (ITT) basis in nine RCTs, with data reported per protocol (PP) in four RCTs, and with the number of SSIs per protocol and the number of randomized patients according to ITT methods in

FIG. 2. PRISMA diagram of study extraction and eligibility selection.

two RCTs (**Step 1**). The funnel plot (Fig. 3) and Egger intercept test (intercept = -0.746 , standard error (SE) = 0.647 , t -value = 1.153 , degrees of freedom (df) = 13 , 2-tailed $p = 0.269$) suggested no publication bias (**Step 2**). The Q-test (**Step 3**) showed no heterogeneity between the RRs of the 15 trials ($Q = 18.572$, $df = 14$, $p = 0.182$) and the I^2 estimated 24.6% of variability because of heterogeneity. Given the heterogeneity between the 15 RCTs in terms of incidence Rates (IR) of SSI ($Q = 123.97$, $df = 14$, $p < 0.00001$, $I^2 = 88.7\%$), the pooled RR continued to be estimated using a random-effects model until ruling out all identified potential confounders and factors of bias.

The overall pooled RR was 0.67 , 95% CI 0.54 – 0.84 , $p = 0.00053$ (**Step 4**). The Forrest plot (Fig. 4) showed the predominant effect to be a lower frequency of SSIs in the TS arm than in the NTS arm. Sensitivity analysis (**Step 5**) (Table 2) showed no indicator of publication bias and the predominant effect was robust to the iterative removal of any single RCT and any pair of RCTs. When removing three RCTs at a time, the only combination to result in the borderline $p = 0.045$ found during the previous meta-analysis was the joint removal of the Thimour-Bergström, Galal, and Justinger trials, with no indicator of associated publication bias [38,26,39].

Pooled RR by blinding subgroup (**Step 6**), showed that nine RCTs were double-blind, four open-label, and two assessor single-blind (Fig. 5). Because of the low number of RRs in the open-label and assessor single-blind subgroups, the mixed-effects analysis ($Q = 2.688$, $df = 2$, $p = 0.261$) was

underpowered to test the heterogeneity of the subgroups and to rule out potential bias because of the blinding method. However, the 95% CI of the pooled RR in the open-label subgroup included one, so open-label RCTs did not cause an overestimation of the predominant effect compared to double-blind RCTs. The assessor-blind subgroup consisted of the two RCTs comparing TS to silk sutures had a lower RR than the other subgroups, but their impact was small because of those RCTs' low relative weights of 1.88 and 0.63 respectively (Fig. 4) [14,15]. The nine double-blind RCTs were the core subgroup that drove the overall pooled RR with no indicator of publication bias, a pooled RR of 0.65 , 95% CI 0.51 – 0.82 ($p < 0.00024$) (Table 3), and conclusions robust to sensitivity analysis with the removal of any single RCT and any pair of RCTs. That subgroup was vulnerable to the joint removal of three RCTs ($p = 0.107$) [16,38,39].

Pooled RR by diagnostic method subgroup (**Step 7**) demonstrated that 10 RCTs had used CDC criteria and five had used other methods. There were thus enough RRs per subgroup to test their heterogeneity with a mixed-effects analysis ($Q = 0.682$, $df = 1$, $p = 0.409$) but no heterogeneity was found. Diagnostic method was thus ruled out as a cause of bias and heterogeneity in the overall pooled RR. It should be noted that, the 95% CI of the pooled RR in the "other methods" subgroup included one, thus did not cause an overestimation of the predominant effect compared to RCTs using CDC criteria.

Pooled RR by suture material (**Step 8**) showed five suture-material subgroups: Polyglactin with versus without triclosan

TABLE 1. CHARACTERISTICS OF ELIGIBLE RCTs

<i>Study</i>	<i>Study design & population</i>	<i>Primary surgical procedures & population</i>	<i>CDCSurgical incision status</i>	<i>Compared sutures - Sutured incision depth -SSI definition</i>	<i>Reported Results Numbers used in meta-analysis</i>
Ford 2005, USA ⁹	prospective, randomized, controlled, open-label with respect to SSI, comparative, single-center study 30-d follow-up	General surgery procedures Patients: Children (151 randomized, 147 treated)	Not reported in publication but details disclosed by the author to reviewers.	-triclosan polyglactin 910, n=98 -polyglactin 910, n=49 SSI secondary endpoint. Defined as observed redness 3–5 mm from incision margins, edema, purulent discharge, pain, and increased skin temperature. Culture not required. SSI occurrence reported up to 80 d.	SSI reported cases PP: TS: 3/98 NTS: 0/49 p value not reported SSI meta-analysis PP
Rozzelle 2008, USA ¹²	Single-center, prospective, randomized, double-blinded, and placebo-controlled study 6-mo follow-up Methodological issue: Data reported in relation to number of procedures	CSF shunt implantation or revision surgery Patients: 61 children and adults, median age 6.3y, 84 randomized shunt procedures	Not reported in publication. Reviewers' assumption: CSF shunt surgery, all CDC class I	- triclosan polyglactin 910, n=46 procedures - polyglactin 910, n=38 procedures Used in: -galea closure & - fascia closure SSI primary endpoint. Defined as culture proven shunt infection from CSF or explanted shunt 84 shunt procedures: 40 implants & 44 reinterventions	SSI reported cases ITT: TS: 2/46 NTS: 8/38 Reported p=0.038 SSI meta-analysis ITT
Mingmalairak 2009, Thailand ¹³	Single-center, prospective, randomized, controlled, double blind, comparative trial 12-mo follow-up	Appendectomy for acute appendicitis and ruptured appendix Patients: 100, age 15–60 y randomized	Not reported in publication but details disclosed by the author to reviewers.	- triclosan polyglactin 910, n=50 - polyglactin 910, n=50 Used in: -closure not described Per publication: SSI diagnostic criteria not described in publication. Culture proof ad hoc. Author confirmed to reviewers using CDC diagnostic criteria	SSI reported cases ITT: TS: 5/50 NTS: 4/50 Reported p=0.727 SSI meta-analysis ITT

(continued)

TABLE 1. (CONTINUED)

<i>Study</i>	<i>Study design & population</i>	<i>Primary surgical procedures & population</i>	<i>CDCSurgical incision status</i>	<i>Compared sutures - Sutured incision depth -SSI definition</i>	<i>Reported Results Numbers used in meta-analysis</i>
Zhuang 2009, China ¹⁴	Single-center, prospective, randomized, assessor-blinded, comparative trial 12 to 14-mo follow-up	Laparotomy (various abdominal operations) Patients: 450 adults randomized	Not reported in publication. No response to reviewers' request.	-triclosan polyglactin 910, n = 150 -polydioxanone, n = 150 -silk, n = 150 Used in: -abdominal wall closure: anadesma, muscle, and peritoneum. -skin excluded in triclosan, polyglactin, and polydioxanone groups. -Translation does not indicate differences in skin closure between group or differences in counting superficial incisional infections SSI secondary endpoint. Diagnostic criteria not described.	SSI reported cases ITT: TS polyglactin 910: 0/150 NTS polydioxanone, 3/150 NTS silk: 15/150 Reported lowest SSI rate in TS: p < 0.01 SSI meta-analysis ITT
Zhang 2011, China ¹⁵	Multi-center, prospective, randomized, assessor-blinded, comparative trial 90-d follow-up	Modified radical mastectomy Patients: 101 adults randomized	Publication reports all CDC class I	-triclosan polyglactin 910, n = 50 -silk, n = 50 Used in: -skin closure SSI secondary endpoint. CDC diagnostic criteria & ASESIS score	CDC SSI reported cases PP: TS: 2/46 NTS: 5/43 SSI meta-analysis ITT
Galal 2011, Egypt ¹⁶	Single-center, prospective, randomized, double-blinded, comparative trial 12-mo follow-up: SSI within 30 d or 1 y if prosthesis	Various operations Patients: 450 adults randomized	Details reported in publication	- triclosan polyglactin 910, n = 230 - polyglactin 910, n = 220 Used in: -broad range of surgical operations SSI primary endpoint. CDC diagnostic criteria.	SSI reported cases ITT: TS: 17/230 NTS: 33/220 Reported p = 0.011 SSI meta-analysis ITT
Rasic 2011, Croatia ¹⁷	Single-center, prospective, randomized, non-blinded, comparative trial Follow-up (unspecified, apparently, minimum 14 d)	Colorectal surgery Patients: 184 adults randomized	Not reported in publication. Author confirmed to reviewers evaluated according to incision presentation but all patients were clean to Reviewers assumed all colorectal operations CDC class II	- triclosan polyglactin 910, n = 91 - polyglactin 910, n = 93 Used in: -midline abdominal wall incision closed with continuous single-layer mass technique (peritoneum, muscle, and fascia) -skin closure with polyamide suture SSI part of primary endpoint. Diagnostic criteria not specified.	SSI reported cases ITT: TS: 4/91 NTS: 12/93 Report p = 0.035 SSI meta-analysis ITT

(continued)

TABLE 1. (CONTINUED)

<i>Study</i>	<i>Study design & population</i>	<i>Primary surgical procedures & population</i>	<i>CDC Surgical incision status</i>	<i>Compared sutures - Sutured incision depth - SSI definition</i>	<i>Reported Results Numbers used in meta-analysis</i>
Baracs 2011, ¹⁹ Hungary	Multi-center, prospective, randomized, comparative trial Blinding not specified 30-d follow-up	Colorectal surgery Patients: 485 adults randomized	Not reported in publication. Author confirmed to reviewers: All CDC class II.	-triclosan polydioxanone, n = 188 -polydioxanone, n = 197 Used in: - abdominal fascia closure with polydioxanone (TS or NTS depending on randomization) -skin closure with triclosan polyglactone 25 in both arms SSI primary endpoint. Author confirmed to reviewers using CDC diagnostic criteria.	SSI reported cases ITT: TS: 23/188 NTS: 24/197 p value reported as non-significant SSI meta-analysis ITT
Williams 2011, ¹⁸ UK	Single-center, prospective, randomized, double-blinded, comparative trial 6-wk follow-up duration	Skin closure after breast cancer surgery Patients: 150 female adults randomized	Not reported in publication. Author confirmed to reviewers: All CDC class I	-triclosan polyglactin 910 & triclosan poliglecaprone 25, n = 75 - polyglactin 910 & poliglecaprone 25 n = 75 Used in: - deep tissue closure with polyglactin 910 -skin closure with poliglecaprone 25 SSI primary endpoint. CDC diagnostic criteria.	CDC SSI cases PP: TS: 10/66 NTS: 14/61 p value reported as non-significant SSI meta-analysis ITT
Turtiainen 2012, ²⁰ Finland	Multi-center, prospective, randomized, double-blinded, comparative trial 1-mo follow-up duration	Incision closure after lower limb revascularization surgery Patients: 276 adults randomized	Not reported in publication. No response to reviewers' request.	-triclosan polyglactin 910 & triclosan poliglecaprone 25, n = 139 - polyglactin 910 & poliglecaprone 25 n = 137 Used in: - subcutaneous closure with polyglactin 910 -intracutaneous closure with poliglecaprone 25 SSI primary endpoint. CDC diagnostic criteria. Culture-proof ad hoc	SSI reported ITT: TS: 31/139 NTS: 30/137 Reported p = 0.94 SSI meta-analysis ITT
Seim 2012, ²¹ Norway	Single-center, prospective, randomized, non-blinded, comparative trial 4-wk follow-up duration	Leg incision after vein harvesting for CABG Patients: 328 adults randomized	Not reported in publication. No response to reviewers' request. Reviewers' assumption CABG all CDC class I	- triclosan polyglactin 910, n = 160 - polyglactin 910, n = 163 Used in: -leg incision closure without further de-tails Leg SSI primary endpoint. Diagnostic criteria not specified	SSI reported PP: TS: 16/160 NTS: 17/163 p value reported as NS SSI meta-analysis PP

(continued)

TABLE 1. (CONTINUED)

<i>Study</i>	<i>Study design & population</i>	<i>Primary surgical procedures & population</i>	<i>CDCSurgical incision status</i>	<i>Compared sutures - Sutured incision depth -SSI definition</i>	<i>Reported Results Numbers used in meta-analysis</i>
Isik 2012, Turkey ²²	Single-center, prospective, randomized, double-blinded, comparative trial 30-d follow-up duration	Sternal incision in cardiac surgery. Patients: 510 adults randomized	Not reported in publication. Author confirmed to reviewers: All CDC class I	- tricosan polyglactin 910, n = 170 - polyglactin 910, n = 340 Used in: -sternal incision closure without further detail Sternal SSI 1mary endpoint. CDC diagnostic criteria.	SSI reported sternal incision ITT: TS: 4/170 NTS: 12/340 Reported p = 0.516 SSI sternal meta-analysis ITT
Nakamura 2013, Japan ²³	Single-center, prospective, randomized, double-blind, comparative trial 30-d follow-up duration	Colorectal surgery Patients: 410 adults randomized	Reported in publication.	- tricosan polyglactin 910, n = 206 - polyglactin 910, n = 204 Used in: -colorectal incision closure without further detail SSI primary endpoint. CDC diagnostic criteria.	SSI reported ITT: TS: 9/206 NTS: 19/204 Report p = 0.047 SSI meta-analysis ITT
Thimour-Bergstöm 2013, Sweden ³⁸	Single-center, prospective, randomized, double-blind, comparative trial 60-d follow-up duration	Leg incision after vein harvesting for CABG Patients: 392 adults randomized	Not reported in publication. Author confirmed to reviewers: All CDC class I	-tricosan polyglactin 910 & triclosan poliglecaprone 25, n = 184 - polyglactin 910 & poliglecaprone 25 n = 190 Used in: - subcutaneous closure with polyglactin 910 -intracutaneous closure with poliglecaprone 25 SSI primary endpoint. CDC criteria. Bacterial culture proof in some patients.	SSI reported PP: TS :23/184 NTS: 38/190 Reported p = 0.0497 SSI meta-analysis PP
Justinger 2013, Germany ³⁹	Single-center, prospective, randomized, double-blind, comparative trial 2-wk follow-up duration Only the primary incision was evaluated.	Laparotomy (various abdominal operations) Patients: 856 adults randomized	Reported in publication	-tricosan polydioxanone, n = 485 -polydioxanone, n = 371 Used in: -abdominal fascia closure. CDC diagnostic criteria.	SSI reported ITT: TS: 31/485 NTS: 42/371 reported p < 0.05 SSI meta-analysis ITT

RCT = randomized controlled trial; TS = tricosan-coated sutures; NTS = non-tricosan-coated sutures; n = number of patients; SSI = surgical site infections; ITT = intention-to-treat; PP = per-protocol; CDC = U.S. Centers for Disease Control and Prevention; CSF = cerebrospinal fluid; CABG = coronary artery bypass grafting.

FIG. 3. Funnel plot of the risk ratios of eligible RCTs.

in eight RCTs, polyglactin and poliglecaprone with versus without triclosan in three RCTs, polydioxanone with versus without triclosan in two RCTs, triclosan-polyglactin versus polydioxanone or versus silk in one RCT, and triclosan-polyglactin versus silk in one RCT. Because of the low number of RRs in all but the first subgroup, the mixed-effects analysis ($Q=4.220$, $df=4$, $p=0.377$) was underpowered to test the heterogeneity of the subgroups and to rule out potential bias or heterogeneity in the overall pooled RR because of comparisons of different suture materials. The 11 RCTs of the two first subgroups studying polyglactin, with versus without triclosan, had no indicator of publication bias (Egger test: $p=0.66$), yielded a pooled RR=0.68, 95% CI [0.52, 0.88], $p=0.003$ and sensitivity analysis showed robustness to the iterative removal of any single RCT. Those results were vulnerable to the joint removal of the Thimour-Bergström and Galal trials ($p=0.066$) [16,38].

Pooled RR by incision contamination (Step 9) showed five subgroups (Fig. 6). Given only clean and clean-contaminated

incisions had at least five RRs to pool, the mixed-effects analysis ($Q=6.560$, $df=4$, $p=0.161$) was underpowered to test the heterogeneity of the five subgroups and to rule out potential bias or heterogeneity in the overall pooled RR because of pooling studies with different incision contamination. However, the 95% CI of the pooled RR in clean, clean-contaminated, and contaminated incisions (borderline number of observations with four RRs to pool) were all significant with p values ranging from 0.001 in clean incisions, 0.010 in clean-contaminated incision and 0.026 in contaminated incisions. The clean incision subgroup presented no indicator of publication bias (Egger test: $p=0.158$) and sensitivity analysis showed conclusions to be robust to the removal of any single RCT, and borderline vulnerable to the joint removal of the Thimour-Bergström and Justinger trials ($p=0.046$) [38,39]. The clean-contaminated and contaminated subgroups were not robust to the iterative removal of one RCT at a time.

Pooled RR by incision depth and organ/space infection (Step 10) could not be conducted because of the inability to extract the data. Pooled RR by operation type (Step 11) was performed using all published data and resulted in 13 subgroups but only 12 could be analyzed because no SSI was reported in either arm in the abdominal vascular surgery subgroup (Fig. 7). Appendectomy was grouped together with colorectal surgery and the resulting subgroup was the only one pooling at least five RRs, with a pooled RR of 0.698, 95% CI 0.476–1.002 ($p=0.0536$). Because of the low number of RRs in all but that subgroup, the mixed-effects analysis ($Q=10.233$, $df=11$, $p=0.510$) was underpowered to test the heterogeneity of the subgroups and to rule out potential bias or heterogeneity in the overall pooled RR because of pooling studies with different operation types.

The heterogeneity of average incidence rate (IR) of SSI in treatment arms described in step three was visible on the Forrest plot (Fig. 8). The meta-regression was conducted

FIG. 4. Forest plot of the risk ratios—Overall.

TABLE 2. SENSITIVITY ANALYSIS: META-ANALYSIS OF ALL ELIGIBLE RCTS

Number of RCTs removed	RCT removed	RR	95% CI LL	95% CI UL	p value	Publication bias		N SSI/N TS	N SSI/N NTS	IR N SSI/N total & % SSI
						Egger p value				
0	None	0.671	0.536	0.841	0.001	0.269		180/2323	273/2477	450/4653, 9.5% [7.2%–12.5%]
1	Zhuang	0.685	0.555	0.846	0.000	0.578		180/2173	258/2177	437/4290, 9.8% [7.5%–12.6%]
1	Rozelle	0.691	0.558	0.857	0.001	0.476		178/2277	265/2439	438/4350, 9.7% [7.4%–12.6%]
1	Rasic	0.690	0.551	0.865	0.001	0.388		176/2232	261/2384	425/4390, 9.2% [6.8%–12.3%]
1	Zhang	0.677	0.537	0.854	0.001	0.353		178/2272	268/2427	446/4699, 9.1% [6.8%–12.2%]
1	Nakamura	0.688	0.543	0.871	0.002	0.317		171/2117	254/2273	380/3944, 9.0% [6.6%–12.2%]
1	Galal	0.696	0.549	0.884	0.003	0.232		163/2093	240/2257	437/4616, 9.0% [6.7%–12.1%]
1	Justinger	0.686	0.534	0.883	0.003	0.155		149/1838	231/2106	444/4700, 9.0% [6.7%–12.0%]
1	Thimour-Bergström	0.673	0.521	0.870	0.002	0.248		157/2139	235/2287	420/4477, 8.9% [6.5%–12.0%]
1	Isik	0.669	0.526	0.849	0.001	0.278		176/2153	261/2137	403/4350, 8.8% [6.4%–11.9%]
1	Williams	0.665	0.520	0.849	0.001	0.283		170/2248	259/2402	443/4716, 8.8% [6.6%–11.8%]
1	Seim	0.648	0.511	0.823	0.000	0.265		164/2163	256/2314	406/4415, 8.7% [6.4%–11.8%]
1	Baracs	0.640	0.507	0.808	0.000	0.315		157/2135	249/2280	429/4650, 8.6% [6.4%–11.5%]
1	Turtainen	0.631	0.505	0.788	0.000	0.447		149/2184	243/2340	392/4426, 8.6% [6.3%–11.5%]
1	Mingmalairak	0.658	0.523	0.829	0.000	0.163		175/2273	269/2427	392/4524, 8.4% [6.5%–10.9%]
1	Ford	0.666	0.532	0.833	0.000	0.079		177/2225	273/2428	450/4653, 9.5% [7.2%–12.5%]
2-RCTs removal: C (15, 2) = 105 possible combinations tested. Combinations with the largest effect displayed.										
2	Justinger & Thimour-Bergström	0.689	0.516	0.920	0.012	0.138		126/1654	193/1916	319/3570, 8.5% [6.0%–11.8%]
2	Justinger & Galal	0.721	0.553	0.940	0.016	0.124		132/1608	198/1886	330/3494, 8.8% [6.2%–12.4%]
2	Thimour-Bergström & Galal	0.702	0.534	0.923	0.011	0.197		140/1909	202/2067	342/3976, 8.3% [5.9%–11.5%]
3-RCTs removal: C (15, 3) = 455 possible combinations tested. Combination with the largest effect displayed.										
3	Thimour-Bergström & Justinger & Galal	0.730	0.538	0.992	0.045	0.069		109/1424	160/1696	269/3120, 8.2% [5.6%–12.0%]

RR = risk ratio; CI = confidence interval; LL = lower limit; UL = upper limit; dF = degrees of freedom; TS = triclosan sutures; NTS = non-triclosan sutures; N = number of patients; SSI = surgical site infections; IR = incidence rate.

FIG. 5. Forest plot of risk ratios—Blinding subgroup.

using the mixed effects method of moments on the 15 RCTs with the IR as independent parameter and $\text{Log}(\text{RR})$ as dependent parameter. The regression slope was 3.35, 95% CI $-0.60-7.31$ ($p=0.096$), thus non-significant (Fig. 9). The null hypothesis that the RR is independent of IR was therefore not rejected (Step 12). IR was ruled out as a potential confounding factor for RR in the studied ranges. One RCT presented a 22.1% IR (61 SSI in 276 patients), which was an outlier according to Tukey's rule [20,44]. That outlier IR had a negligible effect on the meta-regression of the 15 RCTs, but led to a borderline significance in the meta-regression of the nine double-blind RCTs (slope 3.74, 95% CI $0.068-7.41$ ($p<0.046$)). Sensitivity analysis with the removal of the outlier IR resulted in a meta-regression of the eight other double-blind RCTs (271 SSI in 2836 patients) with a non-significant slope (slope 3.05, 95% CI $-4.83-7.11$ ($p=0.710$)).

Given the above results, it was concluded that results were significant with robust conclusions, and that triclosan was the only known cause of difference in SSI frequency between treatment arms (Step 13). Certain risk factors could not be ruled out as potential factors of heterogeneity and bias, so effect size should continue to be regarded as potentially heterogeneous across study populations.

Discussion

Internal validity

The two newly included trials increased the number of double-blind RCTs to nine studies compared with seven in the previous meta-analysis [11]. The double-blind RCTs, being designed to ensure equipoise of the patient sample and comparability of study arms from baseline through follow-up completion (given TS and NTS made of the same material), could not be distinguished. All double-blind RCTs used CDC criteria of SSI except the Rozelle trial, which used microbiologic culture proof of shunt infection [12]. The core of double-blind studies yielded a highly robust predominant effect, with a small proportion of heterogeneity (I^2 close to 0), no indicator of publication bias and robustness of results to the removal of up to three RCTs.

Certain potential heterogeneity factors that could have affected the pooled RR were not ruled out owing to an insufficient number of estimates per subgroup. Such was the case with suture material and operation type. However, the proportion of heterogeneity among the nine double-blind RCTs was close to zero (percentage calculated using the I^2 statistic). The incidence rate of SSI, which is the expression of the differences in risk through the trials' clinical settings, was ruled out as factor of bias in the double-blind RCTs below outlier IR of 22% and in the 15 RCTs including the outlier [20]. Finally, the pooled RR was significant throughout three incision classes, clean, clean-contaminated, and contaminated classes, both among the double-blind RCTs and all 15 RCTs, which included approximately 80% of enrolled patients. This distribution of a predominant effect throughout the incision contamination class case-mix indicated that the overall predominant effect was not biased by that case-mix. Therefore, the core of double-blind RCTs enabled this updated SLR to meet the many rigorous criteria of internal validity that were not met by previous SLRs on this topic. In addition, no indicator suggested that the conclusions of double-blind RCTs should not apply to the entire set of 15 RCTs [11,45,46].

The results of this analysis suggest strongly that the presence of triclosan-coated sutures within the surgical incision bed appears to be the predominant factor associated with a reduction in the incidence of SSIs reported in the 15 RCTs included in this meta-analysis. Given the insufficient number of trials that included the size of effect according to operation type, definition of SSI incisional and organ/space sites, it would be prudent to continue to report the predominant effect size obtained with a random-effects model, to account for potentially underestimated heterogeneity across study populations.

The two trials where RR was calculated with the number of SSIs reported per protocol and the number of randomized patients for whom demographics were also reported, according to ITT method avoided decreasing the sample size by 11.9% (12 in 101 patients) yielding an RR of 0.39 ITT instead of 0.37 PP in the Zhang trial, and avoided decreasing the sample size by 15.3% (23 in 150 patients) yielding an RR of

TABLE 3. SENSITIVITY ANALYSIS: META-ANALYSIS OF THE DOUBLE-BLIND RCTs

Number of RCTs removed	RCT removed	RR	95% CI LL	95% CI UL	p-value	Publication bias		N SSI/N TS	N SSI/N NTS	IR N SSI/N total & % SSI	
						Egger p-value					
	None	0.645	0.511	0.815	0.00024	0.521		132 / 1,585	200 / 1,625	332/3,210, 10.6%	[7.4%, 15.0%]
1	Galal	0.675	0.523	0.870	0.002	0.504		115 / 1,355	167 / 1,405	282/2,760, 10.5%	[6.8%, 15.7%]
1	Isik	0.641	0.495	0.829	0.001	0.498		128 / 1,415	188 / 1,285	316/2,700, 12.2%	[8.9%, 16.4%]
1	Justinger	0.662	0.500	0.876	0.004	0.403		101 / 1,100	158 / 1,254	259/2,354, 10.9%	[7.2%, 16.1%]
1	Mingmalairak	0.631	0.498	0.800	0.00015	0.204		127 / 1,535	196 / 1,575	323/3,110, 10.7%	[7.3%, 15.5%]
1	Nakamura	0.663	0.516	0.853	0.001	0.644		123 / 1,379	181 / 1,421	304/2,800, 11.2%	[7.6%, 16.1%]
1	Rozelle	0.666	0.539	0.822	0.00016	0.926		130 / 1,539	192 / 1,587	322/3,126, 10.4%	[7.1%, 15.2%]
1	Thimour-Bergström	0.644	0.484	0.857	0.003	0.523		109 / 1,401	162 / 1,435	271/2,836, 9.9%	[6.6%, 14.7%]
1	Turtiainen	0.574	0.454	0.726	0.000004	0.941		101 / 1,446	170 / 1,488	271/2,934, 9.5%	[6.8%, 13.2%]
1	Williams	0.635	0.487	0.828	0.001	0.515		122 / 1,510	186 / 1,550	308/3,060, 10.0%	[6.7%, 14.7%]
2-RCTs removal C (9, 2) = 36 possible combinations tested. Combinations with the largest effect displayed.											
2	Justinger & Galal	0.707	0.521	0.960	0.026	0.331		84 / 870	125 / 1,034	209/1,904, 10.8%	[6.5%, 17.3%]
2	Thimour-Bergström	0.661	0.464	0.943	0.022	0.362		78 / 916	120 / 1,064	198/1,980, 10.1%	[6.2%, 16.2%]
2	Thimour-Bergström & Justinger	0.680	0.493	0.938	0.019	0.482		92 / 1,171	129 / 1,215	221/2,386, 9.7%	[5.9%, 15.6%]
3-RCTs removal: C (9, 3) = 84 possible combinations tested. Combination with the largest effect displayed.											
3	Thimour-Bergström & Justinger & Galal	0.721	0.484	1.073	0.107	0.214		109/1,424	160 / 1,696	148/1,530, 9.9%	[5.2%, 18.0%]

RR = risk ratio; CI = confidence interval; LL = lower limit; UL = upper limit; dF = degrees of freedom; TS = triclosan sutures; NTS = non-triclosan sutures; N = number of patients; SSI = surgical site infections; IR = incidence rate.

FIG. 6. Forest plot of risk ratios—CDC incision class subgroup

0.71 ITT instead of 0.66 PP in the Williams trial.[15,18] This approach thus maintained sample size without overestimating treatment effect.

External validity

The applicability of these pooled results to future patient care depends on the similarity of the case-mix of future surgical patient populations based upon the case-mix reported in the 15 RCTs, where case-mix is defined in terms of

incision contamination class. External validity should be quantifiable when examining patients with incisions classified as clean, clean-contaminated, and contaminated, with high confidence that the risk reduction of SSIs found in clean incisions should apply, given the robustness of that subgroup but with a lower confidence for clean-contaminated and contaminated incisions. External validity could not be established for the effect of triclosan coated sutures to close dirty incisions or for operations where incision contamination was not described.

FIG. 7. Forest plot of risk ratios—Operation type subgroup.

Random-Effects Pooled Incidence Rate (IR) of SSI - 15 RCTs

FIG. 8. Forest plot of the incidence rates of SSI–Overall.

Conclusion

This systematic literature review (SLR) and meta-analysis identified 15 randomized controlled trials and produced a risk ratio of 0.67, 95% CI 0.54–0.84 ($p = < 0.00053$), demonstrating a highly statistically significant, lower risk of SSI following operative procedures in incisions which were closed with triclosan-coated sutures compared to non-antimicrobial closure technology. This result was robust to the removal of up to three trials. Similar results, yielded by the core of nine double-blind randomized controlled trials, supported the hypothesis that triclosan coated sutures were responsible for the reported reduction in SSI.

Trials comparing triclosan polyglactin to polyglactin demonstrated a robust risk ratio in favor of triclosan. Subgroup analysis was underpowered to determine heterogeneity between studies comparing different suture materials.

Diagnostic method was ruled out as a cause of bias and the incidence of SSIs, which ranged between 2% and 22%, had no significant effect on risk reduction effect of the triclosan coated sutures and was therefore ruled out as a cause of bias. Based upon this analysis there was a high concentration of confidence that a 20% to 50% reduction in SSIs should be expected in surgical procedures involving clean surgical incisions. A smaller and statistically significant reduction could be expected in clean-contaminated and contaminated incisions but these results were not robust when considered separately from the clean incisions. No conclusions could be drawn based upon this analysis on the impact of triclosan sutures as a risk reduction strategy for SSIs involving dirty incisions or surgical procedures where the composite incidence rate of infections exceeded 17%. The two additional peer-reviewed double-blind RCTs reinforced the evidence level of this SLR (CEBM level Ia).

FIG. 9. Meta-regression of (Log) risk ratio depending on incidence rate.

Acknowledgments

FCD, CEE, and DJL thank the authors of the compiled studies who responded to our request for additional information: Henri R. Ford, MD (University of Southern California Keck School of Medicine, USA); Chatchai Mingmalairak, MD (Thammasat University Hospital, Pathumthani, Thailand); Jozsef Baracs, MD (University of Pecs, Pecs, Hungary); Toru Nakamura, MD (Graduate School of Medicine, Sapporo, Hokkaido, Japan); Anders Jeppsson, MD, on behalf of Thimour-Bergström L. et al. (Sahlgrenska University Hospital, Gothenburg, Sweden); Sahin Sahey, MD on behalf of Isik I. et al., (Marmara University, Istanbul, Turkey); Marko Sever, MD, on behalf of Rasic Z. et al. (University Hospital Sveti Duh, Zagreb, Croatia); and David Leaper, MD, FACS, (Imperial College, London, UK) on behalf of Williams et al.; who were co-authors of the studies included in the meta-analysis.

Author Disclosure Statement

This systematic analysis was conducted independent of any external funding support from non-profit or for-profit entities. The design, execution, analysis, and conclusions of this analysis are the sole responsibility of authors Daoud, Edmiston, and Leaper.

References

1. Jones RD, Jampani HB, Newmann JL, et al. Triclosan; a review of effectiveness and safety in healthcare settings. *Am J Infect Control* 2000;28:184–196.
2. Bhargava HN, Leonard PA. Triclosan: application and safety. *Am J Infect Control* 1996; 24: 209–218.
3. Rothenburger S, Spangler D, Bhende S, et al. In-vitro antimicrobial evaluation of coated Vicryl Plus antibacterial sutures (coated Polyglactin 910 with Triclosan) using zone of inhibition assay. *Surg Infect* 2002; 3: S79–S87.
4. Storch ML, Rothenburger SJ, Jacinto G. Experimental efficacy study of coated VICRYL plus antibacterial suture in guinea pigs challenged with *Staphylococcus aureus*. *Surg Infect* 2004;5:281–288.
5. In Ming X. et. al. “In vivo and in vitro antibacterial efficacy of PDS plus (polidioxanone with triclosan) suture. *Surg Infect* 2008;9:451–457.
6. Ethicon instruction for use: VICRYL Plus – 389595.R04
7. Ethicon instruction for use: PDS Plus – 389688.R02
8. Ethicon instruction for use: MONOCRYL Plus – 389680.R02
9. Ford HR, Jones P, Gaines B, et al. Intraoperative handling and incision healing: controlled clinical trial comparing coated Vicryl Plus antibacterial suture (coated polyglactin 910 with triclosan) with coated Vicryl suture (coated Polyglactin 910 suture). *Surg Infect* 2005;6:313–321.
10. Storch ML, Scalzo H, Van Lue S, et al. Physical and Functional Comparison of Coated VICRYL *Plus Antibacterial Suture (Coated Polyglactin 910 Suture with Triclosan) with Coated VICRYL*Suture (Coated Polyglactin 910 Suture) *Surg Infect* 2002;3 Suppl 1:S65–77.
11. Edmiston CE Jr, Daoud FC, Leaper D.: Is there an evidence-based argument for embracing an antimicrobial (triclosan)-coated suture technology to reduce the risk for surgical-site infections?: A meta-analysis. *Surgery* 2013;154:89–100.
12. Rozzelle CJ, Leonardo J, Li V. Antimicrobial suture incision closure for cerebrospinal fluid shunt surgery: a prospective, double-blinded, randomized controlled trial. *J Neurosurg Pediatr* 2008;2:111–117.
13. Mingmalairak C, Ungbhakorn P, Paocharoen V. Efficacy of antimicrobial coating suture coated polyglactin 910 with triclosan (Vicryl plus) compared with polyglactin 910 (Vicryl) in reduced surgical site infection of appendicitis, double blind randomized control trial, preliminary safety report. *J Med Assoc Thai* 2009;92:770–775.
14. Zhuang CP, Cai GY, Wang YQ. Comparison of two absorbable sutures in abdominal wall incision. *Journal of Clinical Rehabilitative Tissue Engineering Research* 2009; 13: 4045–4048. (www.crter.cn CRTER website group Liao ICP 05011357 CRTER, submission: <http://oa.crter.org/zgclckfen/ch/index.aspx>). Embase dx.doi.org/10.3969.
15. Zhang ZT, Zhang HW, Fang XD, et al. Pilot evaluation of cosmetic outcome and surgical site infection rates of coated vicryl* plus antibacterial (polyglactin 910) suture compared to Chinese silk in scheduled breast cancer surgery. *Chin Med J* 2011;124:719–724.
16. Galal I, El-Hindawy K. Impact of using triclosan-antibacterial sutures on incidence of surgical site infection. *Am J Surg* 2011;202:133–138.
17. Rasic Z, Schwarz D, Adam VN, et al. Efficacy of antimicrobial triclosan-coated polyglactin 910 (Vicryl* Plus) suture for closure of the abdominal wall after colorectal surgery. *Collegium Antropologicum* 2011;35:439–443.
18. Williams N, Sweetland H, Goyal S, et al. Randomized trial of antimicrobial-coated sutures to prevent surgical site infection after breast cancer surgery. *Surg Infect* 2011;12: 469–474.
19. Baracs J, Huszár O, Sajjadi SG, et al. Surgical site infections after abdominal closure in colorectal surgery using triclosan-coated absorbable suture (PDS Plus) vs. uncoated sutures (PDS II): a randomized multicenter study. *Surg Infect* 2011;12:483–489.
20. Turtiainen J, Saimanen EI, Mäkinen KT, et al. Effect of triclosan-coated sutures on the incidence of surgical incision infection after lower limb revascularization surgery: A randomized controlled trial. *World J Surg.* 2012;36:2528–2534.
21. Seim BE, Tønnessen T, Woldbaek PR. Triclosan-coated sutures do not reduce leg incision infections after coronary artery bypass grafting. *Interact Cardiovasc Thorac Surg* 2012;15:411–415.
22. Isik I, Selimen D, Senay S, et al. Efficiency of Antibacterial Suture Material in Cardiac Surgery: A Double-Blind Randomized Prospective Study. *Heart Surg Forum.* 2012;15:E40–45.
23. Nakamura T, Kashimura N, Noji T, et al. Triclosan-coated sutures reduce the incidence of incision infection and cost after colorectal surgery: A randomized controlled trial. *Surgery* 2013;153:576–583.
24. Moher D, Liberati A, Tetzlaff J, et al. The PRISMA Group. Preferred reporting items for systematic reviews and meta-analyses: the PRISMA statement. *Open Med* 2009;3:123–130.
25. Mangram AJ, Horan TC, Pearson ML, Silver LC, Jarvis WR. Guideline for prevention of surgical site infection, 1999. Hospital Infection Control Practices Advisory Committee. *Infect Control Hosp. Epidemiol* 1999;20:250–278.
26. Horan TC, Gaynes RP, Martone WJ, et al. CDC definitions of nosocomial surgical site infections, 1992: a modification of CDC definitions of surgical incision infections. *Infect Control Hosp Epidemiol* 1992;13: 606–608.
27. Centre for Evidence-Based Medicine, University of Oxford, UK. www.cebm.net. Last accessed July 30. 2013.

28. Higgins JPT, Green S, eds. Cochrane Handbook for Systematic Reviews of Interventions version 5.1.0 updated March 2011. www.cochrane-handbook.org.
29. Egger M, Davey Smith G, Schneider M, et al. Bias in meta-analysis detected by a simple, graphical test. *Br Med J* 1997;315:629–634.
30. Egger M, Davey Smith G. Bias in location and selection of studies. *Br Med J* 1998;316:61–66.
31. Sterne JAC, Egger M, Davey Smith G. Investigating and dealing with publication and other biases. In: Egger M, Davey Smith G, Altman DG (eds). *Systematic Reviews in Health Care: Meta-Analysis in Context*. London: BMJ Publishing Group, 2001:189–208.
32. Gavaghan DJ, Moore AR, McQay HJ. An evaluation of homogeneity tests in meta-analysis in pain using simulations of patient data. *Pain* 2000;85:415–424.
33. Higgins JPT, Thompson SG. Quantifying heterogeneity in a meta-analysis. *Statistics in Medicine* 2002;21:1539–1558.
34. Higgins JPT, Thompson SG, Deeks JJ, Altman DG. Measuring inconsistency in meta-analyses. *Brit Med J* 2003;327:557–560.
35. Borenstein M, Larry V Hedges L, et al. A basic introduction to fixed-effect and random-effects models for meta-analysis. *Res Syn Meth* 2010;1:97–111.
36. DerSimonian R, Laird N. Meta-analysis in Clinical Trials. *Control Clin Trials* 1986;7:177–188.
37. Mangram AJ, Horan TC, Pearson ML, Silver LC, Jarvis WR. Guideline for prevention of surgical site infection, 1999. Hospital Infection Control Practices Advisory Committee. *Infect Control Hosp. Epidemiol* 1999;20:250–278.
38. Thimour-Bergström L, Roman-Emanuel C, Scherstén H, Friberg O, Gudbjartsson T, Jeppsson A. Triclosan-coated sutures reduce surgical site infection after open vein harvesting in coronary artery bypass grafting patients: a randomized controlled trial. *Eur J Cardiothorac Surg* 2013;44:931–938.
39. Justinger C, Slotta JE, Ningel S, Gräber S, Kollmar O, Schilling MK. Surgical-site infection after abdominal wall closure with triclosan-impregnated polydioxanone sutures: Results of a randomized clinical pathway facilitated trial (NCT00998907). *Surgery* 2013;154:589–595.
40. Defazio A, Datta M, Nezhat C. Does the use of Vicryl Plus antibacterial suture decrease the incidence of umbilical infection when compared to Vicryl suture? *Fertil Steril* 2005;84(suppl 1):S161;P–29.
41. Mattavelli I, Nespoli L, Alfieri S, et al. Effect of triclosan-coated suture on surgical site infection after colorectal surgery: Final results of a multicenter, prospective, randomized trial. *Surg Infect* 2013;14(2):A9.
42. Singh H, Emmert MY, Sakaguchi H, et al. Antibacterial suture reduces surgical site infections in coronary artery bypass grafting. *Heart Surgery Forum*. 2010;13(suppl 1):S85.
43. Khachatryan N, Dibirov M, Omelyanovsky V, et al. Prevention of postoperative infections in abdominal surgery using reabsorbable suture with antibacterial activity (vicryl plus) versus reabsorbable standard sutures. *Surgical Infect* 2013;12(2):A13–A14.
44. Hoaglin D, Frederick D, Mosteller F, Tukey J, eds. *Understanding robust and exploratory data analysis*. New York. John Wiley & Sons, 1983:39,54,62,223.
45. Wang ZX, Jiang CP, Cao Y, et al. Systematic review and meta-analysis of triclosan-coated sutures for the prevention of surgical-site infection. *Brit J Surg* 2013;100:465–473.
46. Chang WK, Srinivasa S, Morton R, Hill AG. Triclosan-impregnated sutures to decrease surgical site infections: systematic review and meta-analysis of randomized trials. *Ann Surg* 2012;255:854–859.

Address correspondence to:

Dr. Frederic C. Daoud

Unit of Biostatistics and Bioinformatics

MEDEXTENS SARL

75 rue de Lourmel

75015, Paris, France

E-mail: fcdaoud@medextens.com