Published in final edited form as:

Phys Med Biol. 2014 April 21; 59(8): 1877–1897. doi:10.1088/0031-9155/59/8/1877.

Noise and Analyzer-Crystal Angular Position Analysis for Analyzer-Based Phase-Contrast Imaging

Keivan Majidi¹, Jun Li², Carol Muehleman², and Jovan G. Brankov¹

Jovan G. Brankov: brankov@iit.edu

¹Department of Electrical and Computer Engineering, Illinois Institute of Technology, Chicago, IL 60616, USA

²Department of Biochemistry Rush University Medical Center, Chicago, IL 60616, USA

Abstract

The analyzer-based phase-contrast X-ray imaging (ABI) method is emerging as a potential alternative to conventional radiography. Like many of the modern imaging techniques, ABI is a computed imaging method (meaning that images are calculated from raw data). ABI can simultaneously generate a number of planar parametric images containing information about absorption, refraction, and scattering properties of an object. These images are estimated from raw data acquired by measuring (sampling) the angular intensity profile (AIP) of the X-ray beam passed through the object at different angular positions of the analyzer crystal. The noise in the estimated ABI parametric images depends upon imaging conditions like the source intensity (flux), measurements angular positions, object properties, and the estimation method. In this paper, we use the Cramér-Rao lower bound (CRLB) to quantify the noise properties in parametric images and to investigate the effect of source intensity, different analyzer-crystal angular positions and object properties on this bound, assuming a fixed radiation dose delivered to an object. The CRLB is the minimum bound for the variance of *an unbiased estimator and defines the best noise performance that one can obtain regardless of which estimation method is used to estimate ABI parametric images*. The main result of this manuscript is that the variance (hence the noise) in parametric images is directly proportional to the source intensity and only a limited number of analyzer-crystal angular measurements (eleven for uniform and three for optimal non-uniform) are required to get the best parametric images. The following angular measurements only spread the total dose to the measurements *without improving or worsening CRLB*, but the added measurements may improve parametric images by reducing estimation bias.

Next, using CRLB we evaluate the Multiple-Image Radiography (MIR), Diffraction Enhanced Imaging (DEI) and Scatter Diffraction Enhanced Imaging (S-DEI) estimation techniques, though the proposed methodology can be used to evaluate *any other* ABI parametric image estimation technique.

1. Introduction

Analyzer-based phase-contrast imaging (ABI) utilizes the Bonse-Hart camera (Bonse and Hart, 1965), which is schematically shown in Figure 1. This system uses a system of diffracting crystals to make precise measurements of the angular content of an X-ray beam after it traverses an object. Specifically, the Bonse-Hart camera allows one to measure the

beam angular intensity profile (AIP), i.e. X-ray beam intensity as a function of the angular direction of propagation at each pixel. To achieve this, the ABI system uses a highly collimated (directional) and quasi-monochromatic imaging beam that is obtained by the use of a *double crystal monochromator*. As the collimated X-ray beam passes through the object (e.g., a patient's body), the beam will be attenuated by absorption processes, deflected due to refraction, and directed into a broadened range of angular directions due to ultra small-angle X-ray scattering (USAXS). After passing through the object, the beam strikes the *analyzer-crystal*, which diffracts the beam traveling in a specific narrow direction according to Bragg's law. This beam is then detected by a pixelated X-ray detector. The imaging beam at the location of the object is usually wide (i.e. size in x axis direction) but only several millimeters in the vertical height (i.e. size in y axis direction). Therefore the object is mounted on a scanning stage driven by a stepping motor, thus allowing object *scanning* in the vertical direction. Next, by changing angular position (*rocking*) of the analyzer-crystal (in short analyzer), a number of raw images are acquired and used to calculate parametric images describing object properties.

For example, in the well-known diffraction enhanced imaging technique (DEI) (Chapman *et al.*, 1997), the acquisition is done at two angular analyzer positions and it can therefore simultaneously estimate two parametric images, which contain information about refraction and absorption properties of the object. In multiple image radiography (MIR) (Wernick *et al.*, 2003), which is an improvement upon the DEI, the acquisition is done at a number - larger than three - of angular analyzer positions; therefore, in addition to the previous two planar parametric images, MIR can also estimate ultra small-angle X-ray scattering (USAXS).

Examples of MIR images for formalin fixed human thumb tissue are shown in Figure 2. The attenuation image shows the absorption and Compton scatter rejection together (Chapman *et al.*, 1997). The refraction image in MIR shows the integrated effect of refractive index variations along the beam path, and the USAXS image represents the sub-pixel textural structure of the object.

Because the contrast of the refraction and USAXS images does not rely on the absorption of X-ray by the object, ABI can be performed at X-ray energies higher than conventional radiography where radiation exposure to the patient can be low. In addition, ABI parametric images are virtually immune to degradation due to Compton scatter at higher angles and have good contrast, which can help to eliminate breast compression as in conventional mammography. For these reasons, ABI is emerging as a potential alternative to conventional radiography (Faulconer *et al.*, 2009; Faulconer *et al.*, 2010; Kitchen *et al.*, 2011; Honnicke *et al.*, 2012; Zhao *et al.*, 2012). As additional reading see (Brankov and Zysk, 2012; Suortti *et al.*, 2013) for a border overview of ABI and (Zhou and Brahme, 2008; Coan *et al.*, 2009; Anastasio and La Riviere, 2012; Bravin *et al.*, 2013; Coan *et al.*, 2013; Lider and Kovalchuk, 2013; Liu *et al.*, 2013) for phase-sensitive imaging in general.

Like many other modern imaging systems, ABI is an indirect imaging method, meaning that one must process the measured data in order to obtain the final parametric images.

Therefore, the noise properties in ABI parametric images depend on the imaging conditions and the estimation method.

In the past, experimental ABI investigations were limited to synchrotron facilities due to the need for a very bright synchrotron X-ray radiation source that allows collimatization and monochromatization of the beam, with double crystal monochromator, while still having sufficient number of photons for imaging. Recently, efforts have been underway to implement ABI with a bench-top system employing conventional X-ray tube sources (Chapman *et al.*, 2006; Vine *et al.*, 2007; Nesch *et al.*, 2009; Connor *et al.*, 2011) in which the source photon flux is the major performance constraint.

Therefore, in this paper, we will quantify the parametric images noise due to two major causes: limited source intensity and analyzer angular positions by using the Cramér-Rao lower bound (CRLB) (Kay, 1993). CRLB is the minimum bound for the variance of *any unbiased estimator* and defines the best noise performance that one can obtain in parameter estimation from the raw data. The CRLB methodology unfortunately does not provide the minimum variance unbiased (MVU) estimator.

The work presented here is an extension of preliminary conference results (Majidi *et al.*, 2008b; Marquet *et al.*, 2006) where we used simulated phantom data. In (Brankov and Zysk, 2012) we presented some theoretical results, but in this paper we take this approach a step further including two experimental data sets, providing the full derivation, and clearly identifying source intensity and analyzer angular position impact on parametric images. In addition to MIR we also provide findings for analyzer angular positions selected in DEI (Chapman *et al.*, 1997) and scatter-DEI (Rigon *et al.*, 2003).

The idea of assessing the quality of the raw data by quantifying the parametric image quality is not new. For example, in (Pagot *et al.*, 2005) the authors evaluated the best angular measurement positions for the refraction and scattering signals, as defined by area contrast, for a scenario where only two images were acquired. In our approach we define a more general statistical approach, based on CRLB, to assess the quality of the raw data. We also go a step further and present methodology to optimize the analyzer angular positions at which we measure beam's AIP. Finally we finish by using CRLB to evaluate the performance of three estimation methods, in this work that is MIR (Wernick *et al.*, 2003), DEI (Chapman *et al.*, 1997) and scatter-DEI (Rigon *et al.*, 2003). The presented evaluation methodology can be directly used on other estimation methods like extended-DEI (Chou *et al.*, 2007; Rigon *et al.*, 2007) and model based MIR (Majidi *et al.*, 2008a). The purpose of this work is not to fully evaluate MIR or any other technique or to prove conclusively that any estimation methods will always perform best. The study is merely an illustration of the proposed procedure for quantifying noise and selecting analyzer angular positions by utilizing CRLB. The illustration reveals pitfalls and gains that one can encounter during this process.

First, we present the ABI image formation model (Khelashvili *et al.*, 2006) based on radiative transfer theory (Ishimaru, 1997), followed by derivation of the CRLB in section 3. In section 4, we use CRLB as a guide to optimize the set of analyzer angular positions while

in section 5 we evaluate the performance of three previously mentioned estimation methods. The conclusions are given in section 6.

2. Image formation model

In (Khelashvili *et al.*, 2006), the image formation model was derived using radiative transfer theory. According to this model the AIP of the beam after passing through the object can be written as the following convolution:

$$g(\theta; \mathbf{v}) = I_0 R(\theta) * f(\theta; \mathbf{v}), \quad (1)$$

where $R(\theta)$ is the AIP of the beam that would be measured in absence of any object on the detector plane, $*$ denotes one-dimensional convolution, and I_0 is the total number of photons per pixel for a preset total radiation exposure. The function $R(\theta)$ is called the *intrinsic rocking curve* as it is an inherent property of the imaging system and can be accurately measured and modeled as a Pearson type VII function which we will show later.

In ABI, the object is characterized by a hypothetical quantity called the *object function*, $f(\theta; \mathbf{v})$, where θ is the angular direction of propagation and $\mathbf{v}$ is a vector containing the parameters that define object interaction with the X-ray beam.

The full object model given in (Khelashvili *et al.*, 2006) and summarized in Appendix A, is derived using radiative transfer theory where the object is modeled as a scattering, stratified medium. The full object function given in Eq A-3 is uniquely described by *eight object parameters*. Fortunately the approximate simplification of the full image formation model, also shown in Appendix A, yields to the following Gaussian curve for the object function defined by only three parameters:

$$f(\theta; \mathbf{v}) = \exp(-\nu_1) \frac{1}{\sqrt{2\pi\nu_3}} \exp\left(-\frac{(\theta - \nu_2)^2}{2\nu_3}\right). \quad (2)$$

Now, at each pixel the object is characterized by a parameter vector, $\mathbf{v} = [\nu_1, \nu_2, \nu_3]^T$, containing the following parameters: ν_1 , which conveys information about total attenuation by the object, ν_2 is the angle of refraction caused by the object, and ν_3 is the angular beam divergence (analogous to variance) induced by the object. This object function is the AIP that would result from illuminating the object with a unit intensity, perfectly collimated beam, i.e., a beam having its initial AIP described by a Dirac delta function $\delta(\theta)$. In this context, the effect of the object on the X-ray beam is a convolution in terms of the propagation direction θ , and in this sense, the object function plays the role of an angular impulse-response (aka Green's) function. By measuring the parameters of this impulse response function, one can obtain attenuation, refraction, and USAXS parameters, which can be displayed in the form of images.

The image formation model given in Eq (1) and object function given in Eq (2) was previously used in (Wernick *et al.*, 2003; Rigon *et al.*, 2003; Rigon *et al.*, 2007) in an ad-hoc fashion. For other image formation models see (Suhonen *et al.*, 2007; Diemoz *et al.*,

2010; Pavlov *et al.*, 2001; Arfelli *et al.*, 1998) where the authors usually assume a bell-shaped rocking curve such as Gaussian or Lorentzian convolved with an object function but no derivation of the object function was provided such as in (Khelashvili *et al.*, 2006). All image formation models listed above are defined on a pixel-by-pixel basis; see (Wernick *et al.*, 2003) for the justification; hence, $\mathbf{v}$ denotes the values for any arbitrary pixel in the image.

In practice, measurements of $g(\theta; \mathbf{v})$ are made on a digital detector for different angular positions of the analyzer crystal that will be denoted as $g[l]$. The expected value of the measured discrete AIP on the detector can be modeled as:

$$g_l \triangleq E[g[l]] = \left[\frac{I_0}{L} R(\theta)^* f(\theta; \mathbf{v}) \right] \Big|_{\theta=\theta_l}, \quad (3)$$

where $l = 1, 2, \dots, L$ represents the index of the analyzer angle θ_l , $E[\cdot]$ denotes the expected value, and I_0 / L represents the division of total exposure dose among L measurements.

In practical implementation of ABI using conventional X-ray tubes, photon flux will be the performance constraint; therefore, we can assume that the measured data is photon-limited, so Poisson noise will be the dominant noise source in raw ABI images (Wernick *et al.*, 2003). Under this noise model, the likelihood function of the raw data is given by:

$$p(\mathbf{g}; \mathbf{v}) = \prod_{l=1}^L \frac{g_l^{g[l]} e^{-g_l}}{g[l]!}, \quad (4)$$

where $\mathbf{g} = [g[1], g[2], \dots, g[l]]^T$ is a vector containing measured data.

3. Cramér-Rao lower bounds

To investigate the fundamental noise properties of ABI, we next consider the theoretical limit on the noise variance in estimation of the parametric images. This can serve as a foundation for optimizing the data acquisition procedure, as well as evaluation and comparisons of estimation algorithms. In ABI, we have three parameters to estimate simultaneously. The CRLB theorem for vector estimations (Kay, 1993) states that the variance of any unbiased estimator of the i^{th} element of the vector $\mathbf{v}$, denoted $\hat{v}_i$, is bounded as follows:

$$\text{var}(\hat{v}_i) \geq \{\mathbf{I}^{-1}(\mathbf{v}; \boldsymbol{\Theta})\}_{i,i}, \quad (5)$$

in which $\boldsymbol{\Theta} = \{\theta_1, \theta_2, \dots, \theta_L\}$ is a set of analyzer angular positions, $\{\cdot\}_{i,i}$ denotes i^{th} diagonal matrix element and $\mathbf{I}(\mathbf{v}; \boldsymbol{\Theta})$ is a Fisher information matrix. This matrix has its i^{th} column and j^{th} row elements defined by:

$$\{\mathbf{I}(\mathbf{v}; \boldsymbol{\Theta})\}_{i,j} = -E \left[\frac{\partial^2}{\partial v_i \partial v_j} \ln p(\mathbf{g}; \mathbf{v}) \right]. \quad (6)$$

Substituting (1)–(4) into (6) yields to the following Fisher information matrix, which is derived in Appendix B:

$$\mathbf{I}(\boldsymbol{\nu};\boldsymbol{\Theta})=I_0\frac{1}{L}\sum_{l=1}^L\mathbf{I}'(\boldsymbol{\nu};\theta_l), \quad (7)$$

where $\mathbf{I}'(\boldsymbol{\nu};\theta_l)=\mathbf{a}_l\mathbf{a}_l^T$, and

$$\mathbf{a}_l=\begin{bmatrix} \sqrt{g'_l}\frac{\nu_2 g'_l - h'_l}{\nu_3 \sqrt{g'_l}} - \frac{\nu_3 g'_l - q'_l}{2\nu^2 \sqrt{g'_l}} \end{bmatrix}^T, \quad (8)$$

where

$$g'_l \triangleq [R(\theta)^* f(\theta; \boldsymbol{\nu})] \Big|_{\theta=\theta_l}, \quad (9)$$

$$h'_l \triangleq [R(\theta)^* \theta f(\theta; \boldsymbol{\nu})] \Big|_{\theta=\theta_l}, \quad (10)$$

$$q'_l \triangleq [R(\theta)^* (\theta - \nu_2)^2 f(\theta; \boldsymbol{\nu})] \Big|_{\theta=\theta_l}. \quad (11)$$

Here, we note that the $\mathbf{I}'(\boldsymbol{\nu};\theta_l)$ matrix has rank one, so the inverse of the Fisher information matrix, $\mathbf{I}^{-1}(\boldsymbol{\nu};\boldsymbol{\Theta})$, exist *only and only if* at least three independent angular measurements are acquired. Though the CRLB expressions are complicated, one can compute these bounds numerically for any given set of analyzer angular measurements.

Here we will rewrite Eq (5) using result in Eq (7) as:

$$\text{var}(\hat{\nu}_i) \geq \text{CRLB}_i(\boldsymbol{\Theta}), \quad (12)$$

where:

$$\text{CRLB}_i(\boldsymbol{\Theta})=\frac{L}{I_0}\left\{\left[\sum_{l=1}^L\mathbf{I}'(\boldsymbol{\nu};\theta_l)\right]^{-1}\right\}_{i,i}, \quad (13)$$

Here one can note that the estimator variance bound is inversely proportional to the total exposure dose I_0 and directly to the number of angular measurements. This expression shows that information given by one angular measurement can be interpreted independently from the total exposure dose and the total number of angular measurements. Therefore the results and conclusions obtained using a high flux source, like synchrotron, can be translated to a low flux source, like a conventional X-ray tube. This equation also points out that the estimators' variance is a function of intrinsic system rocking curve and the object function

and *not only system rocking curve* as sometimes assumed in literature (Kitchen *et al.*, 2010; Pagot *et al.*, 2005).

Now we would like to point out that Eq (13) is strictly valid only for equal acquisition time at each analyzer angular position. If one desires to modify acquisition times this can be easily incorporated into CRLB as:

$$CRLB_i(\Theta) = \frac{T}{I_0} \left\{ \left[\sum_{l=1}^L t_l \mathbf{I}'(\nu; \theta_l)_l \right]^{-1} \right\}_{i,i}, \quad (14)$$

where t_l is acquisition time for each individual angular position and $T = \sum_{l=1}^L t_l$ is the total acquisition time needed to acquire I_0 photons. Optimization of t_l is left for future research.

Note that the specific form of rocking curve and object function does not affect the CRLB equations derived in (13); therefore, the CRLB equations are general. Also note that considerations in this paper do not take into account the effect of a finite source spot size. This secondary effect is also left for future research.

If one considers methods like DEI (Chapman *et al.*, 1997) or scatter-DEI (Rigon *et al.*, 2003) that only estimate two parameters and use a specific analyzer angular positions one needs to modify bounds as follows. For DEI CRLB will become:

$$CRLB_i^{DEI}([\theta_H, \theta_L]) = \frac{2}{I_0} \left\{ \left[\left\{ \mathbf{I}'_{DEI}(\nu_{DEI}; \theta_H) + \mathbf{I}'_{DEI}(\nu_{DEI}; \theta_L) \right\} \right]^{-1} \right\}_{i,i}, \quad (15)$$

where $\mathbf{I}'_{DEI}(\nu; \theta_l) = \mathbf{a}_l^{DEI} (\mathbf{a}_l^{DEI})^T$,

$$\mathbf{a}_l^{DEI} = \begin{bmatrix} \sqrt{g'_l} \nu_2 g'_l - h'_l \\ \nu_3 \sqrt{g'_l} \end{bmatrix}^T, \quad (16)$$

and $\mathbf{v}_{DEI} = [\nu_1, \nu_2]$ assuming $\nu_3 = 0$.

Here we note that DEI is using angular measurements obtained at half of the rocking curve full with at the half maximum (FWHM) as noted by θ_H and θ_L in later figures.

For scatter-DEI we will have:

$$CRLB_i^{S-DEI}([\theta_{TOP}, \theta_{TOE}]) = \frac{2}{I_0} \left\{ \left[\left\{ \mathbf{I}'_{S-DEI}(\nu_{S-DEI}; \theta_{TOP}) + \mathbf{I}'_{S-DEI}(\nu_{S-DEI}; \theta_{TOE}) \right\} \right]^{-1} \right\}_{i,i}, \quad (17)$$

where $\mathbf{I}'_{S-DEI}(\nu; \theta_l) = \mathbf{a}_l^{S-DEI} (\mathbf{a}_l^{S-DEI})^T$,

$$\mathbf{a}_l^{S-DEI} = \left[\sqrt{g_l'} \frac{\nu_3 g_l' - q_l'}{2\nu^2_3 \sqrt{g_l'}} \right]^T, \quad (18)$$

and $\mathbf{v}_{S-DEI} = [\nu_1, \nu_3]$ assuming $\nu_2 = 0$.

S-DEI is using angular measurements obtained at top and at the toe of the rocking curve as noted by θ_{TOP} , and θ_{TOE} in later figures.

It is well known that the DEI and S-DEI methods are prone to artifact and bias. However, DEI produces reasonably good images if the object is weakly scattering and S-DEI if the object is weakly refracting.

A. Effect of the analyzer crystal angular position on CRLB

In imaging an object, the most important factor affecting the parametric image noise, after the beam intensity I_0 and the number of angular measurements (as noted in the previous section), is the angular position of the analyzer crystal. In this section, we will analyze the effect of the analyzer angular position θ_l on CRLB. To gain intuition about the effect of the angular position, we approximate the true bounds, i.e. $\{\mathbf{I}^{-1}(\boldsymbol{\nu}; \boldsymbol{\Theta})\}_{i,i}$, with the inverse of the diagonal elements of the Fisher information matrix, i.e. $\{\mathbf{I}(\boldsymbol{\nu}; \boldsymbol{\Theta})\}_{i,i}^{-1}$. This would be equivalent to the estimation of each parameter individually, while the other two parameters are known. Now we can define the following information functions:

$$V_1(\theta_l) = \{\mathbf{I}'(\boldsymbol{\nu}; \theta_l)\}_{1,1} = g_l', \quad (19)$$

$$V_2(\theta_l) = \{\mathbf{I}'(\boldsymbol{\nu}; \theta_l)\}_{2,2} = \left(\frac{\nu_2 g_l' - h_l'}{\nu_3 \sqrt{g_l'}} \right)^2, \quad (20)$$

$$V_3(\theta_l) = \{\mathbf{I}'(\boldsymbol{\nu}; \theta_l)\}_{3,3} = \left(\frac{\nu_3 g_l' - q_l'}{2\nu^2_3 \sqrt{g_l'}} \right)^2. \quad (21)$$

These functions show the amount of information we can acquire from a measurement at different analyzer angular positions. Since to visualize these functions one needs to know object parameters, we will examine these functions, for parameters obtained experimentally, in the experimental section.

B. CRLB for uniform set of analyzer angular positions

Next we will provide methodology to calculate and analyze the behavior of CRLB as we increase the number of angular measurements, i.e. analyzer angular positions. We can see in (3) that by increasing the number of measurements, L , we are decreasing the signal-to-noise ratio in each angular measurement, which represents *the fixed total exposure dose to the*

tissue regardless of how many angular measurement are collected and equal acquisition time at each analyzer angular position.

First, we consider a uniform set of analyzer angular positions in the range of $[\theta_{\min}, \theta_{\max}]$ so the angular positions are defined as following:

$$\theta_l = \left(l - \frac{L+1}{2} \right) \frac{\theta_{\max} - \theta_{\min}}{L}, l=1, 2, \dots, L. \quad (22)$$

Now for a given set of analyzer angular positions, $\Theta_u = \{\theta_1, \theta_2, \dots, \theta_L\}$, we can estimate the expected CRLB as:

$$\overline{\text{CRLB}}_i(\Theta_u) \simeq E \left[\left\{ \mathbf{I}^{-1}(\boldsymbol{\nu}; \Theta_u) \right\}_{i,i} \right]_{\text{w.r.t. } \boldsymbol{\nu}}, \quad (23)$$

where $E[\cdot]_{\text{w.r.t. } \boldsymbol{\nu}}$ represents the expectation over the probability density function (pdf) of the parameter vector $\boldsymbol{\nu}$. Taking the expectation in (23) is an intractable problem. Therefore, we approximate the expectation by the arithmetic average over the parameters obtained from the experimental study:

$$\overline{\text{CRLB}}_i(\Theta_u) \simeq \frac{1}{C} \sum_{c=1}^C \left\{ \mathbf{I}^{-1}(\boldsymbol{\nu}^c; \Theta_u) \right\}_{i,i}, \quad (24)$$

where $\boldsymbol{\nu}^c, c=1, 2, \dots, C$ represents the values of the parameter vector $\boldsymbol{\nu}$, at the c^{th} pixel in a pilot experimental study (e.g. human thumb or whole breast tissue study).

Now Eq (24) can serve as a basis for comparison of different sets of analyzer angular positions. It allows us to examine what is the impact of increasing the number of angular measurements, L , on estimated parametric images noise.

C. Optimal set of analyzer angular positions

In this section, we propose a method to investigate the possibility of further lowering the CRLB and hence achieving better estimation performance by using non-uniform analyzer angular positions.

To find a better set of non-uniform analyzer angular positions (denoted Θ^*) with respect to uniform angular positions, Θ_u , we define the following minimization problem:

$$\Theta^* = \underset{\Theta}{\text{argmin}} J(\Theta), \quad (25)$$

where

$$J(\Theta) = \sum_{i=1}^3 \frac{\overline{\text{CRLB}}_i(\Theta)}{\overline{\text{CRLB}}_i(\Theta_u)}. \quad (26)$$

The main idea is: since the magnitude scales of the parametric images average CRLB are different, we normalize each $\overline{\text{CRLB}}_i(\Theta)$ by $\overline{\text{CRLB}}_i(\Theta_u)$, which is the average CRLB obtained for the uniform set of analyzer angular positions Θ_u with the same number of angular measurements.

4. Experimental analysis of CRLB

As we pointed out earlier, to visualize Eqs (19)–(21) or to take the expectation in Eq (23) or (26) one needs the probability density function (pdf) of the parameter vector. In general, this is unmanageable, however one can estimate these quantities from an experimental study.

We want to stress that the purpose of experimental studies is not to fully evaluate ABI imaging or to fully quantify noise and select a set of analyzer angular positions in general. The study is purely an example to illustrate the proposed procedure of utilizing CRLB. The results presented in this paper will be valid only for a similar imaging system, beam energy and tissue types as in our experiments.

A. Experimental studies

Experimental data was acquired using a synchrotron light source at Brookhaven National Laboratory; X15A beamline. A double Si (333) crystal was used for beam monochromatization, and one Si (333) crystal was used as the analyzer.

- a. A formalin fixed human thumb was imaged with the beam energy of 40keV and angularly oversampled at 25 analyzer positions ranging from -4.8 to $4.8 \mu\text{rad}$ with $0.4 \mu\text{rad}$ increments.
- b. An uncompressed, formalin fixed medium size breast was imaged with the beam energy of 60keV and angularly oversampled at 21 analyzer positions ranging from -2.4 to $2.4 \mu\text{rad}$ with $0.2 \mu\text{rad}$ increments.

Both data sets are measured with long exposure time approximately 10 hours per tissue; therefore, achieving a high signal-to-noise ratio ($>35\text{dB}$) in the raw data. Next, using an MIR estimation method (Wernick *et al.*, 2003), we extract three parametric images.

B. Intrinsic rocking curve – Pearson type VII approximation

In (Suhonen *et al.*, 2007; Diemoz *et al.*, 2010; Pavlov *et al.*; Arfelli *et al.*, 1998) the authors assume a bell-shaped rocking curve such as Guassian or Lorentzian. In (Hall *et al.*, 1977; Kitchen *et al.*, 2010; Kitchen *et al.*, 2007) the authors have determined experimentally that for the ABI system as in Figure 1 the rocking curve can be accurately modeled better using a Pearson type VII function as:

$$R(\theta) = c \left(1 + \frac{\theta^2}{ma^2} \right)^{-m}, \quad (27)$$

where in our experiments using three Si (333) crystals and for the incident photon energy of 40keV, $a = 0.7146$, and $m = 2.3737$, and of 60keV, $a = 0.5355$, and $m = 4.3005$. In both cases we assumed $c = 1.0$. Here c defines the curve amplitude, m governs the rate of decay

of the tails, and both a and m determine the profile width. This Pearson type VII function has limits of the Gaussian ($m = 1$), modified Lorentzian ($m = 2$) and Cauchy ($m = \infty$) distributions (Hall et al 1977). Given the variety of functions the Pearson type VII function encompasses, it can provide a superior fit to near-perfect crystal reflections rocking curves. The shapes of these rocking curves can be seen in Figure 3.

In Figure 3 we also indicated analyzer angular positions used in DEI (Chapman *et al.*, 1997) θ_L, θ_H and scatter-DEI (Rigon *et al.*, 2003) $\theta_{TOP}, \theta_{TOE}$ defined earlier.

C. Effect of the analyzer angular position

Examples of functions, defining the effect of the analyzer angular position θ_i on CRLB, as defined by Eqs (19)–(21) for four different object parameter vectors are given in Figure 4. The parameter vectors in use are extracted from two formalin-preserved human tissue samples (details are given in the experimental section): a human thumb (shown in Figure 2) and a whole breast. Each point at these functions shows the relative information at that analyzer angular position.

In short, if we desire to minimize CRLB (i.e. noise) in v_i parameter estimation, using L angular measurements, we should acquire these measurement at analyzer positions, $\Theta =$

$\{\theta_1, \theta_2, \dots, \theta_L\}$, so that $\sum_{i=1}^L V_i(\theta_i)$ is maximized.

In addition to $V_i(\theta)$ functions, Figure 4 shows analyzer angular positions as defined in DEI, θ_H, θ_L , and S-DEI, $\theta_{TOP}, \theta_{TOE}$. If we look at the $V_1(\theta)$ function of the absorption parameter in Figure 4 a) and b), we can see that the maximum amount of information is at $\theta = 0 \mu rad$ in both graphs. If we are concerned with the refraction parameter, we should look at $V_2(\theta)$ function for the average object AIP, where it is easy to observe that the maximum for different imaging setup and tissues combination occurs at different angular positions. For the thumb sample imaged at 40keV this is at $\theta = \pm 1.9 \mu rad$, and for the breast sample imaged at 60keV this is at $\theta = \pm 0.9 \mu rad$. Here it is worth noting that the “No object” $V_2(\theta)$ function would indicate θ_H, θ_L , as the best measuring positions. Finally, for the USAXS parameter we should consider the $V_3(\theta)$ function. Two graphs indicate three regions with almost equal importance for the thumb sample: this is at $\theta = 0$, and $\pm 2.9 \mu rad$ and for the breast tissue at $\theta = 0$, and $\pm 1.6 \mu rad$. Here it is worth noting that the “No object” $V_3(\theta)$ function would approximately give $\theta_{TOP}, \theta_{TOE}$ as the best measuring positions.

Next in Figure 4 c) and d) we show information functions for extreme object parameters. These graphs indicate that different regions in the object may provide a very different object AIP, and therefore, optimal set of analyzer angular positions. More sophisticated angular position selection methods that can take this variability into account are left for future research.

From the considerations above, it is clear that in order to optimize the set of analyzer angular positions, one needs to consider and know in advance: 1) which parameter or combination of parameters are of interest, 2) the tissue type, 3) tissue region of interest, 4) imaging energy and intrinsic system rocking curve.

Let us stress one additional result evident from Figure 4, that is, if one would optimize the analyzer angular positions only considering intrinsic system rocking curves (“No object” curves), suboptimal angular positions would be chosen. To correctly choose analyzer angular positions one needs to incorporate the object effect. To our knowledge, this paper is the first to point this out obviously. Other authors considered parametric image information content and quality based on the rocking curve only (Rigon *et al.*, 2003; Pagot *et al.*, 2005; Pagot *et al.*, 2003; Chapman *et al.*, 1997).

D. Experimental optimization of the number of analyzer angular measurements

The behavior of CRLB as we increase the number of angular measurements is given by Eq (24). As this equation indicates, we will use an average over the object parameters for the whole sample (thumb or breast), however it is important to note that different regions in the object may provide very different parameters (see Figure 4 for example), and therefore, results. For example, the structure’s boundaries will give much larger refraction than other points in the image. This means that very different results could be obtained in the various areas. Therefore, as we pointed out earlier, the results presented in this paper could be directly used only for a similar imaging system and tissue properties, however the present methodology can be used for optimizing any ABI system/imaging setup.

In the presented experiments we will consider a uniform set of analyzer angular positions in the range of $[-5, 5]\mu rad$ for thumb tissue and $[-2.4, 2.4]\mu rad$ for breast tissue (see Eq. (22) and Figure 7a) and b) to visualize the proposed set of analyzer angular positions).

First, from Figure 5, we can see the improvement in average CRLB as we increase the number of analyzer angular measurements. We can observe that for lower numbers of angular measurements, the attenuation and USAXS parameter bounds are higher for L even numbers than for L odd numbers. This is because the maximum amount of information for attenuation and USAXS, as measured by $V_1(\theta)$ and $V_3(\theta)$ (see Figure 4), is at $\theta = 0 \mu rad$, which is not sampled when using an even number of measurements. On the other hand, measuring at $\theta = 0 \mu rad$ does not contain any information about refraction (see $V_2(\theta)$ in Figure 4) and that is why the average CRLB for the refraction image lies on almost the same curve for odd and even numbers of angular measurements.

Another observation that can be made from Figure 5 is that after a certain number of angular measurements (11 for both data sets), there is almost no improvement or loss in the average CRLB bounds. *Therefore, in the case of a Minimum Variance Unbiased (MVU) estimator, there is neither benefit nor loss in acquiring more than 11 images at uniformly spaced analyzer angular positions for a fixed total radiation dose.*

At this point we would like to stress that the Cramér-Rao lower bound (CRLB) is used to quantify the noise properties in parametric images of an unbiased estimator and defines the best noise performance that one can obtain regardless of which method is used to estimate ABI parametric images.

E. Experimental optimization of the analyzer angular positions

Here we will investigate the possibility of lowering the CRLB and hence achieving better estimation performance by using a set of non-uniform analyzer angular positions Θ^* according Eq. (25). For Θ_u , we used a uniform angular positions in the same range of $[-5,5]\mu\text{rad}$ and $[-2.4,2.4]\mu\text{rad}$ as before depending on the tissue type.

To optimize the objective function in (26), we used the exhaustive search method. For every number of angular measurements, we evaluated $J(\Theta)$ at every possible combination of analyzer angular positions (with minimum spacing of $0.2\mu\text{rad}$) and we defined the best set of positions as the one that has the smallest value of $J(\Theta)$.

In order to decrease computation time, we limited the number of possible patterns by imposing two constraints: We always measure at $\theta = 0$, and our angular-sampling positions are always symmetrical around zero since $V_i(\theta_l)$, on average, is always symmetrical around zero. We selected an exhaustive search method for its simplicity, and also since it guarantees a globally optimal solution. The needed computational time needed for this optimization was 8 hours on a iMAC with Intel Core i5 2.9 GHz with 16 GB RAM. Other gradient type of optimization methods (Chong and Ak, 2008) may be used in future.

The resulting average CRLBs of optimal non-uniform and uniform set of analyzer angular positions are shown in Figure 6. One can observe that the attenuation and refraction CRLBs can be significantly improved over uniform angular positions while the USAXS parameter CRLB shows minor improvement. The presented results indicate that one can reduce estimator variance considerably by using a non-uniform set of analyzer angular positions that is optimized for the tissue of interest and a given ABI system.

Figure 6 also shows CRLBs for analyzer angular positions as proposed in the DEI and S-DEI methods. Here we can see that for a weakly scattering and refracting object, like breast tissue, DEI and S-DEI angular positions seem to achieve optimal CRLB as one would expect bearing in mind the curves shown in Figure 4 b) and d).

The key result of this section is that only three angular measurements are required to get the best parametric images. As before, the second intriguing observation is that the successive angular measurements only spread the total dose to the additional measurements *without improving or worsening CRLB*.

The optimal set of analyzer angular positions for different numbers of angular positions and the two tissue samples are shown in Figure 7. As we can see, the angular measurements tend to converge to three dominant angular positions. These positions approximately correspond to the locations with the maximum information measured by $V_i(\theta_l)$, for the averaged AIP, in Figure 4.

5. DEI, S-DEI and MIR estimator comparison to CRLB

The graphs in Figure 5 and Figure 6 can be used as a benchmark, that is a lower bound, to evaluate the performance of ANY estimator. In this section, we will investigate the performance of the multiple-image radiography estimator (MIR) as described in (Brankov *et*

al., 2006), DEI (Chapman *et al.*, 1997) and scatter-DEI (Rigon *et al.*, 2003) with respect to CRLB. The presented evaluation methodology can be directly used on other estimation methods like extended-DEI (Chou *et al.*, 2007; Rigon *et al.*, 2007) and model based MIR (Majidi *et al.*, 2008a).

Specifically we will evaluate the MIR, DEI and S-DEI estimators by calculating estimators' variance, squared bias, and Mean Square Error (MSE) and comparing it to CRLB. The purpose of this section is not to fully evaluate MIR, DEI or S-DEI but it is merely an example to illustrate the proposed evaluation procedure.

It is worth noting that the images estimated using the DEI and S-DEI method are prone to artifacts that are, most of the time, mitigated in the MIR method (as shown in (Wernick *et al.*, 2003)). Similar methods to MIR have also been proposed independently in (Pagot *et al.*, 2003) and (Rigon *et al.*, 2003). Other ABI methods like extended-DEI (Chou *et al.*, 2007; Rigon *et al.*, 2007) use different assumptions about image formation, different number and location of the angular measurements.

A. Simulated data

To evaluate the estimation methods' variance, squared bias, and Mean Square Error (MSE) we need to use simulated data for which true values of the parametric images are known. First, parametric images are estimated using experimental data described in the *Experimental study* section. Note that these raw images are acquired by angular oversampling and under high beam intensity, and as such, they result in accurate, low noise and bias, parametric images. Now we use these parametric images as true values to simulate AIPs according to the image formation model as given in (3). After simulating the angular sampling, we simulate the effect of noise by generating Poisson distributed values, i.e.

$$g[l] \sim \text{Poisson} \{g(\theta_l)\}, \quad (28)$$

where $\sim$ denotes distributed according and we assume that $I_0 = 1800$ photons, which is a reasonable number of photons. To reduce statistical variations we generated 1000 independent noise realizations. Note that, according CRLB equation given in (13), the number assumed photons will only change graphs scale and not general conclusions.

B. Image acquisition using a uniform set of analyzer angular positions

The results for the human thumb study are shown in Figure 8 a). We can see that DEI and SDEI estimator are highly biased. The same can be said for MIR for a low number of angular measurements (less than 7) where MIR occasionally even has variance lower than the average CRLB. However after 11 angular measurements, the bias of the estimation becomes negligible compared to the variance and MIR becomes approximately, for tested tissue, an MVU estimator, meaning that it achieves the best performance possible as defined by the CRLB for an unbiased estimator. Similar results are obtained for the breast study, which are shown in Figure 8 b).

Note that in (Huang *et al.*, 2007; Diemoz *et al.*, 2010) the authors showed that if the object scatter and refraction is sufficiently larger MIR methods still could have unwanted artifacts

and, as such, may not becoming MVU. The origin of this is probably two fold: 1) small range of the analyzer angular-sampling and 2) insufficient angular-sampling density. This consideration is left for future research.

In order to evaluate the above results visually, we presented images for six different numbers of angular measurements in Figure 9. Visual inspection of images also confirms that 11 angular measurements seem sufficient for good parametric image estimation.

Note that the above concepts of bias, variances or MSE are hard to demonstrate visually by the images as in Figure 9, as for example a potential bias is likely not to be noticed by the eye. However, since images are usually evaluated visually, their visual appearance is important.

One should note that we did not specify which estimator could achieve average CRLB for a non-uniform set of analyzer angular positions. The MIR estimator in (Brankov *et al.*, 2006) is designed for a uniform set of analyzer angular positions and cannot be directly applied to a nonuniform set of positions. The problem of finding an MVU estimator for the non-uniform set of analyzer angular positions is left for future research.

6. Conclusion

In this work we derived the Cramér-Rao lower bound (CRLB) for analyzer-based phase-contrast imaging using the image formation model based on radiative transfer theory (Khelashvili *et al.*, 2006). The CRLB is the minimum bound for the variance of *any unbiased estimator* and defines the best noise performance that one can obtain. We analyzed a situation in which the total exposure dose to the imaged object remains fixed regardless of the number of angular measurements. The results in this paper show that for a minimum-variance unbiased (MVU) estimator, only a limited number of angular measurements (eleven for uniform and three for optimal non-uniform set of analyzer angular positions) are required to get the best parametric images. The second intriguing observation is that the successive angular measurements only spread the total radiation dose to the additional measurements without improving or worsening the CRLB, however the added measurements usually can improve parametric images by reducing estimation bias.

The presented analysis is performed using data from two tissue samples. The findings obtained for these two particular tissue samples and imaging energies cannot be generalized, as the results are dependent on the tissue type, on the crystal reflection, energy, analyzer angular-sampling range, etc. Very different results could, in principle, be obtained in other conditions. For example, tissue with very large refraction and scattering angles could require a larger number of angular measurement steps. However the purpose of this work is not to fully evaluate any particular ABI technique or to prove conclusively that any methods or a set of analyzer angular positions will always perform the best. This study is merely an example to illustrate the proposed procedure for a given imaging system setup and tissue type.

We also showed that it is possible to minimize average CRLB (average over the whole image) further by deforming the set of analyzer angular positions from uniform. However,

the problem of finding an MVU for the non-uniform analyzer angular positions is still an open research problem.

Last, we evaluated the MIR (Brankov *et al.*, 2006) DEI (Chapman *et al.*, 1997) and scatter-DEI (Rigon *et al.*, 2003) estimators. MIR becomes an MVU estimator, which achieves average CRLB, when more than 11 uniformly spaced angular measurements are used.

Acknowledgments

This research was supported by NIH/NIBIB grant EB009715, NIH/NCI grants CA111976 and NIH/NHLBI grants HL091017.

Appendix

A. Approximation of the Object Function

Here we review the image formation model proposed in (Khelashvili *et al.*, 2006). The image formation model is based on radiative transfer theory and the object is modeled as a stratified medium where the stratum is characterized by a linear absorption coefficient $\mu(r) = \mu(z)$ and by a spatially varying refractive index $n(r) = n_0 + n_y(z)y$ where $r = (x, y, z)^T$ is the spatial coordinate of the system and the z coordinate is the direction of X-ray beam propagation. Furthermore, the object is assumed to have identical scattering centers distributed in the medium with the density $\rho_n(r) = \rho_n(z)$. Each scatter center is characterized by its extinction cross section σ_{ext} , which is defined as the sum of the scattering cross section σ_s and the absorption cross section σ_a . Next, we use a quantity called *albedo* (Khelashvili *et al.*, 2006), defined as

$$W_0 \triangleq \frac{\sigma_s}{\sigma_a + \sigma_s}. \quad (\text{A-1})$$

And for each scattering center we also define:

$$\alpha_p = \left[4\delta^2 \ln \left(\frac{2}{\delta} + 1 \right) \right]^{-1}, \quad (\text{A-2})$$

where α_p is the beam broadening due to a single scattering center (von Nardroff, 1926) and δ is the difference between the refractive indices of the medium and the scatters. Now, by using the equation presented in (Khelashvili *et al.*, 2006), we can write the object function for this stratum as

$$f(\theta) = e^{-\int_0^z \bar{\mu}(\tau) d\tau} \sum_{k=0}^{\infty} \left[\frac{(\tau(Z)W_0)^k}{k!} e^{-\tau(Z)W_0} \sqrt{\frac{\alpha_p}{k\pi}} e^{-\frac{\alpha_p}{k}(\theta - \Delta\theta)^2} \right], \quad (\text{A-3})$$

where $\bar{\mu}(z) \triangleq \rho_n(z)\sigma_a + \mu(z)$, and

$$\tau(Z) = \int_0^Z (\rho_n(z)\sigma_{ext} + \mu(z)) dz, \quad (A-4)$$

$$\Delta\theta(\vec{\rho}) = \int_0^Z \frac{\partial}{\partial y} \ln(n(x, y, z)) dz = \int_0^Z \frac{n_y(z)}{n_0} dz, \quad (A-5)$$

where Z is the thickness of the medium. We can relate the terms in the summation of Eq. (A-3) to an expectation of a Gaussian curve with k/α_p variance over a Poisson probability distribution function of k . Interestingly, the summation in Eq (A-3) can be derived by statistical argument in which the random distribution of scatterers is modeled by the Poisson process. *Here is it interesting to note that the full model is defined by eight model parameters $\sigma_s, \sigma_a, \rho_n, \mu, \alpha_p, n_0, n_y, Z$.*

For large values of $\tau(Z)W_0$, we can accurately approximate this object function by the dominant term that is a Gaussian curve, where $k = \tau(Z)W_0$. Then we can rewrite (A-3) in the following form:

$$f(\theta; \nu) = \exp(-\nu_1) \frac{1}{\sqrt{2\pi\nu_3}} \exp\left(-\frac{(\theta - \nu_2)^2}{2\nu_3}\right), \quad (A-6)$$

where:

$$\nu_1 = \int_0^Z (\rho_n(z)\sigma_a + \mu(z)) dz, \quad (A-7)$$

$$\nu_2 = \int_0^Z \frac{n_y(z)}{n_0} dz, \quad (A-8)$$

$$\nu_3 = \frac{1}{2} \frac{\tau(Z)W_0}{\alpha_p} = \frac{1}{2} \int_0^Z \frac{\rho_n(z)\sigma_s}{\alpha_p} dz. \quad (A-9)$$

and ν_1 describes the net linear absorption of the medium (including absorption in scattering centers), ν_2 is the scaled refractive index gradient, and ν_3 can be interpreted as *USAXS* parameter of the medium.

B. Derivation of Cramér-Rao Lower Bounds

From (8) we have:

$$\{\mathbf{I}(\boldsymbol{\nu}; \boldsymbol{\Theta})\}_{i,j} = -E \left[\frac{\partial^2}{\partial \nu_i \partial \nu_j} \ln p(\mathbf{g}; \boldsymbol{\nu}) \right]. \quad (\text{B-1})$$

The log-likelihood function of the data is given by

$$\ln p(\mathbf{g}; \boldsymbol{\nu}) = \sum_{l=1}^L [g[l] \ln g_l - g_l - \ln(g[l]!)] . \quad (\text{B-2})$$

By taking the derivatives we have:

$$\frac{\partial}{\partial \nu_i} \ln p(\mathbf{g}; \boldsymbol{\nu}) = \sum_{l=1}^L \left[\left(\frac{g[l]}{g_l} - 1 \right) \frac{\partial g_l}{\partial \nu_i} \right], \quad (\text{B-3})$$

$$\frac{\partial^2}{\partial \nu_i \partial \nu_j} \ln p(\mathbf{g}; \boldsymbol{\nu}) = \sum_{l=1}^L \left[\left(\frac{g[l]}{g_l} - 1 \right) \frac{\partial^2 g_l}{\partial \nu_i \partial \nu_j} - \frac{g[l]}{g_l^2} \frac{\partial g_l}{\partial \nu_i} \frac{\partial g_l}{\partial \nu_j} \right]. \quad (\text{B-4})$$

By taking the expected value of (B-4) and substituting (3) and (4) into the result, noticing that $g_l \triangleq E[g[l]]$, we get:

$$\mathbf{I}(\boldsymbol{\nu}; \boldsymbol{\Theta}) = \sum_{l=1}^L \mathbf{I}(\boldsymbol{\nu}; \theta_l), \quad (\text{B-5})$$

$$\mathbf{I}(\boldsymbol{\nu}; \theta_l) = \begin{bmatrix} g_l & \frac{\nu_2 g_l - h_l}{\nu_3} & \frac{\nu_3 g_l - q_l}{2\nu_3^2} \\ \frac{\nu_2 g_l - h_l}{\nu_3} & \frac{(\nu_2 g_l - h_l)^2}{\nu_3^2 g_l} & \frac{(\nu_2 g_l - h_l)(\nu_3 g_l - q_l)}{2\nu_3^3 g_l} \\ \frac{\nu_3 g_l - q_l}{2\nu_3^2} & \frac{(\nu_2 g_l - h_l)(\nu_3 g_l - q_l)}{2\nu_3^3 g_l} & \frac{(\nu_3 g_l - q_l)^2}{4\nu_3^4 g_l} \end{bmatrix}, \quad (\text{B-6})$$

where:

$$g_l \triangleq E[g[l]] = \left[\frac{I_0}{L} R(\theta)^* f(\theta; \boldsymbol{\nu}) \right] \Big|_{\theta=\theta_l}, \quad (\text{B-7})$$

$$h_l \triangleq \left[\frac{I_0}{L} R(\theta)^* \theta f(\theta; \boldsymbol{\nu}) \right] \Big|_{\theta=\theta_l}, \quad (\text{B-8})$$

$$q_l \triangleq \left[\frac{I_0}{L} R(\theta)^* (\theta - \nu_2)^2 f(\theta; \boldsymbol{\nu}) \right] \Big|_{\theta=\theta_l}. \quad (\text{B-9})$$

Alternatively, we can rewrite (B-6) in normalized form as:

$$\mathbf{I}(\boldsymbol{\nu};\theta_l) = \frac{I_0}{L} \mathbf{I}'(\boldsymbol{\nu};\theta_l), \quad (\text{B-10})$$

where $\mathbf{I}'(\boldsymbol{\nu};\theta_l) = \mathbf{a}_l \mathbf{a}_l^T$, and

$$\mathbf{a} = \begin{bmatrix} \sqrt{g_l} \nu_2 g'_l - h'_l \nu_3 g'_l - q'_l \\ \nu_3 \sqrt{g'_l} \quad 2\nu^2_3 \sqrt{g'_l} \end{bmatrix}^T, \quad (\text{B-11})$$

where

$$g'_l \triangleq [R(\theta)^* \theta f(\theta; \boldsymbol{\nu})] \Big|_{\theta=\theta_l}, \quad (\text{B-12})$$

$$h'_l \triangleq [R(\theta)^* \theta f(\theta; \boldsymbol{\nu})] \Big|_{\theta=\theta_l}, \quad (\text{B-13})$$

$$q'_l \triangleq [R(\theta)^* (\theta - \nu_2)^2 f(\theta; \boldsymbol{\nu})] \Big|_{\theta=\theta_l}. \quad (\text{B-14})$$

References

- Anastasio, MA.; La Riviere, P. Emerging imaging technologies in medicine. Boca Raton, FL: CRC Press; 2012.
- Arfelli F, Assante M, Bonvicini V, Bravin A, Cantatore G, Castelli E, Dalla Palma L, Di Michiel M, Longo R, Olivo A, Pani S, Pontoni D, Poropat P, Prest M, Rashevsky A, Tromba G, Vacchi A, Vallazza E, Zanconati F. Low-dose phase contrast x-ray medical imaging. *Phys Med Biol.* 1998; 43:2845–2852. [PubMed: 9814522]
- Bonse U, Hart M. Tailless X-ray single-crystal reflection curves obtained by multiple reflection. *Appl. Phys. Lett.* 1965; 7:238.
- Brankov JG, Wernick MN, Yang YY, Li J, Muehleman C, Zhong Z, Anastasio MA. Computed tomography implementation of multiple-image radiography. *Med Phys.* 2006; 33:278–289. [PubMed: 16532932]
- Brankov, JG.; Zysk, AM. Emerging Imaging Technologies in Medicine. Anastasio, MA.; Riviere, PL., editors. Taylor & Francis; 2012. p. 368
- Bravin A, Coan P, Suortti P. X-ray phase-contrast imaging: from pre-clinical applications towards clinics. *Phys Med Biol.* 2013; 58:R1–R35. [PubMed: 23220766]
- Chapman D, Nesch I, Hasnah MO, Morrison TI. X-ray optics for emission line X-ray source diffraction enhanced systems. *Nucl Instrum Meth A.* 2006; 562:461–467.
- Chapman D, Thomlinson W, Johnston RE, Washburn D, Pisano E, Gmur N, Zhong Z, Menk R, Arfelli F, Sayers D. Diffraction enhanced x-ray imaging. *Phys Med Biol.* 1997; 42:2015–2025. [PubMed: 9394394]
- Chong, EKP.; ak, SH. An introduction to optimization. Hoboken, N.J.: Wiley-Interscience; 2008.
- Chou CY, Anastasio MA, Brankov JG, Wernick MN, Brey EM, Connor DM, Zhong Z. An extended diffraction-enhanced imaging method for implementing multiple-image radiography. *Phys Med Biol.* 2007; 52:1923–1945. [PubMed: 17374920]
- Coan P, Bravin A, Tromba G. Phase-contrast x-ray imaging of the breast: recent developments towards clinics. *J Phys D Appl Phys.* 2013; 46

- Coan P, Gruener F, Glaser C, Schneider T, Bravin A, Reiser M, Habs D. Phase contrast medical imaging with compact X-ray sources at the Munich-Centre for Advance Photonics (MAP). *Nucl Instrum Meth A*. 2009; 608:S44–S46.
- Connor, DM.; Pisano, E.; Zhong, Z.; Parham, CA.; Hinshaw, W.; Cole, E. Radiological Society of North America. Chicago: 2011. Second Generation Diffraction Enhanced Imaging Prototype: Progress towards a Clinical Imaging System.
- Diemoz PC, Coan P, Glaser C, Bravin A. Absorption, refraction and scattering in analyzer-based imaging: comparison of different algorithms. *Opt Express*. 2010; 18:3494–3509. [PubMed: 20389359]
- Faulconer L, Parham C, Connor DM, Zhong Z, Kim E, Zeng DL, Livasy C, Cole E, Kuzmiak C, Koomen M, Pavic D, Pisano E. Radiologist Evaluation of an X-ray Tube-Based Diffraction-Enhanced Imaging Prototype Using Full-Thickness Breast Specimens. *Acad Radiol*. 2009; 16:1329–1337. [PubMed: 19596593]
- Faulconer LS, Parham CA, Connor DM, Kuzmiak C, Koomen M, Lee Y, Cho KR, Rafoth J, Livasy CA, Kim E, Zeng DL, Cole E, Zhong Z, Pisano ED. Effect of Breast Compression on Lesion Characteristic Visibility with Diffraction-Enhanced Imaging. *Acad Radiol*. 2010; 17:433–440. [PubMed: 20036586]
- Hall MM Jnr, Veeraraghavan VG, Rubin H, Winchell PG. The approximation of symmetric X-ray peaks by Pearson type VII distributions. *J Appl Crystallogr*. 1977; 10:66–68.
- Honnicke MG, Manica J, Mazzaro I, Cusatis C, Huang XR. Source effects in analyzer-based X-ray phase contrast imaging with conventional sources. *Rev Sci Instrum*. 2012; 83
- Huang ZF, Kang KJ, Yang YG. Extraction methods of phase information for X-ray diffraction enhanced imaging. *Nucl Instrum Meth A*. 2007; 579:218–222.
- Ishimaru, A. Wave propagation and scattering in random media. New York Oxford; New York: IEEE Press; Oxford University Press; 1997.
- Kay, SM. Fundamentals of statistical signal processing. Englewood Cliffs, N.J.: Prentice-Hall PTR; 1993.
- Khelashvili G, Brankov JG, Chapman D, Anastasio MA, Yang YY, Zhong Z, Wernick MN. A physical model of multiple-image radiography. *Phys Med Biol*. 2006; 51:221–236. [PubMed: 16394335]
- Kitchen MJ, Paganin DM, Uesugi K, Allison BJ, Lewis RA, Hooper SB, Pavlov KM. X-ray phase, absorption and scatter retrieval using two or more phase contrast images. *Opt Express*. 2010; 18:19994–20012. [PubMed: 20940891]
- Kitchen MJ, Paganin DM, Uesugi K, Allison BJ, Lewis RA, Hooper SB, Pavlov KM. Phase contrast image segmentation using a Laue analyser crystal. *Phys Med Biol*. 2011; 56:515–534. [PubMed: 21212473]
- Kitchen MJ, Pavlov KM, Siu KKW, Menk RH, Tromba G, Lewis RA. Analyser-based phase contrast image reconstruction using geometrical optics. *Phys Med Biol*. 2007; 52:4171–4187. [PubMed: 17664601]
- Lider VV, Kovalchuk MV. X-ray phase-contrast methods. *Crystallogr Rep+*. 2013; 58:769–787.
- Liu Y, Nelson J, Holzner C, Andrews JC, Pianetta P. Recent advances in synchrotron-based hard x-ray phase contrast imaging. *J Phys D Appl Phys*. 2013; 46
- Majidi, K.; Brankov, JG.; Li, J.; Muehleman, C.; Wernick, MN. Parameter estimation in Multiple-Image Radiography; Nuclear Science Symposium Conference Record, 2008. NSS '08. IEEE; 2008a. p. 4214-4217.
- Majidi, K.; Brankov, JG.; Wernick, MN. Sampling strategies in multiple-image radiography; 5th IEEE International Symposium on Biomedical Imaging: From Nano to Macro, 2008; 2008b. p. 688-691.
- Marquet, B.; Brankov, JG.; Wernick, MN. Noise and sampling analysis for multiple-image radiography; Biomedical Imaging: Nano to Macro, 2006. 3rd IEEE International Symposium on; 2006. p. 1232-1235.
- Nesch I, Fogarty DP, Tzvetkov T, Reinhart B, Walus AC, Khelashvili G, Muehleman C, Chapman D. The design and application of an in-laboratory diffraction-enhanced x-ray imaging instrument. *Rev Sci Instrum*. 2009; 80:093702. [PubMed: 19791939]

- Pagot E, Cloetens P, Fiedler S, Bravin A, Coan P, Baruchel J, Hartwig J, Thomlinson W. A method to extract quantitative information in analyzer-based x-ray phase contrast imaging. *Appl Phys Lett*. 2003; 82:3421–3423.
- Pagot E, Fiedler S, Cloetens P, Bravin A, Coan P, Fezzaa K, Baruchel J, Hartwig J. Quantitative comparison between two phase contrast techniques: diffraction enhanced imaging and phase propagation imaging. *Phys Med Biol*. 2005; 50:709–724. [PubMed: 15773629]
- Pavlov KM, Kewish CM, Davis JR, Morgan MJ. A variant on the geometrical optics approximation in diffraction enhanced tomography. *J Phys D Appl Phys*. 2001; 34:A168–A172.
- Rigon L, Arfelli F, Menk RH. Generalized diffraction enhanced imaging to retrieve absorption, refraction and scattering effects. *J Phys D Appl Phys*. 2007; 40:3077–3089.
- Rigon L, Besch HJ, Arfelli F, Menk RH, Heitner G, Plathow-Besch H. A new DEI algorithm capable of investigating sub-pixel structures. *J Phys D Appl Phys*. 2003; 36:A107–A112.
- Suhonen H, Fernandez M, Bravin A, Keyrilainen J, Suortti P. Refraction and scattering of X-rays in analyzer-based imaging. *J Synchrotron Radiat*. 2007; 14:512–521. [PubMed: 17960034]
- Suortti P, Keyrilainen J, Thomlinson A. Analyser-based x-ray imaging for biomedical research. *J Phys D Appl Phys*. 2013; 46
- Vine DJ, Paganin DM, Pavlov KM, Krausslich J, Wehrhan O, Uschmann I, Forster E. Analyzer-based phase contrast imaging and phase retrieval using a rotating anode x-ray source. *Appl Phys Lett*. 2007; 91:254110.
- von Nardroff R. Refraction of X-Rays by Small Particles. *Physical Review*. 1926; 28:240–246.
- Wernick MN, Wirjadi O, Chapman D, Zhong Z, Galatsanos NP, Yang YY, Brankov JG, Oltulu O, Anastasio MA, Muehleman C. Multiple-image radiography. *Phys Med Biol*. 2003; 48:3875–3895. [PubMed: 14703164]
- Zhao YZ, Brun E, Coan P, Huang ZF, Sztrokay A, Diemoz PC, Liebhardt S, Mittone A, Gasilov S, Miao JW, Bravin A. High-resolution, low-dose phase contrast X-ray tomography for 3D diagnosis of human breast cancers. *P Natl Acad Sci USA*. 2012; 109:18290–18294.
- Zhou SA, Brahme A. Development of phase-contrast X-ray imaging techniques and potential medical applications. *Phys Medica*. 2008; 24:129–148.

Figure 1. Schematic diagram of ABI imaging system (not drawn to scale). In this paper, we investigate the effect of the analyzer position on the noise performance.

Figure 2. MIR projections of a human thumb. The images are acquired with the beam energy of 40keV at 25 analyzer positions ranging from -4.8 to 4.8 μrad with 0.4 μrad increments (see experimental sections for acquisition details).

Figure 3.
Example of a Rocking Curve for Si (333) Crystal at beam energy of (a) 40 keV ($FWHM = 1.4 \mu rad$); (b) 60 keV ($FWHM = 1.0 \mu rad$).

Figure 4.

Examples of information functions for parameter plotted on the logarithmic scale. Each function indicates the relative value of obtaining an angular measurement at a given angular analyzer position θ_i in regard to estimation of one of the object parameters. (a) For thumb sample imaged at 40 keV, (b) for breast data imaged at 60keV, in c) and d) we show information functions for extreme object parameters.

Figure 5.

CRLB for an unbiased estimator vs. number of angular measurements using a uniform set of analyzer angular positions. (a) For thumb sample imaged at 40 keV, (b) for breast data imaged at 60keV.

Figure 6. CRLB for an unbiased estimator vs. number of angular measurements for both optimal and uniform set of analyzer angular positions. (a) For thumb sample imaged at 40 keV, (b) for breast data imaged at 60keV.

Figure 7.

Analyzer angular positions in optimal and uniform set of analyzer angular positions for different number of angular measurements. (a) For thumb sample imaged at 40 keV, (b) for breast data imaged at 60keV.

Figure 8.

Variance, MSE, squared bias, and CRLB vs. number of angular measurements for optimized set of analyzer angular positions. (a) For thumb sample imaged at 40 keV, (b) for breast data imaged at 60keV.

Figure 9.

Estimated MIR images using uniform set of analyzer angular positions with different number of angular measurements. Observe that there is little or no improvement after $L>11$.