Low dose dynamic CT myocardial perfusion imaging using a statistical iterative reconstruction method

Yinghua Tao

Department of Medical Physics, University of Wisconsin-Madison, Madison, Wisconsin 53705

Guang-Hong Chen

Department of Medical Physics and Department of Radiology, University of Wisconsin-Madison, Madison, Wisconsin 53705

Timothy A. Hacker and Amish N. Raval

Department of Medicine, University of Wisconsin-Madison, Madison, Wisconsin 53792

Michael S. Van Lysel and Michael A. Speidel^{a)}

Department of Medical Physics and Department of Medicine, University of Wisconsin-Madison, Madison, Wisconsin 53705

(Received 16 December 2013; revised 20 May 2014; accepted for publication 1 June 2014; published 24 June 2014)

Purpose: Dynamic CT myocardial perfusion imaging has the potential to provide both functional and anatomical information regarding coronary artery stenosis. However, radiation dose can be potentially high due to repeated scanning of the same region. The purpose of this study is to investigate the use of statistical iterative reconstruction to improve parametric maps of myocardial perfusion derived from a low tube current dynamic CT acquisition.

Methods: Four pigs underwent high (500 mA) and low (25 mA) dose dynamic CT myocardial perfusion scans with and without coronary occlusion. To delineate the affected myocardial territory, an N-13 ammonia PET perfusion scan was performed for each animal in each occlusion state. Filtered backprojection (FBP) reconstruction was first applied to all CT data sets. Then, a statistical iterative reconstruction (SIR) method was applied to data sets acquired at low dose. Image voxel noise was matched between the low dose SIR and high dose FBP reconstructions. CT perfusion maps were compared among the low dose FBP, low dose SIR and high dose FBP reconstructions. Numerical simulations of a dynamic CT scan at high and low dose (20:1 ratio) were performed to quantitatively evaluate SIR and FBP performance in terms of flow map accuracy, precision, dose efficiency, and spatial resolution.

Results: For *in vivo* studies, the 500 mA FBP maps gave -88.4% , -96.0% , -76.7% , and -65.8% flow change in the occluded anterior region compared to the open-coronary scans (four animals). The percent changes in the 25 mA SIR maps were in good agreement, measuring -94.7% , -81.6% , -84.0% , and -72.2% . The 25 mA FBP maps gave unreliable flow measurements due to streaks caused by photon starvation (percent changes of $+137.4\%$, $+71.0\%$, -11.8% , and -3.5%). Agreement between 25 mA SIR and 500 mA FBP global flow was -9.7% , 8.8% , -3.1% , and 26.4% . The average variability of flow measurements in a nonoccluded region was 16.3% , 24.1% , and 937.9% for the 500 mA FBP, 25 mA SIR, and 25 mA FBP, respectively. In numerical simulations, SIR mitigated streak artifacts in the low dose data and yielded flow maps with mean error $<7\%$ and standard deviation $<9\%$ of mean, for 30×30 pixel ROIs (12.9×12.9 mm²). In comparison, low dose FBP flow errors were -38% to $+258\%$, and standard deviation was 6% – 93% . Additionally, low dose SIR achieved 4.6 times improvement in flow map CNR² per unit input dose compared to low dose FBP.

Conclusions: SIR reconstruction can reduce image noise and mitigate streaking artifacts caused by photon starvation in dynamic CT myocardial perfusion data sets acquired at low dose (low tube current), and improve perfusion map quality in comparison to FBP reconstruction at the same dose.

© 2014 American Association of Physicists in Medicine. [<http://dx.doi.org/10.1118/1.4884023>]

Key words: CT, myocardial perfusion, statistical iterative reconstruction, low dose

1. INTRODUCTION

Patients with stable atypical chest pain and intermediate probability of coronary artery disease (CAD) present a difficult diagnostic challenge. Coronary CT angiography (CTCA) is capable of accurately ruling out CAD based on the absence of stenosis,^{1,2} however, the functional significance of a stenosis found by CTA may be unclear. Case-by-case variations in myocardial perfusion reserve for a given percent stenosis can result from fine details of the stenosis geometry, vessel collateralization, prior myocardial infarction, and the presence of diffuse disease, among other factors.^{3–6} Furthermore, stenosis severity may be inaccurately graded due to coronary calcium or motion artifacts.⁷ Myocardial perfusion imaging

sis found by CTA may be unclear. Case-by-case variations in myocardial perfusion reserve for a given percent stenosis can result from fine details of the stenosis geometry, vessel collateralization, prior myocardial infarction, and the presence of diffuse disease, among other factors.^{3–6} Furthermore, stenosis severity may be inaccurately graded due to coronary calcium or motion artifacts.⁷ Myocardial perfusion imaging

(e.g., SPECT) can identify functionally significant CAD and is widely employed in the noninvasive workup of CAD.^{8,9} However, radionuclide uptake images lack anatomic detail. When a perfusion defect is found, there may be a desire to confirm the result and determine whether the functional reduction is due to a treatable stenosis. A noninvasive technique that provides both function and anatomy would be desirable.

Recently, it has been shown that the hemodynamic significance of a stenosis may be investigated using CT myocardial perfusion imaging (CT-MPI).^{10–12} A CT-MPI scan may be either a single static image volume acquired at some time point after contrast injection,^{13–18} or a dynamic scan in which a series of volumes is acquired as the contrast agent flows through the myocardium.^{10–12,19–21} The dynamic approach has potential to provide quantitative flow, but also represents a potentially large increment (~ 10 mSv) to total CT exam dose.^{10–12} Effective dose in CTCA has been reduced to 2–4 mSv through the application of prospectively gated step-and-shoot scanning, or whole-heart scanning.^{22–24}

Conventional means for minimizing dose include use of prospective ECG gating, optimization of kVp, and lowering the mAs to achieve the maximum tolerable noise level. Advanced reconstruction and image processing methods that reduce the noise compared to conventional CT reconstruction offer a way to operate at even lower tube current, while maintaining noise levels. A variety of methods have been proposed in recent years. Image processing methods such as bilateral filtering,^{25,26} wavelet filtering,^{27–31} and anisotropic diffusion,^{32–34} are designed to reduce image noise while preserving edge information. These are straightforward to implement, however they do not correct existing image artifacts arising from low dose scanning, in particular noise streaks caused by photon starvation. Alternatively, iterative CT reconstruction methods which operate on the projection data and which account for the statistics in the different projection measurements have potential to both reduce noise and avoid streak artifacts.^{35–37}

In this work, we apply statistical iterative reconstruction (SIR) to low-dose (i.e., low tube current) dynamic CT myocardial perfusion imaging. The SIR method reconstructs an image consistent with data and smoothness constraints, while penalizing the contributions of noisy projections. In a previous study, this low dose SIR method was shown to reduce spatial nonuniformity in image noise, lower overall image noise compared to filtered back projection (FBP), and improve correspondence between low dose and high dose perfusion metrics derived from time-attenuation curves.³⁷ In the present work, we focus the ability of SIR to improve the quality of CT-derived parametric maps of blood flow, volume, and transit time when using a low tube current scan. Simulations are used to evaluate perfusion map accuracy and precision, evaluate dose efficiency, and to examine the trade-off between map noise and spatial resolution. Imaging in an animal model of coronary occlusion is used to evaluate the reduction of streaking artifacts and the detectability of regional flow reduction, in comparison to low-mA FBP imaging.

2. METHODS

2.A. Statistical iterative reconstruction method

The SIR method used in this work is based on the quadratic approximation to the maximum *a posteriori* estimation^{37–40}

$$\operatorname{argmin}_{\mathbf{x}} \left[\frac{\lambda}{2} (\mathbf{Ax} - \mathbf{y})^T \mathbf{D} (\mathbf{Ax} - \mathbf{y}) + R(\mathbf{x}) \right], \quad (1)$$

where $\mathbf{x}$ is a reconstructed image, $\mathbf{A}$ is the system matrix that describes CT data acquisition in a pixel representation, $\mathbf{y}$ is the measured projection data, and $\mathbf{D} = \operatorname{diag}\{d_1, d_2, \dots\}$ is a diagonal matrix with elements d_i that are equal to the inverse of the noise variance in the projection data. Each inverse noise variance value was estimated by setting d_i equal to the photon counts measured in the corresponding detector element. The regularizer $R(\mathbf{x})$ is used to incorporate the prior knowledge of the target image object such as the noise level and smoothness of the image. A variety of regularizers are being investigated in the literature and different regularizers may be used to highlight different features in the target image, but there is no consensus on which regularizer is the best. In this work, we selected the total variation (TV) of the image as an example regularizer, to facilitate the presentation and discussion of the paper. Other regularizers may also be used to generate similar results. The weighting parameter λ controls the relative weight between the data fidelity term and the regularizer term in Eq. (1). In general, a larger λ value gives higher weight to the data term, and the reconstructed image has higher noise and greater sharpness. The optimization problem in Eq. (1) was solved using a nonlinear conjugate gradient method.^{41,42} For the dynamic CT-MPI scan, the above procedure was applied for each cardiac cycle using a shortscan projection dataset from each cycle. Note that since the method does not require sharing of data across cardiac cycles, it would also be applicable to static CT-MPI.

2.B. Numerical simulations

2.B.1. Study setup

Numerical simulations of a simple dynamic phantom were performed to evaluate the relative performance of SIR and FBP reconstruction in generating parametric maps of myocardial perfusion. The phantom geometry [Fig. 1(a)] was designed to mimic the porcine model used in the animal studies of Sec. 2.C. Elliptical and cylindrical structures representing healthy myocardium, ischemic myocardium, left ventricle, lung, soft tissue, and bone were simulated. Contrast dynamics of the healthy myocardium, ischemic myocardium, and left ventricle were represented by gamma variate functions [Fig. 1(b)]. The simulated CT scanner geometry had a 949 mm source-to-detector distance, 541 mm source-to-isocenter distance, 55° fan angle, 888 detector elements, 984 views per rotation, and a rotation time of 0.4 s.

Ideal (noise free) projection data were generated using analytical ray-tracing. Projections with quantum noise were then produced by sampling from a Poisson distribution, where the mean of the photon counts distribution for a ray was obtained from the noise-free simulation. Next, to incorporate

FIG. 1. Phantom and contrast dynamics for ventricle, healthy myocardium and ischemic myocardium: (a) Dynamic phantom, (b) programmed time attenuation curves. The major semiaxis of the body ellipse was 185 mm, and minor semiaxis was 130 mm. Contrast dynamics were in the form of gamma variate functions. t_0 was 5 s for all curves. t_{peak} was 16 s for left ventricle, 19 s for the healthy myocardium, and 29 s for the two ischemic myocardium regions. Maximum values were 400, 80, 56, and 24 HU for the ventricle, healthy myocardium, ischemic myocardium (left), and ischemic myocardium (anterior), respectively.

the effects of negative values and dark noise that can arise from electronic noise at low detector exposure, Gaussian-distributed random values were added to the projection data. Simulations were performed at both high and low doses, with a flux ratio that matched the 20:1 tube current ratio used in the animal study. The high dose and low dose simulations used 1.0×10^6 photons per ray (“1000k”) and 5.0×10^4 photons/ray (“50k”), respectively. The high flux level was selected to produce a voxel noise in the FBP images (37 HU) similar to that measured in high dose animal study FBP images (35 HU). The width of the electronic noise Gaussian distribution was adjusted such that in the low-dose simulation, the number of nonpositive transmission values approximately matched the number of artifactually high ray-sums observed in the low dose (25 mA) experimental data, as evidenced by spikes in the sinogram. The standard deviation of the Gaussian distribution was 15 photons/ray. When nonpositive values were encountered, a low transmission value was assigned (1.5×10^{-8}) to the detector element, similar to the behavior observed in the experimental data.

2.B.2. Image reconstruction and data analysis

The simulated low and high dose scans were each 50 s long, enabling reconstruction of 50 cardiac cycles/scan with a time interval of 1 s between images. FBP reconstructions were first performed for both low and high dose datasets using the Ram-Lak filter. Then SIR [Eq. (1)] was applied to the low dose datasets with the parameter λ selected to achieve equal or lower voxel noise level than was measured in the myocardium region of the high dose FBP reconstruction. For both FBP and SIR images, short-scan reconstructions were performed. The reconstructed image dimensions were 44×44 cm (1024×1024 pixels), and the pixel width was 0.43 mm.

Baseline subtracted time attenuation curves (TACs) were constructed from the reconstructed time series images. Baseline subtraction was accomplished by subtracting the mean value of the first five precontrast beats. Perfusion maps were calculated by applying the following equations on a pixel by

pixel basis:

$$\text{Blood volume} = \frac{\int_{t_0}^T C(t)dt}{\int_{t_0}^T C_a(t)dt}, \quad (2)$$

$$\text{Transit time} = \frac{\int_{t_0}^T C(t)t dt}{\int_{t_0}^T C(t)dt}, \quad (3)$$

$$\text{Blood flow} = \frac{\text{Blood volume}}{\text{Transit time}}, \quad (4)$$

where C is the baseline-subtracted time attenuation curve measured in the myocardium, C_a is the baseline-subtracted time attenuation curve measured in the ventricle, T is total imaging time, and t_0 is contrast arrival time (defined as the time point with 10% of the peak enhancement).

These relatively simple calculation methods were used here mainly for two reasons. First, the equations are straightforward and free of complicating factors, such as curve fitting to a specific model. Second, these calculations have been used in literature and have been shown to be related to physiology.^{43,44} These quantities are not absolute, however since our goal was to compare the relative performance of reconstruction methods, they are considered sufficient for the present work.

For each dose level, ten noisy simulations were conducted. Each noisy simulation and reconstruction method yielded a parametric map (e.g., flow map) with some realization of stochastic noise and bias relative to the ground truth. Map accuracy (bias) for a given ROI size and location was defined as the mean difference between the map value within the ROI and the ground truth calculated directly from the simulation time-attenuation curve

$$\text{Accuracy} = \frac{\bar{X} - X_{\text{ref}}}{X_{\text{ref}}} \times 100\%, \quad (5)$$

where $\bar{X}$ is the average value measured within a ROI from a reconstruction (e.g., FBP or SIR) for the ten simulations, and X_{ref} is the value calculated from the simulation curve. Map precision at a given location was defined as the standard

deviation in the ROI means at that location, measured across the ten simulations, relative to reference value

$$\text{Precision} = \frac{\sigma_X}{X_{\text{ref}}} \times 100\%. \quad (6)$$

Accuracy and precision were characterized as a function of ROI size in four different regions of the simulated myocardium: anterior (reduced flow), posterior (normal), left (reduced flow), and right (normal). In each region, accuracy and precision were calculated for ROIs ranging from 30×30 pixels (166.4 mm^2) down to 1 pixel (0.18 mm^2). For the smaller ROIs, measurements were made for all possible locations within the largest 30×30 pixel ROI, and the results were averaged across location.

In addition, the performance of high dose FBP, low dose FBP, and low dose SIR were quantitatively compared using a perfusion map figure-of-merit (FOM), defined as the squared contrast-to-noise ratio in the perfusion map divided by the mean number of photons per input ray

$$\text{FOM}_{\text{map}} = \frac{\text{CNR}_{\text{map}}^2}{I_0} = \left(\frac{\bar{S}_{\text{ROI1}} - \bar{S}_{\text{ROI2}}}{0.5(\sigma_{S1} + \sigma_{S2})} \right)^2 / I_0. \quad (7)$$

Contrast was the difference between the right (normal flow) and left (reduced flow) regions of the myocardium. Noise was the average standard deviation in the two regions. As with accuracy and precision, FOM was calculated as a function of ROI size. The figure of merit normalizes the map standard deviation to the flow differences portrayed in the map, and provides a measure of how efficiently a particular method utilizes x-ray dose in the context of perfusion estimation.

2.C. Animal studies

2.C.1. Experimental setup

Four healthy swine (49–52 kg) were included in this study under an IACUC-approved protocol. Each animal was prepared with an inflatable balloon cuff around the left anterior descending artery (LAD) to enable imaging with and without total arterial occlusion. All animals were mechanically ventilated with isoflurane anesthesia throughout the study. Blood pressure, heart rate, ECG, ETCO₂ level, and temperature were continuously monitored. Zatebradine was administered to lower heart rate, with a target level of <65 bpm.

Low and high dose dynamic CT myocardial perfusion scans were performed on a 64-slice PET/CT scanner (GE Healthcare, Waukesha, WI) operated in cine mode with the following parameters: 120 kV, 0.4 s/rotation, 50 s scan duration. High dose scans were performed at 500 mA and the low dose scans were performed at 25 mA. The value of 25 mA was selected based on a previous porcine study which indicated that this tube current would result in significant photon starvation artifacts using FBP reconstruction,³⁷ and therefore provide a challenging dataset for the present work.

ECG signal was recorded during all scans for retrospective gating purposes. Iodinated contrast (iodixanol 320 mg/ml) with saline chase was injected through a catheter positioned in the inferior vena cava, beginning 3 s into each scan. The con-

trast volume was adjusted by weight (0.88 ml/kg), the saline volume was 85 ml, and the injection rate was 10 ml/s. To simulate breath holding for CT scanning, 45 s overbreathing was performed prior to each scan, and then the ventilator was suspended at end-expiration during CT scanning. Pancuronium was administered to achieve effective paralysis prior to scanning.

At the beginning of each study, two CT coronary angiograms were performed with the LAD in the “open” and “closed” states in order to confirm proper operation of the balloon occluder and assist in positioning the 4-cm thick CT perfusion scan volume. The CT perfusion scan volume was placed midway between the position of the LAD occluder and the apex of the heart. A low dose CTP scan, a high dose CTP scan, and an N-13 ammonia PET scan were performed in succession, for both the open and closed LAD states. The PET scans served to delineate the myocardial territory affected by the occlusion. Approximately 45 mCi was injected via the femoral vein at the beginning of each PET scan, and scan duration was 20 min. A 15 min waiting period was employed between the low dose and high dose CT scans. “Open” scans were performed first in three animals and “closed” scanning was performed first in one animal.

2.C.2. Image reconstruction and data analysis

For CT scans, FBP reconstructions were first performed for both low and high dose datasets. Then SIR was applied to the low dose datasets, and parameter λ was selected to match the noise level measured in the corresponding high dose FBP reconstruction. For this purpose, the noise level was the average image voxel standard deviation measured in uniform regions of the lateral and anterior myocardium, in both pre-contrast and peak-contrast cardiac cycles. In all reconstructions, the ECG signal was used to select projections centered on 75% R-R for each cardiac cycle. Reconstruction pixel size was 0.58 mm, and slice thickness was 2.5 mm. For PET scans, summed activity images were obtained using the Dynamic Vue package (GE Healthcare, Waukesha, WI), available at the scanner console, with pixel size of 3.3 mm, and slice thickness of 3.3 mm.

For each animal, parametric maps of blood flow, blood volume, and transit time were generated for the high dose FBP, low dose FBP, and low dose SIR reconstructions, for both “open” and “closed” scans, using Eqs. (4), (2), and (3). The arterial input function was placed in the left ventricle. For the mean perfusion value measurement, two large ROIs (300 pixels) were placed on each “closed” map in the occluded anterior and normal lateral regions, respectively. Then, the same ROIs were placed on the corresponding “open” map for measurement. To evaluate the ability of a reconstruction method to portray changes in flow, the percent change in the flow value for a given region (anterior or lateral) was calculated between the “closed” and “open” scans (e.g., 25 mA closed vs 25 mA open). To estimate the accuracy of global flow values relative to 500 mA FBP, the mean value measured in the combined anterior and lateral regions of the open scan were compared (e.g., 25 mA SIR vs 500 mA FBP). To provide a measure of

FIG. 2. (a)–(f) Example images of low dose FBP, low dose SIR, and high dose FBP reconstructions. Images are displayed at $[-300, 300]$ HU.

the variability of the map values within a uniform flow region, the average value was calculated for ten nonoverlapping 4×4 pixel ROIs placed within the larger lateral wall ROI, and the standard deviation of these values was calculated.

3. RESULTS

3.A. Numerical simulations

Figure 2 shows the low dose FBP ($I_0 = 50k$), low dose ($I_0 = 50k$) SIR and high dose ($I_0 = 1000k$) FBP reconstructions for a representative beat in the time series. The low dose FBP image was noisy and showed vertically oriented noise streaks due to low x-ray transmission along the anterior-posterior direction [Fig. 2(a)]. When the same data set was reconstructed using the SIR method, the streaks were mitigated [Fig. 2(b)]. In this investigation, the parameter λ ($\lambda = 20$) in the SIR reconstructions was selected to achieve a voxel noise less than or equal to that of the high dose FBP images [Fig. 2(c)]. Voxel noise in the myocardial region measured 37.2, 29.4, and 256.8 HU, respectively, in the high dose FBP, low dose SIR, and low dose FBP images. Later, in Sec. 3.B, we discuss the impact of different λ values on the SIR images and perfusion maps.

Perfusion maps generated on a pixel-by-pixel basis using Eqs. (2)–(4) are shown in Fig. 3. Qualitatively, the low dose FBP ($I_0 = 50k$) maps were too noisy to identify the

two ischemic regions. Maps of blood flow and transit time generated from the SIR reconstruction with same dose level ($I_0 = 50k$) demonstrated improved visibility of the abnormal regions. These low dose SIR ($I_0 = 50k$) maps showed a similar appearance to the high dose FBP ($I_0 = 1000k$) maps. Note that in the blood volume maps, the left ischemic region was not readily observable by any reconstruction method, including high dose FBP. This was because the difference in the volume values between the healthy and the left ischemic region was small by design [11.4% difference in area under the curve in Fig. 1(b)]. The differences in transit time and flow values were larger (29.5% and 37.0%).

Quantitative measurements of flow map accuracy (% error in the mean) and flow map precision (standard deviation, as % of mean) are presented in Fig. 4, as a function of ROI size in the anterior region. For ROIs from 81 to 900 pixels (15.0 to 166.4 mm²) in this region, the high dose FBP ($I_0 = 1000k$) accuracy was -1.7% or better, and the low dose SIR ($I_0 = 50k$) accuracy was -2.9% or better [Fig. 4(a)]. For ROIs less than 81 pixels (<15.0 mm²), high dose FBP accuracy ranged from -2.2% to $+2.0\%$ and low dose SIR accuracy was -1.2% to $+4.4\%$. The low dose FBP ($I_0 = 50k$) map had the largest relative errors, averaging $+280.2\%$ across all ROI sizes [Fig. 4(b)], due to the presence of severe photon starvation streaks. Similarly, the fluctuations in high dose FBP and low dose SIR flow measurements were much smaller than

FIG. 3. (a)–(i) Blood flow, blood volume, and transit time maps for low dose FBP ($I_0 = 50k$), low dose SIR ($I_0 = 50k$), and high dose FBP ($I_0 = 1000k$) simulations. Maps were generated on a pixel by pixel basis using Eqs. (2)–(4).

observed with low dose FBP. At the largest ROI size, the high dose FBP, low dose SIR, and low dose FBP precisions were 3.1%, 8.7%, and 93.0%, respectively. For all three methods, the measurements became less precise as the ROI size decreased [see Figs. 4(c) and 4(d)].

Table I summarizes the relative error and standard deviation of blood flow, blood volume, and transit time maps, measured with a 900 pixel ROI in each of the four myocardial regions. For this ROI size, the high dose FBP map errors were less than 1.3%. The low dose FBP maps demonstrated strong bias, especially in the streak-contaminated anterior region where map errors were up to +258%. With low dose SIR, the noisy projections in the anterior-posterior direction

were strongly penalized during reconstruction, and the resulting parametric map errors were much lower, ranging from -0.9% to $+4.4\%$ in the anterior region. Considering all regions and all map types, SIR errors were -6.6% to $+8.2\%$. SIR map precisions were 1.1%–9.4%. Table I also demonstrates that for all three reconstruction methods, flow errors were mainly due to errors in estimating blood volume.

Figure 5 shows the FOM (squared flow map contrast-to-noise divided by input photons/ray) for each scenario, as a function of ROI size. As expected, all reconstruction types demonstrate improving FOM as ROI size increases, due to increasing map CNR. The SIR figure-of-merit was highest at each ROI size, indicating the method yielded the greatest map

FIG. 4. (a)–(d) Relative error and standard deviation vs ROI size for blood flow measurements from simulations. The ROIs were placed in the anterior region of the myocardium.

CNR per unit dose. Examining the average FOM ratio across all ROI sizes with respect to low dose FBP for the blood flow, blood volume, and transit time maps, the low dose SIR case was 4.6–7.0 times more dose efficient compared to FBP at

TABLE I. Mean error (%) and standard deviation (%) relative to simulation ground truth for a 900 pixel ROI in the anterior (A), posterior (P), right (R), and left (L) myocardium.

		FBP 1000K	SIR 50K	FBP 50K
Blood flow	A (low flow)	-0.7 ± 3.1	-0.9 ± 8.7	258.4 ± 93.0
	P (normal)	0.6 ± 1.4	-6.6 ± 4.4	45.2 ± 22.2
	R (normal)	0.7 ± 1.0	7.3 ± 3.9	-33.8 ± 5.8
	L (low flow)	1.3 ± 1.5	6.9 ± 7.2	-37.8 ± 9.3
Blood volume	A (low flow)	-0.7 ± 3.2	3.3 ± 9.4	169.3 ± 76.1
	P (normal)	0.4 ± 1.7	-5.9 ± 5.6	36.6 ± 24.2
	R (normal)	0.5 ± 1.3	8.2 ± 4.8	-31.3 ± 7.0
	L (low flow)	1.3 ± 1.4	6.1 ± 6.5	-32.1 ± 8.8
Transit time	A (low flow)	-0.1 ± 0.7	4.4 ± 2.2	-24.6 ± 7.7
	P (normal)	-0.1 ± 0.5	0.7 ± 1.9	-5.8 ± 6.2
	R (normal)	-0.1 ± 0.4	3.7 ± 2.9	0.9 ± 1.2
	L (low flow)	1.0 ± 0.4	-0.6 ± 1.1	9.5 ± 3.1

the same dose, and 1.4–2.1 times more dose efficient than the high dose FBP case. High dose FBP was approximately 3.3 times more dose efficient than low dose FBP. The low FOM observed for the low-dose FBP case is believed to be due to the strong influence of electronic noise, which lowers contrast-to-noise in the parametric maps.

3.B. Impact of λ on perfusion maps

The weighting parameter λ in the SIR method [Eq. (1)] controls the tradeoff between spatial resolution and image noise. Section 3.A presented SIR reconstructions with $\lambda = 20$, which was selected to achieve voxel noise similar to the high dose FBP case. In this section, the impact of the λ value on perfusion maps is investigated.

Measurements of CT image voxel noise and image spatial resolution for different λ values are shown in Fig. 6(a). Spatial resolution was determined from profiles measured across object interfaces. Following the method of Lauzier et al.,³⁷ the spatial resolution was characterized as the width (σ) of the Gaussian kernel, which, when convolved with the reference (FBP) profile, generated the measured SIR profile.

FIG. 5. Flow map FOM vs ROI size in simulations. The FOM is calculated as the squared map CNR divided by the entrance photons per ray.

Since SIR spatial resolution is contrast-dependent, interfaces with high contrast (900 HU) and lower contrast (300 HU) were evaluated. The voxel noise in the low dose FBP image was 256.8 HU in the myocardium region. Even at very aggressive noise reduction levels (-92%), the SIR reconstructions showed minimal degradation of spatial resolution in the low contrast edges (i.e., myocardium/ventricle interface, 300 HU), with Gaussian blur width of less than 3 pixels. The high contrast edge (i.e., myocardium/lung interface, 900 HU) was even better maintained, with blur width of less than 1 pixel.

A reduction in CT image noise was accompanied by reduction in perfusion map noise. Figure 6(b) shows that for blood flow maps, there was a nearly linear relationship ($R^2 = 0.98$) between image voxel noise and map voxel noise. A similar linear relationship was observed in blood volume and transit time maps, with R^2 values larger than 0.94. For visual demonstration of the noise/resolution tradeoff, blood flow maps generated from the SIR reconstructions with different λ values are shown in Fig. 7. Smaller λ value resulted in maps with lower noise, because reducing λ gives less weight to the data term in Eq. (1), and hence more relative weight to the regularizer term.

(a)

(b)

FIG. 6. (a) Noise and spatial resolution trade off in SIR reconstructions ($I_0 = 50k$). The low contrast edge was measured at the myocardium/ventricle interface (300 HU) and the high contrast edge was measured at the myocardium/lung interface (900 HU). (b) Relationship between image noise and noise in the blood flow maps. Closed circles represent measurements with λ values of 100, 20, and 10.

3.C. Animal studies

Figure 8 shows example images of 25 mA FBP, 25 mA SIR, and 500 mA FBP reconstructions from the animal study. The 25 mA FBP images [Fig. 8(a)] were severely contaminated by photon starvation streaks, which corresponded to rays passing through the thickest projection of the animal and the thickest projection of the contrast-filled left ventricle. With 25 mA SIR [Fig. 8(b)], the streaks were mostly removed and the reconstructed image demonstrated a similar appearance to the 500 mA FBP image [Fig. 8(c)].

Qualitative flow maps generated using Eq. (4) are presented for each animal in Figs. 9–12. The location of an anterior myocardial infarction created by total LAD occlusion can be seen by comparing the “open” and “closed” PET images in each animal. The “closed” 25 mA SIR and “closed” 500 mA FBP perfusion maps show reduced values in the same infarcted territory, whereas the corresponding “open” maps show relatively uniform values. However, in the 25 mA FBP case, photon starvation streaks dominated the both “open” and “closed” maps and it was impossible to visualize the infarction. The elimination of photon starvation artifacts in SIR images offered major improvement in infarct visualization when imaging at low mA. Note that Fig. 12 corresponds to the animal which had “closed” scans performed prior to “open” scans. In this animal, a small residual perfusion defect is visible in the “open” PET scan [Fig. 12(d)] as well as CT.

Quantitative flow change between the “open” and “closed” scans in anterior wall and lateral wall ROIs are shown in Table II. In the 500 mA scans, coronary occlusion caused -88.4% , -96.0% , -76.7% , and -65.8% percent change in the anterior region value (four animals). The 25 mA SIR results were in good agreement, with the percent changes measuring -94.7% , -81.6% , -84.0% , and -72.2% . In contrast, the FBP-based results from the same 25 mA scan data showed unreliable changes in the anterior region ($+137.4\%$, $+71.0\%$, -11.8% , -3.5%). A similar trend can be observed for the nonoccluded lateral regions.

FIG. 7. Impact of λ on blood flow maps. These maps were generated on a pixel by pixel basis using Eq. (4). $I_0 = 50$ k. Maps are displayed at $[0, 80]$ ml/100 g/min. (a) $\lambda = 100$, (b) $\lambda = 20$, and (c) $\lambda = 10$.

Table III summarizes the standard deviation of flow measurements made within 4×4 pixel ROIs in a uniform lateral region, and the agreement in global flow between the 25 mA reconstructions and 500 mA FBP. The 500 mA FBP and 25 mA SIR standard deviations were 12%–20% and 20%–30%, respectively. Similar to the simulation findings, the 25 mA SIR reconstructions with voxel noise matched to 500 mA FBP yielded slightly higher variability in flow maps, relative to FBP. As expected, the 25 mA FBP flow map vari-

ability was very high, with standard deviation up to 24 times the mean. The global flow value for 25 mA SIR agreed with 500 mA FBP to within 10% for the first three animals, and within 26% for the fourth animal. In the fourth animal, the 25 mA scan had beat-to-beat variations in cardiac shape that were higher than normal, which in turn reduced the thickness of myocardium that could be analyzed. For the 25 mA FBP case, global flow differed from 500 mA FBP by 51% on average.

FIG. 8. (a)–(f) Example reconstructed images of low dose (25 mA) FBP, low dose (25 mA) SIR, and high dose (500 mA) FBP. Images are displayed at $[-600, 600]$ HU. The weight of the animal was 50.8 kg.

FIG. 9. CT blood flow maps and PET images for nonoccluded (a)–(d) and occluded (e)–(h) scans of animal 1. The occluded region was in the anterior wall due to the blockage of the LAD. The weight of the animal was 49.5 kg. The CT maps are displayed at $[0, 60]$ ml/100 g/min. The heart rate was 63.08 ± 1.7 , 64.2 ± 1.9 , 69.2 ± 5.5 , and 67.8 ± 2.6 bpm in the low dose “open,” high dose “open,” low dose “closed,” and high dose “closed” scans.

4. DISCUSSION

Iterative CT reconstruction methods generally seek an image solution which is consistent with acquired projection data and which also minimizes image roughness (noise). The key idea of SIR is to selectively penalize the contributions of the noisy projection rays (i.e., those with low detected photons). Since photon fluctuations follow a Poisson distribution, each measurement is a noisy realization around a “true” value. The higher the detected signal, the smaller the relative fluctuations and therefore the more likely the measurement is near

the truth. Conversely, smaller signals have greater uncertainty and may be farther from the truth. Compared to conventional FBP, which gives equal weight to all measurements, SIR images are less susceptible to noise-induced streaking artifacts that can arise in low-dose scanning scenarios. The regularization term in SIR [Eq. (1)] can also be used to incorporate prior knowledge of the object, such as smoothness. In this study, we chose to use total variation as the regularizer, which promotes uniform regions. When a higher weight is given to the regularizer (i.e., smaller λ), the reconstructed images have lower noise, as shown in Fig. 7. The SIR technique is applied

FIG. 10. CT blood flow maps and PET images for nonoccluded (a)–(d) and occluded (e)–(h) scans of animal 2. The occluded region was in the anterior wall due to the blockage of the LAD. The weight of the animal was 52.2 kg. The CT maps are displayed at $[0, 60]$ ml/100 g/min. The heart rate was 71.0 ± 1.3 , 70.4 ± 0.8 , 67.0 ± 0.6 , and 67.4 ± 0.6 bpm in the low dose “open,” high dose “open,” low dose “closed,” and high dose “closed” scans.

FIG. 11. CT blood flow maps and PET images for nonoccluded (a)–(d) and occluded (e)–(h) scans of animal 3. The occluded region was in the anterior wall due to the blockage of the LAD. The weight of the animal was 50.8 kg. The CT maps are displayed at $[0, 60]$ ml/100 g/min. The heart rate was 61.5 ± 0.8 , 62.4 ± 0.7 , 60.8 ± 1.1 , and 57.2 ± 0.6 bpm in the low dose “open,” high dose “open,” low dose “closed,” and high dose “closed” scans.

to individual image volumes in a sequence. As such, it is not susceptible to blurring that can arise from sharing data across time points.⁴⁵

This study investigates the ability of SIR reconstruction to improve parametric maps of myocardial perfusion derived from a low tube current dynamic CT scan. Numerical simulations were conducted with a low x-ray flux and electronic noise level intended to approximate the conditions of low tube current (25 mA) scanning. This resulted in conventional FBP reconstructions with severe structured noise (i.e.,

streaks) which prevented reliable perfusion analysis. Statistical iterative reconstruction of the low dose simulated data was able to largely mitigate the streaks caused by photon starvation, yielding more accurate and precise perfusion maps, with accuracy $<7\%$ and precision $<9\%$ for large ROI measurements. In comparison, the FBP simulation at the same dose yielded accuracy of -38% to $+258\%$ and precisions from 6% to 93%. FBP with 20 times higher x-ray flux yielded accuracy and precision within 1%. However, measurements of the squared contrast-to-noise ratio per unit dose demonstrated

FIG. 12. CT blood flow maps and PET images for nonoccluded (a)–(d) and occluded (e)–(h) scans of animal 4. The occluded region was in the anterior wall due to the blockage of the LAD. There was a residual perfusion defect in the nonoccluded scan, because the nonoccluded scan was performed after the occluded scan. The weight of the animal was 52.7 kg. The CT maps are displayed at $[0, 60]$ ml/100 g/min. The heart rate was 94.6 ± 6.9 , 95.9 ± 4.6 , 82.6 ± 5.6 , and 84.7 ± 6.3 bpm in the low dose “open,” high dose “open,” low dose “closed,” and high dose “closed” scans.

TABLE II. Change in CT flow metric for occluded vs open states (%).

Myocardial region		500 mA FBP	25 mA SIR	25 mA FBP
Animal 1	Occluded anterior	− 88.4	− 94.7	137.4
	lateral	− 7.7	− 11.6	688.4
Animal 2	Occluded anterior	− 96.0	− 81.6	71.0
	lateral	46.0	36.1	− 174.4
Animal 3	Occluded anterior	− 76.7	− 84.0	− 11.8
	lateral	1.9	− 13.2	− 68.5
Animal 4	Occluded anterior	− 65.8	− 72.2	− 3.5
	lateral	7.5	1.3	− 69.3

that SIR produced higher map contrast-to-noise for a given dose level, relative to FBP. The degree of SIR image noise and perfusion map noise reduction depends on the adjustable parameter λ . In this work, the value was selected so that the low-dose SIR voxel noise was similar to the high-dose FBP case.

In the animal studies performed at 25 mA and 500 mA, it was also found that the low tube current scans created major streak artifacts in FBP reconstruction which were mitigated by SIR reconstruction. Qualitatively, the 25 mA SIR images portrayed a regional change in flow caused by coronary occlusion which was well-correlated with the 500 mA FBP images and the affected territory indicated by PET. Quantitatively, the 25 mA SIR-based measurements of percent change in flow agreed with the corresponding 500 mA FBP-based measurements to within 15%. Agreement in global flow value between 25 mA SIR and 500 mA FBP was within 10%, with the exception of one animal.

This study has several limitations. In the animal studies, comparisons of low dose and high dose CT maps used data acquired at different time points. Since flow changes due to residual iodine and/or temporal changes in hemodynamics cannot be fully excluded, the high dose FBP data only represent an estimate of ground truth. Beam hardening correction was not employed with the animal data. Although no obvious beam hardening artifacts were observed when visually comparing flow in the anterior and lateral regions of the open scans, subtle effects cannot be excluded. Partial scan artifact (PSA) correction^{46–48} was also not employed. In a previous study,⁴⁹ it was found that partial scan artifact was mainly due

to detected x-ray scatter and that these artifacts are manifested as periodic temporal variations in CT number measured in a large ROI. An examination of large-ROI time-attenuation curves in these animals did not reveal obvious partial scan artifact. The numerical simulations were not affected by beam hardening or scatter because a monochromatic x-ray spectrum was used, and scatter was not simulated. Cardiac motion can blur myocardial borders, and beat-to-beat variations in cardiac shape can blur local details in a flow map. Finally, the CT flow metric used in this study was a relative measure, and the study only investigated a single low tube current that was known to create photon starvation with conventional FBP. Studies in an animal model of varying degrees of dose reduction with an absolute CT flow metric would further delineate the capabilities of this technique.

5. CONCLUSIONS

SIR reconstruction can reduce image noise and mitigate streaking artifacts caused by photon starvation in dynamic CT myocardial perfusion data sets acquired at low dose (low tube current), and improve perfusion map quality in comparison to FBP reconstruction at the same dose level.

ACKNOWLEDGMENTS

Financial support for this work was provided by National Institutes of Health (NIH) Grant No. R01 HL090776. The authors wish to thank Jill Koch, Guoqing Song, Christine Jaskowiak, Kari Pulfer, Brooke Peters, Mark McNall, Todd Barnhart, Jerry Nickles, Brad Christian, Ansel Hillmer, Jennifer Meudt, Christian Krueger, and Dhanu Shanmuganayagam for their assistance with the animal studies, and GE Healthcare, Inc., for assistance with preprocessing of CT projection data.

^aAuthor to whom correspondence should be addressed. Electronic mail: speidel@wisc.edu

¹M. J. Budoff et al., “Assessment of coronary artery disease by cardiac computed tomography: A scientific statement from the American Heart Association Committee on Cardiovascular Imaging and Intervention, Council on Cardiovascular Radiology and Intervention, and Committee on Cardiac Imaging, Council on Clinical Cardiology,” *Circulation* **114**, 1761–1791 (2006).

²D. C. Paech and A. R. Weston, “A systematic review of the clinical effectiveness of 64-slice or higher computed tomography angiography as an alternative to invasive coronary angiography in the investigation of suspected coronary artery disease,” *BMC Cardiovasc. Disorders* **11**, 32–42 (2011).

³J. Bartunek, S. U. Sys, G. R. Heyndrickx, N. H. Pijls, and B. De Bruyne, “Quantitative coronary angiography in predicting functional significance of stenoses in an unselected patient cohort,” *J. Am. Coll. Cardiol.* **26**, 328–334 (1995).

⁴A. G. Wasserman, *Cardiac Application of Digital Angiography* (Futura, Mount Kisco, NY, 1989).

⁵N. Pijls, *Maximal Myocardial Perfusion as a Measure of the Functional Significance of Coronary Artery Disease: From a Pathoanatomic to a Pathophysiologic... (Developments in Cardiovascular Medicine)* (Kluwer, London, 1991).

⁶E. J. Topol and S. E. Nissen, “Our preoccupation with coronary luminology: The dissociation between clinical and angiographic findings in ischemic heart disease,” *Circulation* **92**, 2333–2342 (1995).

⁷W. B. Meijboom et al., “Diagnostic accuracy of 64-slice computed tomography coronary angiography: A prospective, multicenter, multivendor study,” *J. Am. Coll. Cardiol.* **52**, 2135–2144 (2008).

TABLE III. Variability of CT flow metrics and comparison of flow means.

	Standard deviation ^a (% of mean)			Global flow ^b relative to 500 mA FBP (% diff.)	
	500 mA FBP	25 mA SIR	25 mA FBP	25 mA SIR	25 mA FBP
Animal 1	16.1	30.1	2435.9	− 9.7	− 56.7
Animal 2	16.7	25.8	1018.2	8.8	24.6
Animal 3	12.0	20.6	161.2	− 3.1	− 24.2
Animal 4	20.3	20.0	136.2	26.4	98.0

^aMeasurements made for ten nonoverlapping 4×4 pixel regions.

^bAverage value in lateral and anterior regions of open scan.

- ⁸F. J. Klocke, M. G. Baird, B. H. Lorell, T. M. Bateman, J. V. Messer, D. S. Berman, P. T. O'Gara, B. A. Carabello, R. O. Russell, Jr., M. D. Cerqueira, M. G. St John Sutton, A. N. DeMaria, J. E. Udelson, J. W. Kennedy, M. S. Verani, K. A. Williams, E. M. Antman, S. C. Smith, Jr., J. S. Alpert, G. Gregoratos, J. L. Anderson, L. F. Hiratzka, D. P. Faxon, S. A. Hunt, V. Fuster, A. K. Jacobs, R. J. Gibbons, and R. O. Russell, American College of Cardiology, American Heart Association Task Force on Practice Guidelines, and American Society of Nuclear Cardiology, "ACC/AHA/ASNC guidelines for the clinical use of cardiac radionuclide imaging—executive summary: A report of the American College of Cardiology/American Heart Association Task Force on Practice Guidelines (ACC/AHA/ASNC Committee to revise the 1995 guidelines for the clinical use of cardiac radionuclide imaging)," *Circulation* **108**, 1404–1418 (2003).
- ⁹R. C. Hendel, D. S. Berman, M. F. Di Carli, P. A. Heidenreich, R. E. Henkin, P. A. Pellikka, G. M. Pohost, and K. A. Williams, "ACCF/ASNC/ACR/AHA/ASE/SCCT/SCMR/SNM 2009 appropriate use criteria for cardiac radionuclide imaging: A report of the American College of Cardiology Foundation Appropriate Use Criteria Task Force, The American Society of Nuclear Cardiology, The American College of Radiology, The American Heart Association, The American Society of Echocardiography, The Society of Cardiovascular Computed Tomography, The Society for Cardiovascular Magnetic Resonance, and The Society of Nuclear Medicine endorsed by the American College of Emergency Physicians," *J. Am. Coll. Cardiol.* **53**, 2201–2229 (2009).
- ¹⁰K.-T. Ho, K.-C. Chua, E. Klotz, and C. Panknin, "Stress and rest dynamic myocardial perfusion imaging by evaluation of complete time-attenuation curves with dual-source CT," *JACC Cardiovasc. Imaging* **3**, 811–820 (2010).
- ¹¹F. Bamberg, A. Becker, F. Schwarz, R. P. Marcus, M. Greif, F. von Ziegler, R. Blankstein, U. Hoffmann, W. H. Sommer, V. S. Hoffmann, T. R. C. Johnson, H.-C. R. Becker, B. J. Wintersperger, M. F. Reiser, and K. Nikolaou, "Detection of hemodynamically significant coronary artery stenosis: Incremental diagnostic value of dynamic CT-based myocardial perfusion imaging," *Radiology* **260**, 689–698 (2011).
- ¹²Y. Wang, L. Qin, X. Shi, Y. Zeng, H. Jing, U. J. Schoepf, and Z. Jin, "Adenosine-stress dynamic myocardial perfusion imaging with second-generation dual-source CT: Comparison with conventional catheter coronary angiography and SPECT nuclear myocardial perfusion imaging," *AJR, Am. J. Roentgenol.* **198**, 521–529 (2012).
- ¹³R. T. George, C. Silva, M. A. S. Cordeiro, A. DiPaula, D. R. Thompson, W. F. McCarthy, T. Ichihara, J. A. C. Lima, and A. C. Lardo, "Multidetector computed tomography myocardial perfusion imaging during adenosine stress," *J. Am. Coll. Cardiol.* **48**, 153–160 (2006).
- ¹⁴R. T. George, A. Arbab-Zadeh, J. M. Miller, K. Kitagawa, H.-J. Chang, D. A. Bluemke, L. Becker, O. Yousuf, J. Texter, A. C. Lardo, and J. A. C. Lima, "Adenosine stress 64- and 256-row detector computed tomography angiography and perfusion imaging: A pilot study evaluating the transmural extent of perfusion abnormalities to predict atherosclerosis causing myocardial ischemia," *Circ. Cardiovasc. Imaging* **2**, 174–182 (2009).
- ¹⁵R. Blankstein, L. D. Shturman, I. S. Rogers, J. A. Rocha-Filho, D. R. Okada, A. Sarwar, A. V. Soni, H. Bezerra, B. B. Ghoshhajra, M. Petranovic, R. Loureiro, G. Feuchtnr, H. Gewirtz, U. Hoffmann, W. S. Mamuya, T. J. Brady, and R. C. Cury, "Adenosine-induced stress myocardial perfusion imaging using dual-source cardiac computed tomography," *J. Am. Coll. Cardiol.* **54**, 1072–1084 (2009).
- ¹⁶J. A. Rocha-Filho, R. Blankstein, L. D. Shturman, H. G. Bezerra, D. R. Okada, I. S. Rogers, B. Ghoshhajra, U. Hoffmann, G. Feuchtnr, W. S. Mamuya, T. J. Brady, and R. C. Cury, "Incremental value of adenosine-induced stress myocardial perfusion imaging with dual-source CT at cardiac CT angiography," *Radiology* **254**, 410–419 (2010).
- ¹⁷R. C. Cury, T. A. Magalhaes, A. T. Paladino, A. A. Shiozaki, M. Perini, T. Senra, P. A. Lemos, R. C. Cury, and C. E. Rochitte, "Dipyridamole stress and rest transmural myocardial perfusion ratio evaluation by 64 detector-row computed tomography," *J. Cardiovasc. Comput. Tomogr.* **5**, 443–448 (2011).
- ¹⁸T. A. Magalhaes, R. C. Cury, A. C. Pereira, V. d. M. Moreira, P. A. Lemos, R. Kalil-Filho, and C. E. Rochitte, "Additional value of dipyridamole stress myocardial perfusion by 64-row computed tomography in patients with coronary stents," *J. Cardiovasc. Comput. Tomogr.* **5**, 449–458 (2011).
- ¹⁹M. Weininger, U. J. Schoepf, A. Ramachandra, C. Fink, G. W. Rowe, P. Costello, and T. Henzler, "Adenosine-stress dynamic real-time myocardial perfusion CT and adenosine-stress first-pass dual-energy myocardial perfusion CT for the assessment of acute chest pain: Initial results," *Eur. J. Radiol.* **81**, 3703–3710 (2012).
- ²⁰A. H. Mahnen, E. Klotz, H. Pietsch, B. Schmidt, T. Allmendinger, U. Haberland, W. A. Kalender, and T. Flohr, "Quantitative whole heart stress perfusion CT imaging as noninvasive assessment of hemodynamics in coronary artery stenosis: Preliminary animal experience," *Invest. Radiol.* **45**, 298–305 (2010).
- ²¹G. Bastarrika, L. Ramos-Duran, M. A. Rosenblum, D. K. Kang, G. W. Rowe, and U. J. Schoepf, "Adenosine-stress dynamic myocardial CT perfusion imaging: Initial clinical experience," *Invest. Radiol.* **45**, 306–313 (2010).
- ²²C. Fink, R. Krissak, T. Henzler, U. Lechel, G. Brix, R. A. Takx, J. W. Nance, J. A. Abro, S. O. Schoenberg, and U. J. Schoepf, "Radiation dose at coronary CT angiography: Second-generation dual-source CT versus single-source 64-MDCT and first-generation dual-source CT," *AJR, Am. J. Roentgenol.* **196**, W550–W557 (2011).
- ²³A. Yerramasu, S. Venuraju, S. Atwal, D. Goodman, D. Lipkin, and A. Lahiri, "Radiation dose of CT coronary angiography in clinical practice: Objective evaluation of strategies for dose optimization," *Eur. J. Radiol.* **81**, 1555–1561 (2012).
- ²⁴M. Y. Chen, S. M. Shanbhag, and A. E. Arai, "Submillisievert median radiation dose for coronary angiography with a second-generation 320-detector row CT scanner in 107 consecutive patients," *Radiology* **267**, 76–85 (2013).
- ²⁵C. Tomasi and R. Manduchi, "Bilateral filtering for gray and color images," in *Proceedings of the Sixth International Conference on Computer Vision*, Bombay, India, 1998 (IEEE Computer Society, Washington DC, 1998), pp. 839–846.
- ²⁶A. M. Mendrik, E.-J. Vonken, B. van Ginneken, H. W. de Jong, A. Riordan, T. van Seeters, E. J. Smit, M. A. Viergever, and M. Prokop, "Tips bilateral noise reduction in 4D CT perfusion scans produces high-quality cerebral blood flow maps," *Phys. Med. Biol.* **56**, 3857–3872 (2011).
- ²⁷S. G. Mallat, "A theory for multiresolution signal decomposition: The wavelet representation," *IEEE Trans. Pattern Anal. Mach. Intell.* **11**, 674–693 (1989).
- ²⁸D. L. Donoho and J. M. Johnstone, "Ideal spatial adaptation by wavelet shrinkage," *Biometrika* **81**, 425–455 (1994).
- ²⁹Y. Xu, J. B. Weaver, D. M. Healy, and J. Lu, "Wavelet transform domain filters: A spatially selective noise filtration technique," *IEEE Trans. Image Process.* **3**, 747–758 (1994).
- ³⁰A. Borsdorf, R. Raupach, T. Flohr, and J. Hornegger, "Wavelet based noise reduction in CT-images using correlation analysis," *IEEE Trans. Med. Imaging* **27**, 1685–1703 (2008).
- ³¹Y. Tao, S. Brunner, J. Tang, M. Speidel, H. Rowley, M. VanLysel, and G.-H. Chen, "Radiation dose reduction in computed tomography (CT) using a new implementation of wavelet denoising in low tube current acquisitions," *Proc. SPIE* **7961**, 79613M (2011).
- ³²P. Perona and J. Malik, "Scale-space and edge detection using anisotropic diffusion," *IEEE Trans. Pattern Anal. Mach. Intell.* **12**, 629–639 (1990).
- ³³J. Weickert, *Anisotropic Diffusion in Image Processing* (Teubner, Stuttgart, 1998), Vol. 1.
- ³⁴A. M. Mendrik, E.-J. Vonken, A. Rutten, M. A. Viergever, and B. van Ginneken, "Noise reduction in computed tomography scans using 3-D anisotropic hybrid diffusion with continuous switch," *IEEE Trans. Med. Imaging* **28**, 1585–1594 (2009).
- ³⁵Y. Vardi, L. Shepp, and L. Kaufman, "A statistical model for positron emission tomography," *J. Am. Stat. Assoc.* **80**, 8–20 (1985).
- ³⁶J. A. Fessler, "Penalized weighted least-squares image reconstruction for positron emission tomography," *IEEE Trans. Med. Imaging* **13**, 290–300 (1994).
- ³⁷P. T. Lauzier, J. Tang, M. A. Speidel, and G.-H. Chen, "Noise spatial nonuniformity and the impact of statistical image reconstruction in CT myocardial perfusion imaging," *Med. Phys.* **39**, 4079–4092 (2012).
- ³⁸J.-B. Thibault, K. D. Sauer, C. A. Bouman, and J. Hsieh, "A three-dimensional statistical approach to improved image quality for multislice helical CT," *Med. Phys.* **34**, 4526–4544 (2007).
- ³⁹K. Sauer and C. Bouman, "A local update strategy for iterative reconstruction from projections," *IEEE Trans. Signal Process.* **41**, 534–548 (1993).
- ⁴⁰C. A. Bouman and K. Sauer, "A unified approach to statistical tomography using coordinate descent optimization," *IEEE Trans. Image Process.* **5**, 480–492 (1996).

- ⁴¹J. Nocedal and S. J. Wright, *Numerical Optimization* (Springer, New York, NY, 1999), Vol. 2.
- ⁴²P. T. Lauzier, J. Tang, and G.-H. Chen, "Prior image constrained compressed sensing: Implementation and performance evaluation," *Med. Phys.* **39**, 66–80 (2012).
- ⁴³A. Konstas, G. Goldmakher, T.-Y. Lee, and M. Lev, "Theoretic basis and technical implementations of CT perfusion in acute ischemic stroke, Part 1: Theoretic basis," *Am. J. Neuroradiol.* **30**, 662–668 (2009).
- ⁴⁴S. Christensen, K. Mouridsen, O. Wu, N. Hjort, H. Karstoft, G. Thomalla, J. Röther, J. Fiehler, T. Kucinski, and L. Østergaard, "Comparison of 10 perfusion MRI parameters in 97 sub-6-hour stroke patients using voxel-based receiver operating characteristics analysis," *Stroke* **40**, 2055–2061 (2009).
- ⁴⁵M. A. Speidel, C. L. Bateman, Y. Tao, A. N. Raval, T. A. Hacker, S. B. Reeder, and M. S. Van Lysel, "Reduction of image noise in low tube current dynamic CT myocardial perfusion imaging using HYPR processing: A time-attenuation curve analysis," *Med. Phys.* **40**, 011904 (8pp.) (2013).
- ⁴⁶J. A. Meinel, E. Hoffman, A. Clough, and G. Wang, "Reduction of half-scan shading artifact based on full-scan correction," *Acad. Radiol.* **13**, 55–62 (2006).
- ⁴⁷P. Stenner, B. Schmidt, H. Bruder, T. Allmendinger, U. Haberland, T. Flohr, and M. Kachelriess, "Partial scan artifact reduction (PSAR) for the assessment of cardiac perfusion in dynamic phase-correlated CT," *Med. Phys.* **36**, 5683–5694 (2009).
- ⁴⁸J. C. Ramirez-Giraldo, L. Yu, B. Kantor, E. L. Ritman, and C. H. McCollough, "A strategy to decrease partial scan reconstruction artifacts in myocardial perfusion CT: Phantom and *in vivo* evaluation," *Med. Phys.* **39**, 214–223 (2012).
- ⁴⁹Y. Tao, M. Speidel, T. P. Szczykutowicz, and G.-H. Chen, "Investigation of the potential causes of partial scan artifacts in dynamic CT myocardial perfusion imaging," *Proc. SPIE* **9033**, 90332B (2014).