Published in final edited form as:

Magn Reson Med. 2014 January ; 71(1): 432–439. doi:10.1002/mrm.24653.

A Fast, Analytically Based Method to Optimize Local Transmit Efficiency for a Transmit Array

Giuseppe Carluccio¹, Christopher M. Collins¹, and Danilo Erricolo^{2,*}

¹Department of Radiology, New York University, New York, New York, USA

²Department of Electrical and Computer Engineering, University of Illinois at Chicago, Chicago, Illinois, USA

Abstract

Purpose—To develop an analytically based algorithm for rapid optimization of the local radiofrequency magnetic (B_1^+) field intensity for a given radiofrequency power through a transmit array. The analytical nature of the method will yield insight to optimization requirements and provides a valuable reference for numerically based searches.

Methods—With the knowledge of the B_1^+ field distribution generated by each single coil of the array, both the phases and the amplitudes of each coil current are optimized to maximize the magnitude of the B_1^+ field in a specific location of the body per unit of power transmitted through the array and, consequently, minimizing the whole body specific absorption rate for a given pulse sequence.

Results—Simulations considering the human body show that the proposed method can reduce the whole-body specific absorption rate for a given B_1^+ magnitude at the location of interest by a factor of about 6.3 compared to the classic birdcage current configuration, and by a factor of 3.2 compared to phase-only shimming in a case with significant coupling between the elements of the array.

Conclusion—The proposed method can rapidly provide valuable information pertinent to the optimization of field distributions from transmit arrays.

Keywords

radiofrequency; magnetic resonance imaging; shimming; power; specific absorption rate; spectroscopy

A current challenge for high-field magnetic resonance imaging is nonuniformity of the radiofrequency magnetic excitation field (B_1^+). Because the frequency of the B_1^+ field is proportional to strength of the static magnetic (B_0) field, at high B_0 fields the B_1^+ field has a relatively short wavelength, resulting in nonhomogeneous flip-angle distributions and

ultimately affecting image quality. Radiofrequency (RF) shimming is the simplest of a variety of approaches using an array of coils in transmission to address this challenge. In RF shimming, a more desirable RF electromagnetic field distribution is achieved with adjustment of the magnitude and/or phase of the currents or voltages driving the elements of the transmit array (1–3). More advanced methods can achieve excitation distributions very different than the RF field distributions (4–6), but in general require significantly longer pulse durations and/or greater total RF energy to achieve a given average flip angle.

In some cases, especially in the human head, reasonably homogeneous excitation of almost the entire volume can be achieved with use of RF shimming (7). In other cases, however, it may not be possible or advantageous to optimize field homogeneity over a large volume. If we are interested either in a single small volume, such as in spectroscopy (8), or in imaging where the region of interest (ROI) is small compared to the sample volume and the sample is large enough that RF shimming cannot readily produce a homogeneous field across its volume (9), local RF shimming may be preferred. In these cases, it is expected that the B_1^+ field across an ROI smaller than about one quarter wavelength will be fairly homogeneous as long as there is constructive interference from the fields of individual arrays there, and attention can be devoted to the efficiency with which B_1^+ is produced in the ROI.

By reducing the amount of power required to create a given B_1^+ field in the ROI, the whole-body (global) specific absorption rate (SAR) is reduced, and there is greater flexibility in the imaging parameters (including imaging time) that can be used. It has been observed that limits on local SAR can often be exceeded before those on average SAR will (10). According to the most recent version of widely used guidelines (11), when an array of transmit coils is used as a volume coil there is no limit on local SAR, providing motivation for considering whole-body SAR. It is also notable that average SAR is more readily monitored than local SAR (12), making methods to reduce it more amenable to verification. Even in cases where local SAR may be the limiting factor, however, rapidly determined shim values that produce optimal overall efficiency and minimal whole-body SAR can provide a valuable reference for other optimization methods designed to consider local SAR.

Although a number of articles have focused on controlling local or average SAR in RF shimming of a large region (13) or in advanced transmit array pulse designs for homogeneous excitation (14), comparatively little work has considered RF shimming on a localized region. Methods for local RF shimming designed to minimize power requirements and whole-body average SAR have included an analytically based approach to adjusting only the phase of array elements for imaging of the human prostate in vivo (9), an approach based on the Rayleigh quotient optimization (15,16) and a numerical optimization of the phase and magnitude of all elements in simulation-based demonstrations (17).

Here, we present a simple, analytically based method to adjust both magnitude and phase of all elements for local RF shimming to minimize power requirements and whole-body SAR.

METHODS

A method that optimizes only the phases of the transmit array elements was shown previously for application to the prostate (9). Indicating with $B_{1,i,m}^+$ the circularly polarized component of the B_1 field generated by the i -th element of the transmit array in the m -th voxel of the ROI when the i -th element is driven with the reference current $I_{i,\text{ref}}$, this phase optimization process consists of acquiring the phases of all $B_{1,i,m}^+$ fields in the M voxels belonging to the ROI with a technique of B_1 phase mapping, and adjusting the phase of the input to the i -th element by an amount equal to the opposite of the measured phase of $B_{1,i,m}^+$ fields in the ROI. The resulting optimal coil current could be written

$$I_i = I_{i,\text{ref}} e^{-j \frac{1}{M} \sum_{m=1}^M \angle B_{1,i,m}^+}, \quad [1]$$

where j is the imaginary unit. After this, all the B_1^+ fields generated by each element of the array will add constructively in the ROI, producing B_1^+ more efficiently. Note that it must be possible to control the current in each element as in Eq. [1] to provide the desired effect on the phase of the field produced.

In the following, we propose and demonstrate a simple method to find the set of currents I_i having both optimal phase and optimal amplitude. This method is developed with the assumption that complex current in each element is known explicitly. In some configurations of transmit arrays, this is indeed the case (18). In others, with adequate measurement of the impedance matrix and knowledge of the input voltage it is technically possible to determine the currents. In any case, this work will provide an intuitive understanding of the requirements for optimizing the efficiency of a transmit array for local excitation.

In the case that $I_{i,\text{ref}}$ is identical for all the elements and equal to I_{ref} , we can write the desired current driving each element of the transmit array as

$$I_i = I_{\text{ref}} C A_i e^{-j \frac{1}{M} \sum_{m=1}^M \angle B_{1,i,m}^+}. \quad [2]$$

where the optimal current amplitudes A_i are dimensionless real positive numbers, and C is a normalization factor equal for all the elements of the transmit array. The value for C can be used to normalize the currents to satisfy, if necessary, some safety requirements such as local average SAR, temperature increase, or to obtain a specific value of flip-angle while still keeping the same efficiency in terms of transmitted field B_1^+ and generated power. Both the magnitudes and phases of B_1^+ can be determined experimentally (8,19). The amplitudes A_i are determined through the optimization of a cost function that attempts to simultaneously maximize the total B_1^+ field at the desired ROI and minimize the transmitted power, while the phases are determined as done in Eq. [2].

The power transmitted through an array can be calculated as

$$P_{Tx} = \frac{1}{2} \sum_{k=1}^N \sum_{i=1}^N \text{Re}\{I_i Z_{ik} I_k^*\}, \quad [3]$$

where Z_{ik} are the elements of the impedance matrix $\mathbf{Z}$ and represent the mutual impedance between the i -th and the k -th element of the array, which can be measured with a network analyzer.

The cost function depends on the observables to be optimized. In particular, in this work, we choose to minimize the square root of the transmitted power over the average B_1^+ field in the ROI:

$$f = \frac{\sqrt{P_{Tx}}}{\left| \frac{1}{M} \sum_{m=1}^M B_{1,m}^+ \right|}, \quad [4]$$

where $B_{1,m}^+ = \sum_{i=1}^N A_i B_{1,i,m}^+$. This will be at a minimum when P_{Tx} is minimized for a given B_1^+ amplitude. There are two motivations to minimize P_{Tx} : (1) the generated power provides an upper bound to the whole-body SAR and (2) P_{Tx} is a measurable parameter in a magnetic resonance imaging system. However, if additional information is available through a more accurate relation between generated power and SAR (12), the cost function could be modified to also take advantage of this. The definition of the function f contains the square root of generated power to avoid a linear dependence with the currents generating the fields. To clarify the explanation of our method, we consider two different cases. In the first one, we examine a simplified situation where there is negligible coupling among the array elements, which causes the impedance matrix $\mathbf{Z}$ to be diagonal, and an exact analytical solution is obtained. In the second case, the more general situation of nonnegligible coupling among array elements is examined and it is solved through a diagonalization of the impedance matrix $\mathbf{Z}$. Keeping these two cases separate allows for evaluation of two different cases (decoupled and coupled arrays) in a natural progression.

Case 1: Negligible Mutual Coupling

When the coupling between different elements of the array is small ($|Z_{ik}| \ll |Z_{ii}|$ for all i and all $k \neq i$), the values of the amplitudes that minimize f can be obtained by finding a set of currents causing the gradient of f to be zero. Specifically, the generated power is approximated as

$$P_{Tx} \approx \frac{1}{2} \sum_{i=1}^N \text{Re}\{Z_{ii}\} |I_i|^2 = \frac{1}{2} C^2 \sum_{i=1}^N \text{Re}\{Z_{ii}\} A_i^2. \quad [5]$$

and the components of the first derivative are set to zero, yielding

$$\frac{\partial f}{\partial A_i} = \frac{\partial \left(\frac{\sqrt{\frac{1}{2}C^2 \sum_{i=1}^N \text{Re}\{Z_{ii}\} A_i^2}}{\sum_{i=1}^N C A_i \left| \frac{1}{M} \sum_{m=1}^M B_{1,i,m}^+ \right|} \right)}{\partial A_i} = \sqrt{\frac{1}{2}} \frac{\partial \left(\frac{\sqrt{\sum_{i=1}^N \text{Re}\{Z_{ii}\} A_i^2}}{\sum_{i=1}^N A_i \left| \frac{1}{M} \sum_{m=1}^M B_{1,i,m}^+ \right|} \right)}{\partial A_i} = 0 \quad [6]$$

or equivalently

$$\frac{\partial f}{\partial A_i} = \sqrt{\frac{1}{2}} \left[\frac{\text{Re}\{Z_{ii}\} A_i \sum_{l=1}^N A_l \left| \frac{1}{M} \sum_{m=1}^M B_{1,i,m}^+ \right| - \left| \frac{1}{M} \sum_{m=1}^M B_{1,i,m}^+ \right| \sum_{l=1}^N \text{Re}\{Z_{ll}\} A_l^2}{\sqrt{\sum_{i=1}^N \text{Re}\{Z_{ii}\} A_i^2} \left(\sum_{i=1}^N A_i \left| \frac{1}{M} \sum_{m=1}^M B_{1,i,m}^+ \right| \right)^2} \right] = 0 \quad [7]$$

By solving Eq. [7] for A_i

$$A_i = \frac{\left| \frac{1}{M} \sum_{m=1}^M B_{1,i,m}^+ \right|}{\text{Re}\{Z_{ii}\}} \frac{\sum_{l=1, l \neq i}^N \text{Re}\{Z_{ll}\} A_l^2}{\sum_{l=1, l \neq i}^N \left| \frac{1}{M} \sum_{m=1}^M B_{1,l,m}^+ \right| A_l} \quad [8]$$

and assuming $A_l = \rho \frac{\left| \frac{1}{M} \sum_{m=1}^M B_{1,l,m}^+ \right|}{\text{Re}\{Z_{ll}\}}$ where $\rho = 1 \frac{\Omega}{T}$ is introduced to keep the terms A_l

dimensionless, the ratio $\sum_{l=1, l \neq i}^N \left| \frac{1}{M} \sum_{m=1}^M B_{1,l,m}^+ \right| A_l$ becomes equal to ρ . Hence, the terms A_i are also given by

$$A_i = \rho \frac{\left| \frac{1}{M} \sum_{m=1}^M B_{1,i,m}^+ \right|}{\text{Re}\{Z_{ii}\}} \quad [9]$$

constituting the solution of Eq. [7].

Therefore, from the measurements of $B_{1,i}^+$ in the ROI and Z_{ii} , the optimal amplitudes A_i that minimize the cost function f at the ROI can be determined immediately. If a value for C that brings B_1^+ back to its original strength is added, a physical interpretation of this solution is seen when observing that its effect is to increase the driving current of the elements that contribute to the average B_1^+ field amplitude at the ROI most efficiently and reduce the driving current of the elements that do so least efficiently.

Case 2: Nonnegligible Coupling

If the coupling among the elements of the array is significant, the impedance matrix $\mathbf{Z}$ is not diagonal as in case 1 and linear algebra operations can be used to solve an equation similar to Eq. [7] of case 1.

Let $\mathbf{A}$ be the currents vector composed of the coefficients A_i , and $\mathbf{B}_1^+$ the vector containing the average values of the circularly polarized field B_1 generated by each element of the array at the location of interest. Then, we can rewrite Eq. [4] as

$$f = \frac{\sqrt{\frac{1}{2} \text{Re}\{\mathbf{A}^* \mathbf{Z} \mathbf{A}\}}}{|\mathbf{B}_1^{+T} \mathbf{A}|} \quad [10]$$

where the i -th element of the vector $\mathbf{B}_1^+$ is equal to $B_{1,i}^+ = \frac{1}{M} \sum_{m=1}^M B_{1,i,m}^+$ and superscripts $*$ and T indicate the complex conjugate and transpose operators, respectively.

Let us write

$$\mathbf{Z} = \mathbf{Z}_R + j \mathbf{Z}_I \quad [11]$$

where

$$\mathbf{Z}_R = \text{Re}\{\mathbf{Z}\} \quad [12]$$

and

$$\mathbf{Z}_I = \text{Im}\{\mathbf{Z}\} \quad [13]$$

With the definitions in Eqs. [12] and [13], we can rewrite Eq. [10] as

$$f = \frac{\sqrt{\frac{1}{2} \text{Re}\{\mathbf{A}^* \mathbf{Z}_R \mathbf{A} + j \mathbf{A}^* \mathbf{Z}_I \mathbf{A}\}}}{|\mathbf{B}_1^{+T} \mathbf{A}|} = \frac{\sqrt{\frac{1}{2} \text{Re}\{\mathbf{A}^* \mathbf{Z}_R \mathbf{A}\} + \frac{1}{2} \text{Re}\{j \mathbf{A}^* \mathbf{Z}_I \mathbf{A}\}}}{|\mathbf{B}_1^{+T} \mathbf{A}|} \quad [14]$$

We can decompose both the matrices $\mathbf{Z}_R$ and $\mathbf{Z}_I$ through the use of eigenvector matrices $\mathbf{Q}_R$ and $\mathbf{Q}_I$

$$\mathbf{Z}_R = \mathbf{Q}_R^{-1} \mathbf{D}_R \mathbf{Q}_R \quad [15]$$

$$\mathbf{Z}_I = \mathbf{Q}_I^{-1} \mathbf{D}_I \mathbf{Q}_I \quad [16]$$

where $\mathbf{D}_R$ and $\mathbf{D}_I$ are diagonal matrices containing eigenvalues of the matrices $\mathbf{Z}_R$ and $\mathbf{Z}_I$.

Because $\mathbf{Z}$ is symmetric, $\mathbf{Z}_R$ and $\mathbf{Z}_I$ are symmetric too, and as both $\mathbf{Z}_R$ and $\mathbf{Z}_I$ have all real elements, $\mathbf{Q}_R^{-1} = \mathbf{Q}_R^{*T}$ and $\mathbf{Q}_I^{-1} = \mathbf{Q}_I^{*T}$. Thus,

$$f = \frac{\sqrt{\frac{1}{2} \text{Re}\{\mathbf{A}^* \mathbf{Z}_R \mathbf{A}\} + \frac{1}{2} \text{Re}\{j \mathbf{A}^* \mathbf{Z}_I \mathbf{A}\}}}{|\mathbf{B}_1^{+T} \mathbf{A}|} = \frac{\sqrt{\frac{1}{2} \text{Re}\{\mathbf{A}^* \mathbf{Q}_R^* \mathbf{D}_R \mathbf{Q}_R \mathbf{A}\} + \frac{1}{2} \text{Re}\{j \mathbf{A}^* \mathbf{Q}_I^* \mathbf{D}_I \mathbf{Q}_I \mathbf{A}\}}}{|\mathbf{B}_1^{+T} \mathbf{A}|} \quad [17]$$

We can write $\mathbf{A}^* \mathbf{Q}_R^{*T} = (\mathbf{Q}_R \mathbf{A})^*{}^T$ and $\mathbf{A}^* \mathbf{Q}_I^{*T} = (\mathbf{Q}_I \mathbf{A})^*{}^T$

$$f = \frac{\sqrt{\frac{1}{2} \text{Re}\{(\mathbf{Q}_R \mathbf{A})^*{}^T \mathbf{D}_R (\mathbf{Q}_R \mathbf{A})\} + \frac{1}{2} \text{Re}\{j (\mathbf{Q}_I \mathbf{A})^*{}^T \mathbf{D}_I (\mathbf{Q}_I \mathbf{A})\}}}{|\mathbf{B}_1^{+T} \mathbf{A}|} = \frac{\sqrt{\frac{1}{2} \text{Re}\{(\mathbf{Q}_R \mathbf{A})^*{}^T \mathbf{D}_R (\mathbf{Q}_R \mathbf{A})\}}}{|\mathbf{B}_1^{+T} \mathbf{A}|} \quad [18]$$

because the product $(\mathbf{Q}_I \mathbf{A})^*{}^T \mathbf{D}_I (\mathbf{Q}_I \mathbf{A})$ in Eq. [17] is real as it is quadratic in form and the eigenvalues of $\mathbf{D}_I$ are real. Thus, $j(\mathbf{Q}_I \mathbf{A})^*{}^T \mathbf{D}_I (\mathbf{Q}_I \mathbf{A})$ is purely imaginary, and $\text{Re}\{j(\mathbf{Q}_I \mathbf{A})^*{}^T \mathbf{D}_I (\mathbf{Q}_I \mathbf{A})\}$ is null.

We can rewrite the denominator of Eq. [18]

$$|\mathbf{B}_1^{+T} \mathbf{A}| = |\mathbf{B}_1^{+T} \mathbf{I} \mathbf{A}| \quad [19]$$

where $\mathbf{I}$ is the identity matrix. By definition of the inverse of a matrix

$$|\mathbf{B}_1^{+T} \mathbf{I} \mathbf{A}| = |\mathbf{B}_1^{+T} \mathbf{Q}_R^{-1} \mathbf{Q}_R \mathbf{A}| \quad [20]$$

By defining $\mathbf{E} = \mathbf{Q}_R \mathbf{A}$ and $\mathbf{F} = \mathbf{B}_1^{+T} \mathbf{Q}_R^{-1}$, we have

$$f = \frac{\sqrt{\frac{1}{2} \text{Re}\{\mathbf{E}^*{}^T \mathbf{D}_R \mathbf{E}\}}}{|\mathbf{F} \mathbf{E}|} \quad [21]$$

The minimization of Eq. [21] is equivalent to that of case 1, provided that the following substitutions are made. The vector $\mathbf{E}$ is the unknown, the impedance matrix is $\mathbf{D}_R$ (equivalent to an impedance matrix with no coupling as $\mathbf{D}_R$ is a diagonal matrix), and $\mathbf{F}$ is the magnetic field vector. With these substitutions, Eq. [9] is used to find the values of $\mathbf{E}$ that minimize Eq. [21].

After $\mathbf{E}$ is obtained, the final current vector $\mathbf{A}$ is computed as

$$\mathbf{A} = \mathbf{Q}_R^{-1} \mathbf{E} \quad [22]$$

Method

The performance of the proposed algorithm for nonnegligible coupling was compared with two other methods to compute the coil currents: (1) the distribution for a birdcage coil in ideal mode 1 resonance and (2) a phase-only optimization published previously (9).

Comparisons included examinations of the magnitude of the B_1^+ field in the ROI for a given P_{Tx} , and also of the P_{Tx} required to produce a given B_1^+ for both the negligible and the nonnegligible coupling cases. In all cases, the field distributions were computed numerically at 300 MHz for a body-sized eight-element array of stripline elements spaced equidistantly on the surface of a cylinder within a large cylindrical shield and loaded with a human body

model (20) positioned with its heart near the center of the array (Fig. 1). The field distribution for each element of the array was computed with all other elements present, but with open circuit at each end to simulate a case of minimal coupling between elements, because coupling between the elements and their fields can be added later. All numerical simulations were performed using a commercially available full-wave electromagnetic field simulator (XFDTD; Remcom, Inc.; State College, PA) and with I_{ref} of 1 A. In the comparisons, the optimized coil currents were normalized by changing the value of the factor C in Eq. [2] so that either P_{Tx} or B_1^+ (as desired) in the ROI was the same for all three cases. For the uncoupled case, fields were computed with each element driven individually and as if the coupling matrix was the identity matrix. For the case with significant coupling, two appendages were applied at the extremities of each stripline to more easily induce fields among other elements of the array. The structure of the impedance matrices of both cases used in this study is reported in Figure 2. Hence, this method could be applied to an experimentally measured impedance matrix. The comparisons were performed considering a cubic ROI 5 mm on each side placed both in the heart (centrally located) and in the shoulder (peripherally located).

RESULTS

Table 1 gives the magnitude of B_1^+ for each target ROI in each of three current distributions normalized to produce a whole-body average SAR of 2 W/Kg with negligible coupling between array elements. Table 2 presents the same for the case with significant coupling. Figure 3 reports for the two ROIs, for each element of the array having negligible mutual coupling, the values of the real part of the self-impedance $\text{Re}\{Z_{ii}\}$, the average absolute value of the circularly polarized magnetic field in the two ROIs, and the optimal amplitude of the currents obtained by applying Eq. [9] scaled by the factor C to produce the fields shown in Table 1. Figures 4 and 5 show $|B_1^+|$ field distributions obtained in the cross section containing the two different ROIs for the two different cases. For an ROI in the heart and given P_{Tx} , the proposed algorithm for optimizing transmit efficiency considering both amplitude and phase of each current element produces an average B_1^+ field having amplitude 5.39 times larger than that of the birdcage coil and 1.20 times larger than that of the phase-only optimization. For an ROI in the shoulder and given P_{Tx} , the proposed algorithm produces a B_1^+ field having amplitude 2.22 times larger than that of the birdcage coil and 1.70 times larger than that of the phase-only optimization.

With the elements of the array having nonnegligible coupling among them, Table 1 gives B_1^+ for the ROI in each of three current distributions normalized to produce the same P_{Tx} . For an ROI in the heart, the proposed algorithm for optimizing transmit efficiency considering both amplitude and phase of each current element produces a B_1^+ field having amplitude 4.73 times larger than that of the birdcage coil and 1.19 times larger than that of the phase-only optimization. For an ROI in the arm, the proposed algorithm produces a B_1^+ field having amplitude 2.57 times larger than that of the birdcage coil and 2.00 times larger than that of the phase-only optimization.

Using these same numbers, it is also possible to determine the power required to produce a given B_1^+ in each case. To produce a given B_1^+ in an ROI in the heart, the proposed algorithm will require 0.034 times the power required by a birdcage coil and 0.694 times the power required by the phase-only optimization for the case of negligible coupling among the elements of the array, while it will require 0.045 times the power required by a birdcage coil and 0.706 times the power required by the phase-only optimization for the case of nonnegligible coupling among the elements of the array. To produce a given B_1^+ in an ROI in the shoulder, the proposed algorithm will require 0.203 times the power required by a birdcage coil and 0.346 times the power required by the phase-only optimization for the case of negligible coupling, while it will require 0.151 times the power required by a birdcage coil and 0.25 times the power required by the phase-only optimization for the nonnegligible coupling. For a given pulse sequence, this would translate to approximately one-fifth the whole-body SAR in the birdcage coil and one-third that in the phase-only optimization.

DISCUSSION

We have presented a simple, analytically based method for optimizing transmit efficiency of exciting a local region considering both magnitude and phase of all elements in a transmit array. For a small ROI, our method provides results that differ by only a few percent from the results obtained with a method (15,16) developed in parallel with ours (21). One advantage of our derivation is that it provides a more explicit relationship between the optimum values of the current amplitudes and the impedances and field distributions of the elements. This is evident in Figure 3, which shows the relationship between impedances, the optimal amplitudes of the currents, and the values of the fields generated by the elements of the array. As in Eq. [9], the optimal current amplitude for each element is proportional to the ratio of the B_1 field it produces in the ROI to its impedance. In the case of exciting a central location, this results in low optimal currents in elements near the arms, which have both relatively high impedance and relatively low B_1^+ in the ROI. When the ROI is in a peripheral location, however, the highest optimal current amplitudes are in the elements near the ROI due to the very low relative B_1^+ fields produced there by elements further away.

In ROIs near the center of the torso, this method is seen to perform slightly better than a previously published analytically based phase-only optimization (9). Away from the center of the array and sample, the improvement over the phase-only optimization is more dramatic. This is to be expected because elements far from the center of the ROI are likely to transmit much less efficiently than others, increasing the value of magnitude-and-phase optimization. Finally, it is interesting to note that when the coupling among the elements of the array is significant the algorithm provides a set of phases different from the ones obtained by a method designed simply to produce constructive interference.

Acknowledgments

Grant sponsor: National Institutes of Health; Grant number: R01 EB000454.

References

1. Hoult DI. The sensitivity and power deposition of the high field imaging experiment. *J Magn Reson Imaging*. 2000; 12:46–67. [PubMed: 10931564]
2. Ibrahim TS, Lee R, Baertlein BA, Abduljalil AM, Zhu H, Robitaille PL. Effect of RF coil excitation on field inhomogeneity at ultra high fields: a field optimized TEM resonator. *Magn Reson Imaging*. 2001; 19:1339–1347. [PubMed: 11804762]
3. Vaughan JT, DelaBarre L, Snyder C, et al. 9.4T Human MRI: preliminary results. *Magn Reson Med*. 2006; 56:1274–1282. [PubMed: 17075852]
4. Katscher U, Boörnert P, Leussler C, van den Brink JS. Transmit SENSE. *Magn Reson Med*. 2003; 49:144–150. [PubMed: 12509830]
5. Zhu Y. Parallel excitation with an array of transmit coils. *Magn Reson Med*. 2004; 51:775–784. [PubMed: 15065251]
6. Saekho S, Yip C-Y, Noll DC, Boada FE, Stenger VA. Fast-kz three-dimensional tailored radiofrequency pulse for reduced B1 inhomogeneity. *Magn Reson Med*. 2006; 55:719–724. [PubMed: 16526012]
7. Mao W, Smith MB, Collins CM. Exploring the limits of RF shimming for high-field MRI of the human head. *Magn Reson Med*. 2006; 56:918–922. [PubMed: 16958070]
8. Versluis MJ, Kan HE, van Buchem MA, Webb AG. Improved signal to noise in proton spectroscopy of the human calf muscle at 7 T using localized B₁ calibration. *Magn Reson Med*. 2010; 63:207–211. [PubMed: 19918906]
9. Metzger GJ, Snyder C, Akgun C, Vaughan T, Ugurbil K, Van de Moortele PF. Local B₁⁺ shimming for prostate imaging with transceiver arrays at 7T based on subject-dependent transmit phase measurements. *Magn Reson Med*. 2008; 59:396–409. [PubMed: 18228604]
10. Wang Z, Lin JC, Mao W, Liu W, Smith MB, Collins CM. SAR and temperature: simulations and comparison to regulatory limits for MRI. *J Magn Reson Imaging*. 2007; 26:437–441. [PubMed: 17654736]
11. International Electrotechnical Commission. IEC60601-2-33. 3.0. Geneva: IEC Press; 2010. Medical electrical equipment—part 2–33: particular requirements for the basic safety and essential performance of magnetic resonance equipment for diagnosis.
12. Zhu, Y. In vivo RF power and SAR calibration for multiport RF transmission. Proceedings of the 17th Annual Meeting of ISMRM; Honolulu, Hawaii, USA. 2009. p. 2585
13. Van den Berg CAT, van den Bergen B, Van de Kamer JB, Raaymakers BW, Kroeze H, Bartels LW, Lagendijk JJW. Simultaneous B₁⁺ homogenization and specific absorption rate hotspot suppression using a magnetic resonance phased array transmit coil. *Magn Reson Med*. 2007; 57:577–586. [PubMed: 17326185]
14. Brunner DO, Pruessmann KP. Optimal design of multiple-channel RF pulses under strict power and SAR constraints. *Magn Reson Med*. 2010; 63:1280–1291. [PubMed: 20432299]
15. Deniz, CM.; Brown, R.; Lattanzi, R.; Alon, L.; Sodickson, DK.; Zhu, Y. Maximum efficiency RF shimming. Proceedings of the 20th Annual Meeting of ISMRM; Melbourne, Australia. 2012. p. 3479
16. Deniz CM, Brown R, Lattanzi R, Alon L, Sodickson DK, Zhu Y. Maximum efficiency RF shimming: theory and initial application for hip imaging at 7 Tesla. *Magn Reson Med*. 2013; 69:1379–1388. [PubMed: 22714835]
17. Abraham R, Ibrahim T. Proposed radiofrequency phased-array excitation scheme for homogenous and localized 7-tesla whole-body imaging based on full-wave numerical simulations. *Magn Reson Med*. 2007; 57:235–242. [PubMed: 17260366]
18. Kurpad, KN.; Boskamp, EB.; Wright, SM. Implementation of coil integrated RF power MOSFET as a voltage controlled current source in a transmit phased array coil. Proceedings of the 12th Annual Meeting of ISMRM; Kyoto, Japan. 2004. p. 1585
19. Yarnykh VL. Actual flip-angle imaging in the pulsed steady state: a method for rapid three-dimensional mapping of the transmitted radiofrequency field. *Magn Reson Med*. 2007; 57:192–200. [PubMed: 17191242]

20. Christ A, Kainz W, Hahn EG, et al. The virtual family—development of surface-based anatomical models of two adults and two children for dosimetric simulations. *Phys Med Biol*. 2010; 55:N23–N38. [PubMed: 20019402]
21. Carluccio, G.; Collins, CM.; Erricolo, D. An analytical method to optimize transmit efficiency for local excitation with a transmit array. *Proceedings of the 20th Annual Meeting of ISMRM*; Melbourne, Australia. 2012. p. 2618

FIG. 1.

Geometry of the model used in simulations: a body-sized eight-element array of stripline elements spaced equidistantly on the surface of a cylinder within a large cylindrical shield and loaded with a human body model positioned with its heart near the center of the array. [Color figure can be viewed in the online issue, which is available at wileyonlinelibrary.com.]

FIG. 2. Plot of the amplitude of the impedance matrix for the transmit array in cases of weak coupling (left) and strong coupling (right). [Color figure can be viewed in the online issue, which is available at wileyonlinelibrary.com.]

FIG. 3.

Graphical representation of the “Optimization with phase and amplitude” case reported in Table 1. For each element of the array having negligible mutual coupling are provided the following. First row: the values of the real part of the self-impedance $Re\{Z_{ii}\}$; second row: the average absolute values of the circularly polarized magnetic field in the two ROI, indicated by the blue circle; and, third row, the amplitudes of the currents scaled to produce the fields shown in Table 1. [Color figure can be viewed in the online issue, which is available at wileyonlinelibrary.com.]

FIG. 4.

Spatial distribution of the magnetic field $|B_1^+|$ obtained with the transmit array having weakly coupled elements driven in the three configurations. For each location of interest (indicated with a black circle), the three $|B_1^+|$ field distributions have been normalized to generate a whole-body average SAR equal to 2 W/kg.

FIG. 5.

Spatial distribution of the magnetic field $|B_1^+|$ obtained with the transmit array having strongly coupled elements driven in the three configurations. For each location of interest (indicated with a black circle), the three $|B_1^+|$ field distributions have been normalized to generate a whole-body average SAR equal to 2 W/kg.

Table 1

Magnitude of the B_1^+ Field Produced by Three Different Current Distributions Including Optimizations for an ROI near the Heart and in the Shoulder for a Transmit Array Having Negligible Coupling Between its Elements

	$\frac{1}{M} \sum_{m=1}^M B_{1,i,m}^+ $ for a ROI near heart (μT)	$\frac{1}{M} \sum_{m=1}^M B_{1,i,m}^+ $ for a ROI near arm (μT)
Birdcage	0.4418	3.5181
Phase-only optimization	1.9738	4.6195
Optimization with phase and amplitude	2.3800	7.8447

In each case, whole-body average SAR is 2 W/kg.

Table 2

Magnitude of the B_1^+ Field Produced by Three Different Current Distributions Including Optimizations for an ROI near the Heart and in the Shoulder for a Transmit Array Having Significant Coupling Between its Elements

	$\frac{1}{M} \sum_{m=1}^M B_{1,i,m}^+ $ for a ROI near heart (μT)	$\frac{1}{M} \sum_{m=1}^M B_{1,i,m}^+ $ for a ROI near arm (μT)
Birdcage	0.4147	3.3904
Phase-only optimization	1.6458	4.3660
Optimization with phase and amplitude	1.9617	8.7185

In each case, whole-body average SAR is 2 W/kg.