Preface

The deleterious effects of feedback is experienced by virtually all who are involved with providing appropriate amplification to their patients. The presence of feedback is an expensive problem for hearing aid and earmold manufacturers who spend an unknown amount annually remaking shells and earmolds because feedback is preventing adequate amplification. Feedback is a problem for spouses and friends who constantly remind the user that their hearing aids are “whistling.” Feedback is a problem for the user because of the “embarrassment” associated when they are reminded that their hearing aids are “whistling.” Feedback is also a problem for the user because the patient is forced to reduce the required amplification in order to “quiet” the hearing aids. This solution typically results in inappropriate amplification. Finally, feedback is, of course, a problem for the dispenser for all of the reasons described above.

The current issue deals with providing the reader with a better understanding of the many causes of the numerous types of feedback. It also offers the clinician a wealth of information on how to resolve feedback using both acoustic and electronic methods. Finally, it summarizes some of the recent attempts by manufacturers to resolve this problem as well as indicates strategies manufacturers may use to address this problem in the future.

This issue is authored by Jeremy Agnew, Ph.D., Director of Product Development for Starkey Laboratories, Colorado Springs, CO. Dr. Agnew has previously been the engineering director of three other hearing aid manufacturers and has been involved in the research and development of hearing aids, tinnitus maskers, audiometric equipment, nerve stimulators, auditory screeners and implantable hearing aids. He has previously been on the advisory and consulting boards of several medical device companies. He is a member of ANSI S3.48 Hearing Aid Standards Committee, ANSI S3.80 Real Ear Measurement Standards Committee, the Joint IEC/ISO Working Group for Real Ear Measurements and is the United States representative to the IEC WG13 Working Group on Hearings Aids. He has authored a textbook on hybrid thick film microelectronics as well as published over 50 publications in electronics, acoustics, audiology and hearing aid technology. In addition, he has made over 30 presentations at professional meetings in the areas of electronics, audiology and hearing aid technology.

Michael Valente, Ph.D.
Editor-in-Chief