RESEARCH ARTICLE

Trabecular bone histomorphometric measurements and contrast-to-noise ratio in CBCT

^{1,2}E Klintström, ^{1,2}Ö Smedby, ²B Klintström, ³T B Brismar and ^{1,2}R Moreno

¹Center for Medical Image Science and Visualization (CMIV), Linköping University, Linköping, Sweden; ²Department of Radiology and Department of Medical and Health Sciences, Linköping University, Linköping, Sweden; ³Department of Clinical Science, Intervention and Technology at Karolinska Institutet and Department of Radiology, Karolinska University Hospital Huddinge, Stockholm, Sweden

Objectives: The aim of this study was to evaluate how imaging parameters at clinical dental CBCT affect the accuracy in quantifying trabecular bone structures, contrast-to-noise ratio (CNR) and radiation dose.

Methods: 15 radius samples were examined using CBCT (Accuitomo FPD; J. Morita Mfg., Kyoto, Japan). Nine imaging protocols were used, differing in current, voltage, rotation degree, voxel size, imaging area and rotation time. Radiation doses were measured using a kerma area product-meter. After segmentation, six bone structure parameters and CNRs were quantified. Micro-CT (μ CT) images with an isotropic resolution of 20 μ m were used as a gold standard.

Results: Structure parameters obtained by CBCT were strongly correlated to those by μ CT, with correlation coefficients >0.90 for all studied parameters. Bone volume and trabecular thickness were not affected by changes in imaging parameters. Increased tube current from 5 to 8 mA, decreased isotropic voxel size from 125 to 80 μ m and decreased rotation angle from 360° to 180° affected correlations for trabecular termini negatively. Decreasing rotation degree also weakened correlations for trabecular separation and trabecular number at 80 μ m voxel size. Changes in the rotation degree and tube current affected CNR significantly. The radiation dose varied between 269 and 1153 mGy cm².

Conclusions: Trabecular bone structure can be accurately quantified by clinical dental CBCT *in vitro*, and the obtained structure parameters are strongly related to those obtained by μ CT. A fair CNR and strong correlations can be obtained with a low radiation dose, indicating the possibility for monitoring trabecular bone structure also *in vivo*.

Dentomaxillofacial Radiology (2014) 43, 20140196. doi: 10.1259/dmfr.20140196

Cite this article as: Klintström E, Smedby Ö, Klintström B, Brismar TB, Moreno R. Trabecular bone histomorphometric measurements and contrast-to-noise ratio in CBCT. *Dentomaxillofac Radiol* 2014; 43: 20140196.

Keywords: CBCT; micro-computed tomography; trabecular bone; histomorphometry; bone segmentation; osteoporosis

Introduction

CBCT is a clinical equipment, often used in mandible, maxillofacial and temporal bone imaging.¹ There have been an increased number of CBCT machines in Europe during the past decade, and the method is considered to be an accurate imaging modality in dental diagnostics such

as orthodontics, implantology and periapical pathology.^{2,3} There has been increasing focus on using mandible images also for osteoporosis research, with the potential for osteoporotic screening. The high resolution and the isotropic voxels of the CBCT (75–400 μ m) makes it particularly suitable for imaging the small structures in the craniofacial region, including the temporal bone.^{4–6}

Correspondence to: Dr Eva Klintström. E-mail: eva.klintstrom@lio.se

Received 8 June 2014; revised 18 August 2014; accepted 27 August 2014

Table 1 Imaging parameters

Name	Tube voltage (kV)	Tube current (mA)	Rotation (degree)	Scanning time (s)	Imaging voxel size (μm)	Field of view (mm)
FPD 170	85	5	360	30.8	125	60×60
FPD 80 scan 1	85	5	360	17.5	125	60×60
FPD 80 scan 2	85	5	360	17.5	80	40×40
FPD 80 scan 3	85	8	360	17.5	125	60×60
FPD 80 scan 4	85	8	360	17.5	80	40×40
FPD 80 scan 5	90	8	360	17.5	80	40×40
FPD 80 scan 6	85	8	360 (from FPD 80 scan 3)	17.5	125 (reconstructed 80)	60×60
FPD 80 scan 7	85	8	180 (from FPD 80 scan 3)	9.0	125	60×60
FPD 80 scan 8	85	8	180 (from FPD 80 scan 4)	9.0	80	40×40

Accutomo FPD; J. Morita Mfg., Kyoto, Japan.

Some of the advantages of CBCT machines are a relatively low investment cost and the low radiation dose⁷ compared with multislice CT machines. There are many manufacturers in the market for CBCT, and devices for patients in sitting, standing and lying positions exist. The equipment used for patients lying down requires more space and is less widely used than devices for patients in the sitting and standing positions (<http://www.sedentext.eu/content/comparison-cbct-machines?sort=asc&order=Position>).

Prior to imaging, the operator/radiologist needs to decide several imaging parameters, such as field of view (FOV), tube current, tube voltage and voxel size. It is also possible to adjust the rotation angle and the rotation speed. All these parameters must be carefully considered since they affect both image quality and radiation dose to the patient.^{8,9}

Studies investigating subjective image quality and visibility of anatomical structures in the mandible comparing CBCT devices and multi-slice CT show that there seems

Figure 1 CBCT images; slice from original image (a), slice from segmented image (b), volume-rendered segmented three-dimensional (3D) image (c) and volume-rendered skeleton of segmented 3D image (d).

Figure 2 Micro-CT images; slice from original image (a), slice from segmented image (b), volume-rendered segmented three-dimensional (3D) image (c) and volume rendered skeleton of segmented 3D image (d).

to be rather big differences between the two types of scanners.¹⁰ However, there are only a few published studies on the influence of imaging parameters such as the voxel size, rotation speed and tube settings on image quality, expressed as the contrast-to noise ratio (CNR) on CBCT.¹¹

Trabecular bone structure is of great importance for bone strength.¹² Quantification of trabecular structure can be of value when determining the risk of osteoporotic bone^{13,14} as well as in the stability of tooth implants.^{15,16} There are, to our knowledge, only a few articles describing the possibility of using CBCT to image trabecular micro-architecture where these measurements have been verified using micro-CT (μ CT),^{17–20} which can be regarded as a gold standard.²¹

The aim of this *in vitro* study was to evaluate how different imaging settings in CBCT may affect the accuracy in quantifying trabecular bone structure by comparing the results to those of a gold standard, μ CT. The study also aimed to identify differences in CNR between the settings and relate them to the radiation dose.

Methods and materials

Material

For this study, 15 radius specimens (human forearm) were taken from human cadavers that were donated

for medical research in accordance with the ethical guidelines regulating such donations. The samples are nearly cubic with a side of approximately 10–12 mm. Each cube included a slab of cortical bone, which facilitated the orientation. The bone samples were defatted and placed in a test tube filled with water. During imaging with CBCT, the different test tubes were placed in the centre of a paraffin cylinder with a diameter of 100 mm, used to mimic soft tissue, to simulate measurements *in vivo*. After the imaging, cubes with a side of approximately 8 mm consisting only of trabecular bone were digitally extracted from each data set. These cubes were used for analysis.

Image acquisition and reconstruction

The specimens were imaged using two different CBCT models, Accuitomo FPD 80 and Accuitomo FPD 170 (J. Morita Mfg., Kyoto, Japan). The imaging parameters were adjusted so that there would be pairs of scans differing in only one of the parameters: tube current, tube voltage, rotation degree angle, voxel size, imaging area or rotation time (Table 1).

The μ CT data were acquired with a small desktop CT used for analysing biopsies and other small specimens (μ CT 40; SCANCO Medical AG, Bassersdorf, Switzerland), where the isotropic voxel size was 20 μ m and the tube current was set to 70 kVp.

Figure 3 Segmentation detail. Original CBCT image (a), enlarged original image (b) and graphical illustration of segmented image (c).

Image processing

In this study, we have used the automated three-dimensional (3D) region growing algorithm (ARG) based on an assessment function²² to segment bone from other tissues. To obtain a binary image, the voxels identified as bone were assigned the value one and the remaining voxels, the value zero. The segmentation method starts with a very restrictive threshold for what is defined as bone, which results in an undersegmented area. The process is then repeated with increasing threshold values and iterated a specified number of times until an over-segmented region is obtained. The segmentation where the assessment function attained its minimum was then used to calculate the parameters. For demonstration, images of raw image slices, segmented image slices, skeletonized images and 3D images from CBCT and μ CT data are shown in Figures 1 and 2.

Six bone structure parameters were measured and calculated using MATLAB[®] (MathWorks[®], Natick, MA) code developed in-house as described in a previous study from our group.¹⁷ The definition of how the six 3D bone structure parameters are defined is explained in Figure 3 and the parameters are listed below (Figure 4):

- Bone volume over total volume (BV/TV) is measured by dividing the number of voxels classified as bone trabeculae by the total number of voxels in the bone sample (Figure 4a).
- Trabecular thickness (Tb.Th) is measured in millimetres, the mean trabecular bone diameter (Figure 4b).
- Trabecular separation (Tb.Sp), also measured in millimetres, the mean distance between the borders of the segmented trabeculae (Figure 4c).
- Trabecular number (Tb.N) is the number of the trabeculae, measured as the inverse of the mean spacing between the midlines of the trabeculae (Figure 4c).

Figure 4 Definition of bone structure parameters. Bone volume over total volume (BV/TV) (a). Trabecular thickness (Tb.Th), number of trabecular nodes (Tb.Nd) and number of trabecular termini (Tb.Tm) (b). Trabecular separation (Tb.Sp) and trabecular number (Tb.N) (c).

Table 2 Measurements of trabecular bone structure parameters

Name	Trabecular separation (mm)	Trabecular thickness (mm)	Trabecular number (mm^{-3})	Bone volume over total volume	Trabecular nodes (mm^{-3})	Trabecular termini (mm^{-3})
FPD 170	0.58 ± 0.10	0.43 ± 0.03	0.90 ± 0.09	0.35 ± 0.10	0.85 ± 0.21	0.37 ± 0.09
FPD 80 scan 1	0.59 ± 0.10	0.46 ± 0.04	0.87 ± 0.09	0.36 ± 0.10	0.77 ± 0.18	0.38 ± 0.11
FPD 80 scan 2	0.46 ± 0.07	0.40 ± 0.05	1.03 ± 0.10	0.38 ± 0.11	1.79 ± 0.43	1.39 ± 0.30
FPD 80 scan 3	0.59 ± 0.11	0.48 ± 0.04	0.87 ± 0.09	0.38 ± 0.12	0.76 ± 0.19	0.41 ± 0.19
FPD 80 scan 4	0.48 ± 0.07	0.42 ± 0.04	0.99 ± 0.10	0.38 ± 0.10	1.66 ± 0.40	1.26 ± 0.31
FPD 80 scan 5	0.51 ± 0.08	0.41 ± 0.04	0.99 ± 0.11	0.37 ± 0.10	1.72 ± 0.44	1.24 ± 0.44
FPD 80 scan 6	0.50 ± 0.08	0.41 ± 0.04	1.00 ± 0.10	0.38 ± 0.10	1.75 ± 0.36	1.28 ± 0.31
FPD 80 scan 7	0.58 ± 0.09	0.44 ± 0.03	0.89 ± 0.09	0.37 ± 0.10	0.86 ± 0.19	0.42 ± 0.10
FPD 80 scan 8	0.42 ± 0.04	0.37 ± 0.03	1.11 ± 0.07	0.38 ± 0.09	2.04 ± 0.38	1.68 ± 0.23
μCT	0.65 ± 0.12	0.13 ± 0.01	1.17 ± 0.16	0.10 ± 0.03	5.11 ± 1.67	0.83 ± 0.27

μCT , micro-CT.

Values are given as mean $\pm$ standard deviations.

Accutomo FPD; J. Morita Mfg., Kyoto, Japan.

- Trabecular nodes (Tb.Nd) is the number of trabecular intersections per volume (Figure 4b).
- Trabecular termini (Tb.Tm) is the number of free ends of trabeculae per volume (Figure 4b).

The CNR was also measured in the 3D volume using MATLAB code developed in-house on the same standard personal computer (PC). The CNR was calculated as the difference in signal from the trabecular bone and the signal in the background (*i.e.* the surrounding water) divided by the standard deviation of the noise in the background. The segmented skeleton represented bone, and the background was eroded with two voxels prior to the calculations to avoid the influence of the partial volume effect.

A standard PC with Intel® Core™ i5 (Intel, Santa Clara, CA), CPU at 2.60 GHz, 4-GB random access memory and 64-bit operating system was used for all the calculations.

Radiation doses were measured using a kerma area product (KAP)-meter (Electrometer Doseguard 100; RTI Electronics AB, Mölndal, Sweden and Ion chamber Vacutec model No 70157) calibrated at the Swedish radiation safety authority at Riksmätplatsen, Stockholm, Sweden.

Statistical methods

Results are presented as mean values with standard deviations. Parameters were related to each other using Pearson correlation with 95% confidence intervals. Differences between measures were tested with the paired Student's *t*-test (two-tailed).

Figure 5 Trabecular termini calculated from six CBCT protocols (excluding resampled protocols) in relationship to micro-CT (μCT). Accutomo FPD; J. Morita Mfg., Kyoto, Japan. Tb.Tm, trabecular termini.

Figure 6 Trabecular number calculated from six CBCT protocols (excluding resampled protocols) in relationship to micro-CT (μCT). Accutomo FPD; J. Morita Mfg., Kyoto, Japan. Tb.N, trabecular number.

Table 3 Correlations with micro-CT

Name	Trabecular separation	Trabecular thickness	Trabecular number	Bone volume over total volume	Trabecular nodes	Trabecular termini
FPD 170	0.92 (0.77;0.97)	0.94 (0.81;0.98)	0.95 (0.84;0.98)	0.97 (0.91;0.99)	0.89 (0.70;0.96)	0.91 (0.75;0.97)
FPD 80 scan 1	0.92 (0.77;0.97)	0.92 (0.77;0.97)	0.93 (0.81;0.98)	0.98 (0.94;0.99)	0.89 (0.70;0.96)	0.92 (0.76;0.97)
FPD 80 scan 2	0.94 (0.82;0.98)	0.94 (0.82;0.98)	0.96 (0.87;0.99)	0.98 (0.93;0.99)	0.92 (0.77;0.97)	0.83 (0.55;0.94)
FPD 80 scan 3	0.89 (0.70;0.96)	0.94 (0.82;0.98)	0.92 (0.77;0.97)	0.96 (0.89;0.99)	0.87 (0.66;0.96)	0.87 (0.64;0.96)
FPD 80 scan 4	0.91 (0.73;0.97)	0.96 (0.87;0.99)	0.97 (0.89;0.99)	0.97 (0.91;0.99)	0.93 (0.80;0.98)	0.74 (0.36;0.91)
FPD 80 scan 5	0.89 (0.68;0.96)	0.94 (0.83;0.98)	0.96 (0.89;0.99)	0.96 (0.90;0.99)	0.95 (0.85;0.98)	0.73 (0.34;0.90)
FPD 80 scan 6	0.82 (0.52;0.94)	0.91 (0.76;0.97)	0.94 (0.82;0.98)	0.95 (0.86;0.98)	0.88 (0.67;0.96)	0.76 (0.41;0.92)
FPD 80 scan 7	0.83 (0.55;0.94)	0.93 (0.81;0.98)	0.91 (0.76;0.97)	0.94 (0.82;0.98)	0.84 (0.57;0.95)	0.76 (0.40;0.91)
FPD 80 scan 8	0.74 (0.37;0.91)	0.92 (0.76;0.97)	0.87 (0.65;0.96)	0.96 (0.88;0.99)	0.95 (0.85;0.98)	0.67 (0.25;0.88)

Values are given as Pearson correlation coefficients (*r*) with 95% confidence limits.

Bold figures denote correlations >0.90.

Accutomo FPD; J. Morita Mfg., Kyoto, Japan.

Results

CBCT overestimated BV/TV and Tb.Th more than three times compared with μ CT; on the other hand, Tb.Nd, Tb.Sp and Tb.N were underestimated (Table 2). Tb.Tm was overestimated by the 80- μ m voxels and underestimated by the 125- μ m voxels (Table 2, Figure 5). Regarding the parameter Tb.N, the underestimation was greater for the 125- μ m voxels than for the 80- μ m voxels (Table 2, Figure 6).

Measurements from CBCT showed strong correlation with μ CT for all studied bone parameters, with two settings having correlation coefficients of 0.89 or higher for all bone parameters (Table 3). The most stable parameters, BV/TV and Tb.Th, had correlations of at least 0.91 regardless of the settings. The parameter Tb.Tm had lower correlations and varied more with changes in the settings (0.92–0.67).

Despite the higher radiation dose and longer scan time (30.8 vs 17.5 s), the FPD 170 high resolution and the FPD 80 scan 1 resulted in almost identical correlation coefficients. However, reducing the radiation dose by decreasing the rotation angle from 360° to 180°, with a shorter scan time (17.5 s vs 9.0 s), resulted in weaker correlations, especially regarding Tb.Sp, Tb.N and Tb.Tm. The parameters BV/TV, Tb.Th and Tb.Nd were not equally affected. Increasing the tube current from 5 to 8 mA led to weaker correlations for Tb.Sp and Tb.Tm. Changing the tube voltage from 85 to 90 kV had nearly no impact on the correlation coefficients. For the 360° rotation angle, the 80- μ m voxel size showed slightly

stronger correlation with μ CT for all parameters except Tb.Tm when compared with the 125- μ m voxel size.

Contrast-to-noise ratio

The CNR values for the different protocols can be seen in Table 4 and the corresponding significance tests in Table 5.

When increasing the tube current from 5 to 8 mA, there was a statistically significant increase in CNR for the 125- μ m voxels but not for the 80- μ m voxels. Reconstructing 180° from a 360° rotation angle scan resulted in significantly lower CNR values for the 80- μ m voxels, but not for the 125- μ m voxels. Differences in resolution mode, reconstruction, tube voltage and voxel sizes did not have any significant impact on the CNR (Table 5).

Radiation dose

The doses measured by the KAP meter, the mean of the correlation values with μ CT for all the studied histomorphometric parameters and the CNR values for the nine different protocols can be seen in Table 6. Radiation doses measured in kerma-area product varied between 269 and 1284 mGy cm².

The FPD 170, FPD 80 scan 1 and FPD 80 scan 2 all had a mean correlation value for the trabecular parameters of 0.93 but showed big differences in radiation dose.

Discussion

CBCT acquisition settings can be adjusted in almost uncountable ways. We use the protocol FPD 80 scan 2 (Table 1) in pre-implant planning and for studying periapical pathology as clinical routine at our institution. By adjusting this protocol in a stepwise order, it was possible to elucidate how the quantification of trabecular bone structure can be optimized without increasing the radiation dose unacceptably. Those settings should also be useful for clinical imaging of small structures in the craniofacial region.

The parameters BV/TV and Tb.Th were the most stable ones and had the strongest correlations with those of μ CT, varying from 0.91 to 0.98, regardless of the settings. Tb.Tm, on the other hand, seemed to be the most sensitive parameter and was the only parameter that with a 360° rotation showed stronger correlations

Table 4 Contrast-to-noise ratio (CNR)

Name	CNR
FPD 170	8.57 ± 1.97
FPD 80 scan 1	9.28 ± 1.92
FPD 80 scan 2	8.43 ± 1.46
FPD 80 scan 3	10.75 ± 2.35
FPD 80 scan 4	9.68 ± 1.82
FPD 80 scan 5	9.76 ± 1.71
FPD 80 scan 6	9.32 ± 1.66
FPD 80 scan 7	8.07 ± 1.37
FPD 80 scan 8	7.06 ± 1.17

Values are given as mean ± standard deviations.

Accutomo FPD; J. Morita Mfg., Kyoto, Japan.

Table 5 Differences in contrast-to-noise ratio values (paired *t*-test)

Compared imaging parameters	Name	t-value	p-value
Tube current (8–5 mA)	FPD 80 scan 3 – FPD 80 scan 1	3.205	0.006
	FPD 80 scan 4 – FPD 80 scan 2	2.090	0.055
Rotation degree (360°–180°)	FPD 80 scan 3 – FPD 80 scan 7	1.970	0.069
	FPD 80 scan 4 – FPD 80 scan 8	2.869	0.012
Voxel size (125–80 μm)	FPD 80 scan 3 – FPD 80 scan 4	1.450	0.169
	FPD 80 scan 7 – FPD 80 scan 8	1.373	0.191
Mode (high resolution—normal)	FPD 170 – FPD 80 scan 1	−0.599	0.559
Tube voltage	FPD 80 scan 5 – FPD 80 scan 4	0.164	0.872
Reconstruction	FPD 80 scan 6 – FPD 80 scan 4	−0.699	0.496
	FPD 80 scan 6 – FPD 80 scan 3	−1.645	0.122

Values provided are *t*-values and *p*-values (two tailed).

Bold numbers signify significant differences (*p* < 0.05).

Accutomo FPD; J. Morita Mfg., Kyoto, Japan.

for the 125-μm voxels than for the 80-μm voxels. The overestimation of Tb.Tm and the weaker correlations regarding Tb.Tm for the 80 μm voxels may be due to noise because of the smaller imaged FOV. To examine this, we reconstructed the raw data from a 60 × 60-mm FOV and 125-μm voxels, to 80-μm voxels. This resulted in correlations comparable with the other scans with 80-μm voxels and the same overestimation. We therefore concluded that those differences were linked to the voxel size rather than to the smaller FOV.

The high-resolution mode of the Accutomo 170 FPD model, which involves a longer scanning time, did not result in stronger correlations despite the increased dose. Increasing the tube current, which also increases the dose, did not have any substantial impact on most of the parameters. In fact, the only impact of increased tube current was slightly weaker correlation coefficients regarding Tb.Tm and Tb.Sp. To decrease radiation dose, it is possible to use only a 180° rotation angle at the scanning. When reconstructing 180° images from a 360° scan, the correlation became clearly weaker for Tb.Sp, Tb.N and Tb.Tm regarding the 80-μm voxels. In conclusion, increasing the dose did not result in stronger correlations, but decreasing it by 50% affected some of the parameters negatively.

Table 6 Mean correlation, radiation dose in kerma–area product (PKA) and contrast-to-noise ratios (CNRs)

Name	Mean correlation	PKA (mGy cm ²)	CNR (mean ± standard deviation)
FPD 170	0.93	1284	8.57 ± 1.97
FPD 80 scan 1	0.93	728	9.28 ± 1.92
FPD 80 scan 2	0.93	331	8.43 ± 1.46
FPD 80 scan 3	0.91	1153	10.75 ± 2.35
FPD 80 scan 4	0.91	527	9.68 ± 1.82
FPD 80 scan 5	0.91	587	9.76 ± 1.71
FPD 80 scan 6	0.88	1153	9.32 ± 1.66
FPD 80 scan 7	0.87	269	8.07 ± 1.37
FPD 80 scan 8	0.85	591	7.06 ± 1.17

The mean correlation denotes the mean value of the correlation coefficients with micro-CT for all the studied histomorphometric parameters (trabecular separation, trabecular thickness, trabecular number, bone volume over total volume, trabecular nodes and trabecular termini). Accutomo FPD; J. Morita Mfg., Kyoto, Japan.

Changing the tube voltage from 85 to 90 kV did hardly at all affect the correlations with μCT for any of the studied parameters. This suggests that the lower tube voltage can safely be used for imaging small structures using CBCT, without compromising image quality.

Using the same machine, CNR increased with increasing dose. If the tube current was kept constant at 8 mA, a higher CNR resulted in higher correlation values (Table 6, Figure 7). However, the protocols using 5 mA, like our clinically used protocol (FPD 80 scan 2) obtained strong correlation with μCT despite comparatively low CNR and kerma–area product values.

Currently, CBCT is an often-used imaging technique in dental implant planning as well as in other pre-operative examinations. In previous years, articles describing the possibility to describe bone density using CBCT^{23–25}

Figure 7 Contrast to noise ratio (CNR) related to mean correlation. The mean correlation denotes the mean value of the correlation coefficients with micro-CT for all the studied histomorphometric parameters (Tb.Sp, Tb.Th, Tb.N, BV/TV, Tb.Nd and Tb.Tm). Accutomo FPD; J. Morita Mfg., Kyoto, Japan.

have been published. Also, bone strength related to trabecular bone structure parameters have been studied,²⁶ and a recent study showed the possibility to identify women at risk of future fractures by observing trabecular patterns in the mandible in panoramic radiographs.²⁷ However, inflammation owing to diseases of the teeth will affect the trabecular parameters and dental status will therefore also influence the accuracy to predict future fractures using bone structure analysis of the mandible.

This study has some limitations. One major limitation is that specimens from the radius and not the mandible were used. Specimens of the radius have, however, previously been used in many studies describing trabecular bone microarchitecture.^{17,28,29} The similarity between the radius and mandible (trabecular bone enclosed in compact bone) suggests that the results may also be applicable for the mandible. This may imply the possibility to use CBCT images for studies on bone structure in the mandible aiming also for osteoporotic screening. Another limitation is the use of only two models from a single manufacturer. We are currently integrating our algorithms into picture archiving and communication systems to provide a framework for multicentre studies with the possibility of including different brands and models of CBCT machines.

Since this study was made *in vitro*, it does not consider movement artefacts. Thus, the influence of increased imaging time (rotation speed and rotation degree angle) has not been investigated. Furthermore, our phantom model into which the bone samples were inserted cannot

perfectly match that of the human neck or skull. For example, artefacts from one side of the mandible may affect the image quality on the examined other side of the bone. Future *in vivo* studies are therefore needed to evaluate these factors.

There are numerous different parameters that can be used to describe trabecular bone structure. In this study, we have analysed six of the most frequently used parameters. In our continued research, we aim to include the analysis of specimens from different skeletal sites and the computation of other descriptors such as the structure model index,³⁰ degree of anisotropy using different fabric tensors³¹ and connectivity density.³²

Conclusion

The protocol most often used in our clinical practice showed strong correlations with μ CT, fair CNR and low radiation dose. Answering the question whether or not this protocol is the most appropriate for studying trabecular bone structure *in vivo* is part of our ongoing research.

Acknowledgments

Andres Laib at Scanco Medical AG in Switzerland performed the μ CT imaging. Michael Sandborg at Linköping University performed the radiation dose measurements.

References

1. Arai Y, Tammsalo E, Iwai K, Hashimoto K, Shinoda K. Development of a compact computed tomographic apparatus for dental use. *Dentomaxillofac Radiol* 1999; **28**: 245–8. doi: [10.1038/sj/dmfr/4600448](https://doi.org/10.1038/sj/dmfr/4600448)
2. Scarfe WC, Farman AG, Sukovic P. Clinical applications of cone-beam computed tomography in dental practice. *J Can Dent Assoc* 2006; **72**: 75–80.
3. Lofthag-Hansen S, Huuonen S, Gröndahl K, Gröndahl HG. Limited cone-beam CT and intraoral radiography for the diagnosis of periapical pathology. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod* 2007; **103**: 114–19. doi: [10.1016/j.tripleo.2006.01.001](https://doi.org/10.1016/j.tripleo.2006.01.001)
4. Miracle AC, Mukherji SK. Conebeam CT of the head and neck, part 2: clinical applications. *AJNR Am J Neuroradiol* 2009; **30**: 1285–92. doi: [10.3174/ajnr.A1654](https://doi.org/10.3174/ajnr.A1654)
5. Miracle AC, Mukherji SK. Conebeam CT of the head and neck, part 1: physical principles. *AJNR Am J Neuroradiol* 2009; **30**: 1088–95. doi: [10.3174/ajnr.A1653](https://doi.org/10.3174/ajnr.A1653)
6. Peltonen LI, Aarnisalo AA, Kortensniemi MK, Suomalainen A, Jero J, Robinson S. Limited cone-beam computed tomography imaging of the middle ear: a comparison with multislice helical computed tomography. *Acta Radiol* 2007; **48**: 207–12. doi: [10.1080/02841850601080465](https://doi.org/10.1080/02841850601080465)
7. Ludlow JB, Ivanovic M. Comparative dosimetry of dental CBCT devices and 64-slice CT for oral and maxillofacial radiology. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod* 2008; **106**: 106–14. doi: [10.1016/j.tripleo.2008.03.018](https://doi.org/10.1016/j.tripleo.2008.03.018)
8. Ibrahim N, Parsa A, Hassan B, van der Stelt P, Aartman IH, Wismeijer D. The effect of scan parameters on cone beam CT trabecular bone microstructural measurements of human mandible. *Dentomaxillofac Radiol* 2013; **42**: 20130206. doi: [10.1259/dmfr.20130206](https://doi.org/10.1259/dmfr.20130206)
9. Pauwels R, Beinsberger J, Collaert B, Theodorakou C, Rogers J, Walker A, et al. Effective dose range for dental cone beam computed tomography scanners. *Eur J Radiol* 2012; **81**: 267–71. doi: [10.1016/j.ejrad.2010.11.028](https://doi.org/10.1016/j.ejrad.2010.11.028)
10. Liang X, Jacobs R, Hassan B, Li L, Pauwels R, Corpas L, et al. A comparative evaluation of cone beam computed tomography (CBCT) and multi-slice CT (MSCT) part I. On subjective image quality. *Eur J Radiol* 2010; **75**: 265–9. doi: [10.1016/j.ejrad.2009.03.042](https://doi.org/10.1016/j.ejrad.2009.03.042)
11. Bechara B, McMahan CA, Moore WS, Noujeim M, Geha H, Teixeira FB. Contrast-to-noise ratio difference in small field of view cone beam computed tomography machines. *J Oral Sci* 2012; **54**: 227–32. doi: [10.2334/josnurd.54.227](https://doi.org/10.2334/josnurd.54.227)
12. Ulrich D, Van Rietbergen B, Laib A, Rüeggsegger P. The ability of three-dimensional structural indices to reflect mechanical aspects of trabecular bone. *Bone* 1999; **25**: 55–60. doi: [10.1016/s8756-3282\(99\)00098-8](https://doi.org/10.1016/s8756-3282(99)00098-8)
13. Link TM, Majumdar S, Lin JC, Newitt D, Augat P, Ouyang X, et al. A comparative study of trabecular bone properties in the spine and femur using high resolution MRI and CT. *J Bone Miner Res* 1998; **13**: 122–32. doi: [10.1359/jbmr.1998.13.1.122](https://doi.org/10.1359/jbmr.1998.13.1.122)
14. Boutry N, Cortet B, Dubois P, Marchandise X, Cotten A. Trabecular bone structure of the calcaneus: preliminary *in vivo* MR imaging assessment in men with osteoporosis. *Radiology* 2003; **227**: 708–17. doi: [10.1148/radiol.2273020420](https://doi.org/10.1148/radiol.2273020420)
15. Sakka S, Coulthard P. Bone quality: a reality for the process of osseointegration. *Implant Dent* 2009; **18**: 480–5. doi: [10.1097/ID.0b013e3181bb840d](https://doi.org/10.1097/ID.0b013e3181bb840d)
16. Fanuscu MI, Chang TL. Three-dimensional morphometric analysis of human cadaver bone: microstructural data from maxilla

- and mandible. *Clin Oral Implants Res* 2004; **15**: 213–18. doi: [10.1111/j.1600-0501.2004.00969.x](https://doi.org/10.1111/j.1600-0501.2004.00969.x)
17. Klintström E, Smedby Ö, Moreno R, Brismar TB. Trabecular bone structure parameters from 3D image processing of clinical multi-slice and cone-beam computed tomography data. *Skeletal Radiol* 2014; **43**: 197–204. doi: [10.1007/s00256-013-1766-5](https://doi.org/10.1007/s00256-013-1766-5)
18. Van Dessel J, Huang Y, Depypere M, Rubira-Bullen I, Maes F, Jacobs R. A comparative evaluation of cone beam CT and micro-CT on trabecular bone structures in the human mandible. *Dentomaxillofac Radiol* 2013; **42**: 20130145. doi: [10.1259/dmfr.20130145](https://doi.org/10.1259/dmfr.20130145)
19. Ibrahim N, Parsa A, Hassan B, van der Stelt P, Aartman IH, Wismeijer D. Accuracy of trabecular bone microstructural measurement at planned dental implant sites using cone-beam CT datasets. *Clin Oral Implants Res* 2014; **25**: 941–5. doi: [10.1111/clr.12163](https://doi.org/10.1111/clr.12163)
20. Ho JT, Wu J, Huang HL, Chen MY, Fuh LJ, Hsu JT. Trabecular bone structural parameters evaluated using dental cone-beam computed tomography: cellular synthetic bones. *Biomed Eng Online* 2013; **12**: 115. doi: [10.1186/1475-925X-12-115](https://doi.org/10.1186/1475-925X-12-115)
21. Thomsen JS, Laib A, Koller B, Prohaska S, Mosekilde L, Gowin W. Stereological measures of trabecular bone structure: comparison of 3D micro computed tomography with 2D histological sections in human proximal tibial bone biopsies. *J Microsc* 2005; **218**: 171–9. doi: [10.1111/j.1365-2818.2005.01469.x](https://doi.org/10.1111/j.1365-2818.2005.01469.x)
22. Revol-Muller C, Peyrin F, Carrillon Y, Odet C. Automated 3D region growing algorithm based on an assessment function. *Pattern Recogn Lett* 2002; **23**: 137–50. doi: [10.1016/S0167-8655\(01\)00116-7](https://doi.org/10.1016/S0167-8655(01)00116-7)
23. Hohlweg-Majert B, Metzger MC, Kummer T, Schulze D. Morphometric analysis—cone beam computed tomography to predict bone quality and quantity. *J Craniomaxillofac Surg* 2011; **39**: 330–4. doi: [10.1016/j.jcms.2010.10.002](https://doi.org/10.1016/j.jcms.2010.10.002)
24. Gonzalez-Garcia R, Monje F. The reliability of cone-beam computed tomography to assess bone density at dental implant recipient sites: a histomorphometric analysis by micro-CT. *Clin Oral Implants Res* 2013; **24**: 871–9. doi: [10.1111/j.1600-0501.2011.02390.x](https://doi.org/10.1111/j.1600-0501.2011.02390.x)
25. Naitoh M, Aimiya H, Hirukawa A, Ariji E. Morphometric analysis of mandibular trabecular bone using cone beam computed tomography: an *in vitro* study. *Int J Oral Maxillofac Implants* 2010; **25**: 1093–8.
26. Hsu JT, Wang SP, Huang HL, Chen YJ, Wu J, Tsai MT. The assessment of trabecular bone parameters and cortical bone strength: a comparison of micro-CT and dental cone-beam CT. *J Biomech* 2013; **46**: 2611–18. doi: [10.1016/j.jbiomech.2013.08.004](https://doi.org/10.1016/j.jbiomech.2013.08.004)
27. Jonasson G, Sundh V, Hakeberg M, Hassani-Nejad A, Lissner L, Ahlqvist M. Mandibular bone changes in 24 years and skeletal fracture prediction. *Clin Oral Investig* 2013; **17**: 565–72. doi: [10.1007/s00784-012-0745-x](https://doi.org/10.1007/s00784-012-0745-x)
28. Nägele E, Kuhn V, Vogt H, Link TM, Müller R, Lochmüller EM, et al. Technical considerations for microstructural analysis of human trabecular bone from specimens excised from various skeletal sites. *Calcif Tissue Int* 2004; **75**: 15–22. doi: [10.1007/s00223-004-0151-8](https://doi.org/10.1007/s00223-004-0151-8)
29. Bauer JS, Link TM. Advances in osteoporosis imaging. *Eur J Radiol* 2009; **71**: 440–9. doi: [10.1016/j.ejrad.2008.04.064](https://doi.org/10.1016/j.ejrad.2008.04.064)
30. Hildebrand T, Rüeggsegger P. Quantification of bone micro-architecture with the structure model index. *Comput Methods Biomech Biomed Eng* 1997; **1**: 15–23. doi: [10.1080/01495739708936692](https://doi.org/10.1080/01495739708936692)
31. Moreno R, Borgia M, Smedby Ö. Techniques for computing fabric tensors: a review. Visualization and processing of tensors and higher order descriptors for multi-valued data. Berlin/Heidelberg, Germany: Springer; 2014. pp. 271–92. doi: [10.1007/978-3-642-54301-2_12](https://doi.org/10.1007/978-3-642-54301-2_12)
32. Gundersen HJ, Boyce RW, Nyengaard JR, Odgaard A. The Connector: unbiased estimation of connectivity using physical disectors under projection. *Bone* 1993; **14**: 217–22. doi: [10.1016/8756-3282\(93\)90144-Y](https://doi.org/10.1016/8756-3282(93)90144-Y)