Centrifugal multiplexing fixed-volume dispenser on a plastic lab-on-a-disk for parallel biochemical single-end-point assays

Moonwoo La, Sang Min Park, and Dong Sung Kim^{a)}

Department of Mechanical Engineering, Pohang University of Science and Technology (POSTECH), Pohang 790-784, South Korea

(Received 4 November 2014; accepted 5 January 2015; published online 13 January 2015)

In this study, a multiple sample dispenser for precisely metered fixed volumes was successfully designed, fabricated, and fully characterized on a plastic centrifugal lab-on-a-disk (LOD) for parallel biochemical single-end-point assays. The dispenser, namely, a centrifugal multiplexing fixed-volume dispenser (C-MUFID) was designed with microfluidic structures based on the theoretical modeling about a centrifugal circumferential filling flow. The designed LODs were fabricated with a polystyrene substrate through micromachining and they were thermally bonded with a flat substrate. Furthermore, six parallel metering and dispensing assays were conducted at the same fixed-volume ($1.27\ \mu\text{l}$) with a relative variation of $\pm 0.02\ \mu\text{l}$. Moreover, the samples were metered and dispensed at different sub-volumes. To visualize the metering and dispensing performances, the C-MUFID was integrated with a serpentine micromixer during parallel centrifugal mixing tests. Parallel biochemical single-end-point assays were successfully conducted on the developed LOD using a standard serum with albumin, glucose, and total protein reagents. The developed LOD could be widely applied to various biochemical single-end-point assays which require different volume ratios of the sample and reagent by controlling the design of the C-MUFID. The proposed LOD is feasible for point-of-care diagnostics because of its mass-producible structures, reliable metering/dispensing performance, and parallel biochemical single-end-point assays, which can identify numerous biochemical. © 2015 AIP Publishing LLC.

[<http://dx.doi.org/10.1063/1.4905940>]

I. INTRODUCTION

Lab-on-a-disk (LOD), a centrifugal force-based microfluidic platform, which has attracted attention in the field of clinical diagnostics because it can be easily miniaturized and automated for biomedical analysis.^{1,2} The centrifugal force in the LOD is the primary driving force that induces the various centrifugal microfluidic functions such as separating, valving, metering, dispensing, mixing, and reacting.^{3–7} An integrated LOD system combines these functions with analytical measurement methods and medical examination techniques. It facilitates high-throughput screening and parallel assays for biomedical diagnostics.

Centrifugal liquid metering and dispensing are essential procedures in conducting parallel assays on LOD. Liquid flow must be accurately controlled through effective centrifugal microfluidic functions to produce a reliable and valid biomedical diagnosis on the LOD. Accordingly, Honda *et al.* carried out simultaneous multiple immunoassays on a LOD that consisted of a complex microchannel network with hydrophobic path-type valves.^{3,8} DNA analysis based on primary amplification and secondary real-time polymerase chain reaction (PCR) was evaluated on several types of LODs.^{9–11} Moreover, simultaneous analysis of biochemistry and immunoassay from whole blood was achieved on an integrated LOD based on laser irradiated ferrowax

^{a)} Author to whom correspondence should be addressed. Electronic mail: smkds@postech.ac.kr

microvalves (LIFM).^{4,12} An enzyme-linked immunosorbent assay (ELISA) was also conducted on a LOD with passive burst valves and magnetic beads.¹³

When developing innovative LODs for the parallel assays, several critical issues are raised as follows: (1) small volumes: a small volume of liquid (i.e., in the order of $1\ \mu\text{l}$) should be controlled precisely. Most biochemical assay protocols indicate that the volume ratios of the sample to the reagents or buffer solutions should be one-tenth or less. When we apply 1–2 drops of reagent (approximately in the order of $10\ \mu\text{l}$) on the miniaturized LOD, the corresponding sample volume inevitably shifts to the order of $1\ \mu\text{l}$. Inganäs *et al.* developed a microfluidic compact disk (CD) device (Gyrolab BioaffyTM) for precise metering of sample volume with $0.2\ \mu\text{l}$. In their study, liquid movement and localization were conducted based on the capillary filling, centrifugal metering, and hydrophobic obstructing, which were realized by local surface modification within radially complex microstructure.³ (2) Precise volume control: both fixed and small volume controls are important because small volumetric error in the sample can limit the accuracy of the biochemical assays. (3) Multiple liquid handling: to diagnose diseases and define health condition, diverse information on the physiological substances is acquired using parallel assays. The primary unit operation in the parallel assays involves the multiple metering and dispensing of the samples. Specifically, parallel assays can be simply and cost-effectively automated through *in situ* centrifugal microfluidic metering and dispensing without external devices such as through the injection of nozzles and pipettes. In line with this information, Mark *et al.* developed sample aliquoting structure based on the centrifuge-pneumatic valve.⁵ In their study, samples can be split into several sub-volumes through a radially sloped feed channel with a series of branching metering chambers. (4) Flexible interconnection with further microfluidic structures: several microfluidic functions should be integrated to miniaturize a biomedical diagnosis on a LOD. For example, a biochemical assay required complicated interconnection of various functional units including sample preparation, sample/reagent reaction, and result investigation.¹ (5) Design guide for microfluidic functions: in the LOD design stage, an appropriate theoretical model that simulates and predicts centrifugal microfluidic functions should be developed. One can then estimate effects of parameters, such as structural dimensions, driving forces, and fluid properties of the sample, on these functions to optimize the design of a LOD under appropriate operating conditions. Such theoretical model also enhances the flexibility of the design in response to potential uncertainty and internal or external changes. (6) Low-cost LOD fabrication: LOD disposability is an indispensable characteristic in insuring safe examination and in preventing contamination. Thus, LODs should be mass-produced through molding processes such as hot embossing and injection molding. Plastics are the most appropriate material for LOD fabrication given their high reproducibility and cost-effectiveness.^{14,15} Productivity is also enhanced by minimizing the need for additional processes, including multilayer assembly and the injection of additional substances.

In this regard, we propose a new multiple sample dispenser for precisely metering and dispensing fixed volumes, namely, a *centrifugal multiplexing fixed-volume dispenser (C-MUFID)*, on a plastic LOD for parallel biochemical single-end-point assays. The theoretical model for the centrifugal circumferential filling flow is established by experimental validation prior to C-MUFID design. The model provides appropriate design guide according to experimental setup (e.g., space, rotational speed, and duration). A LOD containing the C-MUFID is then designed on a single layer and fabricated via computer numerical control (CNC) micromachining on a polystyrene (PS) substrate. The C-MUFID facilitates multiple samples metering and dispensing at $1\ \mu\text{l}$ -volume with $\pm 0.02\ \mu\text{l}$ -relative variation based on centrifugal liquid filling followed by centrifugal doctor-blading. As the C-MUFID has circumferential design with passive microvalves and ventilation chambers, it enables flexible interconnection with micromixers. Finally, parallel biochemical single-end-point assays are conducted on the C-MUFID combination with a centrifugal serpentine micromixer (CSM).¹⁷

II. DESIGN AND CONCEPT

A. Design of LOD

Fig. 1(a) shows a first type of LOD design having C-MUFID, namely, a dispensing LOD which was designed to evaluate the liquid metering and dispensing performances of the LOD.

FIG. 1. Dispensing LOD containing C-MUFID for sample metering and dispensing: (a) image of the fabricated PS LOD and (b) schematic of the structure. Integrated LOD with C-MUFID and CSM for parallel biochemical single-end-point assays: (c) image of the fabricated PS LOD and (b) schematic of structure. Each structure is labelled as follows: (1) inlet hole, (2) inlet chamber, (3) feeding channel, (4) metering chamber, (5) waste chamber, (6) vent hole, (7) passive valve, (8) dispensing chamber, (9) reagent inlet chamber, (10) serpentine micromixer, and (11) reaction chamber.

The dispensing LOD consists of two C-MUFIDs: one was used to dispensing six liquid samples with the same volume ($\sim 1.2 \mu\text{l}$), whereas the other was utilized to dispense six samples with different volumes ($4.0, 3.0, 2.3, 1.6, 1.1$, and $0.6 \mu\text{l}$). Each C-MUFID was composed of the following elements: an inlet chamber (volume of $120 \mu\text{l}$), a circumferential feeding channel at a radial distance of 16.5 mm from the center (width of $700 \mu\text{m}$ and height of 1 mm), six metering chambers to store the desired dispensing volumes, six passive valves (width of $200 \mu\text{m}$, height of $200 \mu\text{m}$, and length of $500 \mu\text{m}$), six dispensing chambers (width of 1 mm , height of 1 mm , and length of 5.5 mm), and a waste chamber (volume of $150 \mu\text{l}$) as indicated in Fig. 1(b). Metering chambers were arranged in the circumferential direction and connected with the circumferential feeding channel in radial direction. Each passive valve between the metering chamber and the dispensing chamber was designed to be burst when the rotational speed of the LOD exceeded 1500 rpm .¹⁶ Fig. 1(c) describes a second type of LOD designs, namely, an integrated LOD that was designed for the parallel biochemical single-end-point assays based on sample dispensing through the C-MUFID followed by the mixing of the samples with reagents. The integrated LOD consisted of a C-MUFID containing reagent inlet chambers (volume of $27 \mu\text{l}$) instead of the dispensing chambers, along with CSMs¹⁷ and reaction chambers (volume of $25 \mu\text{l}$) (Fig. 1(d)). Parallel biochemical single-end-point assays generally achieved by starting with the multiple precise metering and dispensing of sample and following single reading of the end result after a fixed reaction period.^{12,18} In this sense, the integrated LOD, consisted of the C-MUFID providing multiple fixed-volumes metering and dispensing of sample and the CSM helping reaction of sample and reagent, could be suitable to conduct the parallel biochemical single-end-point assays.

B. Conceptual operations

Fig. 2(a) depicts the conceptual procedures for the multiple sample metering and dispensing process that uses the C-MUFID on the dispensing LOD. A series of processes was controlled by regulating the rotational speed of this dispensing LOD. At the stationary state, $120 \mu\text{l}$ of the liquid sample was pipetted into the inlet chamber (Fig. 2(a-i)). The liquid sample was then

FIG. 2. Conceptual procedures for (a) the multiple sample metering and dispensing through C-MUFID and (b) parallel biochemical single-end-point assays on the integrated LOD. Each step is indicated as follows: (i) initial sample loading, (ii) sample feeding, (iii) sample metering, (iv) sample dispensing, (v) reagent loading, (vi) centrifugal mixing of the sample and the reagent, and (vii) reaction of the mixture.

delivered to the circumferential feeding channel and to the connecting metering chambers at 500 rpm (Fig. 2(a-ii)). The capillary force caused the passive valve to hold the flow front at the end of each metering chamber. When the rotational speed was increased to 800 rpm, the sample in the feeding channel passed into the waste chamber, whereas the sub-volumes of the sample in the metering chambers remained stationary because of structural restriction (i.e., the “doctor-blading” process) in the sample metering stage. During the sample feeding and metering stages, the flow front at the end of each metering chamber was held constant by the passive valve (Fig. 2(a-iii)). Each metered sample could be dispensed and delivered to each connected chamber by bursting the passive valve at 1,500 rpm (Fig. 2(a-iv)).

Fig. 2(b) illustrates the conceptual procedures for the parallel biochemical single-end-point assays. The sample feeding, metering, and dispensing processes were identical to those of the dispensing LOD (Figs. 2(b-i)–2(b-iv)). After sample dispensing (Fig. 2(b-iv)), 25 μ l of the reagents were loaded at the stationary state by pipetting (PIPETMAN P20, Gilson, Inc.) into each reagent inlet chamber, the top of which had half-opened surfaces (Fig. 2(b-v)). Metering the reagent at 25 μ l volume using the C-MUFID required additional area more than 25 mm² on the restricted area of the LOD. This condition indicated that the number of parallel assays should be reduced to enlarge the area per unit assay. Thus, we loaded externally metered reagent into LOD to conduct as many assays as possible instead of using the C-MUFID. Furthermore, one could employ a laboratory automation device to load the reagent. You *et al.* achieved direct loading of the reagent into the LOD with 0.4 μ l volumetric resolution by developing a laboratory automation device based on a diaphragm pump.¹⁹ Combination of the laboratory automation device and the proposed LOD is expected to carry out automatic point-of-care testing based on rapid processing of large numbers of parallel assays. The dispensed sample and the loaded reagent were mixed effectively when they passed through the CSM as the integrated LOD was rotated again at 2,000 rpm (Fig. 2(b-vi)). At this moment, the half-closing at the top of the chamber easily prevented the overflow of the liquids from the reagent inlet chamber as induced by centrifugal force (see supplementary Fig. S1, Ref. 28). Finally, the mixture promoted the reaction (i.e., color change) of the sample with the reagent (Fig. 2(b-vii)). Consequently, we can determine the effects of specific biochemical substances through a colorimetric analysis of the mixtures. Furthermore, we can use the integrated LOD to investigate several biochemical substances, such as albumin (ALB), cholesterol, total protein

(TP), uric acid, triglyceride, inorganic phosphorus, and glucose (GLU), through the parallel single end-point assays. In this study, we conducted parallel biochemical single-end-point assays on the integrated LOD with ALB, GLU, and TP as the objective substances.

III. MATERIALS AND METHODS

A. LOD fabrication

The dispensed and integrated LODs were fabricated on PS (GP0000, Kumho Petrochemical) substrates as presented in Figs. 1(a) and 1(c), respectively. The chamber and channel structures were engraved precisely on a disk-shaped PS substrate using a CNC micro-machine (EGX360, Roland DGA Corp.). Subsequently, the structured substrate was covered with a flat PS substrate via collapse-free thermal bonding technique.²⁰ To verify moldability of the LOD designs, injection molding using metal mold has been conducted to fabricate the preliminary LOD containing the C-MUFID and the CSM on a trial basis. Consequently, injection molding of the preliminary LOD was successfully achieved as described in supplementary Fig. S2 (Ref. 28). The LODs in the present study possessed moldable structures on a single-layer substrate; hence, they were expected to be mass-producible through the use of plastic molding technologies such as hot embossing and injection molding.¹⁵

B. Sample and reagent preparation

Deionized (DI) water that was dyed purple, denoted by purple-dyed water, was used for experiment to verify the centrifugal fixed-volume metering and dispensing performance of the dispensing LOD. Yellow-dyed water and Blue-dyed water were prepared to be used as the sample and the reagent, respectively, in the parallel centrifugal mixing tests. Finally, parallel biochemical single-end-point assays were conducted using standard serum (HiSens BioTrol level-1; HBi Co. Ltd.) as a sample and three types of biochemical reagents (HiSens ALB/CHO/TP reagent; HBi Co., Ltd.) in room temperature.

C. Operation and visualization system

Experimental setup including the operation and visualization components is schematically described in supplementary Fig. S3 (Ref. 28). The operation part was composed of a motor (SGMJV-01AD21; Yaskawa Electric., Co.) and a control box (SGDV-R90A01A; Yaskawa Electric., Co.) to rotate the LODs at the prescribed rotational speed. The visualization part consisted of a stroboscope (DX-525S; Serom Electronics Lab.), which is a lighting device that generates periodic flashes, and a charge-coupled device (CCD) camera (VH-310C; Vieworks, Co., Ltd) that is used to observe the behavior of the centrifugal fluid in the microfluidic channels of the LOD. The periodic flashes were synchronized with the rotational speed (in rpm) of the LOD and regulated by transmitting the trigger signal to the stroboscope in each LOD revolution. Consequently, one can continuously monitor the objective position of the LOD through the CCD camera in 264 fps frame rate with 640×480 pixels ($7.4 \mu\text{m}$ per unit pixel). The captured images or recoded videos were used to either measure sample volume or to identify colorimetric changes in the mixture of the sample with the reagent using an own-produced image processing program. Supplementary Fig. S4 indicates the measurement procedures for the volumes of the metered or dispensed samples in an image processing program (Ref. 28). First, a target image that presents the result of metering or dispensing was derived from the obtained video. The area occupied by the sample was then specified by analyzing the distribution of color intensity. Finally, the sample volume was calculated based on the number of pixels within the specified area. A single pixel on the obtained image corresponded to $6 \times 10^{-4} \mu\text{l}$ in the image processing program.

IV. RESULTS AND DISCUSSION

A. Theoretical model for centrifugal circumferential filling flow

Theoretical modeling and experimental verification about a centrifugal circumferential filling flow were carried out to provide a simple design guideline for the C-MUFID geometry determination. We defined a model problem to represent the centrifugal circumferential filling flow that originates from the inlet chamber and passes through the circumferential feeding channel on the LOD. Fig. 3(a) shows the schematic of the centrifugal circumferential filling flow in a conceptual LOD design. The dimensional variables and constants of the inlet chamber and of the circumferential channel were represented in the cylindrical coordinate system. A sample fluid (e.g., water or serum with the physical fluid properties of density ρ and viscosity μ) was loaded and filled the inlet chamber completely at the initial state (at $t = 0$). Fluid flow was then initiated as the LOD began to rotate at a constant speed of ω . The fluid flowed out of the inlet chamber and ran through the feeding channel because of the hydraulic pressure generated by the centrifugal force, which was exerted on the fluid in the inlet chamber. The flow front gradually advanced along the circumferential channel path of the LOD. We aimed to examine this advancement over time, i.e., as a function of time t , denoted by $L(t)$. To deliver the sample fluid to a specific location on the LOD within the desired duration, the dimensions and the position of the inlet chamber and the circumferential channel were set in consideration of the rotational speed of the LOD. However, obtaining a full and exact 3D solution to the transient centrifugal circumferential filling flow is generally difficult. In this regard, we attempted to generate a simple analytical solution for $L(t)$ as a function of the design parameters, including fluid properties, chamber and channel geometries, and the processing condition under several assumptions. These assumptions are as follows: (1) we assumed that the fluid was an incompressible Newtonian fluid with a simple constitutive equation; (2) we assumed that the centrifugal force affected only the fluid in the inlet chamber and not that in the circumferential channel; (3) we

FIG. 3. (a) Schematic of the LOD for the experiment on centrifugal circumferential filling flow with (b) geometrical parameters. (c) Experimental images of flow front advancement (clockwise direction) in terms of time at a rotational speed of 600 rpm. (d) Graphs of flow front advancement in terms of time according to the rotational speed of the LOD. Each solid line represents the theoretical profile at the corresponding rotational speed based on Eq. (5).

disregarded the Coriolis force to ignore the transversal secondary behaviors of the fluid flow and to simplify the model; (4) we approximated the centrifugal circumferential filling flow at a specific time as a pseudo-static flow to apply the fully developed flow rate within the rectangular channel; and (5) we neglected the effect of surface tension on the flow front because we considered a PS substrate whose water contact angle is almost 90° .^{21,22}

The mass conservation of the entire control volume can be determined as follows:

$$\alpha r_c(t)H \frac{dr_c(t)}{dt} = WH \frac{dL(t)}{dt}, \text{ with initial condition of } r_c(t=0) = r_0, \quad (1)$$

where $r_c(t)$ is the distance from the center of the disk to the upper free surface of the fluid in the inlet chamber, which is a function of time t ; r_0 is the distance from the center to the top surface of the inlet chamber; α is the circumferential angle of the inlet chamber; W is the width of the circumferential channel; and H denotes the heights of the chamber and the channel.

The pressure difference developed in the inlet chamber can be represented using Eq. (1)

$$\Delta P = \frac{1}{2} \rho \omega^2 (r_d^2 - r_c^2(t)) = \frac{1}{2} \rho \omega^2 \left[r_d^2 - r_0^2 - \frac{2W}{\alpha} L(t) \right], \quad (2)$$

where r_d is the distance from the center to the channel.

The fully developed flow rate in the rectangular channel was approximated as follows to represent the flow rate of the pseudo-static filling flow in this study:²³

$$Q = \frac{H^3 W}{12 \mu L} \Delta P F_p, \quad (3)$$

where F_p represents a shape factor that considers the aspect ratio of the rectangular channel, which is defined as follows:

$$F_p = 1 - \frac{192H}{\pi^5 W} \sum_{p=1,3,5,\dots} \frac{1}{p^5} \tanh\left(\frac{p\pi W}{2H}\right). \quad (4)$$

As depicted in the Appendix, we can determine the non-dimensionalized solutions to the preceding Eqs. (1)–(4)

$$\tilde{L}(\tilde{t}) = \begin{cases} \sqrt{2\tilde{t}} & \text{for } \tilde{t} \ll 1 \\ 1 - A \exp(-\tilde{t}) & \text{for } \tilde{t} \gg 1 \end{cases}, \quad (5)$$

where $\tilde{L}$ and $\tilde{t}$ are dimensionless filling length and time, respectively, and A is an undetermined constant.

One can derive a guide for the design of circumferential channel length from these solutions. This guide corresponds to rotational speed by applying fluid properties, geometrical dimensions, and operating conditions (Figs. 3(a) and 3(b), see Appendix). Finally, we introduced a semi-circular path for the circumferential channel into the design of the C-MUFID by considering the reasonable total filling time because the time can be lengthened significantly by the exponential saturation behavior given an increased channel length.

B. Centrifugal metering and dispensing performance of C-MUFID

Experiments on centrifugal fixed-volume metering and dispensing were conducted on the dispensing LOD using purple-dyed water. Movie 1 presents the experimental results of sample feeding, metering, and dispensing at rotation speeds of 500, 800, and 1,500 rpm, respectively [see Fig. 4 (Multimedia view)]. In the feeding stage, the circumferential filling time was approximately 3 min, which indicated a greater delay than theoretical period derived from the model for

FIG. 4. Results of sample metering and dispensing under (a) similar and (b) different sub-volumes. Images of (i) metered and (ii) dispensed samples; and (iii) quantitatively measured data derived from image processing. (Multimedia view) [URL: <http://dx.doi.org/10.1063/1.4905940.1>]

centrifugal circumferential filling flow. As the flow front approached the edge, its advancement was slowed down by the pressured barrier induced by capillary force during the sudden expansion in the width of the feeding channel that is connected to the metering chamber.^{24–26} The sample in the feeding channel approached the waste chamber. By contrast, the samples in each metering chamber were isolated because of structural restrictions. The multiple sub-volumes from the first loaded sample were metered at the same volume in the corresponding metering chambers based on the “doctor-blading” process. The metered samples were then dispensed immediately to the dispensing chambers by increasing the rotational speed to 1500 rpm. The overall processes were automatically performed on the dispensing LOD according to the assigned rotational speed, with the exception of the sample loading process by pipetting.

A quantitative analysis was then conducted to evaluate the metering and dispensing performance of the dispensing LOD based on image processing. Fig. 4(a) displays the results of sample metering and dispensing given the same sub-volumes. Based on these findings, we can measure the volumes of metered and dispensed samples at all of the indicated positions in the dispensing chambers, as described in Fig. 4(a-iii). The average volume of the metered and dispensed samples at all positions was approximately $1.27 \mu\text{l}$, and the relative difference by position was $\pm 0.02 \mu\text{l}$. Fig. 4(b) exhibits results of sample metering and dispensing given different sub-volumes. The volumetric capacities of chamber Nos. 1–6 were designed and fabricated for metering and dispensing at volumes of 4.0, 3.0, 2.3, 1.6, 1.1, and $0.6 \mu\text{l}$. Fig. 4(b-iii) describes the overall quantitative results, and the difference in volume between each metered and dispensed sample and the target volume was within $\pm 0.02 \mu\text{l}$. Moreover, the results indicated that the dispensing LOD facilitated precise sample metering at a $1 \mu\text{l}$ -scale and a fixed-volume. The metered sample transitioned to the next step effectively in the process. In addition, we can control the final dispensed volume of the sample precisely by varying the size of the metering chamber.

C. Centrifugal mixing performance

Prior to the parallel biochemical single-end-point assays, parallel centrifugal mixing tests were conducted to determine quantitative mixing performance through the combination of the

C-MUFID with the CSM on the integrated LOD. The processes for the feeding, metering, and dispensing of the sample were similar when investigated using the dispensing LOD [movie 2; see Fig. 5 (Multimedia view)]. After sample dispensing, the reagent was loaded into the reagent inlet chambers at intervals of $25\ \mu\text{l}$ by pipetting at the stationary state of the integrated LOD. The two liquids were then mixed using the CSM at a rotation speed of 2000 rpm. The mixing performance was quantitatively analyzed by measuring the colorimetric distributions of the final mixtures. Fig. 5 shows the results of the parallel centrifugal mixing tests on the integrated LOD. The “A-half” of the integrated LOD was subject to six similar mixing tests with $1.6\ \mu\text{l}$ of the sample and $25\ \mu\text{l}$ of the reagent (Fig. 5(a-ii)). By contrast, the “B-half” of the integrated LOD was subject to six different mixing tests (Fig. 5(a-iii)). On this side, $4.0, 3.0, 2.3, 1.6, 1.1,$ and $0.6\ \mu\text{l}$ of the samples were dispensed in chamber Nos. 1–6. The reagent was loaded into the reagent inlet chambers at intervals of $25\ \mu\text{l}$. Thus, a total of 12 parallel centrifugal mixing tests were conducted on the integrated LOD simultaneously.

The color grades of the final mixtures were quantified through colorimetric analysis using the resultant images at the end of the mixing tests. The color of each pixel in the obtained mixture image can be described according to the additive primary colors, i.e., red (R), green (G), and blue (B), each with a value of 0–255. Furthermore, the normalized intensities of the representative colors were determined from the R , G , and B data in of all the pixels. Once the brightness in the obtained images was compensated with a constant R -index value, we measured the average values of the G - and B - indices to determine the quantitative grades of the final mixtures. The measured average values were normalized based on the maximum value (i.e., 255). Figs. 5(b-i) and 5(b-ii) depict the results of the colorimetric analysis of the mixtures generated in the centrifugal mixing tests given both similar or different volume ratios of the sample to the reagent. The normalized intensities of the G - and B - indices were constant regardless of the test number as per the results of the mixing test under a uniform volume ratio. This finding indicated that each mixture displayed physically similar color levels. Thus, the integrated LOD

FIG. 5. (a) Parallel centrifugal mixing tests on the integrated LOD. (i) Overall image: mixtures from the centrifugal mixing at (ii) the same volume ratio and (iii) different volume ratios of the sample to the reagent. (b) Result graphs from the colorimetric analysis of the mixtures derived from the parallel centrifugal mixing tests at (i) the same volume ratio and (ii) different volume ratios of the sample to the reagent. (Multimedia view) [URL: <http://dx.doi.org/10.1063/1.4905940.2>]

FIG. 6. Results of the parallel biochemical single-end-point assays: (a) overall assays on the integrated LOD and (b) magnified view after the reaction of the reagent itself or of the reagent with the serum.

reported a reliable parallel centrifugal mixing performance according to the consistent results of the centrifugal mixing tests given a constant volume ratio of the sample and reagent. Moreover, the results of mixing with different volume ratios indicate that the dispensed volumes of yellow-dyed water decreased as they flowed from chambers no. 1 to 6; in response, the normalized intensity of the *B*-index increased gradually given a similar *G*-index value. Therefore, one may conduct the various biochemical single-end-point assays that require the volume ratios of the sample and the reagent to vary. These ratios can be regulated by precisely controlling the volume of the sample metering chamber of the C-MUFID. In addition, the parallel assays can feasibly identify numerous biochemicals at the same time by introducing serum or plasma and biochemical reagents into the integrated LOD.

D. Parallel biochemical single-end-point assays

Parallel biochemical single-end-point assays were conducted using standard serum as a sample. These assays followed the procedures in the centrifugal mixing tests, and three types of biochemical reagents were used for the ALB, GLU, and TP assays. Each reagent (with a volume of $25\ \mu\text{l}$) was loaded into the reagent inlet chambers after dispensing $1.6\ \mu\text{l}$ of the serum on the A-half side (Fig. 6(a)). As a control, only reagents without dispensed serum were loaded into the reagent inlet chambers on the B-half side (Fig. 6(b)). Following the overall operation of the integrated LOD, the color of each reagent (ALB, yellow; GLU, light orange; and TP, light blue) changed to green, pink, and purple on the A-half side as a result of the reactions of the reagent with the serum (Figs. 6(b) and 6(c)). Therefore, 12 parallel biochemical single-end-point assays were successfully conducted on the integrated LOD. Furthermore, the existence of ALB, GLU, and TP in the serum was confirmed by the color changes in the corresponding reagents. Nonetheless, the levels of ALB, CHO, and TP should be quantitatively calibrated depending on the serum and reagents in future studies. Thus, the developed plastic LOD that is composed of C-MUFID and CSM is suitable for point-of-care testing based on parallel biochemical single-end-point assays.

V. CONCLUSIONS

A C-MUFID for the multiple dispensing of precisely metered sample volumes was successfully designed, fabricated, and fully characterized on a plastic centrifugal LOD, which was applied to parallel biochemical single-end-point assays in combination with CSM. Moreover, a design guideline for the circumferential channel on the LOD, which defines the feeding channel of the C-MUFID, was successfully established based on the theoretical modeling and experimental validation of a centrifugal circumferential filling flow. In addition, two different LODs

(i.e., dispensing and integrated LODs) were fabricated on PS substrates via CNC micromachining, followed by thermal bonding with a flat substrate under a collapse-free condition. Extensive experiments on the centrifugal fixed-volume metering and dispensing of samples were conducted, and the findings indicated that the C-MUFID facilitated the reliable metering and dispensing of samples whose volumes were at the $1\ \mu\text{l}$ -scale with the relative difference of $\pm 0.02\ \mu\text{l}$. Furthermore, the integrated LOD with C-MUFID and CSM enabled metering and dispensing at the prescribed sample volume. It also mixed the dispensed sample and the loaded reagent reliably. Parallel biochemical single-end-point assays were also conducted as a proof-of-concept study using standard serum and biochemical reagents to determine ALB, GLU, and TP. Given its mass-producible structures and its reliable metering and dispensing performance, the proposed LOD can be widely applied to various biochemical assays that require precise volume metering and reliable mixing.

ACKNOWLEDGMENTS

This work was supported by the National Research Foundation of Korea (NRF) grant funded by the Korea government (MSIP) (Grant Nos. 2011-0030075 and 2014R1A2A1A01006527).

APPENDIX: ANALYTICAL SOLUTION TO CENTRIFUGAL CIRCUMFERENTIAL FLOW

The governing equation for $L(t)$ and the corresponding conditions are provided in Eqs. (1)–(4). The substitution of Eqs. (1) and (2) into Eq. (3) yields

$$\frac{dL(t)}{dt} = \frac{\rho\omega^2 H^2}{24\mu} F_p \left(\frac{r_d^2 - r_0^2}{L(t)} - \frac{2W}{\alpha} \right) \text{ with } L(0) = 0, \quad (\text{A1})$$

which describes the change rate in the flow front advancement of the centrifugal circumferential filling flow. We introduce characteristic length L_c , time t_c , and the corresponding dimensionless variables as below

$$L_c = \frac{\alpha(r_d^2 - r_0^2)}{2W}, \quad (\text{A2})$$

$$t_c = \frac{12\mu\alpha L_c}{\rho\omega^2 H^2 W F_p}, \quad (\text{A3})$$

with

$$\tilde{L}(t) = \frac{L(t)}{L_c} \text{ and } \tilde{t} = \frac{t}{t_c}, \quad (\text{A4})$$

where $\tilde{L}$ and $\tilde{t}$ are dimensionless filling length and time, respectively. The non-dimensionalization of Eq. (A1) with Eqs. (A2)–(A4) gives a dimensionless first-order differential equation that governs flow front advancement. This equation is presented as follows:

$$\frac{d\tilde{L}(\tilde{t})}{d\tilde{t}} = \frac{1}{\tilde{L}(\tilde{t})} - 1 \text{ with } \tilde{L}(0) = 0. \quad (\text{A5})$$

Equation (A5) is difficult to solve analytically. Nonetheless, we can approximate the behavior of the filling process in both small and large time regions. Moreover, the values of $\tilde{L}$ is definitely less than one at the initial stage of the filling process (i.e., at the short times, $\tilde{t} \ll 1$). Therefore, we can approximate Eq. (A5) as $d\tilde{L}/d\tilde{t} \approx 1/\tilde{L}$, and the solution becomes

$$\tilde{L}(\tilde{t}) = \sqrt{2\tilde{t}}, \text{ for } \tilde{t} \ll 1. \quad (\text{A6})$$

$\tilde{L} \rightarrow 1$ under asymptotic filling behavior (i.e., at large times, $\tilde{t} \rightarrow \infty$). Therefore, the asymptotic solution of Eq. (A3) is expressed as follows:²⁷

$$\tilde{L}(\tilde{t}) = 1 - A \exp(-\tilde{t}), \text{ for } \tilde{t} \gg 1, \quad (\text{A7})$$

where A is an undetermined constant. Based on the solutions of Eqs. (A4) and (A5), the flow front advances with the square root of the time at the initial stage of the filling process until the characteristic time ($t \ll t_c$). Nonetheless, the filling behavior changes to approach the equilibrium length L_c asymptotically given the exponential saturation at $t \gg t_c$. To accelerate assay times, the circumferential channel length should be designed such that it is filled with the sample within the characteristic time.

This study presents a typical example of how the analytical results can be applied to a real design problem. We suppose the following example in which L_c and t_c are determined according to the given condition:

- Fluid properties: $\rho = 10^3 \text{ kg m}^{-3}$ and $\mu = 10^{-3} \text{ Pa s}$ (typical properties of water)
- Chamber and channel geometry: $W = 0.7 \text{ mm}$, $H = 1 \text{ mm}$, $r_d = 16.5 \text{ mm}$, $r_0 = 10 \text{ mm}$, and $\alpha = 100^\circ$.
- Rotational speeds of the LOD: $\omega = 500 \sim 1,200 \text{ rpm}$.

The centrifugal circumferential filling flow was experimentally investigated at different LOD rotational speeds to verify the theoretical approximate solution for flow front advancement (Fig. 3(c)). The experimental data on the flow front advancement at the given initial period match well with the corresponding theoretical profiles based on the square root function, as shown in Fig. 3(d). This result indicates that Eq. (A4) interprets centrifugal circumferential filling flow validly. When the rotational speed of the LOD was 500 rpm, $L_c = 21.5 \text{ cm}$. This value was both greater and less than that of the circular path for a single revolution at $r_d = 16.5 \text{ mm}$ and $L_t = 10.4 \text{ cm}$. These results implies that the flow front advancement on the designed LOD was expected to follow the square root of time dependency under the characteristic time (i.e., $t_c = 3.28$). Therefore, centrifugal circumferential filling time can be determined in approximately 1 s under a rotational speed of 500 rpm to fill the circular path during a single revolution of the circumferential channel.

The centrifugal circumferential filling time can be delayed because of the conditions that were disregarded in the present theoretical model, such as surface roughness and geometrical pinning effect. Therefore, we introduced a semi-circular path for the circumferential channel into the C-MUFID design by considering the reasonable total filling time. The required filling time can significantly lengthened because of the exponential saturation behavior at an increased channel length.

¹M. Madou, J. Zoval, G. Y. Jia, H. Kido, J. Kim, and N. Kim, *Annu. Rev. Biomed. Eng.* **8**, 601 (2006).

²J. Ducree, S. Haeberle, S. Lutz, S. Pausch, F. von Stetten, and R. Zengerle, *J. Micromech. Microeng.* **17**, S103 (2007).

³M. Inganäs, H. Derand, A. Eckersten, G. Ekstrand, A. K. Honerud, G. Jesson, G. Thorsen, T. Soderman, and P. Andersson, *Clin. Chem.* **51**, 1985 (2005).

⁴J. M. Park, Y. K. Cho, B. S. Lee, J. G. Lee, and C. Ko, *Lab Chip* **7**, 557 (2007).

⁵D. Mark, P. Weber, S. Lutz, M. Focke, R. Zengerle, and F. von Stetten, *Microfluid. Nanofluid.* **10**, 1279 (2011).

⁶C. H. Shih, C. H. Lu, W. L. Yuan, W. L. Chiang, and C. H. Lin, *Biomicrofluidics* **5**, 13414 (2011).

⁷Y. Deng, J. Fan, S. Zhou, T. Zhou, J. Wu, Y. Li, Z. Liu, M. Xuan, and Y. Wu, *Biomicrofluidics* **8**, 024101 (2014).

⁸N. Honda, U. Lindberg, P. Andersson, S. Hoffman, and H. Takei, *Clin. Chem.* **51**, 1955 (2005).

⁹M. Focke, F. Stumpf, G. Roth, R. Zengerle, and F. von Stetten, *Lab Chip* **10**, 3210 (2010).

¹⁰D. Sugumar, L. X. Kong, A. Ismail, M. Ravichandran, and L. Su Yin, *Biomicrofluidics* **6**, 14119 (2012).

¹¹M. Amasia, S. W. Kang, D. Banerjee, and M. Madou, *Biomicrofluidics* **7**, 14106 (2013).

¹²B. S. Lee, Y. U. Lee, H. S. Kim, T. H. Kim, J. Park, J. G. Lee, J. Kim, H. Kim, W. G. Lee, and Y. K. Cho, *Lab Chip* **11**, 70 (2011).

¹³C.-H. Shih, H.-C. Wu, C.-Y. Chang, W.-H. Huang, and Y.-F. Yang, *Biomicrofluidics* **8**, 052110 (2014).

¹⁴M. Hecke and W. K. Schomburg, *J. Micromech. Microeng.* **14**, R1 (2004).

¹⁵D. S. Kim, S. H. Lee, C. H. Ahn, J. Y. Lee, and T. H. Kwon, *Lab Chip* **6**, 794 (2006).

¹⁶J. M. Chen, P. C. Huang, and M. G. Lin, *Microfluid. Nanofluid.* **4**, 427 (2008).

¹⁷M. La, S. J. Park, H. W. Kim, J. J. Park, K. T. Ahn, S. M. Ryew, and D. S. Kim, *Microfluid. Nanofluid.* **15**, 87 (2013).

- ¹⁸J. P. Goddard and J. L. Reymond, *Trends Biotechnol.* **22**, 363 (2004).
- ¹⁹W. S. You, J. J. Park, S. M. Jin, S. M. Ryew, and H. R. Choi, "Point-of-care test equipment for flexible laboratory automation," *J. Lab Autom.* **19**, 403–412 (2014).
- ²⁰D. S. Kim, H. S. Lee, J. Y. Han, S. H. Lee, C. H. Ahn, and T. H. Kwon, *Microsyst. Technol.* **14**, 179 (2008).
- ²¹D. S. Kim and T. H. Kwon, *Microsyst. Technol.* **12**, 822 (2006).
- ²²K. J. Cha, M. H. Na, H. W. Kim, and D. S. Kim, *J. Micromech. Microeng.* **24**, 055002 (2014).
- ²³D. S. Kim, K. C. Lee, T. H. Kwon, and S. S. Lee, *J. Micromech. Microeng.* **12**, 236 (2002).
- ²⁴D. W. Langbein, *Capillary Surfaces: Shape-Stability-Dynamics, in Particular Under Weightlessness* (Springer, Berlin, New York, 2002), pp. XVIII.
- ²⁵H. Kusumaatmaja, C. M. Pooley, S. Girardo, D. Pisignano, and J. M. Yeomans, *Phys. Rev. E* **77**, 067301 (2008).
- ²⁶B. Hagmeyer, F. Zechall, and M. Stelzle, *Biomicrofluidics* **8**, 056501 (2014).
- ²⁷H. Bruus, *Theoretical Microfluidics* (Oxford University Press, Oxford, New York, 2008), p. 346.
- ²⁸See supplementary material at <http://dx.doi.org/10.1063/1.4905940> for schematic of the half-closing at the top of the reagent inlet chamber, schematic of experimental setup for the operation of the LOD and visualization of the centrifugal microfluidic phenomena, and measurement procedures based on image processing program.